

FROM : LEAD INSPECTIONS

PHONE NO. : 914 497 7930

Oct. 16 2007 08:16PM P5

ENVIRONMENTAL HAZARDS SERVICES, L.L.C.

7489 WHITE PINE ROAD - RICHMOND, VA 23237

804-275-4788 FAX 804-275-4907

WIPE LEAD ANALYSIS SUMMARY

CLIENT: House Inspections
 Chauncey S. Sage
 P.O. Box 282
 Blooming Grove, NY 10914

DATE OF SAMPLING: 15 Oct 2007
 DATE OF RECEIPT: 16 Oct 2007
 DATE OF ANALYSIS: 18 Oct 2007
 DATE OF REPORT: 18 Oct 2007

CLIENT NUMBER: 33-2200 D
 EHS PROJECT #: 2007-10-1702
 PROJECT: Deerpark Town Hall; 420 Route 209; Deerpark, NY

EHS SAMPLE#	CLIENT SAMPLE#	WIPE AREA (ft2)	TOTAL LEAD (ug)	CONCENTRATION (ug/ft2)
01	1	1.00	<20	<20
02	2	1.00	<20	<20
03	3	0.250	<20	<80
04	4	0.259	44	170
05	5	1.00	28	28
06	6	0.250	<20	<80
07	7	0.250	25	100
08	8	1.70	<20	<12
09	9	1.00	<20	<20
10	10	0.250	<20	<80
11	11	1.00	120	120
12	12	-----	<20	-----

QUALITY CONTROL DATA

BATCH#:	101607W-1
INCLUSIVE EHS SAMPLE NUMBERS:	01-12
Initial Calibration Verification (1.00ppm Pb)	99.0% Recovery
Continuing Calibration Verification 10 (10.0ppm Pb)	97.0% Recovery
Continuing Calibration Verification 5 (5.00ppm Pb)	96.6% Recovery
Laboratory Control Standard	95.9% Recovery
Matrix Spike	100% Recovery
Duplicate Relative Percent Difference	4.19 RPD
Reporting Limit	20ug
Method Detection Limit	8.0ug

FROM : LEAD INSPECTIONS

PHONE NO. : 914 497 7930

Oct. 16 2007 08:14PM P2

ENVIRONMENTAL HAZARDS SERVICES, L.L.C.7469 WHITE PINE ROAD - RICHMOND, VA 23237
804-275-4788 FAX 804-275-4907**LEAD IN SOIL ANALYSIS SUMMARY**

CLIENT: House Inspections
Chauncey S. Sage
P.O. Box 282
Blooming Grove, NY 10914

DATE OF SAMPLING: 15 Oct 2007
DATE OF RECEIPT: 16 Oct 2007
DATE OF ANALYSIS: 16 Oct 2007
DATE OF REPORT: 16 Oct 2007

CLIENT NUMBER: 33-2200 D
EHS PROJECT #: 2007-10-1703
PROJECT: Deerpark Town Hall; 420 Route 209; Deerpark, NY

EHS SAMPLE#	CLIENT SAMPLE#	CONCENTRATION PPM (mg/kg)
01	12	<48
02	18	100
03	14	62
04	15	280
05	16	91
06	17	1800
07	18	270
08	19	<42

QUALITY CONTROL DATA

BATCH#:	101607S-1
INCLUSIVE EHS SAMPLE NUMBERS:	01-08
Initial Calibration Verification (5.00ppm Pb)	105% Recovery
Continuing Calibration Verification 10 (10.0ppm Pb)	101% Recovery
Continuing Calibration Verification 5 (5.00ppm Pb)	103% Recovery
Laboratory Control Standard	113% Recovery
Matrix Spike	92.8% Recovery
Duplicate Relative Percent Difference	89.7 RPD
Reporting Limit	20ug
Method Detection Limit	2.9ug

PREPARATION METHOD: EPA 600/R-93/200
ANALYSIS METHOD: EPA SW846 7420

ANALYST: Heather Zdobysz

Reviewed By Authorized Signatory:

Irma Faszelowski
Michael A. Mueller, MPH, Laboratory Director
Howard Varner, General Manager
Irma Faszelowski, Quality Assurance Coordinator
David Xu, MS, Senior Chemist

