



Tuesday, November 18, 2014

Attn:
NYSDEC, Region 4
1150 North Westcott Road
Schenectady, NY 12306

Project ID: Katzman Recycling
Sample ID#s: BH39516

This laboratory is in compliance with the NELAC requirements of procedures used except where indicated.

This report contains results for the parameters tested, under the sampling conditions described on the Chain Of Custody, as received by the laboratory.

All soils, solids and sludges are reported on a dry weight basis unless otherwise noted in the sample comments.

A scanned version of the COC form accompanies the analytical report and is an exact duplicate of the original.

If you have any questions concerning this testing, please do not hesitate to contact Phoenix Client Services at ext. 200.

Sincerely yours,

A handwritten signature in black ink that reads "Phyllis Shiller". The signature is written in a cursive style.

Phyllis Shiller
Laboratory Director

NELAC - #NY11301
CT Lab Registration #PH-0618
MA Lab Registration #MA-CT-007
ME Lab Registration #CT-007
NH Lab Registration #213693-A,B

NJ Lab Registration #CT-003
NY Lab Registration #11301
PA Lab Registration #68-03530
RI Lab Registration #63
VT Lab Registration #VT11301



Environmental Laboratories, Inc.
 587 East Middle Turnpike, P.O.Box 370, Manchester, CT 06045
 Tel. (860) 645-1102 Fax (860) 645-0823

Analysis Report
 November 18, 2014

FOR: Attn: NYSDEC, Region 4
 1150 North Westcott Road
 Schenectady, NY 12306

Sample Information

Matrix: SOIL
 Location Code: NYSDEC
 Rush Request: 24 Hour
 P.O.#:

Custody Information

Collected by:
 Received by: LPB
 Analyzed by: see "By" below

Date Time
 11/11/14 12:50
 11/11/14 20:40

Laboratory Data

SDG ID: GBH39516
 Phoenix ID: BH39516

Project ID: Katzman Recycling
 Client ID: ACOMP

Parameter	Result	RL/ PQL	Units	Date/Time	By	Reference
TCLP Silver	< 0.10	0.10	mg/L	11/13/14	LK	SW6010
TCLP Arsenic	< 0.10	0.10	mg/L	11/13/14	LK	SW6010
TCLP Barium	5.10	0.10	mg/L	11/13/14	LK	SW6010
TCLP Cadmium	0.386	0.050	mg/L	11/13/14	LK	SW6010
TCLP Chromium	< 0.10	0.10	mg/L	11/13/14	LK	SW6010
TCLP Mercury	< 0.0002	0.0002	mg/L	11/13/14	RS	SW7470
TCLP Lead	16.7	0.10	mg/L	11/13/14	LK	SW6010
TCLP Selenium	< 0.10	0.10	mg/L	11/13/14	LK	SW6010
TCLP Metals Digestion	Completed			11/13/14	I/I	SW3005
Percent Solid	77		%	11/14/14	I	E160.3
Soil Extraction for PCB	Completed			11/14/14	CC/H	SW3545
TCLP Digestion Mercury	Completed			11/13/14	I/I	E1311/7470
TCLP Herbicides Extraction	Completed			11/13/14	W/D	SW8150 Mod
TCLP Extraction for Metals	Completed			11/12/14	i	EPA 1311
TCLP Extraction for Organics	Completed			11/12/14	i	1311
TCLP Pesticides Extraction	Completed			11/13/14	W/W	SW3510
TCLP Semi-Volatile Extraction	Completed			11/13/14	W/W	SW3510
TCLP Extraction Volatiles	Completed			11/12/14	Y	EPA 1311

Polychlorinated Biphenyls

PCB-1016	ND	2100	ug/Kg	11/17/14	AW	SW 8082
PCB-1221	ND	2100	ug/Kg	11/17/14	AW	SW 8082
PCB-1232	ND	2100	ug/Kg	11/17/14	AW	SW 8082
PCB-1242	ND	2100	ug/Kg	11/17/14	AW	SW 8082
PCB-1248	*	2100	ug/Kg	11/17/14	AW	SW 8082
PCB-1254	ND	2100	ug/Kg	11/17/14	AW	SW 8082
PCB-1260	*	2100	ug/Kg	11/17/14	AW	SW 8082
PCB-1262	ND	2100	ug/Kg	11/17/14	AW	SW 8082

Parameter	Result	RL/ PQL	Units	Date/Time	By	Reference
PCB-1268	ND	2100	ug/Kg	11/17/14	AW	SW 8082
Total PCBs	19000	2100	ug/Kg	11/17/14	AW	SW 8082
<u>QA/QC Surrogates</u>						
% DCBP	Diluted Out		%	11/17/14	AW	30 - 150 %
% TCMX	Diluted Out		%	11/17/14	AW	30 - 150 %
<u>TCLP Herbicides</u>						
2,4,5-TP (Silvex)	ND	4.2	ug/L	11/14/14	BB	SW8151
2,4-D	ND	4.2	ug/L	11/14/14	BB	SW8151
<u>QA/QC Surrogates</u>						
% DCAA	89		%	11/14/14	BB	30 - 150 %
<u>TCLP Pesticides</u>						
4,4' -DDD	ND	1.0	ug/L	11/13/14	CE	SW 8081
4,4' -DDE	ND	1.0	ug/L	11/13/14	CE	SW 8081
4,4' -DDT	ND	1.0	ug/L	11/13/14	CE	SW 8081
a-BHC	ND	0.50	ug/L	11/13/14	CE	SW 8081
Alachlor	ND	0.50	ug/L	11/13/14	CE	SW 8081
Aldrin	ND	0.50	ug/L	11/13/14	CE	SW 8081
b-BHC	ND	0.50	ug/L	11/13/14	CE	SW 8081
Chlordane	ND	5.0	ug/L	11/13/14	CE	SW 8081
d-BHC	ND	0.50	ug/L	11/13/14	CE	SW 8081
Dieldrin	ND	1.0	ug/L	11/13/14	CE	SW 8081
Endosulfan I	ND	0.50	ug/L	11/13/14	CE	SW 8081
Endosulfan II	ND	1.0	ug/L	11/13/14	CE	SW 8081
Endosulfan Sulfate	ND	1.0	ug/L	11/13/14	CE	SW 8081
Endrin	ND	1.0	ug/L	11/13/14	CE	SW 8081
Endrin Aldehyde	ND	1.0	ug/L	11/13/14	CE	SW 8081
g-BHC (Lindane)	ND	0.50	ug/L	11/13/14	CE	SW 8081
Heptachlor	ND	0.50	ug/L	11/13/14	CE	SW 8081
Heptachlor epoxide	ND	0.50	ug/L	11/13/14	CE	SW 8081
Methoxychlor	ND	0.50	ug/L	11/13/14	CE	SW 8081
Toxaphene	ND	20	ug/L	11/13/14	CE	SW 8081
<u>QA/QC Surrogates</u>						
%DCBP (Surrogate Rec)	92		%	11/13/14	CE	30 - 150 %
%TCMX (Surrogate Rec)	85		%	11/13/14	CE	30 - 150 %
<u>TCLP Volatiles</u>						
1,1-Dichloroethene	ND	25	ug/L	11/13/14	MH	SW8260
1,2-Dichloroethane	ND	25	ug/L	11/13/14	MH	SW8260
Benzene	ND	25	ug/L	11/13/14	MH	SW8260
Carbon tetrachloride	ND	25	ug/L	11/13/14	MH	SW8260
Chlorobenzene	ND	25	ug/L	11/13/14	MH	SW8260
Chloroform	ND	25	ug/L	11/13/14	MH	SW8260
Methyl ethyl ketone	ND	25	ug/L	11/13/14	MH	SW8260
Tetrachloroethene	ND	25	ug/L	11/13/14	MH	SW8260
Trichloroethene	ND	25	ug/L	11/13/14	MH	SW8260
Vinyl chloride	ND	25	ug/L	11/13/14	MH	SW8260
<u>QA/QC Surrogates</u>						
% 1,2-dichlorobenzene-d4	100		%	11/13/14	MH	70 - 130 %

Parameter	Result	RL/ PQL	Units	Date/Time	By	Reference
% Bromofluorobenzene	93		%	11/13/14	MH	70 - 130 %
% Dibromofluoromethane	95		%	11/13/14	MH	70 - 130 %
% Toluene-d8	101		%	11/13/14	MH	70 - 130 %
<u>TCLP Acid/Base-Neutral</u>						
1,4-Dichlorobenzene	ND	83	ug/L	11/13/14	DD	SW 8270
2,4,5-Trichlorophenol	ND	83	ug/L	11/13/14	DD	SW 8270
2,4,6-Trichlorophenol	ND	83	ug/L	11/13/14	DD	SW 8270
2,4-Dinitrotoluene	ND	83	ug/L	11/13/14	DD	SW 8270
2-Methylphenol (o-cresol)	ND	83	ug/L	11/13/14	DD	SW 8270
3&4-Methylphenol (m&p-Cresol)	ND	83	ug/L	11/13/14	DD	SW 8270
Hexachlorobenzene	ND	83	ug/L	11/13/14	DD	SW 8270
Hexachlorobutadiene	ND	83	ug/L	11/13/14	DD	SW 8270
Hexachloroethane	ND	83	ug/L	11/13/14	DD	SW 8270
Nitrobenzene	ND	83	ug/L	11/13/14	DD	SW 8270
Pentachlorophenol	ND	83	ug/L	11/13/14	DD	SW 8270
Pyridine	ND	83	ug/L	11/13/14	DD	SW 8270
<u>QA/QC Surrogates</u>						
% 2,4,6-Tribromophenol	77		%	11/13/14	DD	15 - 110 %
% 2-Fluorobiphenyl	76		%	11/13/14	DD	30 - 130 %
% 2-Fluorophenol	51		%	11/13/14	DD	15 - 110 %
% Nitrobenzene-d5	78		%	11/13/14	DD	30 - 130 %
% Phenol-d5	43		%	11/13/14	DD	15 - 110 %
% Terphenyl-d14	80		%	11/13/14	DD	30 - 130 %

1 = This parameter is not certified by NY NELAC for this matrix. NY NELAC does not offer certification for all parameters at this time.
 B = Present in blank, no bias suspected.

RL/PQL=Reporting/Practical Quantitation Level (Equivalent to NELAC LOQ, Limit of Quantitation) ND=Not Detected
 BRL=Below Reporting Level


Comments:

PCB Comment:

* For PCBs, as per section 11.9.3, when multiple Aroclor's of PCBs are present and the aroclor is no longer recognizable, quantitation may be performed by comparing the total area of the PCB pattern to that of the aroclor it mostly resembles. The PCB pattern did not resemble any of the standards, but most closely resembles a mixture of the Aroclors 1248 and 1260.

All soils, solids and sludges are reported on a dry weight basis unless otherwise noted in the sample comments.

If there are any questions regarding this data, please call Phoenix Client Services at extension 200.
 This report must not be reproduced except in full as defined by the attached chain of custody.



Phyllis Shiller, Laboratory Director

November 18, 2014

Reviewed and Released by: Bobbi Aloisa, Vice President



Environmental Laboratories, Inc.
 587 East Middle Turnpike, P.O.Box 370, Manchester, CT 06045
 Tel. (860) 645-1102 Fax (860) 645-0823



QA/QC Report

November 18, 2014

QA/QC Data

SDG I.D.: GBH39516

Parameter	Blank	Sample Result	Dup Result	Dup RPD	LCS %	LCSD %	LCS RPD	MS %	MSD %	MS RPD	% Rec Limits	% RPD Limits	
QA/QC Batch 291889, QC Sample No: BH38558 (BH39516)													
<u>ICP Metals - TCLP Extraction</u>													
Arsenic	BRL	0.01	0.01	NC	121	112	7.7	121	118	2.5	75 - 125	20	
Barium		0.02	0.42	0.45	6.90	108	100	7.7	106	104	1.9	75 - 125	20
Cadmium	BRL	1.86	1.96	5.20	114	106	7.3	104	102	1.9	75 - 125	20	
Chromium	BRL	<0.010	<0.010	NC	114	105	8.2	108	105	2.8	75 - 125	20	
Lead	BRL	0.600	0.638	6.10	114	106	7.3	109	106	2.8	75 - 125	20	
Selenium	BRL	0.02	0.02	NC	121	112	7.7	118	114	3.4	75 - 125	20	
Silver	BRL	<0.010	<0.010	NC	115	107	7.2	117	114	2.6	75 - 125	20	
QA/QC Batch 292142, QC Sample No: BH39787 (BH39516)													
Mercury - Water	BRL	<0.0002	<0.0002	NC	109	98.8	9.8	98.4	96.4	2.1	70 - 130	20	
Comment:													
Additional Mercury criteria: LCS acceptance range for waters is 80-120% and for soils is 70-130%.													



Environmental Laboratories, Inc.
 587 East Middle Turnpike, P.O. Box 370, Manchester, CT 06045
 Tel. (860) 645-1102 Fax (860) 645-0823



QA/QC Report

November 18, 2014

QA/QC Data

SDG I.D.: GBH39516

Parameter	Blank	LCS %	LCSD %	LCS RPD	MS %	MSD %	MS RPD	% Rec Limits	% RPD Limits
-----------	-------	-------	--------	---------	------	-------	--------	--------------	--------------

QA/QC Batch 292049, QC Sample No: BH38497 (BH39516)

Chlorinated Herbicides

2,4,5-TP (Silvex)	ND	90	81	10.5				40 - 140	20
2,4-D	ND	98	84	15.4				40 - 140	20
% DCAA (Surrogate Rec)	73	85	70	19.4				30 - 150	20

QA/QC Batch 292119, QC Sample No: BH39171 (BH39516)

Semivolatiles

1,4-Dichlorobenzene	ND	67	66	1.5				30 - 130	20
2,4,5-Trichlorophenol	ND	99	97	2.0				30 - 130	20
2,4,6-Trichlorophenol	ND	92	89	3.3				30 - 130	20
2,4-Dinitrotoluene	ND	92	96	4.3				30 - 130	20
2-Methylphenol (o-cresol)	ND	82	81	1.2				30 - 130	20
3&4-Methylphenol (m&p-cresol)	ND	80	79	1.3				30 - 130	20
Hexachlorobenzene	ND	106	96	9.9				30 - 130	20
Hexachlorobutadiene	ND	79	75	5.2				30 - 130	20
Hexachloroethane	ND	67	65	3.0				30 - 130	20
Nitrobenzene	ND	87	87	0.0				30 - 130	20
Pentachlorophenol	ND	83	84	1.2				30 - 130	20
Pyridine	ND	31	31	0.0				30 - 130	20
% 2,4,6-Tribromophenol	92	83	78	6.2				15 - 110	20
% 2-Fluorobiphenyl	79	76	74	2.7				30 - 130	20
% 2-Fluorophenol	60	55	54	1.8				15 - 110	20
% Nitrobenzene-d5	80	77	77	0.0				30 - 130	20
% Phenol-d5	50	49	48	2.1				15 - 110	20
% Terphenyl-d14	71	72	81	11.8				30 - 130	20

Comment:

A LCS and LCS Duplicate were performed instead of a matrix spike and matrix spike duplicate.

Additional 8270 criteria: 20% of compounds can be outside of acceptance criteria as long as recovery is at least 10%. (Acid surrogates acceptance range for aqueous samples: 15-110%, for soils 30-130%)

QA/QC Batch 292117, QC Sample No: BH39529 (BH39516)

Pesticides

4,4' -DDD	ND	80	90	11.8				40 - 140	20
4,4' -DDE	ND	80	91	12.9				40 - 140	20
4,4' -DDT	ND	76	87	13.5				40 - 140	20
a-BHC	ND	82	98	17.8				40 - 140	20
a-Chlordane	ND	79	89	11.9				40 - 140	20
Alachlor	ND	NA	NA	NC				40 - 140	20
Aldrin	ND	74	86	15.0				40 - 140	20
b-BHC	ND	87	98	11.9				40 - 140	20
Chlordane	ND	80	90	11.8				40 - 140	20
d-BHC	ND	81	90	10.5				40 - 140	20
Dieldrin	ND	95	107	11.9				40 - 140	20

QA/QC Data

SDG I.D.: GBH39516

Parameter	Blank	LCS %	LCSD %	LCS RPD	MS %	MSD %	MS RPD	% Rec Limits	% RPD Limits
Endosulfan I	ND	80	89	10.7				40 - 140	20
Endosulfan II	ND	85	101	17.2				40 - 140	20
Endosulfan sulfate	ND	81	88	8.3				40 - 140	20
Endrin	ND	78	89	13.2				40 - 140	20
Endrin aldehyde	ND	99	101	2.0				40 - 140	20
g-BHC	ND	87	102	15.9				40 - 140	20
g-Chlordane	ND	80	90	11.8				40 - 140	20
Heptachlor	ND	80	99	21.2				40 - 140	20
Heptachlor epoxide	ND	81	94	14.9				40 - 140	20
Methoxychlor	ND	86	102	17.0				40 - 140	20
Toxaphene	ND	NA	NA	NC				40 - 140	20
% DCBP	102	88	93	5.5				40 - 150	20
% TCMX	83	65	80	20.7				40 - 150	20

Comment:

A LCS and LCS duplicate were performed instead of a MS and MSD. Alpha and gamma chlordane were spiked and analyzed instead of technical chlordane. Gamma chlordane recovery is reported as chlordane in the LCS and LCSD

QA/QC Batch 292338, QC Sample No: BH39658 (BH39516 (5X))

Volatiles - TCLP

1,1-Dichloroethene	ND	110	100	9.5	109	111	1.8	70 - 130	30
1,2-Dichloroethane	ND	98	100	2.0	93	100	7.3	70 - 130	30
Benzene	ND	103	104	1.0	99	106	6.8	70 - 130	30
Carbon tetrachloride	ND	106	99	6.8	101	108	6.7	70 - 130	30
Chlorobenzene	ND	100	103	3.0	99	103	4.0	70 - 130	30
Chloroform	ND	101	105	3.9	95	101	6.1	70 - 130	30
Methyl ethyl ketone	ND	91	91	0.0	91	93	2.2	70 - 130	30
Tetrachloroethene	ND	103	99	4.0	103	109	5.7	70 - 130	30
Trichloroethene	ND	103	103	0.0	101	106	4.8	70 - 130	30
Vinyl chloride	ND	111	124	11.1	109	118	7.9	70 - 130	30
% 1,2-dichlorobenzene-d4	99	98	99	1.0	98	99	1.0	70 - 130	30
% Bromofluorobenzene	95	98	99	1.0	100	99	1.0	70 - 130	30
% Dibromofluoromethane	98	101	98	3.0	97	97	0.0	70 - 130	30
% Toluene-d8	101	101	100	1.0	99	100	1.0	70 - 130	30

Comment:

Additional 8260 criteria: 10% of LCS/LCSD compounds can be outside of acceptance criteria as long as recovery is 40-160%.

QA/QC Batch 292389, QC Sample No: BH41592 (BH39516)

Polychlorinated Biphenyls - Soil

PCB-1016	ND	112	117	4.4	118	103	13.6	40 - 140	30
PCB-1221	ND							40 - 140	30
PCB-1232	ND							40 - 140	30
PCB-1242	ND							40 - 140	30
PCB-1248	ND							40 - 140	30
PCB-1254	ND							40 - 140	30
PCB-1260	ND	100	104	3.9	101	101	0.0	40 - 140	30
PCB-1262	ND							40 - 140	30
PCB-1268	ND							40 - 140	30
% DCBP (Surrogate Rec)	91	95	102	7.1	96	96	0.0	30 - 150	30
% TCMX (Surrogate Rec)	101	102	106	3.8	106	96	9.9	30 - 150	30

r = This parameter is outside laboratory rpd specified recovery limits.


QA/QC Data

SDG I.D.: GBH39516

Parameter	Blank	LCS %	LCSD %	LCS RPD	MS %	MSD %	MS RPD	% Rec Limits	% RPD Limits
-----------	-------	----------	-----------	------------	---------	----------	-----------	--------------------	--------------------

If there are any questions regarding this data, please call Phoenix Client Services at extension 200.

- RPD - Relative Percent Difference
- LCS - Laboratory Control Sample
- LCSD - Laboratory Control Sample Duplicate
- MS - Matrix Spike
- MS Dup - Matrix Spike Duplicate
- NC - No Criteria
- Intf - Interference



Phyllis Shiller, Laboratory Director
November 18, 2014

Criteria: None

State: NY

Sample Criteria Exceedences Report

GBH39516 - NYSDEC

SampNo	Acode	Phoenix Analyte	Criteria	Result	RL	Criteria	RL Criteria	Analysis Units
BH39516	TCLP-PB	TCLP Lead	EPA / 40 CFR 261.24 / Toxicity Characteristics	16.7	0.10	5	5	mg/L

Phoenix Laboratories does not assume responsibility for the data contained in this report. It is provided as an additional tool to identify requested criteria exceedences. All efforts are made to ensure the accuracy of the data (obtained from appropriate agencies). A lack of exceedence information does not necessarily suggest conformance to the criteria. It is ultimately the site professional's responsibility to determine appropriate compliance.



Environmental Laboratories, Inc.
587 East Middle Turnpike, P.O.Box 370, Manchester, CT 06045
Tel. (860) 645-1102 Fax (860) 645-0823



NY Temperature Narration

November 18, 2014

SDG I.D.: GBH39516

The samples in this delivery group were received at 4°C.
(Note acceptance criteria is above freezing up to 6°C)

Coolant: PK Yes No
 ICE No

Temp: ° C Pg of

NY/NJ CHAIN OF CUSTODY RECORD



587 East Middle Turnpike, P.O. Box 370, Manchester, CT 06040
 Email: info@phoenixlabs.com Fax (860) 645-0823
 Client Services (860) 645-8726

Contact Options:

Fax:
 Phone:
 Email:

Customer: NYSDDEC-5 via Precision Env. Project P.O.: Tobacco Recycling Co.
 Address: 831 Rt 67 Lot 38 Report to: A. Grogan, NYSDDEC
Bullston Ave NY 12020 Invoice to: NYSDDEC # 558035

Sampler's Signature: [Signature] Date: 11/11/17
 Client Sample - Information - Identification
 Matrix Code: GW=Ground Water SW=Surface Water MW=Waste Water
 DW=Drinking Water RW=Raw Water SE=Sludge S=Soil SD=Solid W=Wipe
 OIL=Oil B=Bulk L=Liquid

PHOENIX USE ONLY SAMPLE #	Customer Sample Identification	Sample Matrix	Date Sampled	Time Sampled	Analysis Request
29516	Acene	S	11/14/17	8:30	[Grid]

Requisitioned by: [Signature] Accepted by: [Signature] Date: 11/11/17 Time: 15:30

Comments, Special Requirements or Regulations:
*Standard IATA to
Report to A. Grogan NYSDDEC
cc. S. Phelps P&ES
Spill # 558035

Turnaround:
 1 Day*
 2 Days*
 3 Days*
 5 Days
 10 Days
 Other

* SURCHARGE APPLIES Standard

NJ Res. Criteria
 Non-Res. Criteria
 Impact to GW Soil Cleanup Criteria
 GW Criteria

NY TAGM 4046 GW
 TAGM 4046 SOIL
 NY375 Unrestricted Use Soil
 NY375 Residential Soil
 Restricted/Residential Commercial Industrial

Data Format
 Phoenix Std Report
 Excel
 PDF
 GIS/Key
 EQUIS
 NJ Hazsite EDD
 NY EZ EDD (ASP)
 Other

Data Package
 NJ Reduced Deliv. *
 NY Enhanced (ASP B) *
 Other

State where samples were collected: NY

Sarah - Phoenixlabs

From: Stephen Phelps [Sphelps@PrecisionEnvironmentalny.com]
Sent: Wednesday, November 12, 2014 12:59 PM
To: 'Sarah - Phoenixlabs'
Subject: RE: Recycling Chains

Can you up the TAT on those 10 samples to 24 hour?

-----Original Message-----

From: Sarah - Phoenixlabs [<mailto:sarah@phoenixlabs.com>]
Sent: Wednesday, November 12, 2014 12:31 PM
To: 'Stephen Phelps'
Cc: 'Purzycki, Alicia J (DEC)'
Subject: RE: Recycling Chains

Thank You so much Steve and Hello Alicia. I will add you to this account Alicia, once the report is ready and released you will get an email from clientservices@phoenixlabs.com which will take you to the report.

Thanks again
Sarah

-----Original Message-----

From: Stephen Phelps [<mailto:Sphelps@PrecisionEnvironmentalny.com>]
Sent: Wednesday, November 12, 2014 12:22 PM
To: 'Sarah - Phoenixlabs'
Cc: 'Purzycki, Alicia J (DEC)'
Subject: RE: Recycling Chains

Hi Sarah - Alicia has been CC'd on this correspondence and I also informed her earlier of the decision to use you folks on this project.

Regards, Steve

-----Original Message-----

From: Sarah - Phoenixlabs [<mailto:sarah@phoenixlabs.com>]
Sent: Wednesday, November 12, 2014 12:19 PM
To: 'Stephen Phelps'
Subject: Recycling Chains

Do you have A. Purzycki email address? We don't have that it appears on this account thanks

Sarah Bell
Client Services Representative
Phoenix Environmental Laboratories
587 East Middle Turnpike
Manchester, CT 06040
Ph: 1-860-645-1102

Linda - Phoenixlabs

From: Stephen Phelps [Sphelps@PrecisionEnvironmentalny.com]
Sent: Friday, November 14, 2014 8:33 AM
To: 'Linda - Phoenixlabs'
Subject: RE: Katzman Recycling results
Attachments: image001.emz; image003.emz; oledata.mso

Hi Linda - Please note that for the other Katzman COC for composite sample where full TCLP was requested we do need PCBs run on that sample as well.

Feel Free to call should you have any questions.

Regards, Steve

Stephen M. Phelps
Precision Environmental Services, Inc.
831 Route 67, Lot 38A
Ballston Spa, NY 12020
Tel: 518.885.4399
Fax: 518.885.4416
Web: www.precisionenvironmentalny.com



From: Linda - Phoenixlabs [mailto:linda@phoenixlabs.com]
Sent: Thursday, November 13, 2014 11:34 AM
To: alicia.purzycki@dec.ny.gov; 'Stephen Phelps'
Subject: Katzman Recycling results

11/14/2014