



## ANALYTICAL REPORT

Lab Number:	L1822216
Client:	CH2M / Dow Chemical Company 299 Madison Ave. Morristown, NJ 07960
ATTN:	David Newman
Phone:	(862) 242-7061
Project Name:	FORMER HAMPSHIRE CHEMICAL CORP
Project Number:	701970.01.SA
Report Date:	06/19/18

The original project report/data package is held by Alpha Analytical. This report/data package is paginated and should be reproduced only in its entirety. Alpha Analytical holds no responsibility for results and/or data that are not consistent with the original.

Certifications & Approvals: MA (M-MA086), NH NELAP (2064), CT (PH-0574), IL (200077), ME (MA00086), MD (348), NJ (MA935), NY (11148), NC (25700/666), PA (68-03671), RI (LAO00065), TX (T104704476), VT (VT-0935), VA (460195), USDA (Permit #P330-17-00196).

---

Eight Walkup Drive, Westborough, MA 01581-1019  
508-898-9220 (Fax) 508-898-9193 800-624-9220 - [www.alphalab.com](http://www.alphalab.com)



**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP  
**Project Number:** 701970.01.SA

**Lab Number:** L1822216  
**Report Date:** 06/19/18

<b>Alpha Sample ID</b>	<b>Client ID</b>	<b>Matrix</b>	<b>Sample Location</b>	<b>Collection Date/Time</b>	<b>Receive Date</b>
L1822216-01	MW23-061318	WATER	WATERLOO, NY	06/13/18 09:20	06/13/18
L1822216-02	PZ06-060818	WATER	WATERLOO, NY	06/08/18 12:00	06/13/18
L1822216-03	PZ06-061318	WATER	WATERLOO, NY	06/13/18 10:25	06/13/18
L1822216-04	TB-061318	WATER	WATERLOO, NY	06/13/18 08:00	06/13/18
L1822216-05	MW24-061318	WATER	WATERLOO, NY	06/13/18 09:50	06/13/18

**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP  
**Project Number:** 701970.01.SA

**Lab Number:** L1822216  
**Report Date:** 06/19/18

### Case Narrative

The samples were received in accordance with the Chain of Custody and no significant deviations were encountered during the preparation or analysis unless otherwise noted. Sample Receipt, Container Information, and the Chain of Custody are located at the back of the report.

Results contained within this report relate only to the samples submitted under this Alpha Lab Number and meet NELAP requirements for all NELAP accredited parameters unless otherwise noted in the following narrative. The data presented in this report is organized by parameter (i.e. VOC, SVOC, etc.). Sample specific Quality Control data (i.e. Surrogate Spike Recovery) is reported at the end of the target analyte list for each individual sample, followed by the Laboratory Batch Quality Control at the end of each parameter. Tentatively Identified Compounds (TICs), if requested, are reported for compounds identified to be present and are not part of the method/program Target Compound List, even if only a subset of the TCL are being reported. If a sample was re-analyzed or re-extracted due to a required quality control corrective action and if both sets of data are reported, the Laboratory ID of the re-analysis or re-extraction is designated with an "R" or "RE", respectively. When multiple Batch Quality Control elements are reported (e.g. more than one LCS), the associated samples for each element are noted in the grey shaded header line of each data table. Any Laboratory Batch, Sample Specific % recovery or RPD value that is outside the listed Acceptance Criteria is bolded in the report. All specific QC information is also incorporated in the Data Usability format of our Data Merger tool where it can be reviewed along with any associated usability implications. Soil/sediments, solids and tissues are reported on a dry weight basis unless otherwise noted. Definitions of all data qualifiers and acronyms used in this report are provided in the Glossary located at the back of the report.

In reference to questions H (CAM) or 4 (RCP) when "NO" is checked, the performance criteria for CAM and RCP methods allow for some quality control failures to occur and still be within method compliance. In these instances the specific failure is not narrated but noted in the associated QC table. The information is also incorporated in the Data Usability format of our Data Merger tool where it can be reviewed along with any associated usability implications.

Please see the associated ADEx data file for a comparison of laboratory reporting limits that were achieved with the regulatory Numerical Standards requested on the Chain of Custody.

#### HOLD POLICY

For samples submitted on hold, Alpha's policy is to hold samples (with the exception of Air canisters) free of charge for 21 calendar days from the date the project is completed. After 21 calendar days, we will dispose of all samples submitted including those put on hold unless you have contacted your Client Service Representative and made arrangements for Alpha to continue to hold the samples. Air canisters will be disposed after 3 business days from the date the project is completed.

Please contact Client Services at 800-624-9220 with any questions.

**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP  
**Project Number:** 701970.01.SA

**Lab Number:** L1822216  
**Report Date:** 06/19/18

### Case Narrative (continued)

#### Report Submission

All non-detect (ND) or estimated concentrations (J-qualified) have been quantitated to the limit noted in the MDL column.

#### Sample Receipt

The project number was specified by the project manager.

#### Volatile Organics

The initial calibration, associated with L1822216-02 and -04, did not meet the method required minimum response factor for the calibration standards for bromodichloromethane, cis-1,3-dichloropropene, bromomethane, chloroethane, trichloroethene, dibromomethane, 2-butanone, 4-methyl-2-pentanone, 2-hexanone, bromochloromethane and 1,4-dioxane.


The continuing calibration, associated with L1822216-02 and -04, did not meet the method required minimum response factor for bromomethane, chloroethane, bromodichloromethane, 2-butanone, 1,4-dioxane, 4-methyl-2-pentanone and 2-hexanone.

#### Nitrogen, Ammonia

The WG1126858-4 MS recovery (82%), performed on L1822216-05, is outside the acceptance criteria; however, the associated LCS recovery is within criteria. No further action was taken.

I, the undersigned, attest under the pains and penalties of perjury that, to the best of my knowledge and belief and based upon my personal inquiry of those responsible for providing the information contained in this analytical report, such information is accurate and complete. This certificate of analysis is not complete unless this page accompanies any and all pages of this report.

Authorized Signature:



Michelle M. Morris

Title: Technical Director/Representative

Date: 06/19/18

# ORGANICS

# **VOLATILES**

**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP  
**Project Number:** 701970.01.SA

**Lab Number:** L1822216  
**Report Date:** 06/19/18

**SAMPLE RESULTS**

**Lab ID:** L1822216-02  
**Client ID:** PZ06-060818  
**Sample Location:** WATERLOO, NY

**Date Collected:** 06/08/18 12:00  
**Date Received:** 06/13/18  
**Field Prep:** Refer to COC

**Sample Depth:**

**Matrix:** Water  
**Analytical Method:** 1,8260C  
**Analytical Date:** 06/16/18 10:05  
**Analyst:** JC

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor
Volatile Organics by GC/MS - Westborough Lab						
Methylene chloride	ND		ug/l	2.5	0.70	1
1,1-Dichloroethane	ND		ug/l	2.5	0.70	1
Chloroform	ND		ug/l	2.5	0.70	1
Carbon tetrachloride	ND		ug/l	0.50	0.13	1
1,2-Dichloropropane	ND		ug/l	1.0	0.14	1
Dibromochloromethane	ND		ug/l	0.50	0.15	1
1,1,2-Trichloroethane	ND		ug/l	1.5	0.50	1
Tetrachloroethene	ND		ug/l	0.50	0.18	1
Chlorobenzene	ND		ug/l	2.5	0.70	1
Trichlorofluoromethane	ND		ug/l	2.5	0.70	1
1,2-Dichloroethane	ND		ug/l	0.50	0.13	1
1,1,1-Trichloroethane	ND		ug/l	2.5	0.70	1
Bromodichloromethane	ND		ug/l	0.50	0.19	1
trans-1,3-Dichloropropene	ND		ug/l	0.50	0.16	1
cis-1,3-Dichloropropene	ND		ug/l	0.50	0.14	1
1,3-Dichloropropene, Total	ND		ug/l	0.50	0.14	1
1,1-Dichloropropene	ND		ug/l	2.5	0.70	1
Bromoform	ND		ug/l	2.0	0.65	1
1,1,2,2-Tetrachloroethane	ND		ug/l	0.50	0.17	1
Benzene	ND		ug/l	0.50	0.16	1
Toluene	ND		ug/l	2.5	0.70	1
Ethylbenzene	ND		ug/l	2.5	0.70	1
Chloromethane	ND		ug/l	2.5	0.70	1
Bromomethane	ND		ug/l	2.5	0.70	1
Vinyl chloride	ND		ug/l	1.0	0.07	1
Chloroethane	ND		ug/l	2.5	0.70	1
1,1-Dichloroethene	ND		ug/l	0.50	0.17	1
trans-1,2-Dichloroethene	ND		ug/l	2.5	0.70	1

**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP  
**Project Number:** 701970.01.SA

**Lab Number:** L1822216  
**Report Date:** 06/19/18

**SAMPLE RESULTS**

**Lab ID:** L1822216-02  
**Client ID:** PZ06-060818  
**Sample Location:** WATERLOO, NY

**Date Collected:** 06/08/18 12:00  
**Date Received:** 06/13/18  
**Field Prep:** Refer to COC

**Sample Depth:**

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor
Volatile Organics by GC/MS - Westborough Lab						
Trichloroethene	ND		ug/l	0.50	0.18	1
1,2-Dichlorobenzene	ND		ug/l	2.5	0.70	1
1,3-Dichlorobenzene	ND		ug/l	2.5	0.70	1
1,4-Dichlorobenzene	ND		ug/l	2.5	0.70	1
Methyl tert butyl ether	ND		ug/l	2.5	0.70	1
p/m-Xylene	ND		ug/l	2.5	0.70	1
o-Xylene	ND		ug/l	2.5	0.70	1
Xylenes, Total	ND		ug/l	2.5	0.70	1
cis-1,2-Dichloroethene	ND		ug/l	2.5	0.70	1
1,2-Dichloroethene, Total	ND		ug/l	2.5	0.70	1
Dibromomethane	ND		ug/l	5.0	1.0	1
1,2,3-Trichloropropane	ND		ug/l	2.5	0.70	1
Styrene	ND		ug/l	2.5	0.70	1
Dichlorodifluoromethane	ND		ug/l	5.0	1.0	1
Acetone	6.8		ug/l	5.0	1.5	1
Carbon disulfide	ND		ug/l	5.0	1.0	1
2-Butanone	ND		ug/l	5.0	1.9	1
Vinyl acetate	ND		ug/l	5.0	1.0	1
4-Methyl-2-pentanone	ND		ug/l	5.0	1.0	1
2-Hexanone	ND		ug/l	5.0	1.0	1
Bromochloromethane	ND		ug/l	2.5	0.70	1
2,2-Dichloropropane	ND		ug/l	2.5	0.70	1
1,2-Dibromoethane	ND		ug/l	2.0	0.65	1
1,3-Dichloropropane	ND		ug/l	2.5	0.70	1
1,1,1,2-Tetrachloroethane	ND		ug/l	2.5	0.70	1
Bromobenzene	ND		ug/l	2.5	0.70	1
n-Butylbenzene	ND		ug/l	2.5	0.70	1
sec-Butylbenzene	ND		ug/l	2.5	0.70	1
tert-Butylbenzene	ND		ug/l	2.5	0.70	1
o-Chlorotoluene	ND		ug/l	2.5	0.70	1
p-Chlorotoluene	ND		ug/l	2.5	0.70	1
1,2-Dibromo-3-chloropropane	ND		ug/l	2.5	0.70	1
Hexachlorobutadiene	ND		ug/l	2.5	0.70	1
Isopropylbenzene	ND		ug/l	2.5	0.70	1
p-Isopropyltoluene	ND		ug/l	2.5	0.70	1
Naphthalene	ND		ug/l	2.5	0.70	1
n-Propylbenzene	ND		ug/l	2.5	0.70	1



**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP  
**Project Number:** 701970.01.SA

**Lab Number:** L1822216  
**Report Date:** 06/19/18

**SAMPLE RESULTS**

**Lab ID:** L1822216-02  
**Client ID:** PZ06-060818  
**Sample Location:** WATERLOO, NY

**Date Collected:** 06/08/18 12:00  
**Date Received:** 06/13/18  
**Field Prep:** Refer to COC

Sample Depth:

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor
Volatile Organics by GC/MS - Westborough Lab						
1,2,3-Trichlorobenzene	ND		ug/l	2.5	0.70	1
1,2,4-Trichlorobenzene	ND		ug/l	2.5	0.70	1
1,3,5-Trimethylbenzene	ND		ug/l	2.5	0.70	1
1,2,4-Trimethylbenzene	ND		ug/l	2.5	0.70	1
1,4-Dioxane	ND		ug/l	250	61.	1

Surrogate	% Recovery	Qualifier	Acceptance Criteria
1,2-Dichloroethane-d4	97		70-130
Toluene-d8	97		70-130
4-Bromofluorobenzene	101		70-130
Dibromofluoromethane	97		70-130

**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP  
**Project Number:** 701970.01.SA

**Lab Number:** L1822216  
**Report Date:** 06/19/18

**SAMPLE RESULTS**

**Lab ID:** L1822216-04  
**Client ID:** TB-061318  
**Sample Location:** WATERLOO, NY

**Date Collected:** 06/13/18 08:00  
**Date Received:** 06/13/18  
**Field Prep:** Not Specified

**Sample Depth:**

**Matrix:** Water  
**Analytical Method:** 1,8260C  
**Analytical Date:** 06/16/18 09:40  
**Analyst:** JC

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor
Volatile Organics by GC/MS - Westborough Lab						
Methylene chloride	ND		ug/l	2.5	0.70	1
1,1-Dichloroethane	ND		ug/l	2.5	0.70	1
Chloroform	ND		ug/l	2.5	0.70	1
Carbon tetrachloride	ND		ug/l	0.50	0.13	1
1,2-Dichloropropane	ND		ug/l	1.0	0.14	1
Dibromochloromethane	ND		ug/l	0.50	0.15	1
1,1,2-Trichloroethane	ND		ug/l	1.5	0.50	1
Tetrachloroethene	ND		ug/l	0.50	0.18	1
Chlorobenzene	ND		ug/l	2.5	0.70	1
Trichlorofluoromethane	ND		ug/l	2.5	0.70	1
1,2-Dichloroethane	ND		ug/l	0.50	0.13	1
1,1,1-Trichloroethane	ND		ug/l	2.5	0.70	1
Bromodichloromethane	ND		ug/l	0.50	0.19	1
trans-1,3-Dichloropropene	ND		ug/l	0.50	0.16	1
cis-1,3-Dichloropropene	ND		ug/l	0.50	0.14	1
1,3-Dichloropropene, Total	ND		ug/l	0.50	0.14	1
1,1-Dichloropropene	ND		ug/l	2.5	0.70	1
Bromoform	ND		ug/l	2.0	0.65	1
1,1,2,2-Tetrachloroethane	ND		ug/l	0.50	0.17	1
Benzene	ND		ug/l	0.50	0.16	1
Toluene	ND		ug/l	2.5	0.70	1
Ethylbenzene	ND		ug/l	2.5	0.70	1
Chloromethane	ND		ug/l	2.5	0.70	1
Bromomethane	ND		ug/l	2.5	0.70	1
Vinyl chloride	ND		ug/l	1.0	0.07	1
Chloroethane	ND		ug/l	2.5	0.70	1
1,1-Dichloroethene	ND		ug/l	0.50	0.17	1
trans-1,2-Dichloroethene	ND		ug/l	2.5	0.70	1

**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP  
**Project Number:** 701970.01.SA

**Lab Number:** L1822216  
**Report Date:** 06/19/18

**SAMPLE RESULTS**

**Lab ID:** L1822216-04  
**Client ID:** TB-061318  
**Sample Location:** WATERLOO, NY

**Date Collected:** 06/13/18 08:00  
**Date Received:** 06/13/18  
**Field Prep:** Not Specified

**Sample Depth:**

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor
Volatile Organics by GC/MS - Westborough Lab						
Trichloroethene	ND		ug/l	0.50	0.18	1
1,2-Dichlorobenzene	ND		ug/l	2.5	0.70	1
1,3-Dichlorobenzene	ND		ug/l	2.5	0.70	1
1,4-Dichlorobenzene	ND		ug/l	2.5	0.70	1
Methyl tert butyl ether	ND		ug/l	2.5	0.70	1
p/m-Xylene	ND		ug/l	2.5	0.70	1
o-Xylene	ND		ug/l	2.5	0.70	1
Xylenes, Total	ND		ug/l	2.5	0.70	1
cis-1,2-Dichloroethene	ND		ug/l	2.5	0.70	1
1,2-Dichloroethene, Total	ND		ug/l	2.5	0.70	1
Dibromomethane	ND		ug/l	5.0	1.0	1
1,2,3-Trichloropropane	ND		ug/l	2.5	0.70	1
Styrene	ND		ug/l	2.5	0.70	1
Dichlorodifluoromethane	ND		ug/l	5.0	1.0	1
Acetone	ND		ug/l	5.0	1.5	1
Carbon disulfide	ND		ug/l	5.0	1.0	1
2-Butanone	ND		ug/l	5.0	1.9	1
Vinyl acetate	ND		ug/l	5.0	1.0	1
4-Methyl-2-pentanone	ND		ug/l	5.0	1.0	1
2-Hexanone	ND		ug/l	5.0	1.0	1
Bromochloromethane	ND		ug/l	2.5	0.70	1
2,2-Dichloropropane	ND		ug/l	2.5	0.70	1
1,2-Dibromoethane	ND		ug/l	2.0	0.65	1
1,3-Dichloropropane	ND		ug/l	2.5	0.70	1
1,1,1,2-Tetrachloroethane	ND		ug/l	2.5	0.70	1
Bromobenzene	ND		ug/l	2.5	0.70	1
n-Butylbenzene	ND		ug/l	2.5	0.70	1
sec-Butylbenzene	ND		ug/l	2.5	0.70	1
tert-Butylbenzene	ND		ug/l	2.5	0.70	1
o-Chlorotoluene	ND		ug/l	2.5	0.70	1
p-Chlorotoluene	ND		ug/l	2.5	0.70	1
1,2-Dibromo-3-chloropropane	ND		ug/l	2.5	0.70	1
Hexachlorobutadiene	ND		ug/l	2.5	0.70	1
Isopropylbenzene	ND		ug/l	2.5	0.70	1
p-Isopropyltoluene	ND		ug/l	2.5	0.70	1
Naphthalene	ND		ug/l	2.5	0.70	1
n-Propylbenzene	ND		ug/l	2.5	0.70	1

**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP  
**Project Number:** 701970.01.SA

**Lab Number:** L1822216  
**Report Date:** 06/19/18

**SAMPLE RESULTS**

**Lab ID:** L1822216-04  
**Client ID:** TB-061318  
**Sample Location:** WATERLOO, NY

**Date Collected:** 06/13/18 08:00  
**Date Received:** 06/13/18  
**Field Prep:** Not Specified

Sample Depth:

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor
Volatile Organics by GC/MS - Westborough Lab						
1,2,3-Trichlorobenzene	ND		ug/l	2.5	0.70	1
1,2,4-Trichlorobenzene	ND		ug/l	2.5	0.70	1
1,3,5-Trimethylbenzene	ND		ug/l	2.5	0.70	1
1,2,4-Trimethylbenzene	ND		ug/l	2.5	0.70	1
1,4-Dioxane	ND		ug/l	250	61.	1

Surrogate	% Recovery	Qualifier	Acceptance Criteria
1,2-Dichloroethane-d4	96		70-130
Toluene-d8	98		70-130
4-Bromofluorobenzene	102		70-130
Dibromofluoromethane	97		70-130

Project Name: FORMER HAMPSHIRE CHEMICAL CORP

Lab Number: L1822216

Project Number: 701970.01.SA

Report Date: 06/19/18

### Method Blank Analysis Batch Quality Control

Analytical Method: 1,8260C  
 Analytical Date: 06/16/18 09:01  
 Analyst: BD

Parameter	Result	Qualifier	Units	RL	MDL
Volatile Organics by GC/MS - Westborough Lab for sample(s): 02,04 Batch: WG1127131-5					
Methylene chloride	ND		ug/l	2.5	0.70
1,1-Dichloroethane	ND		ug/l	2.5	0.70
Chloroform	ND		ug/l	2.5	0.70
Carbon tetrachloride	ND		ug/l	0.50	0.13
1,2-Dichloropropane	ND		ug/l	1.0	0.14
Dibromochloromethane	ND		ug/l	0.50	0.15
1,1,2-Trichloroethane	ND		ug/l	1.5	0.50
Tetrachloroethene	ND		ug/l	0.50	0.18
Chlorobenzene	ND		ug/l	2.5	0.70
Trichlorofluoromethane	ND		ug/l	2.5	0.70
1,2-Dichloroethane	ND		ug/l	0.50	0.13
1,1,1-Trichloroethane	ND		ug/l	2.5	0.70
Bromodichloromethane	ND		ug/l	0.50	0.19
trans-1,3-Dichloropropene	ND		ug/l	0.50	0.16
cis-1,3-Dichloropropene	ND		ug/l	0.50	0.14
1,3-Dichloropropene, Total	ND		ug/l	0.50	0.14
1,1-Dichloropropene	ND		ug/l	2.5	0.70
Bromoform	ND		ug/l	2.0	0.65
1,1,2,2-Tetrachloroethane	ND		ug/l	0.50	0.17
Benzene	ND		ug/l	0.50	0.16
Toluene	ND		ug/l	2.5	0.70
Ethylbenzene	ND		ug/l	2.5	0.70
Chloromethane	ND		ug/l	2.5	0.70
Bromomethane	ND		ug/l	2.5	0.70
Vinyl chloride	ND		ug/l	1.0	0.07
Chloroethane	ND		ug/l	2.5	0.70
1,1-Dichloroethene	ND		ug/l	0.50	0.17
trans-1,2-Dichloroethene	ND		ug/l	2.5	0.70
Trichloroethene	ND		ug/l	0.50	0.18

Project Name: FORMER HAMPSHIRE CHEMICAL CORP

Lab Number: L1822216

Project Number: 701970.01.SA

Report Date: 06/19/18

### Method Blank Analysis Batch Quality Control

Analytical Method: 1,8260C  
 Analytical Date: 06/16/18 09:01  
 Analyst: BD

Parameter	Result	Qualifier	Units	RL	MDL
Volatile Organics by GC/MS - Westborough Lab for sample(s): 02,04 Batch: WG1127131-5					
1,2-Dichlorobenzene	ND		ug/l	2.5	0.70
1,3-Dichlorobenzene	ND		ug/l	2.5	0.70
1,4-Dichlorobenzene	ND		ug/l	2.5	0.70
Methyl tert butyl ether	ND		ug/l	2.5	0.70
p/m-Xylene	ND		ug/l	2.5	0.70
o-Xylene	ND		ug/l	2.5	0.70
Xylenes, Total	ND		ug/l	2.5	0.70
cis-1,2-Dichloroethene	ND		ug/l	2.5	0.70
1,2-Dichloroethene, Total	ND		ug/l	2.5	0.70
Dibromomethane	ND		ug/l	5.0	1.0
1,2,3-Trichloropropane	ND		ug/l	2.5	0.70
Styrene	ND		ug/l	2.5	0.70
Dichlorodifluoromethane	ND		ug/l	5.0	1.0
Acetone	ND		ug/l	5.0	1.5
Carbon disulfide	ND		ug/l	5.0	1.0
2-Butanone	ND		ug/l	5.0	1.9
Vinyl acetate	ND		ug/l	5.0	1.0
4-Methyl-2-pentanone	ND		ug/l	5.0	1.0
2-Hexanone	ND		ug/l	5.0	1.0
Bromochloromethane	ND		ug/l	2.5	0.70
2,2-Dichloropropane	ND		ug/l	2.5	0.70
1,2-Dibromoethane	ND		ug/l	2.0	0.65
1,3-Dichloropropane	ND		ug/l	2.5	0.70
1,1,1,2-Tetrachloroethane	ND		ug/l	2.5	0.70
Bromobenzene	ND		ug/l	2.5	0.70
n-Butylbenzene	ND		ug/l	2.5	0.70
sec-Butylbenzene	ND		ug/l	2.5	0.70
tert-Butylbenzene	ND		ug/l	2.5	0.70
o-Chlorotoluene	ND		ug/l	2.5	0.70

Project Name: FORMER HAMPSHIRE CHEMICAL CORP

Lab Number: L1822216

Project Number: 701970.01.SA

Report Date: 06/19/18

### Method Blank Analysis Batch Quality Control

Analytical Method: 1,8260C  
 Analytical Date: 06/16/18 09:01  
 Analyst: BD

Parameter	Result	Qualifier	Units	RL	MDL
Volatile Organics by GC/MS - Westborough Lab for sample(s): 02,04 Batch: WG1127131-5					
p-Chlorotoluene	ND		ug/l	2.5	0.70
1,2-Dibromo-3-chloropropane	ND		ug/l	2.5	0.70
Hexachlorobutadiene	ND		ug/l	2.5	0.70
Isopropylbenzene	ND		ug/l	2.5	0.70
p-Isopropyltoluene	ND		ug/l	2.5	0.70
Naphthalene	0.70	J	ug/l	2.5	0.70
n-Propylbenzene	ND		ug/l	2.5	0.70
1,2,3-Trichlorobenzene	ND		ug/l	2.5	0.70
1,2,4-Trichlorobenzene	ND		ug/l	2.5	0.70
1,3,5-Trimethylbenzene	ND		ug/l	2.5	0.70
1,2,4-Trimethylbenzene	ND		ug/l	2.5	0.70
1,4-Dioxane	ND		ug/l	250	61.

#### Tentatively Identified Compounds

No Tentatively Identified Compounds ND ug/l

Surrogate	%Recovery	Qualifier	Acceptance Criteria
1,2-Dichloroethane-d4	95		70-130
Toluene-d8	98		70-130
4-Bromofluorobenzene	100		70-130
Dibromofluoromethane	97		70-130

# Lab Control Sample Analysis

## Batch Quality Control

**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP

**Lab Number:** L1822216

**Project Number:** 701970.01.SA

**Report Date:** 06/19/18

Parameter	LCS %Recovery	Qual	LCSD %Recovery	Qual	%Recovery Limits	RPD	Qual	RPD Limits
Volatile Organics by GC/MS - Westborough Lab Associated sample(s): 02,04 Batch: WG1127131-3 WG1127131-4								
Methylene chloride	97		97		70-130	0		20
1,1-Dichloroethane	98		96		70-130	2		20
Chloroform	96		94		70-130	2		20
Carbon tetrachloride	96		95		63-132	1		20
1,2-Dichloropropane	98		98		70-130	0		20
Dibromochloromethane	81		81		63-130	0		20
1,1,2-Trichloroethane	99		98		70-130	1		20
Tetrachloroethene	93		90		70-130	3		20
Chlorobenzene	95		94		75-130	1		20
Trichlorofluoromethane	97		94		62-150	3		20
1,2-Dichloroethane	94		94		70-130	0		20
1,1,1-Trichloroethane	95		94		67-130	1		20
Bromodichloromethane	94		95		67-130	1		20
trans-1,3-Dichloropropene	86		85		70-130	1		20
cis-1,3-Dichloropropene	98		99		70-130	1		20
1,1-Dichloropropene	96		94		70-130	2		20
Bromoform	77		77		54-136	0		20
1,1,2,2-Tetrachloroethane	93		95		67-130	2		20
Benzene	98		97		70-130	1		20
Toluene	95		92		70-130	3		20
Ethylbenzene	95		92		70-130	3		20
Chloromethane	110		110		64-130	0		20
Bromomethane	110		100		39-139	10		20



# Lab Control Sample Analysis

## Batch Quality Control

**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP

**Lab Number:** L1822216

**Project Number:** 701970.01.SA

**Report Date:** 06/19/18

Parameter	LCS %Recovery	Qual	LCSD %Recovery	Qual	%Recovery Limits	RPD	Qual	RPD Limits
Volatile Organics by GC/MS - Westborough Lab Associated sample(s): 02,04 Batch: WG1127131-3 WG1127131-4								
Vinyl chloride	100		100		55-140	0		20
Chloroethane	110		100		55-138	10		20
1,1-Dichloroethene	99		95		61-145	4		20
trans-1,2-Dichloroethene	100		98		70-130	2		20
Trichloroethene	91		92		70-130	1		20
1,2-Dichlorobenzene	98		96		70-130	2		20
1,3-Dichlorobenzene	97		96		70-130	1		20
1,4-Dichlorobenzene	97		95		70-130	2		20
Methyl tert butyl ether	95		97		63-130	2		20
p/m-Xylene	100		95		70-130	5		20
o-Xylene	100		95		70-130	5		20
cis-1,2-Dichloroethene	100		97		70-130	3		20
Dibromomethane	95		96		70-130	1		20
1,2,3-Trichloropropane	88		95		64-130	8		20
Styrene	100		100		70-130	0		20
Dichlorodifluoromethane	100		100		36-147	0		20
Acetone	88		86		58-148	2		20
Carbon disulfide	110		100		51-130	10		20
2-Butanone	90		94		63-138	4		20
Vinyl acetate	97		98		70-130	1		20
4-Methyl-2-pentanone	85		88		59-130	3		20
2-Hexanone	84		85		57-130	1		20
Bromochloromethane	96		99		70-130	3		20

# Lab Control Sample Analysis

## Batch Quality Control

**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP

**Lab Number:** L1822216

**Project Number:** 701970.01.SA

**Report Date:** 06/19/18

Parameter	LCS %Recovery	Qual	LCSD %Recovery	Qual	%Recovery Limits	RPD	Qual	RPD Limits
Volatile Organics by GC/MS - Westborough Lab Associated sample(s): 02,04 Batch: WG1127131-3 WG1127131-4								
2,2-Dichloropropane	100		99		63-133	1		20
1,2-Dibromoethane	94		95		70-130	1		20
1,3-Dichloropropane	95		96		70-130	1		20
1,1,1,2-Tetrachloroethane	95		95		64-130	0		20
Bromobenzene	96		94		70-130	2		20
n-Butylbenzene	100		99		53-136	1		20
sec-Butylbenzene	100		96		70-130	4		20
tert-Butylbenzene	100		96		70-130	4		20
o-Chlorotoluene	120		120		70-130	0		20
p-Chlorotoluene	100		96		70-130	4		20
1,2-Dibromo-3-chloropropane	77		78		41-144	1		20
Hexachlorobutadiene	130		110		63-130	17		20
Isopropylbenzene	98		95		70-130	3		20
p-Isopropyltoluene	100		98		70-130	2		20
Naphthalene	94		87		70-130	8		20
n-Propylbenzene	98		94		69-130	4		20
1,2,3-Trichlorobenzene	120		100		70-130	18		20
1,2,4-Trichlorobenzene	110		99		70-130	11		20
1,3,5-Trimethylbenzene	100		96		64-130	4		20
1,2,4-Trimethylbenzene	100		97		70-130	3		20
1,4-Dioxane	96		102		56-162	6		20

**Lab Control Sample Analysis****Batch Quality Control****Project Name:** FORMER HAMPSHIRE CHEMICAL CORP**Lab Number:** L1822216**Project Number:** 701970.01.SA**Report Date:** 06/19/18

<b>Parameter</b>	<b>LCS %Recovery</b>	<b>Qual</b>	<b>LCSD %Recovery</b>	<b>Qual</b>	<b>%Recovery Limits</b>	<b>RPD</b>	<b>Qual</b>	<b>RPD Limits</b>
------------------	--------------------------	-------------	---------------------------	-------------	-----------------------------	------------	-------------	-----------------------

Volatile Organics by GC/MS - Westborough Lab Associated sample(s): 02,04 Batch: WG1127131-3 WG1127131-4

<b>Surrogate</b>	<b>LCS %Recovery</b>	<b>Qual</b>	<b>LCSD %Recovery</b>	<b>Qual</b>	<b>Acceptance Criteria</b>
1,2-Dichloroethane-d4	96		98		70-130
Toluene-d8	99		98		70-130
4-Bromofluorobenzene	103		101		70-130
Dibromofluoromethane	98		99		70-130

## METALS

**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP**Lab Number:** L1822216**Project Number:** 701970.01.SA**Report Date:** 06/19/18**SAMPLE RESULTS**

Lab ID: L1822216-01

Date Collected: 06/13/18 09:20

Client ID: MW23-061318

Date Received: 06/13/18

Sample Location: WATERLOO, NY

Field Prep: Refer to COC

Sample Depth:

Matrix: Water

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor	Date Prepared	Date Analyzed	Prep Method	Analytical Method	Analyst
<b>Total Metals - Mansfield Lab</b>											
Aluminum, Total	0.0366		mg/l	0.0100	0.00327	1	06/18/18 09:10	06/19/18 12:12	EPA 3005A	1,6020A	AM
Arsenic, Total	0.07332		mg/l	0.00050	0.00016	1	06/18/18 09:10	06/19/18 12:12	EPA 3005A	1,6020A	AM
Calcium, Total	138.		mg/l	0.100	0.0394	1	06/18/18 09:10	06/19/18 12:12	EPA 3005A	1,6020A	AM
Iron, Total	4.81		mg/l	0.0500	0.0191	1	06/18/18 09:10	06/19/18 12:12	EPA 3005A	1,6020A	AM
Magnesium, Total	13.7		mg/l	0.0700	0.0242	1	06/18/18 09:10	06/19/18 12:12	EPA 3005A	1,6020A	AM
Manganese, Total	0.1815		mg/l	0.00100	0.00044	1	06/18/18 09:10	06/19/18 12:12	EPA 3005A	1,6020A	AM
Potassium, Total	6.57		mg/l	0.100	0.0309	1	06/18/18 09:10	06/19/18 12:12	EPA 3005A	1,6020A	AM
Silica, Total	30.2		mg/l	0.500	0.007	1	06/18/18 09:10	06/18/18 16:18	EPA 3005A	19,200.7	LC
Sodium, Total	1050		mg/l	5.00	1.46	50	06/18/18 09:10	06/19/18 12:56	EPA 3005A	1,6020A	AM
<b>Dissolved Metals - Mansfield Lab</b>											
Aluminum, Dissolved	0.0185		mg/l	0.0100	0.00327	1	06/18/18 13:30	06/19/18 12:08	EPA 3005A	1,6020A	AM
Arsenic, Dissolved	0.06287		mg/l	0.00050	0.00016	1	06/18/18 13:30	06/19/18 12:08	EPA 3005A	1,6020A	AM
Iron, Dissolved	0.193		mg/l	0.0500	0.0191	1	06/18/18 13:30	06/19/18 12:08	EPA 3005A	1,6020A	AM
Manganese, Dissolved	0.1704		mg/l	0.00100	0.00044	1	06/18/18 13:30	06/19/18 12:08	EPA 3005A	1,6020A	AM



**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP**Lab Number:** L1822216**Project Number:** 701970.01.SA**Report Date:** 06/19/18**SAMPLE RESULTS**

Lab ID: L1822216-02

Date Collected: 06/08/18 12:00

Client ID: PZ06-060818

Date Received: 06/13/18

Sample Location: WATERLOO, NY

Field Prep: Refer to COC

Sample Depth:

Matrix: Water

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor	Date Prepared	Date Analyzed	Prep Method	Analytical Method	Analyst
Total Metals - Mansfield Lab											
Aluminum, Total	0.319		mg/l	0.0100	0.00327	1	06/18/18 09:10	06/19/18 12:16	EPA 3005A	1,6020A	AM
Arsenic, Total	0.00462		mg/l	0.00050	0.00016	1	06/18/18 09:10	06/19/18 12:16	EPA 3005A	1,6020A	AM
Calcium, Total	25.2		mg/l	0.100	0.0394	1	06/18/18 09:10	06/19/18 12:16	EPA 3005A	1,6020A	AM
Chromium, Total	0.00254		mg/l	0.00100	0.00017	1	06/18/18 09:10	06/19/18 12:16	EPA 3005A	1,6020A	AM
Iron, Total	0.502		mg/l	0.0500	0.0191	1	06/18/18 09:10	06/19/18 12:16	EPA 3005A	1,6020A	AM
Magnesium, Total	4.12		mg/l	0.0700	0.0242	1	06/18/18 09:10	06/19/18 12:16	EPA 3005A	1,6020A	AM
Manganese, Total	0.01688		mg/l	0.00100	0.00044	1	06/18/18 09:10	06/19/18 12:16	EPA 3005A	1,6020A	AM
Potassium, Total	3.87		mg/l	0.100	0.0309	1	06/18/18 09:10	06/19/18 12:16	EPA 3005A	1,6020A	AM
Silica, Total	14.6		mg/l	0.500	0.007	1	06/18/18 09:10	06/18/18 16:23	EPA 3005A	19,200.7	LC
Sodium, Total	240.		mg/l	0.100	0.0293	1	06/18/18 09:10	06/19/18 12:16	EPA 3005A	1,6020A	AM
Dissolved Metals - Mansfield Lab											
Aluminum, Dissolved	0.0321		mg/l	0.0100	0.00327	1	06/18/18 13:30	06/19/18 12:45	EPA 3005A	1,6020A	AM
Arsenic, Dissolved	0.00431		mg/l	0.00050	0.00016	1	06/18/18 13:30	06/19/18 12:45	EPA 3005A	1,6020A	AM
Chromium, Dissolved	0.00250		mg/l	0.00100	0.00017	1	06/18/18 13:30	06/19/18 12:45	EPA 3005A	1,6020A	AM
Iron, Dissolved	0.0505		mg/l	0.0500	0.0191	1	06/18/18 13:30	06/19/18 12:45	EPA 3005A	1,6020A	AM
Manganese, Dissolved	0.00561		mg/l	0.00100	0.00044	1	06/18/18 13:30	06/19/18 12:45	EPA 3005A	1,6020A	AM



**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP**Lab Number:** L1822216**Project Number:** 701970.01.SA**Report Date:** 06/19/18**SAMPLE RESULTS**

Lab ID: L1822216-05

Date Collected: 06/13/18 09:50

Client ID: MW24-061318

Date Received: 06/13/18

Sample Location: WATERLOO, NY

Field Prep: Refer to COC

Sample Depth:

Matrix: Water

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor	Date Prepared	Date Analyzed	Prep Method	Analytical Method	Analyst
<b>Total Metals - Mansfield Lab</b>											
Aluminum, Total	0.0696		mg/l	0.0100	0.00327	1	06/18/18 09:10	06/19/18 12:20	EPA 3005A	1,6020A	AM
Arsenic, Total	0.00804		mg/l	0.00050	0.00016	1	06/18/18 09:10	06/19/18 12:20	EPA 3005A	1,6020A	AM
Calcium, Total	192.		mg/l	0.100	0.0394	1	06/18/18 09:10	06/19/18 12:20	EPA 3005A	1,6020A	AM
Iron, Total	10.8		mg/l	0.0500	0.0191	1	06/18/18 09:10	06/19/18 12:20	EPA 3005A	1,6020A	AM
Magnesium, Total	81.8		mg/l	0.0700	0.0242	1	06/18/18 09:10	06/19/18 12:20	EPA 3005A	1,6020A	AM
Manganese, Total	0.1620		mg/l	0.00100	0.00044	1	06/18/18 09:10	06/19/18 12:20	EPA 3005A	1,6020A	AM
Potassium, Total	5.66		mg/l	0.100	0.0309	1	06/18/18 09:10	06/19/18 12:20	EPA 3005A	1,6020A	AM
Silica, Total	27.9		mg/l	0.500	0.007	1	06/18/18 09:10	06/18/18 16:27	EPA 3005A	19,200.7	LC
Sodium, Total	790.		mg/l	5.00	1.46	50	06/18/18 09:10	06/19/18 13:00	EPA 3005A	1,6020A	AM
<b>Dissolved Metals - Mansfield Lab</b>											
Aluminum, Dissolved	0.0118		mg/l	0.0100	0.00327	1	06/18/18 13:30	06/19/18 12:49	EPA 3005A	1,6020A	AM
Arsenic, Dissolved	0.00538		mg/l	0.00050	0.00016	1	06/18/18 13:30	06/19/18 12:49	EPA 3005A	1,6020A	AM
Iron, Dissolved	5.20		mg/l	0.0500	0.0191	1	06/18/18 13:30	06/19/18 12:49	EPA 3005A	1,6020A	AM
Manganese, Dissolved	0.2038		mg/l	0.00100	0.00044	1	06/18/18 13:30	06/19/18 12:49	EPA 3005A	1,6020A	AM



Project Name: FORMER HAMPSHIRE CHEMICAL CORP

Lab Number: L1822216

Project Number: 701970.01.SA

Report Date: 06/19/18

## Method Blank Analysis Batch Quality Control

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor	Date Prepared	Date Analyzed	Analytical Method	Analyst
Total Metals - Mansfield Lab for sample(s): 01-02,05 Batch: WG1126960-1										
Silica, Total	0.025	J	mg/l	0.500	0.007	1	06/18/18 09:10	06/18/18 15:27	19,200.7	LC

### Prep Information

Digestion Method: EPA 3005A

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor	Date Prepared	Date Analyzed	Analytical Method	Analyst
Total Metals - Mansfield Lab for sample(s): 01-02,05 Batch: WG1126964-1										
Aluminum, Total	ND		mg/l	0.0100	0.00327	1	06/18/18 09:10	06/19/18 10:12	1,6020A	AM
Arsenic, Total	ND		mg/l	0.00050	0.00016	1	06/18/18 09:10	06/19/18 10:12	1,6020A	AM
Calcium, Total	ND		mg/l	0.100	0.0394	1	06/18/18 09:10	06/19/18 10:12	1,6020A	AM
Chromium, Total	ND		mg/l	0.00100	0.00017	1	06/18/18 09:10	06/19/18 10:12	1,6020A	AM
Iron, Total	0.0199	J	mg/l	0.0500	0.0191	1	06/18/18 09:10	06/19/18 10:12	1,6020A	AM
Magnesium, Total	ND		mg/l	0.0700	0.0242	1	06/18/18 09:10	06/19/18 10:12	1,6020A	AM
Manganese, Total	ND		mg/l	0.00100	0.00044	1	06/18/18 09:10	06/19/18 10:12	1,6020A	AM
Potassium, Total	ND		mg/l	0.100	0.0309	1	06/18/18 09:10	06/19/18 10:12	1,6020A	AM
Sodium, Total	ND		mg/l	0.100	0.0293	1	06/18/18 09:10	06/19/18 10:12	1,6020A	AM

### Prep Information

Digestion Method: EPA 3005A

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor	Date Prepared	Date Analyzed	Analytical Method	Analyst
Dissolved Metals - Mansfield Lab for sample(s): 01-02,05 Batch: WG1127029-1										
Aluminum, Dissolved	ND		mg/l	0.0100	0.00327	1	06/18/18 13:30	06/19/18 11:49	1,6020A	AM
Arsenic, Dissolved	ND		mg/l	0.00050	0.00016	1	06/18/18 13:30	06/19/18 11:49	1,6020A	AM
Chromium, Dissolved	ND		mg/l	0.00100	0.00017	1	06/18/18 13:30	06/19/18 11:49	1,6020A	AM
Iron, Dissolved	ND		mg/l	0.0500	0.0191	1	06/18/18 13:30	06/19/18 11:49	1,6020A	AM
Manganese, Dissolved	ND		mg/l	0.00100	0.00044	1	06/18/18 13:30	06/19/18 11:49	1,6020A	AM





**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP

**Lab Number:** L1822216

**Project Number:** 701970.01.SA

**Report Date:** 06/19/18

## **Method Blank Analysis Batch Quality Control**

### **Prep Information**

---

Digestion Method: EPA 3005A

# Lab Control Sample Analysis

## Batch Quality Control

**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP

**Lab Number:** L1822216

**Project Number:** 701970.01.SA

**Report Date:** 06/19/18

Parameter	LCS %Recovery	Qual	LCSD %Recovery	Qual	%Recovery Limits	RPD	Qual	RPD Limits
Total Metals - Mansfield Lab Associated sample(s): 01-02,05 Batch: WG1126960-2								
Silica, Total	99		-		85-115	-		
Total Metals - Mansfield Lab Associated sample(s): 01-02,05 Batch: WG1126964-2								
Aluminum, Total	105		-		80-120	-		
Arsenic, Total	112		-		80-120	-		
Calcium, Total	102		-		80-120	-		
Chromium, Total	103		-		80-120	-		
Iron, Total	115		-		80-120	-		
Magnesium, Total	107		-		80-120	-		
Manganese, Total	103		-		80-120	-		
Potassium, Total	104		-		80-120	-		
Sodium, Total	104		-		80-120	-		
Dissolved Metals - Mansfield Lab Associated sample(s): 01-02,05 Batch: WG1127029-2								
Aluminum, Dissolved	104		-		80-120	-		
Arsenic, Dissolved	109		-		80-120	-		
Chromium, Dissolved	103		-		80-120	-		
Iron, Dissolved	118		-		80-120	-		
Manganese, Dissolved	102		-		80-120	-		

# **Matrix Spike Analysis** **Batch Quality Control**

**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP  
**Project Number:** 701970.01.SA

**Lab Number:** L1822216  
**Report Date:** 06/19/18

Parameter	Native Sample	MS Added	MS Found	MS %Recovery	Qual	MSD Found	MSD %Recovery	Qual	Recovery Limits	RPD	Qual	RPD Limits
-----------	---------------	----------	----------	--------------	------	-----------	---------------	------	-----------------	-----	------	------------

Total Metals - Mansfield Lab Associated sample(s): 01-02,05 QC Batch ID: WG1126960-3 QC Sample: L1821563-01 Client ID: MS Sample

Silica, Total	1.50	2.14	5.58	191	Q	-	-		75-125	-		20
---------------	------	------	------	-----	---	---	---	--	--------	---	--	----

Total Metals - Mansfield Lab Associated sample(s): 01-02,05 QC Batch ID: WG1126964-3 WG1126964-4 QC Sample: L1821907-01 Client ID: MS Sample

Aluminum, Total	0.931	2	3.21	114		3.20	113		75-125	0		20
Arsenic, Total	0.00062	0.12	0.1412	117		0.1384	115		75-125	2		20
Calcium, Total	55.8	10	69.2	134	Q	68.2	124		75-125	1		20
Chromium, Total	0.00290	0.2	0.2186	108		0.2180	108		75-125	0		20
Iron, Total	3.57	1	4.70	113		4.77	120		75-125	1		20
Magnesium, Total	21.0	10	34.0	130	Q	33.2	122		75-125	2		20
Manganese, Total	4.926	0.5	5.599	135	Q	5.618	138	Q	75-125	0		20
Potassium, Total	3.08	10	13.5	104		14.0	109		75-125	4		20
Sodium, Total	19.2	10	31.0	118		30.4	112		75-125	2		20

Dissolved Metals - Mansfield Lab Associated sample(s): 01-02,05 QC Batch ID: WG1127029-3 QC Sample: L1822216-01 Client ID: MW23-061318

Aluminum, Dissolved	0.0185	2	1.99	98		-	-		75-125	-		20
Arsenic, Dissolved	0.06287	0.12	0.1961	111		-	-		75-125	-		20
Chromium, Dissolved	0.0056	0.2	0.2018	98		-	-		75-125	-		20
Iron, Dissolved	0.193	1	1.31	112		-	-		75-125	-		20
Manganese, Dissolved	0.1704	0.5	0.6654	99		-	-		75-125	-		20

**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP

**Project Number:** 701970.01.SA

## Lab Duplicate Analysis

*Batch Quality Control*

**Lab Number:** L1822216

**Report Date:** 06/19/18

Parameter	Native Sample	Duplicate Sample	Units	RPD	Qual	RPD Limits
Dissolved Metals - Mansfield Lab Associated sample(s): 01-02,05 QC Batch ID: WG1127029-4 QC Sample: L1822216-01 Client ID: MW23-061318						
Aluminum, Dissolved	0.0185	0.0194	mg/l	5		20
Arsenic, Dissolved	0.06287	0.06467	mg/l	3		20
Iron, Dissolved	0.193	0.209	mg/l	8		20
Manganese, Dissolved	0.1704	0.1750	mg/l	3		20

# **INORGANICS & MISCELLANEOUS**

**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP  
**Project Number:** 701970.01.SA

**Lab Number:** L1822216  
**Report Date:** 06/19/18

### SAMPLE RESULTS

**Lab ID:** L1822216-01  
**Client ID:** MW23-061318  
**Sample Location:** WATERLOO, NY

**Date Collected:** 06/13/18 09:20  
**Date Received:** 06/13/18  
**Field Prep:** Refer to COC

**Sample Depth:**  
**Matrix:** Water

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor	Date Prepared	Date Analyzed	Analytical Method	Analyst
<b>General Chemistry - Westborough Lab</b>										
Alkalinity, Total	1050		mg CaCO3/L	10.0	NA	5	-	06/15/18 08:35	121,2320B	BR
Solids, Total Dissolved	3200		mg/l	10	3.1	1	-	06/14/18 12:10	121,2540C	SD
Nitrogen, Ammonia	4.65		mg/l	0.075	0.024	1	06/17/18 16:00	06/18/18 22:01	44,350.1	AT
Nitrogen, Nitrate	0.035	J	mg/l	0.10	0.033	1	-	06/14/18 21:07	44,353.2	MR
Nitrogen, Total Kjeldahl	5.29		mg/l	0.300	0.066	1	06/14/18 16:30	06/19/18 00:05	4,351.3/.1 (M)	AT
Phosphorus, Total	1.44		mg/l	0.020	0.006	2	06/18/18 09:40	06/18/18 14:20	121,4500P-E	SD
Phosphorus, Orthophosphate	0.413		mg/l	0.005	0.001	1	-	06/14/18 22:41	121,4500P-E	AS
Sulfide	5.8		mg/l	1.0	1.0	10	06/14/18 19:30	06/15/18 03:23	121,4500S2-AD	CW
Total Organic Carbon	13.0		mg/l	2.50	0.570	5	-	06/17/18 11:16	121,5310C	AG
<b>Anions by Ion Chromatography - Westborough Lab</b>										
Chloride	246.		mg/l	50.0	8.39	100	-	06/14/18 21:07	44,300.0	AU
Sulfate	1140		mg/l	100	16.0	100	-	06/14/18 21:07	44,300.0	AU



**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP  
**Project Number:** 701970.01.SA

**Lab Number:** L1822216  
**Report Date:** 06/19/18

### SAMPLE RESULTS

**Lab ID:** L1822216-02  
**Client ID:** PZ06-060818  
**Sample Location:** WATERLOO, NY

**Date Collected:** 06/08/18 12:00  
**Date Received:** 06/13/18  
**Field Prep:** Refer to COC

**Sample Depth:**  
**Matrix:** Water

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor	Date Prepared	Date Analyzed	Analytical Method	Analyst
General Chemistry - Westborough Lab										
Alkalinity, Total	295.		mg CaCO3/L	2.00	NA	1	-	06/15/18 08:35	121,2320B	BR
Solids, Total Dissolved	950		mg/l	10	3.1	1	-	06/14/18 12:10	121,2540C	SD
Nitrogen, Ammonia	0.372		mg/l	0.075	0.024	1	06/17/18 16:00	06/18/18 22:02	44,350.1	AT
Nitrogen, Total Kjeldahl	0.755		mg/l	0.300	0.066	1	06/14/18 16:30	06/19/18 00:06	4,351.3/.1 (M)	AT
Phosphorus, Total	0.230		mg/l	0.020	0.006	2	06/18/18 09:40	06/18/18 14:21	121,4500P-E	SD
Sulfide	0.15		mg/l	0.10	0.10	1	06/14/18 19:30	06/15/18 03:24	121,4500S2-AD	CW
Total Organic Carbon	6.79		mg/l	2.50	0.570	5	-	06/17/18 11:16	121,5310C	AG



**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP  
**Project Number:** 701970.01.SA

**Lab Number:** L1822216  
**Report Date:** 06/19/18

### SAMPLE RESULTS

**Lab ID:** L1822216-03  
**Client ID:** PZ06-061318  
**Sample Location:** WATERLOO, NY

**Date Collected:** 06/13/18 10:25  
**Date Received:** 06/13/18  
**Field Prep:** Not Specified

**Sample Depth:**  
**Matrix:** Water

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor	Date Prepared	Date Analyzed	Analytical Method	Analyst
General Chemistry - Westborough Lab										
Nitrogen, Nitrate	0.17		mg/l	0.10	0.033	1	-	06/14/18 21:11	44,353.2	MR
Phosphorus, Orthophosphate	0.118		mg/l	0.005	0.001	1	-	06/14/18 22:41	121,4500P-E	AS
Anions by Ion Chromatography - Westborough Lab										
Chloride	254.		mg/l	12.5	2.10	25	-	06/15/18 00:55	44,300.0	AU
Sulfate	93.5		mg/l	25.0	4.00	25	-	06/15/18 00:55	44,300.0	AU





**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP  
**Project Number:** 701970.01.SA

**Lab Number:** L1822216  
**Report Date:** 06/19/18

### SAMPLE RESULTS

**Lab ID:** L1822216-05  
**Client ID:** MW24-061318  
**Sample Location:** WATERLOO, NY

**Date Collected:** 06/13/18 09:50  
**Date Received:** 06/13/18  
**Field Prep:** Refer to COC

**Sample Depth:**  
**Matrix:** Water

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor	Date Prepared	Date Analyzed	Analytical Method	Analyst
<b>General Chemistry - Westborough Lab</b>										
Alkalinity, Total	1270		mg CaCO <sub>3</sub> /L	10.0	NA	5	-	06/15/18 08:35	121,2320B	BR
Solids, Total Dissolved	2900		mg/l	10	3.1	1	-	06/14/18 12:10	121,2540C	SD
Nitrogen, Ammonia	2.17		mg/l	0.075	0.024	1	06/17/18 16:00	06/18/18 22:03	44,350.1	AT
Nitrogen, Nitrate	0.047	J	mg/l	0.10	0.033	1	-	06/14/18 21:12	44,353.2	MR
Nitrogen, Total Kjeldahl	2.17		mg/l	0.300	0.066	1	06/14/18 16:30	06/19/18 00:07	4,351.3/.1 (M)	AT
Phosphorus, Total	0.101		mg/l	0.010	0.003	1	06/18/18 09:40	06/18/18 14:22	121,4500P-E	SD
Phosphorus, Orthophosphate	0.012		mg/l	0.005	0.001	1	-	06/14/18 22:42	121,4500P-E	AS
Sulfide	ND		mg/l	0.10	0.10	1	06/14/18 19:30	06/15/18 03:24	121,4500S2-AD	CW
Total Organic Carbon	11.3		mg/l	2.50	0.570	5	-	06/18/18 13:57	121,5310C	DW
<b>Anions by Ion Chromatography - Westborough Lab</b>										
Chloride	593.		mg/l	50.0	8.39	100	-	06/14/18 21:31	44,300.0	AU
Sulfate	213.		mg/l	100	16.0	100	-	06/14/18 21:31	44,300.0	AU



Project Name: FORMER HAMPSHIRE CHEMICAL CO

Lab Number: L1822216

Project Number: 701970.01.SA

Report Date: 06/19/18

### Method Blank Analysis Batch Quality Control

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor	Date Prepared	Date Analyzed	Analytical Method	Analyst
General Chemistry - Westborough Lab for sample(s): 02 Batch: WG1125825-1										
Solids, Total Dissolved	ND		mg/l	10	3.1	1	-	06/14/18 12:10	121,2540C	SD
General Chemistry - Westborough Lab for sample(s): 01,05 Batch: WG1125828-1										
Solids, Total Dissolved	ND		mg/l	10	3.1	1	-	06/14/18 12:10	121,2540C	SD
General Chemistry - Westborough Lab for sample(s): 01,03,05 Batch: WG1126082-1										
Nitrogen, Nitrate	ND		mg/l	0.10	0.033	1	-	06/14/18 20:12	44,353.2	MR
General Chemistry - Westborough Lab for sample(s): 01-02,05 Batch: WG1126118-1										
Sulfide	ND		mg/l	0.10	0.10	1	06/14/18 19:30	06/15/18 03:18	121,4500S2-AD	CW
General Chemistry - Westborough Lab for sample(s): 01,03,05 Batch: WG1126155-1										
Phosphorus, Orthophosphate	0.002	J	mg/l	0.005	0.001	1	-	06/14/18 22:40	121,4500P-E	AS
General Chemistry - Westborough Lab for sample(s): 01-02,05 Batch: WG1126293-1										
Alkalinity, Total	ND		mg CaCO3/L	2.00	NA	1	-	06/15/18 08:35	121,2320B	BR
General Chemistry - Westborough Lab for sample(s): 01-02,05 Batch: WG1126328-1										
Nitrogen, Total Kjeldahl	0.134	J	mg/l	0.300	0.022	1	06/14/18 16:30	06/18/18 23:59	4,351.3/.1 (M)	AT
Anions by Ion Chromatography - Westborough Lab for sample(s): 01,03,05 Batch: WG1126596-1										
Chloride	ND		mg/l	0.500	0.083	1	-	06/14/18 18:07	44,300.0	AU
Sulfate	ND		mg/l	1.00	0.160	1	-	06/14/18 18:07	44,300.0	AU
General Chemistry - Westborough Lab for sample(s): 01-02 Batch: WG1126827-1										
Total Organic Carbon	0.160	J	mg/l	0.500	0.114	1	-	06/17/18 11:16	121,5310C	AG
General Chemistry - Westborough Lab for sample(s): 01-02,05 Batch: WG1126858-1										
Nitrogen, Ammonia	0.024	J	mg/l	0.075	0.024	1	06/17/18 16:00	06/18/18 21:57	44,350.1	AT
General Chemistry - Westborough Lab for sample(s): 01-02,05 Batch: WG1126983-1										
Phosphorus, Total	0.004	J	mg/l	0.010	0.003	1	06/18/18 09:40	06/18/18 14:20	121,4500P-E	SD
General Chemistry - Westborough Lab for sample(s): 05 Batch: WG1126993-1										
Total Organic Carbon	ND		mg/l	0.500	0.114	1	-	06/18/18 13:57	121,5310C	DW

**Lab Control Sample Analysis****Batch Quality Control****Project Name:** FORMER HAMPSHIRE CHEMICAL CORP**Lab Number:** L1822216**Project Number:** 701970.01.SA**Report Date:** 06/19/18

Parameter	LCS %Recovery	Qual	LCSD %Recovery	Qual	%Recovery Limits	RPD	Qual	RPD Limits
General Chemistry - Westborough Lab Associated sample(s): 02 Batch: WG1125825-2								
Solids, Total Dissolved	106		-		80-120	-		
General Chemistry - Westborough Lab Associated sample(s): 01,05 Batch: WG1125828-2								
Solids, Total Dissolved	106		-		80-120	-		
General Chemistry - Westborough Lab Associated sample(s): 01,03,05 Batch: WG1126082-2								
Nitrogen, Nitrate	96		-		90-110	-		
General Chemistry - Westborough Lab Associated sample(s): 01-02,05 Batch: WG1126118-2								
Sulfide	104		-		75-125	-		
General Chemistry - Westborough Lab Associated sample(s): 01,03,05 Batch: WG1126155-2								
Phosphorus, Orthophosphate	100		-		90-110	-		
General Chemistry - Westborough Lab Associated sample(s): 01-02,05 Batch: WG1126293-2								
Alkalinity, Total	103		-		90-110	-		10
General Chemistry - Westborough Lab Associated sample(s): 01-02,05 Batch: WG1126328-2								
Nitrogen, Total Kjeldahl	97		-		78-122	-		

# Lab Control Sample Analysis

## Batch Quality Control

**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP

**Lab Number:** L1822216

**Project Number:** 701970.01.SA

**Report Date:** 06/19/18

Parameter	LCS %Recovery	LCSD %Recovery	%Recovery Limits	RPD	RPD Limits
Anions by Ion Chromatography - Westborough Lab Associated sample(s): 01,03,05 Batch: WG1126596-2					
Chloride	102	-	90-110	-	
Sulfate	101	-	90-110	-	
General Chemistry - Westborough Lab Associated sample(s): 01-02 Batch: WG1126827-2					
Total Organic Carbon	95	-	90-110	-	
General Chemistry - Westborough Lab Associated sample(s): 01-02,05 Batch: WG1126858-2					
Nitrogen, Ammonia	92	-	90-110	-	20
General Chemistry - Westborough Lab Associated sample(s): 01-02,05 Batch: WG1126983-2					
Phosphorus, Total	98	-	80-120	-	
General Chemistry - Westborough Lab Associated sample(s): 05 Batch: WG1126993-2					
Total Organic Carbon	98	-	90-110	-	

# **Matrix Spike Analysis** **Batch Quality Control**

**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP  
**Project Number:** 701970.01.SA

**Lab Number:** L1822216  
**Report Date:** 06/19/18

Parameter	Native Sample	MS Added	MS Found	MS %Recovery	Qual	MSD Found	MSD %Recovery	Qual	Recovery Limits	RPD	Qual	RPD Limits
General Chemistry - Westborough Lab Associated sample(s): 01,03,05				QC Batch ID: WG1126082-4			QC Sample: L1822216-01			Client ID: MW23-061318		
Nitrogen, Nitrate	0.035J	4	3.7	92		-	-		83-113	-		6
General Chemistry - Westborough Lab Associated sample(s): 01-02,05				QC Batch ID: WG1126118-4			QC Sample: L1800006-100			Client ID: MS Sample		
Sulfide	ND	0.55	ND	0	Q	-	-		70-130	-		20
General Chemistry - Westborough Lab Associated sample(s): 01,03,05				QC Batch ID: WG1126155-4			QC Sample: L1822216-05			Client ID: MW24-061318		
Phosphorus, Orthophosphate	0.012	2.5	2.63	105		-	-		80-120	-		20
General Chemistry - Westborough Lab Associated sample(s): 01-02,05				QC Batch ID: WG1126293-4			QC Sample: L1821989-01			Client ID: MS Sample		
Alkalinity, Total	42.9	100	144	101		-	-		86-116	-		10
General Chemistry - Westborough Lab Associated sample(s): 01-02,05				QC Batch ID: WG1126328-4			QC Sample: L1821948-23			Client ID: MS Sample		
Nitrogen, Total Kjeldahl	1.71	8	9.15	93		-	-		77-111	-		24
Anions by Ion Chromatography - Westborough Lab Associated sample(s): 01,03,05				QC Batch ID: WG1126596-3			QC Sample: L1821978-05			Client ID: MS Sample		
Chloride	618.	100	726	108		-	-		90-110	-		18
Sulfate	85.5	200	303	109		-	-		90-110	-		20
General Chemistry - Westborough Lab Associated sample(s): 01-02				QC Batch ID: WG1126827-4			QC Sample: L1821991-02			Client ID: MS Sample		
Total Organic Carbon	1.38	4	5.00	90		-	-		80-120	-		20
General Chemistry - Westborough Lab Associated sample(s): 01-02,05				QC Batch ID: WG1126858-4			QC Sample: L1822216-05			Client ID: MW24-061318		
Nitrogen, Ammonia	2.17	4	5.43	82	Q	-	-		90-110	-		20
General Chemistry - Westborough Lab Associated sample(s): 01-02,05				QC Batch ID: WG1126983-3			QC Sample: L1822216-01			Client ID: MW23-061318		
Phosphorus, Total	1.44	1	2.36	92		-	-		75-125	-		20

**Matrix Spike Analysis**  
Batch Quality Control

**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP  
**Project Number:** 701970.01.SA

**Lab Number:** L1822216  
**Report Date:** 06/19/18

Parameter	Native Sample	MS Added	MS Found	MS %Recovery	MSD Found	MSD %Recovery	Recovery Limits	RPD	RPD Limits
General Chemistry - Westborough Lab Associated sample(s): 05 QC Batch ID: WG1126993-4 QC Sample: L1822216-05 Client ID: MW24-061318									
Total Organic Carbon	11.3	40	50.5	98	-	-	80-120	-	20

**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP  
**Project Number:** 701970.01.SA

**Lab Duplicate Analysis**  
*Batch Quality Control*

**Lab Number:** L1822216  
**Report Date:** 06/19/18

Parameter	Native Sample	Duplicate Sample	Units	RPD	Qual	RPD Limits
General Chemistry - Westborough Lab Associated sample(s): 02 QC Batch ID: WG1125825-3 QC Sample: L1821665-02 Client ID: DUP Sample						
Solids, Total Dissolved	150	240	mg/l	46	Q	10
General Chemistry - Westborough Lab Associated sample(s): 01,05 QC Batch ID: WG1125828-3 QC Sample: L1821999-08 Client ID: DUP Sample						
Solids, Total Dissolved	490	490	mg/l	0		10
General Chemistry - Westborough Lab Associated sample(s): 01,03,05 QC Batch ID: WG1126082-3 QC Sample: L1822216-01 Client ID: MW23-061318						
Nitrogen, Nitrate	0.035J	0.062J	mg/l	NC		6
General Chemistry - Westborough Lab Associated sample(s): 01-02,05 QC Batch ID: WG1126118-3 QC Sample: L1800006-100 Client ID: DUP Sample						
Sulfide	ND	ND	mg/l	NC		20
General Chemistry - Westborough Lab Associated sample(s): 01,03,05 QC Batch ID: WG1126155-3 QC Sample: L1822216-01 Client ID: MW23-061318						
Phosphorus, Orthophosphate	0.413	0.412	mg/l	0		20
General Chemistry - Westborough Lab Associated sample(s): 01-02,05 QC Batch ID: WG1126293-3 QC Sample: L1821989-02 Client ID: DUP Sample						
Alkalinity, Total	63.2	62.7	mg CaCO <sub>3</sub> /L	1		10
General Chemistry - Westborough Lab Associated sample(s): 01-02,05 QC Batch ID: WG1126328-3 QC Sample: L1821948-23 Client ID: DUP Sample						
Nitrogen, Total Kjeldahl	1.71	1.84	mg/l	7		24
Anions by Ion Chromatography - Westborough Lab Associated sample(s): 01,03,05 QC Batch ID: WG1126596-4 QC Sample: L1821978-05 Client ID: DUP Sample						
Chloride	618.	618	mg/l	0		18
Sulfate	85.5	87.0	mg/l	2		20

# **Lab Duplicate Analysis** *Batch Quality Control*

**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP

**Project Number:** 701970.01.SA

**Lab Number:** L1822216

**Report Date:** 06/19/18

Parameter	Native Sample	Duplicate Sample	Units	RPD	RPD Limits
General Chemistry - Westborough Lab Associated sample(s): 01-02 QC Batch ID: WG1126827-3 QC Sample: L1821991-01 Client ID: DUP Sample					
Total Organic Carbon	1.82	1.79	mg/l	2	20
General Chemistry - Westborough Lab Associated sample(s): 01-02,05 QC Batch ID: WG1126858-3 QC Sample: L1822216-05 Client ID: MW24-061318					
Nitrogen, Ammonia	2.17	2.19	mg/l	1	20
General Chemistry - Westborough Lab Associated sample(s): 01-02,05 QC Batch ID: WG1126983-4 QC Sample: L1822216-01 Client ID: MW23-061318					
Phosphorus, Total	1.44	1.52	mg/l	5	20
General Chemistry - Westborough Lab Associated sample(s): 05 QC Batch ID: WG1126993-3 QC Sample: L1822216-05 Client ID: MW24-061318					
Total Organic Carbon	11.3	11.4	mg/l	1	20



**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP**Lab Number:** L1822216**Project Number:** 701970.01.SA**Report Date:** 06/19/18**Sample Receipt and Container Information**

Were project specific reporting limits specified?

YES

**Cooler Information**

Cooler	Custody Seal
A	Absent
B	Absent

**Container Information**

Container ID	Container Type	Cooler	Initial pH	Final pH	Temp deg C	Pres	Seal	Frozen Date/Time	Analysis(*)
L1822216-01D	Vial H2SO4 preserved	B	NA		3.3	Y	Absent		TOC-5310(28)
L1822216-01E	Vial H2SO4 preserved	B	NA		3.3	Y	Absent		TOC-5310(28)
L1822216-01F	Vial H2SO4 preserved	B	NA		3.3	Y	Absent		TOC-5310(28)
L1822216-01G	Plastic 250ml unpreserved/No Headspace	B	NA		3.3	Y	Absent		ALK-T-2320(14)
L1822216-01H	Plastic 250ml unpreserved	B	7	7	3.3	Y	Absent		OPHOS-4500(2),SO4-300(28),CL-300(28),NO3-353(2),TDS-2540(7)
L1822216-01I	Plastic 250ml unpreserved	B	7	7	3.3	Y	Absent		OPHOS-4500(2),SO4-300(28),CL-300(28),NO3-353(2),TDS-2540(7)
L1822216-01J	Plastic 250ml HNO3 preserved	B	<2	<2	3.3	Y	Absent		MN-6020S(180),FE-6020S(180),AS-6020S(180),AL-6020S(180)
L1822216-01K	Plastic 250ml HNO3 preserved	B	<2	<2	3.3	Y	Absent		FE-6020T(180),CA-6020T(180),K-6020T(180),NA-6020T(180),MN-6020T(180),AS-6020T(180),SO-UI(180),AL-6020T(180),MG-6020T(180)
L1822216-01L	Plastic 250ml Zn Acetate/NaOH preserved	B	>9	>9	3.3	Y	Absent		SULFIDE-4500(7)
L1822216-01M	Plastic 250ml Zn Acetate/NaOH preserved	B	>9	>9	3.3	Y	Absent		SULFIDE-4500(7)
L1822216-01N	Plastic 500ml unpreserved	B	7	7	3.3	Y	Absent		OPHOS-4500(2),SO4-300(28),CL-300(28),NO3-353(2),TDS-2540(7)
L1822216-01O	Plastic 500ml H2SO4 preserved	B	<2	<2	3.3	Y	Absent		TKN-351(28),TPHOS-4500(28),NH3-350(28)
L1822216-01P	Plastic 500ml H2SO4 preserved	B	<2	<2	3.3	Y	Absent		TKN-351(28),TPHOS-4500(28),NH3-350(28)
L1822216-02A	Vial HCl preserved	B	NA		3.3	Y	Absent		NYTCL-8260(14)
L1822216-02B	Vial HCl preserved	B	NA		3.3	Y	Absent		NYTCL-8260(14)
L1822216-02C	Vial HCl preserved	B	NA		3.3	Y	Absent		NYTCL-8260(14)
L1822216-02D	Vial H2SO4 preserved	B	NA		3.3	Y	Absent		TOC-5310(28)
L1822216-02E	Vial H2SO4 preserved	B	NA		3.3	Y	Absent		TOC-5310(28)
L1822216-02F	Vial H2SO4 preserved	B	NA		3.3	Y	Absent		TOC-5310(28)

**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP  
**Project Number:** 701970.01.SA

**Serial\_No:** 06191821:16  
**Lab Number:** L1822216  
**Report Date:** 06/19/18

**Container Information**

Container ID	Container Type	Cooler	Initial pH	Final pH	Temp deg C	Pres	Seal	Frozen Date/Time	Analysis(*)
L1822216-02G	Plastic 250ml unpreserved/No Headspace	B	NA		3.3	Y	Absent		ALK-T-2320(14)
L1822216-02H	Plastic 250ml unpreserved	B	7	7	3.3	Y	Absent		TDS-2540(7)
L1822216-02J	Plastic 250ml HNO3 preserved	B	<2	<2	3.3	Y	Absent		MN-6020S(180),CR-6020S(180),FE-6020S(180),AS-6020S(180),AL-6020S(180)
L1822216-02K	Plastic 250ml HNO3 preserved	B	<2	<2	3.3	Y	Absent		FE-6020T(180),CA-6020T(180),CR-6020T(180),K-6020T(180),NA-6020T(180),MN-6020T(180),AS-6020T(180),SO-UI(180),AL-6020T(180),MG-6020T(180)
L1822216-02L	Plastic 250ml Zn Acetate/NaOH preserved	B	>9	>9	3.3	Y	Absent		SULFIDE-4500(7)
L1822216-02M	Plastic 250ml Zn Acetate/NaOH preserved	B	>9	>9	3.3	Y	Absent		SULFIDE-4500(7)
L1822216-02O	Plastic 500ml H2SO4 preserved	B	<2	<2	3.3	Y	Absent		TKN-351(28),TPHOS-4500(28),NH3-350(28)
L1822216-02P	Plastic 500ml H2SO4 preserved	B	<2	<2	3.3	Y	Absent		TKN-351(28),TPHOS-4500(28),NH3-350(28)
L1822216-03A	Plastic 250ml unpreserved	B	7	7	3.3	Y	Absent		OPHOS-4500(2)
L1822216-03B	Plastic 500ml unpreserved	B	7	7	3.3	Y	Absent		SO4-300(28),CL-300(28),NO3-353(2)
L1822216-04A	Vial HCl preserved	B	NA		3.3	Y	Absent		NYTCL-8260(14)
L1822216-04B	Vial HCl preserved	B	NA		3.3	Y	Absent		NYTCL-8260(14)
L1822216-05D	Vial H2SO4 preserved	B	NA		3.3	Y	Absent		TOC-5310(28)
L1822216-05E	Vial H2SO4 preserved	B	NA		3.3	Y	Absent		TOC-5310(28)
L1822216-05F	Vial H2SO4 preserved	B	NA		3.3	Y	Absent		TOC-5310(28)
L1822216-05G	Plastic 250ml unpreserved/No Headspace	B	NA		3.3	Y	Absent		ALK-T-2320(14)
L1822216-05H	Plastic 250ml unpreserved	B	7	7	3.3	Y	Absent		OPHOS-4500(2),SO4-300(28),CL-300(28),NO3-353(2),TDS-2540(7)
L1822216-05I	Plastic 250ml unpreserved	A	7	7	3.9	Y	Absent		OPHOS-4500(2),SO4-300(28),CL-300(28),NO3-353(2),TDS-2540(7)
L1822216-05J	Plastic 250ml HNO3 preserved	A	<2	<2	3.9	Y	Absent		MN-6020S(180),FE-6020S(180),AS-6020S(180),AL-6020S(180)
L1822216-05K	Plastic 250ml HNO3 preserved	B	<2	<2	3.3	Y	Absent		FE-6020T(180),CA-6020T(180),K-6020T(180),NA-6020T(180),MN-6020T(180),AS-6020T(180),SO-UI(180),AL-6020T(180),MG-6020T(180)
L1822216-05L	Plastic 250ml Zn Acetate/NaOH preserved	A	>9	>9	3.9	Y	Absent		SULFIDE-4500(7)
L1822216-05M	Plastic 250ml Zn Acetate/NaOH preserved	A	>9	>9	3.9	Y	Absent		SULFIDE-4500(7)
L1822216-05N	Plastic 500ml unpreserved	B	7	7	3.3	Y	Absent		OPHOS-4500(2),SO4-300(28),CL-300(28),NO3-353(2),TDS-2540(7)
L1822216-05O	Plastic 500ml H2SO4 preserved	A	<2	<2	3.9	Y	Absent		TKN-351(28),TPHOS-4500(28),NH3-350(28)

**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP  
**Project Number:** 701970.01.SA

Serial\_No:06191821:16  
**Lab Number:** L1822216  
**Report Date:** 06/19/18

**Container Information**

<b>Container ID</b>	<b>Container Type</b>	<b>Cooler</b>	<b>Initial pH</b>	<b>Final pH</b>	<b>Temp deg C</b>	<b>Pres</b>	<b>Seal</b>	<b>Frozen Date/Time</b>	<b>Analysis(*)</b>
L1822216-05P	Plastic 500ml H2SO4 preserved	A	<2	<2	3.9	Y	Absent		TKN-351(28),TPHOS-4500(28),NH3-350(28)

**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP  
**Project Number:** 701970.01.SA

**Lab Number:** L1822216  
**Report Date:** 06/19/18

## GLOSSARY

### Acronyms

EDL	- Estimated Detection Limit: This value represents the level to which target analyte concentrations are reported as estimated values, when those target analyte concentrations are quantified below the reporting limit (RL). The EDL includes any adjustments from dilutions, concentrations or moisture content, where applicable. The use of EDLs is specific to the analysis of PAHs using Solid-Phase Microextraction (SPME).
EPA	- Environmental Protection Agency.
LCS	- Laboratory Control Sample: A sample matrix, free from the analytes of interest, spiked with verified known amounts of analytes or a material containing known and verified amounts of analytes.
LCSD	- Laboratory Control Sample Duplicate: Refer to LCS.
LFB	- Laboratory Fortified Blank: A sample matrix, free from the analytes of interest, spiked with verified known amounts of analytes or a material containing known and verified amounts of analytes.
MDL	- Method Detection Limit: This value represents the level to which target analyte concentrations are reported as estimated values, when those target analyte concentrations are quantified below the reporting limit (RL). The MDL includes any adjustments from dilutions, concentrations or moisture content, where applicable.
MS	- Matrix Spike Sample: A sample prepared by adding a known mass of target analyte to a specified amount of matrix sample for which an independent estimate of target analyte concentration is available.
MSD	- Matrix Spike Sample Duplicate: Refer to MS.
NA	- Not Applicable.
NC	- Not Calculated: Term is utilized when one or more of the results utilized in the calculation are non-detect at the parameter's reporting unit.
NDPA/DPA	- N-Nitrosodiphenylamine/Diphenylamine.
NI	- Not Ignitable.
NP	- Non-Plastic: Term is utilized for the analysis of Atterberg Limits in soil.
RL	- Reporting Limit: The value at which an instrument can accurately measure an analyte at a specific concentration. The RL includes any adjustments from dilutions, concentrations or moisture content, where applicable.
RPD	- Relative Percent Difference: The results from matrix and/or matrix spike duplicates are primarily designed to assess the precision of analytical results in a given matrix and are expressed as relative percent difference (RPD). Values which are less than five times the reporting limit for any individual parameter are evaluated by utilizing the absolute difference between the values; although the RPD value will be provided in the report.
SRM	- Standard Reference Material: A reference sample of a known or certified value that is of the same or similar matrix as the associated field samples.
STLP	- Semi-dynamic Tank Leaching Procedure per EPA Method 1315.
TIC	- Tentatively Identified Compound: A compound that has been identified to be present and is not part of the target compound list (TCL) for the method and/or program. All TICs are qualitatively identified and reported as estimated concentrations.

### Footnotes

- 1 - The reference for this analyte should be considered modified since this analyte is absent from the target analyte list of the original method.

### Terms

**Analytical Method:** Both the document from which the method originates and the analytical reference method. (Example: EPA 8260B is shown as 1,8260B.) The codes for the reference method documents are provided in the References section of the Addendum.

**Final pH:** As it pertains to Sample Receipt & Container Information section of the report, Final pH reflects pH of container determined after adjustment at the laboratory, if applicable. If no adjustment required, value reflects Initial pH.

**Frozen Date/Time:** With respect to Volatile Organics in soil, Frozen Date/Time reflects the date/time at which associated Reagent Water-preserved vials were initially frozen. Note: If frozen date/time is beyond 48 hours from sample collection, value will be reflected in 'bold'.

**Initial pH:** As it pertains to Sample Receipt & Container Information section of the report, Initial pH reflects pH of container determined upon receipt, if applicable.

**Total:** With respect to Organic analyses, a 'Total' result is defined as the summation of results for individual isomers or Aroclors. If a 'Total' result is requested, the results of its individual components will also be reported. This is applicable to 'Total' results for methods 8260, 8081 and 8082.

### Data Qualifiers

- A** - Spectra identified as "Aldol Condensation Product".
- B** - The analyte was detected above the reporting limit in the associated method blank. Flag only applies to associated field samples that have detectable concentrations of the analyte at less than ten times (10x) the concentration found in the blank. For MCP-related

**Report Format:** DU Report with 'J' Qualifiers



**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP  
**Project Number:** 701970.01.SA

**Lab Number:** L1822216  
**Report Date:** 06/19/18

#### Data Qualifiers

projects, flag only applies to associated field samples that have detectable concentrations of the analyte at less than ten times (10x) the concentration found in the blank. For DOD-related projects, flag only applies to associated field samples that have detectable concentrations of the analyte at less than ten times (10x) the concentration found in the blank AND the analyte was detected above one-half the reporting limit (or above the reporting limit for common lab contaminants) in the associated method blank. For NJ-Air-related projects, flag only applies to associated field samples that have detectable concentrations of the analyte above the reporting limit. For NJ-related projects (excluding Air), flag only applies to associated field samples that have detectable concentrations of the analyte, which was detected above the reporting limit in the associated method blank or above five times the reporting limit for common lab contaminants (Phthalates, Acetone, Methylene Chloride, 2-Butanone).

- C** - Co-elution: The target analyte co-elutes with a known lab standard (i.e. surrogate, internal standards, etc.) for co-extracted analyses.
- D** - Concentration of analyte was quantified from diluted analysis. Flag only applies to field samples that have detectable concentrations of the analyte.
- E** - Concentration of analyte exceeds the range of the calibration curve and/or linear range of the instrument.
- G** - The concentration may be biased high due to matrix interferences (i.e. co-elution) with non-target compound(s). The result should be considered estimated.
- H** - The analysis of pH was performed beyond the regulatory-required holding time of 15 minutes from the time of sample collection.
- I** - The lower value for the two columns has been reported due to obvious interference.
- M** - Reporting Limit (RL) exceeds the MCP CAM Reporting Limit for this analyte.
- NJ** - Presumptive evidence of compound. This represents an estimated concentration for Tentatively Identified Compounds (TICs), where the identification is based on a mass spectral library search.
- P** - The RPD between the results for the two columns exceeds the method-specified criteria.
- Q** - The quality control sample exceeds the associated acceptance criteria. For DOD-related projects, LCS and/or Continuing Calibration Standard exceedences are also qualified on all associated sample results. Note: This flag is not applicable for matrix spike recoveries when the sample concentration is greater than 4x the spike added or for batch duplicate RPD when the sample concentrations are less than 5x the RL. (Metals only.)
- R** - Analytical results are from sample re-analysis.
- RE** - Analytical results are from sample re-extraction.
- S** - Analytical results are from modified screening analysis.
- J** - Estimated value. The Target analyte concentration is below the quantitation limit (RL), but above the Method Detection Limit (MDL) or Estimated Detection Limit (EDL) for SPME-related analyses. This represents an estimated concentration for Tentatively Identified Compounds (TICs).
- ND** - Not detected at the method detection limit (MDL) for the sample, or estimated detection limit (EDL) for SPME-related analyses.

Report Format: DU Report with 'J' Qualifiers



**Project Name:** FORMER HAMPSHIRE CHEMICAL CORP  
**Project Number:** 701970.01.SA

**Lab Number:** L1822216  
**Report Date:** 06/19/18

## REFERENCES

- 1 Test Methods for Evaluating Solid Waste: Physical/Chemical Methods. EPA SW-846. Third Edition. Updates I - IV, 2007.
- 4 Methods for Chemical Analysis of Water and Wastes. EPA 600/4-79-020. Revised March 1983.
- 19 Inductively Coupled Plasma Atomic Emission Spectrometric Method for Trace Element Analysis of Water and Wastes. Appendix C, Part 136, 40 CFR (Code of Federal Regulations). July 1, 1999 edition.
- 44 Methods for the Determination of Inorganic Substances in Environmental Samples, EPA/600/R-93/100, August 1993.
- 121 Standard Methods for the Examination of Water and Wastewater. APHA-AWWA-WEF. Standard Methods Online.

## LIMITATION OF LIABILITIES

Alpha Analytical performs services with reasonable care and diligence normal to the analytical testing laboratory industry. In the event of an error, the sole and exclusive responsibility of Alpha Analytical shall be to re-perform the work at it's own expense. In no event shall Alpha Analytical be held liable for any incidental, consequential or special damages, including but not limited to, damages in any way connected with the use of, interpretation of, information or analysis provided by Alpha Analytical.

We strongly urge our clients to comply with EPA protocol regarding sample volume, preservation, cooling, containers, sampling procedures, holding time and splitting of samples in the field.



## Certification Information

The following analytes are not included in our Primary NELAP Scope of Accreditation:

### Westborough Facility

**EPA 624:** m/p-xylene, o-xylene

**EPA 8260C:** NPW: 1,2,4,5-Tetramethylbenzene; 4-Ethyltoluene, Azobenzene; SCM: Iodomethane (methyl iodide), Methyl methacrylate, 1,2,4,5-Tetramethylbenzene; 4-Ethyltoluene.

**EPA 8270D:** NPW: Dimethylnaphthalene, 1,4-Diphenylhydrazine; SCM: Dimethylnaphthalene, 1,4-Diphenylhydrazine.

**EPA 300:** DW: Bromide

**EPA 6860:** SCM: Perchlorate

**EPA 9010:** NPW and SCM: Amenable Cyanide Distillation

**SM4500:** NPW: Amenable Cyanide, Dissolved Oxygen; SCM: Total Phosphorus, TKN, NO<sub>2</sub>, NO<sub>3</sub>.

### Mansfield Facility

**SM 2540D:** TSS

**EPA 8082A:** NPW: PCB: 1, 5, 31, 87,101, 110, 141, 151, 153, 180, 183, 187.

**EPA TO-15:** Halothane, 2,4,4-Trimethyl-2-pentene, 2,4,4-Trimethyl-1-pentene, Thiophene, 2-Methylthiophene, 3-Methylthiophene, 2-Ethylthiophene, 1,2,3-Trimethylbenzene, Indan, Indene, 1,2,4,5-Tetramethylbenzene, Benzothiophene, 1-Methylnaphthalene.

**Biological Tissue Matrix:** EPA 3050B

The following analytes are included in our Massachusetts DEP Scope of Accreditation

### Westborough Facility:

#### Drinking Water

**EPA 300.0:** Chloride, Nitrate-N, Fluoride, Sulfate; **EPA 353.2:** Nitrate-N, Nitrite-N; **SM4500NO3-F:** Nitrate-N, Nitrite-N; **SM4500F-C, SM4500CN-CE,**

**EPA 180.1, SM2130B, SM4500CI-D, SM2320B, SM2540C, SM4500H-B**

**EPA 332:** Perchlorate; **EPA 524.2:** THMs and VOCs; **EPA 504.1:** EDB, DBCP.

**Microbiology:** **SM9215B; SM9223-P/A, SM9223B-Colilert-QT, SM9222D.**

#### Non-Potable Water

**SM4500H,B, EPA 120.1, SM2510B, SM2540C, SM2320B, SM4500CL-E, SM4500F-BC, SM4500NH3-BH:** Ammonia-N and Kjeldahl-N, **EPA 350.1:**

Ammonia-N, **LACHAT 10-107-06-1-B:** Ammonia-N, **EPA 351.1, SM4500NO3-F, EPA 353.2:** Nitrate-N, **EPA 351.1, SM4500P-E, SM4500P-B, E,**

**SM4500SO4-E, SM5220D, EPA 410.4, SM5210B, SM5310C, SM4500CL-D, EPA 1664, EPA 420.1, SM4500-CN-CE, SM2540D.**

**EPA 624:** Volatile Halocarbons & Aromatics,

**EPA 608:** Chlordane, Toxaphene, Aldrin, alpha-BHC, beta-BHC, gamma-BHC, delta-BHC, Dieldrin, DDD, DDE, DDT, Endosulfan I, Endosulfan II, Endosulfan sulfate, Endrin, Endrin Aldehyde, Heptachlor, Heptachlor Epoxide, PCBs

**EPA 625:** SVOC (Acid/Base/Neutral Extractables), **EPA 600/4-81-045:** PCB-Oil.

**Microbiology:** **SM9223B-Colilert-QT; Enterolert-QT, SM9221E, SM9222D.**

### Mansfield Facility:

#### Drinking Water

**EPA 200.7:** Al, Ba, Be, Cd, Cr, Cu, Mn, Ni, Na, Ag, Ca, Zn. **EPA 200.8:** Al, Sb, As, Ba, Be, Cd, Cr, Cu, Pb, Mn, Ni, Se, Ag, TL, Zn. **EPA 245.1 Hg.**

**EPA 522.**

#### Non-Potable Water

**EPA 200.7:** Al, Sb, As, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Mo, Ni, K, Se, Ag, Na, Sr, TL, Ti, V, Zn.

**EPA 200.8:** Al, Sb, As, Be, Cd, Cr, Cu, Pb, Mn, Ni, Se, Ag, TL, Zn.

**EPA 245.1 Hg.**

**SM2340B**

For a complete listing of analytes and methods, please contact your Alpha Project Manager.

PAGE 1 of 1