

ANALYTICAL REPORT

TestAmerica Laboratories, Inc.

TestAmerica Buffalo

10 Hazelwood Drive

Amherst, NY 14228-2298

Tel: (716)691-2600

TestAmerica Job ID: 480-10739-1

Client Project/Site: NYSDEC - Patton's Busy Bee:site #902014

For:

New York State D.E.C.

270 Michigan Avenue

Buffalo, New York 14203

Attn: Mr. Brian Sadowski



Authorized for release by:

10/19/2011 03:27:55 PM

Brian Fischer

Project Manager II

brian.fischer@testamericainc.com

LINKS

Review your project
results through

TotalAccess

Have a Question?



Visit us at:

www.testamericainc.com

Results relate only to the items tested and the sample(s) as received by the laboratory. The test results in this report meet all 2003 NELAC and 2009 TNI requirements for accredited parameters, exceptions are noted in this report. This report may not be reproduced except in full, and with written approval from the laboratory. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

This report has been electronically signed and authorized by the signatory. Electronic signature is intended to be the legally binding equivalent of a traditionally handwritten signature.

I certify that this data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed within the body of this report. Release of the data contained in this sample data package and in the electronic data deliverable has been authorized by the Laboratory Manager or his/her designee, as verified by the following signature.



Brian Fischer
Project Manager II
10/19/2011 03:27:55 PM

Table of Contents

Cover Page	1
Table of Contents	3
Definitions	4
Case Narrative	5
Client Sample Results	6
Chronicle	34
Certification Summary	38
Method Summary	39
Sample Summary	40
Chain of Custody	41
Receipt Checklists	43



Definitions/Glossary

Client: New York State D.E.C.

TestAmerica Job ID: 480-10739-1

Project/Site: NYSDEC - Patton's Busy Bee:site #902014

Qualifiers

GC/MS VOA

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Metals

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.
B	Compound was found in the blank and sample.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CNF	Contains no Free Liquid
DL, RA, RE, IN	Indicates a Dilution, Reanalysis, Re-extraction, or additional Initial metals/anion analysis of the sample
EDL	Estimated Detection Limit
EPA	United States Environmental Protection Agency
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
ND	Not detected at the reporting limit (or MDL or EDL if shown)
PQL	Practical Quantitation Limit
RL	Reporting Limit
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)

Case Narrative

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Job ID: 480-10739-1

Laboratory: TestAmerica Buffalo

Narrative

Job Narrative 480-10739-1

Comments

No additional comments.

Receipt

All samples were received in good condition within temperature requirements.

GC/MS VOA

Method(s) 8260B: The following sample was analyzed with headspace in the sample vial due to multiple reanalysis and limited sample volume: TRIP BLANK (480-10739-14).

No other analytical or quality issues were noted.

Metals

Method(s) 3005A: Due to the matrix, the initial volume(s) used for the following sample(s) deviated from the standard procedure: MW-107I (480-10739-3). The reporting limits (RLs) have been adjusted proportionately.

Method(s) 6010B: The Method Blank for batch 480-34291 contained total manganese above the method detection limit. This target analyte concentration was less than the reporting limit (RL); therefore, re-extraction and/or re-analysis of samples MW-101D (480-10739-5), MW-101I (480-10739-4), MW-102D (480-10739-10), MW-103D (480-10739-9), MW-103I (480-10739-8), MW-104D (480-10739-7), MW-104I (480-10739-6), MW-107I (480-10739-3), MW-107S (480-10739-2), MW-108D (480-10739-13), MW-108I (480-10739-12), MW-109 (480-10739-11), MW-113 (480-10739-1) was not performed.

Method(s) 6010B: The Matrix Spike / Matrix Spike Duplicate (MS/MSD) precision for batch 480-34291 was outside control limits for total zinc. Non-homogeneity of the sample matrix is suspected. The associated Laboratory Control Sample met acceptance criteria, therefore, no corrective action was necessary.

No other analytical or quality issues were noted.

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-113

Date Collected: 10/04/11 13:35

Date Received: 10/05/11 09:35

Lab Sample ID: 480-10739-1

Matrix: Water

Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/14/11 17:14	1
1,1,2,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/14/11 17:14	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/14/11 17:14	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/14/11 17:14	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/14/11 17:14	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/14/11 17:14	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/14/11 17:14	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/14/11 17:14	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/14/11 17:14	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/14/11 17:14	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/14/11 17:14	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/14/11 17:14	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/14/11 17:14	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/14/11 17:14	1
2-Hexanone	ND		5.0	1.2	ug/L			10/14/11 17:14	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/14/11 17:14	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/14/11 17:14	1
Acetone	ND		10	3.0	ug/L			10/14/11 17:14	1
Benzene	ND		1.0	0.41	ug/L			10/14/11 17:14	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/14/11 17:14	1
Bromoform	ND		1.0	0.26	ug/L			10/14/11 17:14	1
Bromomethane	ND		1.0	0.69	ug/L			10/14/11 17:14	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/14/11 17:14	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/14/11 17:14	1
Chlorobenzene	ND		1.0	0.75	ug/L			10/14/11 17:14	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/14/11 17:14	1
Chloroethane	ND		1.0	0.32	ug/L			10/14/11 17:14	1
Chloroform	ND		1.0	0.34	ug/L			10/14/11 17:14	1
Chloromethane	ND		1.0	0.35	ug/L			10/14/11 17:14	1
cis-1,2-Dichloroethene	ND		1.0	0.81	ug/L			10/14/11 17:14	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/14/11 17:14	1
Cyclohexane	ND		1.0	0.18	ug/L			10/14/11 17:14	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/14/11 17:14	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/14/11 17:14	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/14/11 17:14	1
Methyl acetate	ND		1.0	0.50	ug/L			10/14/11 17:14	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/14/11 17:14	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/14/11 17:14	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/14/11 17:14	1
Styrene	ND		1.0	0.73	ug/L			10/14/11 17:14	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/14/11 17:14	1
Toluene	ND		1.0	0.51	ug/L			10/14/11 17:14	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/14/11 17:14	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/14/11 17:14	1
Trichloroethene	ND		1.0	0.46	ug/L			10/14/11 17:14	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/14/11 17:14	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/14/11 17:14	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/14/11 17:14	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-113

Lab Sample ID: 480-10739-1

Date Collected: 10/04/11 13:35

Matrix: Water

Date Received: 10/05/11 09:35

Surrogate	% Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	126		66 - 137		10/14/11 17:14	1
Toluene-d8 (Surr)	115		71 - 126		10/14/11 17:14	1
4-Bromofluorobenzene (Surr)	108		73 - 120		10/14/11 17:14	1

Method: 6010B - Metals (ICP)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Aluminum	38.2		0.20	0.060	mg/L		10/07/11 08:30	10/07/11 16:09	1
Antimony	ND		0.020	0.0068	mg/L		10/07/11 08:30	10/07/11 16:09	1
Arsenic	0.045		0.010	0.0056	mg/L		10/07/11 08:30	10/07/11 16:09	1
Barium	0.29		0.0020	0.00050	mg/L		10/07/11 08:30	10/07/11 16:09	1
Beryllium	0.0017	J	0.0020	0.00030	mg/L		10/07/11 08:30	10/07/11 16:09	1
Cadmium	0.017		0.0010	0.00033	mg/L		10/07/11 08:30	10/07/11 16:09	1
Calcium	62.0		0.50	0.10	mg/L		10/07/11 08:30	10/07/11 16:09	1
Chromium	0.049		0.0040	0.00087	mg/L		10/07/11 08:30	10/07/11 16:09	1
Cobalt	0.031		0.0040	0.00063	mg/L		10/07/11 08:30	10/07/11 16:09	1
Copper	0.054		0.010	0.0015	mg/L		10/07/11 08:30	10/07/11 16:09	1
Iron	60.3		0.050	0.019	mg/L		10/07/11 08:30	10/07/11 16:09	1
Lead	0.041		0.0050	0.0030	mg/L		10/07/11 08:30	10/07/11 16:09	1
Magnesium	29.9		0.20	0.043	mg/L		10/07/11 08:30	10/07/11 16:09	1
Manganese	1.9	B	0.0030	0.00030	mg/L		10/07/11 08:30	10/07/11 16:09	1
Nickel	0.063		0.010	0.0013	mg/L		10/07/11 08:30	10/07/11 16:09	1
Potassium	15.2		0.50	0.20	mg/L		10/07/11 08:30	10/07/11 16:09	1
Selenium	ND		0.015	0.0087	mg/L		10/07/11 08:30	10/07/11 16:09	1
Silver	ND		0.0030	0.0017	mg/L		10/07/11 08:30	10/07/11 16:09	1
Sodium	4.4		1.0	0.32	mg/L		10/07/11 08:30	10/07/11 16:09	1
Thallium	ND		0.020	0.010	mg/L		10/07/11 08:30	10/07/11 16:09	1
Vanadium	0.049		0.0050	0.0011	mg/L		10/07/11 08:30	10/07/11 16:09	1
Zinc	0.13		0.010	0.0017	mg/L		10/07/11 08:30	10/07/11 16:09	1

Method: 7470A - Mercury (CVAA)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Mercury	ND		0.00020	0.00012	mg/L		10/06/11 11:40	10/06/11 16:49	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-107S

Date Collected: 10/04/11 13:25

Date Received: 10/05/11 09:35

Lab Sample ID: 480-10739-2

Matrix: Water

Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/14/11 17:36	1
1,1,2,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/14/11 17:36	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/14/11 17:36	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/14/11 17:36	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/14/11 17:36	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/14/11 17:36	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/14/11 17:36	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/14/11 17:36	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/14/11 17:36	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/14/11 17:36	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/14/11 17:36	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/14/11 17:36	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/14/11 17:36	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/14/11 17:36	1
2-Hexanone	ND		5.0	1.2	ug/L			10/14/11 17:36	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/14/11 17:36	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/14/11 17:36	1
Acetone	ND		10	3.0	ug/L			10/14/11 17:36	1
Benzene	ND		1.0	0.41	ug/L			10/14/11 17:36	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/14/11 17:36	1
Bromoform	ND		1.0	0.26	ug/L			10/14/11 17:36	1
Bromomethane	ND		1.0	0.69	ug/L			10/14/11 17:36	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/14/11 17:36	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/14/11 17:36	1
Chlorobenzene	ND		1.0	0.75	ug/L			10/14/11 17:36	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/14/11 17:36	1
Chloroethane	ND		1.0	0.32	ug/L			10/14/11 17:36	1
Chloroform	ND		1.0	0.34	ug/L			10/14/11 17:36	1
Chloromethane	ND		1.0	0.35	ug/L			10/14/11 17:36	1
cis-1,2-Dichloroethene	ND		1.0	0.81	ug/L			10/14/11 17:36	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/14/11 17:36	1
Cyclohexane	ND		1.0	0.18	ug/L			10/14/11 17:36	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/14/11 17:36	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/14/11 17:36	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/14/11 17:36	1
Methyl acetate	ND		1.0	0.50	ug/L			10/14/11 17:36	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/14/11 17:36	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/14/11 17:36	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/14/11 17:36	1
Styrene	ND		1.0	0.73	ug/L			10/14/11 17:36	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/14/11 17:36	1
Toluene	ND		1.0	0.51	ug/L			10/14/11 17:36	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/14/11 17:36	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/14/11 17:36	1
Trichloroethene	ND		1.0	0.46	ug/L			10/14/11 17:36	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/14/11 17:36	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/14/11 17:36	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/14/11 17:36	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-107S

Lab Sample ID: 480-10739-2

Date Collected: 10/04/11 13:25

Matrix: Water

Date Received: 10/05/11 09:35

Surrogate	% Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	125		66 - 137		10/14/11 17:36	1
Toluene-d8 (Surr)	114		71 - 126		10/14/11 17:36	1
4-Bromofluorobenzene (Surr)	107		73 - 120		10/14/11 17:36	1

Method: 6010B - Metals (ICP)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Aluminum	0.15	J	0.20	0.060	mg/L		10/07/11 08:30	10/07/11 16:11	1
Antimony	ND		0.020	0.0068	mg/L		10/07/11 08:30	10/07/11 16:11	1
Arsenic	ND		0.010	0.0056	mg/L		10/07/11 08:30	10/07/11 16:11	1
Barium	0.059		0.0020	0.00050	mg/L		10/07/11 08:30	10/07/11 16:11	1
Beryllium	ND		0.0020	0.00030	mg/L		10/07/11 08:30	10/07/11 16:11	1
Cadmium	0.00040	J	0.0010	0.00033	mg/L		10/07/11 08:30	10/07/11 16:11	1
Calcium	30.1		0.50	0.10	mg/L		10/07/11 08:30	10/07/11 16:11	1
Chromium	ND		0.0040	0.00087	mg/L		10/07/11 08:30	10/07/11 16:11	1
Cobalt	ND		0.0040	0.00063	mg/L		10/07/11 08:30	10/07/11 16:11	1
Copper	0.0026	J	0.010	0.0015	mg/L		10/07/11 08:30	10/07/11 16:11	1
Iron	0.13		0.050	0.019	mg/L		10/07/11 08:30	10/07/11 16:11	1
Lead	ND		0.0050	0.0030	mg/L		10/07/11 08:30	10/07/11 16:11	1
Magnesium	8.7		0.20	0.043	mg/L		10/07/11 08:30	10/07/11 16:11	1
Manganese	0.70	B	0.0030	0.00030	mg/L		10/07/11 08:30	10/07/11 16:11	1
Nickel	0.0053	J	0.010	0.0013	mg/L		10/07/11 08:30	10/07/11 16:11	1
Potassium	1.4		0.50	0.20	mg/L		10/07/11 08:30	10/07/11 16:11	1
Selenium	ND		0.015	0.0087	mg/L		10/07/11 08:30	10/07/11 16:11	1
Silver	ND		0.0030	0.0017	mg/L		10/07/11 08:30	10/07/11 16:11	1
Sodium	18.0		1.0	0.32	mg/L		10/07/11 08:30	10/07/11 16:11	1
Thallium	ND		0.020	0.010	mg/L		10/07/11 08:30	10/07/11 16:11	1
Vanadium	ND		0.0050	0.0011	mg/L		10/07/11 08:30	10/07/11 16:11	1
Zinc	ND		0.010	0.0017	mg/L		10/07/11 08:30	10/07/11 16:11	1

Method: 7470A - Mercury (CVAA)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Mercury	0.00030		0.00020	0.00012	mg/L		10/06/11 11:40	10/06/11 16:51	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-1071

Date Collected: 10/04/11 13:20

Date Received: 10/05/11 09:35

Lab Sample ID: 480-10739-3

Matrix: Water

Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/14/11 17:58	1
1,1,2,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/14/11 17:58	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/14/11 17:58	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/14/11 17:58	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/14/11 17:58	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/14/11 17:58	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/14/11 17:58	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/14/11 17:58	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/14/11 17:58	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/14/11 17:58	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/14/11 17:58	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/14/11 17:58	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/14/11 17:58	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/14/11 17:58	1
2-Hexanone	ND		5.0	1.2	ug/L			10/14/11 17:58	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/14/11 17:58	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/14/11 17:58	1
Acetone	ND		10	3.0	ug/L			10/14/11 17:58	1
Benzene	ND		1.0	0.41	ug/L			10/14/11 17:58	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/14/11 17:58	1
Bromoform	ND		1.0	0.26	ug/L			10/14/11 17:58	1
Bromomethane	ND		1.0	0.69	ug/L			10/14/11 17:58	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/14/11 17:58	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/14/11 17:58	1
Chlorobenzene	ND		1.0	0.75	ug/L			10/14/11 17:58	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/14/11 17:58	1
Chloroethane	ND		1.0	0.32	ug/L			10/14/11 17:58	1
Chloroform	ND		1.0	0.34	ug/L			10/14/11 17:58	1
Chloromethane	ND		1.0	0.35	ug/L			10/14/11 17:58	1
cis-1,2-Dichloroethene	ND		1.0	0.81	ug/L			10/14/11 17:58	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/14/11 17:58	1
Cyclohexane	ND		1.0	0.18	ug/L			10/14/11 17:58	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/14/11 17:58	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/14/11 17:58	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/14/11 17:58	1
Methyl acetate	ND		1.0	0.50	ug/L			10/14/11 17:58	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/14/11 17:58	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/14/11 17:58	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/14/11 17:58	1
Styrene	ND		1.0	0.73	ug/L			10/14/11 17:58	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/14/11 17:58	1
Toluene	ND		1.0	0.51	ug/L			10/14/11 17:58	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/14/11 17:58	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/14/11 17:58	1
Trichloroethene	ND		1.0	0.46	ug/L			10/14/11 17:58	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/14/11 17:58	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/14/11 17:58	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/14/11 17:58	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-1071

Lab Sample ID: 480-10739-3

Date Collected: 10/04/11 13:20

Matrix: Water

Date Received: 10/05/11 09:35

Surrogate	% Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	128		66 - 137		10/14/11 17:58	1
Toluene-d8 (Surr)	114		71 - 126		10/14/11 17:58	1
4-Bromofluorobenzene (Surr)	111		73 - 120		10/14/11 17:58	1

Method: 6010B - Metals (ICP)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Aluminum	41.7		0.25	0.075	mg/L		10/07/11 08:30	10/07/11 16:13	1
Antimony	ND		0.025	0.0085	mg/L		10/07/11 08:30	10/07/11 16:13	1
Arsenic	0.023		0.013	0.0069	mg/L		10/07/11 08:30	10/07/11 16:13	1
Barium	0.42		0.0025	0.00063	mg/L		10/07/11 08:30	10/07/11 16:13	1
Beryllium	0.0019	J	0.0025	0.00038	mg/L		10/07/11 08:30	10/07/11 16:13	1
Cadmium	0.0093		0.0013	0.00041	mg/L		10/07/11 08:30	10/07/11 16:13	1
Calcium	87.4		0.63	0.13	mg/L		10/07/11 08:30	10/07/11 16:13	1
Chromium	0.062		0.0050	0.0011	mg/L		10/07/11 08:30	10/07/11 16:13	1
Cobalt	0.033		0.0050	0.00079	mg/L		10/07/11 08:30	10/07/11 16:13	1
Copper	0.046		0.013	0.0019	mg/L		10/07/11 08:30	10/07/11 16:13	1
Iron	56.9		0.063	0.024	mg/L		10/07/11 08:30	10/07/11 16:13	1
Lead	0.077		0.0063	0.0038	mg/L		10/07/11 08:30	10/07/11 16:13	1
Magnesium	31.4		0.25	0.054	mg/L		10/07/11 08:30	10/07/11 16:13	1
Manganese	1.6	B	0.0038	0.00038	mg/L		10/07/11 08:30	10/07/11 16:13	1
Nickel	0.064		0.013	0.0016	mg/L		10/07/11 08:30	10/07/11 16:13	1
Potassium	15.9		0.63	0.25	mg/L		10/07/11 08:30	10/07/11 16:13	1
Selenium	ND		0.019	0.011	mg/L		10/07/11 08:30	10/07/11 16:13	1
Silver	ND		0.0038	0.0021	mg/L		10/07/11 08:30	10/07/11 16:13	1
Sodium	8.8		1.3	0.41	mg/L		10/07/11 08:30	10/07/11 16:13	1
Thallium	ND		0.025	0.013	mg/L		10/07/11 08:30	10/07/11 16:13	1
Vanadium	0.054		0.0063	0.0014	mg/L		10/07/11 08:30	10/07/11 16:13	1
Zinc	0.42		0.013	0.0021	mg/L		10/07/11 08:30	10/07/11 16:13	1

Method: 7470A - Mercury (CVAA)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Mercury	ND		0.00020	0.00012	mg/L		10/06/11 11:40	10/06/11 16:52	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-1011

Date Collected: 10/04/11 13:00

Date Received: 10/05/11 09:35

Lab Sample ID: 480-10739-4

Matrix: Water

Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/14/11 18:20	1
1,1,2,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/14/11 18:20	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/14/11 18:20	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/14/11 18:20	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/14/11 18:20	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/14/11 18:20	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/14/11 18:20	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/14/11 18:20	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/14/11 18:20	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/14/11 18:20	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/14/11 18:20	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/14/11 18:20	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/14/11 18:20	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/14/11 18:20	1
2-Hexanone	ND		5.0	1.2	ug/L			10/14/11 18:20	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/14/11 18:20	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/14/11 18:20	1
Acetone	ND		10	3.0	ug/L			10/14/11 18:20	1
Benzene	ND		1.0	0.41	ug/L			10/14/11 18:20	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/14/11 18:20	1
Bromoform	ND		1.0	0.26	ug/L			10/14/11 18:20	1
Bromomethane	ND		1.0	0.69	ug/L			10/14/11 18:20	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/14/11 18:20	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/14/11 18:20	1
Chlorobenzene	ND		1.0	0.75	ug/L			10/14/11 18:20	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/14/11 18:20	1
Chloroethane	ND		1.0	0.32	ug/L			10/14/11 18:20	1
Chloroform	ND		1.0	0.34	ug/L			10/14/11 18:20	1
Chloromethane	ND		1.0	0.35	ug/L			10/14/11 18:20	1
cis-1,2-Dichloroethene	ND		1.0	0.81	ug/L			10/14/11 18:20	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/14/11 18:20	1
Cyclohexane	ND		1.0	0.18	ug/L			10/14/11 18:20	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/14/11 18:20	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/14/11 18:20	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/14/11 18:20	1
Methyl acetate	ND		1.0	0.50	ug/L			10/14/11 18:20	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/14/11 18:20	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/14/11 18:20	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/14/11 18:20	1
Styrene	ND		1.0	0.73	ug/L			10/14/11 18:20	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/14/11 18:20	1
Toluene	ND		1.0	0.51	ug/L			10/14/11 18:20	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/14/11 18:20	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/14/11 18:20	1
Trichloroethene	ND		1.0	0.46	ug/L			10/14/11 18:20	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/14/11 18:20	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/14/11 18:20	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/14/11 18:20	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-1011

Lab Sample ID: 480-10739-4

Date Collected: 10/04/11 13:00

Matrix: Water

Date Received: 10/05/11 09:35

Surrogate	% Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	128		66 - 137		10/14/11 18:20	1
Toluene-d8 (Surr)	115		71 - 126		10/14/11 18:20	1
4-Bromofluorobenzene (Surr)	109		73 - 120		10/14/11 18:20	1

Method: 6010B - Metals (ICP)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Aluminum	90.9		0.20	0.060	mg/L		10/07/11 08:30	10/07/11 16:16	1
Antimony	ND		0.020	0.0068	mg/L		10/07/11 08:30	10/07/11 16:16	1
Arsenic	0.059		0.010	0.0056	mg/L		10/07/11 08:30	10/07/11 16:16	1
Barium	0.73		0.0020	0.00050	mg/L		10/07/11 08:30	10/07/11 16:16	1
Beryllium	0.0052		0.0020	0.00030	mg/L		10/07/11 08:30	10/07/11 16:16	1
Cadmium	0.0022		0.0010	0.00033	mg/L		10/07/11 08:30	10/07/11 16:16	1
Calcium	106		0.50	0.10	mg/L		10/07/11 08:30	10/07/11 16:16	1
Chromium	0.10		0.0040	0.00087	mg/L		10/07/11 08:30	10/07/11 16:16	1
Cobalt	0.071		0.0040	0.00063	mg/L		10/07/11 08:30	10/07/11 16:16	1
Copper	0.11		0.010	0.0015	mg/L		10/07/11 08:30	10/07/11 16:16	1
Iron	140		0.050	0.019	mg/L		10/07/11 08:30	10/07/11 16:16	1
Lead	0.046		0.0050	0.0030	mg/L		10/07/11 08:30	10/07/11 16:16	1
Magnesium	69.4		0.20	0.043	mg/L		10/07/11 08:30	10/07/11 16:16	1
Manganese	2.4	B	0.0030	0.00030	mg/L		10/07/11 08:30	10/07/11 16:16	1
Nickel	0.14		0.010	0.0013	mg/L		10/07/11 08:30	10/07/11 16:16	1
Potassium	32.6		0.50	0.20	mg/L		10/07/11 08:30	10/07/11 16:16	1
Selenium	ND		0.015	0.0087	mg/L		10/07/11 08:30	10/07/11 16:16	1
Silver	ND		0.0030	0.0017	mg/L		10/07/11 08:30	10/07/11 16:16	1
Sodium	7.5		1.0	0.32	mg/L		10/07/11 08:30	10/07/11 16:16	1
Thallium	ND		0.020	0.010	mg/L		10/07/11 08:30	10/07/11 16:16	1
Vanadium	0.12		0.0050	0.0011	mg/L		10/07/11 08:30	10/07/11 16:16	1
Zinc	0.77		0.010	0.0017	mg/L		10/07/11 08:30	10/07/11 16:16	1

Method: 7470A - Mercury (CVAA)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Mercury	ND		0.00020	0.00012	mg/L		10/06/11 11:40	10/06/11 16:54	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-101D

Date Collected: 10/04/11 13:05

Date Received: 10/05/11 09:35

Lab Sample ID: 480-10739-5

Matrix: Water

Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/14/11 18:42	1
1,1,2,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/14/11 18:42	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/14/11 18:42	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/14/11 18:42	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/14/11 18:42	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/14/11 18:42	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/14/11 18:42	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/14/11 18:42	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/14/11 18:42	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/14/11 18:42	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/14/11 18:42	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/14/11 18:42	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/14/11 18:42	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/14/11 18:42	1
2-Hexanone	ND		5.0	1.2	ug/L			10/14/11 18:42	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/14/11 18:42	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/14/11 18:42	1
Acetone	ND		10	3.0	ug/L			10/14/11 18:42	1
Benzene	0.83	J	1.0	0.41	ug/L			10/14/11 18:42	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/14/11 18:42	1
Bromoform	ND		1.0	0.26	ug/L			10/14/11 18:42	1
Bromomethane	ND		1.0	0.69	ug/L			10/14/11 18:42	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/14/11 18:42	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/14/11 18:42	1
Chlorobenzene	4.6		1.0	0.75	ug/L			10/14/11 18:42	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/14/11 18:42	1
Chloroethane	ND		1.0	0.32	ug/L			10/14/11 18:42	1
Chloroform	ND		1.0	0.34	ug/L			10/14/11 18:42	1
Chloromethane	ND		1.0	0.35	ug/L			10/14/11 18:42	1
cis-1,2-Dichloroethene	ND		1.0	0.81	ug/L			10/14/11 18:42	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/14/11 18:42	1
Cyclohexane	ND		1.0	0.18	ug/L			10/14/11 18:42	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/14/11 18:42	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/14/11 18:42	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/14/11 18:42	1
Methyl acetate	ND		1.0	0.50	ug/L			10/14/11 18:42	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/14/11 18:42	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/14/11 18:42	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/14/11 18:42	1
Styrene	ND		1.0	0.73	ug/L			10/14/11 18:42	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/14/11 18:42	1
Toluene	ND		1.0	0.51	ug/L			10/14/11 18:42	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/14/11 18:42	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/14/11 18:42	1
Trichloroethene	ND		1.0	0.46	ug/L			10/14/11 18:42	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/14/11 18:42	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/14/11 18:42	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/14/11 18:42	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-101D

Lab Sample ID: 480-10739-5

Date Collected: 10/04/11 13:05

Matrix: Water

Date Received: 10/05/11 09:35

Surrogate	% Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	134		66 - 137		10/14/11 18:42	1
Toluene-d8 (Surr)	116		71 - 126		10/14/11 18:42	1
4-Bromofluorobenzene (Surr)	112		73 - 120		10/14/11 18:42	1

Method: 6010B - Metals (ICP)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Aluminum	0.35		0.20	0.060	mg/L		10/07/11 08:30	10/07/11 16:22	1
Antimony	ND		0.020	0.0068	mg/L		10/07/11 08:30	10/07/11 16:22	1
Arsenic	0.0073	J	0.010	0.0056	mg/L		10/07/11 08:30	10/07/11 16:22	1
Barium	0.67		0.0020	0.00050	mg/L		10/07/11 08:30	10/07/11 16:22	1
Beryllium	ND		0.0020	0.00030	mg/L		10/07/11 08:30	10/07/11 16:22	1
Cadmium	0.0061		0.0010	0.00033	mg/L		10/07/11 08:30	10/07/11 16:22	1
Calcium	134		0.50	0.10	mg/L		10/07/11 08:30	10/07/11 16:22	1
Chromium	0.0030	J	0.0040	0.00087	mg/L		10/07/11 08:30	10/07/11 16:22	1
Cobalt	0.0046		0.0040	0.00063	mg/L		10/07/11 08:30	10/07/11 16:22	1
Copper	0.0093	J	0.010	0.0015	mg/L		10/07/11 08:30	10/07/11 16:22	1
Iron	1.4		0.050	0.019	mg/L		10/07/11 08:30	10/07/11 16:22	1
Lead	0.018		0.0050	0.0030	mg/L		10/07/11 08:30	10/07/11 16:22	1
Magnesium	84.2		0.20	0.043	mg/L		10/07/11 08:30	10/07/11 16:22	1
Manganese	5.5	B	0.0030	0.00030	mg/L		10/07/11 08:30	10/07/11 16:22	1
Nickel	0.012		0.010	0.0013	mg/L		10/07/11 08:30	10/07/11 16:22	1
Potassium	11.9		0.50	0.20	mg/L		10/07/11 08:30	10/07/11 16:22	1
Selenium	ND		0.015	0.0087	mg/L		10/07/11 08:30	10/07/11 16:22	1
Silver	ND		0.0030	0.0017	mg/L		10/07/11 08:30	10/07/11 16:22	1
Sodium	42.0		1.0	0.32	mg/L		10/07/11 08:30	10/07/11 16:22	1
Thallium	ND		0.020	0.010	mg/L		10/07/11 08:30	10/07/11 16:22	1
Vanadium	ND		0.0050	0.0011	mg/L		10/07/11 08:30	10/07/11 16:22	1
Zinc	0.017		0.010	0.0017	mg/L		10/07/11 08:30	10/07/11 16:22	1

Method: 7470A - Mercury (CVAA)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Mercury	ND		0.00020	0.00012	mg/L		10/06/11 11:40	10/06/11 16:56	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-104I

Date Collected: 10/04/11 12:20

Date Received: 10/05/11 09:35

Lab Sample ID: 480-10739-6

Matrix: Water

Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/14/11 19:04	1
1,1,2,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/14/11 19:04	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/14/11 19:04	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/14/11 19:04	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/14/11 19:04	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/14/11 19:04	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/14/11 19:04	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/14/11 19:04	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/14/11 19:04	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/14/11 19:04	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/14/11 19:04	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/14/11 19:04	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/14/11 19:04	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/14/11 19:04	1
2-Hexanone	ND		5.0	1.2	ug/L			10/14/11 19:04	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/14/11 19:04	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/14/11 19:04	1
Acetone	ND		10	3.0	ug/L			10/14/11 19:04	1
Benzene	ND		1.0	0.41	ug/L			10/14/11 19:04	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/14/11 19:04	1
Bromoform	ND		1.0	0.26	ug/L			10/14/11 19:04	1
Bromomethane	ND		1.0	0.69	ug/L			10/14/11 19:04	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/14/11 19:04	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/14/11 19:04	1
Chlorobenzene	ND		1.0	0.75	ug/L			10/14/11 19:04	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/14/11 19:04	1
Chloroethane	ND		1.0	0.32	ug/L			10/14/11 19:04	1
Chloroform	1.5		1.0	0.34	ug/L			10/14/11 19:04	1
Chloromethane	ND		1.0	0.35	ug/L			10/14/11 19:04	1
cis-1,2-Dichloroethene	1.5		1.0	0.81	ug/L			10/14/11 19:04	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/14/11 19:04	1
Cyclohexane	ND		1.0	0.18	ug/L			10/14/11 19:04	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/14/11 19:04	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/14/11 19:04	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/14/11 19:04	1
Methyl acetate	ND		1.0	0.50	ug/L			10/14/11 19:04	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/14/11 19:04	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/14/11 19:04	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/14/11 19:04	1
Styrene	ND		1.0	0.73	ug/L			10/14/11 19:04	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/14/11 19:04	1
Toluene	ND		1.0	0.51	ug/L			10/14/11 19:04	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/14/11 19:04	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/14/11 19:04	1
Trichloroethene	2.6		1.0	0.46	ug/L			10/14/11 19:04	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/14/11 19:04	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/14/11 19:04	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/14/11 19:04	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-104I

Lab Sample ID: 480-10739-6

Date Collected: 10/04/11 12:20

Matrix: Water

Date Received: 10/05/11 09:35

Surrogate	% Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	128		66 - 137		10/14/11 19:04	1
Toluene-d8 (Surr)	116		71 - 126		10/14/11 19:04	1
4-Bromofluorobenzene (Surr)	109		73 - 120		10/14/11 19:04	1

Method: 6010B - Metals (ICP)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Aluminum	0.72		0.20	0.060	mg/L		10/07/11 08:30	10/07/11 16:24	1
Antimony	ND		0.020	0.0068	mg/L		10/07/11 08:30	10/07/11 16:24	1
Arsenic	ND		0.010	0.0056	mg/L		10/07/11 08:30	10/07/11 16:24	1
Barium	0.028		0.0020	0.00050	mg/L		10/07/11 08:30	10/07/11 16:24	1
Beryllium	ND		0.0020	0.00030	mg/L		10/07/11 08:30	10/07/11 16:24	1
Cadmium	0.0017		0.0010	0.00033	mg/L		10/07/11 08:30	10/07/11 16:24	1
Calcium	6.8		0.50	0.10	mg/L		10/07/11 08:30	10/07/11 16:24	1
Chromium	0.0013	J	0.0040	0.00087	mg/L		10/07/11 08:30	10/07/11 16:24	1
Cobalt	ND		0.0040	0.00063	mg/L		10/07/11 08:30	10/07/11 16:24	1
Copper	0.0017	J	0.010	0.0015	mg/L		10/07/11 08:30	10/07/11 16:24	1
Iron	0.73		0.050	0.019	mg/L		10/07/11 08:30	10/07/11 16:24	1
Lead	ND		0.0050	0.0030	mg/L		10/07/11 08:30	10/07/11 16:24	1
Magnesium	2.2		0.20	0.043	mg/L		10/07/11 08:30	10/07/11 16:24	1
Manganese	0.014	B	0.0030	0.00030	mg/L		10/07/11 08:30	10/07/11 16:24	1
Nickel	0.0022	J	0.010	0.0013	mg/L		10/07/11 08:30	10/07/11 16:24	1
Potassium	1.2		0.50	0.20	mg/L		10/07/11 08:30	10/07/11 16:24	1
Selenium	ND		0.015	0.0087	mg/L		10/07/11 08:30	10/07/11 16:24	1
Silver	ND		0.0030	0.0017	mg/L		10/07/11 08:30	10/07/11 16:24	1
Sodium	5.4		1.0	0.32	mg/L		10/07/11 08:30	10/07/11 16:24	1
Thallium	ND		0.020	0.010	mg/L		10/07/11 08:30	10/07/11 16:24	1
Vanadium	0.0017	J	0.0050	0.0011	mg/L		10/07/11 08:30	10/07/11 16:24	1
Zinc	0.0029	J	0.010	0.0017	mg/L		10/07/11 08:30	10/07/11 16:24	1

Method: 7470A - Mercury (CVAA)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Mercury	ND		0.00020	0.00012	mg/L		10/06/11 11:40	10/06/11 17:00	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-104D

Date Collected: 10/04/11 12:15

Date Received: 10/05/11 09:35

Lab Sample ID: 480-10739-7

Matrix: Water

Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/14/11 19:26	1
1,1,2,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/14/11 19:26	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/14/11 19:26	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/14/11 19:26	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/14/11 19:26	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/14/11 19:26	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/14/11 19:26	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/14/11 19:26	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/14/11 19:26	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/14/11 19:26	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/14/11 19:26	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/14/11 19:26	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/14/11 19:26	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/14/11 19:26	1
2-Hexanone	ND		5.0	1.2	ug/L			10/14/11 19:26	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/14/11 19:26	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/14/11 19:26	1
Acetone	ND		10	3.0	ug/L			10/14/11 19:26	1
Benzene	ND		1.0	0.41	ug/L			10/14/11 19:26	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/14/11 19:26	1
Bromoform	ND		1.0	0.26	ug/L			10/14/11 19:26	1
Bromomethane	ND		1.0	0.69	ug/L			10/14/11 19:26	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/14/11 19:26	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/14/11 19:26	1
Chlorobenzene	ND		1.0	0.75	ug/L			10/14/11 19:26	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/14/11 19:26	1
Chloroethane	ND		1.0	0.32	ug/L			10/14/11 19:26	1
Chloroform	ND		1.0	0.34	ug/L			10/14/11 19:26	1
Chloromethane	ND		1.0	0.35	ug/L			10/14/11 19:26	1
cis-1,2-Dichloroethene	ND		1.0	0.81	ug/L			10/14/11 19:26	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/14/11 19:26	1
Cyclohexane	ND		1.0	0.18	ug/L			10/14/11 19:26	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/14/11 19:26	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/14/11 19:26	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/14/11 19:26	1
Methyl acetate	ND		1.0	0.50	ug/L			10/14/11 19:26	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/14/11 19:26	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/14/11 19:26	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/14/11 19:26	1
Styrene	ND		1.0	0.73	ug/L			10/14/11 19:26	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/14/11 19:26	1
Toluene	ND		1.0	0.51	ug/L			10/14/11 19:26	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/14/11 19:26	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/14/11 19:26	1
Trichloroethene	ND		1.0	0.46	ug/L			10/14/11 19:26	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/14/11 19:26	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/14/11 19:26	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/14/11 19:26	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-104D

Lab Sample ID: 480-10739-7

Date Collected: 10/04/11 12:15

Matrix: Water

Date Received: 10/05/11 09:35

Surrogate	% Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	128		66 - 137		10/14/11 19:26	1
Toluene-d8 (Surr)	117		71 - 126		10/14/11 19:26	1
4-Bromofluorobenzene (Surr)	109		73 - 120		10/14/11 19:26	1

Method: 6010B - Metals (ICP)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Aluminum	17.2		0.20	0.060	mg/L		10/07/11 08:30	10/07/11 16:27	1
Antimony	ND		0.020	0.0068	mg/L		10/07/11 08:30	10/07/11 16:27	1
Arsenic	0.042		0.010	0.0056	mg/L		10/07/11 08:30	10/07/11 16:27	1
Barium	0.17		0.0020	0.00050	mg/L		10/07/11 08:30	10/07/11 16:27	1
Beryllium	0.00073	J	0.0020	0.00030	mg/L		10/07/11 08:30	10/07/11 16:27	1
Cadmium	0.0083		0.0010	0.00033	mg/L		10/07/11 08:30	10/07/11 16:27	1
Calcium	48.2		0.50	0.10	mg/L		10/07/11 08:30	10/07/11 16:27	1
Chromium	0.019		0.0040	0.00087	mg/L		10/07/11 08:30	10/07/11 16:27	1
Cobalt	0.012		0.0040	0.00063	mg/L		10/07/11 08:30	10/07/11 16:27	1
Copper	0.024		0.010	0.0015	mg/L		10/07/11 08:30	10/07/11 16:27	1
Iron	22.2		0.050	0.019	mg/L		10/07/11 08:30	10/07/11 16:27	1
Lead	0.023		0.0050	0.0030	mg/L		10/07/11 08:30	10/07/11 16:27	1
Magnesium	29.4		0.20	0.043	mg/L		10/07/11 08:30	10/07/11 16:27	1
Manganese	0.44	B	0.0030	0.00030	mg/L		10/07/11 08:30	10/07/11 16:27	1
Nickel	0.022		0.010	0.0013	mg/L		10/07/11 08:30	10/07/11 16:27	1
Potassium	8.6		0.50	0.20	mg/L		10/07/11 08:30	10/07/11 16:27	1
Selenium	ND		0.015	0.0087	mg/L		10/07/11 08:30	10/07/11 16:27	1
Silver	ND		0.0030	0.0017	mg/L		10/07/11 08:30	10/07/11 16:27	1
Sodium	5.5		1.0	0.32	mg/L		10/07/11 08:30	10/07/11 16:27	1
Thallium	ND		0.020	0.010	mg/L		10/07/11 08:30	10/07/11 16:27	1
Vanadium	0.021		0.0050	0.0011	mg/L		10/07/11 08:30	10/07/11 16:27	1
Zinc	0.053		0.010	0.0017	mg/L		10/07/11 08:30	10/07/11 16:27	1

Method: 7470A - Mercury (CVAA)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Mercury	ND		0.00020	0.00012	mg/L		10/06/11 11:40	10/06/11 17:02	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-1031

Date Collected: 10/04/11 12:05

Date Received: 10/05/11 09:35

Lab Sample ID: 480-10739-8

Matrix: Water

Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/14/11 19:48	1
1,1,2,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/14/11 19:48	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/14/11 19:48	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/14/11 19:48	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/14/11 19:48	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/14/11 19:48	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/14/11 19:48	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/14/11 19:48	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/14/11 19:48	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/14/11 19:48	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/14/11 19:48	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/14/11 19:48	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/14/11 19:48	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/14/11 19:48	1
2-Hexanone	ND		5.0	1.2	ug/L			10/14/11 19:48	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/14/11 19:48	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/14/11 19:48	1
Acetone	ND		10	3.0	ug/L			10/14/11 19:48	1
Benzene	ND		1.0	0.41	ug/L			10/14/11 19:48	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/14/11 19:48	1
Bromoform	ND		1.0	0.26	ug/L			10/14/11 19:48	1
Bromomethane	ND		1.0	0.69	ug/L			10/14/11 19:48	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/14/11 19:48	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/14/11 19:48	1
Chlorobenzene	ND		1.0	0.75	ug/L			10/14/11 19:48	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/14/11 19:48	1
Chloroethane	ND		1.0	0.32	ug/L			10/14/11 19:48	1
Chloroform	ND		1.0	0.34	ug/L			10/14/11 19:48	1
Chloromethane	ND		1.0	0.35	ug/L			10/14/11 19:48	1
cis-1,2-Dichloroethene	4.6		1.0	0.81	ug/L			10/14/11 19:48	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/14/11 19:48	1
Cyclohexane	ND		1.0	0.18	ug/L			10/14/11 19:48	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/14/11 19:48	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/14/11 19:48	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/14/11 19:48	1
Methyl acetate	ND		1.0	0.50	ug/L			10/14/11 19:48	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/14/11 19:48	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/14/11 19:48	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/14/11 19:48	1
Styrene	ND		1.0	0.73	ug/L			10/14/11 19:48	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/14/11 19:48	1
Toluene	ND		1.0	0.51	ug/L			10/14/11 19:48	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/14/11 19:48	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/14/11 19:48	1
Trichloroethene	4.4		1.0	0.46	ug/L			10/14/11 19:48	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/14/11 19:48	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/14/11 19:48	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/14/11 19:48	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-103I

Lab Sample ID: 480-10739-8

Date Collected: 10/04/11 12:05

Matrix: Water

Date Received: 10/05/11 09:35

Surrogate	% Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	126		66 - 137		10/14/11 19:48	1
Toluene-d8 (Surr)	113		71 - 126		10/14/11 19:48	1
4-Bromofluorobenzene (Surr)	108		73 - 120		10/14/11 19:48	1

Method: 6010B - Metals (ICP)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Aluminum	0.93		0.20	0.060	mg/L		10/07/11 08:30	10/07/11 16:29	1
Antimony	ND		0.020	0.0068	mg/L		10/07/11 08:30	10/07/11 16:29	1
Arsenic	ND		0.010	0.0056	mg/L		10/07/11 08:30	10/07/11 16:29	1
Barium	0.044		0.0020	0.00050	mg/L		10/07/11 08:30	10/07/11 16:29	1
Beryllium	ND		0.0020	0.00030	mg/L		10/07/11 08:30	10/07/11 16:29	1
Cadmium	0.00096	J	0.0010	0.00033	mg/L		10/07/11 08:30	10/07/11 16:29	1
Calcium	13.1		0.50	0.10	mg/L		10/07/11 08:30	10/07/11 16:29	1
Chromium	0.0013	J	0.0040	0.00087	mg/L		10/07/11 08:30	10/07/11 16:29	1
Cobalt	ND		0.0040	0.00063	mg/L		10/07/11 08:30	10/07/11 16:29	1
Copper	0.0018	J	0.010	0.0015	mg/L		10/07/11 08:30	10/07/11 16:29	1
Iron	0.84		0.050	0.019	mg/L		10/07/11 08:30	10/07/11 16:29	1
Lead	ND		0.0050	0.0030	mg/L		10/07/11 08:30	10/07/11 16:29	1
Magnesium	4.5		0.20	0.043	mg/L		10/07/11 08:30	10/07/11 16:29	1
Manganese	0.028	B	0.0030	0.00030	mg/L		10/07/11 08:30	10/07/11 16:29	1
Nickel	ND		0.010	0.0013	mg/L		10/07/11 08:30	10/07/11 16:29	1
Potassium	1.7		0.50	0.20	mg/L		10/07/11 08:30	10/07/11 16:29	1
Selenium	ND		0.015	0.0087	mg/L		10/07/11 08:30	10/07/11 16:29	1
Silver	ND		0.0030	0.0017	mg/L		10/07/11 08:30	10/07/11 16:29	1
Sodium	2.8		1.0	0.32	mg/L		10/07/11 08:30	10/07/11 16:29	1
Thallium	ND		0.020	0.010	mg/L		10/07/11 08:30	10/07/11 16:29	1
Vanadium	ND		0.0050	0.0011	mg/L		10/07/11 08:30	10/07/11 16:29	1
Zinc	0.0031	J	0.010	0.0017	mg/L		10/07/11 08:30	10/07/11 16:29	1

Method: 7470A - Mercury (CVAA)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Mercury	ND		0.00020	0.00012	mg/L		10/06/11 11:40	10/06/11 17:04	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-103D

Date Collected: 10/04/11 12:00

Date Received: 10/05/11 09:35

Lab Sample ID: 480-10739-9

Matrix: Water

Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/14/11 20:10	1
1,1,2,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/14/11 20:10	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/14/11 20:10	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/14/11 20:10	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/14/11 20:10	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/14/11 20:10	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/14/11 20:10	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/14/11 20:10	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/14/11 20:10	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/14/11 20:10	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/14/11 20:10	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/14/11 20:10	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/14/11 20:10	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/14/11 20:10	1
2-Hexanone	ND		5.0	1.2	ug/L			10/14/11 20:10	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/14/11 20:10	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/14/11 20:10	1
Acetone	ND		10	3.0	ug/L			10/14/11 20:10	1
Benzene	ND		1.0	0.41	ug/L			10/14/11 20:10	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/14/11 20:10	1
Bromoform	ND		1.0	0.26	ug/L			10/14/11 20:10	1
Bromomethane	ND		1.0	0.69	ug/L			10/14/11 20:10	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/14/11 20:10	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/14/11 20:10	1
Chlorobenzene	ND		1.0	0.75	ug/L			10/14/11 20:10	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/14/11 20:10	1
Chloroethane	ND		1.0	0.32	ug/L			10/14/11 20:10	1
Chloroform	ND		1.0	0.34	ug/L			10/14/11 20:10	1
Chloromethane	ND		1.0	0.35	ug/L			10/14/11 20:10	1
cis-1,2-Dichloroethene	3.3		1.0	0.81	ug/L			10/14/11 20:10	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/14/11 20:10	1
Cyclohexane	ND		1.0	0.18	ug/L			10/14/11 20:10	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/14/11 20:10	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/14/11 20:10	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/14/11 20:10	1
Methyl acetate	ND		1.0	0.50	ug/L			10/14/11 20:10	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/14/11 20:10	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/14/11 20:10	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/14/11 20:10	1
Styrene	ND		1.0	0.73	ug/L			10/14/11 20:10	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/14/11 20:10	1
Toluene	ND		1.0	0.51	ug/L			10/14/11 20:10	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/14/11 20:10	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/14/11 20:10	1
Trichloroethene	2.5		1.0	0.46	ug/L			10/14/11 20:10	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/14/11 20:10	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/14/11 20:10	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/14/11 20:10	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-103D

Lab Sample ID: 480-10739-9

Date Collected: 10/04/11 12:00

Matrix: Water

Date Received: 10/05/11 09:35

Surrogate	% Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	128		66 - 137		10/14/11 20:10	1
Toluene-d8 (Surr)	118		71 - 126		10/14/11 20:10	1
4-Bromofluorobenzene (Surr)	108		73 - 120		10/14/11 20:10	1

Method: 6010B - Metals (ICP)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Aluminum	0.18	J	0.20	0.060	mg/L		10/07/11 08:30	10/07/11 16:31	1
Antimony	ND		0.020	0.0068	mg/L		10/07/11 08:30	10/07/11 16:31	1
Arsenic	ND		0.010	0.0056	mg/L		10/07/11 08:30	10/07/11 16:31	1
Barium	0.044		0.0020	0.00050	mg/L		10/07/11 08:30	10/07/11 16:31	1
Beryllium	ND		0.0020	0.00030	mg/L		10/07/11 08:30	10/07/11 16:31	1
Cadmium	0.00069	J	0.0010	0.00033	mg/L		10/07/11 08:30	10/07/11 16:31	1
Calcium	16.3		0.50	0.10	mg/L		10/07/11 08:30	10/07/11 16:31	1
Chromium	ND		0.0040	0.00087	mg/L		10/07/11 08:30	10/07/11 16:31	1
Cobalt	ND		0.0040	0.00063	mg/L		10/07/11 08:30	10/07/11 16:31	1
Copper	ND		0.010	0.0015	mg/L		10/07/11 08:30	10/07/11 16:31	1
Iron	0.14		0.050	0.019	mg/L		10/07/11 08:30	10/07/11 16:31	1
Lead	ND		0.0050	0.0030	mg/L		10/07/11 08:30	10/07/11 16:31	1
Magnesium	6.2		0.20	0.043	mg/L		10/07/11 08:30	10/07/11 16:31	1
Manganese	0.0026	J B	0.0030	0.00030	mg/L		10/07/11 08:30	10/07/11 16:31	1
Nickel	ND		0.010	0.0013	mg/L		10/07/11 08:30	10/07/11 16:31	1
Potassium	1.5		0.50	0.20	mg/L		10/07/11 08:30	10/07/11 16:31	1
Selenium	ND		0.015	0.0087	mg/L		10/07/11 08:30	10/07/11 16:31	1
Silver	ND		0.0030	0.0017	mg/L		10/07/11 08:30	10/07/11 16:31	1
Sodium	3.0		1.0	0.32	mg/L		10/07/11 08:30	10/07/11 16:31	1
Thallium	ND		0.020	0.010	mg/L		10/07/11 08:30	10/07/11 16:31	1
Vanadium	ND		0.0050	0.0011	mg/L		10/07/11 08:30	10/07/11 16:31	1
Zinc	0.0021	J	0.010	0.0017	mg/L		10/07/11 08:30	10/07/11 16:31	1

Method: 7470A - Mercury (CVAA)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Mercury	ND		0.00020	0.00012	mg/L		10/06/11 11:40	10/06/11 17:06	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-102D

Lab Sample ID: 480-10739-10

Date Collected: 10/04/11 11:00

Matrix: Water

Date Received: 10/05/11 09:35

Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/14/11 20:32	1
1,1,2,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/14/11 20:32	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/14/11 20:32	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/14/11 20:32	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/14/11 20:32	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/14/11 20:32	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/14/11 20:32	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/14/11 20:32	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/14/11 20:32	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/14/11 20:32	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/14/11 20:32	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/14/11 20:32	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/14/11 20:32	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/14/11 20:32	1
2-Hexanone	ND		5.0	1.2	ug/L			10/14/11 20:32	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/14/11 20:32	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/14/11 20:32	1
Acetone	ND		10	3.0	ug/L			10/14/11 20:32	1
Benzene	ND		1.0	0.41	ug/L			10/14/11 20:32	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/14/11 20:32	1
Bromoform	ND		1.0	0.26	ug/L			10/14/11 20:32	1
Bromomethane	ND		1.0	0.69	ug/L			10/14/11 20:32	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/14/11 20:32	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/14/11 20:32	1
Chlorobenzene	ND		1.0	0.75	ug/L			10/14/11 20:32	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/14/11 20:32	1
Chloroethane	ND		1.0	0.32	ug/L			10/14/11 20:32	1
Chloroform	ND		1.0	0.34	ug/L			10/14/11 20:32	1
Chloromethane	ND		1.0	0.35	ug/L			10/14/11 20:32	1
cis-1,2-Dichloroethene	ND		1.0	0.81	ug/L			10/14/11 20:32	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/14/11 20:32	1
Cyclohexane	ND		1.0	0.18	ug/L			10/14/11 20:32	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/14/11 20:32	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/14/11 20:32	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/14/11 20:32	1
Methyl acetate	ND		1.0	0.50	ug/L			10/14/11 20:32	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/14/11 20:32	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/14/11 20:32	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/14/11 20:32	1
Styrene	ND		1.0	0.73	ug/L			10/14/11 20:32	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/14/11 20:32	1
Toluene	ND		1.0	0.51	ug/L			10/14/11 20:32	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/14/11 20:32	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/14/11 20:32	1
Trichloroethene	ND		1.0	0.46	ug/L			10/14/11 20:32	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/14/11 20:32	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/14/11 20:32	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/14/11 20:32	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-102D

Lab Sample ID: 480-10739-10

Date Collected: 10/04/11 11:00

Matrix: Water

Date Received: 10/05/11 09:35

Surrogate	% Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	128		66 - 137		10/14/11 20:32	1
Toluene-d8 (Surr)	116		71 - 126		10/14/11 20:32	1
4-Bromofluorobenzene (Surr)	110		73 - 120		10/14/11 20:32	1

Method: 6010B - Metals (ICP)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Aluminum	ND		0.20	0.060	mg/L		10/07/11 08:30	10/07/11 16:33	1
Antimony	ND		0.020	0.0068	mg/L		10/07/11 08:30	10/07/11 16:33	1
Arsenic	ND		0.010	0.0056	mg/L		10/07/11 08:30	10/07/11 16:33	1
Barium	0.087		0.0020	0.00050	mg/L		10/07/11 08:30	10/07/11 16:33	1
Beryllium	ND		0.0020	0.00030	mg/L		10/07/11 08:30	10/07/11 16:33	1
Cadmium	0.012		0.0010	0.00033	mg/L		10/07/11 08:30	10/07/11 16:33	1
Calcium	46.4		0.50	0.10	mg/L		10/07/11 08:30	10/07/11 16:33	1
Chromium	0.00094	J	0.0040	0.00087	mg/L		10/07/11 08:30	10/07/11 16:33	1
Cobalt	ND		0.0040	0.00063	mg/L		10/07/11 08:30	10/07/11 16:33	1
Copper	0.0061	J	0.010	0.0015	mg/L		10/07/11 08:30	10/07/11 16:33	1
Iron	0.13		0.050	0.019	mg/L		10/07/11 08:30	10/07/11 16:33	1
Lead	ND		0.0050	0.0030	mg/L		10/07/11 08:30	10/07/11 16:33	1
Magnesium	15.8		0.20	0.043	mg/L		10/07/11 08:30	10/07/11 16:33	1
Manganese	0.0035	B	0.0030	0.00030	mg/L		10/07/11 08:30	10/07/11 16:33	1
Nickel	0.0018	J	0.010	0.0013	mg/L		10/07/11 08:30	10/07/11 16:33	1
Potassium	2.3		0.50	0.20	mg/L		10/07/11 08:30	10/07/11 16:33	1
Selenium	ND		0.015	0.0087	mg/L		10/07/11 08:30	10/07/11 16:33	1
Silver	ND		0.0030	0.0017	mg/L		10/07/11 08:30	10/07/11 16:33	1
Sodium	3.8		1.0	0.32	mg/L		10/07/11 08:30	10/07/11 16:33	1
Thallium	ND		0.020	0.010	mg/L		10/07/11 08:30	10/07/11 16:33	1
Vanadium	ND		0.0050	0.0011	mg/L		10/07/11 08:30	10/07/11 16:33	1
Zinc	0.0027	J	0.010	0.0017	mg/L		10/07/11 08:30	10/07/11 16:33	1

Method: 7470A - Mercury (CVAA)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Mercury	ND		0.00020	0.00012	mg/L		10/06/11 11:40	10/06/11 17:07	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-109

Lab Sample ID: 480-10739-11

Date Collected: 10/04/11 11:15

Matrix: Water

Date Received: 10/05/11 09:35

Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/14/11 20:54	1
1,1,2,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/14/11 20:54	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/14/11 20:54	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/14/11 20:54	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/14/11 20:54	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/14/11 20:54	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/14/11 20:54	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/14/11 20:54	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/14/11 20:54	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/14/11 20:54	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/14/11 20:54	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/14/11 20:54	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/14/11 20:54	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/14/11 20:54	1
2-Hexanone	ND		5.0	1.2	ug/L			10/14/11 20:54	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/14/11 20:54	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/14/11 20:54	1
Acetone	ND		10	3.0	ug/L			10/14/11 20:54	1
Benzene	ND		1.0	0.41	ug/L			10/14/11 20:54	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/14/11 20:54	1
Bromoform	ND		1.0	0.26	ug/L			10/14/11 20:54	1
Bromomethane	ND		1.0	0.69	ug/L			10/14/11 20:54	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/14/11 20:54	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/14/11 20:54	1
Chlorobenzene	ND		1.0	0.75	ug/L			10/14/11 20:54	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/14/11 20:54	1
Chloroethane	ND		1.0	0.32	ug/L			10/14/11 20:54	1
Chloroform	ND		1.0	0.34	ug/L			10/14/11 20:54	1
Chloromethane	ND		1.0	0.35	ug/L			10/14/11 20:54	1
cis-1,2-Dichloroethene	ND		1.0	0.81	ug/L			10/14/11 20:54	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/14/11 20:54	1
Cyclohexane	ND		1.0	0.18	ug/L			10/14/11 20:54	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/14/11 20:54	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/14/11 20:54	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/14/11 20:54	1
Methyl acetate	ND		1.0	0.50	ug/L			10/14/11 20:54	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/14/11 20:54	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/14/11 20:54	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/14/11 20:54	1
Styrene	ND		1.0	0.73	ug/L			10/14/11 20:54	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/14/11 20:54	1
Toluene	ND		1.0	0.51	ug/L			10/14/11 20:54	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/14/11 20:54	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/14/11 20:54	1
Trichloroethene	ND		1.0	0.46	ug/L			10/14/11 20:54	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/14/11 20:54	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/14/11 20:54	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/14/11 20:54	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-109

Lab Sample ID: 480-10739-11

Date Collected: 10/04/11 11:15

Matrix: Water

Date Received: 10/05/11 09:35

Surrogate	% Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	129		66 - 137		10/14/11 20:54	1
Toluene-d8 (Surr)	114		71 - 126		10/14/11 20:54	1
4-Bromofluorobenzene (Surr)	109		73 - 120		10/14/11 20:54	1

Method: 6010B - Metals (ICP)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Aluminum	1.9		0.20	0.060	mg/L		10/07/11 08:30	10/07/11 16:36	1
Antimony	ND		0.020	0.0068	mg/L		10/07/11 08:30	10/07/11 16:36	1
Arsenic	ND		0.010	0.0056	mg/L		10/07/11 08:30	10/07/11 16:36	1
Barium	0.22		0.0020	0.00050	mg/L		10/07/11 08:30	10/07/11 16:36	1
Beryllium	ND		0.0020	0.00030	mg/L		10/07/11 08:30	10/07/11 16:36	1
Cadmium	0.013		0.0010	0.00033	mg/L		10/07/11 08:30	10/07/11 16:36	1
Calcium	36.2		0.50	0.10	mg/L		10/07/11 08:30	10/07/11 16:36	1
Chromium	0.0039	J	0.0040	0.00087	mg/L		10/07/11 08:30	10/07/11 16:36	1
Cobalt	0.0015	J	0.0040	0.00063	mg/L		10/07/11 08:30	10/07/11 16:36	1
Copper	0.012		0.010	0.0015	mg/L		10/07/11 08:30	10/07/11 16:36	1
Iron	4.5		0.050	0.019	mg/L		10/07/11 08:30	10/07/11 16:36	1
Lead	ND		0.0050	0.0030	mg/L		10/07/11 08:30	10/07/11 16:36	1
Magnesium	11.9		0.20	0.043	mg/L		10/07/11 08:30	10/07/11 16:36	1
Manganese	0.19	B	0.0030	0.00030	mg/L		10/07/11 08:30	10/07/11 16:36	1
Nickel	0.0052	J	0.010	0.0013	mg/L		10/07/11 08:30	10/07/11 16:36	1
Potassium	3.4		0.50	0.20	mg/L		10/07/11 08:30	10/07/11 16:36	1
Selenium	ND		0.015	0.0087	mg/L		10/07/11 08:30	10/07/11 16:36	1
Silver	ND		0.0030	0.0017	mg/L		10/07/11 08:30	10/07/11 16:36	1
Sodium	3.8		1.0	0.32	mg/L		10/07/11 08:30	10/07/11 16:36	1
Thallium	ND		0.020	0.010	mg/L		10/07/11 08:30	10/07/11 16:36	1
Vanadium	0.0029	J	0.0050	0.0011	mg/L		10/07/11 08:30	10/07/11 16:36	1
Zinc	0.010		0.010	0.0017	mg/L		10/07/11 08:30	10/07/11 16:36	1

Method: 7470A - Mercury (CVAA)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Mercury	ND		0.00020	0.00012	mg/L		10/06/11 11:40	10/06/11 17:09	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-108I

Lab Sample ID: 480-10739-12

Date Collected: 10/04/11 11:40

Matrix: Water

Date Received: 10/05/11 09:35

Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/14/11 21:16	1
1,1,2,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/14/11 21:16	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/14/11 21:16	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/14/11 21:16	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/14/11 21:16	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/14/11 21:16	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/14/11 21:16	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/14/11 21:16	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/14/11 21:16	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/14/11 21:16	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/14/11 21:16	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/14/11 21:16	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/14/11 21:16	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/14/11 21:16	1
2-Hexanone	ND		5.0	1.2	ug/L			10/14/11 21:16	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/14/11 21:16	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/14/11 21:16	1
Acetone	ND		10	3.0	ug/L			10/14/11 21:16	1
Benzene	ND		1.0	0.41	ug/L			10/14/11 21:16	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/14/11 21:16	1
Bromoform	ND		1.0	0.26	ug/L			10/14/11 21:16	1
Bromomethane	ND		1.0	0.69	ug/L			10/14/11 21:16	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/14/11 21:16	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/14/11 21:16	1
Chlorobenzene	ND		1.0	0.75	ug/L			10/14/11 21:16	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/14/11 21:16	1
Chloroethane	ND		1.0	0.32	ug/L			10/14/11 21:16	1
Chloroform	ND		1.0	0.34	ug/L			10/14/11 21:16	1
Chloromethane	ND		1.0	0.35	ug/L			10/14/11 21:16	1
cis-1,2-Dichloroethene	ND		1.0	0.81	ug/L			10/14/11 21:16	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/14/11 21:16	1
Cyclohexane	ND		1.0	0.18	ug/L			10/14/11 21:16	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/14/11 21:16	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/14/11 21:16	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/14/11 21:16	1
Methyl acetate	ND		1.0	0.50	ug/L			10/14/11 21:16	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/14/11 21:16	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/14/11 21:16	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/14/11 21:16	1
Styrene	ND		1.0	0.73	ug/L			10/14/11 21:16	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/14/11 21:16	1
Toluene	ND		1.0	0.51	ug/L			10/14/11 21:16	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/14/11 21:16	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/14/11 21:16	1
Trichloroethene	ND		1.0	0.46	ug/L			10/14/11 21:16	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/14/11 21:16	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/14/11 21:16	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/14/11 21:16	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-108I

Lab Sample ID: 480-10739-12

Date Collected: 10/04/11 11:40

Matrix: Water

Date Received: 10/05/11 09:35

Surrogate	% Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	130		66 - 137		10/14/11 21:16	1
Toluene-d8 (Surr)	117		71 - 126		10/14/11 21:16	1
4-Bromofluorobenzene (Surr)	107		73 - 120		10/14/11 21:16	1

Method: 6010B - Metals (ICP)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Aluminum	0.45		0.20	0.060	mg/L		10/07/11 08:30	10/07/11 16:38	1
Antimony	ND		0.020	0.0068	mg/L		10/07/11 08:30	10/07/11 16:38	1
Arsenic	ND		0.010	0.0056	mg/L		10/07/11 08:30	10/07/11 16:38	1
Barium	0.075		0.0020	0.00050	mg/L		10/07/11 08:30	10/07/11 16:38	1
Beryllium	ND		0.0020	0.00030	mg/L		10/07/11 08:30	10/07/11 16:38	1
Cadmium	ND		0.0010	0.00033	mg/L		10/07/11 08:30	10/07/11 16:38	1
Calcium	48.9		0.50	0.10	mg/L		10/07/11 08:30	10/07/11 16:38	1
Chromium	0.0012	J	0.0040	0.00087	mg/L		10/07/11 08:30	10/07/11 16:38	1
Cobalt	ND		0.0040	0.00063	mg/L		10/07/11 08:30	10/07/11 16:38	1
Copper	0.0041	J	0.010	0.0015	mg/L		10/07/11 08:30	10/07/11 16:38	1
Iron	0.49		0.050	0.019	mg/L		10/07/11 08:30	10/07/11 16:38	1
Lead	ND		0.0050	0.0030	mg/L		10/07/11 08:30	10/07/11 16:38	1
Magnesium	29.5		0.20	0.043	mg/L		10/07/11 08:30	10/07/11 16:38	1
Manganese	0.024	B	0.0030	0.00030	mg/L		10/07/11 08:30	10/07/11 16:38	1
Nickel	ND		0.010	0.0013	mg/L		10/07/11 08:30	10/07/11 16:38	1
Potassium	4.0		0.50	0.20	mg/L		10/07/11 08:30	10/07/11 16:38	1
Selenium	ND		0.015	0.0087	mg/L		10/07/11 08:30	10/07/11 16:38	1
Silver	ND		0.0030	0.0017	mg/L		10/07/11 08:30	10/07/11 16:38	1
Sodium	4.1		1.0	0.32	mg/L		10/07/11 08:30	10/07/11 16:38	1
Thallium	ND		0.020	0.010	mg/L		10/07/11 08:30	10/07/11 16:38	1
Vanadium	0.0013	J	0.0050	0.0011	mg/L		10/07/11 08:30	10/07/11 16:38	1
Zinc	0.0026	J	0.010	0.0017	mg/L		10/07/11 08:30	10/07/11 16:38	1

Method: 7470A - Mercury (CVAA)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Mercury	ND		0.00020	0.00012	mg/L		10/06/11 11:40	10/06/11 17:11	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-108D

Lab Sample ID: 480-10739-13

Date Collected: 10/04/11 11:30

Matrix: Water

Date Received: 10/05/11 09:35

Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/16/11 18:49	1
1,1,2,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/16/11 18:49	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/16/11 18:49	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/16/11 18:49	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/16/11 18:49	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/16/11 18:49	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/16/11 18:49	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/16/11 18:49	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/16/11 18:49	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/16/11 18:49	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/16/11 18:49	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/16/11 18:49	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/16/11 18:49	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/16/11 18:49	1
2-Hexanone	ND		5.0	1.2	ug/L			10/16/11 18:49	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/16/11 18:49	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/16/11 18:49	1
Acetone	ND		10	3.0	ug/L			10/16/11 18:49	1
Benzene	ND		1.0	0.41	ug/L			10/16/11 18:49	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/16/11 18:49	1
Bromoform	ND		1.0	0.26	ug/L			10/16/11 18:49	1
Bromomethane	ND		1.0	0.69	ug/L			10/16/11 18:49	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/16/11 18:49	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/16/11 18:49	1
Chlorobenzene	ND		1.0	0.75	ug/L			10/16/11 18:49	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/16/11 18:49	1
Chloroethane	ND		1.0	0.32	ug/L			10/16/11 18:49	1
Chloroform	ND		1.0	0.34	ug/L			10/16/11 18:49	1
Chloromethane	ND		1.0	0.35	ug/L			10/16/11 18:49	1
cis-1,2-Dichloroethene	ND		1.0	0.81	ug/L			10/16/11 18:49	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/16/11 18:49	1
Cyclohexane	ND		1.0	0.18	ug/L			10/16/11 18:49	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/16/11 18:49	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/16/11 18:49	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/16/11 18:49	1
Methyl acetate	ND		1.0	0.50	ug/L			10/16/11 18:49	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/16/11 18:49	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/16/11 18:49	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/16/11 18:49	1
Styrene	ND		1.0	0.73	ug/L			10/16/11 18:49	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/16/11 18:49	1
Toluene	ND		1.0	0.51	ug/L			10/16/11 18:49	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/16/11 18:49	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/16/11 18:49	1
Trichloroethene	ND		1.0	0.46	ug/L			10/16/11 18:49	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/16/11 18:49	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/16/11 18:49	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/16/11 18:49	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-108D

Lab Sample ID: 480-10739-13

Date Collected: 10/04/11 11:30

Matrix: Water

Date Received: 10/05/11 09:35

Surrogate	% Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	134		66 - 137		10/16/11 18:49	1
Toluene-d8 (Surr)	120		71 - 126		10/16/11 18:49	1
4-Bromofluorobenzene (Surr)	111		73 - 120		10/16/11 18:49	1

Method: 6010B - Metals (ICP)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Aluminum	0.51		0.20	0.060	mg/L		10/07/11 08:30	10/07/11 16:40	1
Antimony	ND		0.020	0.0068	mg/L		10/07/11 08:30	10/07/11 16:40	1
Arsenic	ND		0.010	0.0056	mg/L		10/07/11 08:30	10/07/11 16:40	1
Barium	0.13		0.0020	0.00050	mg/L		10/07/11 08:30	10/07/11 16:40	1
Beryllium	ND		0.0020	0.00030	mg/L		10/07/11 08:30	10/07/11 16:40	1
Cadmium	0.015		0.0010	0.00033	mg/L		10/07/11 08:30	10/07/11 16:40	1
Calcium	29.9		0.50	0.10	mg/L		10/07/11 08:30	10/07/11 16:40	1
Chromium	0.0018	J	0.0040	0.00087	mg/L		10/07/11 08:30	10/07/11 16:40	1
Cobalt	ND		0.0040	0.00063	mg/L		10/07/11 08:30	10/07/11 16:40	1
Copper	0.020		0.010	0.0015	mg/L		10/07/11 08:30	10/07/11 16:40	1
Iron	1.2		0.050	0.019	mg/L		10/07/11 08:30	10/07/11 16:40	1
Lead	ND		0.0050	0.0030	mg/L		10/07/11 08:30	10/07/11 16:40	1
Magnesium	12.1		0.20	0.043	mg/L		10/07/11 08:30	10/07/11 16:40	1
Manganese	0.33	B	0.0030	0.00030	mg/L		10/07/11 08:30	10/07/11 16:40	1
Nickel	0.0089	J	0.010	0.0013	mg/L		10/07/11 08:30	10/07/11 16:40	1
Potassium	2.2		0.50	0.20	mg/L		10/07/11 08:30	10/07/11 16:40	1
Selenium	ND		0.015	0.0087	mg/L		10/07/11 08:30	10/07/11 16:40	1
Silver	ND		0.0030	0.0017	mg/L		10/07/11 08:30	10/07/11 16:40	1
Sodium	3.7		1.0	0.32	mg/L		10/07/11 08:30	10/07/11 16:40	1
Thallium	ND		0.020	0.010	mg/L		10/07/11 08:30	10/07/11 16:40	1
Vanadium	0.0011	J	0.0050	0.0011	mg/L		10/07/11 08:30	10/07/11 16:40	1
Zinc	0.89		0.010	0.0017	mg/L		10/07/11 08:30	10/07/11 16:40	1

Method: 7470A - Mercury (CVAA)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Mercury	ND		0.00020	0.00012	mg/L		10/06/11 11:40	10/06/11 17:16	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: TRIP BLANK

Lab Sample ID: 480-10739-14

Date Collected: 10/04/11 00:00

Matrix: Water

Date Received: 10/05/11 09:35

Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/16/11 19:55	1
1,1,2,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/16/11 19:55	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/16/11 19:55	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/16/11 19:55	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/16/11 19:55	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/16/11 19:55	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/16/11 19:55	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/16/11 19:55	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/16/11 19:55	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/16/11 19:55	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/16/11 19:55	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/16/11 19:55	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/16/11 19:55	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/16/11 19:55	1
2-Hexanone	ND		5.0	1.2	ug/L			10/16/11 19:55	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/16/11 19:55	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/16/11 19:55	1
Acetone	ND		10	3.0	ug/L			10/16/11 19:55	1
Benzene	ND		1.0	0.41	ug/L			10/16/11 19:55	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/16/11 19:55	1
Bromoform	ND		1.0	0.26	ug/L			10/16/11 19:55	1
Bromomethane	ND		1.0	0.69	ug/L			10/16/11 19:55	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/16/11 19:55	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/16/11 19:55	1
Chlorobenzene	ND		1.0	0.75	ug/L			10/16/11 19:55	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/16/11 19:55	1
Chloroethane	ND		1.0	0.32	ug/L			10/16/11 19:55	1
Chloroform	ND		1.0	0.34	ug/L			10/16/11 19:55	1
Chloromethane	ND		1.0	0.35	ug/L			10/16/11 19:55	1
cis-1,2-Dichloroethene	ND		1.0	0.81	ug/L			10/16/11 19:55	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/16/11 19:55	1
Cyclohexane	ND		1.0	0.18	ug/L			10/16/11 19:55	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/16/11 19:55	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/16/11 19:55	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/16/11 19:55	1
Methyl acetate	ND		1.0	0.50	ug/L			10/16/11 19:55	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/16/11 19:55	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/16/11 19:55	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/16/11 19:55	1
Styrene	ND		1.0	0.73	ug/L			10/16/11 19:55	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/16/11 19:55	1
Toluene	ND		1.0	0.51	ug/L			10/16/11 19:55	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/16/11 19:55	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/16/11 19:55	1
Trichloroethene	ND		1.0	0.46	ug/L			10/16/11 19:55	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/16/11 19:55	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/16/11 19:55	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/16/11 19:55	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: TRIP BLANK

Lab Sample ID: 480-10739-14

Date Collected: 10/04/11 00:00

Matrix: Water

Date Received: 10/05/11 09:35

<i>Surrogate</i>	<i>% Recovery</i>	<i>Qualifier</i>	<i>Limits</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
1,2-Dichloroethane-d4 (Surr)	128		66 - 137		10/16/11 19:55	1
Toluene-d8 (Surr)	118		71 - 126		10/16/11 19:55	1
4-Bromofluorobenzene (Surr)	107		73 - 120		10/16/11 19:55	1

Lab Chronicle

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-113

Date Collected: 10/04/11 13:35

Date Received: 10/05/11 09:35

Lab Sample ID: 480-10739-1

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared Or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	35423	10/14/11 17:14	DC	TAL BUF
Total/NA	Prep	7470A			34236	10/06/11 11:40	MM	TAL BUF
Total/NA	Analysis	7470A		1	34333	10/06/11 16:49	MM	TAL BUF
Total/NA	Prep	3005A			34291	10/07/11 08:30	JM	TAL BUF
Total/NA	Analysis	6010B		1	34669	10/07/11 16:09	LH	TAL BUF

Client Sample ID: MW-107S

Date Collected: 10/04/11 13:25

Date Received: 10/05/11 09:35

Lab Sample ID: 480-10739-2

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared Or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	35423	10/14/11 17:36	DC	TAL BUF
Total/NA	Prep	7470A			34236	10/06/11 11:40	MM	TAL BUF
Total/NA	Analysis	7470A		1	34333	10/06/11 16:51	MM	TAL BUF
Total/NA	Prep	3005A			34291	10/07/11 08:30	JM	TAL BUF
Total/NA	Analysis	6010B		1	34669	10/07/11 16:11	LH	TAL BUF

Client Sample ID: MW-107I

Date Collected: 10/04/11 13:20

Date Received: 10/05/11 09:35

Lab Sample ID: 480-10739-3

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared Or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	35423	10/14/11 17:58	DC	TAL BUF
Total/NA	Prep	7470A			34236	10/06/11 11:40	MM	TAL BUF
Total/NA	Analysis	7470A		1	34333	10/06/11 16:52	MM	TAL BUF
Total/NA	Prep	3005A			34291	10/07/11 08:30	JM	TAL BUF
Total/NA	Analysis	6010B		1	34669	10/07/11 16:13	LH	TAL BUF

Client Sample ID: MW-101I

Date Collected: 10/04/11 13:00

Date Received: 10/05/11 09:35

Lab Sample ID: 480-10739-4

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared Or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	35423	10/14/11 18:20	DC	TAL BUF
Total/NA	Prep	7470A			34236	10/06/11 11:40	MM	TAL BUF
Total/NA	Analysis	7470A		1	34333	10/06/11 16:54	MM	TAL BUF
Total/NA	Prep	3005A			34291	10/07/11 08:30	JM	TAL BUF
Total/NA	Analysis	6010B		1	34669	10/07/11 16:16	LH	TAL BUF

Lab Chronicle

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-101D

Lab Sample ID: 480-10739-5

Date Collected: 10/04/11 13:05

Matrix: Water

Date Received: 10/05/11 09:35

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared Or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	35423	10/14/11 18:42	DC	TAL BUF
Total/NA	Prep	7470A			34236	10/06/11 11:40	MM	TAL BUF
Total/NA	Analysis	7470A		1	34333	10/06/11 16:56	MM	TAL BUF
Total/NA	Prep	3005A			34291	10/07/11 08:30	JM	TAL BUF
Total/NA	Analysis	6010B		1	34669	10/07/11 16:22	LH	TAL BUF

Client Sample ID: MW-104I

Lab Sample ID: 480-10739-6

Date Collected: 10/04/11 12:20

Matrix: Water

Date Received: 10/05/11 09:35

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared Or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	35423	10/14/11 19:04	DC	TAL BUF
Total/NA	Prep	7470A			34236	10/06/11 11:40	MM	TAL BUF
Total/NA	Analysis	7470A		1	34333	10/06/11 17:00	MM	TAL BUF
Total/NA	Prep	3005A			34291	10/07/11 08:30	JM	TAL BUF
Total/NA	Analysis	6010B		1	34669	10/07/11 16:24	LH	TAL BUF

Client Sample ID: MW-104D

Lab Sample ID: 480-10739-7

Date Collected: 10/04/11 12:15

Matrix: Water

Date Received: 10/05/11 09:35

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared Or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	35423	10/14/11 19:26	DC	TAL BUF
Total/NA	Prep	7470A			34236	10/06/11 11:40	MM	TAL BUF
Total/NA	Analysis	7470A		1	34333	10/06/11 17:02	MM	TAL BUF
Total/NA	Prep	3005A			34291	10/07/11 08:30	JM	TAL BUF
Total/NA	Analysis	6010B		1	34669	10/07/11 16:27	LH	TAL BUF

Client Sample ID: MW-103I

Lab Sample ID: 480-10739-8

Date Collected: 10/04/11 12:05

Matrix: Water

Date Received: 10/05/11 09:35

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared Or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	35423	10/14/11 19:48	DC	TAL BUF
Total/NA	Prep	7470A			34236	10/06/11 11:40	MM	TAL BUF
Total/NA	Analysis	7470A		1	34333	10/06/11 17:04	MM	TAL BUF
Total/NA	Prep	3005A			34291	10/07/11 08:30	JM	TAL BUF
Total/NA	Analysis	6010B		1	34669	10/07/11 16:29	LH	TAL BUF

Lab Chronicle

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-103D

Lab Sample ID: 480-10739-9

Date Collected: 10/04/11 12:00

Matrix: Water

Date Received: 10/05/11 09:35

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared Or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	35423	10/14/11 20:10	DC	TAL BUF
Total/NA	Prep	7470A			34236	10/06/11 11:40	MM	TAL BUF
Total/NA	Analysis	7470A		1	34333	10/06/11 17:06	MM	TAL BUF
Total/NA	Prep	3005A			34291	10/07/11 08:30	JM	TAL BUF
Total/NA	Analysis	6010B		1	34669	10/07/11 16:31	LH	TAL BUF

Client Sample ID: MW-102D

Lab Sample ID: 480-10739-10

Date Collected: 10/04/11 11:00

Matrix: Water

Date Received: 10/05/11 09:35

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared Or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	35423	10/14/11 20:32	DC	TAL BUF
Total/NA	Prep	7470A			34236	10/06/11 11:40	MM	TAL BUF
Total/NA	Analysis	7470A		1	34333	10/06/11 17:07	MM	TAL BUF
Total/NA	Prep	3005A			34291	10/07/11 08:30	JM	TAL BUF
Total/NA	Analysis	6010B		1	34669	10/07/11 16:33	LH	TAL BUF

Client Sample ID: MW-109

Lab Sample ID: 480-10739-11

Date Collected: 10/04/11 11:15

Matrix: Water

Date Received: 10/05/11 09:35

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared Or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	35423	10/14/11 20:54	DC	TAL BUF
Total/NA	Prep	7470A			34236	10/06/11 11:40	MM	TAL BUF
Total/NA	Analysis	7470A		1	34333	10/06/11 17:09	MM	TAL BUF
Total/NA	Prep	3005A			34291	10/07/11 08:30	JM	TAL BUF
Total/NA	Analysis	6010B		1	34669	10/07/11 16:36	LH	TAL BUF

Client Sample ID: MW-108I

Lab Sample ID: 480-10739-12

Date Collected: 10/04/11 11:40

Matrix: Water

Date Received: 10/05/11 09:35

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared Or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	35423	10/14/11 21:16	DC	TAL BUF
Total/NA	Prep	7470A			34236	10/06/11 11:40	MM	TAL BUF
Total/NA	Analysis	7470A		1	34333	10/06/11 17:11	MM	TAL BUF
Total/NA	Prep	3005A			34291	10/07/11 08:30	JM	TAL BUF
Total/NA	Analysis	6010B		1	34669	10/07/11 16:38	LH	TAL BUF

Lab Chronicle

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-10739-1

Client Sample ID: MW-108D

Lab Sample ID: 480-10739-13

Date Collected: 10/04/11 11:30

Matrix: Water

Date Received: 10/05/11 09:35

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared Or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	35647	10/16/11 18:49	DC	TAL BUF
Total/NA	Prep	7470A			34236	10/06/11 11:40	MM	TAL BUF
Total/NA	Analysis	7470A		1	34333	10/06/11 17:16	MM	TAL BUF
Total/NA	Prep	3005A			34291	10/07/11 08:30	JM	TAL BUF
Total/NA	Analysis	6010B		1	34669	10/07/11 16:40	LH	TAL BUF

Client Sample ID: TRIP BLANK

Lab Sample ID: 480-10739-14

Date Collected: 10/04/11 00:00

Matrix: Water

Date Received: 10/05/11 09:35

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared Or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	35647	10/16/11 19:55	DC	TAL BUF

Laboratory References:

TAL BUF = TestAmerica Buffalo, 10 Hazelwood Drive, Amherst, NY 14228-2298, TEL (716)691-2600

Certification Summary

Client: New York State D.E.C.

TestAmerica Job ID: 480-10739-1

Project/Site: NYSDEC - Patton's Busy Bee:site #902014

Laboratory	Authority	Program	EPA Region	Certification ID
TestAmerica Buffalo	Arkansas	State Program	6	88-0686
TestAmerica Buffalo	California	NELAC	9	1169CA
TestAmerica Buffalo	Connecticut	State Program	1	PH-0568
TestAmerica Buffalo	Florida	NELAC	4	E87672
TestAmerica Buffalo	Georgia	Georgia EPD	4	N/A
TestAmerica Buffalo	Georgia	State Program	4	956
TestAmerica Buffalo	Illinois	NELAC	5	100325 / 200003
TestAmerica Buffalo	Iowa	State Program	7	374
TestAmerica Buffalo	Kansas	NELAC	7	E-10187
TestAmerica Buffalo	Kentucky	Kentucky UST	4	30
TestAmerica Buffalo	Kentucky	State Program	4	90029
TestAmerica Buffalo	Louisiana	NELAC	6	02031
TestAmerica Buffalo	Maine	State Program	1	NY0044
TestAmerica Buffalo	Maryland	State Program	3	294
TestAmerica Buffalo	Massachusetts	State Program	1	M-NY044
TestAmerica Buffalo	Michigan	State Program	5	9937
TestAmerica Buffalo	Minnesota	NELAC	5	036-999-337
TestAmerica Buffalo	New Hampshire	NELAC	1	2337
TestAmerica Buffalo	New Hampshire	NELAC	1	68-00281
TestAmerica Buffalo	New Jersey	NELAC	2	NY455
TestAmerica Buffalo	New York	NELAC	2	10026
TestAmerica Buffalo	North Dakota	State Program	8	R-176
TestAmerica Buffalo	Oklahoma	State Program	6	9421
TestAmerica Buffalo	Oregon	NELAC	10	NY200003
TestAmerica Buffalo	Pennsylvania	NELAC	3	68-00281
TestAmerica Buffalo	Tennessee	State Program	4	TN02970
TestAmerica Buffalo	Texas	NELAC	6	T104704412-08-TX
TestAmerica Buffalo	USDA	USDA		P330-08-00242
TestAmerica Buffalo	Virginia	NELAC Secondary AB	3	460185
TestAmerica Buffalo	Virginia	State Program	3	278
TestAmerica Buffalo	Washington	State Program	10	C1677
TestAmerica Buffalo	Wisconsin	State Program	5	998310390

Accreditation may not be offered or required for all methods and analytes reported in this package. Please contact your project manager for the laboratory's current list of certified methods and analytes.

Method Summary

Client: New York State D.E.C.

TestAmerica Job ID: 480-10739-1

Project/Site: NYSDEC - Patton's Busy Bee:site #902014

Method	Method Description	Protocol	Laboratory
8260B	Volatile Organic Compounds (GC/MS)	SW846	TAL BUF
6010B	Metals (ICP)	SW846	TAL BUF
7470A	Mercury (CVAA)	SW846	TAL BUF

Protocol References:

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

Laboratory References:

TAL BUF = TestAmerica Buffalo, 10 Hazelwood Drive, Amherst, NY 14228-2298, TEL (716)691-2600

Sample Summary

Client: New York State D.E.C.

TestAmerica Job ID: 480-10739-1

Project/Site: NYSDEC - Patton's Busy Bee:site #902014

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
480-10739-1	MW-113	Water	10/04/11 13:35	10/05/11 09:35
480-10739-2	MW-107S	Water	10/04/11 13:25	10/05/11 09:35
480-10739-3	MW-107I	Water	10/04/11 13:20	10/05/11 09:35
480-10739-4	MW-101I	Water	10/04/11 13:00	10/05/11 09:35
480-10739-5	MW-101D	Water	10/04/11 13:05	10/05/11 09:35
480-10739-6	MW-104I	Water	10/04/11 12:20	10/05/11 09:35
480-10739-7	MW-104D	Water	10/04/11 12:15	10/05/11 09:35
480-10739-8	MW-103I	Water	10/04/11 12:05	10/05/11 09:35
480-10739-9	MW-103D	Water	10/04/11 12:00	10/05/11 09:35
480-10739-10	MW-102D	Water	10/04/11 11:00	10/05/11 09:35
480-10739-11	MW-109	Water	10/04/11 11:15	10/05/11 09:35
480-10739-12	MW-108I	Water	10/04/11 11:40	10/05/11 09:35
480-10739-13	MW-108D	Water	10/04/11 11:30	10/05/11 09:35
480-10739-14	TRIP BLANK	Water	10/04/11 00:00	10/05/11 09:35

TestAmerica

THE LEADER IN ENVIRONMENTAL TESTING

Temperature on Receipt _____

Drinking Water? Yes ☐ No ☒

Chain of Custody Record

TAL-4124 (1007)

Client: HYSDEC Project Manager: BRIAN SADOWSKI Date: 10/4/11 Chain of Custody Number: 191006

Address: 270 MICHIGAN AVE. Telephone Number (Area Code)/Fax Number: 716-851-7220 Page: 1 of 2

City: BUFFALO State: NY Zip Code: 14203

Project Name and Location (State): FACTORIS BUSY BEE SITE

Contract/Purchase Order/Quote No.:

City	State	Zip Code	Site Contact	Lab Contact	Analysis (Attach list if more space is needed)	Special Instructions/ Conditions of Receipt									
Buffalo	NY	14203		B. FISCHER											
Project Name and Location (State)			Carrier/Waybill Number												
PATTON'S Bury Bore Site															
Contract/Purchase Order/Quote No.			Matrix		Containers & Preservatives										
Sample I.D. No. and Description (Containers for each sample may be combined on one line)	Date	Time	Air	Aqueous	Sed	Soil	Unpres.	H2SO4	HNO3	HCl	NaOH	ZnAc/NaOH			
MW-113	10/4/11	1335	✓						1	3			✓	✓	LOW VOL. IN METALS SCRAP IF CANT RUN
MW-107 S		1325	✓						1	3			✓	✓	
MW-107 I *		1320	✓						1	3			✓	✓	
MW-101 I		1300	✓						1	3			✓	✓	
MW-101 d		1305	✓						1	3			✓	✓	
MW-104 I		1220	✓						1	3			✓	✓	
MW-104 d		1215	✓						1	3			✓	✓	
MW-103 I		1205	✓						1	3			✓	✓	
MW-103 d		1200	✓						1	3			✓	✓	
MW-102 d		1100	✓						1	3			✓	✓	
MW-109		1115	✓						1	3			✓	✓	

Possible Hazard Identification: ☐ Non-Hazard ☐ Flammable ☐ Skin Irritant ☐ Poison B ☐ Unknown

Sample Disposal: ☐ Return To Client ☐ Disposal By Lab ☐ Archive For _____ Months (A fee may be assessed if samples are retained longer than 1 month)

Turn Around Time Required: ☐ 24 Hours ☐ 48 Hours ☐ 7 Days ☐ 14 Days ☐ 21 Days ☒ Other: STD.

1. Relinquished By: Way Johnson Date: 10/5/11 Time: 9:35

2. Relinquished By: Way Johnson Date: 10/5/11 Time: 0935

3. Relinquished By: _____ Date: _____ Time: _____

Comments: 3.9°C

DISTRIBUTION: WHITE - Returned to Client with Report; CANARY - Stays with the Sample; PINK - Field Copy

1
2
3
4
5
6
7
8
9
10
11

Login Sample Receipt Checklist

Client: New York State D.E.C.

Job Number: 480-10739-1

Login Number: 10739

List Source: TestAmerica Buffalo

List Number: 1

Creator: Wienke, Robert

Question	Answer	Comment
Radioactivity either was not measured or, if measured, is at or below background	True	
The cooler's custody seal, if present, is intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	True	
There are no discrepancies between the sample IDs on the containers and the COC.	True	
Samples are received within Holding Time.	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
VOA sample vials do not have headspace or bubble is <6mm (1/4") in diameter.	True	
If necessary, staff have been informed of any short hold time or quick TAT needs	True	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Sampling Company provided.	False	
Samples received within 48 hours of sampling.	True	
Samples requiring field filtration have been filtered in the field.	N/A	
Chlorine Residual checked.	N/A	