

TestAmerica

THE LEADER IN ENVIRONMENTAL TESTING

ANALYTICAL REPORT

TestAmerica Laboratories, Inc.

TestAmerica Buffalo

10 Hazelwood Drive

Amherst, NY 14228-2298

Tel: (716)691-2600

TestAmerica Job ID: 480-47249-1

Client Project/Site: NYSDEC - Patton's Busy Bee:site #902014


For:

New York State D.E.C.

270 Michigan Avenue

Buffalo, New York 14203

Attn: Mr. Brian Sadowski



Authorized for release by:

10/18/2013 11:18:55 AM

Joe Giacomazza, Project Administrator

joe.giacomazza@testamericainc.com

Designee for

Brian Fischer, Project Manager II

(716)504-9835

brian.fischer@testamericainc.com

LINKS

Review your project
results through

TotalAccess

Have a Question?



Visit us at:

www.testamericainc.com

The test results in this report meet all 2003 NELAC and 2009 TNI requirements for accredited parameters, exceptions are noted in this report. This report may not be reproduced except in full, and with written approval from the laboratory. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

This report has been electronically signed and authorized by the signatory. Electronic signature is intended to be the legally binding equivalent of a traditionally handwritten signature.

Results relate only to the items tested and the sample(s) as received by the laboratory.

1

2

3

4

5

6

7


8

9

10

11

I certify that this data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed within the body of this report. Release of the data contained in this sample data package and in the electronic data deliverable has been authorized by the Laboratory Manager or his/her designee, as verified by the following signature.



Joe Giacomazza
Project Administrator
10/18/2013 11:18:55 AM



Table of Contents

Cover Page	1
Table of Contents	3
Definitions	4
Case Narrative	5
Client Sample Results	6
Chronicle	32
Certification Summary	36
Method Summary	37
Sample Summary	38
Chain of Custody	39
Receipt Checklists	41

Definitions/Glossary

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Qualifiers

GC/MS VOA

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Metals

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
α	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CNF	Contains no Free Liquid
DER	Duplicate error ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision level concentration
MDA	Minimum detectable activity
EDL	Estimated Detection Limit
MDC	Minimum detectable concentration
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
NC	Not Calculated
ND	Not detected at the reporting limit (or MDL or EDL if shown)
PQL	Practical Quantitation Limit
QC	Quality Control
RER	Relative error ratio
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)

Case Narrative

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Job ID: 480-47249-1

Laboratory: TestAmerica Buffalo

Narrative

Job Narrative
480-47249-1

Receipt

The samples were received on 10/4/2013 9:50 AM; the samples arrived in good condition, properly preserved and, where required, on ice. The temperature of the cooler at receipt was 3.1° C.

GC/MS VOA

Method(s) 8260C: The large number of analytes included in the continuing calibration verification (CCV) for batch 145298, gives a high probability that one or more analytes will be outside acceptance criteria. As indicated in the reference method, analysis may proceed as long as no more than 20% of the analytes of interest are outside the method-defined %D criteria.

Method(s) 8260C: The matrix spike / matrix spike duplicate (MS/MSD) precision for batch 145298 was outside control limits.

No other analytical or quality issues were noted.

Metals

No analytical or quality issues were noted.



Client Sample Results

Client: New York State D.E.C.
 Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-107S

Lab Sample ID: 480-47249-1

Date Collected: 10/03/13 09:55

Matrix: Water

Date Received: 10/04/13 09:50

Method: 8260C - Volatile Organic Compounds by GC/MS

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/16/13 14:01	1
1,1,2,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/16/13 14:01	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/16/13 14:01	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/16/13 14:01	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/16/13 14:01	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/16/13 14:01	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/16/13 14:01	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/16/13 14:01	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/16/13 14:01	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/16/13 14:01	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/16/13 14:01	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/16/13 14:01	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/16/13 14:01	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/16/13 14:01	1
2-Hexanone	ND		5.0	1.2	ug/L			10/16/13 14:01	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/16/13 14:01	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/16/13 14:01	1
Acetone	7.6	J	10	3.0	ug/L			10/16/13 14:01	1
Benzene	ND		1.0	0.41	ug/L			10/16/13 14:01	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/16/13 14:01	1
Bromoform	ND		1.0	0.26	ug/L			10/16/13 14:01	1
Bromomethane	ND		1.0	0.69	ug/L			10/16/13 14:01	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/16/13 14:01	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/16/13 14:01	1
Chlorobenzene	ND		1.0	0.75	ug/L			10/16/13 14:01	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/16/13 14:01	1
Chloroethane	ND		1.0	0.32	ug/L			10/16/13 14:01	1
Chloroform	ND		1.0	0.34	ug/L			10/16/13 14:01	1
Chloromethane	ND		1.0	0.35	ug/L			10/16/13 14:01	1
cis-1,2-Dichloroethene	ND		1.0	0.81	ug/L			10/16/13 14:01	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/16/13 14:01	1
Cyclohexane	ND		1.0	0.18	ug/L			10/16/13 14:01	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/16/13 14:01	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/16/13 14:01	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/16/13 14:01	1
Methyl acetate	ND		1.0	0.50	ug/L			10/16/13 14:01	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/16/13 14:01	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/16/13 14:01	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/16/13 14:01	1
Styrene	ND		1.0	0.73	ug/L			10/16/13 14:01	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/16/13 14:01	1
Toluene	ND		1.0	0.51	ug/L			10/16/13 14:01	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/16/13 14:01	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/16/13 14:01	1
Trichloroethene	ND		1.0	0.46	ug/L			10/16/13 14:01	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/16/13 14:01	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/16/13 14:01	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/16/13 14:01	1

TestAmerica Buffalo

Client Sample Results

Client: New York State D.E.C.
 Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-107S

Lab Sample ID: 480-47249-1

Date Collected: 10/03/13 09:55

Matrix: Water

Date Received: 10/04/13 09:50

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	96		66 - 137		10/16/13 14:01	1
Toluene-d8 (Surr)	101		71 - 126		10/16/13 14:01	1
4-Bromofluorobenzene (Surr)	99		73 - 120		10/16/13 14:01	1

Method: 6010C - Metals (ICP)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Aluminum	ND		0.20	0.060	mg/L		10/07/13 08:25	10/07/13 23:48	1
Antimony	ND		0.020	0.0068	mg/L		10/07/13 08:25	10/07/13 23:48	1
Arsenic	ND		0.010	0.0056	mg/L		10/07/13 08:25	10/09/13 12:20	1
Barium	0.059		0.0020	0.00070	mg/L		10/07/13 08:25	10/07/13 23:48	1
Beryllium	ND		0.0020	0.00030	mg/L		10/07/13 08:25	10/07/13 23:48	1
Cadmium	0.0023		0.0010	0.00050	mg/L		10/07/13 08:25	10/07/13 23:48	1
Calcium	42.4		0.50	0.10	mg/L		10/07/13 08:25	10/07/13 23:48	1
Chromium	ND		0.0040	0.0010	mg/L		10/07/13 08:25	10/07/13 23:48	1
Cobalt	ND		0.0040	0.00063	mg/L		10/07/13 08:25	10/07/13 23:48	1
Copper	0.0032	J	0.010	0.0016	mg/L		10/07/13 08:25	10/07/13 23:48	1
Iron	0.038	J	0.050	0.019	mg/L		10/07/13 08:25	10/07/13 23:48	1
Lead	ND		0.0050	0.0030	mg/L		10/07/13 08:25	10/07/13 23:48	1
Magnesium	11.0		0.20	0.043	mg/L		10/07/13 08:25	10/07/13 23:48	1
Manganese	0.92		0.0030	0.00040	mg/L		10/07/13 08:25	10/07/13 23:48	1
Nickel	0.0044	J	0.010	0.0013	mg/L		10/07/13 08:25	10/07/13 23:48	1
Potassium	1.6		0.50	0.10	mg/L		10/07/13 08:25	10/07/13 23:48	1
Selenium	ND		0.015	0.0087	mg/L		10/07/13 08:25	10/07/13 23:48	1
Silver	ND		0.0030	0.0017	mg/L		10/07/13 08:25	10/07/13 23:48	1
Sodium	13.7		1.0	0.32	mg/L		10/07/13 08:25	10/07/13 23:48	1
Thallium	ND		0.020	0.010	mg/L		10/07/13 08:25	10/07/13 23:48	1
Vanadium	ND		0.0050	0.0015	mg/L		10/07/13 08:25	10/07/13 23:48	1
Zinc	0.0015	J	0.010	0.0015	mg/L		10/07/13 08:25	10/07/13 23:48	1

Method: 7470A - Mercury (CVAA)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Mercury	ND		0.00020	0.00012	mg/L		10/07/13 09:30	10/07/13 13:21	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-1071

Lab Sample ID: 480-47249-2

Date Collected: 10/03/13 10:00

Matrix: Water

Date Received: 10/04/13 09:50

Method: 8260C - Volatile Organic Compounds by GC/MS

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/16/13 14:29	1
1,1,2,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/16/13 14:29	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/16/13 14:29	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/16/13 14:29	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/16/13 14:29	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/16/13 14:29	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/16/13 14:29	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/16/13 14:29	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/16/13 14:29	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/16/13 14:29	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/16/13 14:29	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/16/13 14:29	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/16/13 14:29	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/16/13 14:29	1
2-Hexanone	ND		5.0	1.2	ug/L			10/16/13 14:29	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/16/13 14:29	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/16/13 14:29	1
Acetone	5.8	J	10	3.0	ug/L			10/16/13 14:29	1
Benzene	ND		1.0	0.41	ug/L			10/16/13 14:29	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/16/13 14:29	1
Bromoform	ND		1.0	0.26	ug/L			10/16/13 14:29	1
Bromomethane	ND		1.0	0.69	ug/L			10/16/13 14:29	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/16/13 14:29	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/16/13 14:29	1
Chlorobenzene	ND		1.0	0.75	ug/L			10/16/13 14:29	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/16/13 14:29	1
Chloroethane	ND		1.0	0.32	ug/L			10/16/13 14:29	1
Chloroform	ND		1.0	0.34	ug/L			10/16/13 14:29	1
Chloromethane	ND		1.0	0.35	ug/L			10/16/13 14:29	1
cis-1,2-Dichloroethene	ND		1.0	0.81	ug/L			10/16/13 14:29	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/16/13 14:29	1
Cyclohexane	ND		1.0	0.18	ug/L			10/16/13 14:29	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/16/13 14:29	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/16/13 14:29	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/16/13 14:29	1
Methyl acetate	ND		1.0	0.50	ug/L			10/16/13 14:29	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/16/13 14:29	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/16/13 14:29	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/16/13 14:29	1
Styrene	ND		1.0	0.73	ug/L			10/16/13 14:29	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/16/13 14:29	1
Toluene	ND		1.0	0.51	ug/L			10/16/13 14:29	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/16/13 14:29	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/16/13 14:29	1
Trichloroethene	ND		1.0	0.46	ug/L			10/16/13 14:29	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/16/13 14:29	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/16/13 14:29	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/16/13 14:29	1

TestAmerica Buffalo

Client Sample Results

Client: New York State D.E.C.
 Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-1071

Lab Sample ID: 480-47249-2

Date Collected: 10/03/13 10:00

Matrix: Water

Date Received: 10/04/13 09:50

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	98		66 - 137		10/16/13 14:29	1
Toluene-d8 (Surr)	100		71 - 126		10/16/13 14:29	1
4-Bromofluorobenzene (Surr)	99		73 - 120		10/16/13 14:29	1

Method: 6010C - Metals (ICP)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Aluminum	12.1		0.20	0.060	mg/L		10/07/13 08:25	10/07/13 23:50	1
Antimony	ND		0.020	0.0068	mg/L		10/07/13 08:25	10/07/13 23:50	1
Arsenic	0.0075	J	0.010	0.0056	mg/L		10/07/13 08:25	10/09/13 12:22	1
Barium	0.24		0.0020	0.00070	mg/L		10/07/13 08:25	10/07/13 23:50	1
Beryllium	0.00072	J	0.0020	0.00030	mg/L		10/07/13 08:25	10/07/13 23:50	1
Cadmium	0.0040		0.0010	0.00050	mg/L		10/07/13 08:25	10/07/13 23:50	1
Calcium	75.8		0.50	0.10	mg/L		10/07/13 08:25	10/07/13 23:50	1
Chromium	0.017		0.0040	0.0010	mg/L		10/07/13 08:25	10/07/13 23:50	1
Cobalt	0.017		0.0040	0.00063	mg/L		10/07/13 08:25	10/07/13 23:50	1
Copper	0.021		0.010	0.0016	mg/L		10/07/13 08:25	10/07/13 23:50	1
Iron	17.6		0.050	0.019	mg/L		10/07/13 08:25	10/07/13 23:50	1
Lead	0.037		0.0050	0.0030	mg/L		10/07/13 08:25	10/07/13 23:50	1
Magnesium	22.9		0.20	0.043	mg/L		10/07/13 08:25	10/07/13 23:50	1
Manganese	0.95		0.0030	0.00040	mg/L		10/07/13 08:25	10/07/13 23:50	1
Nickel	0.021		0.010	0.0013	mg/L		10/07/13 08:25	10/07/13 23:50	1
Potassium	7.2		0.50	0.10	mg/L		10/07/13 08:25	10/07/13 23:50	1
Selenium	ND		0.015	0.0087	mg/L		10/07/13 08:25	10/07/13 23:50	1
Silver	ND		0.0030	0.0017	mg/L		10/07/13 08:25	10/07/13 23:50	1
Sodium	8.3		1.0	0.32	mg/L		10/07/13 08:25	10/07/13 23:50	1
Thallium	ND		0.020	0.010	mg/L		10/07/13 08:25	10/07/13 23:50	1
Vanadium	0.015		0.0050	0.0015	mg/L		10/07/13 08:25	10/07/13 23:50	1
Zinc	0.16		0.010	0.0015	mg/L		10/07/13 08:25	10/07/13 23:50	1

Method: 7470A - Mercury (CVAA)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Mercury	ND		0.00020	0.00012	mg/L		10/07/13 09:30	10/07/13 13:23	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-101D

Lab Sample ID: 480-47249-3

Date Collected: 10/03/13 11:00

Matrix: Water

Date Received: 10/04/13 09:50

Method: 8260C - Volatile Organic Compounds by GC/MS

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/16/13 14:57	1
1,1,2,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/16/13 14:57	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/16/13 14:57	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/16/13 14:57	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/16/13 14:57	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/16/13 14:57	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/16/13 14:57	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/16/13 14:57	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/16/13 14:57	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/16/13 14:57	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/16/13 14:57	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/16/13 14:57	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/16/13 14:57	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/16/13 14:57	1
2-Hexanone	ND		5.0	1.2	ug/L			10/16/13 14:57	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/16/13 14:57	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/16/13 14:57	1
Acetone	5.5	J	10	3.0	ug/L			10/16/13 14:57	1
Benzene	0.47	J	1.0	0.41	ug/L			10/16/13 14:57	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/16/13 14:57	1
Bromoform	ND		1.0	0.26	ug/L			10/16/13 14:57	1
Bromomethane	ND		1.0	0.69	ug/L			10/16/13 14:57	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/16/13 14:57	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/16/13 14:57	1
Chlorobenzene	1.9		1.0	0.75	ug/L			10/16/13 14:57	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/16/13 14:57	1
Chloroethane	ND		1.0	0.32	ug/L			10/16/13 14:57	1
Chloroform	ND		1.0	0.34	ug/L			10/16/13 14:57	1
Chloromethane	ND		1.0	0.35	ug/L			10/16/13 14:57	1
cis-1,2-Dichloroethene	ND		1.0	0.81	ug/L			10/16/13 14:57	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/16/13 14:57	1
Cyclohexane	ND		1.0	0.18	ug/L			10/16/13 14:57	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/16/13 14:57	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/16/13 14:57	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/16/13 14:57	1
Methyl acetate	ND		1.0	0.50	ug/L			10/16/13 14:57	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/16/13 14:57	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/16/13 14:57	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/16/13 14:57	1
Styrene	ND		1.0	0.73	ug/L			10/16/13 14:57	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/16/13 14:57	1
Toluene	ND		1.0	0.51	ug/L			10/16/13 14:57	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/16/13 14:57	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/16/13 14:57	1
Trichloroethene	ND		1.0	0.46	ug/L			10/16/13 14:57	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/16/13 14:57	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/16/13 14:57	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/16/13 14:57	1

TestAmerica Buffalo

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-101D

Lab Sample ID: 480-47249-3

Date Collected: 10/03/13 11:00

Matrix: Water

Date Received: 10/04/13 09:50

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	97		66 - 137		10/16/13 14:57	1
Toluene-d8 (Surr)	99		71 - 126		10/16/13 14:57	1
4-Bromofluorobenzene (Surr)	97		73 - 120		10/16/13 14:57	1

Method: 6010C - Metals (ICP)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Aluminum	ND		0.20	0.060	mg/L		10/07/13 08:25	10/07/13 23:53	1
Antimony	ND		0.020	0.0068	mg/L		10/07/13 08:25	10/07/13 23:53	1
Arsenic	ND		0.010	0.0056	mg/L		10/07/13 08:25	10/09/13 12:25	1
Barium	0.48		0.0020	0.00070	mg/L		10/07/13 08:25	10/07/13 23:53	1
Beryllium	ND		0.0020	0.00030	mg/L		10/07/13 08:25	10/07/13 23:53	1
Cadmium	0.00059	J	0.0010	0.00050	mg/L		10/07/13 08:25	10/07/13 23:53	1
Calcium	133		0.50	0.10	mg/L		10/07/13 08:25	10/07/13 23:53	1
Chromium	0.0014	J	0.0040	0.0010	mg/L		10/07/13 08:25	10/07/13 23:53	1
Cobalt	0.0011	J	0.0040	0.00063	mg/L		10/07/13 08:25	10/07/13 23:53	1
Copper	0.0034	J	0.010	0.0016	mg/L		10/07/13 08:25	10/07/13 23:53	1
Iron	0.099		0.050	0.019	mg/L		10/07/13 08:25	10/07/13 23:53	1
Lead	ND		0.0050	0.0030	mg/L		10/07/13 08:25	10/07/13 23:53	1
Magnesium	77.4		0.20	0.043	mg/L		10/07/13 08:25	10/07/13 23:53	1
Manganese	3.0		0.0030	0.00040	mg/L		10/07/13 08:25	10/07/13 23:53	1
Nickel	0.0057	J	0.010	0.0013	mg/L		10/07/13 08:25	10/07/13 23:53	1
Potassium	11.3		0.50	0.10	mg/L		10/07/13 08:25	10/07/13 23:53	1
Selenium	ND		0.015	0.0087	mg/L		10/07/13 08:25	10/07/13 23:53	1
Silver	ND		0.0030	0.0017	mg/L		10/07/13 08:25	10/07/13 23:53	1
Sodium	39.6		1.0	0.32	mg/L		10/07/13 08:25	10/07/13 23:53	1
Thallium	ND		0.020	0.010	mg/L		10/07/13 08:25	10/07/13 23:53	1
Vanadium	ND		0.0050	0.0015	mg/L		10/07/13 08:25	10/07/13 23:53	1
Zinc	0.0024	J	0.010	0.0015	mg/L		10/07/13 08:25	10/07/13 23:53	1

Method: 7470A - Mercury (CVAA)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Mercury	ND		0.00020	0.00012	mg/L		10/07/13 09:30	10/07/13 13:25	1

Client Sample Results

Client: New York State D.E.C.
 Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-1011

Lab Sample ID: 480-47249-4

Date Collected: 10/03/13 11:05

Matrix: Water

Date Received: 10/04/13 09:50

Method: 8260C - Volatile Organic Compounds by GC/MS

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/16/13 15:25	1
1,1,2,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/16/13 15:25	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/16/13 15:25	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/16/13 15:25	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/16/13 15:25	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/16/13 15:25	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/16/13 15:25	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/16/13 15:25	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/16/13 15:25	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/16/13 15:25	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/16/13 15:25	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/16/13 15:25	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/16/13 15:25	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/16/13 15:25	1
2-Hexanone	ND		5.0	1.2	ug/L			10/16/13 15:25	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/16/13 15:25	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/16/13 15:25	1
Acetone	7.8	J	10	3.0	ug/L			10/16/13 15:25	1
Benzene	ND		1.0	0.41	ug/L			10/16/13 15:25	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/16/13 15:25	1
Bromoform	ND		1.0	0.26	ug/L			10/16/13 15:25	1
Bromomethane	ND		1.0	0.69	ug/L			10/16/13 15:25	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/16/13 15:25	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/16/13 15:25	1
Chlorobenzene	ND		1.0	0.75	ug/L			10/16/13 15:25	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/16/13 15:25	1
Chloroethane	ND		1.0	0.32	ug/L			10/16/13 15:25	1
Chloroform	ND		1.0	0.34	ug/L			10/16/13 15:25	1
Chloromethane	ND		1.0	0.35	ug/L			10/16/13 15:25	1
cis-1,2-Dichloroethene	ND		1.0	0.81	ug/L			10/16/13 15:25	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/16/13 15:25	1
Cyclohexane	ND		1.0	0.18	ug/L			10/16/13 15:25	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/16/13 15:25	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/16/13 15:25	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/16/13 15:25	1
Methyl acetate	ND		1.0	0.50	ug/L			10/16/13 15:25	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/16/13 15:25	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/16/13 15:25	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/16/13 15:25	1
Styrene	ND		1.0	0.73	ug/L			10/16/13 15:25	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/16/13 15:25	1
Toluene	ND		1.0	0.51	ug/L			10/16/13 15:25	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/16/13 15:25	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/16/13 15:25	1
Trichloroethene	ND		1.0	0.46	ug/L			10/16/13 15:25	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/16/13 15:25	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/16/13 15:25	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/16/13 15:25	1

TestAmerica Buffalo

Client Sample Results

Client: New York State D.E.C.
 Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-1011

Lab Sample ID: 480-47249-4

Date Collected: 10/03/13 11:05

Matrix: Water

Date Received: 10/04/13 09:50

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	99		66 - 137		10/16/13 15:25	1
Toluene-d8 (Surr)	101		71 - 126		10/16/13 15:25	1
4-Bromofluorobenzene (Surr)	98		73 - 120		10/16/13 15:25	1

Method: 6010C - Metals (ICP)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Aluminum	123		0.20	0.060	mg/L		10/07/13 08:25	10/07/13 23:55	1
Antimony	ND		0.020	0.0068	mg/L		10/07/13 08:25	10/07/13 23:55	1
Arsenic	0.077		0.010	0.0056	mg/L		10/07/13 08:25	10/09/13 12:27	1
Barium	0.99		0.0020	0.00070	mg/L		10/07/13 08:25	10/07/13 23:55	1
Beryllium	0.0074		0.0020	0.00030	mg/L		10/07/13 08:25	10/07/13 23:55	1
Cadmium	0.0033		0.0010	0.00050	mg/L		10/07/13 08:25	10/07/13 23:55	1
Calcium	112		0.50	0.10	mg/L		10/07/13 08:25	10/07/13 23:55	1
Chromium	0.15		0.0040	0.0010	mg/L		10/07/13 08:25	10/07/13 23:55	1
Cobalt	0.12		0.0040	0.00063	mg/L		10/07/13 08:25	10/07/13 23:55	1
Copper	0.20		0.010	0.0016	mg/L		10/07/13 08:25	10/07/13 23:55	1
Iron	220		0.050	0.019	mg/L		10/07/13 08:25	10/07/13 23:55	1
Lead	0.10		0.0050	0.0030	mg/L		10/07/13 08:25	10/07/13 23:55	1
Magnesium	81.6		0.20	0.043	mg/L		10/07/13 08:25	10/07/13 23:55	1
Manganese	3.4		0.0030	0.00040	mg/L		10/07/13 08:25	10/07/13 23:55	1
Nickel	0.23		0.010	0.0013	mg/L		10/07/13 08:25	10/07/13 23:55	1
Potassium	34.8		0.50	0.10	mg/L		10/07/13 08:25	10/07/13 23:55	1
Selenium	ND		0.015	0.0087	mg/L		10/07/13 08:25	10/07/13 23:55	1
Silver	ND		0.0030	0.0017	mg/L		10/07/13 08:25	10/07/13 23:55	1
Sodium	7.6		1.0	0.32	mg/L		10/07/13 08:25	10/07/13 23:55	1
Thallium	ND		0.020	0.010	mg/L		10/07/13 08:25	10/07/13 23:55	1
Vanadium	0.14		0.0050	0.0015	mg/L		10/07/13 08:25	10/07/13 23:55	1
Zinc	0.65		0.010	0.0015	mg/L		10/07/13 08:25	10/07/13 23:55	1

Method: 7470A - Mercury (CVAA)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Mercury	ND		0.00020	0.00012	mg/L		10/07/13 09:30	10/07/13 13:29	1

Client Sample Results

Client: New York State D.E.C.
 Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-103I

Lab Sample ID: 480-47249-5

Date Collected: 10/03/13 11:35

Matrix: Water

Date Received: 10/04/13 09:50

Method: 8260C - Volatile Organic Compounds by GC/MS

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/16/13 15:53	1
1,1,1,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/16/13 15:53	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/16/13 15:53	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/16/13 15:53	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/16/13 15:53	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/16/13 15:53	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/16/13 15:53	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/16/13 15:53	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/16/13 15:53	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/16/13 15:53	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/16/13 15:53	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/16/13 15:53	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/16/13 15:53	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/16/13 15:53	1
2-Hexanone	ND		5.0	1.2	ug/L			10/16/13 15:53	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/16/13 15:53	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/16/13 15:53	1
Acetone	3.4	J	10	3.0	ug/L			10/16/13 15:53	1
Benzene	ND		1.0	0.41	ug/L			10/16/13 15:53	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/16/13 15:53	1
Bromoform	ND		1.0	0.26	ug/L			10/16/13 15:53	1
Bromomethane	ND		1.0	0.69	ug/L			10/16/13 15:53	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/16/13 15:53	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/16/13 15:53	1
Chlorobenzene	ND		1.0	0.75	ug/L			10/16/13 15:53	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/16/13 15:53	1
Chloroethane	ND		1.0	0.32	ug/L			10/16/13 15:53	1
Chloroform	ND		1.0	0.34	ug/L			10/16/13 15:53	1
Chloromethane	ND		1.0	0.35	ug/L			10/16/13 15:53	1
cis-1,2-Dichloroethene	7.6		1.0	0.81	ug/L			10/16/13 15:53	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/16/13 15:53	1
Cyclohexane	ND		1.0	0.18	ug/L			10/16/13 15:53	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/16/13 15:53	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/16/13 15:53	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/16/13 15:53	1
Methyl acetate	ND		1.0	0.50	ug/L			10/16/13 15:53	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/16/13 15:53	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/16/13 15:53	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/16/13 15:53	1
Styrene	ND		1.0	0.73	ug/L			10/16/13 15:53	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/16/13 15:53	1
Toluene	ND		1.0	0.51	ug/L			10/16/13 15:53	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/16/13 15:53	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/16/13 15:53	1
Trichloroethene	5.2		1.0	0.46	ug/L			10/16/13 15:53	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/16/13 15:53	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/16/13 15:53	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/16/13 15:53	1

TestAmerica Buffalo

Client Sample Results

Client: New York State D.E.C.
 Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-103I

Lab Sample ID: 480-47249-5

Date Collected: 10/03/13 11:35

Matrix: Water

Date Received: 10/04/13 09:50

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	103		66 - 137		10/16/13 15:53	1
Toluene-d8 (Surr)	102		71 - 126		10/16/13 15:53	1
4-Bromofluorobenzene (Surr)	99		73 - 120		10/16/13 15:53	1

Method: 6010C - Metals (ICP)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Aluminum	4.0		0.20	0.060	mg/L		10/07/13 08:25	10/07/13 23:58	1
Antimony	ND		0.020	0.0068	mg/L		10/07/13 08:25	10/07/13 23:58	1
Arsenic	ND		0.010	0.0056	mg/L		10/07/13 08:25	10/09/13 12:30	1
Barium	0.089		0.0020	0.00070	mg/L		10/07/13 08:25	10/07/13 23:58	1
Beryllium	ND		0.0020	0.00030	mg/L		10/07/13 08:25	10/07/13 23:58	1
Cadmium	0.0043		0.0010	0.00050	mg/L		10/07/13 08:25	10/07/13 23:58	1
Calcium	18.6		0.50	0.10	mg/L		10/07/13 08:25	10/07/13 23:58	1
Chromium	0.0039	J	0.0040	0.0010	mg/L		10/07/13 08:25	10/07/13 23:58	1
Cobalt	0.0020	J	0.0040	0.00063	mg/L		10/07/13 08:25	10/07/13 23:58	1
Copper	0.0054	J	0.010	0.0016	mg/L		10/07/13 08:25	10/07/13 23:58	1
Iron	4.7		0.050	0.019	mg/L		10/07/13 08:25	10/07/13 23:58	1
Lead	ND		0.0050	0.0030	mg/L		10/07/13 08:25	10/07/13 23:58	1
Magnesium	6.4		0.20	0.043	mg/L		10/07/13 08:25	10/07/13 23:58	1
Manganese	0.17		0.0030	0.00040	mg/L		10/07/13 08:25	10/07/13 23:58	1
Nickel	0.0042	J	0.010	0.0013	mg/L		10/07/13 08:25	10/07/13 23:58	1
Potassium	3.1		0.50	0.10	mg/L		10/07/13 08:25	10/07/13 23:58	1
Selenium	ND		0.015	0.0087	mg/L		10/07/13 08:25	10/07/13 23:58	1
Silver	ND		0.0030	0.0017	mg/L		10/07/13 08:25	10/07/13 23:58	1
Sodium	4.0		1.0	0.32	mg/L		10/07/13 08:25	10/07/13 23:58	1
Thallium	ND		0.020	0.010	mg/L		10/07/13 08:25	10/07/13 23:58	1
Vanadium	0.0051		0.0050	0.0015	mg/L		10/07/13 08:25	10/07/13 23:58	1
Zinc	0.010		0.010	0.0015	mg/L		10/07/13 08:25	10/07/13 23:58	1

Method: 7470A - Mercury (CVAA)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Mercury	ND		0.00020	0.00012	mg/L		10/07/13 09:30	10/07/13 13:31	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-103D

Lab Sample ID: 480-47249-6

Date Collected: 10/03/13 11:40

Matrix: Water

Date Received: 10/04/13 09:50

Method: 8260C - Volatile Organic Compounds by GC/MS

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/16/13 16:20	1
1,1,2,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/16/13 16:20	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/16/13 16:20	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/16/13 16:20	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/16/13 16:20	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/16/13 16:20	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/16/13 16:20	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/16/13 16:20	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/16/13 16:20	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/16/13 16:20	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/16/13 16:20	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/16/13 16:20	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/16/13 16:20	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/16/13 16:20	1
2-Hexanone	ND		5.0	1.2	ug/L			10/16/13 16:20	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/16/13 16:20	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/16/13 16:20	1
Acetone	ND		10	3.0	ug/L			10/16/13 16:20	1
Benzene	ND		1.0	0.41	ug/L			10/16/13 16:20	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/16/13 16:20	1
Bromoform	ND		1.0	0.26	ug/L			10/16/13 16:20	1
Bromomethane	ND		1.0	0.69	ug/L			10/16/13 16:20	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/16/13 16:20	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/16/13 16:20	1
Chlorobenzene	ND		1.0	0.75	ug/L			10/16/13 16:20	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/16/13 16:20	1
Chloroethane	ND		1.0	0.32	ug/L			10/16/13 16:20	1
Chloroform	ND		1.0	0.34	ug/L			10/16/13 16:20	1
Chloromethane	ND		1.0	0.35	ug/L			10/16/13 16:20	1
cis-1,2-Dichloroethene	5.3		1.0	0.81	ug/L			10/16/13 16:20	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/16/13 16:20	1
Cyclohexane	ND		1.0	0.18	ug/L			10/16/13 16:20	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/16/13 16:20	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/16/13 16:20	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/16/13 16:20	1
Methyl acetate	ND		1.0	0.50	ug/L			10/16/13 16:20	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/16/13 16:20	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/16/13 16:20	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/16/13 16:20	1
Styrene	ND		1.0	0.73	ug/L			10/16/13 16:20	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/16/13 16:20	1
Toluene	ND		1.0	0.51	ug/L			10/16/13 16:20	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/16/13 16:20	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/16/13 16:20	1
Trichloroethene	3.5		1.0	0.46	ug/L			10/16/13 16:20	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/16/13 16:20	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/16/13 16:20	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/16/13 16:20	1

TestAmerica Buffalo

Client Sample Results

Client: New York State D.E.C.
 Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-103D

Lab Sample ID: 480-47249-6

Date Collected: 10/03/13 11:40

Matrix: Water

Date Received: 10/04/13 09:50

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	100		66 - 137		10/16/13 16:20	1
Toluene-d8 (Surr)	100		71 - 126		10/16/13 16:20	1
4-Bromofluorobenzene (Surr)	97		73 - 120		10/16/13 16:20	1

Method: 6010C - Metals (ICP)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Aluminum	1.4		0.20	0.060	mg/L		10/07/13 08:25	10/08/13 00:00	1
Antimony	ND		0.020	0.0068	mg/L		10/07/13 08:25	10/08/13 00:00	1
Arsenic	ND		0.010	0.0056	mg/L		10/07/13 08:25	10/09/13 12:32	1
Barium	0.065		0.0020	0.00070	mg/L		10/07/13 08:25	10/08/13 00:00	1
Beryllium	ND		0.0020	0.00030	mg/L		10/07/13 08:25	10/08/13 00:00	1
Cadmium	0.010		0.0010	0.00050	mg/L		10/07/13 08:25	10/08/13 00:00	1
Calcium	21.1		0.50	0.10	mg/L		10/07/13 08:25	10/08/13 00:00	1
Chromium	0.0016	J	0.0040	0.0010	mg/L		10/07/13 08:25	10/08/13 00:00	1
Cobalt	ND		0.0040	0.00063	mg/L		10/07/13 08:25	10/08/13 00:00	1
Copper	0.0052	J	0.010	0.0016	mg/L		10/07/13 08:25	10/08/13 00:00	1
Iron	1.3		0.050	0.019	mg/L		10/07/13 08:25	10/08/13 00:00	1
Lead	ND		0.0050	0.0030	mg/L		10/07/13 08:25	10/08/13 00:00	1
Magnesium	7.8		0.20	0.043	mg/L		10/07/13 08:25	10/08/13 00:00	1
Manganese	0.054		0.0030	0.00040	mg/L		10/07/13 08:25	10/08/13 00:00	1
Nickel	0.0014	J	0.010	0.0013	mg/L		10/07/13 08:25	10/08/13 00:00	1
Potassium	2.2		0.50	0.10	mg/L		10/07/13 08:25	10/08/13 00:00	1
Selenium	ND		0.015	0.0087	mg/L		10/07/13 08:25	10/08/13 00:00	1
Silver	ND		0.0030	0.0017	mg/L		10/07/13 08:25	10/08/13 00:00	1
Sodium	3.8		1.0	0.32	mg/L		10/07/13 08:25	10/08/13 00:00	1
Thallium	ND		0.020	0.010	mg/L		10/07/13 08:25	10/08/13 00:00	1
Vanadium	0.0019	J	0.0050	0.0015	mg/L		10/07/13 08:25	10/08/13 00:00	1
Zinc	0.0056	J	0.010	0.0015	mg/L		10/07/13 08:25	10/08/13 00:00	1

Method: 7470A - Mercury (CVAA)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Mercury	ND		0.00020	0.00012	mg/L		10/07/13 09:30	10/07/13 13:33	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-104I

Lab Sample ID: 480-47249-7

Date Collected: 10/03/13 11:45

Matrix: Water

Date Received: 10/04/13 09:50

Method: 8260C - Volatile Organic Compounds by GC/MS

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/16/13 16:48	1
1,1,2,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/16/13 16:48	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/16/13 16:48	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/16/13 16:48	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/16/13 16:48	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/16/13 16:48	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/16/13 16:48	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/16/13 16:48	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/16/13 16:48	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/16/13 16:48	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/16/13 16:48	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/16/13 16:48	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/16/13 16:48	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/16/13 16:48	1
2-Hexanone	ND		5.0	1.2	ug/L			10/16/13 16:48	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/16/13 16:48	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/16/13 16:48	1
Acetone	16		10	3.0	ug/L			10/16/13 16:48	1
Benzene	ND		1.0	0.41	ug/L			10/16/13 16:48	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/16/13 16:48	1
Bromoform	ND		1.0	0.26	ug/L			10/16/13 16:48	1
Bromomethane	ND		1.0	0.69	ug/L			10/16/13 16:48	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/16/13 16:48	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/16/13 16:48	1
Chlorobenzene	ND		1.0	0.75	ug/L			10/16/13 16:48	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/16/13 16:48	1
Chloroethane	ND		1.0	0.32	ug/L			10/16/13 16:48	1
Chloroform	0.61	J	1.0	0.34	ug/L			10/16/13 16:48	1
Chloromethane	ND		1.0	0.35	ug/L			10/16/13 16:48	1
cis-1,2-Dichloroethene	6.3		1.0	0.81	ug/L			10/16/13 16:48	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/16/13 16:48	1
Cyclohexane	ND		1.0	0.18	ug/L			10/16/13 16:48	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/16/13 16:48	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/16/13 16:48	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/16/13 16:48	1
Methyl acetate	ND		1.0	0.50	ug/L			10/16/13 16:48	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/16/13 16:48	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/16/13 16:48	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/16/13 16:48	1
Styrene	ND		1.0	0.73	ug/L			10/16/13 16:48	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/16/13 16:48	1
Toluene	ND		1.0	0.51	ug/L			10/16/13 16:48	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/16/13 16:48	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/16/13 16:48	1
Trichloroethene	4.0		1.0	0.46	ug/L			10/16/13 16:48	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/16/13 16:48	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/16/13 16:48	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/16/13 16:48	1

TestAmerica Buffalo

Client Sample Results

Client: New York State D.E.C.
 Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-104I

Lab Sample ID: 480-47249-7

Date Collected: 10/03/13 11:45

Matrix: Water

Date Received: 10/04/13 09:50

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	99		66 - 137		10/16/13 16:48	1
Toluene-d8 (Surr)	101		71 - 126		10/16/13 16:48	1
4-Bromofluorobenzene (Surr)	98		73 - 120		10/16/13 16:48	1

Method: 6010C - Metals (ICP)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Aluminum	1.3		0.20	0.060	mg/L		10/07/13 08:25	10/08/13 00:02	1
Antimony	ND		0.020	0.0068	mg/L		10/07/13 08:25	10/08/13 00:02	1
Arsenic	ND		0.010	0.0056	mg/L		10/07/13 08:25	10/09/13 12:35	1
Barium	0.077		0.0020	0.00070	mg/L		10/07/13 08:25	10/08/13 00:02	1
Beryllium	ND		0.0020	0.00030	mg/L		10/07/13 08:25	10/08/13 00:02	1
Cadmium	0.0025		0.0010	0.00050	mg/L		10/07/13 08:25	10/08/13 00:02	1
Calcium	24.4		0.50	0.10	mg/L		10/07/13 08:25	10/08/13 00:02	1
Chromium	0.0017	J	0.0040	0.0010	mg/L		10/07/13 08:25	10/08/13 00:02	1
Cobalt	ND		0.0040	0.00063	mg/L		10/07/13 08:25	10/08/13 00:02	1
Copper	0.0042	J	0.010	0.0016	mg/L		10/07/13 08:25	10/08/13 00:02	1
Iron	1.6		0.050	0.019	mg/L		10/07/13 08:25	10/08/13 00:02	1
Lead	ND		0.0050	0.0030	mg/L		10/07/13 08:25	10/08/13 00:02	1
Magnesium	6.1		0.20	0.043	mg/L		10/07/13 08:25	10/08/13 00:02	1
Manganese	0.025		0.0030	0.00040	mg/L		10/07/13 08:25	10/08/13 00:02	1
Nickel	0.0021	J	0.010	0.0013	mg/L		10/07/13 08:25	10/08/13 00:02	1
Potassium	1.9		0.50	0.10	mg/L		10/07/13 08:25	10/08/13 00:02	1
Selenium	ND		0.015	0.0087	mg/L		10/07/13 08:25	10/08/13 00:02	1
Silver	ND		0.0030	0.0017	mg/L		10/07/13 08:25	10/08/13 00:02	1
Sodium	7.2		1.0	0.32	mg/L		10/07/13 08:25	10/08/13 00:02	1
Thallium	ND		0.020	0.010	mg/L		10/07/13 08:25	10/08/13 00:02	1
Vanadium	0.0017	J	0.0050	0.0015	mg/L		10/07/13 08:25	10/08/13 00:02	1
Zinc	0.0055	J	0.010	0.0015	mg/L		10/07/13 08:25	10/08/13 00:02	1

Method: 7470A - Mercury (CVAA)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Mercury	ND		0.00020	0.00012	mg/L		10/07/13 09:30	10/07/13 13:34	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-104D

Lab Sample ID: 480-47249-8

Date Collected: 10/03/13 11:50

Matrix: Water

Date Received: 10/04/13 09:50

Method: 8260C - Volatile Organic Compounds by GC/MS

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/16/13 17:16	1
1,1,2,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/16/13 17:16	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/16/13 17:16	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/16/13 17:16	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/16/13 17:16	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/16/13 17:16	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/16/13 17:16	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/16/13 17:16	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/16/13 17:16	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/16/13 17:16	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/16/13 17:16	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/16/13 17:16	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/16/13 17:16	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/16/13 17:16	1
2-Hexanone	ND		5.0	1.2	ug/L			10/16/13 17:16	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/16/13 17:16	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/16/13 17:16	1
Acetone	ND		10	3.0	ug/L			10/16/13 17:16	1
Benzene	ND		1.0	0.41	ug/L			10/16/13 17:16	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/16/13 17:16	1
Bromoform	ND		1.0	0.26	ug/L			10/16/13 17:16	1
Bromomethane	ND		1.0	0.69	ug/L			10/16/13 17:16	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/16/13 17:16	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/16/13 17:16	1
Chlorobenzene	ND		1.0	0.75	ug/L			10/16/13 17:16	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/16/13 17:16	1
Chloroethane	ND		1.0	0.32	ug/L			10/16/13 17:16	1
Chloroform	ND		1.0	0.34	ug/L			10/16/13 17:16	1
Chloromethane	ND		1.0	0.35	ug/L			10/16/13 17:16	1
cis-1,2-Dichloroethene	ND		1.0	0.81	ug/L			10/16/13 17:16	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/16/13 17:16	1
Cyclohexane	ND		1.0	0.18	ug/L			10/16/13 17:16	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/16/13 17:16	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/16/13 17:16	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/16/13 17:16	1
Methyl acetate	ND		1.0	0.50	ug/L			10/16/13 17:16	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/16/13 17:16	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/16/13 17:16	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/16/13 17:16	1
Styrene	ND		1.0	0.73	ug/L			10/16/13 17:16	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/16/13 17:16	1
Toluene	ND		1.0	0.51	ug/L			10/16/13 17:16	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/16/13 17:16	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/16/13 17:16	1
Trichloroethene	ND		1.0	0.46	ug/L			10/16/13 17:16	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/16/13 17:16	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/16/13 17:16	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/16/13 17:16	1

TestAmerica Buffalo

Client Sample Results

Client: New York State D.E.C.
 Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-104D

Lab Sample ID: 480-47249-8

Date Collected: 10/03/13 11:50

Matrix: Water

Date Received: 10/04/13 09:50

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	103		66 - 137		10/16/13 17:16	1
Toluene-d8 (Surr)	100		71 - 126		10/16/13 17:16	1
4-Bromofluorobenzene (Surr)	97		73 - 120		10/16/13 17:16	1

Method: 6010C - Metals (ICP)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Aluminum	10.0		0.20	0.060	mg/L		10/07/13 08:25	10/08/13 00:12	1
Antimony	ND		0.020	0.0068	mg/L		10/07/13 08:25	10/08/13 00:12	1
Arsenic	0.018		0.010	0.0056	mg/L		10/07/13 08:25	10/09/13 12:37	1
Barium	0.11		0.0020	0.00070	mg/L		10/07/13 08:25	10/08/13 00:12	1
Beryllium	0.00049	J	0.0020	0.00030	mg/L		10/07/13 08:25	10/08/13 00:12	1
Cadmium	0.0092		0.0010	0.00050	mg/L		10/07/13 08:25	10/08/13 00:12	1
Calcium	48.8		0.50	0.10	mg/L		10/07/13 08:25	10/08/13 00:12	1
Chromium	0.011		0.0040	0.0010	mg/L		10/07/13 08:25	10/08/13 00:12	1
Cobalt	0.0063		0.0040	0.00063	mg/L		10/07/13 08:25	10/08/13 00:12	1
Copper	0.015		0.010	0.0016	mg/L		10/07/13 08:25	10/08/13 00:12	1
Iron	13.9		0.050	0.019	mg/L		10/07/13 08:25	10/08/13 00:12	1
Lead	0.0062		0.0050	0.0030	mg/L		10/07/13 08:25	10/08/13 00:12	1
Magnesium	25.0		0.20	0.043	mg/L		10/07/13 08:25	10/08/13 00:12	1
Manganese	0.20		0.0030	0.00040	mg/L		10/07/13 08:25	10/08/13 00:12	1
Nickel	0.014		0.010	0.0013	mg/L		10/07/13 08:25	10/08/13 00:12	1
Potassium	6.2		0.50	0.10	mg/L		10/07/13 08:25	10/08/13 00:12	1
Selenium	ND		0.015	0.0087	mg/L		10/07/13 08:25	10/08/13 00:12	1
Silver	ND		0.0030	0.0017	mg/L		10/07/13 08:25	10/08/13 00:12	1
Sodium	5.5		1.0	0.32	mg/L		10/07/13 08:25	10/08/13 00:12	1
Thallium	ND		0.020	0.010	mg/L		10/07/13 08:25	10/08/13 00:12	1
Vanadium	0.011		0.0050	0.0015	mg/L		10/07/13 08:25	10/08/13 00:12	1
Zinc	0.035		0.010	0.0015	mg/L		10/07/13 08:25	10/08/13 00:12	1

Method: 7470A - Mercury (CVAA)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Mercury	ND		0.00020	0.00012	mg/L		10/07/13 09:30	10/07/13 13:36	1

Client Sample Results

Client: New York State D.E.C.
 Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-102D

Lab Sample ID: 480-47249-9

Date Collected: 10/03/13 12:40

Matrix: Water

Date Received: 10/04/13 09:50

Method: 8260C - Volatile Organic Compounds by GC/MS

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/16/13 17:43	1
1,1,2,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/16/13 17:43	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/16/13 17:43	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/16/13 17:43	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/16/13 17:43	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/16/13 17:43	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/16/13 17:43	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/16/13 17:43	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/16/13 17:43	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/16/13 17:43	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/16/13 17:43	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/16/13 17:43	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/16/13 17:43	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/16/13 17:43	1
2-Hexanone	ND		5.0	1.2	ug/L			10/16/13 17:43	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/16/13 17:43	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/16/13 17:43	1
Acetone	ND		10	3.0	ug/L			10/16/13 17:43	1
Benzene	ND		1.0	0.41	ug/L			10/16/13 17:43	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/16/13 17:43	1
Bromoform	ND		1.0	0.26	ug/L			10/16/13 17:43	1
Bromomethane	ND		1.0	0.69	ug/L			10/16/13 17:43	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/16/13 17:43	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/16/13 17:43	1
Chlorobenzene	ND		1.0	0.75	ug/L			10/16/13 17:43	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/16/13 17:43	1
Chloroethane	ND		1.0	0.32	ug/L			10/16/13 17:43	1
Chloroform	ND		1.0	0.34	ug/L			10/16/13 17:43	1
Chloromethane	ND		1.0	0.35	ug/L			10/16/13 17:43	1
cis-1,2-Dichloroethene	ND		1.0	0.81	ug/L			10/16/13 17:43	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/16/13 17:43	1
Cyclohexane	ND		1.0	0.18	ug/L			10/16/13 17:43	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/16/13 17:43	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/16/13 17:43	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/16/13 17:43	1
Methyl acetate	ND		1.0	0.50	ug/L			10/16/13 17:43	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/16/13 17:43	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/16/13 17:43	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/16/13 17:43	1
Styrene	ND		1.0	0.73	ug/L			10/16/13 17:43	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/16/13 17:43	1
Toluene	ND		1.0	0.51	ug/L			10/16/13 17:43	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/16/13 17:43	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/16/13 17:43	1
Trichloroethene	ND		1.0	0.46	ug/L			10/16/13 17:43	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/16/13 17:43	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/16/13 17:43	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/16/13 17:43	1

TestAmerica Buffalo

Client Sample Results

Client: New York State D.E.C.
 Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-102D

Lab Sample ID: 480-47249-9

Date Collected: 10/03/13 12:40

Matrix: Water

Date Received: 10/04/13 09:50

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	101		66 - 137		10/16/13 17:43	1
Toluene-d8 (Surr)	100		71 - 126		10/16/13 17:43	1
4-Bromofluorobenzene (Surr)	98		73 - 120		10/16/13 17:43	1

Method: 6010C - Metals (ICP)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Aluminum	0.16	J	0.20	0.060	mg/L		10/07/13 08:25	10/08/13 00:14	1
Antimony	ND		0.020	0.0068	mg/L		10/07/13 08:25	10/08/13 00:14	1
Arsenic	ND		0.010	0.0056	mg/L		10/07/13 08:25	10/09/13 12:47	1
Barium	0.094		0.0020	0.00070	mg/L		10/07/13 08:25	10/08/13 00:14	1
Beryllium	ND		0.0020	0.00030	mg/L		10/07/13 08:25	10/08/13 00:14	1
Cadmium	0.022		0.0010	0.00050	mg/L		10/07/13 08:25	10/08/13 00:14	1
Calcium	49.6		0.50	0.10	mg/L		10/07/13 08:25	10/08/13 00:14	1
Chromium	ND		0.0040	0.0010	mg/L		10/07/13 08:25	10/08/13 00:14	1
Cobalt	ND		0.0040	0.00063	mg/L		10/07/13 08:25	10/08/13 00:14	1
Copper	0.0065	J	0.010	0.0016	mg/L		10/07/13 08:25	10/08/13 00:14	1
Iron	0.16		0.050	0.019	mg/L		10/07/13 08:25	10/08/13 00:14	1
Lead	ND		0.0050	0.0030	mg/L		10/07/13 08:25	10/08/13 00:14	1
Magnesium	15.7		0.20	0.043	mg/L		10/07/13 08:25	10/08/13 00:14	1
Manganese	0.0034		0.0030	0.00040	mg/L		10/07/13 08:25	10/08/13 00:14	1
Nickel	ND		0.010	0.0013	mg/L		10/07/13 08:25	10/08/13 00:14	1
Potassium	2.4		0.50	0.10	mg/L		10/07/13 08:25	10/08/13 00:14	1
Selenium	ND		0.015	0.0087	mg/L		10/07/13 08:25	10/08/13 00:14	1
Silver	ND		0.0030	0.0017	mg/L		10/07/13 08:25	10/08/13 00:14	1
Sodium	4.1		1.0	0.32	mg/L		10/07/13 08:25	10/08/13 00:14	1
Thallium	ND		0.020	0.010	mg/L		10/07/13 08:25	10/08/13 00:14	1
Vanadium	ND		0.0050	0.0015	mg/L		10/07/13 08:25	10/08/13 00:14	1
Zinc	0.0060	J	0.010	0.0015	mg/L		10/07/13 08:25	10/08/13 00:14	1

Method: 7470A - Mercury (CVAA)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Mercury	ND		0.00020	0.00012	mg/L		10/07/13 09:30	10/07/13 13:38	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-109

Lab Sample ID: 480-47249-10

Date Collected: 10/03/13 13:05

Matrix: Water

Date Received: 10/04/13 09:50

Method: 8260C - Volatile Organic Compounds by GC/MS

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/16/13 18:11	1
1,1,1,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/16/13 18:11	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/16/13 18:11	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/16/13 18:11	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/16/13 18:11	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/16/13 18:11	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/16/13 18:11	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/16/13 18:11	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/16/13 18:11	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/16/13 18:11	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/16/13 18:11	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/16/13 18:11	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/16/13 18:11	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/16/13 18:11	1
2-Hexanone	ND		5.0	1.2	ug/L			10/16/13 18:11	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/16/13 18:11	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/16/13 18:11	1
Acetone	3.5	J	10	3.0	ug/L			10/16/13 18:11	1
Benzene	ND		1.0	0.41	ug/L			10/16/13 18:11	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/16/13 18:11	1
Bromoform	ND		1.0	0.26	ug/L			10/16/13 18:11	1
Bromomethane	ND		1.0	0.69	ug/L			10/16/13 18:11	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/16/13 18:11	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/16/13 18:11	1
Chlorobenzene	ND		1.0	0.75	ug/L			10/16/13 18:11	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/16/13 18:11	1
Chloroethane	ND		1.0	0.32	ug/L			10/16/13 18:11	1
Chloroform	ND		1.0	0.34	ug/L			10/16/13 18:11	1
Chloromethane	ND		1.0	0.35	ug/L			10/16/13 18:11	1
cis-1,2-Dichloroethene	ND		1.0	0.81	ug/L			10/16/13 18:11	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/16/13 18:11	1
Cyclohexane	ND		1.0	0.18	ug/L			10/16/13 18:11	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/16/13 18:11	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/16/13 18:11	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/16/13 18:11	1
Methyl acetate	ND		1.0	0.50	ug/L			10/16/13 18:11	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/16/13 18:11	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/16/13 18:11	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/16/13 18:11	1
Styrene	ND		1.0	0.73	ug/L			10/16/13 18:11	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/16/13 18:11	1
Toluene	ND		1.0	0.51	ug/L			10/16/13 18:11	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/16/13 18:11	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/16/13 18:11	1
Trichloroethene	ND		1.0	0.46	ug/L			10/16/13 18:11	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/16/13 18:11	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/16/13 18:11	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/16/13 18:11	1

TestAmerica Buffalo

Client Sample Results

Client: New York State D.E.C.
 Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-109

Lab Sample ID: 480-47249-10

Date Collected: 10/03/13 13:05

Matrix: Water

Date Received: 10/04/13 09:50

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	102		66 - 137		10/16/13 18:11	1
Toluene-d8 (Surr)	100		71 - 126		10/16/13 18:11	1
4-Bromofluorobenzene (Surr)	96		73 - 120		10/16/13 18:11	1

Method: 6010C - Metals (ICP)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Aluminum	0.16	J	0.20	0.060	mg/L		10/07/13 08:25	10/08/13 00:26	1
Antimony	ND		0.020	0.0068	mg/L		10/07/13 08:25	10/08/13 00:26	1
Arsenic	ND		0.010	0.0056	mg/L		10/07/13 08:25	10/09/13 12:58	1
Barium	0.13		0.0020	0.00070	mg/L		10/07/13 08:25	10/08/13 00:26	1
Beryllium	ND		0.0020	0.00030	mg/L		10/07/13 08:25	10/08/13 00:26	1
Cadmium	0.028		0.0010	0.00050	mg/L		10/07/13 08:25	10/08/13 00:26	1
Calcium	24.7		0.50	0.10	mg/L		10/07/13 08:25	10/08/13 00:26	1
Chromium	0.0030	J	0.0040	0.0010	mg/L		10/07/13 08:25	10/08/13 00:26	1
Cobalt	0.011		0.0040	0.00063	mg/L		10/07/13 08:25	10/08/13 00:26	1
Copper	0.0071	J	0.010	0.0016	mg/L		10/07/13 08:25	10/08/13 00:26	1
Iron	2.3		0.050	0.019	mg/L		10/07/13 08:25	10/08/13 00:26	1
Lead	ND		0.0050	0.0030	mg/L		10/07/13 08:25	10/08/13 00:26	1
Magnesium	9.0		0.20	0.043	mg/L		10/07/13 08:25	10/08/13 00:26	1
Manganese	0.12		0.0030	0.00040	mg/L		10/07/13 08:25	10/08/13 00:26	1
Nickel	0.0026	J	0.010	0.0013	mg/L		10/07/13 08:25	10/08/13 00:26	1
Potassium	2.8		0.50	0.10	mg/L		10/07/13 08:25	10/08/13 00:26	1
Selenium	ND		0.015	0.0087	mg/L		10/07/13 08:25	10/08/13 00:26	1
Silver	ND		0.0030	0.0017	mg/L		10/07/13 08:25	10/08/13 00:26	1
Sodium	4.5		1.0	0.32	mg/L		10/07/13 08:25	10/08/13 00:26	1
Thallium	ND		0.020	0.010	mg/L		10/07/13 08:25	10/08/13 00:26	1
Vanadium	ND		0.0050	0.0015	mg/L		10/07/13 08:25	10/08/13 00:26	1
Zinc	0.0065	J	0.010	0.0015	mg/L		10/07/13 08:25	10/08/13 00:26	1

Method: 7470A - Mercury (CVAA)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Mercury	ND		0.00020	0.00012	mg/L		10/07/13 09:30	10/07/13 13:45	1

Client Sample Results

Client: New York State D.E.C.
 Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-108I

Lab Sample ID: 480-47249-11

Date Collected: 10/03/13 13:15

Matrix: Water

Date Received: 10/04/13 09:50

Method: 8260C - Volatile Organic Compounds by GC/MS

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/16/13 18:39	1
1,1,1,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/16/13 18:39	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/16/13 18:39	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/16/13 18:39	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/16/13 18:39	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/16/13 18:39	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/16/13 18:39	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/16/13 18:39	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/16/13 18:39	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/16/13 18:39	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/16/13 18:39	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/16/13 18:39	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/16/13 18:39	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/16/13 18:39	1
2-Hexanone	ND		5.0	1.2	ug/L			10/16/13 18:39	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/16/13 18:39	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/16/13 18:39	1
Acetone	ND		10	3.0	ug/L			10/16/13 18:39	1
Benzene	ND		1.0	0.41	ug/L			10/16/13 18:39	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/16/13 18:39	1
Bromoform	ND		1.0	0.26	ug/L			10/16/13 18:39	1
Bromomethane	ND		1.0	0.69	ug/L			10/16/13 18:39	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/16/13 18:39	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/16/13 18:39	1
Chlorobenzene	ND		1.0	0.75	ug/L			10/16/13 18:39	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/16/13 18:39	1
Chloroethane	ND		1.0	0.32	ug/L			10/16/13 18:39	1
Chloroform	ND		1.0	0.34	ug/L			10/16/13 18:39	1
Chloromethane	ND		1.0	0.35	ug/L			10/16/13 18:39	1
cis-1,2-Dichloroethene	ND		1.0	0.81	ug/L			10/16/13 18:39	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/16/13 18:39	1
Cyclohexane	ND		1.0	0.18	ug/L			10/16/13 18:39	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/16/13 18:39	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/16/13 18:39	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/16/13 18:39	1
Methyl acetate	ND		1.0	0.50	ug/L			10/16/13 18:39	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/16/13 18:39	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/16/13 18:39	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/16/13 18:39	1
Styrene	ND		1.0	0.73	ug/L			10/16/13 18:39	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/16/13 18:39	1
Toluene	ND		1.0	0.51	ug/L			10/16/13 18:39	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/16/13 18:39	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/16/13 18:39	1
Trichloroethene	ND		1.0	0.46	ug/L			10/16/13 18:39	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/16/13 18:39	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/16/13 18:39	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/16/13 18:39	1

TestAmerica Buffalo

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-108I

Lab Sample ID: 480-47249-11

Date Collected: 10/03/13 13:15

Matrix: Water

Date Received: 10/04/13 09:50

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	101		66 - 137		10/16/13 18:39	1
Toluene-d8 (Surr)	99		71 - 126		10/16/13 18:39	1
4-Bromofluorobenzene (Surr)	96		73 - 120		10/16/13 18:39	1

Method: 6010C - Metals (ICP)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Aluminum	0.67		0.20	0.060	mg/L		10/07/13 08:25	10/08/13 00:28	1
Antimony	ND		0.020	0.0068	mg/L		10/07/13 08:25	10/08/13 00:28	1
Arsenic	ND		0.010	0.0056	mg/L		10/07/13 08:25	10/09/13 13:01	1
Barium	0.077		0.0020	0.00070	mg/L		10/07/13 08:25	10/08/13 00:28	1
Beryllium	ND		0.0020	0.00030	mg/L		10/07/13 08:25	10/08/13 00:28	1
Cadmium	ND		0.0010	0.00050	mg/L		10/07/13 08:25	10/08/13 00:28	1
Calcium	52.9		0.50	0.10	mg/L		10/07/13 08:25	10/08/13 00:28	1
Chromium	ND		0.0040	0.0010	mg/L		10/07/13 08:25	10/08/13 00:28	1
Cobalt	ND		0.0040	0.00063	mg/L		10/07/13 08:25	10/08/13 00:28	1
Copper	0.0044	J	0.010	0.0016	mg/L		10/07/13 08:25	10/08/13 00:28	1
Iron	0.70		0.050	0.019	mg/L		10/07/13 08:25	10/08/13 00:28	1
Lead	ND		0.0050	0.0030	mg/L		10/07/13 08:25	10/08/13 00:28	1
Magnesium	29.3		0.20	0.043	mg/L		10/07/13 08:25	10/08/13 00:28	1
Manganese	0.022		0.0030	0.00040	mg/L		10/07/13 08:25	10/08/13 00:28	1
Nickel	ND		0.010	0.0013	mg/L		10/07/13 08:25	10/08/13 00:28	1
Potassium	4.3		0.50	0.10	mg/L		10/07/13 08:25	10/08/13 00:28	1
Selenium	ND		0.015	0.0087	mg/L		10/07/13 08:25	10/08/13 00:28	1
Silver	ND		0.0030	0.0017	mg/L		10/07/13 08:25	10/08/13 00:28	1
Sodium	4.5		1.0	0.32	mg/L		10/07/13 08:25	10/08/13 00:28	1
Thallium	ND		0.020	0.010	mg/L		10/07/13 08:25	10/08/13 00:28	1
Vanadium	ND		0.0050	0.0015	mg/L		10/07/13 08:25	10/08/13 00:28	1
Zinc	0.0026	J	0.010	0.0015	mg/L		10/07/13 08:25	10/08/13 00:28	1

Method: 7470A - Mercury (CVAA)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Mercury	ND		0.00020	0.00012	mg/L		10/07/13 09:30	10/07/13 13:50	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-108D

Lab Sample ID: 480-47249-12

Date Collected: 10/03/13 13:20

Matrix: Water

Date Received: 10/04/13 09:50

Method: 8260C - Volatile Organic Compounds by GC/MS

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/16/13 19:07	1
1,1,2,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/16/13 19:07	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/16/13 19:07	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/16/13 19:07	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/16/13 19:07	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/16/13 19:07	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/16/13 19:07	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/16/13 19:07	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/16/13 19:07	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/16/13 19:07	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/16/13 19:07	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/16/13 19:07	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/16/13 19:07	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/16/13 19:07	1
2-Hexanone	ND		5.0	1.2	ug/L			10/16/13 19:07	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/16/13 19:07	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/16/13 19:07	1
Acetone	4.8	J	10	3.0	ug/L			10/16/13 19:07	1
Benzene	ND		1.0	0.41	ug/L			10/16/13 19:07	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/16/13 19:07	1
Bromoform	ND		1.0	0.26	ug/L			10/16/13 19:07	1
Bromomethane	ND		1.0	0.69	ug/L			10/16/13 19:07	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/16/13 19:07	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/16/13 19:07	1
Chlorobenzene	ND		1.0	0.75	ug/L			10/16/13 19:07	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/16/13 19:07	1
Chloroethane	ND		1.0	0.32	ug/L			10/16/13 19:07	1
Chloroform	ND		1.0	0.34	ug/L			10/16/13 19:07	1
Chloromethane	ND		1.0	0.35	ug/L			10/16/13 19:07	1
cis-1,2-Dichloroethene	ND		1.0	0.81	ug/L			10/16/13 19:07	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/16/13 19:07	1
Cyclohexane	ND		1.0	0.18	ug/L			10/16/13 19:07	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/16/13 19:07	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/16/13 19:07	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/16/13 19:07	1
Methyl acetate	ND		1.0	0.50	ug/L			10/16/13 19:07	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/16/13 19:07	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/16/13 19:07	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/16/13 19:07	1
Styrene	ND		1.0	0.73	ug/L			10/16/13 19:07	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/16/13 19:07	1
Toluene	ND		1.0	0.51	ug/L			10/16/13 19:07	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/16/13 19:07	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/16/13 19:07	1
Trichloroethene	ND		1.0	0.46	ug/L			10/16/13 19:07	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/16/13 19:07	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/16/13 19:07	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/16/13 19:07	1

TestAmerica Buffalo

Client Sample Results

Client: New York State D.E.C.
 Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-108D

Lab Sample ID: 480-47249-12

Date Collected: 10/03/13 13:20

Matrix: Water

Date Received: 10/04/13 09:50

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	103		66 - 137		10/16/13 19:07	1
Toluene-d8 (Surr)	100		71 - 126		10/16/13 19:07	1
4-Bromofluorobenzene (Surr)	94		73 - 120		10/16/13 19:07	1

Method: 6010C - Metals (ICP)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Aluminum	0.75		0.20	0.060	mg/L		10/07/13 08:25	10/08/13 00:31	1
Antimony	ND		0.020	0.0068	mg/L		10/07/13 08:25	10/08/13 00:31	1
Arsenic	ND		0.010	0.0056	mg/L		10/07/13 08:25	10/09/13 13:03	1
Barium	0.12		0.0020	0.00070	mg/L		10/07/13 08:25	10/08/13 00:31	1
Beryllium	ND		0.0020	0.00030	mg/L		10/07/13 08:25	10/08/13 00:31	1
Cadmium	0.0068		0.0010	0.00050	mg/L		10/07/13 08:25	10/08/13 00:31	1
Calcium	29.2		0.50	0.10	mg/L		10/07/13 08:25	10/08/13 00:31	1
Chromium	0.0018	J	0.0040	0.0010	mg/L		10/07/13 08:25	10/08/13 00:31	1
Cobalt	0.00074	J	0.0040	0.00063	mg/L		10/07/13 08:25	10/08/13 00:31	1
Copper	0.0055	J	0.010	0.0016	mg/L		10/07/13 08:25	10/08/13 00:31	1
Iron	2.2		0.050	0.019	mg/L		10/07/13 08:25	10/08/13 00:31	1
Lead	ND		0.0050	0.0030	mg/L		10/07/13 08:25	10/08/13 00:31	1
Magnesium	11.3		0.20	0.043	mg/L		10/07/13 08:25	10/08/13 00:31	1
Manganese	0.24		0.0030	0.00040	mg/L		10/07/13 08:25	10/08/13 00:31	1
Nickel	0.0028	J	0.010	0.0013	mg/L		10/07/13 08:25	10/08/13 00:31	1
Potassium	2.3		0.50	0.10	mg/L		10/07/13 08:25	10/08/13 00:31	1
Selenium	ND		0.015	0.0087	mg/L		10/07/13 08:25	10/08/13 00:31	1
Silver	ND		0.0030	0.0017	mg/L		10/07/13 08:25	10/08/13 00:31	1
Sodium	3.9		1.0	0.32	mg/L		10/07/13 08:25	10/08/13 00:31	1
Thallium	ND		0.020	0.010	mg/L		10/07/13 08:25	10/08/13 00:31	1
Vanadium	ND		0.0050	0.0015	mg/L		10/07/13 08:25	10/08/13 00:31	1
Zinc	0.11		0.010	0.0015	mg/L		10/07/13 08:25	10/08/13 00:31	1

Method: 7470A - Mercury (CVAA)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Mercury	ND		0.00020	0.00012	mg/L		10/07/13 09:30	10/07/13 13:52	1

Client Sample Results

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-113

Lab Sample ID: 480-47249-13

Date Collected: 10/03/13 13:50

Matrix: Water

Date Received: 10/04/13 09:50

Method: 8260C - Volatile Organic Compounds by GC/MS

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1-Trichloroethane	ND		1.0	0.82	ug/L			10/16/13 19:34	1
1,1,2,2-Tetrachloroethane	ND		1.0	0.21	ug/L			10/16/13 19:34	1
1,1,2-Trichloroethane	ND		1.0	0.23	ug/L			10/16/13 19:34	1
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		1.0	0.31	ug/L			10/16/13 19:34	1
1,1-Dichloroethane	ND		1.0	0.38	ug/L			10/16/13 19:34	1
1,1-Dichloroethene	ND		1.0	0.29	ug/L			10/16/13 19:34	1
1,2,4-Trichlorobenzene	ND		1.0	0.41	ug/L			10/16/13 19:34	1
1,2-Dibromo-3-Chloropropane	ND		1.0	0.39	ug/L			10/16/13 19:34	1
1,2-Dibromoethane	ND		1.0	0.73	ug/L			10/16/13 19:34	1
1,2-Dichlorobenzene	ND		1.0	0.79	ug/L			10/16/13 19:34	1
1,2-Dichloroethane	ND		1.0	0.21	ug/L			10/16/13 19:34	1
1,2-Dichloropropane	ND		1.0	0.72	ug/L			10/16/13 19:34	1
1,3-Dichlorobenzene	ND		1.0	0.78	ug/L			10/16/13 19:34	1
1,4-Dichlorobenzene	ND		1.0	0.84	ug/L			10/16/13 19:34	1
2-Hexanone	ND		5.0	1.2	ug/L			10/16/13 19:34	1
2-Butanone (MEK)	ND		10	1.3	ug/L			10/16/13 19:34	1
4-Methyl-2-pentanone (MIBK)	ND		5.0	2.1	ug/L			10/16/13 19:34	1
Acetone	ND		10	3.0	ug/L			10/16/13 19:34	1
Benzene	ND		1.0	0.41	ug/L			10/16/13 19:34	1
Bromodichloromethane	ND		1.0	0.39	ug/L			10/16/13 19:34	1
Bromoform	ND		1.0	0.26	ug/L			10/16/13 19:34	1
Bromomethane	ND		1.0	0.69	ug/L			10/16/13 19:34	1
Carbon disulfide	ND		1.0	0.19	ug/L			10/16/13 19:34	1
Carbon tetrachloride	ND		1.0	0.27	ug/L			10/16/13 19:34	1
Chlorobenzene	ND		1.0	0.75	ug/L			10/16/13 19:34	1
Dibromochloromethane	ND		1.0	0.32	ug/L			10/16/13 19:34	1
Chloroethane	ND		1.0	0.32	ug/L			10/16/13 19:34	1
Chloroform	ND		1.0	0.34	ug/L			10/16/13 19:34	1
Chloromethane	ND		1.0	0.35	ug/L			10/16/13 19:34	1
cis-1,2-Dichloroethene	ND		1.0	0.81	ug/L			10/16/13 19:34	1
cis-1,3-Dichloropropene	ND		1.0	0.36	ug/L			10/16/13 19:34	1
Cyclohexane	ND		1.0	0.18	ug/L			10/16/13 19:34	1
Dichlorodifluoromethane	ND		1.0	0.68	ug/L			10/16/13 19:34	1
Ethylbenzene	ND		1.0	0.74	ug/L			10/16/13 19:34	1
Isopropylbenzene	ND		1.0	0.79	ug/L			10/16/13 19:34	1
Methyl acetate	ND		1.0	0.50	ug/L			10/16/13 19:34	1
Methyl tert-butyl ether	ND		1.0	0.16	ug/L			10/16/13 19:34	1
Methylcyclohexane	ND		1.0	0.16	ug/L			10/16/13 19:34	1
Methylene Chloride	ND		1.0	0.44	ug/L			10/16/13 19:34	1
Styrene	ND		1.0	0.73	ug/L			10/16/13 19:34	1
Tetrachloroethene	ND		1.0	0.36	ug/L			10/16/13 19:34	1
Toluene	ND		1.0	0.51	ug/L			10/16/13 19:34	1
trans-1,2-Dichloroethene	ND		1.0	0.90	ug/L			10/16/13 19:34	1
trans-1,3-Dichloropropene	ND		1.0	0.37	ug/L			10/16/13 19:34	1
Trichloroethene	ND		1.0	0.46	ug/L			10/16/13 19:34	1
Trichlorofluoromethane	ND		1.0	0.88	ug/L			10/16/13 19:34	1
Vinyl chloride	ND		1.0	0.90	ug/L			10/16/13 19:34	1
Xylenes, Total	ND		2.0	0.66	ug/L			10/16/13 19:34	1

TestAmerica Buffalo

Client Sample Results

Client: New York State D.E.C.
 Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-113

Lab Sample ID: 480-47249-13

Date Collected: 10/03/13 13:50

Matrix: Water

Date Received: 10/04/13 09:50

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	101		66 - 137		10/16/13 19:34	1
Toluene-d8 (Surr)	99		71 - 126		10/16/13 19:34	1
4-Bromofluorobenzene (Surr)	95		73 - 120		10/16/13 19:34	1

Method: 6010C - Metals (ICP)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Aluminum	6.2		0.20	0.060	mg/L		10/07/13 08:25	10/08/13 00:40	1
Antimony	ND		0.020	0.0068	mg/L		10/07/13 08:25	10/08/13 00:40	1
Arsenic	0.010		0.010	0.0056	mg/L		10/07/13 08:25	10/09/13 13:05	1
Barium	0.062		0.0020	0.00070	mg/L		10/07/13 08:25	10/08/13 00:40	1
Beryllium	0.00032	J	0.0020	0.00030	mg/L		10/07/13 08:25	10/08/13 00:40	1
Cadmium	0.0021		0.0010	0.00050	mg/L		10/07/13 08:25	10/08/13 00:40	1
Calcium	54.1		0.50	0.10	mg/L		10/07/13 08:25	10/08/13 00:40	1
Chromium	0.0073		0.0040	0.0010	mg/L		10/07/13 08:25	10/08/13 00:40	1
Cobalt	0.0052		0.0040	0.00063	mg/L		10/07/13 08:25	10/08/13 00:40	1
Copper	0.011		0.010	0.0016	mg/L		10/07/13 08:25	10/08/13 00:40	1
Iron	8.8		0.050	0.019	mg/L		10/07/13 08:25	10/08/13 00:40	1
Lead	0.0045	J	0.0050	0.0030	mg/L		10/07/13 08:25	10/08/13 00:40	1
Magnesium	22.3		0.20	0.043	mg/L		10/07/13 08:25	10/08/13 00:40	1
Manganese	0.23		0.0030	0.00040	mg/L		10/07/13 08:25	10/08/13 00:40	1
Nickel	0.0082	J	0.010	0.0013	mg/L		10/07/13 08:25	10/08/13 00:40	1
Potassium	5.1		0.50	0.10	mg/L		10/07/13 08:25	10/08/13 00:40	1
Selenium	ND		0.015	0.0087	mg/L		10/07/13 08:25	10/08/13 00:40	1
Silver	ND		0.0030	0.0017	mg/L		10/07/13 08:25	10/08/13 00:40	1
Sodium	4.3		1.0	0.32	mg/L		10/07/13 08:25	10/08/13 00:40	1
Thallium	ND		0.020	0.010	mg/L		10/07/13 08:25	10/08/13 00:40	1
Vanadium	0.0070		0.0050	0.0015	mg/L		10/07/13 08:25	10/08/13 00:40	1
Zinc	0.019		0.010	0.0015	mg/L		10/07/13 08:25	10/08/13 00:40	1

Method: 7470A - Mercury (CVAA)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Mercury	ND		0.00020	0.00012	mg/L		10/07/13 09:30	10/07/13 13:53	1

Lab Chronicle

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-107S

Lab Sample ID: 480-47249-1

Date Collected: 10/03/13 09:55

Matrix: Water

Date Received: 10/04/13 09:50

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260C		1	145298	10/16/13 14:01	NMD1	TAL BUF
Total/NA	Prep	7470A			143124	10/07/13 09:30	JRK	TAL BUF
Total/NA	Analysis	7470A		1	143242	10/07/13 13:21	JRK	TAL BUF
Total/NA	Analysis	6010C		1	143381	10/07/13 23:48	LMH	TAL BUF
Total/NA	Prep	3005A			143000	10/07/13 08:25	SS1	TAL BUF
Total/NA	Analysis	6010C		1	144523	10/09/13 12:20	LMH	TAL BUF

Client Sample ID: MW-107I

Lab Sample ID: 480-47249-2

Date Collected: 10/03/13 10:00

Matrix: Water

Date Received: 10/04/13 09:50

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260C		1	145298	10/16/13 14:29	NMD1	TAL BUF
Total/NA	Prep	7470A			143124	10/07/13 09:30	JRK	TAL BUF
Total/NA	Analysis	7470A		1	143242	10/07/13 13:23	JRK	TAL BUF
Total/NA	Prep	3005A			143000	10/07/13 08:25	SS1	TAL BUF
Total/NA	Analysis	6010C		1	143381	10/07/13 23:50	LMH	TAL BUF
Total/NA	Analysis	6010C		1	144523	10/09/13 12:22	LMH	TAL BUF

Client Sample ID: MW-101D

Lab Sample ID: 480-47249-3

Date Collected: 10/03/13 11:00

Matrix: Water

Date Received: 10/04/13 09:50

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260C		1	145298	10/16/13 14:57	NMD1	TAL BUF
Total/NA	Prep	7470A			143124	10/07/13 09:30	JRK	TAL BUF
Total/NA	Analysis	7470A		1	143242	10/07/13 13:25	JRK	TAL BUF
Total/NA	Analysis	6010C		1	143381	10/07/13 23:53	LMH	TAL BUF
Total/NA	Prep	3005A			143000	10/07/13 08:25	SS1	TAL BUF
Total/NA	Analysis	6010C		1	144523	10/09/13 12:25	LMH	TAL BUF

Client Sample ID: MW-101I

Lab Sample ID: 480-47249-4

Date Collected: 10/03/13 11:05

Matrix: Water

Date Received: 10/04/13 09:50

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260C		1	145298	10/16/13 15:25	NMD1	TAL BUF
Total/NA	Prep	7470A			143124	10/07/13 09:30	JRK	TAL BUF
Total/NA	Analysis	7470A		1	143242	10/07/13 13:29	JRK	TAL BUF
Total/NA	Prep	3005A			143000	10/07/13 08:25	SS1	TAL BUF
Total/NA	Analysis	6010C		1	143381	10/07/13 23:55	LMH	TAL BUF
Total/NA	Analysis	6010C		1	144523	10/09/13 12:27	LMH	TAL BUF

TestAmerica Buffalo

Lab Chronicle

Client: New York State D.E.C.
 Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-103I

Lab Sample ID: 480-47249-5

Date Collected: 10/03/13 11:35

Matrix: Water

Date Received: 10/04/13 09:50

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260C		1	145298	10/16/13 15:53	NMD1	TAL BUF
Total/NA	Prep	7470A			143124	10/07/13 09:30	JRK	TAL BUF
Total/NA	Analysis	7470A		1	143242	10/07/13 13:31	JRK	TAL BUF
Total/NA	Prep	3005A			143000	10/07/13 08:25	SS1	TAL BUF
Total/NA	Analysis	6010C		1	143381	10/07/13 23:58	LMH	TAL BUF
Total/NA	Analysis	6010C		1	144523	10/09/13 12:30	LMH	TAL BUF

Client Sample ID: MW-103D

Lab Sample ID: 480-47249-6

Date Collected: 10/03/13 11:40

Matrix: Water

Date Received: 10/04/13 09:50

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260C		1	145298	10/16/13 16:20	NMD1	TAL BUF
Total/NA	Prep	7470A			143124	10/07/13 09:30	JRK	TAL BUF
Total/NA	Analysis	7470A		1	143242	10/07/13 13:33	JRK	TAL BUF
Total/NA	Prep	3005A			143000	10/07/13 08:25	SS1	TAL BUF
Total/NA	Analysis	6010C		1	143381	10/08/13 00:00	LMH	TAL BUF
Total/NA	Analysis	6010C		1	144523	10/09/13 12:32	LMH	TAL BUF

Client Sample ID: MW-104I

Lab Sample ID: 480-47249-7

Date Collected: 10/03/13 11:45

Matrix: Water

Date Received: 10/04/13 09:50

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260C		1	145298	10/16/13 16:48	NMD1	TAL BUF
Total/NA	Prep	7470A			143124	10/07/13 09:30	JRK	TAL BUF
Total/NA	Analysis	7470A		1	143242	10/07/13 13:34	JRK	TAL BUF
Total/NA	Prep	3005A			143000	10/07/13 08:25	SS1	TAL BUF
Total/NA	Analysis	6010C		1	143381	10/08/13 00:02	LMH	TAL BUF
Total/NA	Analysis	6010C		1	144523	10/09/13 12:35	LMH	TAL BUF

Client Sample ID: MW-104D

Lab Sample ID: 480-47249-8

Date Collected: 10/03/13 11:50

Matrix: Water

Date Received: 10/04/13 09:50

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260C		1	145298	10/16/13 17:16	NMD1	TAL BUF
Total/NA	Prep	7470A			143124	10/07/13 09:30	JRK	TAL BUF
Total/NA	Analysis	7470A		1	143242	10/07/13 13:36	JRK	TAL BUF
Total/NA	Prep	3005A			143000	10/07/13 08:25	SS1	TAL BUF
Total/NA	Analysis	6010C		1	143381	10/08/13 00:12	LMH	TAL BUF
Total/NA	Analysis	6010C		1	144523	10/09/13 12:37	LMH	TAL BUF

TestAmerica Buffalo

Lab Chronicle

Client: New York State D.E.C.
 Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-102D

Lab Sample ID: 480-47249-9

Date Collected: 10/03/13 12:40

Matrix: Water

Date Received: 10/04/13 09:50

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260C		1	145298	10/16/13 17:43	NMD1	TAL BUF
Total/NA	Prep	7470A			143124	10/07/13 09:30	JRK	TAL BUF
Total/NA	Analysis	7470A		1	143242	10/07/13 13:38	JRK	TAL BUF
Total/NA	Prep	3005A			143000	10/07/13 08:25	SS1	TAL BUF
Total/NA	Analysis	6010C		1	143381	10/08/13 00:14	LMH	TAL BUF
Total/NA	Analysis	6010C		1	144523	10/09/13 12:47	LMH	TAL BUF

Client Sample ID: MW-109

Lab Sample ID: 480-47249-10

Date Collected: 10/03/13 13:05

Matrix: Water

Date Received: 10/04/13 09:50

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260C		1	145298	10/16/13 18:11	NMD1	TAL BUF
Total/NA	Prep	7470A			143124	10/07/13 09:30	JRK	TAL BUF
Total/NA	Analysis	7470A		1	143242	10/07/13 13:45	JRK	TAL BUF
Total/NA	Prep	3005A			143000	10/07/13 08:25	SS1	TAL BUF
Total/NA	Analysis	6010C		1	143381	10/08/13 00:26	LMH	TAL BUF
Total/NA	Analysis	6010C		1	144523	10/09/13 12:58	LMH	TAL BUF

Client Sample ID: MW-108I

Lab Sample ID: 480-47249-11

Date Collected: 10/03/13 13:15

Matrix: Water

Date Received: 10/04/13 09:50

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260C		1	145298	10/16/13 18:39	NMD1	TAL BUF
Total/NA	Prep	7470A			143124	10/07/13 09:30	JRK	TAL BUF
Total/NA	Analysis	7470A		1	143242	10/07/13 13:50	JRK	TAL BUF
Total/NA	Prep	3005A			143000	10/07/13 08:25	SS1	TAL BUF
Total/NA	Analysis	6010C		1	143381	10/08/13 00:28	LMH	TAL BUF
Total/NA	Analysis	6010C		1	144523	10/09/13 13:01	LMH	TAL BUF

Client Sample ID: MW-108D

Lab Sample ID: 480-47249-12

Date Collected: 10/03/13 13:20

Matrix: Water

Date Received: 10/04/13 09:50

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260C		1	145298	10/16/13 19:07	NMD1	TAL BUF
Total/NA	Prep	7470A			143124	10/07/13 09:30	JRK	TAL BUF
Total/NA	Analysis	7470A		1	143242	10/07/13 13:52	JRK	TAL BUF
Total/NA	Prep	3005A			143000	10/07/13 08:25	SS1	TAL BUF
Total/NA	Analysis	6010C		1	143381	10/08/13 00:31	LMH	TAL BUF
Total/NA	Analysis	6010C		1	144523	10/09/13 13:03	LMH	TAL BUF

TestAmerica Buffalo

Lab Chronicle

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Client Sample ID: MW-113

Lab Sample ID: 480-47249-13

Date Collected: 10/03/13 13:50

Matrix: Water

Date Received: 10/04/13 09:50

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260C		1	145298	10/16/13 19:34	NMD1	TAL BUF
Total/NA	Prep	7470A			143124	10/07/13 09:30	JRK	TAL BUF
Total/NA	Analysis	7470A		1	143242	10/07/13 13:53	JRK	TAL BUF
Total/NA	Prep	3005A			143000	10/07/13 08:25	SS1	TAL BUF
Total/NA	Analysis	6010C		1	143381	10/08/13 00:40	LMH	TAL BUF
Total/NA	Analysis	6010C		1	144523	10/09/13 13:05	LMH	TAL BUF

Laboratory References:

TAL BUF = TestAmerica Buffalo, 10 Hazelwood Drive, Amherst, NY 14228-2298, TEL (716)691-2600



Certification Summary

Client: New York State D.E.C.
 Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Laboratory: TestAmerica Buffalo

All certifications held by this laboratory are listed. Not all certifications are applicable to this report.

Authority	Program	EPA Region	Certification ID	Expiration Date
Arkansas DEQ	State Program	6	88-0686	07-06-13 *
California	NELAP	9	1169CA	09-30-14
Connecticut	State Program	1	PH-0568	09-30-14
Florida	NELAP	4	E87672	06-30-14
Georgia	State Program	4	N/A	03-31-14
Iowa	State Program	7	374	03-15-15
Kansas	NELAP	7	E-10187	01-31-14
Kentucky	State Program	4	90029	12-31-13
Kentucky (UST)	State Program	4	30	04-01-14
Louisiana	NELAP	6	02031	06-30-14
Maine	State Program	1	NY00044	12-04-14
Maryland	State Program	3	294	03-31-14
Massachusetts	State Program	1	M-NY044	06-30-14
Michigan	State Program	5	9937	04-01-14
Minnesota	NELAP	5	036-999-337	12-31-13
New Hampshire	NELAP	1	2973	09-11-14
New Jersey	NELAP	2	NY455	06-30-14
New York	NELAP	2	10026	04-01-14
North Dakota	State Program	8	R-176	03-31-14
Oklahoma	State Program	6	9421	08-31-14
Oregon	NELAP	10	NY200003	06-09-14
Pennsylvania	NELAP	3	68-00281	07-31-14
Rhode Island	State Program	1	LAO00328	12-31-13
Tennessee	State Program	4	TN02970	04-01-14
Texas	NELAP	6	T104704412-11-2	07-31-14
USDA	Federal		P330-11-00386	11-22-14
Virginia	NELAP	3	460185	09-14-14
Washington	State Program	10	C784	02-10-14
West Virginia DEP	State Program	3	252	12-31-13
Wisconsin	State Program	5	998310390	08-31-14

* Expired certification is currently pending renewal and is considered valid.



Method Summary

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Method	Method Description	Protocol	Laboratory
8260C	Volatile Organic Compounds by GC/MS	SW846	TAL BUF
6010C	Metals (ICP)	SW846	TAL BUF
7470A	Mercury (CVAA)	SW846	TAL BUF

Protocol References:

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

Laboratory References:

TAL BUF = TestAmerica Buffalo, 10 Hazelwood Drive, Amherst, NY 14228-2298, TEL (716)691-2600



Sample Summary

Client: New York State D.E.C.
Project/Site: NYSDEC - Patton's Busy Bee:site #902014

TestAmerica Job ID: 480-47249-1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
480-47249-1	MW-107S	Water	10/03/13 09:55	10/04/13 09:50
480-47249-2	MW-107I	Water	10/03/13 10:00	10/04/13 09:50
480-47249-3	MW-101D	Water	10/03/13 11:00	10/04/13 09:50
480-47249-4	MW-101I	Water	10/03/13 11:05	10/04/13 09:50
480-47249-5	MW-103I	Water	10/03/13 11:35	10/04/13 09:50
480-47249-6	MW-103D	Water	10/03/13 11:40	10/04/13 09:50
480-47249-7	MW-104I	Water	10/03/13 11:45	10/04/13 09:50
480-47249-8	MW-104D	Water	10/03/13 11:50	10/04/13 09:50
480-47249-9	MW-102D	Water	10/03/13 12:40	10/04/13 09:50
480-47249-10	MW-109	Water	10/03/13 13:05	10/04/13 09:50
480-47249-11	MW-108I	Water	10/03/13 13:15	10/04/13 09:50
480-47249-12	MW-108D	Water	10/03/13 13:20	10/04/13 09:50
480-47249-13	MW-113	Water	10/03/13 13:50	10/04/13 09:50

TestAmerica

THE LEADER IN ENVIRONMENTAL TESTING

Temperature on Receipt _____
 Drinking Water? Yes No

Chain of Custody Record

TAL-4124 (1007)
 Client: **NY State**
 Address: **270 MICHIGAN AVE.**
 City: **ROCHESTER** State: **NY** Zip Code: **146203**
 Project Name and Location (State): **FAYTON'S BUSY BUS**
 Contract/Purchase Order/Quote No.: _____
 Project Manager: **BRIAN JADOWSKI** Date: **10/3/13**
 Telephone Number (Area Code)/Fax Number: **716-851-7220** Lab Number: _____
 Site Contact: **B. FISCHER** Page: **1** of **2**
 Carrier/Waybill Number: _____

Sample I.D. No. and Description (Containers for each sample may be combined on one line)	Date	Time	Matrix			Containers & Preservatives					Analysis (Attach list if more space is needed)	Special Instructions/ Conditions of Receipt		
			Air	Aqueous	Sed	Soil	Unpres.	H2SO4	HNO3	HCl			NaOH	ZnAc/ NaOH
MW-107 S	10/3/13	0955	✓						1	3		✓	VOL TEL BRKDR	
MW-107 I		1000	✓						1	3		✓	✓	
MW-101 D		1100	✓						1	3		✓	✓	
MW-101 I		1105	✓						1	3		✓	✓	
MW-103 I		1135	✓						1	3		✓	✓	LOW VOL. ON METALS
MW-103 D		1140	✓						1	3		✓	✓	
MW-104 I		1145	✓						1	3		✓	✓	
MW-104 D		1150	✓						1	3		✓	✓	
MW-102 D		1240	✓						3	9		✓	✓	ms/msd
MW-109		1305	✓						1	3		✓	✓	
MW-108 I		1315	✓						1	3		✓	✓	
MW-108 D		1320	✓						1	3		✓	✓	

Possible Hazard Identification
 Non-Hazard Flammable Skin Irritant Poison B Unknown Other: **STD.**
 Return To Client Disposal By Lab Archive For _____ Months
 (A fee may be assessed if samples are retained longer than 1 month)

Turn Around Time Required
 24 Hours 48 Hours 7 Days 14 Days 21 Days

1. Relinquished By: **Wayne Fischer** Date: **10/4/13** Time: **0950**
 2. Relinquished By: **WJ** Date: **10/4/13** Time: **0950**
 3. Relinquished By: _____ Date: _____ Time: _____

Comments: **31A**



Login Sample Receipt Checklist

Client: New York State D.E.C.

Job Number: 480-47249-1

Login Number: 47249

List Source: TestAmerica Buffalo

List Number: 1

Creator: Janish, Carl M

Question	Answer	Comment
Radioactivity either was not measured or, if measured, is at or below background	True	
The cooler's custody seal, if present, is intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	True	
There are no discrepancies between the sample IDs on the containers and the COC.	True	
Samples are received within Holding Time.	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
VOA sample vials do not have headspace or bubble is <6mm (1/4") in diameter.	True	
If necessary, staff have been informed of any short hold time or quick TAT needs	True	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Sampling Company provided.	True	dec
Samples received within 48 hours of sampling.	True	
Samples requiring field filtration have been filtered in the field.	N/A	
Chlorine Residual checked.	N/A	

