

**DATA USEABILITY SUMMARY REPORT (DUSR)
COURTLANDT CORNERS I
364 EAST 161ST STREET, BRONX, NEW YORK
ENDPOINT SOIL SAMPLING
ENVIRONMENTAL RESOURCES MANAGEMENT (ERM)
ACCUTEST LABORATORIES JOB NUMBERS
JA22090, JA23207, JA23638, JA23843 JA24059,
JA24443, JA24692, JA25087, JA25246, JA25522, AND JA25638**

Deliverables

The above referenced data packages for thirty-five (35) soil samples, one (1) blind field duplicate sample, one (1) field blank, one (1) trip blank, and one (1) set of matrix spike/matrix spike duplicate (MS/MSD) sample contains all required deliverables as stipulated under the 2005 New York State Analytical Services Protocols (ASP) for Category B deliverables. The sample specific analysis included Target Compound List (TCL) Volatile Organic Compounds (VOC), analyzed in accordance with United States Environmental Protection Agency (USEPA) SW-846 Method 8260B, TCL Semivolatile Organic Compounds (SVOCs) analyzed in accordance with USEPA SW-846 Method 8270C, TCL Pesticides analyzed in accordance with USEPA SW-846 Method 8081A, and Target Analyte List (TAL) Metals analyzed in accordance with USEPA SW-846 Methods 6010B and 7471B. The TCL and TAL are consistent with Exhibit C of the USEPA Contract Laboratory Program Statement of Work (CLP SOW) for organics analysis Multi-Media, Multi-Concentration OLM04.2, May 1999 and inorganics analysis Multi-Media, Multi-Concentration ILM05.3, February 2004.

The data have been validated according to the protocols and quality control (QC) requirements of the following documents and the reviewer's professional judgment:

- the NYSDEC ASP;
- the USEPA CLP National Functional Guidelines for Organic Data Review (October 1999);
- the USEPA Region II Data Review Standard Operating Procedure (SOP) Number HW-24, Revision 2, October 2006: Validating Volatile Organic Compounds by SW-846 Method 8260B;
- the USEPA Region II Data Review SOP Number HW-22, Revision 3, October 2006: Validating Semivolatile Organic Compounds by SW-846 Method 8270C;
- the USEPA Region II Data Review SOP Number HW-44, Revision 1: Data Validation SOP of Organochlorine Pesticides by Gas

- Chromatography SW-846 Method 8081B - October 2006;
- the USEPA CLP National Functional Guidelines for Inorganic Data Review (October 2004); and
- the USEPA Region II Data Review SOP Number HW-2, Revision 13, September 2006: Evaluation of Metals Data for the CLP Program.

This DUSR pertains to the following soil samples collected from June 30, 2009 through August 14, 2009:

Samples

<u>JA22090</u>	<u>JA22307</u>	<u>JA23638</u>		<u>JA23843</u>
B-CCI-02(08)	B-CCI-003(20') B-CCI-004(20')	SW-CCI-001(18.5') SW-CCI-002(18.5') SW-CCI-003(18.5') SW-CCI-004(19') SW-CCI-005(18.5')	SW-CCI-006(19') SW-CCI-007(19') SW-CCI-008(13.5') SW-CCI-009(18.5')	B-CCI-006(15') B-CCI-007(20')
<u>JA24059</u>	<u>JA24443</u>	<u>JA24692</u>	<u>JA25087</u>	<u>JA25246</u>
B-CCI-010(22') B-CCI-013(10')	B-CCI-016(23') B-CCI-017(23') B-CCI-018(26') B-CCI-019(26')	B-CCI-022 B-CCI-023 B-CCI-024 B-CCI-025	B-CCI-026 (2') B-CCI-027 (2')	SW-CCI-011(24.5)
	<u>JA25446</u>	<u>JA25522</u>	<u>JA25638</u>	
	B-CCI-028(2') B-CCI-029(20') B-CCI-030(4') B-CCI-031(10')	SW-CCI-012 SW-CCI-013 SW-CCI-014	SW-CCI-015	

QC Samples

DUPLICATE (blind field duplicate of sample B-CCI-007(20'))
SW-CCI-005(18.5') MS/MSD
Batch MS/MSD
FB072009 (field blank)
TRIP BLANK (shipped on June 30, 2009)

Chain-of-Custody

- The Chains-of-Custody (COC) were reviewed for completeness and accuracy. There were no discrepancies observed with the samples presented on the COC except those listed below. All tests specified on the COC were performed for the designated samples.

- Samples B-CCI-01, B-CCI-005(20'), B-CCI-008(20'), B-CCI-009(20'), B-CCI-011(22'), B-CCI-014(22'), B-CCI-020(02'), and B-CCI-021(08'), collected on June 25, 2009, July 15, 2009, July 22, 2009, July 22, 2009, July 27, 2009, July 27, 2009, August 3, 2009, and August 3, 2009 respectively are not part of this review. They do appear on the COCs for Accutest Job Numbers JA21788, JA23207, JA23843, JA23843, JA24059, JA24059, JA24619, and JA24619 respectively. These samples were recollected. No qualification of the sample data is required.
- The COC for the samples collected on July 27, 2009 included samples B-CCI-012(24') and B-CCI-015(24'). The analysis for these samples was cancelled per request.

Organics

The following items/criteria were reviewed:

- Case narrative and deliverables compliance
- Chains-of-Custody
- Holding times both technical and procedural and sample preservation (including pH and temperature)
- Surrogate Compound recoveries, summary and data
- Matrix Spike/Matrix Spike Duplicate (MS/MSD) recoveries, summary and data
- Blank Spike Sample (BSS) recoveries, summary and data
- Method blank summary and data
- Gas Chromatography/Mass Spectroscopy (GC/MS) tuning and performance
- Initial and continuing calibration summaries
- Internal standard areas, retention times, summary and data
- Organic analysis data sheets (Form I)
- Blind Field Duplicate sample results
- Field Blank results
- Trip Blank results
- GC/MS and GC chromatograms, mass spectra and quantitation reports
- Quantitation/detection limits
- Qualitative and quantitative compound identification

The items listed above were in compliance with the analytical methods

methods and with the ASP and USEPA criteria with the exceptions discussed in the text below. The data have been validated according to the procedures outlined above and qualified accordingly.

Volatiles

- The laboratory has reported m+p-Xylene and o-Xylene as well as Xylene (total) on the Form Is. The TCL only requires Xylene (total) be reported. The m+p-Xylene and o-Xylene compounds have not been reviewed and will not be included in the summary tables or any electronic deliverable. Results for m+p-Xylene and o-Xylene have not been modified on the Form Is.
- The percent recovery (%R) for the surrogate compound 4-bromofluorobenzene (BFB) was slightly below QC criteria in the initial analysis of sample B-CCI-019(26') (57%; QC criteria 62-138%). The sample was reanalyzed further diluted and all surrogate %R were within QC criteria. All results have been reported from the initial analysis. All results are possibly biased low. Positive results are considered estimated and qualified "J" while non-detects are also considered estimated and qualified "UJ".
- The collection of MS/MSD was tracked in the field. Typically an MS/MSD set was collected and sent to the laboratory per twenty field samples collected. In many cases no MS/MSD was submitted to the laboratory with a particular group of samples. In these instances, the laboratory provided batch QC data to fulfill the protocol obligations. In some instances the laboratory analyzed an MS/MSD on a sample from this project. The batch QC is not utilized to qualify sample data. The blank spike sample serves as a better indicator of spike compound recovery.
- The %R for methyl acetate in the MS/MSD analysis of sample B-CCI-013(10') was above QC criteria. No qualification of the sample data is required as methyl acetate was not positively identified in sample B-CCI-013(10'). Results are valid and useable.
- The %R for 2-butanone, chloromethane, and vinyl chloride in the MSD analysis and vinyl chloride in the MS analysis of sample B-CCI-028(2') were below QC criteria. Qualification of sample data

data is not performed based on MS/MSD results alone. These compounds are possibly biased low. These compounds were not positively identified in sample B-CCI-028(2'). Results for these compounds in the unspiked portion of the sample only are considered estimated and qualified "UJ".

- The %R of 1,1,2-trichloroethane was slightly above QC criteria for the blank spike sample in Accutest Job Number JA24692 applicable to samples B-CCI-022, B-CCI-023, and B-CCI-024. Results for 1,1,2-trichloroethane for these samples may be biased high. No qualification of the sample data is required as 1,1,2-trichloroethane was not positively identified in these samples. Results are valid and useable.
- The %R of chloroform was slightly above QC criteria for the blank spike sample in Accutest Job Number JA25522 applicable to sample SW-CCI-013. Results for chloroform for sample SW-CCI-013 may be biased high. No qualification of the sample data is required as chloroform was not positively identified in sample SW-CCI-013. Results are valid and useable.
- Methylene chloride was detected in several method blanks at relatively low concentrations. Concentrations of methylene chloride in associated samples less than ten (10) times the concentration in the associated method blank are considered to be attributable to lab contamination and have been negated and qualified "U". Results are valid and useable as non-detects.
- The following table includes samples that were either analyzed initially at a dilution based on screening data and samples that were reanalyzed at a dilution due to the concentration of target compounds exceeding the calibration range of the instrument in the initial analysis. Samples that were reanalyzed at a dilution have been listed in this table with a "DL" suffix after the sample identification (ID). If the initial analysis was performed at a dilution and the dilution was justified and all target compounds were within the calibration range, then no further action was required and no qualification of the sample data is necessary. No compounds have been listed in the table for this occurrence. If the sample required reanalysis at a dilution, the laboratory has reported the final result for all compounds on the Form I. Only the target

target compounds listed in the table below are reported from the diluted analysis. Again, no qualification of the sample data is required.

Sample ID	Compound (s) Requiring Dilution	Dilution Factor
SW-CCI-002(18.5')	NA	medium level
SW-CCI-002(18.5') DL	1,2,4-trimethylbenzene 1,3,5-trimethylbenzene xylene (total)	medium level 20x
SW-CCI-003(18.5')	NA	medium level
B-CCI-007(20')	NA	medium level
DUPLICATE	NA	medium level
B-CCI-019(26')	NA	5x
B-CCI-019(26') DL	NA *	medium level
B-CCI-025	NA	medium level

* - no compounds reported from this analysis. Used only for confirmation due to surrogate deficiencies.

Semivolatiles

- The collection of MS/MSD was tracked in the field. Typically an MS/MSD set was collected and sent to the laboratory per twenty field samples collected. In many cases no MS/MSD was submitted to the laboratory with a particular group of samples. In these instances, the laboratory provided batch QC data to fulfill the protocol obligations. In some instances the laboratory analyzed an MS/MSD on a sample from this project. The batch QC is not utilized to qualify sample data. The blank spike sample serves as a better indicator of spike compound recovery.
- The %R for bis(2-ethylhexyl)phthalate was above QC criteria in the BSS associated to the field blank in Accutest Job Number JA23638. Bis(2-ethylhexyl)phthalate was not positively identified in the field blank therefore no qualification of the sample data is required.
- Bis(2-ethylhexyl)phthalate was detected in the method blank associated to the field blank in Accutest Job Number JA23638. Bis(2-ethylhexyl)phthalate was not positively identified in the field blank therefore no qualification of the sample data is required.

- All samples in Accutest Job Number JA25522 were reanalyzed to confirm positive detections due to a QC issue with the batch MS/MSD. No qualification of the sample data is required and results are valid and useable.
- Several method blanks contained system artifacts as Tentatively Identified Compounds (TICs). TICs were not reported for the samples and the TICs identified in the method blanks did not interfere with the reporting of any target compound in the samples. No qualification of the sample data is required.
- Naphthalene was positively identified in sample B-CCI-007(20), but not in the associated blind field duplicate sample, DUPLICATE. Benzo(b)fluoranthene and Indeno(1,2,3-cd)pyrene were not positively identified in sample B-CCI-007(20), but were in the associated blind field duplicate sample, DUPLICATE. Results for Naphthalene, Benzo(b)fluoranthene, and Indeno(1,2,3-cd)pyrene in both samples are considered possibly biased and have been qualified "J" for positive results and "UJ" for non-detects. The results are still valid and useable.

Pesticides

- The collection of MS/MSD was tracked in the field. Typically an MS/MSD set was collected and sent to the laboratory per twenty field samples collected. In many cases no MS/MSD was submitted to the laboratory with a particular group of samples. In these instances, the laboratory provided batch QC data to fulfill the protocol obligations. In some instances the laboratory analyzed an MS/MSD on a sample from this project. The batch QC is not utilized to qualify sample data. The blank spike sample serves as a better indicator of spike compound recovery.
- The %R for the surrogate compound decachlorobiphenyl (DCB) was above QC criteria for sample B-CCI-027(2'). The sample was reanalyzed at a dilution and exhibited a similar %R. Positive results for all compounds in sample B-CCI-027(2') are considered estimated and qualified "J".
- The %R for 4,4'-DDT, heptachlor, and methoxychlor were above QC criteria in the BSS applicable to all samples in Accutest Job Number

Number JA23207. Results for these compounds in all samples in Accutest Job Number JA23207 are possibly biased high. No qualification of the sample data is required as these compounds were not positively identified in any sample in Accutest Job Number JA23207.

- The relative percent difference (RPD) for 4,4'-DDE, endosulfan sulfate, endrin aldehyde, endosulfan I, and endrin ketone were above QC criteria in the MS/MSD analyzed on sample SW-CCI-005(18.5). Qualification of sample data is not performed based on MS/MSD results alone. The result for these compounds in the unspiked sample only are considered estimated and qualified "J" for positive detects and "UJ" for non-detects.
- The following table includes samples that were reanalyzed at a dilution due to the concentration of target compounds exceeding the calibration range of the instrument in the initial analysis. The laboratory has reported the final result for all compounds on the Form I. Only the target compounds listed in the table below are reported from the diluted analysis. No qualification of the sample data is required.

Sample ID	Compound (s) Requiring Dilution
B-CCI-027(2')	4,4'-DDE 4,4'-DDT

- The RPD of the concentration of the compounds listed in the table below differed by more than 25% on the two analytical GC columns for the following samples. The results for these compounds are possibly biased in the samples listed and have been qualified with a "J". The results are valid and useable.

Sample ID	Compound
B-CCI-02(08)	dieldrin
SW-CCI-007(19')	4,4'-DDD
B-CCI-013(10')	4,4'-DDD
B-CCI-027(2')	dieldrin, heptachlor epoxide
B-CCI-030(4')	dieldrin

Sample ID	Compound
SW-CCI-012	4,4'-DDT

- The concentration for 4,4'-DDT in sample B-CCI-02(08) has been reported from a two-fold (2x) diluted analysis. This analysis was required due to the concentration of 4,4'-DDT exceeding the calibration range of the instrument in the initial analysis of sample B-CCI-02(08). Only the result for 4,4'-DDT has been reported from this diluted analysis. All other results for sample B-CCI-02(08) have been reported from the initial analysis. No qualification of the sample data is required. Results are valid and useable.
- The laboratory has noted in several instances that results have been reported from the second analytical GC column due to percent difference (%D) issues with the first GC column. No qualification of the sample data is required as the %D on the second GC column meet QC criteria. Results are valid and useable.

Inorganics

The following items/criteria were reviewed:

- Case narrative and deliverable requirements
- Holding times and sample preservation
- Detection limits
- Inorganic analysis data sheets (Form I)
- Initial and continuing calibration verifications
- Contract Required Detection Limit (CRDL) standard analysis
- Lab blank data
- ICP interference check sample analysis
- Matrix spike and matrix spike duplicate analysis
- Laboratory control sample (LCS) results
- ICP serial dilution analysis
- Blind Field Duplicate sample results
- Field Blank results

The items listed above were in compliance with the analytical methods and with the ASP and USEPA criteria with the exceptions discussed in the text below. The data have been validated according to the procedures outlined above and qualified accordingly.

Metals

- The collection of MS/MSD was tracked in the field. Typically an MS/MSD set was collected and sent to the laboratory per twenty field samples collected. In many cases no MS/MSD was submitted to the laboratory with a particular group of samples. In these instances, the laboratory provided batch QC data to fulfill the protocol obligations. In some instances the laboratory analyzed an MS/MSD on a sample from this project. The batch QC is not utilized to qualify sample data. The blank spike sample serves as a better indicator of spike compound recovery.
- The following table lists samples whose associated matrix spike and/or matrix spike duplicate contained analytes where the percent recovery (%R) was outside QC criteria (75-125%). For a %R above criteria the results are possibly biased high and positive detections are qualified "J" while non-detects do not require qualification. For a %R below criteria the results are possibly biased low and positive detections are qualified "J" while non-detects are qualified "UJ". All data are still valid and useable for project objectives.

Accutest Job Number	Metal	%R (high/low)	Associated Samples
JA22090	antimony	low	B-CCI-02(08)
JA23207	antimony iron manganese	low low low	B-CCI-003(20'), B-CCI-004(20')
JA23638	antimony manganese	low low	SW-CCI-001(18.5'), SW-CCI-006(19'), SW-CCI-002(18.5'), SW-CCI-007(19'), SW-CCI-003(18.5'), SW-CCI-008(13.5'), SW-CCI-004(19'), SW-CCI-009(18.5'), SW-CCI-005(18.5')
JA23843	antimony	low	B-CCI-006(15'), B-CCI-007(20'), DUPLICATE
JA24059	antimony cobalt copper iron manganese potassium	low low low low low low	B-CCI-010(22'), B-CCI-013(10')

Accutest Job Number	Metal	%R (high/low)	Associated Samples
	selenium	low	
	sodium	low	
	vanadium	low	
JA24692	antimony	low	B-CCI-022, B-CCI-023, B-CCI-024,
	calcium	low	B-CCI-025
	mercury	high	
JA25087	aluminum	low	B-CCI-026 (2'), B-CCI-027 (2')
	antimony	low	
	chromium	low	
	potassium	low	
	vanadium	low	
JA25246	antimony	low	SW-CCI-011(24.5)
	chromium	low	
	manganese	low	
JA25446	antimony	low	B-CCI-028(2'), B-CCI-029(20'),
	calcium	low	B-CCI-030(4'), B-CCI-031(10')
	iron	low	
	magnesium	low	
	manganese	low	
JA25522	aluminum	low	SW-CCI-012, SW-CCI-013,
	antimony	low	SW-CCI-014
	iron	low	
	magnesium	high	
	manganese	low	
JA25638	aluminum	high	SW-CCI-015
	antimony	low	

- The following table lists samples whose associated serial dilution contained analytes where the percent difference (%D) was above QC criteria (20%). Results are possibly biased. Positive detections are qualified "J" while non-detects are qualified "UJ". All data are still valid and useable for project objectives.

Accutest Job Number	Metal	%R	Associated Samples
JA22090	potassium	58.0	B-CCI-02(08)
JA23207	beryllium	16.2	B-CCI-003(20'), B-CCI-004(20')
	nickel	13.0	
	zinc	15.3	
JA23843	calcium	15.3	B-CCI-006(15'), B-CCI-007(20'),

Accutest Job Number	Metal	%R	Associated Samples
			DUPLICATE
JA24059	aluminum	20.2	B-CCI-010(22'), B-CCI-013(10')
	barium	21.2	
	calcium	26.2	
	chromium	18.3	
	cobalt	17.2	
	copper	21.5	
	iron	20.8	
	magnesium	23.4	
	manganese	20.5	
	potassium	57.0	
	vanadium	19.4	
	zinc	19.2	
JA24692	beryllium	56.3	B-CCI-022, B-CCI-023, B-CCI-024,
	cobalt	23.3	B-CCI-025
	nickel	12.5	
JA25087	potassium	42.4	B-CCI-026 (2'), B-CCI-027 (2')
	nickel	20.6	
	zinc	16.2	
JA25446	calcium	10.1	B-CCI-028(2'), B-CCI-029(20'),
	cobalt	12.2	B-CCI-030(4'), B-CCI-031(10')
	copper	57.5	
	iron	11.0	
	magnesium	12.4	
	manganese	10.9	
	nickel	23.9	
	potassium	50.3	
	vanadium	11.1	
	zinc	15.0	
JA25522	potassium	39.6	SW-CCI-012, SW-CCI-013, SW-CCI-014

- The following table presents samples that required an analyte to be analyzed at a dilution due to the concentration of the analyte exceeding the calibration range in the initial analysis. The associated analyte and dilution factor are also included. The laboratory has reported the final result only on the Form I. No qualification of the sample data is required. All results are valid and useable.

Sample	Analyte	Dilution Factor
B-CCI-003(20')	calcium	20x

Sample	Analyte	Dilution Factor
B-CCI-004(20')	aluminum	10x
	calcium	10x
	magnesium	5x
SW-CCI-004(19')	calcium	2x
	lead	2x
	selenium	2x
SW-CCI-006(19')	calcium	3x
	lead	3x
	selenium	3x
SW-CCI-009(18.5')	calcium	5x
	lead	5x
	selenium	5x
B-CCI-010(22')	antimony *	5x
	calcium	5x
	lead *	5x
	selenium *	5x
	thallium *	5x
SW-CCI-011(24.5')	calcium	2x
	lead *	2x
	selenium *	2x
	thallium *	2x
B-CCI-028(2')	antimony *	2x
	calcium	2x
	lead	2x
	selenium *	2x
	thallium *	2x

* - dilution required due to interfering element. Analytes are non-detects and are considered possibly biased and qualified "UJ". Analytes have elevated reporting limits.

Accutest Laboratories

Report of Analysis

Page 1 of 2

Client Sample ID:	B-CCI-02(08)	Date Sampled:	06/30/09
Lab Sample ID:	JA22090-1	Date Received:	06/30/09
Matrix:	SO - Soil	Percent Solids:	83.0
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	Y89150.D	1	07/01/09	YXC	n/a	n/a	VY3732

Run #1	Initial Weight
Run #2	5.1 g

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	12	2.6	ug/kg	
71-43-2	Benzene	ND	1.2	0.40	ug/kg	
75-27-4	Bromodichloromethane	ND	5.9	0.30	ug/kg	
75-25-2	Bromoform	ND	5.9	0.18	ug/kg	
74-83-9	Bromomethane	ND	5.9	0.48	ug/kg	
78-93-3	2-Butanone (MEK)	ND	12	2.3	ug/kg	
104-51-8	n-Butylbenzene	ND	5.9	0.45	ug/kg	
135-98-8	sec-Butylbenzene	ND	5.9	0.58	ug/kg	
98-06-6	tert-Butylbenzene	ND	5.9	0.56	ug/kg	
75-15-0	Carbon disulfide	ND	5.9	0.36	ug/kg	
56-23-5	Carbon tetrachloride	ND	5.9	0.66	ug/kg	
108-90-7	Chlorobenzene	ND	5.9	0.40	ug/kg	
75-00-3	Chloroethane	ND	5.9	1.3	ug/kg	
67-66-3	Chloroform	ND	5.9	0.38	ug/kg	
74-87-3	Chloromethane	ND	5.9	0.19	ug/kg	
110-82-7	Cyclohexane	ND	5.9	0.18	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	12	0.64	ug/kg	
124-48-1	Dibromochloromethane	ND	5.9	0.13	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.2	0.16	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	5.9	0.32	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	5.9	0.32	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	5.9	0.40	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	5.9	1.1	ug/kg	
75-34-3	1,1-Dichloroethane	ND	5.9	0.16	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.2	0.41	ug/kg	
75-35-4	1,1-Dichloroethene	ND	5.9	0.78	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	5.9	0.28	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	5.9	0.53	ug/kg	
78-87-5	1,2-Dichloropropane	ND	5.9	0.15	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	5.9	0.16	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	5.9	0.11	ug/kg	
123-91-1	1,4-Dioxane	ND	150	100	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-02(08)	Date Sampled:	06/30/09
Lab Sample ID:	JA22090-1	Date Received:	06/30/09
Matrix:	SO - Soil	Percent Solids:	83.0
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	1.2	0.44	ug/kg	
76-13-1	Freon 113	ND	5.9	0.66	ug/kg	
591-78-6	2-Hexanone	ND	5.9	1.1	ug/kg	
98-82-8	Isopropylbenzene	ND	5.9	0.61	ug/kg	
79-20-9	Methyl Acetate	ND	5.9	0.97	ug/kg	
108-87-2	Methylcyclohexane	ND	5.9	0.77	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	ND	1.2	0.33	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	5.9	0.96	ug/kg	
75-09-2	Methylene chloride	ND	5.9	0.26	ug/kg	
103-65-1	n-Propylbenzene	ND	5.9	0.30	ug/kg	
100-42-5	Styrene	ND	5.9	0.13	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	5.9	0.35	ug/kg	
127-18-4	Tetrachloroethene	ND	5.9	0.17	ug/kg	
108-88-3	Toluene	ND	1.2	0.34	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	5.9	0.41	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	5.9	0.15	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	5.9	0.22	ug/kg	
79-01-6	Trichloroethene	ND	5.9	0.62	ug/kg	
75-69-4	Trichlorofluoromethane	ND	5.9	0.27	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	ND	5.9	0.51	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	ND	5.9	0.42	ug/kg	
75-01-4	Vinyl chloride	ND	5.9	0.21	ug/kg	
	m,p-Xylene	ND	2.4	0.55	ug/kg	
95-47-6	o-Xylene	ND	1.2	0.55	ug/kg	
1330-20-7	Xylene (total)	ND	2.4	0.55	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	108%		67-127%
17060-07-0	1,2-Dichloroethane-D4	100%		65-132%
2037-26-5	Toluene-D8	112%		74-129%
460-00-4	4-Bromofluorobenzene	107%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 2

3.1
3

Client Sample ID:	B-CCI-003 (20')	Date Sampled:	07/15/09
Lab Sample ID:	JA23207-1	Date Received:	07/15/09
Matrix:	SO - Soil	Percent Solids:	91.6
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	Y89661.D	1	07/16/09	YXC	n/a	n/a	VY3755

Run #1	Initial Weight
Run #2	5.3 g

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	13.5	10	2.3	ug/kg	
71-43-2	Benzene	0.55	1.0	0.35	ug/kg	J
75-27-4	Bromodichloromethane	ND	5.1	0.26	ug/kg	
75-25-2	Bromoform	ND	5.1	0.16	ug/kg	
74-83-9	Bromomethane	ND	5.1	0.42	ug/kg	
78-93-3	2-Butanone (MEK)	ND	10	2.0	ug/kg	
104-51-8	n-Butylbenzene	3.4	5.1	0.39	ug/kg	J
135-98-8	sec-Butylbenzene	1.5	5.1	0.50	ug/kg	J
98-06-6	tert-Butylbenzene	ND	5.1	0.49	ug/kg	
75-15-0	Carbon disulfide	ND	5.1	0.31	ug/kg	
56-23-5	Carbon tetrachloride	ND	5.1	0.57	ug/kg	
108-90-7	Chlorobenzene	ND	5.1	0.35	ug/kg	
75-00-3	Chloroethane	ND	5.1	1.2	ug/kg	
67-66-3	Chloroform	ND	5.1	0.33	ug/kg	
74-87-3	Chloromethane	ND	5.1	0.17	ug/kg	
110-82-7	Cyclohexane	ND	5.1	0.16	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	10	0.56	ug/kg	
124-48-1	Dibromochloromethane	ND	5.1	0.11	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.0	0.14	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	5.1	0.28	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	5.1	0.28	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	5.1	0.35	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	5.1	0.97	ug/kg	
75-34-3	1,1-Dichloroethane	ND	5.1	0.14	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.0	0.36	ug/kg	
75-35-4	1,1-Dichloroethene	ND	5.1	0.68	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	5.1	0.25	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	5.1	0.46	ug/kg	
78-87-5	1,2-Dichloropropane	ND	5.1	0.13	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	5.1	0.14	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	5.1	0.099	ug/kg	
123-91-1	1,4-Dioxane	ND	130	89	ug/kg	

ND = Not detected MDL - Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-003 (20')	Date Sampled:	07/15/09
Lab Sample ID:	JA23207-1	Date Received:	07/15/09
Matrix:	SO - Soil	Percent Solids:	91.6
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	6.2	1.0	0.38	ug/kg	
76-13-1	Freon 113	ND	5.1	0.58	ug/kg	
591-78-6	2-Hexanone	ND	5.1	0.99	ug/kg	
98-82-8	Isopropylbenzene	2.0	5.1	0.53	ug/kg	J
79-20-9	Methyl Acetate	ND	5.1	0.85	ug/kg	
108-87-2	Methylcyclohexane	1.5	5.1	0.67	ug/kg	J
1634-04-4	Methyl Tert Butyl Ether	22.6	1.0	0.29	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	5.1	0.84	ug/kg	
75-09-2	Methylene chloride	ND	5.1	0.23	ug/kg	
103-65-1	n-Propylbenzene	6.6	5.1	0.26	ug/kg	
100-42-5	Styrene	ND	5.1	0.11	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	5.1	0.30	ug/kg	
127-18-4	Tetrachloroethene	ND	5.1	0.15	ug/kg	
108-88-3	Toluene	2.8	1.0	0.30	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	5.1	0.36	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	5.1	0.13	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	5.1	0.19	ug/kg	
79-01-6	Trichloroethene	ND	5.1	0.54	ug/kg	
75-69-4	Trichlorofluoromethane	ND	5.1	0.24	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	53.1	5.1	0.44	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	24.2	5.1	0.37	ug/kg	
75-01-4	Vinyl chloride	ND	5.1	0.18	ug/kg	
	m,p-Xylene	24.6	2.1	0.48	ug/kg	
95-47-6	o-Xylene	20.0	1.0	0.48	ug/kg	
1330-20-7	Xylene (total)	44.5	2.1	0.48	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	71%		67-127%
17060-07-0	1,2-Dichloroethane-D4	104%		65-132%
2037-26-5	Toluene-D8	110%		74-129%
460-00-4	4-Bromofluorobenzene	101%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 2

Client Sample ID:	B-CCI-004 (20')	Date Sampled:	07/15/09
Lab Sample ID:	JA23207-2	Date Received:	07/15/09
Matrix:	SO - Soil	Percent Solids:	91.6
Method:	SW846 8260B	Project: Courtlandt Corners I, 364 East 161st Street, Bronx, NY	

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	Y89660.D	1	07/16/09	YXC	n/a	n/a	VY3755

Run #1	Initial Weight
Run #2	5.2 g

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	10	2.3	ug/kg	
71-43-2	Benzene	ND	1.0	0.36	ug/kg	
75-27-4	Bromodichloromethane	ND	5.2	0.27	ug/kg	
75-25-2	Bromoform	ND	5.2	0.16	ug/kg	
74-83-9	Bromomethane	ND	5.2	0.42	ug/kg	
78-93-3	2-Butanone (MEK)	ND	10	2.1	ug/kg	
104-51-8	n-Butylbenzene	ND	5.2	0.40	ug/kg	
135-98-8	sec-Butylbenzene	ND	5.2	0.51	ug/kg	
98-06-6	tert-Butylbenzene	ND	5.2	0.50	ug/kg	
75-15-0	Carbon disulfide	ND	5.2	0.32	ug/kg	
56-23-5	Carbon tetrachloride	ND	5.2	0.58	ug/kg	
108-90-7	Chlorobenzene	ND	5.2	0.36	ug/kg	
75-00-3	Chloroethane	ND	5.2	1.2	ug/kg	
67-66-3	Chloroform	ND	5.2	0.33	ug/kg	
74-87-3	Chloromethane	ND	5.2	0.17	ug/kg	
110-82-7	Cyclohexane	ND	5.2	0.16	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	10	0.57	ug/kg	
124-48-1	Dibromochloromethane	ND	5.2	0.12	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.0	0.14	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	5.2	0.28	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	5.2	0.29	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	5.2	0.35	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	5.2	0.99	ug/kg	
75-34-3	1,1-Dichloroethane	ND	5.2	0.14	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.0	0.36	ug/kg	
75-35-4	1,1-Dichloroethene	ND	5.2	0.69	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	5.2	0.25	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	5.2	0.47	ug/kg	
78-87-5	1,2-Dichloropropane	ND	5.2	0.14	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	5.2	0.14	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	5.2	0.10	ug/kg	
123-91-1	1,4-Dioxane	ND	130	91	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-004 (20')	Date Sampled:	07/15/09
Lab Sample ID:	JA23207-2	Date Received:	07/15/09
Matrix:	SO - Soil	Percent Solids:	91.6
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	1.0	0.39	ug/kg	
76-13-1	Freon 113	ND	5.2	0.59	ug/kg	
591-78-6	2-Hexanone	ND	5.2	1.0	ug/kg	
98-82-8	Isopropylbenzene	ND	5.2	0.54	ug/kg	
79-20-9	Methyl Acetate	ND	5.2	0.86	ug/kg	
108-87-2	Methylcyclohexane	ND	5.2	0.69	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	ND	1.0	0.30	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	5.2	0.85	ug/kg	
75-09-2	Methylene chloride	ND	5.2	0.23	ug/kg	
103-65-1	n-Propylbenzene	ND	5.2	0.27	ug/kg	
100-42-5	Styrene	ND	5.2	0.11	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	5.2	0.31	ug/kg	
127-18-4	Tetrachloroethene	ND	5.2	0.15	ug/kg	
108-88-3	Toluene	ND	1.0	0.31	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	5.2	0.36	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	5.2	0.13	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	5.2	0.19	ug/kg	
79-01-6	Trichloroethene	ND	5.2	0.55	ug/kg	
75-69-4	Trichlorofluoromethane	ND	5.2	0.24	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	ND	5.2	0.45	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	ND	5.2	0.38	ug/kg	
75-01-4	Vinyl chloride	ND	5.2	0.19	ug/kg	
	m,p-Xylene	ND	2.1	0.49	ug/kg	
95-47-6	o-Xylene	ND	1.0	0.49	ug/kg	
1330-20-7	Xylene (total)	ND	2.1	0.49	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	104%		67-127%
17060-07-0	1,2-Dichloroethane-D4	97%		65-132%
2037-26-5	Toluene-D8	110%		74-129%
460-00-4	4-Bromofluorobenzene	105%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 2

Client Sample ID:	B-CCI-006 (15')	Date Sampled:	07/22/09
Lab Sample ID:	JA23843-1	Date Received:	07/23/09
Matrix:	SO - Soil	Percent Solids:	90.6
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	G122196.D	1	07/23/09	SJM	n/a	n/a	VG5829

Run #1	Initial Weight
Run #2	4.5 g

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	12.9	12	2.7	ug/kg	
71-43-2	Benzene	ND	1.2	0.42	ug/kg	
75-27-4	Bromodichloromethane	ND	6.1	0.32	ug/kg	
75-25-2	Bromoform	ND	6.1	0.19	ug/kg	
74-83-9	Bromomethane	ND	6.1	0.50	ug/kg	
78-93-3	2-Butanone (MEK)	ND	12	2.4	ug/kg	
104-51-8	n-Butylbenzene	ND	6.1	0.47	ug/kg	
135-98-8	sec-Butylbenzene	ND	6.1	0.60	ug/kg	
98-06-6	tert-Butylbenzene	ND	6.1	0.58	ug/kg	
75-15-0	Carbon disulfide	ND	6.1	0.37	ug/kg	
56-23-5	Carbon tetrachloride	ND	6.1	0.68	ug/kg	
108-90-7	Chlorobenzene	ND	6.1	0.42	ug/kg	
75-00-3	Chloroethane	ND	6.1	1.4	ug/kg	
67-66-3	Chloroform	ND	6.1	0.39	ug/kg	
74-87-3	Chloromethane	ND	6.1	0.20	ug/kg	
110-82-7	Cyclohexane	ND	6.1	0.19	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	12	0.66	ug/kg	
124-48-1	Dibromochloromethane	ND	6.1	0.13	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.2	0.17	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	6.1	0.33	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	6.1	0.34	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	6.1	0.41	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	6.1	1.2	ug/kg	
75-34-3	1,1-Dichloroethane	ND	6.1	0.17	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.2	0.42	ug/kg	
75-35-4	1,1-Dichloroethene	ND	6.1	0.81	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	6.1	0.29	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	6.1	0.55	ug/kg	
78-87-5	1,2-Dichloropropane	ND	6.1	0.16	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	6.1	0.16	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	6.1	0.12	ug/kg	
123-91-1	1,4-Dioxane	ND	150	110	ug/kg	

ND = Not detected MDL - Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-006 (15')	Date Sampled:	07/22/09
Lab Sample ID:	JA23843-1	Date Received:	07/23/09
Matrix:	SO - Soil	Percent Solids:	90.6
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	1.2	0.45	ug/kg	
76-13-1	Freon 113	ND	6.1	0.69	ug/kg	
591-78-6	2-Hexanone	ND	6.1	1.2	ug/kg	
98-82-8	Isopropylbenzene	ND	6.1	0.64	ug/kg	
79-20-9	Methyl Acetate	ND	6.1	1.0	ug/kg	
108-87-2	Methylcyclohexane	ND	6.1	0.80	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	ND	1.2	0.35	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	6.1	0.99	ug/kg	
75-09-2	Methylene chloride	3.5 ND	6.1	0.27	ug/kg	JN
103-65-1	n-Propylbenzene	ND	6.1	0.31	ug/kg	
100-42-5	Styrene	ND	6.1	0.13	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	6.1	0.36	ug/kg	
127-18-4	Tetrachloroethene	ND	6.1	0.18	ug/kg	
108-88-3	Toluene	ND	1.2	0.36	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	6.1	0.42	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	6.1	0.16	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	6.1	0.23	ug/kg	
79-01-6	Trichloroethene	ND	6.1	0.65	ug/kg	
75-69-4	Trichlorofluoromethane	ND	6.1	0.28	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	ND	6.1	0.53	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	ND	6.1	0.44	ug/kg	
75-01-4	Vinyl chloride	ND	6.1	0.22	ug/kg	
	m,p-Xylene	ND	2.5	0.58	ug/kg	
95-47-6	o-Xylene	4.2	1.2	0.58	ug/kg	
1330-20-7	Xylene (total)	4.2	2.5	0.58	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	84%		67-127%
17060-07-0	1,2-Dichloroethane-D4	80%		65-132%
2037-26-5	Toluene-D8	97%		74-129%
460-00-4	4-Bromofluorobenzene	97%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 2

Client Sample ID:	B-CCI-007 (20')	Date Sampled:	07/22/09
Lab Sample ID:	JA23843-2	Date Received:	07/23/09
Matrix:	SO - Soil	Percent Solids:	94.5
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1 ^a	D157509.D	1	07/24/09	TDN	n/a	n/a	VD6331
Run #2							

Run #	Initial Weight	Final Volume	Methanol Aliquot
Run #1	10.0 g	10.0 ml	100 ul
Run #2			

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	560	120	ug/kg	
71-43-2	Benzene	ND	56	19	ug/kg	
75-27-4	Bromodichloromethane	ND	280	14	ug/kg	
75-25-2	Bromoform	ND	280	8.4	ug/kg	
74-83-9	Bromomethane	ND	280	23	ug/kg	
78-93-3	2-Butanone (MEK)	ND	560	110	ug/kg	
104-51-8	n-Butylbenzene	56.6	280	21	ug/kg	J
135-98-8	sec-Butylbenzene	ND	280	27	ug/kg	
98-06-6	tert-Butylbenzene	ND	280	27	ug/kg	
75-15-0	Carbon disulfide	ND	280	17	ug/kg	
56-23-5	Carbon tetrachloride	ND	280	31	ug/kg	
108-90-7	Chlorobenzene	ND	280	19	ug/kg	
75-00-3	Chloroethane	ND	280	64	ug/kg	
67-66-3	Chloroform	ND	280	18	ug/kg	
74-87-3	Chloromethane	ND	280	9.2	ug/kg	
110-82-7	Cyclohexane	ND	280	8.4	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	560	30	ug/kg	
124-48-1	Dibromochloromethane	ND	280	6.1	ug/kg	
106-93-4	1,2-Dibromoethane	ND	56	7.6	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	280	15	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	280	15	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	280	19	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	280	53	ug/kg	
75-34-3	1,1-Dichloroethane	ND	280	7.7	ug/kg	
107-06-2	1,2-Dichloroethane	ND	56	19	ug/kg	
75-35-4	1,1-Dichloroethene	ND	280	37	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	280	13	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	280	25	ug/kg	
78-87-5	1,2-Dichloropropane	ND	280	7.3	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	280	7.4	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	280	5.4	ug/kg	
123-91-1	1,4-Dioxane	ND	7000	4800	ug/kg	

ND = Not detected MDL - Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-007 (20')	Date Sampled:	07/22/09
Lab Sample ID:	JA23843-2	Date Received:	07/23/09
Matrix:	SO - Soil	Percent Solids:	94.5
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	56	21	ug/kg	
76-13-1	Freon 113	ND	280	31	ug/kg	
591-78-6	2-Hexanone	ND	280	54	ug/kg	
98-82-8	Isopropylbenzene	ND	280	29	ug/kg	
79-20-9	Methyl Acetate	ND	280	46	ug/kg	
108-87-2	Methylcyclohexane	ND	280	37	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	ND	56	16	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	280	45	ug/kg	
75-09-2	Methylene chloride	ND	280	12	ug/kg	
103-65-1	n-Propylbenzene	25.5	280	14	ug/kg	J
100-42-5	Styrene	ND	280	6.0	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	280	16	ug/kg	
127-18-4	Tetrachloroethene	ND	280	8.1	ug/kg	
108-88-3	Toluene	ND	56	16	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	280	19	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	280	7.1	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	280	10	ug/kg	
79-01-6	Trichloroethene	ND	280	29	ug/kg	
75-69-4	Trichlorofluoromethane	ND	280	13	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	146	280	24	ug/kg	J
108-67-8	1,3,5-Trimethylbenzene	237	280	20	ug/kg	J
75-01-4	Vinyl chloride	ND	280	9.9	ug/kg	
	m,p-Xylene	37.6	110	26	ug/kg	J
95-47-6	o-Xylene	84.9	56	26	ug/kg	
1330-20-7	Xylene (total)	123	110	26	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	105%		67-127%
17060-07-0	1,2-Dichloroethane-D4	111%		65-132%
2037-26-5	Toluene-D8	105%		74-129%
460-00-4	4-Bromofluorobenzene	106%		62-138%

(a) Dilution required due to matrix interference.

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 2

Client Sample ID:	DUPLICATE	Date Sampled:	07/22/09
Lab Sample ID:	JA23843-5	Date Received:	07/23/09
Matrix:	SO - Soil	Percent Solids:	94.2
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1 ^a	D157508.D	1	07/24/09	TDN	n/a	n/a	VD6331
Run #2							

Run #	Initial Weight	Final Volume	Methanol Aliquot
Run #1	10.0 g	10.0 ml	100 ul
Run #2			

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	560	130	ug/kg	
71-43-2	Benzene	ND	56	19	ug/kg	
75-27-4	Bromodichloromethane	ND	280	14	ug/kg	
75-25-2	Bromoform	ND	280	8.5	ug/kg	
74-83-9	Bromomethane	ND	280	23	ug/kg	
78-93-3	2-Butanone (MEK)	ND	560	110	ug/kg	
104-51-8	n-Butylbenzene	63.6	280	21	ug/kg	J
135-98-8	sec-Butylbenzene	ND	280	27	ug/kg	
98-06-6	tert-Butylbenzene	ND	280	27	ug/kg	
75-15-0	Carbon disulfide	ND	280	17	ug/kg	
56-23-5	Carbon tetrachloride	ND	280	31	ug/kg	
108-90-7	Chlorobenzene	ND	280	19	ug/kg	
75-00-3	Chloroethane	ND	280	64	ug/kg	
67-66-3	Chloroform	ND	280	18	ug/kg	
74-87-3	Chloromethane	ND	280	9.3	ug/kg	
110-82-7	Cyclohexane	ND	280	8.5	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	560	30	ug/kg	
124-48-1	Dibromochloromethane	ND	280	6.2	ug/kg	
106-93-4	1,2-Dibromoethane	ND	56	7.7	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	280	15	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	280	15	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	280	19	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	280	53	ug/kg	
75-34-3	1,1-Dichloroethane	ND	280	7.7	ug/kg	
107-06-2	1,2-Dichloroethane	ND	56	19	ug/kg	
75-35-4	1,1-Dichloroethene	ND	280	37	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	280	13	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	280	25	ug/kg	
78-87-5	1,2-Dichloropropane	ND	280	7.3	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	280	7.5	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	280	5.4	ug/kg	
123-91-1	1,4-Dioxane	ND	7000	4900	ug/kg	

ND = Not detected MDL - Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	DUPLICATE	Date Sampled:	07/22/09
Lab Sample ID:	JA23843-5	Date Received:	07/23/09
Matrix:	SO - Soil	Percent Solids:	94.2
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	56	21	ug/kg	
76-13-1	Freon 113	ND	280	32	ug/kg	
591-78-6	2-Hexanone	ND	280	54	ug/kg	
98-82-8	Isopropylbenzene	ND	280	29	ug/kg	
79-20-9	Methyl Acetate	ND	280	46	ug/kg	
108-87-2	Methylcyclohexane	ND	280	37	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	ND	56	16	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	280	46	ug/kg	
75-09-2	Methylene chloride	ND	280	13	ug/kg	
103-65-1	n-Propylbenzene	33.0	280	14	ug/kg	J
100-42-5	Styrene	ND	280	6.0	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	280	16	ug/kg	
127-18-4	Tetrachloroethene	ND	280	8.1	ug/kg	
108-88-3	Toluene	ND	56	16	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	280	19	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	280	7.2	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	280	10	ug/kg	
79-01-6	Trichloroethene	ND	280	30	ug/kg	
75-69-4	Trichlorofluoromethane	ND	280	13	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	216	280	24	ug/kg	J
108-67-8	1,3,5-Trimethylbenzene	307	280	20	ug/kg	
75-01-4	Vinyl chloride	ND	280	10	ug/kg	
	m,p-Xylene	ND	110	26	ug/kg	
95-47-6	o-Xylene	123	56	26	ug/kg	
1330-20-7	Xylene (total)	148	110	26	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	105%		67-127%
17060-07-0	1,2-Dichloroethane-D4	113%		65-132%
2037-26-5	Toluene-D8	101%		74-129%
460-00-4	4-Bromofluorobenzene	105%		62-138%

(a) Dilution required due to matrix interference.

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 2

Client Sample ID:	B-CCI-010(22')	Date Sampled:	07/24/09
Lab Sample ID:	JA24059-1	Date Received:	07/27/09
Matrix:	SO - Soil	Percent Solids:	82.1
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	G122316.D	1	07/28/09	SJM	n/a	n/a	VG5835

Run #1	Initial Weight
Run #2	4.6 g

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	13	3.0	ug/kg	
71-43-2	Benzene	ND	1.3	0.45	ug/kg	
75-27-4	Bromodichloromethane	ND	6.6	0.34	ug/kg	
75-25-2	Bromoform	ND	6.6	0.20	ug/kg	
74-83-9	Bromomethane	ND	6.6	0.53	ug/kg	
78-93-3	2-Butanone (MEK)	ND	13	2.6	ug/kg	
104-51-8	n-Butylbenzene	ND	6.6	0.50	ug/kg	
135-98-8	sec-Butylbenzene	1.6	6.6	0.65	ug/kg	J
98-06-6	tert-Butylbenzene	ND	6.6	0.63	ug/kg	
75-15-0	Carbon disulfide	0.85	6.6	0.40	ug/kg	J
56-23-5	Carbon tetrachloride	ND	6.6	0.73	ug/kg	
108-90-7	Chlorobenzene	ND	6.6	0.45	ug/kg	
75-00-3	Chloroethane	ND	6.6	1.5	ug/kg	
67-66-3	Chloroform	ND	6.6	0.42	ug/kg	
74-87-3	Chloromethane	ND	6.6	0.22	ug/kg	
110-82-7	Cyclohexane	ND	6.6	0.20	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	13	0.72	ug/kg	
124-48-1	Dibromochloromethane	ND	6.6	0.15	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.3	0.18	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	6.6	0.36	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	6.6	0.36	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	6.6	0.45	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	6.6	1.3	ug/kg	
75-34-3	1,1-Dichloroethane	ND	6.6	0.18	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.3	0.46	ug/kg	
75-35-4	1,1-Dichloroethene	ND	6.6	0.88	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	6.6	0.32	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	6.6	0.59	ug/kg	
78-87-5	1,2-Dichloropropane	ND	6.6	0.17	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	6.6	0.18	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	6.6	0.13	ug/kg	
123-91-1	1,4-Dioxane	ND	170	110	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

3.1
3

Client Sample ID:	B-CCI-010(22')	Date Sampled:	07/24/09
Lab Sample ID:	JA24059-1	Date Received:	07/27/09
Matrix:	SO - Soil	Percent Solids:	82.1
Method:	SW846 8260B	Project: Courtlandt Corners I, 364 East 161st Street, Bronx, NY	

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	1.3	0.49	ug/kg	
76-13-1	Freon 113	ND	6.6	0.74	ug/kg	
591-78-6	2-Hexanone	ND	6.6	1.3	ug/kg	
98-82-8	Isopropylbenzene	ND	6.6	0.69	ug/kg	
79-20-9	Methyl Acetate	ND	6.6	1.1	ug/kg	
108-87-2	Methylcyclohexane	5.6	6.6	0.87	ug/kg	J
1634-04-4	Methyl Tert Butyl Ether	ND	1.3	0.37	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	6.6	1.1	ug/kg	
75-09-2	Methylene chloride	ND	6.6	0.30	ug/kg	
103-65-1	n-Propylbenzene	ND	6.6	0.34	ug/kg	
100-42-5	Styrene	ND	6.6	0.14	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	6.6	0.39	ug/kg	
127-18-4	Tetrachloroethene	ND	6.6	0.19	ug/kg	
108-88-3	Toluene	ND	1.3	0.39	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	6.6	0.46	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	6.6	0.17	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	6.6	0.24	ug/kg	
79-01-6	Trichloroethene	ND	6.6	0.70	ug/kg	
75-69-4	Trichlorofluoromethane	ND	6.6	0.30	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	ND	6.6	0.57	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	9.5	6.6	0.47	ug/kg	
75-01-4	Vinyl chloride	ND	6.6	0.24	ug/kg	
	m,p-Xylene	ND	2.6	0.62	ug/kg	
95-47-6	o-Xylene	ND	1.3	0.62	ug/kg	
1330-20-7	Xylene (total)	ND	2.6	0.62	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	71%		67-127%
17060-07-0	1,2-Dichloroethane-D4	73%		65-132%
2037-26-5	Toluene-D8	94%		74-129%
460-00-4	4-Bromofluorobenzene	101%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 2

Client Sample ID:	B-CCI-013(10')	Date Sampled:	07/27/09
Lab Sample ID:	JA24059-3	Date Received:	07/27/09
Matrix:	SO - Soil	Percent Solids:	82.9
Method:	SW846 8260B	Project: Courtlandt Corners I, 364 East 161st Street, Bronx, NY	

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	G122317.D	1	07/28/09	SJM	n/a	n/a	VG5835

Run #1	Initial Weight
Run #2	4.7 g

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	13	2.9	ug/kg	
71-43-2	Benzene	ND	1.3	0.44	ug/kg	
75-27-4	Bromodichloromethane	ND	6.4	0.33	ug/kg	
75-25-2	Bromoform	ND	6.4	0.19	ug/kg	
74-83-9	Bromomethane	ND	6.4	0.52	ug/kg	
78-93-3	2-Butanone (MEK)	ND	13	2.5	ug/kg	
104-51-8	n-Butylbenzene	ND	6.4	0.49	ug/kg	
135-98-8	sec-Butylbenzene	ND	6.4	0.63	ug/kg	
98-06-6	tert-Butylbenzene	ND	6.4	0.61	ug/kg	
75-15-0	Carbon disulfide	ND	6.4	0.39	ug/kg	
56-23-5	Carbon tetrachloride	ND	6.4	0.71	ug/kg	
108-90-7	Chlorobenzene	ND	6.4	0.44	ug/kg	
75-00-3	Chloroethane	ND	6.4	1.5	ug/kg	
67-66-3	Chloroform	ND	6.4	0.41	ug/kg	
74-87-3	Chloromethane	ND	6.4	0.21	ug/kg	
110-82-7	Cyclohexane	ND	6.4	0.19	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	13	0.69	ug/kg	
124-48-1	Dibromochloromethane	ND	6.4	0.14	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.3	0.18	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	6.4	0.35	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	6.4	0.35	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	6.4	0.43	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	6.4	1.2	ug/kg	
75-34-3	1,1-Dichloroethane	ND	6.4	0.18	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.3	0.44	ug/kg	
75-35-4	1,1-Dichloroethene	ND	6.4	0.85	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	6.4	0.31	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	6.4	0.58	ug/kg	
78-87-5	1,2-Dichloropropane	ND	6.4	0.17	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	6.4	0.17	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	6.4	0.12	ug/kg	
123-91-1	1,4-Dioxane	ND	160	110	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-013(10')	Date Sampled:	07/27/09
Lab Sample ID:	JA24059-3	Date Received:	07/27/09
Matrix:	SO - Soil	Percent Solids:	82.9
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	0.93	1.3	0.48	ug/kg	J
76-13-1	Freon 113	ND	6.4	0.72	ug/kg	
591-78-6	2-Hexanone	ND	6.4	1.2	ug/kg	
98-82-8	Isopropylbenzene	ND	6.4	0.66	ug/kg	
79-20-9	Methyl Acetate	ND	6.4	1.1	ug/kg	
108-87-2	Methylcyclohexane	ND	6.4	0.84	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	ND	1.3	0.36	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	6.4	1.0	ug/kg	
75-09-2	Methylene chloride	ND	6.4	0.29	ug/kg	
103-65-1	n-Propylbenzene	ND	6.4	0.33	ug/kg	
100-42-5	Styrene	ND	6.4	0.14	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	6.4	0.38	ug/kg	
127-18-4	Tetrachloroethene	ND	6.4	0.19	ug/kg	
108-88-3	Toluene	1.7	1.3	0.37	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	6.4	0.44	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	6.4	0.16	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	6.4	0.24	ug/kg	
79-01-6	Trichloroethene	ND	6.4	0.68	ug/kg	
75-69-4	Trichlorofluoromethane	ND	6.4	0.29	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	1.2	6.4	0.55	ug/kg	J
108-67-8	1,3,5-Trimethylbenzene	0.85	6.4	0.46	ug/kg	J
75-01-4	Vinyl chloride	ND	6.4	0.23	ug/kg	
	m,p-Xylene	2.6	2.6	0.60	ug/kg	
95-47-6	o-Xylene	1.3	1.3	0.60	ug/kg	
1330-20-7	Xylene (total)	3.9	2.6	0.60	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	85%		67-127%
17060-07-0	1,2-Dichloroethane-D4	90%		65-132%
2037-26-5	Toluene-D8	96%		74-129%
460-00-4	4-Bromofluorobenzene	102%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 2

Client Sample ID:	B-CCI-016(23')	Date Sampled:	07/31/09
Lab Sample ID:	JA24443-1	Date Received:	07/31/09
Matrix:	SO - Soil	Percent Solids:	94.6
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	X94201.D	1	07/31/09	JTP	n/a	n/a	VX3946

Run #1	Initial Weight
Run #2	4.7 g

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	11	2.5	ug/kg	
71-43-2	Benzene	ND	1.1	0.38	ug/kg	
75-27-4	Bromodichloromethane	ND	5.6	0.29	ug/kg	
75-25-2	Bromoform	ND	5.6	0.17	ug/kg	
74-83-9	Bromomethane	ND	5.6	0.45	ug/kg	
78-93-3	2-Butanone (MEK)	ND	11	2.2	ug/kg	
104-51-8	n-Butylbenzene	ND	5.6	0.43	ug/kg	
135-98-8	sec-Butylbenzene	ND	5.6	0.55	ug/kg	
98-06-6	tert-Butylbenzene	ND	5.6	0.53	ug/kg	
75-15-0	Carbon disulfide	ND	5.6	0.34	ug/kg	
56-23-5	Carbon tetrachloride	ND	5.6	0.62	ug/kg	
108-90-7	Chlorobenzene	ND	5.6	0.38	ug/kg	
75-00-3	Chloroethane	ND	5.6	1.3	ug/kg	
67-66-3	Chloroform	ND	5.6	0.36	ug/kg	
74-87-3	Chloromethane	ND	5.6	0.19	ug/kg	
110-82-7	Cyclohexane	ND	5.6	0.17	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	11	0.61	ug/kg	
124-48-1	Dibromochloromethane	ND	5.6	0.12	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.1	0.15	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	5.6	0.30	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	5.6	0.31	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	5.6	0.38	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	5.6	1.1	ug/kg	
75-34-3	1,1-Dichloroethane	ND	5.6	0.16	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.1	0.39	ug/kg	
75-35-4	1,1-Dichloroethene	ND	5.6	0.74	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	5.6	0.27	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	5.6	0.50	ug/kg	
78-87-5	1,2-Dichloropropane	ND	5.6	0.15	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	5.6	0.15	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	5.6	0.11	ug/kg	
123-91-1	1,4-Dioxane	ND	140	97	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-016(23')	Date Sampled:	07/31/09
Lab Sample ID:	JA24443-1	Date Received:	07/31/09
Matrix:	SO - Soil	Percent Solids:	94.6
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	1.1	0.42	ug/kg	
76-13-1	Freon 113	ND	5.6	0.63	ug/kg	
591-78-6	2-Hexanone	ND	5.6	1.1	ug/kg	
98-82-8	Isopropylbenzene	ND	5.6	0.58	ug/kg	
79-20-9	Methyl Acetate	ND	5.6	0.93	ug/kg	
108-87-2	Methylcyclohexane	ND	5.6	0.74	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	9.7	1.1	0.32	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	5.6	0.91	ug/kg	
75-09-2	Methylene chloride	ND	5.6	0.25	ug/kg	
103-65-1	n-Propylbenzene	ND	5.6	0.29	ug/kg	
100-42-5	Styrene	ND	5.6	0.12	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	5.6	0.33	ug/kg	
127-18-4	Tetrachloroethene	ND	5.6	0.16	ug/kg	
108-88-3	Toluene	ND	1.1	0.33	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	5.6	0.39	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	5.6	0.14	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	5.6	0.21	ug/kg	
79-01-6	Trichloroethene	ND	5.6	0.59	ug/kg	
75-69-4	Trichlorofluoromethane	ND	5.6	0.26	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	ND	5.6	0.48	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	ND	5.6	0.40	ug/kg	
75-01-4	Vinyl chloride	ND	5.6	0.20	ug/kg	
	m,p-Xylene	ND	2.2	0.53	ug/kg	
95-47-6	o-Xylene	ND	1.1	0.53	ug/kg	
1330-20-7	Xylene (total)	ND	2.2	0.53	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	95%		67-127%
17060-07-0	1,2-Dichloroethane-D4	93%		65-132%
2037-26-5	Toluene-D8	104%		74-129%
460-00-4	4-Bromofluorobenzene	84%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 2

Client Sample ID:	B-CCI-017(23')	Date Sampled:	07/31/09
Lab Sample ID:	JA24443-2	Date Received:	07/31/09
Matrix:	SO - Soil	Percent Solids:	92.3
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	V93026.D	1	07/31/09	JLI	n/a	n/a	VV3824

Run #1	Initial Weight
Run #2	5.2 g

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	7.1	10	2.3	ug/kg	J
71-43-2	Benzene	ND	1.0	0.36	ug/kg	
75-27-4	Bromodichloromethane	ND	5.2	0.27	ug/kg	
75-25-2	Bromoform	ND	5.2	0.16	ug/kg	
74-83-9	Bromomethane	ND	5.2	0.42	ug/kg	
78-93-3	2-Butanone (MEK)	ND	10	2.1	ug/kg	
104-51-8	n-Butylbenzene	ND	5.2	0.40	ug/kg	
135-98-8	sec-Butylbenzene	ND	5.2	0.51	ug/kg	
98-06-6	tert-Butylbenzene	ND	5.2	0.49	ug/kg	
75-15-0	Carbon disulfide	ND	5.2	0.32	ug/kg	
56-23-5	Carbon tetrachloride	ND	5.2	0.58	ug/kg	
108-90-7	Chlorobenzene	ND	5.2	0.35	ug/kg	
75-00-3	Chloroethane	ND	5.2	1.2	ug/kg	
67-66-3	Chloroform	ND	5.2	0.33	ug/kg	
74-87-3	Chloromethane	ND	5.2	0.17	ug/kg	
110-82-7	Cyclohexane	ND	5.2	0.16	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	10	0.56	ug/kg	
124-48-1	Dibromochloromethane	ND	5.2	0.11	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.0	0.14	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	5.2	0.28	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	5.2	0.29	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	5.2	0.35	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	5.2	0.98	ug/kg	
75-34-3	1,1-Dichloroethane	ND	5.2	0.14	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.0	0.36	ug/kg	
75-35-4	1,1-Dichloroethene	ND	5.2	0.69	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	5.2	0.25	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	5.2	0.47	ug/kg	
78-87-5	1,2-Dichloropropane	ND	5.2	0.14	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	5.2	0.14	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	5.2	0.10	ug/kg	
123-91-1	1,4-Dioxane	ND	130	90	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

3.2
3

Client Sample ID:	B-CCI-017(23')	Date Sampled:	07/31/09
Lab Sample ID:	JA24443-2	Date Received:	07/31/09
Matrix:	SO - Soil	Percent Solids:	92.3
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	1.0	0.39	ug/kg	
76-13-1	Freon 113	ND	5.2	0.59	ug/kg	
591-78-6	2-Hexanone	ND	5.2	1.0	ug/kg	
98-82-8	Isopropylbenzene	ND	5.2	0.54	ug/kg	
79-20-9	Methyl Acetate	ND	5.2	0.86	ug/kg	
108-87-2	Methylcyclohexane	ND	5.2	0.68	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	0.44	1.0	0.29	ug/kg	J
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	5.2	0.84	ug/kg	
75-09-2	Methylene chloride	ND	5.2	0.23	ug/kg	
103-65-1	n-Propylbenzene	ND	5.2	0.27	ug/kg	
100-42-5	Styrene	ND	5.2	0.11	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	5.2	0.31	ug/kg	
127-18-4	Tetrachloroethene	ND	5.2	0.15	ug/kg	
108-88-3	Toluene	ND	1.0	0.30	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	5.2	0.36	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	5.2	0.13	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	5.2	0.19	ug/kg	
79-01-6	Trichloroethene	ND	5.2	0.55	ug/kg	
75-69-4	Trichlorofluoromethane	ND	5.2	0.24	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	ND	5.2	0.45	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	ND	5.2	0.37	ug/kg	
75-01-4	Vinyl chloride	ND	5.2	0.19	ug/kg	
	m,p-Xylene	ND	2.1	0.49	ug/kg	
95-47-6	o-Xylene	ND	1.0	0.49	ug/kg	
1330-20-7	Xylene (total)	ND	2.1	0.49	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	106%		67-127%
17060-07-0	1,2-Dichloroethane-D4	107%		65-132%
2037-26-5	Toluene-D8	98%		74-129%
460-00-4	4-Bromofluorobenzene	76%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 2

Client Sample ID:	B-CCI-018(26')	Date Sampled:	07/31/09
Lab Sample ID:	JA24443-3	Date Received:	07/31/09
Matrix:	SO - Soil	Percent Solids:	96.2
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	X94202A.D	1	07/31/09	JTP	n/a	n/a	VX3946
Run #2							

Run #	Initial Weight
Run #1	4.8 g
Run #2	

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	11.5	11	2.4	ug/kg	
71-43-2	Benzene	0.67	1.1	0.37	ug/kg	J
75-27-4	Bromodichloromethane	ND	5.4	0.28	ug/kg	
75-25-2	Bromoform	ND	5.4	0.16	ug/kg	
74-83-9	Bromomethane	ND	5.4	0.44	ug/kg	
78-93-3	2-Butanone (MEK)	ND	11	2.1	ug/kg	
104-51-8	n-Butylbenzene	8.1	5.4	0.41	ug/kg	
135-98-8	sec-Butylbenzene	3.5	5.4	0.53	ug/kg	J
98-06-6	tert-Butylbenzene	ND	5.4	0.51	ug/kg	
75-15-0	Carbon disulfide	ND	5.4	0.33	ug/kg	
56-23-5	Carbon tetrachloride	ND	5.4	0.60	ug/kg	
108-90-7	Chlorobenzene	ND	5.4	0.37	ug/kg	
75-00-3	Chloroethane	ND	5.4	1.2	ug/kg	
67-66-3	Chloroform	ND	5.4	0.34	ug/kg	
74-87-3	Chloromethane	ND	5.4	0.18	ug/kg	
110-82-7	Cyclohexane	1.4	5.4	0.16	ug/kg	J
96-12-8	1,2-Dibromo-3-chloropropane	ND	11	0.59	ug/kg	
124-48-1	Dibromochloromethane	ND	5.4	0.12	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.1	0.15	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	5.4	0.29	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	5.4	0.30	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	5.4	0.36	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	5.4	1.0	ug/kg	
75-34-3	1,1-Dichloroethane	ND	5.4	0.15	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.1	0.37	ug/kg	
75-35-4	1,1-Dichloroethene	ND	5.4	0.72	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	5.4	0.26	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	5.4	0.49	ug/kg	
78-87-5	1,2-Dichloropropane	ND	5.4	0.14	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	5.4	0.14	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	5.4	0.10	ug/kg	
123-91-1	1,4-Dioxane	ND	140	94	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-018(26')	Date Sampled:	07/31/09
Lab Sample ID:	JA24443-3	Date Received:	07/31/09
Matrix:	SO - Soil	Percent Solids:	96.2
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	12.2	1.1	0.40	ug/kg	
76-13-1	Freon 113	ND	5.4	0.61	ug/kg	
591-78-6	2-Hexanone	ND	5.4	1.0	ug/kg	
98-82-8	Isopropylbenzene	4.3	5.4	0.56	ug/kg	J
79-20-9	Methyl Acetate	ND	5.4	0.89	ug/kg	
108-87-2	Methylcyclohexane	23.8	5.4	0.71	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	29.7	1.1	0.31	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	5.4	0.88	ug/kg	
75-09-2	Methylene chloride	ND	5.4	0.24	ug/kg	
103-65-1	n-Propylbenzene	13.5	5.4	0.28	ug/kg	
100-42-5	Styrene	ND	5.4	0.12	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	5.4	0.32	ug/kg	
127-18-4	Tetrachloroethene	ND	5.4	0.16	ug/kg	
108-88-3	Toluene	10.2	1.1	0.32	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	5.4	0.37	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	5.4	0.14	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	5.4	0.20	ug/kg	
79-01-6	Trichloroethene	ND	5.4	0.57	ug/kg	
75-69-4	Trichlorofluoromethane	ND	5.4	0.25	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	117	5.4	0.47	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	41.5	5.4	0.39	ug/kg	
75-01-4	Vinyl chloride	ND	5.4	0.19	ug/kg	
	m,p-Xylene	60.9	2.2	0.51	ug/kg	
95-47-6	o-Xylene	32.6	1.1	0.51	ug/kg	
1330-20-7	Xylene (total)	93.5	2.2	0.51	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	97%		67-127%
17060-07-0	1,2-Dichloroethane-D4	95%		65-132%
2037-26-5	Toluene-D8	98%		74-129%
460-00-4	4-Bromofluorobenzene	86%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

3.4
3

Client Sample ID:	B-CCI-019(26')	Date Sampled:	07/31/09
Lab Sample ID:	JA24443-4	Date Received:	07/31/09
Matrix:	SO - Soil	Percent Solids:	99.4
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	V93030.D	1	07/31/09	JLI	n/a	n/a	VV3824
Run #2 ^a	3C53249.D	1	07/31/09	NDJ	n/a	n/a	V3C2329

Run #	Initial Weight	Final Volume	Methanol Aliquot
Run #1	1.1 g		
Run #2	11.0 g	10.0 ml	100 ul

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	46	10	ug/kg	J
71-43-2	Benzene	ND	4.6	1.6	ug/kg	
75-27-4	Bromodichloromethane	ND	23	1.2	ug/kg	
75-25-2	Bromoform	ND	23	0.69	ug/kg	
74-83-9	Bromomethane	ND	23	1.8	ug/kg	
78-93-3	2-Butanone (MEK)	ND	46	9.0	ug/kg	
104-51-8	n-Butylbenzene	52.1	23	1.7	ug/kg	
135-98-8	sec-Butylbenzene	185	23	2.2	ug/kg	
98-06-6	tert-Butylbenzene	19.8	23	2.2	ug/kg	J
75-15-0	Carbon disulfide	ND	23	1.4	ug/kg	
56-23-5	Carbon tetrachloride	ND	23	2.5	ug/kg	
108-90-7	Chlorobenzene	ND	23	1.6	ug/kg	
75-00-3	Chloroethane	ND	23	5.2	ug/kg	
67-66-3	Chloroform	ND	23	1.5	ug/kg	
74-87-3	Chloromethane	ND	23	0.75	ug/kg	
110-82-7	Cyclohexane	ND	23	0.69	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	46	2.5	ug/kg	
124-48-1	Dibromochloromethane	ND	23	0.50	ug/kg	
106-93-4	1,2-Dibromoethane	ND	4.6	0.63	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	23	1.2	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	23	1.3	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	23	1.5	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	23	4.3	ug/kg	
75-34-3	1,1-Dichloroethane	ND	23	0.63	ug/kg	
107-06-2	1,2-Dichloroethane	ND	4.6	1.6	ug/kg	
75-35-4	1,1-Dichloroethene	ND	23	3.0	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	23	1.1	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	23	2.1	ug/kg	
78-87-5	1,2-Dichloropropane	ND	23	0.59	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	23	0.61	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	23	0.44	ug/kg	
123-91-1	1,4-Dioxane	ND	570	400	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

3.4
3

Client Sample ID:	B-CCI-019(26')	Date Sampled:	07/31/09
Lab Sample ID:	JA24443-4	Date Received:	07/31/09
Matrix:	SO - Soil	Percent Solids:	99.4
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	5.3	4.6	1.7	ug/kg	J
76-13-1	Freon 113	ND	23	2.6	ug/kg	
591-78-6	2-Hexanone	ND	23	4.4	ug/kg	
98-82-8	Isopropylbenzene	34.8	23	2.4	ug/kg	
79-20-9	Methyl Acetate	ND	23	3.8	ug/kg	
108-87-2	Methylcyclohexane	ND	23	3.0	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	ND	4.6	1.3	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	23	3.7	ug/kg	
75-09-2	Methylene chloride	ND	23	1.0	ug/kg	
103-65-1	n-Propylbenzene	53.0	23	1.2	ug/kg	
100-42-5	Styrene	ND	23	0.49	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	23	1.3	ug/kg	
127-18-4	Tetrachloroethene	ND	23	0.66	ug/kg	
108-88-3	Toluene	ND	4.6	1.3	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	23	1.6	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	23	0.59	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	23	0.85	ug/kg	
79-01-6	Trichloroethene	ND	23	2.4	ug/kg	
75-69-4	Trichlorofluoromethane	ND	23	1.0	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	45.1	23	2.0	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	14.7	23	1.6	ug/kg	J
75-01-4	Vinyl chloride	ND	23	0.81	ug/kg	
	m,p-Xylene	14.5	9.1	2.1	ug/kg	
95-47-6	o-Xylene	6.6	4.6	2.1	ug/kg	
1330-20-7	Xylene (total)	21.1	9.1	2.1	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	92%	99%	67-127%
17060-07-0	1,2-Dichloroethane-D4	83%	106%	65-132%
2037-26-5	Toluene-D8	97%	100%	74-129%
460-00-4	4-Bromofluorobenzene	57% ^b	82%	62-138%

- (a) Confirmation run for surrogate recoveries.
- (b) Outside control limits due to matrix interference.

ND = Not detected MDL - Method Detection Limit J = Indicates an estimated value
 RL = Reporting Limit B = Indicates analyte found in associated method blank
 E = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 2

Client Sample ID:	B-CCI-022	Date Sampled:	08/04/09
Lab Sample ID:	JA24692-1	Date Received:	08/04/09
Matrix:	SO - Soil	Percent Solids:	87.0
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	Y90403.D	1	08/05/09	YXC	n/a	n/a	VY3788

Run #1	Initial Weight
Run #2	4.9 g

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	12	2.6	ug/kg	
71-43-2	Benzene	ND	1.2	0.40	ug/kg	
75-27-4	Bromodichloromethane	ND	5.9	0.30	ug/kg	
75-25-2	Bromoform	ND	5.9	0.18	ug/kg	
74-83-9	Bromomethane	ND	5.9	0.47	ug/kg	
78-93-3	2-Butanone (MEK)	ND	12	2.3	ug/kg	
104-51-8	n-Butylbenzene	ND	5.9	0.45	ug/kg	
135-98-8	sec-Butylbenzene	ND	5.9	0.57	ug/kg	
98-06-6	tert-Butylbenzene	ND	5.9	0.56	ug/kg	
75-15-0	Carbon disulfide	ND	5.9	0.36	ug/kg	
56-23-5	Carbon tetrachloride	ND	5.9	0.65	ug/kg	
108-90-7	Chlorobenzene	ND	5.9	0.40	ug/kg	
75-00-3	Chloroethane	ND	5.9	1.3	ug/kg	
67-66-3	Chloroform	ND	5.9	0.37	ug/kg	
74-87-3	Chloromethane	ND	5.9	0.19	ug/kg	
110-82-7	Cyclohexane	ND	5.9	0.18	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	12	0.63	ug/kg	
124-48-1	Dibromochloromethane	ND	5.9	0.13	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.2	0.16	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	5.9	0.32	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	5.9	0.32	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	5.9	0.40	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	5.9	1.1	ug/kg	
75-34-3	1,1-Dichloroethane	ND	5.9	0.16	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.2	0.40	ug/kg	
75-35-4	1,1-Dichloroethene	ND	5.9	0.78	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	5.9	0.28	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	5.9	0.53	ug/kg	
78-87-5	1,2-Dichloropropane	ND	5.9	0.15	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	5.9	0.16	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	5.9	0.11	ug/kg	
123-91-1	1,4-Dioxane	ND	150	100	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-022	Date Sampled:	08/04/09
Lab Sample ID:	JA24692-1	Date Received:	08/04/09
Matrix:	SO - Soil	Percent Solids:	87.0
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	1.2	0.44	ug/kg	
76-13-1	Freon 113	ND	5.9	0.66	ug/kg	
591-78-6	2-Hexanone	ND	5.9	1.1	ug/kg	
98-82-8	Isopropylbenzene	ND	5.9	0.61	ug/kg	
79-20-9	Methyl Acetate	ND	5.9	0.97	ug/kg	
108-87-2	Methylcyclohexane	ND	5.9	0.77	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	ND	1.2	0.33	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	5.9	0.95	ug/kg	
75-09-2	Methylene chloride	ND	5.9	0.26	ug/kg	
103-65-1	n-Propylbenzene	ND	5.9	0.30	ug/kg	
100-42-5	Styrene	ND	5.9	0.13	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	5.9	0.34	ug/kg	
127-18-4	Tetrachloroethene	ND	5.9	0.17	ug/kg	
108-88-3	Toluene	ND	1.2	0.34	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	5.9	0.40	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	5.9	0.15	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	5.9	0.22	ug/kg	
79-01-6	Trichloroethene	ND	5.9	0.62	ug/kg	
75-69-4	Trichlorofluoromethane	ND	5.9	0.27	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	ND	5.9	0.50	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	ND	5.9	0.42	ug/kg	
75-01-4	Vinyl chloride	ND	5.9	0.21	ug/kg	
	m,p-Xylene	ND	2.3	0.55	ug/kg	
95-47-6	o-Xylene	ND	1.2	0.55	ug/kg	
1330-20-7	Xylene (total)	ND	2.3	0.55	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	96%		67-127%
17060-07-0	1,2-Dichloroethane-D4	90%		65-132%
2037-26-5	Toluene-D8	102%		74-129%
460-00-4	4-Bromofluorobenzene	94%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 2

Client Sample ID:	B-CCI-023	Date Sampled:	08/04/09
Lab Sample ID:	JA24692-2	Date Received:	08/04/09
Matrix:	SO - Soil	Percent Solids:	84.1
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	Y90404.D	1	08/05/09	YXC	n/a	n/a	VY3788

Run #1	Initial Weight
Run #2	4.8 g

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	12	2.8	ug/kg	
71-43-2	Benzene	ND	1.2	0.42	ug/kg	
75-27-4	Bromodichloromethane	ND	6.2	0.32	ug/kg	
75-25-2	Bromoform	ND	6.2	0.19	ug/kg	
74-83-9	Bromomethane	ND	6.2	0.50	ug/kg	
78-93-3	2-Butanone (MEK)	ND	12	2.4	ug/kg	
104-51-8	n-Butylbenzene	ND	6.2	0.47	ug/kg	
135-98-8	sec-Butylbenzene	ND	6.2	0.60	ug/kg	
98-06-6	tert-Butylbenzene	ND	6.2	0.59	ug/kg	
75-15-0	Carbon disulfide	ND	6.2	0.38	ug/kg	
56-23-5	Carbon tetrachloride	ND	6.2	0.69	ug/kg	
108-90-7	Chlorobenzene	ND	6.2	0.42	ug/kg	
75-00-3	Chloroethane	ND	6.2	1.4	ug/kg	
67-66-3	Chloroform	ND	6.2	0.39	ug/kg	
74-87-3	Chloromethane	ND	6.2	0.20	ug/kg	
110-82-7	Cyclohexane	ND	6.2	0.19	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	12	0.67	ug/kg	
124-48-1	Dibromochloromethane	ND	6.2	0.14	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.2	0.17	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	6.2	0.33	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	6.2	0.34	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	6.2	0.42	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	6.2	1.2	ug/kg	
75-34-3	1,1-Dichloroethane	ND	6.2	0.17	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.2	0.43	ug/kg	
75-35-4	1,1-Dichloroethene	ND	6.2	0.82	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	6.2	0.30	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	6.2	0.56	ug/kg	
78-87-5	1,2-Dichloropropane	ND	6.2	0.16	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	6.2	0.16	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	6.2	0.12	ug/kg	
123-91-1	1,4-Dioxane	ND	150	110	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-023	Date Sampled:	08/04/09
Lab Sample ID:	JA24692-2	Date Received:	08/04/09
Matrix:	SO - Soil	Percent Solids:	84.1
Method:	SW846 8260B	Project: Courtlandt Corners I, 364 East 161st Street, Bronx, NY	

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	1.2	0.46	ug/kg	
76-13-1	Freon 113	ND	6.2	0.70	ug/kg	
591-78-6	2-Hexanone	ND	6.2	1.2	ug/kg	
98-82-8	Isopropylbenzene	ND	6.2	0.64	ug/kg	
79-20-9	Methyl Acetate	ND	6.2	1.0	ug/kg	
108-87-2	Methylcyclohexane	ND	6.2	0.81	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	ND	1.2	0.35	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	6.2	1.0	ug/kg	
75-09-2	Methylene chloride	ND	6.2	0.28	ug/kg	
103-65-1	n-Propylbenzene	ND	6.2	0.32	ug/kg	
100-42-5	Styrene	ND	6.2	0.13	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	6.2	0.36	ug/kg	
127-18-4	Tetrachloroethene	ND	6.2	0.18	ug/kg	
108-88-3	Toluene	ND	1.2	0.36	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	6.2	0.43	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	6.2	0.16	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	6.2	0.23	ug/kg	
79-01-6	Trichloroethene	ND	6.2	0.65	ug/kg	
75-69-4	Trichlorofluoromethane	ND	6.2	0.28	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	ND	6.2	0.53	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	ND	6.2	0.44	ug/kg	
75-01-4	Vinyl chloride	ND	6.2	0.22	ug/kg	
	m,p-Xylene	ND	2.5	0.58	ug/kg	
95-47-6	o-Xylene	ND	1.2	0.58	ug/kg	
1330-20-7	Xylene (total)	ND	2.5	0.58	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	97%		67-127%
17060-07-0	1,2-Dichloroethane-D4	93%		65-132%
2037-26-5	Toluene-D8	102%		74-129%
460-00-4	4-Bromofluorobenzene	93%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 2

Client Sample ID:	B-CCI-024	Date Sampled:	08/04/09
Lab Sample ID:	JA24692-3	Date Received:	08/04/09
Matrix:	SO - Soil	Percent Solids:	89.0
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	Y90405.D	1	08/05/09	YXC	n/a	n/a	VY3788
Run #2							

Run #	Initial Weight
Run #1	4.5 g
Run #2	

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	12	2.8	ug/kg	
71-43-2	Benzene	ND	1.2	0.43	ug/kg	
75-27-4	Bromodichloromethane	ND	6.2	0.32	ug/kg	
75-25-2	Bromoform	ND	6.2	0.19	ug/kg	
74-83-9	Bromomethane	ND	6.2	0.50	ug/kg	
78-93-3	2-Butanone (MEK)	ND	12	2.5	ug/kg	
104-51-8	n-Butylbenzene	ND	6.2	0.47	ug/kg	
135-98-8	sec-Butylbenzene	ND	6.2	0.61	ug/kg	
98-06-6	tert-Butylbenzene	ND	6.2	0.59	ug/kg	
75-15-0	Carbon disulfide	ND	6.2	0.38	ug/kg	
56-23-5	Carbon tetrachloride	ND	6.2	0.69	ug/kg	
108-90-7	Chlorobenzene	ND	6.2	0.42	ug/kg	
75-00-3	Chloroethane	ND	6.2	1.4	ug/kg	
67-66-3	Chloroform	ND	6.2	0.40	ug/kg	
74-87-3	Chloromethane	ND	6.2	0.21	ug/kg	
110-82-7	Cyclohexane	ND	6.2	0.19	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	12	0.68	ug/kg	
124-48-1	Dibromochloromethane	ND	6.2	0.14	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.2	0.17	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	6.2	0.34	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	6.2	0.34	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	6.2	0.42	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	6.2	1.2	ug/kg	
75-34-3	1,1-Dichloroethane	ND	6.2	0.17	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.2	0.43	ug/kg	
75-35-4	1,1-Dichloroethene	ND	6.2	0.83	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	6.2	0.30	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	6.2	0.56	ug/kg	
78-87-5	1,2-Dichloropropane	ND	6.2	0.16	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	6.2	0.17	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	6.2	0.12	ug/kg	
123-91-1	1,4-Dioxane	ND	160	110	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-024	Date Sampled:	08/04/09
Lab Sample ID:	JA24692-3	Date Received:	08/04/09
Matrix:	SO - Soil	Percent Solids:	89.0
Method:	SW846 8260B	Project: Courtlandt Corners I, 364 East 161st Street, Bronx, NY	

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	1.2	0.46	ug/kg	
76-13-1	Freon 113	ND	6.2	0.70	ug/kg	
591-78-6	2-Hexanone	ND	6.2	1.2	ug/kg	
98-82-8	Isopropylbenzene	ND	6.2	0.65	ug/kg	
79-20-9	Methyl Acetate	ND	6.2	1.0	ug/kg	
108-87-2	Methylcyclohexane	ND	6.2	0.82	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	ND	1.2	0.35	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	6.2	1.0	ug/kg	
75-09-2	Methylene chloride	ND	6.2	0.28	ug/kg	
103-65-1	n-Propylbenzene	ND	6.2	0.32	ug/kg	
100-42-5	Styrene	ND	6.2	0.13	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	6.2	0.37	ug/kg	
127-18-4	Tetrachloroethene	ND	6.2	0.18	ug/kg	
108-88-3	Toluene	ND	1.2	0.36	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	6.2	0.43	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	6.2	0.16	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	6.2	0.23	ug/kg	
79-01-6	Trichloroethene	ND	6.2	0.66	ug/kg	
75-69-4	Trichlorofluoromethane	ND	6.2	0.29	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	ND	6.2	0.54	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	ND	6.2	0.45	ug/kg	
75-01-4	Vinyl chloride	ND	6.2	0.22	ug/kg	
	m,p-Xylene	ND	2.5	0.59	ug/kg	
95-47-6	o-Xylene	ND	1.2	0.59	ug/kg	
1330-20-7	Xylene (total)	ND	2.5	0.59	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	95%		67-127%
17060-07-0	1,2-Dichloroethane-D4	87%		65-132%
2037-26-5	Toluene-D8	103%		74-129%
460-00-4	4-Bromofluorobenzene	94%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 2

3.4
3

Client Sample ID:	B-CCI-025	Date Sampled:	08/04/09
Lab Sample ID:	JA24692-4	Date Received:	08/04/09
Matrix:	SO - Soil	Percent Solids:	94.7
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1 ^a	E154494.D	1	08/05/09	OTR	n/a	n/a	VE6808
Run #2							

Run #	Initial Weight	Final Volume	Methanol Aliquot
Run #1	10.1 g	10.0 ml	100 ul
Run #2			

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	550	120	ug/kg	
71-43-2	Benzene	ND	55	19	ug/kg	
75-27-4	Bromodichloromethane	ND	280	14	ug/kg	
75-25-2	Bromoform	ND	280	8.3	ug/kg	
74-83-9	Bromomethane	ND	280	22	ug/kg	
78-93-3	2-Butanone (MEK)	ND	550	110	ug/kg	
104-51-8	n-Butylbenzene	406	280	21	ug/kg	
135-98-8	sec-Butylbenzene	181	280	27	ug/kg	J
98-06-6	tert-Butylbenzene	ND	280	26	ug/kg	
75-15-0	Carbon disulfide	ND	280	17	ug/kg	
56-23-5	Carbon tetrachloride	ND	280	31	ug/kg	
108-90-7	Chlorobenzene	ND	280	19	ug/kg	
75-00-3	Chloroethane	ND	280	63	ug/kg	
67-66-3	Chloroform	ND	280	18	ug/kg	
74-87-3	Chloromethane	ND	280	9.1	ug/kg	
110-82-7	Cyclohexane	ND	280	8.3	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	550	30	ug/kg	
124-48-1	Dibromochloromethane	ND	280	6.1	ug/kg	
106-93-4	1,2-Dibromoethane	ND	55	7.5	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	280	15	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	280	15	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	280	19	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	280	52	ug/kg	
75-34-3	1,1-Dichloroethane	ND	280	7.6	ug/kg	
107-06-2	1,2-Dichloroethane	ND	55	19	ug/kg	
75-35-4	1,1-Dichloroethene	ND	280	36	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	280	13	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	280	25	ug/kg	
78-87-5	1,2-Dichloropropane	ND	280	7.2	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	280	7.3	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	280	5.3	ug/kg	
123-91-1	1,4-Dioxane	ND	6900	4800	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-025	Date Sampled:	08/04/09
Lab Sample ID:	JA24692-4	Date Received:	08/04/09
Matrix:	SO - Soil	Percent Solids:	94.7
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	101	55	20	ug/kg	
76-13-1	Freon 113	ND	280	31	ug/kg	
591-78-6	2-Hexanone	ND	280	53	ug/kg	
98-82-8	Isopropylbenzene	73.0	280	29	ug/kg	J
79-20-9	Methyl Acetate	ND	280	45	ug/kg	
108-87-2	Methylcyclohexane	104	280	36	ug/kg	J
1634-04-4	Methyl Tert Butyl Ether	ND	55	16	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	280	45	ug/kg	
75-09-2	Methylene chloride	ND	280	12	ug/kg	
103-65-1	n-Propylbenzene	232	280	14	ug/kg	J
100-42-5	Styrene	ND	280	5.9	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	280	16	ug/kg	
127-18-4	Tetrachloroethene	ND	280	8.0	ug/kg	
108-88-3	Toluene	19.9	55	16	ug/kg	J
120-82-1	1,2,4-Trichlorobenzene	ND	280	19	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	280	7.0	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	280	10	ug/kg	
79-01-6	Trichloroethene	ND	280	29	ug/kg	
75-69-4	Trichlorofluoromethane	ND	280	13	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	2430	280	24	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	267	280	20	ug/kg	J
75-01-4	Vinyl chloride	ND	280	9.8	ug/kg	
	m,p-Xylene	237	110	26	ug/kg	
95-47-6	o-Xylene	85.1	55	26	ug/kg	
1330-20-7	Xylene (total)	322	110	26	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	107%		67-127%
17060-07-0	1,2-Dichloroethane-D4	115%		65-132%
2037-26-5	Toluene-D8	111%		74-129%
460-00-4	4-Bromofluorobenzene	107%		62-138%

(a) Dilution required due to matrix interference.

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 2

Client Sample ID:	B-CCI-026 (2')	Date Sampled:	08/06/09
Lab Sample ID:	JA25087-1	Date Received:	08/07/09
Matrix:	SO - Soil	Percent Solids:	83.0
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	V93345.D	1	08/09/09	JLI	n/a	n/a	VV3842
Run #2							

Run #	Initial Weight
Run #1	5.3 g
Run #2	

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	11	2.5	ug/kg	
71-43-2	Benzene	ND	1.1	0.39	ug/kg	
75-27-4	Bromodichloromethane	ND	5.7	0.29	ug/kg	
75-25-2	Bromoform	ND	5.7	0.17	ug/kg	
74-83-9	Bromomethane	ND	5.7	0.46	ug/kg	
78-93-3	2-Butanone (MEK)	ND	11	2.2	ug/kg	
104-51-8	n-Butylbenzene	ND	5.7	0.43	ug/kg	
135-98-8	sec-Butylbenzene	ND	5.7	0.55	ug/kg	
98-06-6	tert-Butylbenzene	ND	5.7	0.54	ug/kg	
75-15-0	Carbon disulfide	ND	5.7	0.35	ug/kg	
56-23-5	Carbon tetrachloride	ND	5.7	0.63	ug/kg	
108-90-7	Chlorobenzene	ND	5.7	0.39	ug/kg	
75-00-3	Chloroethane	ND	5.7	1.3	ug/kg	
67-66-3	Chloroform	ND	5.7	0.36	ug/kg	
74-87-3	Chloromethane	ND	5.7	0.19	ug/kg	
110-82-7	Cyclohexane	ND	5.7	0.17	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	11	0.61	ug/kg	
124-48-1	Dibromochloromethane	ND	5.7	0.13	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.1	0.16	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	5.7	0.31	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	5.7	0.31	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	5.7	0.38	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	5.7	1.1	ug/kg	
75-34-3	1,1-Dichloroethane	ND	5.7	0.16	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.1	0.39	ug/kg	
75-35-4	1,1-Dichloroethene	ND	5.7	0.75	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	5.7	0.27	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	5.7	0.51	ug/kg	
78-87-5	1,2-Dichloropropane	ND	5.7	0.15	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	5.7	0.15	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	5.7	0.11	ug/kg	
123-91-1	1,4-Dioxane	ND	140	98	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-026 (2')	Date Sampled:	08/06/09
Lab Sample ID:	JA25087-1	Date Received:	08/07/09
Matrix:	SO - Soil	Percent Solids:	83.0
Method:	SW846 8260B	Project: Courtlandt Corners I, 364 East 161st Street, Bronx, NY	

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	1.1	0.42	ug/kg	
76-13-1	Freon 113	ND	5.7	0.64	ug/kg	
591-78-6	2-Hexanone	ND	5.7	1.1	ug/kg	
98-82-8	Isopropylbenzene	ND	5.7	0.59	ug/kg	
79-20-9	Methyl Acetate	ND	5.7	0.94	ug/kg	
108-87-2	Methylcyclohexane	ND	5.7	0.74	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	ND	1.1	0.32	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	5.7	0.92	ug/kg	
75-09-2	Methylene chloride	ND	5.7	0.25	ug/kg	
103-65-1	n-Propylbenzene	ND	5.7	0.29	ug/kg	
100-42-5	Styrene	ND	5.7	0.12	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	5.7	0.33	ug/kg	
127-18-4	Tetrachloroethene	ND	5.7	0.16	ug/kg	
108-88-3	Toluene	ND	1.1	0.33	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	5.7	0.39	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	5.7	0.15	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	5.7	0.21	ug/kg	
79-01-6	Trichloroethene	ND	5.7	0.60	ug/kg	
75-69-4	Trichlorofluoromethane	ND	5.7	0.26	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	ND	5.7	0.49	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	ND	5.7	0.41	ug/kg	
75-01-4	Vinyl chloride	ND	5.7	0.20	ug/kg	
	m,p-Xylene	ND	2.3	0.53	ug/kg	
95-47-6	o-Xylene	ND	1.1	0.53	ug/kg	
1330-20-7	Xylene (total)	ND	2.3	0.53	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	87%		67-127%
17060-07-0	1,2-Dichloroethane-D4	78%		65-132%
2037-26-5	Toluene-D8	86%		74-129%
460-00-4	4-Bromofluorobenzene	66%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 2

Client Sample ID:	B-CCI-027 (2')	Date Sampled:	08/06/09
Lab Sample ID:	JA25087-2	Date Received:	08/07/09
Matrix:	SO - Soil	Percent Solids:	88.1
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	V93346.D	1	08/09/09	JLI	n/a	n/a	VV3842

Run #1	Initial Weight
Run #2	4.6 g

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	12	2.8	ug/kg	
71-43-2	Benzene	ND	1.2	0.42	ug/kg	
75-27-4	Bromodichloromethane	ND	6.2	0.32	ug/kg	
75-25-2	Bromoform	ND	6.2	0.19	ug/kg	
74-83-9	Bromomethane	ND	6.2	0.50	ug/kg	
78-93-3	2-Butanone (MEK)	ND	12	2.4	ug/kg	
104-51-8	n-Butylbenzene	ND	6.2	0.47	ug/kg	
135-98-8	sec-Butylbenzene	ND	6.2	0.60	ug/kg	
98-06-6	tert-Butylbenzene	ND	6.2	0.59	ug/kg	
75-15-0	Carbon disulfide	ND	6.2	0.38	ug/kg	
56-23-5	Carbon tetrachloride	ND	6.2	0.68	ug/kg	
108-90-7	Chlorobenzene	ND	6.2	0.42	ug/kg	
75-00-3	Chloroethane	ND	6.2	1.4	ug/kg	
67-66-3	Chloroform	ND	6.2	0.39	ug/kg	
74-87-3	Chloromethane	ND	6.2	0.20	ug/kg	
110-82-7	Cyclohexane	ND	6.2	0.19	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	12	0.67	ug/kg	
124-48-1	Dibromochloromethane	ND	6.2	0.14	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.2	0.17	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	6.2	0.33	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	6.2	0.34	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	6.2	0.42	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	6.2	1.2	ug/kg	
75-34-3	1,1-Dichloroethane	ND	6.2	0.17	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.2	0.43	ug/kg	
75-35-4	1,1-Dichloroethene	ND	6.2	0.82	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	6.2	0.29	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	6.2	0.55	ug/kg	
78-87-5	1,2-Dichloropropane	ND	6.2	0.16	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	6.2	0.16	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	6.2	0.12	ug/kg	
123-91-1	1,4-Dioxane	ND	150	110	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-027 (2')	Date Sampled:	08/06/09
Lab Sample ID:	JA25087-2	Date Received:	08/07/09
Matrix:	SO - Soil	Percent Solids:	88.1
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	1.2	0.46	ug/kg	
76-13-1	Freon 113	ND	6.2	0.69	ug/kg	
591-78-6	2-Hexanone	ND	6.2	1.2	ug/kg	
98-82-8	Isopropylbenzene	ND	6.2	0.64	ug/kg	
79-20-9	Methyl Acetate	ND	6.2	1.0	ug/kg	
108-87-2	Methylcyclohexane	ND	6.2	0.81	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	ND	1.2	0.35	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	6.2	1.0	ug/kg	
75-09-2	Methylene chloride	ND	6.2	0.28	ug/kg	
103-65-1	n-Propylbenzene	ND	6.2	0.31	ug/kg	
100-42-5	Styrene	ND	6.2	0.13	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	6.2	0.36	ug/kg	
127-18-4	Tetrachloroethene	ND	6.2	0.18	ug/kg	
108-88-3	Toluene	ND	1.2	0.36	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	6.2	0.43	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	6.2	0.16	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	6.2	0.23	ug/kg	
79-01-6	Trichloroethene	ND	6.2	0.65	ug/kg	
75-69-4	Trichlorofluoromethane	ND	6.2	0.28	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	ND	6.2	0.53	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	ND	6.2	0.44	ug/kg	
75-01-4	Vinyl chloride	ND	6.2	0.22	ug/kg	
	m,p-Xylene	ND	2.5	0.58	ug/kg	
95-47-6	o-Xylene	ND	1.2	0.58	ug/kg	
1330-20-7	Xylene (total)	ND	2.5	0.58	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	89%		67-127%
17060-07-0	1,2-Dichloroethane-D4	77%		65-132%
2037-26-5	Toluene-D8	85%		74-129%
460-00-4	4-Bromofluorobenzene	65%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 2

Client Sample ID:	B-CCI-028(2')	Date Sampled:	08/11/09
Lab Sample ID:	JA25446-1	Date Received:	08/12/09
Matrix:	SO - Soil	Percent Solids:	79.6
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	G122712.D	1	08/13/09	SJM	n/a	n/a	VG5851

Run #1	Initial Weight
Run #2	5.0 g

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	13	2.8	ug/kg	
71-43-2	Benzene	ND	1.3	0.43	ug/kg	
75-27-4	Bromodichloromethane	ND	6.3	0.32	ug/kg	
75-25-2	Bromoform	ND	6.3	0.19	ug/kg	
74-83-9	Bromomethane	ND	6.3	0.51	ug/kg	
78-93-3	2-Butanone (MEK)	ND	13	2.5	ug/kg	J
104-51-8	n-Butylbenzene	ND	6.3	0.48	ug/kg	
135-98-8	sec-Butylbenzene	ND	6.3	0.61	ug/kg	
98-06-6	tert-Butylbenzene	ND	6.3	0.60	ug/kg	
75-15-0	Carbon disulfide	ND	6.3	0.38	ug/kg	
56-23-5	Carbon tetrachloride	ND	6.3	0.70	ug/kg	
108-90-7	Chlorobenzene	ND	6.3	0.43	ug/kg	
75-00-3	Chloroethane	ND	6.3	1.4	ug/kg	
67-66-3	Chloroform	ND	6.3	0.40	ug/kg	
74-87-3	Chloromethane	ND	6.3	0.21	ug/kg	J
110-82-7	Cyclohexane	ND	6.3	0.19	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	13	0.68	ug/kg	
124-48-1	Dibromochloromethane	ND	6.3	0.14	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.3	0.17	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	6.3	0.34	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	6.3	0.35	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	6.3	0.42	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	6.3	1.2	ug/kg	
75-34-3	1,1-Dichloroethane	ND	6.3	0.17	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.3	0.43	ug/kg	
75-35-4	1,1-Dichloroethene	ND	6.3	0.83	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	6.3	0.30	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	6.3	0.56	ug/kg	
78-87-5	1,2-Dichloropropane	ND	6.3	0.16	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	6.3	0.17	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	6.3	0.12	ug/kg	
123-91-1	1,4-Dioxane	ND	160	110	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-028(2')	Date Sampled:	08/11/09
Lab Sample ID:	JA25446-1	Date Received:	08/12/09
Matrix:	SO - Soil	Percent Solids:	79.6
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	1.3	0.47	ug/kg	
76-13-1	Freon 113	ND	6.3	0.71	ug/kg	
591-78-6	2-Hexanone	ND	6.3	1.2	ug/kg	
98-82-8	Isopropylbenzene	ND	6.3	0.65	ug/kg	
79-20-9	Methyl Acetate	ND	6.3	1.0	ug/kg	
108-87-2	Methylcyclohexane	ND	6.3	0.82	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	ND	1.3	0.35	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	6.3	1.0	ug/kg	
75-09-2	Methylene chloride	ND	6.3	0.28	ug/kg	
103-65-1	n-Propylbenzene	ND	6.3	0.32	ug/kg	
100-42-5	Styrene	ND	6.3	0.13	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	6.3	0.37	ug/kg	
127-18-4	Tetrachloroethene	ND	6.3	0.18	ug/kg	
108-88-3	Toluene	ND	1.3	0.37	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	6.3	0.43	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	6.3	0.16	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	6.3	0.23	ug/kg	
79-01-6	Trichloroethene	ND	6.3	0.66	ug/kg	
75-69-4	Trichlorofluoromethane	ND	6.3	0.29	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	ND	6.3	0.54	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	ND	6.3	0.45	ug/kg	
75-01-4	Vinyl chloride	ND	6.3	0.22	ug/kg	J
	m,p-Xylene	ND	2.5	0.59	ug/kg	
95-47-6	o-Xylene	ND	1.3	0.59	ug/kg	
1330-20-7	Xylene (total)	ND	2.5	0.59	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	96%		67-127%
17060-07-0	1,2-Dichloroethane-D4	89%		65-132%
2037-26-5	Toluene-D8	105%		74-129%
460-00-4	4-Bromofluorobenzene	100%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 2

Client Sample ID:	B-CCI-029(20')	Date Sampled:	08/12/09
Lab Sample ID:	JA25446-2	Date Received:	08/12/09
Matrix:	SO - Soil	Percent Solids:	79.4
Method:	SW846 8260B	Project: Courtlandt Corners I, 364 East 161st Street, Bronx, NY	

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	G122715.D	1	08/13/09	SJM	n/a	n/a	VG5851
Run #2							

Run #	Initial Weight
Run #1	4.1 g
Run #2	

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	15	3.4	ug/kg	
71-43-2	Benzene	7.1	1.5	0.52	ug/kg	
75-27-4	Bromodichloromethane	ND	7.7	0.39	ug/kg	
75-25-2	Bromoform	ND	7.7	0.23	ug/kg	
74-83-9	Bromomethane	ND	7.7	0.62	ug/kg	
78-93-3	2-Butanone (MEK)	ND	15	3.0	ug/kg	
104-51-8	n-Butylbenzene	ND	7.7	0.58	ug/kg	
135-98-8	sec-Butylbenzene	4.8	7.7	0.75	ug/kg	J
98-06-6	tert-Butylbenzene	ND	7.7	0.73	ug/kg	
75-15-0	Carbon disulfide	ND	7.7	0.47	ug/kg	
56-23-5	Carbon tetrachloride	ND	7.7	0.85	ug/kg	
108-90-7	Chlorobenzene	ND	7.7	0.52	ug/kg	
75-00-3	Chloroethane	ND	7.7	1.8	ug/kg	
67-66-3	Chloroform	ND	7.7	0.49	ug/kg	
74-87-3	Chloromethane	ND	7.7	0.25	ug/kg	
110-82-7	Cyclohexane	ND	7.7	0.23	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	15	0.83	ug/kg	
124-48-1	Dibromochloromethane	ND	7.7	0.17	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.5	0.21	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	7.7	0.41	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	7.7	0.42	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	7.7	0.52	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	7.7	1.5	ug/kg	
75-34-3	1,1-Dichloroethane	ND	7.7	0.21	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.5	0.53	ug/kg	
75-35-4	1,1-Dichloroethene	ND	7.7	1.0	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	7.7	0.37	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	7.7	0.69	ug/kg	
78-87-5	1,2-Dichloropropane	ND	7.7	0.20	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	7.7	0.20	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	7.7	0.15	ug/kg	
123-91-1	1,4-Dioxane	ND	190	130	ug/kg	

ND = Not detected MDL - Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-029(20')	Date Sampled:	08/12/09
Lab Sample ID:	JA25446-2	Date Received:	08/12/09
Matrix:	SO - Soil	Percent Solids:	79.4
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	5.5	1.5	0.57	ug/kg	
76-13-1	Freon 113	ND	7.7	0.86	ug/kg	
591-78-6	2-Hexanone	ND	7.7	1.5	ug/kg	
98-82-8	Isopropylbenzene	7.2	7.7	0.80	ug/kg	J
79-20-9	Methyl Acetate	ND	7.7	1.3	ug/kg	
108-87-2	Methylcyclohexane	90.6	7.7	1.0	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	2.5	1.5	0.43	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	7.7	1.2	ug/kg	
75-09-2	Methylene chloride	ND	7.7	0.34	ug/kg	
103-65-1	n-Propylbenzene	6.9	7.7	0.39	ug/kg	J
100-42-5	Styrene	ND	7.7	0.16	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	7.7	0.45	ug/kg	
127-18-4	Tetrachloroethene	ND	7.7	0.22	ug/kg	
108-88-3	Toluene	12.2	1.5	0.45	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	7.7	0.53	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	7.7	0.20	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	7.7	0.28	ug/kg	
79-01-6	Trichloroethene	ND	7.7	0.81	ug/kg	
75-69-4	Trichlorofluoromethane	ND	7.7	0.35	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	153	7.7	0.66	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	117	7.7	0.55	ug/kg	
75-01-4	Vinyl chloride	ND	7.7	0.27	ug/kg	
	m,p-Xylene	86.0	3.1	0.72	ug/kg	
95-47-6	o-Xylene	65.0	1.5	0.72	ug/kg	
1330-20-7	Xylene (total)	151	3.1	0.72	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	93%		67-127%
17060-07-0	1,2-Dichloroethane-D4	84%		65-132%
2037-26-5	Toluene-D8	98%		74-129%
460-00-4	4-Bromofluorobenzene	105%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 2

Client Sample ID:	B-CCI-030(4')	Date Sampled:	08/12/09
Lab Sample ID:	JA25446-3	Date Received:	08/12/09
Matrix:	SO - Soil	Percent Solids:	89.1
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	G122714.D	1	08/13/09	SJM	n/a	n/a	VG5851

Run #1	Initial Weight
Run #2	4.2 g

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	13	3.0	ug/kg	
71-43-2	Benzene	ND	1.3	0.46	ug/kg	
75-27-4	Bromodichloromethane	ND	6.7	0.34	ug/kg	
75-25-2	Bromoform	ND	6.7	0.20	ug/kg	
74-83-9	Bromomethane	ND	6.7	0.54	ug/kg	
78-93-3	2-Butanone (MEK)	ND	13	2.6	ug/kg	
104-51-8	n-Butylbenzene	ND	6.7	0.51	ug/kg	
135-98-8	sec-Butylbenzene	ND	6.7	0.65	ug/kg	
98-06-6	tert-Butylbenzene	ND	6.7	0.63	ug/kg	
75-15-0	Carbon disulfide	ND	6.7	0.41	ug/kg	
56-23-5	Carbon tetrachloride	ND	6.7	0.74	ug/kg	
108-90-7	Chlorobenzene	ND	6.7	0.45	ug/kg	
75-00-3	Chloroethane	ND	6.7	1.5	ug/kg	
67-66-3	Chloroform	ND	6.7	0.42	ug/kg	
74-87-3	Chloromethane	ND	6.7	0.22	ug/kg	
110-82-7	Cyclohexane	ND	6.7	0.20	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	13	0.72	ug/kg	
124-48-1	Dibromochloromethane	ND	6.7	0.15	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.3	0.18	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	6.7	0.36	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	6.7	0.37	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	6.7	0.45	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	6.7	1.3	ug/kg	
75-34-3	1,1-Dichloroethane	ND	6.7	0.18	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.3	0.46	ug/kg	
75-35-4	1,1-Dichloroethene	ND	6.7	0.88	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	6.7	0.32	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	6.7	0.60	ug/kg	
78-87-5	1,2-Dichloropropane	ND	6.7	0.17	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	6.7	0.18	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	6.7	0.13	ug/kg	
123-91-1	1,4-Dioxane	ND	170	120	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-030(4')	Date Sampled:	08/12/09
Lab Sample ID:	JA25446-3	Date Received:	08/12/09
Matrix:	SO - Soil	Percent Solids:	89.1
Method:	SW846 8260B	Project: Courtlandt Corners I, 364 East 161st Street, Bronx, NY	

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	1.3	0.50	ug/kg	
76-13-1	Freon 113	ND	6.7	0.75	ug/kg	
591-78-6	2-Hexanone	ND	6.7	1.3	ug/kg	
98-82-8	Isopropylbenzene	ND	6.7	0.69	ug/kg	
79-20-9	Methyl Acetate	ND	6.7	1.1	ug/kg	
108-87-2	Methylcyclohexane	29.0	6.7	0.87	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	ND	1.3	0.38	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	6.7	1.1	ug/kg	
75-09-2	Methylene chloride	ND	6.7	0.30	ug/kg	
103-65-1	n-Propylbenzene	ND	6.7	0.34	ug/kg	
100-42-5	Styrene	ND	6.7	0.14	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	6.7	0.39	ug/kg	
127-18-4	Tetrachloroethene	ND	6.7	0.19	ug/kg	
108-88-3	Toluene	3.2	1.3	0.39	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	6.7	0.46	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	6.7	0.17	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	6.7	0.25	ug/kg	
79-01-6	Trichloroethene	ND	6.7	0.70	ug/kg	
75-69-4	Trichlorofluoromethane	ND	6.7	0.31	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	21.2	6.7	0.57	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	52.3	6.7	0.48	ug/kg	
75-01-4	Vinyl chloride	ND	6.7	0.24	ug/kg	
	m,p-Xylene	21.7	2.7	0.63	ug/kg	
95-47-6	o-Xylene	30.9	1.3	0.63	ug/kg	
1330-20-7	Xylene (total)	52.6	2.7	0.63	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	95%		67-127%
17060-07-0	1,2-Dichloroethane-D4	88%		65-132%
2037-26-5	Toluene-D8	102%		74-129%
460-00-4	4-Bromofluorobenzene	94%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 2

3.4
3

Client Sample ID:	B-CCI-031(10')	Date Sampled:	08/12/09
Lab Sample ID:	JA25446-4	Date Received:	08/12/09
Matrix:	SO - Soil	Percent Solids:	89.4
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	G122713.D	1	08/13/09	SJM	n/a	n/a	VC5851
Run #2							

Run #1	Initial Weight
Run #1	4.2 g
Run #2	

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	13	3.0	ug/kg	
71-43-2	Benzene	ND	1.3	0.45	ug/kg	
75-27-4	Bromodichloromethane	ND	6.7	0.34	ug/kg	
75-25-2	Bromoform	ND	6.7	0.20	ug/kg	
74-83-9	Bromomethane	ND	6.7	0.54	ug/kg	
78-93-3	2-Butanone (MEK)	ND	13	2.6	ug/kg	
104-51-8	n-Butylbenzene	ND	6.7	0.51	ug/kg	
135-98-8	sec-Butylbenzene	ND	6.7	0.65	ug/kg	
98-06-6	tert-Butylbenzene	ND	6.7	0.63	ug/kg	
75-15-0	Carbon disulfide	ND	6.7	0.41	ug/kg	
56-23-5	Carbon tetrachloride	ND	6.7	0.74	ug/kg	
108-90-7	Chlorobenzene	ND	6.7	0.45	ug/kg	
75-00-3	Chloroethane	ND	6.7	1.5	ug/kg	
67-66-3	Chloroform	ND	6.7	0.42	ug/kg	
74-87-3	Chloromethane	ND	6.7	0.22	ug/kg	
110-82-7	Cyclohexane	ND	6.7	0.20	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	13	0.72	ug/kg	
124-48-1	Dibromochloromethane	ND	6.7	0.15	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.3	0.18	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	6.7	0.36	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	6.7	0.37	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	6.7	0.45	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	6.7	1.3	ug/kg	
75-34-3	1,1-Dichloroethane	ND	6.7	0.18	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.3	0.46	ug/kg	
75-35-4	1,1-Dichloroethene	ND	6.7	0.88	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	6.7	0.32	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	6.7	0.60	ug/kg	
78-87-5	1,2-Dichloropropane	ND	6.7	0.17	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	6.7	0.18	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	6.7	0.13	ug/kg	
123-91-1	1,4-Dioxane	ND	170	120	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-031(10')	Date Sampled:	08/12/09
Lab Sample ID:	JA25446-4	Date Received:	08/12/09
Matrix:	SO - Soil	Percent Solids:	89.4
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	1.3	0.49	ug/kg	
76-13-1	Freon 113	ND	6.7	0.75	ug/kg	
591-78-6	2-Hexanone	ND	6.7	1.3	ug/kg	
98-82-8	Isopropylbenzene	ND	6.7	0.69	ug/kg	
79-20-9	Methyl Acetate	ND	6.7	1.1	ug/kg	
108-87-2	Methylcyclohexane	ND	6.7	0.87	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	ND	1.3	0.38	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	6.7	1.1	ug/kg	
75-09-2	Methylene chloride	ND	6.7	0.30	ug/kg	
103-65-1	n-Propylbenzene	ND	6.7	0.34	ug/kg	
100-42-5	Styrene	ND	6.7	0.14	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	6.7	0.39	ug/kg	
127-18-4	Tetrachloroethene	ND	6.7	0.19	ug/kg	
108-88-3	Toluene	ND	1.3	0.39	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	6.7	0.46	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	6.7	0.17	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	6.7	0.25	ug/kg	
79-01-6	Trichloroethene	ND	6.7	0.70	ug/kg	
75-69-4	Trichlorofluoromethane	ND	6.7	0.30	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	ND	6.7	0.57	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	ND	6.7	0.48	ug/kg	
75-01-4	Vinyl chloride	ND	6.7	0.24	ug/kg	
	m,p-Xylene	ND	2.7	0.62	ug/kg	
95-47-6	o-Xylene	ND	1.3	0.62	ug/kg	
1330-20-7	Xylene (total)	ND	2.7	0.62	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	96%		67-127%
17060-07-0	1,2-Dichloroethane-D4	85%		65-132%
2037-26-5	Toluene-D8	105%		74-129%
460-00-4	4-Bromofluorobenzene	94%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

3.2
3

Client Sample ID:	SW-CCI-001(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-2	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	81.1
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	V92783.D	1	07/24/09	JLI	n/a	n/a	VV3814
Run #2							

Run #	Initial Weight
Run #1	4.9 g
Run #2	

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	13	2.8	ug/kg	
71-43-2	Benzene	ND	1.3	0.43	ug/kg	
75-27-4	Bromodichloromethane	ND	6.3	0.32	ug/kg	
75-25-2	Bromoform	ND	6.3	0.19	ug/kg	
74-83-9	Bromomethane	ND	6.3	0.51	ug/kg	
78-93-3	2-Butanone (MEK)	ND	13	2.5	ug/kg	
104-51-8	n-Butylbenzene	ND	6.3	0.48	ug/kg	
135-98-8	sec-Butylbenzene	ND	6.3	0.61	ug/kg	
98-06-6	tert-Butylbenzene	ND	6.3	0.60	ug/kg	
75-15-0	Carbon disulfide	ND	6.3	0.38	ug/kg	
56-23-5	Carbon tetrachloride	ND	6.3	0.70	ug/kg	
108-90-7	Chlorobenzene	ND	6.3	0.43	ug/kg	
75-00-3	Chloroethane	ND	6.3	1.4	ug/kg	
67-66-3	Chloroform	ND	6.3	0.40	ug/kg	
74-87-3	Chloromethane	ND	6.3	0.21	ug/kg	
110-82-7	Cyclohexane	ND	6.3	0.19	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	13	0.68	ug/kg	
124-48-1	Dibromochloromethane	ND	6.3	0.14	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.3	0.17	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	6.3	0.34	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	6.3	0.35	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	6.3	0.42	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	6.3	1.2	ug/kg	
75-34-3	1,1-Dichloroethane	ND	6.3	0.17	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.3	0.43	ug/kg	
75-35-4	1,1-Dichloroethene	ND	6.3	0.83	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	6.3	0.30	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	6.3	0.56	ug/kg	
78-87-5	1,2-Dichloropropane	ND	6.3	0.16	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	6.3	0.17	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	6.3	0.12	ug/kg	
123-91-1	1,4-Dioxane	ND	160	110	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-001(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-2	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	81.1
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	1.3	0.47	ug/kg	
76-13-1	Freon 113	1.6	6.3	0.71	ug/kg	J
591-78-6	2-Hexanone	ND	6.3	1.2	ug/kg	
98-82-8	Isopropylbenzene	ND	6.3	0.65	ug/kg	
79-20-9	Methyl Acetate	ND	6.3	1.0	ug/kg	
108-87-2	Methylcyclohexane	ND	6.3	0.82	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	ND	1.3	0.35	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	6.3	1.0	ug/kg	
75-09-2	Methylene chloride	ND	6.3	0.28	ug/kg	
103-65-1	n-Propylbenzene	ND	6.3	0.32	ug/kg	
100-42-5	Styrene	ND	6.3	0.13	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	6.3	0.37	ug/kg	
127-18-4	Tetrachloroethene	ND	6.3	0.18	ug/kg	
108-88-3	Toluene	ND	1.3	0.37	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	6.3	0.43	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	6.3	0.16	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	6.3	0.23	ug/kg	
79-01-6	Trichloroethene	ND	6.3	0.66	ug/kg	
75-69-4	Trichlorofluoromethane	ND	6.3	0.29	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	ND	6.3	0.54	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	ND	6.3	0.45	ug/kg	
75-01-4	Vinyl chloride	ND	6.3	0.22	ug/kg	
	m,p-Xylene	ND	2.5	0.59	ug/kg	
95-47-6	o-Xylene	ND	1.3	0.59	ug/kg	
1330-20-7	Xylene (total)	ND	2.5	0.59	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	95%		67-127%
17060-07-0	1,2-Dichloroethane-D4	91%		65-132%
2037-26-5	Toluene-D8	92%		74-129%
460-00-4	4-Bromofluorobenzene	75%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest LabLink@514900 10:37 13-Aug-2009

Report of Analysis

Client Sample ID:	SW-CCI-002(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-3	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	86.1
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	D157556.D	1	07/27/09	TDN	n/a	n/a	VD6333
Run #2	D157558.D	1	07/27/09	TDN	n/a	n/a	VD6333

Run #	Initial Weight	Final Volume	Methanol Aliquot
Run #1	11.0 g	10.0 ml	100 ul
Run #2	11.0 g	10.0 ml	5.0 ul

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	610	140	ug/kg	
71-43-2	Benzene	35.7	61	21	ug/kg	J
75-27-4	Bromodichloromethane	ND	300	16	ug/kg	
75-25-2	Bromoform	ND	300	9.2	ug/kg	
74-83-9	Bromomethane	ND	300	25	ug/kg	
78-93-3	2-Butanone (MEK)	ND	610	120	ug/kg	
104-51-8	n-Butylbenzene	2870	300	23	ug/kg	
135-98-8	sec-Butylbenzene	1320	300	30	ug/kg	
98-06-6	tert-Butylbenzene	ND	300	29	ug/kg	
75-15-0	Carbon disulfide	ND	300	19	ug/kg	
56-23-5	Carbon tetrachloride	ND	300	34	ug/kg	
108-90-7	Chlorobenzene	ND	300	21	ug/kg	
75-00-3	Chloroethane	ND	300	69	ug/kg	
67-66-3	Chloroform	ND	300	19	ug/kg	
74-87-3	Chloromethane	ND	300	10	ug/kg	
110-82-7	Cyclohexane	713	300	9.2	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	610	33	ug/kg	
124-48-1	Dibromochloromethane	ND	300	6.7	ug/kg	
106-93-4	1,2-Dibromoethane	ND	61	8.3	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	300	16	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	300	17	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	300	21	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	300	58	ug/kg	
75-34-3	1,1-Dichloroethane	ND	300	8.4	ug/kg	
107-06-2	1,2-Dichloroethane	ND	61	21	ug/kg	
75-35-4	1,1-Dichloroethene	ND	300	40	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	300	15	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	300	27	ug/kg	
78-87-5	1,2-Dichloropropane	ND	300	7.9	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	300	8.1	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	300	5.8	ug/kg	
123-91-1	1,4-Dioxane	ND	7600	5300	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-002(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-3	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	86.1
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	4090	61	23	ug/kg	
76-13-1	Freon 113	ND	300	34	ug/kg	
591-78-6	2-Hexanone	ND	300	59	ug/kg	
98-82-8	Isopropylbenzene	3260	300	32	ug/kg	
79-20-9	Methyl Acetate	ND	300	50	ug/kg	
108-87-2	Methylcyclohexane	5470	300	40	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	ND	61	17	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	300	49	ug/kg	
75-09-2	Methylene chloride	ND	300	14	ug/kg	
103-65-1	n-Propylbenzene	8050	300	16	ug/kg	
100-42-5	Styrene	ND	300	6.5	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	300	18	ug/kg	
127-18-4	Tetrachloroethene	ND	300	8.8	ug/kg	
108-88-3	Toluene	5570	61	18	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	300	21	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	300	7.8	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	300	11	ug/kg	
79-01-6	Trichloroethene	ND	300	32	ug/kg	
75-69-4	Trichlorofluoromethane	ND	300	14	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	99900 ^a	6100	520	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	30200 ^a	6100	440	ug/kg	
75-01-4	Vinyl chloride	ND	300	11	ug/kg	
	m,p-Xylene	98700 ^a	2400	570	ug/kg	
95-47-6	o-Xylene	44500 ^a	1200	570	ug/kg	
1330-20-7	Xylene (total)	143000 ^a	2400	570	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	94%	95%	67-127%
17060-07-0	1,2-Dichloroethane-D4	92%	94%	65-132%
2037-26-5	Toluene-D8	105%	100%	74-129%
460-00-4	4-Bromofluorobenzene	128%	107%	62-138%

(a) Result is from Run# 2

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest LabLink@514900 10:37 13-Aug-2009

Report of Analysis

Page 1 of 2

Client Sample ID:	SW-CCI-003(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-4	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	88.4
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1 ^a	D157559.D	1	07/27/09	TDN	n/a	n/a	VD6333
Run #2							

Run #	Initial Weight	Final Volume	Methanol Aliquot
Run #1	10.8 g	10.0 ml	100 ul
Run #2			

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	590	130	ug/kg	
71-43-2	Benzene	ND	59	20	ug/kg	
75-27-4	Bromodichloromethane	ND	290	15	ug/kg	
75-25-2	Bromoform	ND	290	8.9	ug/kg	
74-83-9	Bromomethane	ND	290	24	ug/kg	
78-93-3	2-Butanone (MEK)	ND	590	120	ug/kg	
104-51-8	n-Butylbenzene	ND	290	22	ug/kg	
135-98-8	sec-Butylbenzene	ND	290	29	ug/kg	
98-06-6	tert-Butylbenzene	ND	290	28	ug/kg	
75-15-0	Carbon disulfide	ND	290	18	ug/kg	
56-23-5	Carbon tetrachloride	ND	290	33	ug/kg	
108-90-7	Chlorobenzene	ND	290	20	ug/kg	
75-00-3	Chloroethane	ND	290	67	ug/kg	
67-66-3	Chloroform	ND	290	19	ug/kg	
74-87-3	Chloromethane	ND	290	9.7	ug/kg	
110-82-7	Cyclohexane	ND	290	8.9	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	590	32	ug/kg	
124-48-1	Dibromochloromethane	ND	290	6.5	ug/kg	
106-93-4	1,2-Dibromoethane	ND	59	8.1	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	290	16	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	290	16	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	290	20	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	290	56	ug/kg	
75-34-3	1,1-Dichloroethane	ND	290	8.1	ug/kg	
107-06-2	1,2-Dichloroethane	ND	59	20	ug/kg	
75-35-4	1,1-Dichloroethene	ND	290	39	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	290	14	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	290	26	ug/kg	
78-87-5	1,2-Dichloropropane	ND	290	7.7	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	290	7.8	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	290	5.7	ug/kg	
123-91-1	1,4-Dioxane	ND	7400	5100	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-003(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-4	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	88.4
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	59	22	ug/kg	
76-13-1	Freon 113	ND	290	33	ug/kg	
591-78-6	2-Hexanone	ND	290	57	ug/kg	
98-82-8	Isopropylbenzene	ND	290	31	ug/kg	
79-20-9	Methyl Acetate	ND	290	49	ug/kg	
108-87-2	Methylcyclohexane	ND	290	39	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	ND	59	17	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	290	48	ug/kg	
75-09-2	Methylene chloride	ND	290	13	ug/kg	
103-65-1	n-Propylbenzene	ND	290	15	ug/kg	
100-42-5	Styrene	ND	290	6.3	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	290	17	ug/kg	
127-18-4	Tetrachloroethene	ND	290	8.5	ug/kg	
108-88-3	Toluene	ND	59	17	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	290	20	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	290	7.5	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	290	11	ug/kg	
79-01-6	Trichloroethene	ND	290	31	ug/kg	
75-69-4	Trichlorofluoromethane	ND	290	13	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	1030	290	25	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	3380	290	21	ug/kg	
75-01-4	Vinyl chloride	ND	290	10	ug/kg	
	m,p-Xylene	111	120	28	ug/kg	J
95-47-6	o-Xylene	309	59	28	ug/kg	
1330-20-7	Xylene (total)	420	120	28	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	94%		67-127%
17060-07-0	1,2-Dichloroethane-D4	93%		65-132%
2037-26-5	Toluene-D8	102%		74-129%
460-00-4	4-Bromofluorobenzene	111%		62-138%

(a) Dilution required due to matrix interference.

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest LabLink@514900 10:37 13-Aug-2009

Report of Analysis

Page 1 of 2

Client Sample ID:	SW-CCI-004(19')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-5	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	87.3
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	V92818.D	1	07/25/09	JLI	n/a	n/a	VV3815
Run #2							

Run #	Initial Weight
Run #1	5.3 g
Run #2	

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	24.9	11	2.4	ug/kg	
71-43-2	Benzene	ND	1.1	0.37	ug/kg	
75-27-4	Bromodichloromethane	ND	5.4	0.28	ug/kg	
75-25-2	Bromoform	ND	5.4	0.16	ug/kg	
74-83-9	Bromomethane	ND	5.4	0.44	ug/kg	
78-93-3	2-Butanone (MEK)	ND	11	2.1	ug/kg	
104-51-8	n-Butylbenzene	2.4	5.4	0.41	ug/kg	J
135-98-8	sec-Butylbenzene	1.0	5.4	0.53	ug/kg	J
98-06-6	tert-Butylbenzene	ND	5.4	0.51	ug/kg	
75-15-0	Carbon disulfide	ND	5.4	0.33	ug/kg	
56-23-5	Carbon tetrachloride	ND	5.4	0.60	ug/kg	
108-90-7	Chlorobenzene	ND	5.4	0.37	ug/kg	
75-00-3	Chloroethane	ND	5.4	1.2	ug/kg	
67-66-3	Chloroform	ND	5.4	0.34	ug/kg	
74-87-3	Chloromethane	ND	5.4	0.18	ug/kg	
110-82-7	Cyclohexane	ND	5.4	0.16	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	11	0.58	ug/kg	
124-48-1	Dibromochloromethane	ND	5.4	0.12	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.1	0.15	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	5.4	0.29	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	5.4	0.30	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	5.4	0.36	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	5.4	1.0	ug/kg	
75-34-3	1,1-Dichloroethane	ND	5.4	0.15	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.1	0.37	ug/kg	
75-35-4	1,1-Dichloroethene	ND	5.4	0.72	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	5.4	0.26	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	5.4	0.49	ug/kg	
78-87-5	1,2-Dichloropropane	ND	5.4	0.14	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	5.4	0.14	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	5.4	0.10	ug/kg	
123-91-1	1,4-Dioxane	ND	140	93	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-004(19')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-5	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	87.3
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	0.87	1.1	0.40	ug/kg	J
76-13-1	Freon 113	ND	5.4	0.61	ug/kg	
591-78-6	2-Hexanone	ND	5.4	1.0	ug/kg	
98-82-8	Isopropylbenzene	0.57	5.4	0.56	ug/kg	J
79-20-9	Methyl Acetate	ND	5.4	0.89	ug/kg	
108-87-2	Methylcyclohexane	1.3	5.4	0.71	ug/kg	J
1634-04-4	Methyl Tert Butyl Ether	ND	1.1	0.30	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	5.4	0.88	ug/kg	
75-09-2	Methylene chloride	ND	5.4	0.24	ug/kg	
103-65-1	n-Propylbenzene	1.8	5.4	0.28	ug/kg	J
100-42-5	Styrene	ND	5.4	0.12	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	5.4	0.32	ug/kg	
127-18-4	Tetrachloroethene	ND	5.4	0.16	ug/kg	
108-88-3	Toluene	0.37	1.1	0.32	ug/kg	J
120-82-1	1,2,4-Trichlorobenzene	ND	5.4	0.37	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	5.4	0.14	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	5.4	0.20	ug/kg	
79-01-6	Trichloroethene	ND	5.4	0.57	ug/kg	
75-69-4	Trichlorofluoromethane	ND	5.4	0.25	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	43.2	5.4	0.46	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	18.6	5.4	0.39	ug/kg	
75-01-4	Vinyl chloride	ND	5.4	0.19	ug/kg	
	m,p-Xylene	8.7	2.2	0.51	ug/kg	
95-47-6	o-Xylene	3.4	1.1	0.51	ug/kg	
1330-20-7	Xylene (total)	12.2	2.2	0.51	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	73%		67-127%
17060-07-0	1,2-Dichloroethane-D4	107%		65-132%
2037-26-5	Toluene-D8	98%		74-129%
460-00-4	4-Bromofluorobenzene	77%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest LabLink@514900 10:37 13-Aug-2009

Report of Analysis

Page 1 of 2

Client Sample ID:	SW-CCI-005(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-6	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	83.6
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	V92782.D	1	07/24/09	JLI	n/a	n/a	VV3814
Run #2							

Run #	Initial Weight
Run #1	5.1 g
Run #2	

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	18.3	12	2.6	ug/kg	
71-43-2	Benzene	ND	1.2	0.40	ug/kg	
75-27-4	Bromodichloromethane	ND	5.9	0.30	ug/kg	
75-25-2	Bromoform	ND	5.9	0.18	ug/kg	
74-83-9	Bromomethane	ND	5.9	0.47	ug/kg	
78-93-3	2-Butanone (MEK)	ND	12	2.3	ug/kg	
104-51-8	n-Butylbenzene	ND	5.9	0.45	ug/kg	
135-98-8	sec-Butylbenzene	ND	5.9	0.57	ug/kg	
98-06-6	tert-Butylbenzene	ND	5.9	0.56	ug/kg	
75-15-0	Carbon disulfide	ND	5.9	0.36	ug/kg	
56-23-5	Carbon tetrachloride	ND	5.9	0.65	ug/kg	
108-90-7	Chlorobenzene	ND	5.9	0.40	ug/kg	
75-00-3	Chloroethane	ND	5.9	1.3	ug/kg	
67-66-3	Chloroform	ND	5.9	0.37	ug/kg	
74-87-3	Chloromethane	ND	5.9	0.19	ug/kg	
110-82-7	Cyclohexane	ND	5.9	0.18	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	12	0.63	ug/kg	
124-48-1	Dibromochloromethane	ND	5.9	0.13	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.2	0.16	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	5.9	0.32	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	5.9	0.32	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	5.9	0.40	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	5.9	1.1	ug/kg	
75-34-3	1,1-Dichloroethane	ND	5.9	0.16	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.2	0.40	ug/kg	
75-35-4	1,1-Dichloroethene	ND	5.9	0.78	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	5.9	0.28	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	5.9	0.53	ug/kg	
78-87-5	1,2-Dichloropropane	ND	5.9	0.15	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	5.9	0.16	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	5.9	0.11	ug/kg	
123-91-1	1,4-Dioxane	ND	150	100	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-005(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-6	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	83.6
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	1.2	0.44	ug/kg	
76-13-1	Freon 113	1.4	5.9	0.66	ug/kg	J
591-78-6	2-Hexanone	ND	5.9	1.1	ug/kg	
98-82-8	Isopropylbenzene	ND	5.9	0.61	ug/kg	
79-20-9	Methyl Acetate	ND	5.9	0.97	ug/kg	
108-87-2	Methylcyclohexane	ND	5.9	0.77	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	ND	1.2	0.33	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	5.9	0.95	ug/kg	
75-09-2	Methylene chloride	ND	5.9	0.26	ug/kg	
103-65-1	n-Propylbenzene	ND	5.9	0.30	ug/kg	
100-42-5	Styrene	ND	5.9	0.13	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	5.9	0.34	ug/kg	
127-18-4	Tetrachloroethene	ND	5.9	0.17	ug/kg	
108-88-3	Toluene	ND	1.2	0.34	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	5.9	0.40	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	5.9	0.15	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	5.9	0.22	ug/kg	
79-01-6	Trichloroethene	ND	5.9	0.62	ug/kg	
75-69-4	Trichlorofluoromethane	ND	5.9	0.27	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	ND	5.9	0.50	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	ND	5.9	0.42	ug/kg	
75-01-4	Vinyl chloride	ND	5.9	0.21	ug/kg	
	m,p-Xylene	ND	2.3	0.55	ug/kg	
95-47-6	o-Xylene	ND	1.2	0.55	ug/kg	
1330-20-7	Xylene (total)	ND	2.3	0.55	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	94%		67-127%
17060-07-0	1,2-Dichloroethane-D4	87%		65-132%
2037-26-5	Toluene-D8	93%		74-129%
460-00-4	4-Bromofluorobenzene	74%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest LabLink@514900 10:37 13-Aug-2009

Report of Analysis

Page 1 of 2

Client Sample ID:	SW-CCI-006(19')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-7	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	91.3
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	V92784.D	1	07/24/09	JLI	n/a	n/a	VV3814
Run #2							

Run #1	Initial Weight
Run #1	5.5 g
Run #2	

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	10	2.2	ug/kg	
71-43-2	Benzene	ND	1.0	0.34	ug/kg	
75-27-4	Bromodichloromethane	ND	5.0	0.26	ug/kg	
75-25-2	Bromoform	ND	5.0	0.15	ug/kg	
74-83-9	Bromomethane	ND	5.0	0.40	ug/kg	
78-93-3	2-Butanone (MEK)	ND	10	2.0	ug/kg	
104-51-8	n-Butylbenzene	ND	5.0	0.38	ug/kg	
135-98-8	sec-Butylbenzene	ND	5.0	0.49	ug/kg	
98-06-6	tert-Butylbenzene	ND	5.0	0.47	ug/kg	
75-15-0	Carbon disulfide	ND	5.0	0.30	ug/kg	
56-23-5	Carbon tetrachloride	ND	5.0	0.55	ug/kg	
108-90-7	Chlorobenzene	ND	5.0	0.34	ug/kg	
75-00-3	Chloroethane	ND	5.0	1.1	ug/kg	
67-66-3	Chloroform	ND	5.0	0.32	ug/kg	
74-87-3	Chloromethane	ND	5.0	0.16	ug/kg	
110-82-7	Cyclohexane	ND	5.0	0.15	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	10	0.54	ug/kg	
124-48-1	Dibromochloromethane	ND	5.0	0.11	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.0	0.14	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	5.0	0.27	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	5.0	0.27	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	5.0	0.34	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	5.0	0.94	ug/kg	
75-34-3	1,1-Dichloroethane	ND	5.0	0.14	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.0	0.34	ug/kg	
75-35-4	1,1-Dichloroethene	ND	5.0	0.66	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	5.0	0.24	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	5.0	0.45	ug/kg	
78-87-5	1,2-Dichloropropane	ND	5.0	0.13	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	5.0	0.13	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	5.0	0.096	ug/kg	
123-91-1	1,4-Dioxane	ND	120	86	ug/kg	

ND = Not detected MDL - Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

Report of Analysis

3.7
3

Client Sample ID:	SW-CCI-006(19')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-7	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	91.3
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	1.0	0.37	ug/kg	
76-13-1	Freon 113	1.3	5.0	0.56	ug/kg	J
591-78-6	2-Hexanone	ND	5.0	0.96	ug/kg	
98-82-8	Isopropylbenzene	ND	5.0	0.52	ug/kg	
79-20-9	Methyl Acetate	ND	5.0	0.82	ug/kg	
108-87-2	Methylcyclohexane	ND	5.0	0.65	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	ND	1.0	0.28	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	5.0	0.81	ug/kg	
75-09-2	Methylene chloride	ND	5.0	0.22	ug/kg	
103-65-1	n-Propylbenzene	ND	5.0	0.25	ug/kg	
100-42-5	Styrene	ND	5.0	0.11	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	5.0	0.29	ug/kg	
127-18-4	Tetrachloroethene	ND	5.0	0.14	ug/kg	
108-88-3	Toluene	ND	1.0	0.29	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	5.0	0.34	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	5.0	0.13	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	5.0	0.18	ug/kg	
79-01-6	Trichloroethene	ND	5.0	0.52	ug/kg	
75-69-4	Trichlorofluoromethane	ND	5.0	0.23	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	ND	5.0	0.43	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	ND	5.0	0.36	ug/kg	
75-01-4	Vinyl chloride	ND	5.0	0.18	ug/kg	
	m,p-Xylene	ND	2.0	0.47	ug/kg	
95-47-6	o-Xylene	ND	1.0	0.47	ug/kg	
1330-20-7	Xylene (total)	ND	2.0	0.47	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	97%		67-127%
17060-07-0	1,2-Dichloroethane-D4	96%		65-132%
2037-26-5	Toluene-D8	94%		74-129%
460-00-4	4-Bromofluorobenzene	74%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest LabLink@514900 10:37 13-Aug-2009

Report of Analysis

Page 1 of 2

Client Sample ID:	SW-CCI-007(19')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-8	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	84.5
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	V92873.D	1	07/27/09	JLI	n/a	n/a	VV3817

Run #1	Initial Weight
Run #2	5.5 g

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	9.5	11	2.4	ug/kg	J
71-43-2	Benzene	ND	1.1	0.37	ug/kg	
75-27-4	Bromodichloromethane	ND	5.4	0.28	ug/kg	
75-25-2	Bromoform	ND	5.4	0.16	ug/kg	
74-83-9	Bromomethane	ND	5.4	0.43	ug/kg	
78-93-3	2-Butanone (MEK)	ND	11	2.1	ug/kg	
104-51-8	n-Butylbenzene	ND	5.4	0.41	ug/kg	
135-98-8	sec-Butylbenzene	ND	5.4	0.53	ug/kg	
98-06-6	tert-Butylbenzene	ND	5.4	0.51	ug/kg	
75-15-0	Carbon disulfide	ND	5.4	0.33	ug/kg	
56-23-5	Carbon tetrachloride	ND	5.4	0.60	ug/kg	
108-90-7	Chlorobenzene	ND	5.4	0.36	ug/kg	
75-00-3	Chloroethane	ND	5.4	1.2	ug/kg	
67-66-3	Chloroform	ND	5.4	0.34	ug/kg	
74-87-3	Chloromethane	ND	5.4	0.18	ug/kg	
110-82-7	Cyclohexane	ND	5.4	0.16	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	11	0.58	ug/kg	
124-48-1	Dibromochloromethane	ND	5.4	0.12	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.1	0.15	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	5.4	0.29	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	5.4	0.30	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	5.4	0.36	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	5.4	1.0	ug/kg	
75-34-3	1,1-Dichloroethane	ND	5.4	0.15	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.1	0.37	ug/kg	
75-35-4	1,1-Dichloroethene	ND	5.4	0.71	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	5.4	0.26	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	5.4	0.48	ug/kg	
78-87-5	1,2-Dichloropropane	ND	5.4	0.14	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	5.4	0.14	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	5.4	0.10	ug/kg	
123-91-1	1,4-Dioxane	ND	130	93	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-007(19')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-8	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	84.5
Method:	SW846 8260B	Project: Courtlandt Corners I, 364 East 161st Street, Bronx, NY	

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	1.1	0.40	ug/kg	
76-13-1	Freon 113	ND	5.4	0.60	ug/kg	
591-78-6	2-Hexanone	ND	5.4	1.0	ug/kg	
98-82-8	Isopropylbenzene	ND	5.4	0.56	ug/kg	
79-20-9	Methyl Acetate	ND	5.4	0.89	ug/kg	
108-87-2	Methylcyclohexane	ND	5.4	0.70	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	ND	1.1	0.30	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	5.4	0.87	ug/kg	
75-09-2	Methylene chloride	ND	5.4	0.24	ug/kg	
103-65-1	n-Propylbenzene	ND	5.4	0.27	ug/kg	
100-42-5	Styrene	ND	5.4	0.12	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	5.4	0.32	ug/kg	
127-18-4	Tetrachloroethene	ND	5.4	0.16	ug/kg	
108-88-3	Toluene	ND	1.1	0.31	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	5.4	0.37	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	5.4	0.14	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	5.4	0.20	ug/kg	
79-01-6	Trichloroethene	ND	5.4	0.57	ug/kg	
75-69-4	Trichlorofluoromethane	ND	5.4	0.25	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	ND	5.4	0.46	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	ND	5.4	0.39	ug/kg	
75-01-4	Vinyl chloride	ND	5.4	0.19	ug/kg	
	m,p-Xylene	ND	2.2	0.50	ug/kg	
95-47-6	o-Xylene	ND	1.1	0.50	ug/kg	
1330-20-7	Xylene (total)	ND	2.2	0.50	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	114%		67-127%
17060-07-0	1,2-Dichloroethane-D4	113%		65-132%
2037-26-5	Toluene-D8	100%		74-129%
460-00-4	4-Bromofluorobenzene	73%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest LabLink@514900 10:37 13-Aug-2009

Report of Analysis

Page 1 of 2

Client Sample ID:	SW-CCI-008(13.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-9	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	85.4
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	V92785.D	1	07/24/09	JLI	n/a	n/a	VV3814
Run #2							

Run #	Initial Weight
Run #1	5.3 g
Run #2	

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	11	2.5	ug/kg	
71-43-2	Benzene	ND	1.1	0.38	ug/kg	
75-27-4	Bromodichloromethane	ND	5.5	0.28	ug/kg	
75-25-2	Bromoform	ND	5.5	0.17	ug/kg	
74-83-9	Bromomethane	ND	5.5	0.45	ug/kg	
78-93-3	2-Butanone (MEK)	ND	11	2.2	ug/kg	
104-51-8	n-Butylbenzene	ND	5.5	0.42	ug/kg	
135-98-8	sec-Butylbenzene	ND	5.5	0.54	ug/kg	
98-06-6	tert-Butylbenzene	ND	5.5	0.52	ug/kg	
75-15-0	Carbon disulfide	ND	5.5	0.34	ug/kg	
56-23-5	Carbon tetrachloride	ND	5.5	0.61	ug/kg	
108-90-7	Chlorobenzene	ND	5.5	0.37	ug/kg	
75-00-3	Chloroethane	ND	5.5	1.3	ug/kg	
67-66-3	Chloroform	ND	5.5	0.35	ug/kg	
74-87-3	Chloromethane	ND	5.5	0.18	ug/kg	
110-82-7	Cyclohexane	ND	5.5	0.17	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	11	0.60	ug/kg	
124-48-1	Dibromochloromethane	ND	5.5	0.12	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.1	0.15	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	5.5	0.30	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	5.5	0.30	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	5.5	0.37	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	5.5	1.0	ug/kg	
75-34-3	1,1-Dichloroethane	ND	5.5	0.15	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.1	0.38	ug/kg	
75-35-4	1,1-Dichloroethene	ND	5.5	0.73	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	5.5	0.26	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	5.5	0.50	ug/kg	
78-87-5	1,2-Dichloropropane	ND	5.5	0.14	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	5.5	0.15	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	5.5	0.11	ug/kg	
123-91-1	1,4-Dioxane	ND	140	95	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-008(13.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-9	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	85.4
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	1.1	0.41	ug/kg	
76-13-1	Freon 113	1.7	5.5	0.62	ug/kg	J
591-78-6	2-Hexanone	ND	5.5	1.1	ug/kg	
98-82-8	Isopropylbenzene	ND	5.5	0.57	ug/kg	
79-20-9	Methyl Acetate	ND	5.5	0.91	ug/kg	
108-87-2	Methylcyclohexane	ND	5.5	0.72	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	ND	1.1	0.31	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	5.5	0.90	ug/kg	
75-09-2	Methylene chloride	ND	5.5	0.25	ug/kg	
103-65-1	n-Propylbenzene	ND	5.5	0.28	ug/kg	
100-42-5	Styrene	ND	5.5	0.12	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	5.5	0.32	ug/kg	
127-18-4	Tetrachloroethene	ND	5.5	0.16	ug/kg	
108-88-3	Toluene	ND	1.1	0.32	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	5.5	0.38	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	5.5	0.14	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	5.5	0.20	ug/kg	
79-01-6	Trichloroethene	ND	5.5	0.58	ug/kg	
75-69-4	Trichlorofluoromethane	ND	5.5	0.25	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	ND	5.5	0.48	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	ND	5.5	0.40	ug/kg	
75-01-4	Vinyl chloride	ND	5.5	0.20	ug/kg	
	m,p-Xylene	ND	2.2	0.52	ug/kg	
95-47-6	o-Xylene	ND	1.1	0.52	ug/kg	
1330-20-7	Xylene (total)	ND	2.2	0.52	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	97%		67-127%
17060-07-0	1,2-Dichloroethane-D4	92%		65-132%
2037-26-5	Toluene-D8	93%		74-129%
460-00-4	4-Bromofluorobenzene	74%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest LabLink@514900 10:37 13-Aug-2009

Report of Analysis

Page 1 of 2

Client Sample ID:	SW-CCI-009(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-10	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	94.6
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	V92786.D	1	07/24/09	JLI	n/a	n/a	VV3814

Run #1	Initial Weight
Run #2	4.9 g

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	11	2.4	ug/kg	
71-43-2	Benzene	ND	1.1	0.37	ug/kg	
75-27-4	Bromodichloromethane	ND	5.4	0.28	ug/kg	
75-25-2	Bromoform	ND	5.4	0.16	ug/kg	
74-83-9	Bromomethane	ND	5.4	0.44	ug/kg	
78-93-3	2-Butanone (MEK)	ND	11	2.1	ug/kg	
104-51-8	n-Butylbenzene	ND	5.4	0.41	ug/kg	
135-98-8	sec-Butylbenzene	ND	5.4	0.53	ug/kg	
98-06-6	tert-Butylbenzene	ND	5.4	0.51	ug/kg	
75-15-0	Carbon disulfide	ND	5.4	0.33	ug/kg	
56-23-5	Carbon tetrachloride	ND	5.4	0.60	ug/kg	
108-90-7	Chlorobenzene	ND	5.4	0.37	ug/kg	
75-00-3	Chloroethane	ND	5.4	1.2	ug/kg	
67-66-3	Chloroform	ND	5.4	0.34	ug/kg	
74-87-3	Chloromethane	ND	5.4	0.18	ug/kg	
110-82-7	Cyclohexane	ND	5.4	0.16	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	11	0.58	ug/kg	
124-48-1	Dibromochloromethane	ND	5.4	0.12	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.1	0.15	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	5.4	0.29	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	5.4	0.30	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	5.4	0.36	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	5.4	1.0	ug/kg	
75-34-3	1,1-Dichloroethane	ND	5.4	0.15	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.1	0.37	ug/kg	
75-35-4	1,1-Dichloroethene	ND	5.4	0.71	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	5.4	0.26	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	5.4	0.48	ug/kg	
78-87-5	1,2-Dichloropropane	ND	5.4	0.14	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	5.4	0.14	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	5.4	0.10	ug/kg	
123-91-1	1,4-Dioxane	ND	130	93	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-009(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-10	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	94.6
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	1.1	0.40	ug/kg	
76-13-1	Freon 113	1.7	5.4	0.61	ug/kg	J
591-78-6	2-Hexanone	ND	5.4	1.0	ug/kg	
98-82-8	Isopropylbenzene	ND	5.4	0.56	ug/kg	
79-20-9	Methyl Acetate	ND	5.4	0.89	ug/kg	
108-87-2	Methylcyclohexane	ND	5.4	0.71	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	ND	1.1	0.30	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	5.4	0.87	ug/kg	
75-09-2	Methylene chloride	ND	5.4	0.24	ug/kg	
103-65-1	n-Propylbenzene	ND	5.4	0.28	ug/kg	
100-42-5	Styrene	ND	5.4	0.12	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	5.4	0.32	ug/kg	
127-18-4	Tetrachloroethene	ND	5.4	0.16	ug/kg	
108-88-3	Toluene	ND	1.1	0.31	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	5.4	0.37	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	5.4	0.14	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	5.4	0.20	ug/kg	
79-01-6	Trichloroethene	ND	5.4	0.57	ug/kg	
75-69-4	Trichlorofluoromethane	ND	5.4	0.25	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	ND	5.4	0.46	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	ND	5.4	0.39	ug/kg	
75-01-4	Vinyl chloride	ND	5.4	0.19	ug/kg	
	m,p-Xylene	ND	2.2	0.51	ug/kg	
95-47-6	o-Xylene	ND	1.1	0.51	ug/kg	
1330-20-7	Xylene (total)	ND	2.2	0.51	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	98%		67-127%
17060-07-0	1,2-Dichloroethane-D4	94%		65-132%
2037-26-5	Toluene-D8	94%		74-129%
460-00-4	4-Bromofluorobenzene	75%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 2

Client Sample ID:	SW-CCI-011(24.5)	Date Sampled:	08/10/09
Lab Sample ID:	JA25246-1	Date Received:	08/10/09
Matrix:	SO - Soil	Percent Solids:	83.6
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	V93437.D	1	08/11/09	JLI	n/a	n/a	VV3845
Run #2							

Run #1	Initial Weight
Run #1	5.3 g
Run #2	

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	11	2.5	ug/kg	
71-43-2	Benzene	ND	1.1	0.38	ug/kg	
75-27-4	Bromodichloromethane	ND	5.6	0.29	ug/kg	
75-25-2	Bromoform	ND	5.6	0.17	ug/kg	
74-83-9	Bromomethane	ND	5.6	0.46	ug/kg	
78-93-3	2-Butanone (MEK)	ND	11	2.2	ug/kg	
104-51-8	n-Butylbenzene	ND	5.6	0.43	ug/kg	
135-98-8	sec-Butylbenzene	ND	5.6	0.55	ug/kg	
98-06-6	tert-Butylbenzene	ND	5.6	0.54	ug/kg	
75-15-0	Carbon disulfide	ND	5.6	0.34	ug/kg	
56-23-5	Carbon tetrachloride	ND	5.6	0.63	ug/kg	
108-90-7	Chlorobenzene	ND	5.6	0.38	ug/kg	
75-00-3	Chloroethane	ND	5.6	1.3	ug/kg	
67-66-3	Chloroform	ND	5.6	0.36	ug/kg	
74-87-3	Chloromethane	ND	5.6	0.19	ug/kg	
110-82-7	Cyclohexane	ND	5.6	0.17	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	11	0.61	ug/kg	
124-48-1	Dibromochloromethane	ND	5.6	0.12	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.1	0.15	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	5.6	0.30	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	5.6	0.31	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	5.6	0.38	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	5.6	1.1	ug/kg	
75-34-3	1,1-Dichloroethane	ND	5.6	0.16	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.1	0.39	ug/kg	
75-35-4	1,1-Dichloroethene	ND	5.6	0.75	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	5.6	0.27	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	5.6	0.51	ug/kg	
78-87-5	1,2-Dichloropropane	ND	5.6	0.15	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	5.6	0.15	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	5.6	0.11	ug/kg	
123-91-1	1,4-Dioxane	ND	140	97	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-011(24.5)	Date Sampled:	08/10/09
Lab Sample ID:	JA25246-1	Date Received:	08/10/09
Matrix:	SO - Soil	Percent Solids:	83.6
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	1.1	0.42	ug/kg	
76-13-1	Freon 113	ND	5.6	0.63	ug/kg	
591-78-6	2-Hexanone	ND	5.6	1.1	ug/kg	
98-82-8	Isopropylbenzene	ND	5.6	0.58	ug/kg	
79-20-9	Methyl Acetate	ND	5.6	0.93	ug/kg	
108-87-2	Methylcyclohexane	4.5	5.6	0.74	ug/kg	J
1634-04-4	Methyl Tert Butyl Ether	9.8	1.1	0.32	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	5.6	0.92	ug/kg	
75-09-2	Methylene chloride	ND	5.6	0.25	ug/kg	
103-65-1	n-Propylbenzene	ND	5.6	0.29	ug/kg	
100-42-5	Styrene	ND	5.6	0.12	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	5.6	0.33	ug/kg	
127-18-4	Tetrachloroethene	ND	5.6	0.16	ug/kg	
108-88-3	Toluene	ND	1.1	0.33	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	5.6	0.39	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	5.6	0.14	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	5.6	0.21	ug/kg	
79-01-6	Trichloroethene	ND	5.6	0.59	ug/kg	
75-69-4	Trichlorofluoromethane	ND	5.6	0.26	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	ND	5.6	0.49	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	1.5	5.6	0.40	ug/kg	J
75-01-4	Vinyl chloride	ND	5.6	0.20	ug/kg	
	m,p-Xylene	ND	2.3	0.53	ug/kg	
95-47-6	o-Xylene	0.54	1.1	0.53	ug/kg	J
1330-20-7	Xylene (total)	0.54	2.3	0.53	ug/kg	J

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	104%		67-127%
17060-07-0	1,2-Dichloroethane-D4	96%		65-132%
2037-26-5	Toluene-D8	97%		74-129%
460-00-4	4-Bromofluorobenzene	75%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 2

Client Sample ID:	SW-CCI-012	Date Sampled:	08/13/09
Lab Sample ID:	JA25522-1	Date Received:	08/13/09
Matrix:	SO - Soil	Percent Solids:	82.7
Method:	SW846 8260B	Project: Courtlandt Corners I, 364 East 161st Street, Bronx, NY	

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	V93710.D	1	08/19/09	JLI	n/a	n/a	VV3859
Run #2							

Run #	Initial Weight
Run #1	5.1 g
Run #2	

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	12	2.6	ug/kg	
71-43-2	Benzene	ND	1.2	0.40	ug/kg	
75-27-4	Bromodichloromethane	ND	5.9	0.30	ug/kg	
75-25-2	Bromoform	ND	5.9	0.18	ug/kg	
74-83-9	Bromomethane	ND	5.9	0.48	ug/kg	
78-93-3	2-Butanone (MEK)	ND	12	2.3	ug/kg	
104-51-8	n-Butylbenzene	ND	5.9	0.45	ug/kg	
135-98-8	sec-Butylbenzene	ND	5.9	0.58	ug/kg	
98-06-6	tert-Butylbenzene	ND	5.9	0.56	ug/kg	
75-15-0	Carbon disulfide	3.3	5.9	0.36	ug/kg	J
56-23-5	Carbon tetrachloride	ND	5.9	0.66	ug/kg	
108-90-7	Chlorobenzene	ND	5.9	0.40	ug/kg	
75-00-3	Chloroethane	ND	5.9	1.4	ug/kg	
67-66-3	Chloroform	ND	5.9	0.38	ug/kg	
74-87-3	Chloromethane	ND	5.9	0.20	ug/kg	
110-82-7	Cyclohexane	ND	5.9	0.18	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	12	0.64	ug/kg	
124-48-1	Dibromochloromethane	ND	5.9	0.13	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.2	0.16	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	5.9	0.32	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	5.9	0.33	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	5.9	0.40	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	5.9	1.1	ug/kg	
75-34-3	1,1-Dichloroethane	ND	5.9	0.16	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.2	0.41	ug/kg	
75-35-4	1,1-Dichloroethene	ND	5.9	0.78	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	5.9	0.28	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	5.9	0.53	ug/kg	
78-87-5	1,2-Dichloropropane	ND	5.9	0.15	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	5.9	0.16	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	5.9	0.11	ug/kg	
123-91-1	1,4-Dioxane	ND	150	100	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-012	Date Sampled:	08/13/09
Lab Sample ID:	JA25522-1	Date Received:	08/13/09
Matrix:	SO - Soil	Percent Solids:	82.7
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	1.2	0.44	ug/kg	
76-13-1	Freon 113	ND	5.9	0.67	ug/kg	
591-78-6	2-Hexanone	ND	5.9	1.1	ug/kg	
98-82-8	Isopropylbenzene	ND	5.9	0.61	ug/kg	
79-20-9	Methyl Acetate	ND	5.9	0.98	ug/kg	
108-87-2	Methylcyclohexane	2.9	5.9	0.78	ug/kg	J
1634-04-4	Methyl Tert Butyl Ether	ND	1.2	0.33	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	5.9	0.96	ug/kg	
75-09-2	Methylene chloride	ND	5.9	0.26	ug/kg	
103-65-1	n-Propylbenzene	ND	5.9	0.30	ug/kg	
100-42-5	Styrene	ND	5.9	0.13	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	5.9	0.35	ug/kg	
127-18-4	Tetrachloroethene	ND	5.9	0.17	ug/kg	
108-88-3	Toluene	ND	1.2	0.35	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	5.9	0.41	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	5.9	0.15	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	5.9	0.22	ug/kg	
79-01-6	Trichloroethene	ND	5.9	0.62	ug/kg	
75-69-4	Trichlorofluoromethane	ND	5.9	0.27	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	ND	5.9	0.51	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	0.56	5.9	0.42	ug/kg	J
75-01-4	Vinyl chloride	ND	5.9	0.21	ug/kg	
	m,p-Xylene	ND	2.4	0.56	ug/kg	
95-47-6	o-Xylene	ND	1.2	0.56	ug/kg	
1330-20-7	Xylene (total)	ND	2.4	0.56	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	91%		67-127%
17060-07-0	1,2-Dichloroethane-D4	75%		65-132%
2037-26-5	Toluene-D8	89%		74-129%
460-00-4	4-Bromofluorobenzene	80%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 2

Client Sample ID:	SW-CCI-013	Date Sampled:	08/13/09
Lab Sample ID:	JA25522-2	Date Received:	08/13/09
Matrix:	SO - Soil	Percent Solids:	88.2
Method:	SW846 8260B	Project: Courtlandt Corners I, 364 East 161st Street, Bronx, NY	

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	V93727.D	1	08/19/09	JLI	n/a	n/a	VV3860

Run #1	Initial Weight
Run #2	5.1 g

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	11	2.5	ug/kg	
71-43-2	Benzene	ND	1.1	0.38	ug/kg	
75-27-4	Bromodichloromethane	ND	5.6	0.29	ug/kg	
75-25-2	Bromoform	ND	5.6	0.17	ug/kg	
74-83-9	Bromomethane	ND	5.6	0.45	ug/kg	
78-93-3	2-Butanone (MEK)	ND	11	2.2	ug/kg	
104-51-8	n-Butylbenzene	ND	5.6	0.42	ug/kg	
135-98-8	sec-Butylbenzene	ND	5.6	0.54	ug/kg	
98-06-6	tert-Butylbenzene	ND	5.6	0.53	ug/kg	
75-15-0	Carbon disulfide	ND	5.6	0.34	ug/kg	
56-23-5	Carbon tetrachloride	ND	5.6	0.62	ug/kg	
108-90-7	Chlorobenzene	ND	5.6	0.38	ug/kg	
75-00-3	Chloroethane	ND	5.6	1.3	ug/kg	
67-66-3	Chloroform	ND	5.6	0.35	ug/kg	
74-87-3	Chloromethane	ND	5.6	0.18	ug/kg	
110-82-7	Cyclohexane	ND	5.6	0.17	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	11	0.60	ug/kg	
124-48-1	Dibromochloromethane	ND	5.6	0.12	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.1	0.15	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	5.6	0.30	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	5.6	0.31	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	5.6	0.37	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	5.6	1.1	ug/kg	
75-34-3	1,1-Dichloroethane	ND	5.6	0.15	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.1	0.38	ug/kg	
75-35-4	1,1-Dichloroethene	ND	5.6	0.74	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	5.6	0.27	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	5.6	0.50	ug/kg	
78-87-5	1,2-Dichloropropane	ND	5.6	0.14	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	5.6	0.15	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	5.6	0.11	ug/kg	
123-91-1	1,4-Dioxane	ND	140	96	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-013	Date Sampled:	08/13/09
Lab Sample ID:	JA25522-2	Date Received:	08/13/09
Matrix:	SO - Soil	Percent Solids:	88.2
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	1.1	0.41	ug/kg	
76-13-1	Freon 113	ND	5.6	0.62	ug/kg	
591-78-6	2-Hexanone	ND	5.6	1.1	ug/kg	
98-82-8	Isopropylbenzene	ND	5.6	0.58	ug/kg	
79-20-9	Methyl Acetate	ND	5.6	0.91	ug/kg	
108-87-2	Methylcyclohexane	ND	5.6	0.73	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	ND	1.1	0.31	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	5.6	0.90	ug/kg	
75-09-2	Methylene chloride	ND	5.6	0.25	ug/kg	
103-65-1	n-Propylbenzene	ND	5.6	0.28	ug/kg	
100-42-5	Styrene	ND	5.6	0.12	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	5.6	0.33	ug/kg	
127-18-4	Tetrachloroethene	ND	5.6	0.16	ug/kg	
108-88-3	Toluene	0.63	1.1	0.32	ug/kg	J
120-82-1	1,2,4-Trichlorobenzene	ND	5.6	0.38	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	5.6	0.14	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	5.6	0.21	ug/kg	
79-01-6	Trichloroethene	ND	5.6	0.58	ug/kg	
75-69-4	Trichlorofluoromethane	ND	5.6	0.25	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	0.78	5.6	0.48	ug/kg	J
108-67-8	1,3,5-Trimethylbenzene	3.2	5.6	0.40	ug/kg	J
75-01-4	Vinyl chloride	ND	5.6	0.20	ug/kg	
	m,p-Xylene	0.62	2.2	0.52	ug/kg	J
95-47-6	o-Xylene	0.59	1.1	0.52	ug/kg	J
1330-20-7	Xylene (total)	1.2	2.2	0.52	ug/kg	J

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	82%		67-127%
17060-07-0	1,2-Dichloroethane-D4	76%		65-132%
2037-26-5	Toluene-D8	87%		74-129%
460-00-4	4-Bromofluorobenzene	75%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 2

Client Sample ID:	SW-CCI-014	Date Sampled:	08/13/09
Lab Sample ID:	JA25522-3	Date Received:	08/13/09
Matrix:	SO - Soil	Percent Solids:	85.6
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	V93712.D	1	08/19/09	JLI	n/a	n/a	VV3859

Run #1	Initial Weight
Run #2	5.1 g

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	7.6	11	2.6	ug/kg	J
71-43-2	Benzene	ND	1.1	0.39	ug/kg	
75-27-4	Bromodichloromethane	ND	5.7	0.29	ug/kg	
75-25-2	Bromoform	ND	5.7	0.17	ug/kg	
74-83-9	Bromomethane	ND	5.7	0.46	ug/kg	
78-93-3	2-Butanone (MEK)	ND	11	2.3	ug/kg	
104-51-8	n-Butylbenzene	ND	5.7	0.44	ug/kg	
135-98-8	sec-Butylbenzene	ND	5.7	0.56	ug/kg	
98-06-6	tert-Butylbenzene	ND	5.7	0.54	ug/kg	
75-15-0	Carbon disulfide	ND	5.7	0.35	ug/kg	
56-23-5	Carbon tetrachloride	ND	5.7	0.64	ug/kg	
108-90-7	Chlorobenzene	ND	5.7	0.39	ug/kg	
75-00-3	Chloroethane	ND	5.7	1.3	ug/kg	
67-66-3	Chloroform	ND	5.7	0.36	ug/kg	
74-87-3	Chloromethane	ND	5.7	0.19	ug/kg	
110-82-7	Cyclohexane	ND	5.7	0.17	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	11	0.62	ug/kg	
124-48-1	Dibromochloromethane	ND	5.7	0.13	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.1	0.16	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	5.7	0.31	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	5.7	0.31	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	5.7	0.39	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	5.7	1.1	ug/kg	
75-34-3	1,1-Dichloroethane	ND	5.7	0.16	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.1	0.40	ug/kg	
75-35-4	1,1-Dichloroethene	ND	5.7	0.76	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	5.7	0.27	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	5.7	0.51	ug/kg	
78-87-5	1,2-Dichloropropane	ND	5.7	0.15	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	5.7	0.15	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	5.7	0.11	ug/kg	
123-91-1	1,4-Dioxane	ND	140	99	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-014	Date Sampled:	08/13/09
Lab Sample ID:	JA25522-3	Date Received:	08/13/09
Matrix:	SO - Soil	Percent Solids:	85.6
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	1.1	0.42	ug/kg	
76-13-1	Freon 113	ND	5.7	0.64	ug/kg	
591-78-6	2-Hexanone	ND	5.7	1.1	ug/kg	
98-82-8	Isopropylbenzene	ND	5.7	0.59	ug/kg	
79-20-9	Methyl Acetate	ND	5.7	0.94	ug/kg	
108-87-2	Methylcyclohexane	0.80	5.7	0.75	ug/kg	J
1634-04-4	Methyl Tert Butyl Ether	ND	1.1	0.32	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	5.7	0.93	ug/kg	
75-09-2	Methylene chloride	ND	5.7	0.26	ug/kg	
103-65-1	n-Propylbenzene	ND	5.7	0.29	ug/kg	
100-42-5	Styrene	ND	5.7	0.12	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	5.7	0.34	ug/kg	
127-18-4	Tetrachloroethene	ND	5.7	0.17	ug/kg	
108-88-3	Toluene	ND	1.1	0.33	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	5.7	0.40	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	5.7	0.15	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	5.7	0.21	ug/kg	
79-01-6	Trichloroethene	ND	5.7	0.60	ug/kg	
75-69-4	Trichlorofluoromethane	0.35	5.7	0.26	ug/kg	J
95-63-6	1,2,4-Trimethylbenzene	ND	5.7	0.49	ug/kg	
108-67-8	1,3,5-Trimethylbenzene	0.98	5.7	0.41	ug/kg	J
75-01-4	Vinyl chloride	ND	5.7	0.20	ug/kg	
	m,p-Xylene	ND	2.3	0.54	ug/kg	
95-47-6	o-Xylene	ND	1.1	0.54	ug/kg	
1330-20-7	Xylene (total)	0.78	2.3	0.54	ug/kg	J

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	83%		67-127%
17060-07-0	1,2-Dichloroethane-D4	66%		65-132%
2037-26-5	Toluene-D8	87%		74-129%
460-00-4	4-Bromofluorobenzene	70%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 2

Client Sample ID:	SW-CCI-015	Date Sampled:	08/14/09
Lab Sample ID:	JA25638-1	Date Received:	08/14/09
Matrix:	SO - Soil	Percent Solids:	81.8
Method:	SW846 8260B	Project: Courtlandt Corners I, 364 East 161st Street, Bronx, NY	

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	V93577.D	1	08/14/09	JLI	n/a	n/a	VV3852
Run #2							

Run #1	Initial Weight
Run #1	5.3 g
Run #2	

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	12	2.6	ug/kg	
71-43-2	Benzene	ND	1.2	0.39	ug/kg	
75-27-4	Bromodichloromethane	ND	5.8	0.30	ug/kg	
75-25-2	Bromoform	ND	5.8	0.17	ug/kg	
74-83-9	Bromomethane	ND	5.8	0.47	ug/kg	
78-93-3	2-Butanone (MEK)	ND	12	2.3	ug/kg	
104-51-8	n-Butylbenzene	ND	5.8	0.44	ug/kg	
135-98-8	sec-Butylbenzene	ND	5.8	0.56	ug/kg	
98-06-6	tert-Butylbenzene	ND	5.8	0.55	ug/kg	
75-15-0	Carbon disulfide	ND	5.8	0.35	ug/kg	
56-23-5	Carbon tetrachloride	ND	5.8	0.64	ug/kg	
108-90-7	Chlorobenzene	ND	5.8	0.39	ug/kg	
75-00-3	Chloroethane	ND	5.8	1.3	ug/kg	
67-66-3	Chloroform	ND	5.8	0.37	ug/kg	
74-87-3	Chloromethane	ND	5.8	0.19	ug/kg	
110-82-7	Cyclohexane	ND	5.8	0.17	ug/kg	
96-12-8	1,2-Dibromo-3-chloropropane	ND	12	0.62	ug/kg	
124-48-1	Dibromochloromethane	ND	5.8	0.13	ug/kg	
106-93-4	1,2-Dibromoethane	ND	1.2	0.16	ug/kg	
95-50-1	1,2-Dichlorobenzene	ND	5.8	0.31	ug/kg	
541-73-1	1,3-Dichlorobenzene	ND	5.8	0.32	ug/kg	
106-46-7	1,4-Dichlorobenzene	ND	5.8	0.39	ug/kg	
75-71-8	Dichlorodifluoromethane	ND	5.8	1.1	ug/kg	
75-34-3	1,1-Dichloroethane	ND	5.8	0.16	ug/kg	
107-06-2	1,2-Dichloroethane	ND	1.2	0.40	ug/kg	
75-35-4	1,1-Dichloroethene	ND	5.8	0.76	ug/kg	
156-59-2	cis-1,2-Dichloroethene	ND	5.8	0.28	ug/kg	
156-60-5	trans-1,2-Dichloroethene	ND	5.8	0.52	ug/kg	
78-87-5	1,2-Dichloropropane	ND	5.8	0.15	ug/kg	
10061-01-5	cis-1,3-Dichloropropene	ND	5.8	0.15	ug/kg	
10061-02-6	trans-1,3-Dichloropropene	ND	5.8	0.11	ug/kg	
123-91-1	1,4-Dioxane	ND	140	100	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

3.1
3

Client Sample ID:	SW-CCI-015	Date Sampled:	08/14/09
Lab Sample ID:	JA25638-1	Date Received:	08/14/09
Matrix:	SO - Soil	Percent Solids:	81.8
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	0.63	1.2	0.43	ug/kg	J
76-13-1	Freon 113	ND	5.8	0.65	ug/kg	
591-78-6	2-Hexanone	ND	5.8	1.1	ug/kg	
98-82-8	Isopropylbenzene	ND	5.8	0.60	ug/kg	
79-20-9	Methyl Acetate	ND	5.8	0.95	ug/kg	
108-87-2	Methylcyclohexane	ND	5.8	0.75	ug/kg	
1634-04-4	Methyl Tert Butyl Ether	1.5	1.2	0.33	ug/kg	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	5.8	0.94	ug/kg	
75-09-2	Methylene chloride	ND	5.8	0.26	ug/kg	
103-65-1	n-Propylbenzene	ND	5.8	0.29	ug/kg	
100-42-5	Styrene	ND	5.8	0.12	ug/kg	
79-34-5	1,1,2,2-Tetrachloroethane	ND	5.8	0.34	ug/kg	
127-18-4	Tetrachloroethene	ND	5.8	0.17	ug/kg	
108-88-3	Toluene	ND	1.2	0.34	ug/kg	
120-82-1	1,2,4-Trichlorobenzene	ND	5.8	0.40	ug/kg	
71-55-6	1,1,1-Trichloroethane	ND	5.8	0.15	ug/kg	
79-00-5	1,1,2-Trichloroethane	ND	5.8	0.21	ug/kg	
79-01-6	Trichloroethene	ND	5.8	0.61	ug/kg	
75-69-4	Trichlorofluoromethane	ND	5.8	0.26	ug/kg	
95-63-6	1,2,4-Trimethylbenzene	2.0	5.8	0.50	ug/kg	J
108-67-8	1,3,5-Trimethylbenzene	1.0	5.8	0.41	ug/kg	J
75-01-4	Vinyl chloride	ND	5.8	0.21	ug/kg	
	m,p-Xylene	2.0	2.3	0.54	ug/kg	J
95-47-6	o-Xylene	0.83	1.2	0.54	ug/kg	J
1330-20-7	Xylene (total)	2.9	2.3	0.54	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	118%		67-127%
17060-07-0	1,2-Dichloroethane-D4	123%		65-132%
2037-26-5	Toluene-D8	93%		74-129%
460-00-4	4-Bromofluorobenzene	70%		62-138%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest LabLink@514900 10:37 13-Aug-2009

Report of Analysis

Page 1 of 2

Client Sample ID:	FB072009	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-1	Date Received:	07/21/09
Matrix:	AQ - Field Blank Soil	Percent Solids:	n/a
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	E154072.D	1	07/23/09	OTR	n/a	n/a	VE6790
Run #2							

Run #	Purge Volume
Run #1	5.0 ml
Run #2	

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	10	2.9	ug/l	
71-43-2	Benzene	ND	1.0	0.23	ug/l	
75-27-4	Bromodichloromethane	ND	1.0	0.22	ug/l	
75-25-2	Bromoform	ND	4.0	0.23	ug/l	
74-83-9	Bromomethane	ND	2.0	0.30	ug/l	
78-93-3	2-Butanone (MEK)	ND	10	1.6	ug/l	
104-51-8	n-Butylbenzene	ND	5.0	0.47	ug/l	
135-98-8	sec-Butylbenzene	ND	5.0	0.22	ug/l	
98-06-6	tert-Butylbenzene	ND	5.0	0.21	ug/l	
75-15-0	Carbon disulfide	ND	2.0	0.74	ug/l	
56-23-5	Carbon tetrachloride	ND	1.0	0.26	ug/l	
108-90-7	Chlorobenzene	ND	1.0	0.39	ug/l	
75-00-3	Chloroethane	ND	1.0	0.37	ug/l	
67-66-3	Chloroform	ND	1.0	0.23	ug/l	
74-87-3	Chloromethane	ND	1.0	0.29	ug/l	
110-82-7	Cyclohexane	ND	5.0	1.9	ug/l	
96-12-8	1,2-Dibromo-3-chloropropane	ND	10	1.1	ug/l	
124-48-1	Dibromochloromethane	ND	1.0	0.22	ug/l	
106-93-4	1,2-Dibromoethane	ND	2.0	0.39	ug/l	
95-50-1	1,2-Dichlorobenzene	ND	1.0	0.26	ug/l	
541-73-1	1,3-Dichlorobenzene	ND	1.0	0.25	ug/l	
106-46-7	1,4-Dichlorobenzene	ND	1.0	0.28	ug/l	
75-71-8	Dichlorodifluoromethane	ND	5.0	0.92	ug/l	
75-34-3	1,1-Dichloroethane	ND	1.0	0.29	ug/l	
107-06-2	1,2-Dichloroethane	ND	1.0	0.33	ug/l	
75-35-4	1,1-Dichloroethene	ND	1.0	0.40	ug/l	
156-59-2	cis-1,2-Dichloroethene	ND	1.0	0.22	ug/l	
156-60-5	trans-1,2-Dichloroethene	ND	1.0	0.25	ug/l	
78-87-5	1,2-Dichloropropane	ND	1.0	0.27	ug/l	
10061-01-5	cis-1,3-Dichloropropene	ND	1.0	0.25	ug/l	
10061-02-6	trans-1,3-Dichloropropene	ND	1.0	0.21	ug/l	
123-91-1	1,4-Dioxane	ND	130	94	ug/l	

ND = Not detected MDL - Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	FB072009	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-1	Date Received:	07/21/09
Matrix:	AQ - Field Blank Soil	Percent Solids:	n/a
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	1.0	0.27	ug/l	
76-13-1	Freon 113	ND	5.0	0.38	ug/l	
591-78-6	2-Hexanone	ND	5.0	1.4	ug/l	
98-82-8	Isopropylbenzene	ND	2.0	0.57	ug/l	
79-20-9	Methyl Acetate	ND	5.0	1.5	ug/l	
108-87-2	Methylcyclohexane	ND	5.0	0.35	ug/l	
1634-04-4	Methyl Tert Butyl Ether	ND	1.0	0.23	ug/l	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	5.0	0.86	ug/l	
75-09-2	Methylene chloride	ND	2.0	0.30	ug/l	
103-65-1	n-Propylbenzene	ND	5.0	0.24	ug/l	
100-42-5	Styrene	ND	5.0	0.58	ug/l	
79-34-5	1,1,2,2-Tetrachloroethane	ND	1.0	0.24	ug/l	
127-18-4	Tetrachloroethene	ND	1.0	0.27	ug/l	
108-88-3	Toluene	ND	1.0	0.30	ug/l	
120-82-1	1,2,4-Trichlorobenzene	ND	5.0	0.56	ug/l	
71-55-6	1,1,1-Trichloroethane	ND	1.0	0.26	ug/l	
79-00-5	1,1,2-Trichloroethane	ND	1.0	0.23	ug/l	
79-01-6	Trichloroethene	ND	1.0	0.24	ug/l	
75-69-4	Trichlorofluoromethane	ND	5.0	0.54	ug/l	
95-63-6	1,2,4-Trimethylbenzene	ND	5.0	0.28	ug/l	
108-67-8	1,3,5-Trimethylbenzene	ND	5.0	0.30	ug/l	
75-01-4	Vinyl chloride	ND	1.0	0.44	ug/l	
	m,p-Xylene	ND	1.0	0.25	ug/l	
95-47-6	o-Xylene	ND	1.0	0.25	ug/l	
1330-20-7	Xylene (total)	ND	1.0	0.25	ug/l	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	98%		76-120%
17060-07-0	1,2-Dichloroethane-D4	97%		64-135%
2037-26-5	Toluene-D8	94%		76-117%
460-00-4	4-Bromofluorobenzene	96%		72-122%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 2

Client Sample ID:	TRIP BLANK	Date Sampled:	06/30/09
Lab Sample ID:	JA22090-2	Date Received:	06/30/09
Matrix:	AQ - Trip Blank Soil	Percent Solids:	n/a
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	2E41013.D	1	07/01/09	DPP	n/a	n/a	V2E1818

Run #1	Purge Volume
Run #2	5.0 ml

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	10	2.9	ug/l	
71-43-2	Benzene	ND	1.0	0.23	ug/l	
75-27-4	Bromodichloromethane	ND	1.0	0.22	ug/l	
75-25-2	Bromoform	ND	4.0	0.23	ug/l	
74-83-9	Bromomethane	ND	2.0	0.30	ug/l	
78-93-3	2-Butanone (MEK)	ND	10	1.6	ug/l	
104-51-8	n-Butylbenzene	ND	5.0	0.47	ug/l	
135-98-8	sec-Butylbenzene	ND	5.0	0.22	ug/l	
98-06-6	tert-Butylbenzene	ND	5.0	0.21	ug/l	
75-15-0	Carbon disulfide	ND	2.0	0.74	ug/l	
56-23-5	Carbon tetrachloride	ND	1.0	0.26	ug/l	
108-90-7	Chlorobenzene	ND	1.0	0.39	ug/l	
75-00-3	Chloroethane	ND	1.0	0.37	ug/l	
67-66-3	Chloroform	ND	1.0	0.23	ug/l	
74-87-3	Chloromethane	ND	1.0	0.29	ug/l	
110-82-7	Cyclohexane	ND	5.0	1.9	ug/l	
96-12-8	1,2-Dibromo-3-chloropropane	ND	10	1.1	ug/l	
124-48-1	Dibromochloromethane	ND	1.0	0.22	ug/l	
106-93-4	1,2-Dibromoethane	ND	2.0	0.39	ug/l	
95-50-1	1,2-Dichlorobenzene	ND	1.0	0.26	ug/l	
541-73-1	1,3-Dichlorobenzene	ND	1.0	0.25	ug/l	
106-46-7	1,4-Dichlorobenzene	ND	1.0	0.28	ug/l	
75-71-8	Dichlorodifluoromethane	ND	5.0	0.92	ug/l	
75-34-3	1,1-Dichloroethane	ND	1.0	0.29	ug/l	
107-06-2	1,2-Dichloroethane	ND	1.0	0.33	ug/l	
75-35-4	1,1-Dichloroethene	ND	1.0	0.40	ug/l	
156-59-2	cis-1,2-Dichloroethene	ND	1.0	0.22	ug/l	
156-60-5	trans-1,2-Dichloroethene	ND	1.0	0.25	ug/l	
78-87-5	1,2-Dichloropropane	ND	1.0	0.27	ug/l	
10061-01-5	cis-1,3-Dichloropropene	ND	1.0	0.25	ug/l	
10061-02-6	trans-1,3-Dichloropropene	ND	1.0	0.21	ug/l	
123-91-1	1,4-Dioxane	ND	130	94	ug/l	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	TRIP BLANK	Date Sampled:	06/30/09
Lab Sample ID:	JA22090-2	Date Received:	06/30/09
Matrix:	AQ - Trip Blank Soil	Percent Solids:	n/a
Method:	SW846 8260B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

VOA TCL List (OLM4.2)

CAS No.	Compound	Result	RL	MDL	Units	Q
100-41-4	Ethylbenzene	ND	1.0	0.27	ug/l	
76-13-1	Freon 113	ND	5.0	0.38	ug/l	
591-78-6	2-Hexanone	ND	5.0	1.4	ug/l	
98-82-8	Isopropylbenzene	ND	2.0	0.57	ug/l	
79-20-9	Methyl Acetate	ND	5.0	1.5	ug/l	
108-87-2	Methylcyclohexane	ND	5.0	0.35	ug/l	
1634-04-4	Methyl Tert Butyl Ether	ND	1.0	0.23	ug/l	
108-10-1	4-Methyl-2-pentanone(MIBK)	ND	5.0	0.86	ug/l	
75-09-2	Methylene chloride	ND	2.0	0.30	ug/l	
103-65-1	n-Propylbenzene	ND	5.0	0.24	ug/l	
100-42-5	Styrene	ND	5.0	0.58	ug/l	
79-34-5	1,1,2,2-Tetrachloroethane	ND	1.0	0.24	ug/l	
127-18-4	Tetrachloroethene	ND	1.0	0.27	ug/l	
108-88-3	Toluene	ND	1.0	0.30	ug/l	
120-82-1	1,2,4-Trichlorobenzene	ND	5.0	0.56	ug/l	
71-55-6	1,1,1-Trichloroethane	ND	1.0	0.26	ug/l	
79-00-5	1,1,2-Trichloroethane	ND	1.0	0.23	ug/l	
79-01-6	Trichloroethene	ND	1.0	0.24	ug/l	
75-69-4	Trichlorofluoromethane	ND	5.0	0.54	ug/l	
95-63-6	1,2,4-Trimethylbenzene	ND	5.0	0.28	ug/l	
108-67-8	1,3,5-Trimethylbenzene	ND	5.0	0.30	ug/l	
75-01-4	Vinyl chloride	ND	1.0	0.44	ug/l	
	m,p-Xylene	ND	1.0	0.25	ug/l	
95-47-6	o-Xylene	ND	1.0	0.25	ug/l	
1330-20-7	Xylene (total)	ND	1.0	0.25	ug/l	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
1868-53-7	Dibromofluoromethane	106%		76-120%
17060-07-0	1,2-Dichloroethane-D4	107%		64-135%
2037-26-5	Toluene-D8	107%		76-117%
460-00-4	4-Bromofluorobenzene	98%		72-122%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-02(08)	Date Sampled:	06/30/09
Lab Sample ID:	JA22090-1	Date Received:	06/30/09
Matrix:	SO - Soil	Percent Solids:	83.0
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	M65638.D	1	07/01/09	LP	07/01/09	OP38711	EM2410
Run #2							

Run #	Initial Weight	Final Volume
Run #1	35.0 g	1.0 ml
Run #2		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	170	28	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	170	42	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	170	36	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	170	42	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	690	370	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	690	36	ug/kg	
95-48-7	2-Methylphenol	ND	69	37	ug/kg	
	3&4-Methylphenol	ND	69	46	ug/kg	
88-75-5	2-Nitrophenol	ND	170	36	ug/kg	
100-02-7	4-Nitrophenol	ND	340	44	ug/kg	
87-86-5	Pentachlorophenol	ND	340	44	ug/kg	
108-95-2	Phenol	ND	69	26	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	170	37	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	170	46	ug/kg	
83-32-9	Acenaphthene	ND	34	18	ug/kg	
208-96-8	Acenaphthylene	ND	34	15	ug/kg	
98-86-2	Acetophenone	ND	170	17	ug/kg	
120-12-7	Anthracene	ND	34	15	ug/kg	
1912-24-9	Atrazine	ND	170	22	ug/kg	
56-55-3	Benzo(a)anthracene	35.7	34	21	ug/kg	
50-32-8	Benzo(a)pyrene	35.4	34	14	ug/kg	
205-99-2	Benzo(b)fluoranthene	44.1	34	19	ug/kg	
191-24-2	Benzo(g,h,i)perylene	29.2	34	17	ug/kg	J
207-08-9	Benzo(k)fluoranthene	17.4	34	17	ug/kg	J
101-55-3	4-Bromophenyl phenyl ether	ND	69	19	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	69	17	ug/kg	
92-52-4	1,1'-Biphenyl	ND	69	17	ug/kg	
100-52-7	Benzaldehyde	ND	170	100	ug/kg	
91-58-7	2-Chloronaphthalene	ND	69	15	ug/kg	
106-47-8	4-Chloroaniline	ND	170	14	ug/kg	
86-74-8	Carbazole	ND	69	15	ug/kg	
105-60-2	Caprolactam	ND	69	28	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-02(08)	Date Sampled:	06/30/09
Lab Sample ID:	JA22090-1	Date Received:	06/30/09
Matrix:	SO - Soil	Percent Solids:	83.0
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	38.6	34	16	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	69	17	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	69	16	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	69	18	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	69	23	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	69	19	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	69	16	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	170	60	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	ND	34	16	ug/kg	
132-64-9	Dibenzofuran	ND	69	16	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	69	21	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	69	15	ug/kg	
84-66-2	Diethyl phthalate	ND	69	16	ug/kg	
131-11-3	Dimethyl phthalate	ND	69	16	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	ND	69	18	ug/kg	
206-44-0	Fluoranthene	60.8	34	16	ug/kg	
86-73-7	Fluorene	ND	34	16	ug/kg	
118-74-1	Hexachlorobenzene	ND	69	20	ug/kg	
87-68-3	Hexachlorobutadiene	ND	34	16	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	690	32	ug/kg	
67-72-1	Hexachloroethane	ND	170	22	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	24.7	34	13	ug/kg	J
78-59-1	Isophorone	ND	69	30	ug/kg	
91-57-6	2-Methylnaphthalene	ND	69	15	ug/kg	
88-74-4	2-Nitroaniline	ND	170	25	ug/kg	
99-09-2	3-Nitroaniline	ND	170	14	ug/kg	
100-01-6	4-Nitroaniline	ND	170	22	ug/kg	
91-20-3	Naphthalene	ND	34	15	ug/kg	
98-95-3	Nitrobenzene	ND	69	16	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	69	21	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	170	23	ug/kg	
85-01-8	Phenanthrene	36.4	34	17	ug/kg	
129-00-0	Pyrene	56.0	34	15	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	70%		30-109%
4165-62-2	Phenol-d5	72%		28-108%
118-79-6	2,4,6-Tribromophenol	87%		28-125%
4165-60-0	Nitrobenzene-d5	87%		28-113%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 3

Client Sample ID:	B-CCI-003 (20')	Date Sampled:	07/15/09
Lab Sample ID:	JA23207-1	Date Received:	07/15/09
Matrix:	SO - Soil	Percent Solids:	91.6
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	3E20902.D	1	07/16/09	OYA	07/16/09	OP38913	E3E950

Run #1	Initial Weight	Final Volume
Run #2	35.0 g	1.0 ml

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	160	26	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	160	38	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	160	32	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	160	38	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	620	340	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	620	33	ug/kg	
95-48-7	2-Methylphenol	ND	62	34	ug/kg	
	3&4-Methylphenol	ND	62	42	ug/kg	
88-75-5	2-Nitrophenol	ND	160	33	ug/kg	
100-02-7	4-Nitrophenol	ND	310	40	ug/kg	
87-86-5	Pentachlorophenol	ND	310	40	ug/kg	
108-95-2	Phenol	ND	62	24	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	160	34	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	160	41	ug/kg	
83-32-9	Acenaphthene	ND	31	16	ug/kg	
208-96-8	Acenaphthylene	ND	31	13	ug/kg	
98-86-2	Acetophenone	ND	160	15	ug/kg	
120-12-7	Anthracene	ND	31	14	ug/kg	
1912-24-9	Atrazine	ND	160	20	ug/kg	
56-55-3	Benzo(a)anthracene	ND	31	19	ug/kg	
50-32-8	Benzo(a)pyrene	ND	31	13	ug/kg	
205-99-2	Benzo(b)fluoranthene	ND	31	17	ug/kg	
191-24-2	Benzo(g,h,i)perylene	ND	31	15	ug/kg	
207-08-9	Benzo(k)fluoranthene	ND	31	16	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	62	17	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	62	16	ug/kg	
92-52-4	1,1'-Biphenyl	ND	62	16	ug/kg	
100-52-7	Benzaldehyde	ND	160	91	ug/kg	
91-58-7	2-Chloronaphthalene	ND	62	14	ug/kg	
106-47-8	4-Chloroaniline	ND	160	13	ug/kg	
86-74-8	Carbazole	ND	62	13	ug/kg	
105-60-2	Caprolactam	ND	62	25	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-003 (20')	Date Sampled:	07/15/09
Lab Sample ID:	JA23207-1	Date Received:	07/15/09
Matrix:	SO - Soil	Percent Solids:	91.6
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	ND	31	14	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	62	16	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	62	15	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	62	16	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	62	21	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	62	17	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	62	14	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	160	54	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	ND	31	15	ug/kg	
132-64-9	Dibenzofuran	ND	62	15	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	62	19	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	62	14	ug/kg	
84-66-2	Diethyl phthalate	ND	62	14	ug/kg	
131-11-3	Dimethyl phthalate	ND	62	15	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	ND	62	16	ug/kg	
206-44-0	Fluoranthene	15.1	31	14	ug/kg	J
86-73-7	Fluorene	ND	31	15	ug/kg	
118-74-1	Hexachlorobenzene	ND	62	18	ug/kg	
87-68-3	Hexachlorobutadiene	ND	31	15	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	620	29	ug/kg	
67-72-1	Hexachloroethane	ND	160	20	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	ND	31	12	ug/kg	
78-59-1	Isophorone	ND	62	27	ug/kg	
91-57-6	2-Methylnaphthalene	ND	62	14	ug/kg	
88-74-4	2-Nitroaniline	ND	160	23	ug/kg	
99-09-2	3-Nitroaniline	ND	160	13	ug/kg	
100-01-6	4-Nitroaniline	ND	160	19	ug/kg	
91-20-3	Naphthalene	ND	31	14	ug/kg	
98-95-3	Nitrobenzene	ND	62	14	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	62	19	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	160	21	ug/kg	
85-01-8	Phenanthrene	15.6	31	15	ug/kg	J
129-00-0	Pyrene	14.1	31	14	ug/kg	J

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	66%		30-109%
4165-62-2	Phenol-d5	74%		28-108%
118-79-6	2,4,6-Tribromophenol	68%		28-125%
4165-60-0	Nitrobenzene-d5	85%		28-113%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 3

Client Sample ID:	B-CCI-004 (20')	Date Sampled:	07/15/09
Lab Sample ID:	JA23207-2	Date Received:	07/15/09
Matrix:	SO - Soil	Percent Solids:	91.6
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	3E20903.D	1	07/16/09	OYA	07/16/09	OP38913	E3E950

Run #1	Initial Weight	Final Volume
Run #2	35.2 g	1.0 ml

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	160	26	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	160	38	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	160	32	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	160	38	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	620	330	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	620	33	ug/kg	
95-48-7	2-Methylphenol	ND	62	33	ug/kg	
	3&4-Methylphenol	ND	62	42	ug/kg	
88-75-5	2-Nitrophenol	ND	160	33	ug/kg	
100-02-7	4-Nitrophenol	ND	310	40	ug/kg	
87-86-5	Pentachlorophenol	ND	310	40	ug/kg	
108-95-2	Phenol	ND	62	23	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	160	33	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	160	41	ug/kg	
83-32-9	Acenaphthene	ND	31	16	ug/kg	
208-96-8	Acenaphthylene	ND	31	13	ug/kg	
98-86-2	Acetophenone	ND	160	15	ug/kg	
120-12-7	Anthracene	ND	31	14	ug/kg	
1912-24-9	Atrazine	ND	160	20	ug/kg	
56-55-3	Benzo(a)anthracene	ND	31	19	ug/kg	
50-32-8	Benzo(a)pyrene	ND	31	13	ug/kg	
205-99-2	Benzo(b)fluoranthene	ND	31	17	ug/kg	
191-24-2	Benzo(g,h,i)perylene	ND	31	15	ug/kg	
207-08-9	Benzo(k)fluoranthene	ND	31	16	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	62	17	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	62	15	ug/kg	
92-52-4	1,1'-Biphenyl	ND	62	16	ug/kg	
100-52-7	Benzaldehyde	ND	160	91	ug/kg	
91-58-7	2-Chloronaphthalene	ND	62	14	ug/kg	
106-47-8	4-Chloroaniline	ND	160	13	ug/kg	
86-74-8	Carbazole	ND	62	13	ug/kg	
105-60-2	Caprolactam	ND	62	25	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-004 (20')	Date Sampled:	07/15/09
Lab Sample ID:	JA23207-2	Date Received:	07/15/09
Matrix:	SO - Soil	Percent Solids:	91.6
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	ND	31	14	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	62	15	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	62	15	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	62	16	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	62	21	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	62	17	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	62	14	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	160	54	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	ND	31	15	ug/kg	
132-64-9	Dibenzofuran	ND	62	14	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	62	19	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	62	14	ug/kg	
84-66-2	Diethyl phthalate	ND	62	14	ug/kg	
131-11-3	Dimethyl phthalate	ND	62	15	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	ND	62	16	ug/kg	
206-44-0	Fluoranthene	ND	31	14	ug/kg	
86-73-7	Fluorene	ND	31	15	ug/kg	
118-74-1	Hexachlorobenzene	ND	62	18	ug/kg	
87-68-3	Hexachlorobutadiene	ND	31	15	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	620	29	ug/kg	
67-72-1	Hexachloroethane	ND	160	20	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	ND	31	11	ug/kg	
78-59-1	Isophorone	ND	62	27	ug/kg	
91-57-6	2-Methylnaphthalene	ND	62	14	ug/kg	
88-74-4	2-Nitroaniline	ND	160	23	ug/kg	
99-09-2	3-Nitroaniline	ND	160	13	ug/kg	
100-01-6	4-Nitroaniline	ND	160	19	ug/kg	
91-20-3	Naphthalene	ND	31	13	ug/kg	
98-95-3	Nitrobenzene	ND	62	14	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	62	19	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	160	21	ug/kg	
85-01-8	Phenanthrene	ND	31	15	ug/kg	
129-00-0	Pyrene	ND	31	14	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	64%		30-109%
4165-62-2	Phenol-d5	72%		28-108%
118-79-6	2,4,6-Tribromophenol	73%		28-125%
4165-60-0	Nitrobenzene-d5	77%		28-113%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 3

Client Sample ID:	B-CCI-006 (15')	Date Sampled:	07/22/09
Lab Sample ID:	JA23843-1	Date Received:	07/23/09
Matrix:	SO - Soil	Percent Solids:	90.6
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	Z49706.D	1	07/24/09	VN	07/24/09	OP39041	EZ1799
Run #2							

Run #1	Initial Weight	Final Volume
Run #1	35.0 g	1.0 ml
Run #2		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	160	26	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	160	38	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	160	33	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	160	38	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	630	340	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	630	33	ug/kg	
95-48-7	2-Methylphenol	ND	63	34	ug/kg	
	3&4-Methylphenol	ND	63	42	ug/kg	
88-75-5	2-Nitrophenol	ND	160	33	ug/kg	
100-02-7	4-Nitrophenol	ND	320	40	ug/kg	
87-86-5	Pentachlorophenol	ND	320	41	ug/kg	
108-95-2	Phenol	ND	63	24	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	160	34	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	160	42	ug/kg	
83-32-9	Acenaphthene	ND	32	17	ug/kg	
208-96-8	Acenaphthylene	ND	32	14	ug/kg	
98-86-2	Acetophenone	ND	160	15	ug/kg	
120-12-7	Anthracene	ND	32	14	ug/kg	
1912-24-9	Atrazine	ND	160	20	ug/kg	
56-55-3	Benzo(a)anthracene	ND	32	19	ug/kg	
50-32-8	Benzo(a)pyrene	ND	32	13	ug/kg	
205-99-2	Benzo(b)fluoranthene	ND	32	17	ug/kg	
191-24-2	Benzo(g,h,i)perylene	ND	32	15	ug/kg	
207-08-9	Benzo(k)fluoranthene	ND	32	16	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	63	17	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	63	16	ug/kg	
92-52-4	1,1'-Biphenyl	ND	63	16	ug/kg	
100-52-7	Benzaldehyde	ND	160	92	ug/kg	
91-58-7	2-Chloronaphthalene	ND	63	14	ug/kg	
106-47-8	4-Chloroaniline	ND	160	13	ug/kg	
86-74-8	Carbazole	ND	63	13	ug/kg	
105-60-2	Caprolactam	ND	63	25	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-006 (15')	Date Sampled:	07/22/09
Lab Sample ID:	JA23843-1	Date Received:	07/23/09
Matrix:	SO - Soil	Percent Solids:	90.6
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	ND	32	15	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	63	16	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	63	15	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	63	16	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	63	21	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	63	17	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	63	14	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	160	55	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	ND	32	15	ug/kg	
132-64-9	Dibenzofuran	ND	63	15	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	63	20	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	63	14	ug/kg	
84-66-2	Diethyl phthalate	ND	63	14	ug/kg	
131-11-3	Dimethyl phthalate	ND	63	15	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	ND	63	16	ug/kg	
206-44-0	Fluoranthene	ND	32	14	ug/kg	
86-73-7	Fluorene	ND	32	15	ug/kg	
118-74-1	Hexachlorobenzene	ND	63	18	ug/kg	
87-68-3	Hexachlorobutadiene	ND	32	15	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	630	30	ug/kg	
67-72-1	Hexachloroethane	ND	160	20	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	ND	32	12	ug/kg	
78-59-1	Isophorone	ND	63	28	ug/kg	
91-57-6	2-Methylnaphthalene	ND	63	14	ug/kg	
88-74-4	2-Nitroaniline	ND	160	23	ug/kg	
99-09-2	3-Nitroaniline	ND	160	13	ug/kg	
100-01-6	4-Nitroaniline	ND	160	20	ug/kg	
91-20-3	Naphthalene	ND	32	14	ug/kg	
98-95-3	Nitrobenzene	ND	63	14	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	63	20	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	160	21	ug/kg	
85-01-8	Phenanthrene	ND	32	16	ug/kg	
129-00-0	Pyrene	ND	32	14	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	59%		30-109%
4165-62-2	Phenol-d5	51%		28-108%
118-79-6	2,4,6-Tribromophenol	63%		28-125%
4165-60-0	Nitrobenzene-d5	66%		28-113%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

3.2
3

Client Sample ID:	B-CCI-007 (20')	Date Sampled:	07/22/09
Lab Sample ID:	JA23843-2	Date Received:	07/23/09
Matrix:	SO - Soil	Percent Solids:	94.5
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	Z49707.D	1	07/24/09	VN	07/24/09	OP39041	EZ1799

Run #1	Initial Weight	Final Volume
Run #2	35.0 g	1.0 ml

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	150	25	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	150	37	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	150	31	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	150	37	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	600	330	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	600	32	ug/kg	
95-48-7	2-Methylphenol	ND	60	33	ug/kg	
	3&4-Methylphenol	ND	60	41	ug/kg	
88-75-5	2-Nitrophenol	ND	150	32	ug/kg	
100-02-7	4-Nitrophenol	ND	300	39	ug/kg	
87-86-5	Pentachlorophenol	ND	300	39	ug/kg	
108-95-2	Phenol	ND	60	23	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	150	33	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	150	40	ug/kg	
83-32-9	Acenaphthene	32.1	30	16	ug/kg	
208-96-8	Acenaphthylene	ND	30	13	ug/kg	
98-86-2	Acetophenone	ND	150	15	ug/kg	
120-12-7	Anthracene	81.5	30	14	ug/kg	
1912-24-9	Atrazine	ND	150	19	ug/kg	
56-55-3	Benzo(a)anthracene	41.9	30	18	ug/kg	
50-32-8	Benzo(a)pyrene	24.3	30	13	ug/kg	J
205-99-2	Benzo(b)fluoranthene	ND	30	16	ug/kg	J
191-24-2	Benzo(g,h,i)perylene	23.3	30	15	ug/kg	J
207-08-9	Benzo(k)fluoranthene	ND	30	15	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	60	16	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	60	15	ug/kg	
92-52-4	1,1'-Biphenyl	30.5	60	15	ug/kg	J
100-52-7	Benzaldehyde	ND	150	88	ug/kg	
91-58-7	2-Chloronaphthalene	ND	60	13	ug/kg	
106-47-8	4-Chloroaniline	ND	150	12	ug/kg	
86-74-8	Carbazole	ND	60	13	ug/kg	
105-60-2	Caprolactam	ND	60	24	ug/kg	

ND = Not detected MDL - Method Detection Limit J = Indicates an estimated value
 RL = Reporting Limit B = Indicates analyte found in associated method blank
 E = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Report of Analysis

3.2
3

Client Sample ID:	B-CCI-007 (20')	Date Sampled:	07/22/09
Lab Sample ID:	JA23843-2	Date Received:	07/23/09
Matrix:	SO - Soil	Percent Solids:	94.5
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	26.9	30	14	ug/kg	J
111-91-1	bis(2-Chloroethoxy)methane	ND	60	15	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	60	14	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	60	16	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	60	20	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	60	16	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	60	14	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	150	52	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	ND	30	14	ug/kg	
132-64-9	Dibenzofuran	ND	60	14	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	60	19	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	60	13	ug/kg	
84-66-2	Diethyl phthalate	ND	60	14	ug/kg	
131-11-3	Dimethyl phthalate	ND	60	14	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	ND	60	16	ug/kg	
206-44-0	Fluoranthene	55.9	30	14	ug/kg	
86-73-7	Fluorene	84.4	30	14	ug/kg	
118-74-1	Hexachlorobenzene	ND	60	17	ug/kg	
87-68-3	Hexachlorobutadiene	ND	30	14	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	600	29	ug/kg	
67-72-1	Hexachloroethane	ND	150	20	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	ND	30	11	ug/kg	J
78-59-1	Isophorone	ND	60	27	ug/kg	
91-57-6	2-Methylnaphthalene	945	60	14	ug/kg	
88-74-4	2-Nitroaniline	ND	150	22	ug/kg	
99-09-2	3-Nitroaniline	ND	150	12	ug/kg	
100-01-6	4-Nitroaniline	ND	150	19	ug/kg	
91-20-3	Naphthalene	118	30	13	ug/kg	J
98-95-3	Nitrobenzene	ND	60	14	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	60	19	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	150	20	ug/kg	
85-01-8	Phenanthrene	191	30	15	ug/kg	
129-00-0	Pyrene	98.3	30	13	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	61%		30-109%
4165-62-2	Phenol-d5	51%		28-108%
118-79-6	2,4,6-Tribromophenol	64%		28-125%
4165-60-0	Nitrobenzene-d5	66%		28-113%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 3

Client Sample ID:	DUPLICATE		Date Sampled:	07/22/09
Lab Sample ID:	JA23843-5	Date Received:	07/23/09	
Matrix:	SO - Soil		Percent Solids:	94.2
Method:	SW846 8270C SW846 3550B			
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY			

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	Z49708.D	1	07/24/09	VN	07/24/09	OP39041	EZ1799
Run #2							

Run #	Initial Weight	Final Volume
Run #1	35.1 g	1.0 ml
Run #2		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	150	25	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	150	37	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	150	31	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	150	37	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	600	330	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	600	32	ug/kg	
95-48-7	2-Methylphenol	ND	60	33	ug/kg	
	3&4-Methylphenol	ND	60	41	ug/kg	
88-75-5	2-Nitrophenol	ND	150	32	ug/kg	
100-02-7	4-Nitrophenol	ND	300	39	ug/kg	
87-86-5	Pentachlorophenol	ND	300	39	ug/kg	
108-95-2	Phenol	ND	60	23	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	150	33	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	150	40	ug/kg	
83-32-9	Acenaphthene	31.5	30	16	ug/kg	
208-96-8	Acenaphthylene	ND	30	13	ug/kg	
98-86-2	Acetophenone	ND	150	15	ug/kg	
120-12-7	Anthracene	84.5	30	14	ug/kg	
1912-24-9	Atrazine	ND	150	19	ug/kg	
56-55-3	Benzo(a)anthracene	43.5	30	18	ug/kg	
50-32-8	Benzo(a)pyrene	28.7	30	13	ug/kg	J
205-99-2	Benzo(b)fluoranthene	18.7	30	16	ug/kg	J
191-24-2	Benzo(g,h,i)perylene	28.8	30	15	ug/kg	J
207-08-9	Benzo(k)fluoranthene	ND	30	15	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	60	16	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	60	15	ug/kg	
92-52-4	1,1'-Biphenyl	27.5	60	15	ug/kg	J
100-52-7	Benzaldehyde	ND	150	88	ug/kg	
91-58-7	2-Chloronaphthalene	ND	60	13	ug/kg	
106-47-8	4-Chloroaniline	ND	150	12	ug/kg	
86-74-8	Carbazole	ND	60	13	ug/kg	
105-60-2	Caprolactam	ND	60	24	ug/kg	

ND = Not detected MDL - Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

Report of Analysis

3.5
3

Client Sample ID:	DUPLICATE	Date Sampled:	07/22/09
Lab Sample ID:	JA23843-5	Date Received:	07/23/09
Matrix:	SO - Soil	Percent Solids:	94.2
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	29.0	30	14	ug/kg	J
111-91-1	bis(2-Chloroethoxy)methane	ND	60	15	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	60	14	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	60	16	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	60	20	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	60	16	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	60	14	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	150	52	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	ND	30	14	ug/kg	
132-64-9	Dibenzofuran	ND	60	14	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	60	19	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	60	13	ug/kg	
84-66-2	Diethyl phthalate	ND	60	14	ug/kg	
131-11-3	Dimethyl phthalate	ND	60	14	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	ND	60	16	ug/kg	
206-44-0	Fluoranthene	59.0	30	14	ug/kg	
86-73-7	Fluorene	87.6	30	14	ug/kg	
118-74-1	Hexachlorobenzene	ND	60	17	ug/kg	
87-68-3	Hexachlorobutadiene	ND	30	14	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	600	29	ug/kg	
67-72-1	Hexachloroethane	ND	150	20	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	11.3	30	11	ug/kg	J
78-59-1	Isophorone	ND	60	27	ug/kg	
91-57-6	2-Methylnaphthalene	400	60	14	ug/kg	
88-74-4	2-Nitroaniline	ND	150	22	ug/kg	
99-09-2	3-Nitroaniline	ND	150	12	ug/kg	
100-01-6	4-Nitroaniline	ND	150	19	ug/kg	
91-20-3	Naphthalene	ND	30	13	ug/kg	J
98-95-3	Nitrobenzene	ND	60	14	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	60	19	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	150	20	ug/kg	
85-01-8	Phenanthrene	195	30	15	ug/kg	
129-00-0	Pyrene	105	30	13	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	55%		30-109%
4165-62-2	Phenol-d5	46%		28-108%
118-79-6	2,4,6-Tribromophenol	54%		28-125%
4165-60-0	Nitrobenzene-d5	61%		28-113%

ND = Not detected MDL - Method Detection Limit J = Indicates an estimated value
 RL = Reporting Limit B = Indicates analyte found in associated method blank
 E = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 3

3.1
3

Client Sample ID:	B-CCI-010(22')	Date Sampled:	07/24/09
Lab Sample ID:	JA24059-1	Date Received:	07/27/09
Matrix:	SO - Soil	Percent Solids:	82.1
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	P44136.D	1	07/28/09	NAP	07/28/09	OP39098	EP1878

Run #1	Initial Weight	Final Volume
Run #2	35.0 g	1.0 ml

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	170	29	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	170	42	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	170	36	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	170	42	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	700	380	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	700	37	ug/kg	
95-48-7	2-Methylphenol	ND	70	38	ug/kg	
	3&4-Methylphenol	ND	70	47	ug/kg	
88-75-5	2-Nitrophenol	ND	170	37	ug/kg	
100-02-7	4-Nitrophenol	ND	350	45	ug/kg	
87-86-5	Pentachlorophenol	ND	350	45	ug/kg	
108-95-2	Phenol	ND	70	26	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	170	38	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	170	46	ug/kg	
83-32-9	Acenaphthene	ND	35	18	ug/kg	
208-96-8	Acenaphthylene	ND	35	15	ug/kg	
98-86-2	Acetophenone	ND	170	17	ug/kg	
120-12-7	Anthracene	ND	35	16	ug/kg	
1912-24-9	Atrazine	ND	170	22	ug/kg	
56-55-3	Benzo(a)anthracene	ND	35	21	ug/kg	
50-32-8	Benzo(a)pyrene	ND	35	14	ug/kg	
205-99-2	Benzo(b)fluoranthene	ND	35	19	ug/kg	
191-24-2	Benzo(g,h,i)perylene	ND	35	17	ug/kg	
207-08-9	Benzo(k)fluoranthene	ND	35	17	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	70	19	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	70	17	ug/kg	
92-52-4	1,1'-Biphenyl	ND	70	18	ug/kg	
100-52-7	Benzaldehyde	ND	170	100	ug/kg	
91-58-7	2-Chloronaphthalene	ND	70	15	ug/kg	
106-47-8	4-Chloroaniline	ND	170	14	ug/kg	
86-74-8	Carbazole	ND	70	15	ug/kg	
105-60-2	Caprolactam	ND	70	28	ug/kg	

ND = Not detected MDL - Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-010(22')	Date Sampled:	07/24/09
Lab Sample ID:	JA24059-1	Date Received:	07/27/09
Matrix:	SO - Soil	Percent Solids:	82.1
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	ND	35	16	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	70	17	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	70	16	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	70	18	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	70	23	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	70	19	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	70	16	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	170	60	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	ND	35	17	ug/kg	
132-64-9	Dibenzofuran	ND	70	16	ug/kg	
84-74-2	Di-n-butyl phthalate	98.6	70	22	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	70	15	ug/kg	
84-66-2	Diethyl phthalate	ND	70	16	ug/kg	
131-11-3	Dimethyl phthalate	ND	70	17	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	46.5	70	18	ug/kg	J
206-44-0	Fluoranthene	ND	35	16	ug/kg	
86-73-7	Fluorene	ND	35	16	ug/kg	
118-74-1	Hexachlorobenzene	ND	70	20	ug/kg	
87-68-3	Hexachlorobutadiene	ND	35	17	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	700	33	ug/kg	
67-72-1	Hexachloroethane	ND	170	22	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	ND	35	13	ug/kg	
78-59-1	Isophorone	ND	70	31	ug/kg	
91-57-6	2-Methylnaphthalene	ND	70	16	ug/kg	
88-74-4	2-Nitroaniline	ND	170	26	ug/kg	
99-09-2	3-Nitroaniline	ND	170	14	ug/kg	
100-01-6	4-Nitroaniline	ND	170	22	ug/kg	
91-20-3	Naphthalene	ND	35	15	ug/kg	
98-95-3	Nitrobenzene	ND	70	16	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	70	22	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	170	23	ug/kg	
85-01-8	Phenanthrene	ND	35	17	ug/kg	
129-00-0	Pyrene	ND	35	15	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	70%		30-109%
4165-62-2	Phenol-d5	75%		28-108%
118-79-6	2,4,6-Tribromophenol	81%		28-125%
4165-60-0	Nitrobenzene-d5	91%		28-113%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 3

Client Sample ID:	B-CCI-013(10')	Date Sampled:	07/27/09
Lab Sample ID:	JA24059-3	Date Received:	07/27/09
Matrix:	SO - Soil	Percent Solids:	82.9
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	P44139.D	1	07/28/09	NAP	07/28/09	OP39098	EP1878
Run #2							

Run #	Initial Weight	Final Volume
Run #1	35.4 g	1.0 ml
Run #2		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	170	28	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	170	41	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	170	35	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	170	42	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	680	370	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	680	36	ug/kg	
95-48-7	2-Methylphenol	ND	68	37	ug/kg	
	3&4-Methylphenol	ND	68	46	ug/kg	
88-75-5	2-Nitrophenol	ND	170	36	ug/kg	
100-02-7	4-Nitrophenol	ND	340	44	ug/kg	
87-86-5	Pentachlorophenol	ND	340	44	ug/kg	
108-95-2	Phenol	ND	68	26	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	170	37	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	170	45	ug/kg	
83-32-9	Acenaphthene	ND	34	18	ug/kg	
208-96-8	Acenaphthylene	ND	34	15	ug/kg	
98-86-2	Acetophenone	ND	170	17	ug/kg	
120-12-7	Anthracene	18.2	34	15	ug/kg	J
1912-24-9	Atrazine	ND	170	21	ug/kg	
56-55-3	Benzo(a)anthracene	105	34	21	ug/kg	
50-32-8	Benzo(a)pyrene	109	34	14	ug/kg	
205-99-2	Benzo(b)fluoranthene	112	34	19	ug/kg	
191-24-2	Benzo(g,h,i)perylene	77.7	34	17	ug/kg	
207-08-9	Benzo(k)fluoranthene	80.6	34	17	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	68	19	ug/kg	
85-68-7	Butyl benzyl phthalate	48.8	68	17	ug/kg	J
92-52-4	1,1'-Biphenyl	ND	68	17	ug/kg	
100-52-7	Benzaldehyde	ND	170	100	ug/kg	
91-58-7	2-Chloronaphthalene	ND	68	15	ug/kg	
106-47-8	4-Chloroaniline	ND	170	14	ug/kg	
86-74-8	Carbazole	16.0	68	15	ug/kg	J
105-60-2	Caprolactam	ND	68	27	ug/kg	

ND = Not detected MDL - Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-013(10')	Date Sampled:	07/27/09
Lab Sample ID:	JA24059-3	Date Received:	07/27/09
Matrix:	SO - Soil	Percent Solids:	82.9
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	114	34	16	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	68	17	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	68	16	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	68	18	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	68	23	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	68	19	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	68	15	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	170	59	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	17.4	34	16	ug/kg	J
132-64-9	Dibenzofuran	ND	68	16	ug/kg	
84-74-2	Di-n-butyl phthalate	125	68	21	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	68	15	ug/kg	
84-66-2	Diethyl phthalate	ND	68	15	ug/kg	
131-11-3	Dimethyl phthalate	ND	68	16	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	81.7	68	18	ug/kg	
206-44-0	Fluoranthene	210	34	16	ug/kg	
86-73-7	Fluorene	ND	34	16	ug/kg	
118-74-1	Hexachlorobenzene	ND	68	19	ug/kg	
87-68-3	Hexachlorobutadiene	ND	34	16	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	680	32	ug/kg	
67-72-1	Hexachloroethane	ND	170	22	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	65.7	34	13	ug/kg	
78-59-1	Isophorone	ND	68	30	ug/kg	
91-57-6	2-Methylnaphthalene	ND	68	15	ug/kg	
88-74-4	2-Nitroaniline	ND	170	25	ug/kg	
99-09-2	3-Nitroaniline	ND	170	14	ug/kg	
100-01-6	4-Nitroaniline	ND	170	21	ug/kg	
91-20-3	Naphthalene	ND	34	15	ug/kg	
98-95-3	Nitrobenzene	ND	68	15	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	68	21	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	170	23	ug/kg	
85-01-8	Phenanthrene	84.8	34	17	ug/kg	
129-00-0	Pyrene	247	34	15	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	66%		30-109%
4165-62-2	Phenol-d5	72%		28-108%
118-79-6	2,4,6-Tribromophenol	80%		28-125%
4165-60-0	Nitrobenzene-d5	87%		28-113%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 3

Client Sample ID:	B-CCI-022	Date Sampled:	08/04/09
Lab Sample ID:	JA24692-1	Date Received:	08/04/09
Matrix:	SO - Soil	Percent Solids:	87.0
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	3E21219.D	1	08/05/09	OYA	08/05/09	OP39251	E3E964

Run #1	Initial Weight	Final Volume
Run #2	35.2 g	1.0 ml

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	160	27	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	160	40	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	160	34	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	160	40	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	650	350	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	650	35	ug/kg	
95-48-7	2-Methylphenol	ND	65	35	ug/kg	
	3&4-Methylphenol	ND	65	44	ug/kg	
88-75-5	2-Nitrophenol	ND	160	34	ug/kg	
100-02-7	4-Nitrophenol	ND	330	42	ug/kg	
87-86-5	Pentachlorophenol	ND	330	42	ug/kg	
108-95-2	Phenol	ND	65	25	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	160	35	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	160	43	ug/kg	
83-32-9	Acenaphthene	ND	33	17	ug/kg	
208-96-8	Acenaphthylene	27.3	33	14	ug/kg	J
98-86-2	Acetophenone	ND	160	16	ug/kg	
120-12-7	Anthracene	26.9	33	15	ug/kg	J
1912-24-9	Atrazine	ND	160	21	ug/kg	
56-55-3	Benzo(a)anthracene	137	33	20	ug/kg	
50-32-8	Benzo(a)pyrene	162	33	14	ug/kg	
205-99-2	Benzo(b)fluoranthene	176	33	18	ug/kg	
191-24-2	Benzo(g,h,i)perylene	111	33	16	ug/kg	
207-08-9	Benzo(k)fluoranthene	75.7	33	16	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	65	18	ug/kg	
85-68-7	Butyl benzyl phthalate	78.6	65	16	ug/kg	
92-52-4	1,1'-Biphenyl	ND	65	16	ug/kg	
100-52-7	Benzaldehyde	ND	160	95	ug/kg	
91-58-7	2-Chloronaphthalene	ND	65	14	ug/kg	
106-47-8	4-Chloroaniline	ND	160	13	ug/kg	
86-74-8	Carbazole	ND	65	14	ug/kg	
105-60-2	Caprolactam	ND	65	26	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-022	Date Sampled:	08/04/09
Lab Sample ID:	JA24692-1	Date Received:	08/04/09
Matrix:	SO - Soil	Percent Solids:	87.0
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	143	33	15	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	65	16	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	65	15	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	65	17	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	65	22	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	65	18	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	65	15	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	160	56	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	32.4	33	16	ug/kg	J
132-64-9	Dibenzofuran	ND	65	15	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	65	20	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	65	14	ug/kg	
84-66-2	Diethyl phthalate	ND	65	15	ug/kg	
131-11-3	Dimethyl phthalate	ND	65	16	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	70.2	65	17	ug/kg	
206-44-0	Fluoranthene	271	33	15	ug/kg	
86-73-7	Fluorene	ND	33	15	ug/kg	
118-74-1	Hexachlorobenzene	ND	65	19	ug/kg	
87-68-3	Hexachlorobutadiene	ND	33	16	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	650	31	ug/kg	
67-72-1	Hexachloroethane	ND	160	21	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	98.8	33	12	ug/kg	
78-59-1	Isophorone	ND	65	29	ug/kg	
91-57-6	2-Methylnaphthalene	ND	65	15	ug/kg	
88-74-4	2-Nitroaniline	ND	160	24	ug/kg	
99-09-2	3-Nitroaniline	ND	160	13	ug/kg	
100-01-6	4-Nitroaniline	ND	160	20	ug/kg	
91-20-3	Naphthalene	ND	33	14	ug/kg	
98-95-3	Nitrobenzene	ND	65	15	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	65	20	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	160	22	ug/kg	
85-01-8	Phenanthrene	141	33	16	ug/kg	
129-00-0	Pyrene	250	33	14	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	60%		30-109%
4165-62-2	Phenol-d5	65%		28-108%
118-79-6	2,4,6-Tribromophenol	75%		28-125%
4165-60-0	Nitrobenzene-d5	90%		28-113%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 3

Client Sample ID:	B-CCI-023	Date Sampled:	08/04/09
Lab Sample ID:	JA24692-2	Date Received:	08/04/09
Matrix:	SO - Soil	Percent Solids:	84.1
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	3E21217.D	1	08/05/09	OYA	08/05/09	OP39251	E3E964
Run #2							

Run #	Initial Weight	Final Volume
Run #1	35.2 g	1.0 ml
Run #2		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	170	28	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	170	41	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	170	35	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	170	41	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	680	360	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	680	36	ug/kg	
95-48-7	2-Methylphenol	ND	68	36	ug/kg	
	3&4-Methylphenol	ND	68	45	ug/kg	
88-75-5	2-Nitrophenol	ND	170	35	ug/kg	
100-02-7	4-Nitrophenol	ND	340	43	ug/kg	
87-86-5	Pentachlorophenol	ND	340	44	ug/kg	
108-95-2	Phenol	ND	68	26	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	170	36	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	170	45	ug/kg	
83-32-9	Acenaphthene	ND	34	18	ug/kg	
208-96-8	Acenaphthylene	ND	34	14	ug/kg	
98-86-2	Acetophenone	ND	170	17	ug/kg	
120-12-7	Anthracene	ND	34	15	ug/kg	
1912-24-9	Atrazine	ND	170	21	ug/kg	
56-55-3	Benzo(a)anthracene	ND	34	20	ug/kg	
50-32-8	Benzo(a)pyrene	ND	34	14	ug/kg	
205-99-2	Benzo(b)fluoranthene	ND	34	18	ug/kg	
191-24-2	Benzo(g,h,i)perylene	ND	34	16	ug/kg	
207-08-9	Benzo(k)fluoranthene	ND	34	17	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	68	18	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	68	17	ug/kg	
92-52-4	1,1'-Biphenyl	ND	68	17	ug/kg	
100-52-7	Benzaldehyde	ND	170	99	ug/kg	
91-58-7	2-Chloronaphthalene	ND	68	15	ug/kg	
106-47-8	4-Chloroaniline	ND	170	14	ug/kg	
86-74-8	Carbazole	ND	68	14	ug/kg	
105-60-2	Caprolactam	ND	68	27	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-023	Date Sampled:	08/04/09
Lab Sample ID:	JA24692-2	Date Received:	08/04/09
Matrix:	SO - Soil	Percent Solids:	84.1
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	ND	34	16	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	68	17	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	68	16	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	68	18	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	68	22	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	68	18	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	68	15	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	170	58	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	ND	34	16	ug/kg	
132-64-9	Dibenzofuran	ND	68	16	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	68	21	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	68	15	ug/kg	
84-66-2	Diethyl phthalate	ND	68	15	ug/kg	
131-11-3	Dimethyl phthalate	ND	68	16	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	39.3	68	18	ug/kg	J
206-44-0	Fluoranthene	ND	34	15	ug/kg	
86-73-7	Fluorene	ND	34	16	ug/kg	
118-74-1	Hexachlorobenzene	ND	68	19	ug/kg	
87-68-3	Hexachlorobutadiene	ND	34	16	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	680	32	ug/kg	
67-72-1	Hexachloroethane	ND	170	22	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	ND	34	12	ug/kg	
78-59-1	Isophorone	ND	68	30	ug/kg	
91-57-6	2-Methylnaphthalene	ND	68	15	ug/kg	
88-74-4	2-Nitroaniline	ND	170	25	ug/kg	
99-09-2	3-Nitroaniline	ND	170	14	ug/kg	
100-01-6	4-Nitroaniline	ND	170	21	ug/kg	
91-20-3	Naphthalene	ND	34	15	ug/kg	
98-95-3	Nitrobenzene	ND	68	15	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	68	21	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	170	23	ug/kg	
85-01-8	Phenanthrene	ND	34	17	ug/kg	
129-00-0	Pyrene	ND	34	15	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	58%		30-109%
4165-62-2	Phenol-d5	60%		28-108%
118-79-6	2,4,6-Tribromophenol	75%		28-125%
4165-60-0	Nitrobenzene-d5	87%		28-113%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 3

Client Sample ID:	B-CCI-024	Date Sampled:	08/04/09
Lab Sample ID:	JA24692-3	Date Received:	08/04/09
Matrix:	SO - Soil	Percent Solids:	89.0
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	3E21218.D	1	08/05/09	OYA	08/05/09	OP39251	E3E964
Run #2							

Run #	Initial Weight	Final Volume
Run #1	35.2 g	1.0 ml
Run #2		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	160	26	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	160	39	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	160	33	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	160	39	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	640	340	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	640	34	ug/kg	
95-48-7	2-Methylphenol	ND	64	34	ug/kg	
	3&4-Methylphenol	ND	64	43	ug/kg	
88-75-5	2-Nitrophenol	ND	160	34	ug/kg	
100-02-7	4-Nitrophenol	ND	320	41	ug/kg	
87-86-5	Pentachlorophenol	ND	320	41	ug/kg	
108-95-2	Phenol	ND	64	24	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	160	34	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	160	42	ug/kg	
83-32-9	Acenaphthene	ND	32	17	ug/kg	
208-96-8	Acenaphthylene	ND	32	14	ug/kg	
98-86-2	Acetophenone	ND	160	16	ug/kg	
120-12-7	Anthracene	ND	32	14	ug/kg	
1912-24-9	Atrazine	ND	160	20	ug/kg	
56-55-3	Benzo(a)anthracene	ND	32	19	ug/kg	
50-32-8	Benzo(a)pyrene	ND	32	13	ug/kg	
205-99-2	Benzo(b)fluoranthene	ND	32	17	ug/kg	
191-24-2	Benzo(g,h,i)perylene	ND	32	15	ug/kg	
207-08-9	Benzo(k)fluoranthene	ND	32	16	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	64	17	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	64	16	ug/kg	
92-52-4	1,1'-Biphenyl	ND	64	16	ug/kg	
100-52-7	Benzaldehyde	ND	160	93	ug/kg	
91-58-7	2-Chloronaphthalene	ND	64	14	ug/kg	
106-47-8	4-Chloroaniline	ND	160	13	ug/kg	
86-74-8	Carbazole	ND	64	14	ug/kg	
105-60-2	Caprolactam	ND	64	26	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-024	Date Sampled:	08/04/09
Lab Sample ID:	JA24692-3	Date Received:	08/04/09
Matrix:	SO - Soil	Percent Solids:	89.0
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	ND	32	15	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	64	16	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	64	15	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	64	17	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	64	21	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	64	17	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	64	14	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	160	55	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	ND	32	15	ug/kg	
132-64-9	Dibenzofuran	ND	64	15	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	64	20	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	64	14	ug/kg	
84-66-2	Diethyl phthalate	ND	64	14	ug/kg	
131-11-3	Dimethyl phthalate	ND	64	15	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	89.7	64	17	ug/kg	
206-44-0	Fluoranthene	ND	32	15	ug/kg	
86-73-7	Fluorene	ND	32	15	ug/kg	
118-74-1	Hexachlorobenzene	ND	64	18	ug/kg	
87-68-3	Hexachlorobutadiene	ND	32	15	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	640	30	ug/kg	
67-72-1	Hexachloroethane	ND	160	21	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	ND	32	12	ug/kg	
78-59-1	Isophorone	ND	64	28	ug/kg	
91-57-6	2-Methylnaphthalene	ND	64	14	ug/kg	
88-74-4	2-Nitroaniline	ND	160	23	ug/kg	
99-09-2	3-Nitroaniline	ND	160	13	ug/kg	
100-01-6	4-Nitroaniline	ND	160	20	ug/kg	
91-20-3	Naphthalene	ND	32	14	ug/kg	
98-95-3	Nitrobenzene	ND	64	14	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	64	20	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	160	22	ug/kg	
85-01-8	Phenanthrene	ND	32	16	ug/kg	
129-00-0	Pyrene	ND	32	14	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	61%		30-109%
4165-62-2	Phenol-d5	64%		28-108%
118-79-6	2,4,6-Tribromophenol	73%		28-125%
4165-60-0	Nitrobenzene-d5	82%		28-113%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-025	Date Sampled:	08/04/09
Lab Sample ID:	JA24692-4	Date Received:	08/04/09
Matrix:	SO - Soil	Percent Solids:	94.7
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	3E21216.D	1	08/05/09	OYA	08/05/09	OP39251	E3E964
Run #2							

Run #	Initial Weight	Final Volume
Run #1	35.0 g	1.0 ml
Run #2		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	150	25	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	150	37	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	150	31	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	150	37	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	600	330	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	600	32	ug/kg	
95-48-7	2-Methylphenol	ND	60	33	ug/kg	
	3&4-Methylphenol	ND	60	40	ug/kg	
88-75-5	2-Nitrophenol	ND	150	32	ug/kg	
100-02-7	4-Nitrophenol	ND	300	39	ug/kg	
87-86-5	Pentachlorophenol	ND	300	39	ug/kg	
108-95-2	Phenol	ND	60	23	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	150	33	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	150	40	ug/kg	
83-32-9	Acenaphthene	ND	30	16	ug/kg	
208-96-8	Acenaphthylene	ND	30	13	ug/kg	
98-86-2	Acetophenone	ND	150	15	ug/kg	
120-12-7	Anthracene	17.2	30	14	ug/kg	J
1912-24-9	Atrazine	ND	150	19	ug/kg	
56-55-3	Benzo(a)anthracene	ND	30	18	ug/kg	
50-32-8	Benzo(a)pyrene	ND	30	12	ug/kg	
205-99-2	Benzo(b)fluoranthene	ND	30	16	ug/kg	
191-24-2	Benzo(g,h,i)perylene	ND	30	15	ug/kg	
207-08-9	Benzo(k)fluoranthene	ND	30	15	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	60	16	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	60	15	ug/kg	
92-52-4	1,1'-Biphenyl	ND	60	15	ug/kg	
100-52-7	Benzaldehyde	ND	150	88	ug/kg	
91-58-7	2-Chloronaphthalene	ND	60	13	ug/kg	
106-47-8	4-Chloroaniline	ND	150	12	ug/kg	
86-74-8	Carbazole	ND	60	13	ug/kg	
105-60-2	Caprolactam	ND	60	24	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-025	Date Sampled:	08/04/09
Lab Sample ID:	JA24692-4	Date Received:	08/04/09
Matrix:	SO - Soil	Percent Solids:	94.7
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	15.3	30	14	ug/kg	J
111-91-1	bis(2-Chloroethoxy)methane	ND	60	15	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	60	14	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	60	16	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	60	20	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	60	16	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	60	14	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	150	52	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	ND	30	14	ug/kg	
132-64-9	Dibenzofuran	ND	60	14	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	60	19	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	60	13	ug/kg	
84-66-2	Diethyl phthalate	ND	60	14	ug/kg	
131-11-3	Dimethyl phthalate	ND	60	14	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	ND	60	16	ug/kg	
206-44-0	Fluoranthene	26.7	30	14	ug/kg	J
86-73-7	Fluorene	ND	30	14	ug/kg	
118-74-1	Hexachlorobenzene	ND	60	17	ug/kg	
87-68-3	Hexachlorobutadiene	ND	30	14	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	600	28	ug/kg	
67-72-1	Hexachloroethane	ND	150	19	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	ND	30	11	ug/kg	
78-59-1	Isophorone	ND	60	26	ug/kg	
91-57-6	2-Methylnaphthalene	167	60	14	ug/kg	
88-74-4	2-Nitroaniline	ND	150	22	ug/kg	
99-09-2	3-Nitroaniline	ND	150	12	ug/kg	
100-01-6	4-Nitroaniline	ND	150	19	ug/kg	
91-20-3	Naphthalene	70.5	30	13	ug/kg	
98-95-3	Nitrobenzene	ND	60	14	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	60	19	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	150	20	ug/kg	
85-01-8	Phenanthrene	47.8	30	15	ug/kg	
129-00-0	Pyrene	30.3	30	13	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	43%		30-109%
4165-62-2	Phenol-d5	44%		28-108%
118-79-6	2,4,6-Tribromophenol	48%		28-125%
4165-60-0	Nitrobenzene-d5	61%		28-113%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 3

Client Sample ID:	B-CCI-026 (2')	Date Sampled:	08/06/09
Lab Sample ID:	JA25087-1	Date Received:	08/07/09
Matrix:	SO - Soil	Percent Solids:	83.0
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	3M13354.D	1	08/10/09	KLS	08/07/09	OP39296	E3M596

Run #1	Initial Weight	Final Volume
Run #2	35.0 g	1.0 ml

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	170	28	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	170	42	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	170	36	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	170	42	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	690	370	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	690	36	ug/kg	
95-48-7	2-Methylphenol	ND	69	37	ug/kg	
	3&4-Methylphenol	ND	69	46	ug/kg	
88-75-5	2-Nitrophenol	ND	170	36	ug/kg	
100-02-7	4-Nitrophenol	ND	340	44	ug/kg	
87-86-5	Pentachlorophenol	ND	340	44	ug/kg	
108-95-2	Phenol	ND	69	26	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	170	37	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	170	46	ug/kg	
83-32-9	Acenaphthene	19.4	34	18	ug/kg	J
208-96-8	Acenaphthylene	316	34	15	ug/kg	
98-86-2	Acetophenone	ND	170	17	ug/kg	
120-12-7	Anthracene	267	34	15	ug/kg	
1912-24-9	Atrazine	ND	170	22	ug/kg	
56-55-3	Benzo(a)anthracene	808	34	21	ug/kg	
50-32-8	Benzo(a)pyrene	864	34	14	ug/kg	
205-99-2	Benzo(b)fluoranthene	1040	34	19	ug/kg	
191-24-2	Benzo(g,h,i)perylene	926	34	17	ug/kg	
207-08-9	Benzo(k)fluoranthene	333	34	17	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	69	19	ug/kg	
85-68-7	Butyl benzyl phthalate	51.6	69	17	ug/kg	J
92-52-4	1,1'-Biphenyl	ND	69	17	ug/kg	
100-52-7	Benzaldehyde	ND	170	100	ug/kg	
91-58-7	2-Chloronaphthalene	ND	69	15	ug/kg	
106-47-8	4-Chloroaniline	ND	170	14	ug/kg	
86-74-8	Carbazole	127	69	15	ug/kg	
105-60-2	Caprolactam	ND	69	28	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

3.1
3

Client Sample ID:	B-CCI-026 (2')	Date Sampled:	08/06/09
Lab Sample ID:	JA25087-1	Date Received:	08/07/09
Matrix:	SO - Soil	Percent Solids:	83.0
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	787	34	16	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	69	17	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	69	16	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	69	18	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	69	23	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	69	19	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	69	16	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	170	60	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	200	34	16	ug/kg	
132-64-9	Dibenzofuran	23.0	69	16	ug/kg	J
84-74-2	Di-n-butyl phthalate	ND	69	21	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	69	15	ug/kg	
84-66-2	Diethyl phthalate	ND	69	16	ug/kg	
131-11-3	Dimethyl phthalate	ND	69	16	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	59.3	69	18	ug/kg	J
206-44-0	Fluoranthene	1390	34	16	ug/kg	
86-73-7	Fluorene	39.8	34	16	ug/kg	
118-74-1	Hexachlorobenzene	ND	69	20	ug/kg	
87-68-3	Hexachlorobutadiene	ND	34	16	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	690	32	ug/kg	
67-72-1	Hexachloroethane	ND	170	22	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	704	34	13	ug/kg	
78-59-1	Isophorone	ND	69	30	ug/kg	
91-57-6	2-Methylnaphthalene	ND	69	15	ug/kg	
88-74-4	2-Nitroaniline	ND	170	25	ug/kg	
99-09-2	3-Nitroaniline	ND	170	14	ug/kg	
100-01-6	4-Nitroaniline	ND	170	22	ug/kg	
91-20-3	Naphthalene	ND	34	15	ug/kg	
98-95-3	Nitrobenzene	ND	69	16	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	69	21	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	170	23	ug/kg	
85-01-8	Phenanthrene	679	34	17	ug/kg	
129-00-0	Pyrene	1240	34	15	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	66%		30-109%
4165-62-2	Phenol-d5	64%		28-108%
118-79-6	2,4,6-Tribromophenol	74%		28-125%
4165-60-0	Nitrobenzene-d5	64%		28-113%

ND = Not detected MDL - Method Detection Limit J = Indicates an estimated value
 RL = Reporting Limit B = Indicates analyte found in associated method blank
 E = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 3

Client Sample ID:	B-CCI-027 (2')	Date Sampled:	08/06/09
Lab Sample ID:	JA25087-2	Date Received:	08/07/09
Matrix:	SO - Soil	Percent Solids:	88.1
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	3M13355.D	1	08/10/09	KLS	08/07/09	OP39296	E3M596

Run #1	Initial Weight	Final Volume
Run #2	35.0 g	1.0 ml

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	160	27	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	160	39	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	160	34	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	160	40	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	650	350	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	650	34	ug/kg	
95-48-7	2-Methylphenol	ND	65	35	ug/kg	
	3&4-Methylphenol	ND	65	43	ug/kg	
88-75-5	2-Nitrophenol	ND	160	34	ug/kg	
100-02-7	4-Nitrophenol	ND	320	42	ug/kg	
87-86-5	Pentachlorophenol	ND	320	42	ug/kg	
108-95-2	Phenol	ND	65	25	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	160	35	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	160	43	ug/kg	
83-32-9	Acenaphthene	17.0	32	17	ug/kg	J
208-96-8	Acenaphthylene	153	32	14	ug/kg	
98-86-2	Acetophenone	ND	160	16	ug/kg	
120-12-7	Anthracene	254	32	15	ug/kg	
1912-24-9	Atrazine	ND	160	20	ug/kg	
56-55-3	Benzo(a)anthracene	713	32	20	ug/kg	
50-32-8	Benzo(a)pyrene	497	32	13	ug/kg	
205-99-2	Benzo(b)fluoranthene	677	32	18	ug/kg	
191-24-2	Benzo(g,h,i)perylene	284	32	16	ug/kg	
207-08-9	Benzo(k)fluoranthene	209	32	16	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	65	18	ug/kg	
85-68-7	Butyl benzyl phthalate	111	65	16	ug/kg	
92-52-4	1,1'-Biphenyl	ND	65	16	ug/kg	
100-52-7	Benzaldehyde	ND	160	95	ug/kg	
91-58-7	2-Chloronaphthalene	ND	65	14	ug/kg	
106-47-8	4-Chloroaniline	ND	160	13	ug/kg	
86-74-8	Carbazole	151	65	14	ug/kg	
105-60-2	Caprolactam	ND	65	26	ug/kg	

ND = Not detected MDL - Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-027 (2')	Date Sampled:	08/06/09
Lab Sample ID:	JA25087-2	Date Received:	08/07/09
Matrix:	SO - Soil	Percent Solids:	88.1
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	671	32	15	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	65	16	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	65	15	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	65	17	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	65	22	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	65	18	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	65	15	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	160	56	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	109	32	15	ug/kg	
132-64-9	Dibenzofuran	89.9	65	15	ug/kg	
84-74-2	Di-n-butyl phthalate	66.3	65	20	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	65	14	ug/kg	
84-66-2	Diethyl phthalate	ND	65	15	ug/kg	
131-11-3	Dimethyl phthalate	ND	65	15	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	417	65	17	ug/kg	
206-44-0	Fluoranthene	1580	32	15	ug/kg	
86-73-7	Fluorene	64.7	32	15	ug/kg	
118-74-1	Hexachlorobenzene	ND	65	18	ug/kg	
87-68-3	Hexachlorobutadiene	ND	32	16	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	650	31	ug/kg	
67-72-1	Hexachloroethane	ND	160	21	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	291	32	12	ug/kg	
78-59-1	Isophorone	ND	65	28	ug/kg	
91-57-6	2-Methylnaphthalene	ND	65	15	ug/kg	
88-74-4	2-Nitroaniline	ND	160	24	ug/kg	
99-09-2	3-Nitroaniline	ND	160	13	ug/kg	
100-01-6	4-Nitroaniline	ND	160	20	ug/kg	
91-20-3	Naphthalene	ND	32	14	ug/kg	
98-95-3	Nitrobenzene	ND	65	15	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	65	20	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	160	22	ug/kg	
85-01-8	Phenanthrene	1610	32	16	ug/kg	
129-00-0	Pyrene	1240	32	14	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	64%		30-109%
4165-62-2	Phenol-d5	60%		28-108%
118-79-6	2,4,6-Tribromophenol	70%		28-125%
4165-60-0	Nitrobenzene-d5	63%		28-113%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 3

Client Sample ID:	B-CCI-028(2')	Date Sampled:	08/11/09
Lab Sample ID:	JA25446-1	Date Received:	08/12/09
Matrix:	SO - Soil	Percent Solids:	79.6
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	2M23516.D	1	08/13/09	OYA	08/13/09	OP39413	E2M991
Run #2							

Run #	Initial Weight	Final Volume
Run #1	35.3 g	1.0 ml
Run #2		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	180	29	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	180	43	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	180	37	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	180	43	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	710	380	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	710	38	ug/kg	
95-48-7	2-Methylphenol	ND	71	38	ug/kg	
	3&4-Methylphenol	ND	71	48	ug/kg	
88-75-5	2-Nitrophenol	ND	180	37	ug/kg	
100-02-7	4-Nitrophenol	ND	360	46	ug/kg	
87-86-5	Pentachlorophenol	ND	360	46	ug/kg	
108-95-2	Phenol	ND	71	27	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	180	38	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	180	47	ug/kg	
83-32-9	Acenaphthene	ND	36	19	ug/kg	
208-96-8	Acenaphthylene	133	36	15	ug/kg	
98-86-2	Acetophenone	ND	180	17	ug/kg	
120-12-7	Anthracene	130	36	16	ug/kg	
1912-24-9	Atrazine	ND	180	22	ug/kg	
56-55-3	Benzo(a)anthracene	714	36	22	ug/kg	
50-32-8	Benzo(a)pyrene	629	36	15	ug/kg	
205-99-2	Benzo(b)fluoranthene	558	36	19	ug/kg	
191-24-2	Benzo(g,h,i)perylene	413	36	17	ug/kg	
207-08-9	Benzo(k)fluoranthene	523	36	18	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	71	19	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	71	18	ug/kg	
92-52-4	1,1'-Biphenyl	ND	71	18	ug/kg	
100-52-7	Benzaldehyde	ND	180	100	ug/kg	
91-58-7	2-Chloronaphthalene	ND	71	16	ug/kg	
106-47-8	4-Chloroaniline	ND	180	14	ug/kg	
86-74-8	Carbazole	33.3	71	15	ug/kg	J
105-60-2	Caprolactam	ND	71	29	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-028(2')	Date Sampled:	08/11/09
Lab Sample ID:	JA25446-1	Date Received:	08/12/09
Matrix:	SO - Soil	Percent Solids:	79.6
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	658	36	16	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	71	18	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	71	17	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	71	19	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	71	24	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	71	19	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	71	16	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	180	62	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	144	36	17	ug/kg	
132-64-9	Dibenzofuran	ND	71	17	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	71	22	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	71	16	ug/kg	
84-66-2	Diethyl phthalate	ND	71	16	ug/kg	
131-11-3	Dimethyl phthalate	ND	71	17	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	ND	71	19	ug/kg	
206-44-0	Fluoranthene	1310	36	16	ug/kg	
86-73-7	Fluorene	20.0	36	17	ug/kg	J
118-74-1	Hexachlorobenzene	ND	71	20	ug/kg	
87-68-3	Hexachlorobutadiene	ND	36	17	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	710	34	ug/kg	
67-72-1	Hexachloroethane	ND	180	23	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	393	36	13	ug/kg	
78-59-1	Isophorone	ND	71	31	ug/kg	
91-57-6	2-Methylnaphthalene	ND	71	16	ug/kg	
88-74-4	2-Nitroaniline	ND	180	26	ug/kg	
99-09-2	3-Nitroaniline	ND	180	14	ug/kg	
100-01-6	4-Nitroaniline	ND	180	22	ug/kg	
91-20-3	Naphthalene	ND	36	15	ug/kg	
98-95-3	Nitrobenzene	ND	71	16	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	71	22	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	180	24	ug/kg	
85-01-8	Phenanthrene	405	36	18	ug/kg	
129-00-0	Pyrene	1160	36	16	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	70%		30-109%
4165-62-2	Phenol-d5	72%		28-108%
118-79-6	2,4,6-Tribromophenol	87%		28-125%
4165-60-0	Nitrobenzene-d5	73%		28-113%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 3

Client Sample ID:	B-CCI-029(20')	
Lab Sample ID:	JA25446-2	Date Sampled: 08/12/09
Matrix:	SO - Soil	Date Received: 08/12/09
Method:	SW846 8270C SW846 3550B	Percent Solids: 79.4
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY	

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	2M23513.D	1	08/13/09	OYA	08/13/09	OP39413	E2M991
Run #2							

Run #1	Initial Weight	Final Volume
Run #1	35.0 g	1.0 ml
Run #2		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	180	30	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	180	44	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	180	37	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	180	44	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	720	390	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	720	38	ug/kg	
95-48-7	2-Methylphenol	ND	72	39	ug/kg	
	3&4-Methylphenol	ND	72	48	ug/kg	
88-75-5	2-Nitrophenol	ND	180	38	ug/kg	
100-02-7	4-Nitrophenol	ND	360	46	ug/kg	
87-86-5	Pentachlorophenol	ND	360	46	ug/kg	
108-95-2	Phenol	ND	72	27	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	180	39	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	180	48	ug/kg	
83-32-9	Acenaphthene	ND	36	19	ug/kg	
208-96-8	Acenaphthylene	ND	36	15	ug/kg	
98-86-2	Acetophenone	ND	180	18	ug/kg	
120-12-7	Anthracene	ND	36	16	ug/kg	
1912-24-9	Atrazine	ND	180	23	ug/kg	
56-55-3	Benzo(a)anthracene	ND	36	22	ug/kg	
50-32-8	Benzo(a)pyrene	ND	36	15	ug/kg	
205-99-2	Benzo(b)fluoranthene	ND	36	20	ug/kg	
191-24-2	Benzo(g,h,i)perylene	ND	36	17	ug/kg	
207-08-9	Benzo(k)fluoranthene	ND	36	18	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	72	20	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	72	18	ug/kg	
92-52-4	1,1'-Biphenyl	ND	72	18	ug/kg	
100-52-7	Benzaldehyde	ND	180	110	ug/kg	
91-58-7	2-Chloronaphthalene	ND	72	16	ug/kg	
106-47-8	4-Chloroaniline	ND	180	15	ug/kg	
86-74-8	Carbazole	ND	72	15	ug/kg	
105-60-2	Caprolactam	ND	72	29	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-029(20')	Date Sampled:	08/12/09
Lab Sample ID:	JA25446-2	Date Received:	08/12/09
Matrix:	SO - Soil	Percent Solids:	79.4
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	ND	36	17	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	72	18	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	72	17	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	72	19	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	72	24	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	72	20	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	72	16	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	180	62	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	ND	36	17	ug/kg	
132-64-9	Dibenzofuran	ND	72	17	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	72	22	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	72	16	ug/kg	
84-66-2	Diethyl phthalate	ND	72	16	ug/kg	
131-11-3	Dimethyl phthalate	ND	72	17	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	ND	72	19	ug/kg	
206-44-0	Fluoranthene	ND	36	16	ug/kg	
86-73-7	Fluorene	ND	36	17	ug/kg	
118-74-1	Hexachlorobenzene	ND	72	20	ug/kg	
87-68-3	Hexachlorobutadiene	ND	36	17	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	720	34	ug/kg	
67-72-1	Hexachloroethane	ND	180	23	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	ND	36	13	ug/kg	
78-59-1	Isophorone	ND	72	32	ug/kg	
91-57-6	2-Methylnaphthalene	28.2	72	16	ug/kg	J
88-74-4	2-Nitroaniline	ND	180	26	ug/kg	
99-09-2	3-Nitroaniline	ND	180	15	ug/kg	
100-01-6	4-Nitroaniline	ND	180	22	ug/kg	
91-20-3	Naphthalene	ND	36	16	ug/kg	
98-95-3	Nitrobenzene	ND	72	16	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	72	22	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	180	24	ug/kg	
85-01-8	Phenanthrene	ND	36	18	ug/kg	
129-00-0	Pyrene	ND	36	16	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	80%		30-109%
4165-62-2	Phenol-d5	80%		28-108%
118-79-6	2,4,6-Tribromophenol	92%		28-125%
4165-60-0	Nitrobenzene-d5	81%		28-113%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 3

Client Sample ID:	B-CCI-030(4')	Date Sampled:	08/12/09
Lab Sample ID:	JA25446-3	Date Received:	08/12/09
Matrix:	SO - Soil	Percent Solids:	89.1
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	2M23515.D	1	08/13/09	OYA	08/13/09	OP39413	E2M991
Run #2							

Run #1	Initial Weight	Final Volume
Run #1	35.1 g	1.0 ml
Run #2		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	160	26	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	160	39	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	160	33	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	160	39	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	640	350	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	640	34	ug/kg	
95-48-7	2-Methylphenol	ND	64	35	ug/kg	
	3&4-Methylphenol	ND	64	43	ug/kg	
88-75-5	2-Nitrophenol	ND	160	34	ug/kg	
100-02-7	4-Nitrophenol	ND	320	41	ug/kg	
87-86-5	Pentachlorophenol	ND	320	41	ug/kg	
108-95-2	Phenol	ND	64	24	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	160	35	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	160	43	ug/kg	
83-32-9	Acenaphthene	ND	32	17	ug/kg	
208-96-8	Acenaphthylene	40.6	32	14	ug/kg	
98-86-2	Acetophenone	ND	160	16	ug/kg	
120-12-7	Anthracene	55.1	32	14	ug/kg	
1912-24-9	Atrazine	ND	160	20	ug/kg	
56-55-3	Benzo(a)anthracene	223	32	19	ug/kg	
50-32-8	Benzo(a)pyrene	235	32	13	ug/kg	
205-99-2	Benzo(b)fluoranthene	190	32	17	ug/kg	
191-24-2	Benzo(g,h,i)perylene	194	32	16	ug/kg	
207-08-9	Benzo(k)fluoranthene	216	32	16	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	64	17	ug/kg	
85-68-7	Butyl benzyl phthalate	66.4	64	16	ug/kg	
92-52-4	1,1'-Biphenyl	ND	64	16	ug/kg	
100-52-7	Benzaldehyde	ND	160	93	ug/kg	
91-58-7	2-Chloronaphthalene	ND	64	14	ug/kg	
106-47-8	4-Chloroaniline	ND	160	13	ug/kg	
86-74-8	Carbazole	31.3	64	14	ug/kg	J
105-60-2	Caprolactam	ND	64	26	ug/kg	

ND = Not detected MDL - Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-030(4')	Date Sampled:	08/12/09
Lab Sample ID:	JA25446-3	Date Received:	08/12/09
Matrix:	SO - Soil	Percent Solids:	89.1
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	243	32	15	ug/kg	
111-91-1	bis(2-Chlorooctoxy)methane	ND	64	16	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	64	15	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	64	17	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	64	21	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	64	17	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	64	14	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	160	55	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	52.0	32	15	ug/kg	
132-64-9	Dibenzofuran	ND	64	15	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	64	20	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	64	14	ug/kg	
84-66-2	Diethyl phthalate	ND	64	14	ug/kg	
131-11-3	Dimethyl phthalate	ND	64	15	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	413	64	17	ug/kg	
206-44-0	Fluoranthene	446	32	15	ug/kg	
86-73-7	Fluorene	23.0	32	15	ug/kg	J
118-74-1	Hexachlorobenzene	ND	64	18	ug/kg	
87-68-3	Hexachlorobutadiene	ND	32	15	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	640	30	ug/kg	
67-72-1	Hexachloroethane	ND	160	21	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	159	32	12	ug/kg	
78-59-1	Isophorone	ND	64	28	ug/kg	
91-57-6	2-Methylnaphthalene	27.4	64	14	ug/kg	J
88-74-4	2-Nitroaniline	ND	160	24	ug/kg	
99-09-2	3-Nitroaniline	ND	160	13	ug/kg	
100-01-6	4-Nitroaniline	ND	160	20	ug/kg	
91-20-3	Naphthalene	20.6	32	14	ug/kg	J
98-95-3	Nitrobenzene	ND	64	14	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	64	20	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	160	22	ug/kg	
85-01-8	Phenanthrene	243	32	16	ug/kg	
129-00-0	Pyrene	422	32	14	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	73%		30-109%
4165-62-2	Phenol-d5	74%		28-108%
118-79-6	2,4,6-Tribromophenol	86%		28-125%
4165-60-0	Nitrobenzene-d5	78%		28-113%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 3

3.4
3

Client Sample ID:	B-CCI-031(10')	Date Sampled:	08/12/09
Lab Sample ID:	JA25446-4	Date Received:	08/12/09
Matrix:	SO - Soil	Percent Solids:	89.4
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	2M23514.D	1	08/13/09	OYA	08/13/09	OP39413	E2M991
Run #2							

Run #	Initial Weight	Final Volume
Run #1	35.0 g	1.0 ml
Run #2		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	160	26	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	160	39	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	160	33	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	160	39	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	640	350	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	640	34	ug/kg	
95-48-7	2-Methylphenol	ND	64	35	ug/kg	
	3&4-Methylphenol	ND	64	43	ug/kg	
88-75-5	2-Nitrophenol	ND	160	34	ug/kg	
100-02-7	4-Nitrophenol	ND	320	41	ug/kg	
87-86-5	Pentachlorophenol	ND	320	41	ug/kg	
108-95-2	Phenol	ND	64	24	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	160	35	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	160	43	ug/kg	
83-32-9	Acenaphthene	ND	32	17	ug/kg	
208-96-8	Acenaphthylene	ND	32	14	ug/kg	
98-86-2	Acetophenone	ND	160	16	ug/kg	
120-12-7	Anthracene	ND	32	14	ug/kg	
1912-24-9	Atrazine	ND	160	20	ug/kg	
56-55-3	Benzo(a)anthracene	ND	32	19	ug/kg	
50-32-8	Benzo(a)pyrene	ND	32	13	ug/kg	
205-99-2	Benzo(b)fluoranthene	ND	32	17	ug/kg	
191-24-2	Benzo(g,h,i)perylene	ND	32	16	ug/kg	
207-08-9	Benzo(k)fluoranthene	ND	32	16	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	64	17	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	64	16	ug/kg	
92-52-4	1,1'-Biphenyl	ND	64	16	ug/kg	
100-52-7	Benzaldehyde	ND	160	93	ug/kg	
91-58-7	2-Chloronaphthalene	ND	64	14	ug/kg	
106-47-8	4-Chloroaniline	ND	160	13	ug/kg	
86-74-8	Carbazole	ND	64	14	ug/kg	
105-60-2	Caprolactam	ND	64	26	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-031(10')	Date Sampled:	08/12/09
Lab Sample ID:	JA25446-4	Date Received:	08/12/09
Matrix:	SO - Soil	Percent Solids:	89.4
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	ND	32	15	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	64	16	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	64	15	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	64	17	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	64	21	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	64	17	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	64	14	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	160	55	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	ND	32	15	ug/kg	
132-64-9	Dibenzofuran	ND	64	15	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	64	20	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	64	14	ug/kg	
84-66-2	Diethyl phthalate	ND	64	14	ug/kg	
131-11-3	Dimethyl phthalate	ND	64	15	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	ND	64	17	ug/kg	
206-44-0	Fluoranthene	ND	32	15	ug/kg	
86-73-7	Fluorene	ND	32	15	ug/kg	
118-74-1	Hexachlorobenzene	ND	64	18	ug/kg	
87-68-3	Hexachlorobutadiene	ND	32	15	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	640	30	ug/kg	
67-72-1	Hexachloroethane	ND	160	21	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	ND	32	12	ug/kg	
78-59-1	Isophorone	ND	64	28	ug/kg	
91-57-6	2-Methylnaphthalene	ND	64	14	ug/kg	
88-74-4	2-Nitroaniline	ND	160	23	ug/kg	
99-09-2	3-Nitroaniline	ND	160	13	ug/kg	
100-01-6	4-Nitroaniline	ND	160	20	ug/kg	
91-20-3	Naphthalene	ND	32	14	ug/kg	
98-95-3	Nitrobenzene	ND	64	14	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	64	20	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	160	22	ug/kg	
85-01-8	Phenanthrene	ND	32	16	ug/kg	
129-00-0	Pyrene	ND	32	14	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	73%		30-109%
4165-62-2	Phenol-d5	73%		28-108%
118-79-6	2,4,6-Tribromophenol	78%		28-125%
4165-60-0	Nitrobenzene-d5	76%		28-113%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest LabLink@514900 10:37 13-Aug-2009

Report of Analysis

Page 1 of 3

Client Sample ID:	SW-CCI-001(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-2	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	81.1
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	3E21094.D	1	07/30/09	OYA	07/29/09	OP39117	E3E960
Run #2							

Run #	Initial Weight	Final Volume
Run #1	35.6 g	1.0 ml
Run #2		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	170	29	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	170	42	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	170	36	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	170	42	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	690	370	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	690	37	ug/kg	
95-48-7	2-Methylphenol	ND	69	37	ug/kg	
	3&4-Methylphenol	ND	69	46	ug/kg	
88-75-5	2-Nitrophenol	ND	170	36	ug/kg	
100-02-7	4-Nitrophenol	ND	350	44	ug/kg	
87-86-5	Pentachlorophenol	ND	350	45	ug/kg	
108-95-2	Phenol	ND	69	26	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	170	37	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	170	46	ug/kg	
83-32-9	Acenaphthene	ND	35	18	ug/kg	
208-96-8	Acenaphthylene	ND	35	15	ug/kg	
98-86-2	Acetophenone	ND	170	17	ug/kg	
120-12-7	Anthracene	ND	35	16	ug/kg	
1912-24-9	Atrazine	ND	170	22	ug/kg	
56-55-3	Benzo(a)anthracene	ND	35	21	ug/kg	
50-32-8	Benzo(a)pyrene	ND	35	14	ug/kg	
205-99-2	Benzo(b)fluoranthene	ND	35	19	ug/kg	
191-24-2	Benzo(g,h,i)perylene	ND	35	17	ug/kg	
207-08-9	Benzo(k)fluoranthene	ND	35	17	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	69	19	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	69	17	ug/kg	
92-52-4	1,1'-Biphenyl	ND	69	17	ug/kg	
100-52-7	Benzaldehyde	ND	170	100	ug/kg	
91-58-7	2-Chloronaphthalene	ND	69	15	ug/kg	
106-47-8	4-Chloroaniline	ND	170	14	ug/kg	
86-74-8	Carbazole	ND	69	15	ug/kg	
105-60-2	Caprolactam	ND	69	28	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-001(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-2	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	81.1
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	ND	35	16	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	69	17	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	69	16	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	69	18	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	69	23	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	69	19	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	69	16	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	170	60	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	ND	35	17	ug/kg	
132-64-9	Dibenzofuran	ND	69	16	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	69	22	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	69	15	ug/kg	
84-66-2	Diethyl phthalate	ND	69	16	ug/kg	
131-11-3	Dimethyl phthalate	ND	69	17	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	ND	69	18	ug/kg	
206-44-0	Fluoranthene	ND	35	16	ug/kg	
86-73-7	Fluorene	ND	35	16	ug/kg	
118-74-1	Hexachlorobenzene	ND	69	20	ug/kg	
87-68-3	Hexachlorobutadiene	ND	35	17	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	690	33	ug/kg	
67-72-1	Hexachloroethane	ND	170	22	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	ND	35	13	ug/kg	
78-59-1	Isophorone	ND	69	30	ug/kg	
91-57-6	2-Methylnaphthalene	ND	69	16	ug/kg	
88-74-4	2-Nitroaniline	ND	170	25	ug/kg	
99-09-2	3-Nitroaniline	ND	170	14	ug/kg	
100-01-6	4-Nitroaniline	ND	170	22	ug/kg	
91-20-3	Naphthalene	ND	35	15	ug/kg	
98-95-3	Nitrobenzene	ND	69	16	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	69	21	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	170	23	ug/kg	
85-01-8	Phenanthrene	ND	35	17	ug/kg	
129-00-0	Pyrene	ND	35	15	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	69%		30-109%
4165-62-2	Phenol-d5	75%		28-108%
118-79-6	2,4,6-Tribromophenol	77%		28-125%
4165-60-0	Nitrobenzene-d5	95%		28-113%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest LabLink@514900 10:37 13-Aug-2009

Report of Analysis

Page 1 of 3

Client Sample ID:	SW-CCI-002(18.5')		Date Sampled:	07/20/09
Lab Sample ID:	JA23638-3		Date Received:	07/21/09
Matrix:	SO - Soil		Percent Solids:	86.1
Method:	SW846 8270C SW846 3550B			
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY			

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	3E21102.D	1	07/30/09	OYA	07/29/09	OP39117	E3E960
Run #2							

Run #	Initial Weight	Final Volume
Run #1	35.2 g	1.0 ml
Run #2		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	160	27	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	160	40	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	160	34	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	160	40	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	660	360	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	660	35	ug/kg	
95-48-7	2-Methylphenol	ND	66	36	ug/kg	
	3&4-Methylphenol	ND	66	44	ug/kg	
88-75-5	2-Nitrophenol	ND	160	35	ug/kg	
100-02-7	4-Nitrophenol	ND	330	42	ug/kg	
87-86-5	Pentachlorophenol	ND	330	43	ug/kg	
108-95-2	Phenol	ND	66	25	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	160	36	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	160	44	ug/kg	
83-32-9	Acenaphthene	21.0	33	17	ug/kg	J
208-96-8	Acenaphthylene	ND	33	14	ug/kg	
98-86-2	Acetophenone	ND	160	16	ug/kg	
120-12-7	Anthracene	24.9	33	15	ug/kg	J
1912-24-9	Atrazine	ND	160	21	ug/kg	
56-55-3	Benzo(a)anthracene	ND	33	20	ug/kg	
50-32-8	Benzo(a)pyrene	ND	33	14	ug/kg	
205-99-2	Benzo(b)fluoranthene	ND	33	18	ug/kg	
191-24-2	Benzo(g,h,i)perylene	ND	33	16	ug/kg	
207-08-9	Benzo(k)fluoranthene	ND	33	17	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	66	18	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	66	16	ug/kg	
92-52-4	1,1'-Biphenyl	19.3	66	17	ug/kg	J
100-52-7	Benzaldehyde	ND	160	96	ug/kg	
91-58-7	2-Chloronaphthalene	ND	66	15	ug/kg	
106-47-8	4-Chloroaniline	ND	160	13	ug/kg	
86-74-8	Carbazole	ND	66	14	ug/kg	
105-60-2	Caprolactam	ND	66	26	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-002(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-3	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	86.1
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	ND	33	15	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	66	16	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	66	16	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	66	17	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	66	22	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	66	18	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	66	15	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	160	57	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	ND	33	16	ug/kg	
132-64-9	Dibenzofuran	ND	66	15	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	66	21	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	66	15	ug/kg	
84-66-2	Diethyl phthalate	ND	66	15	ug/kg	
131-11-3	Dimethyl phthalate	ND	66	16	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	ND	66	17	ug/kg	
206-44-0	Fluoranthene	17.3	33	15	ug/kg	J
86-73-7	Fluorene	34.3	33	16	ug/kg	
118-74-1	Hexachlorobenzene	ND	66	19	ug/kg	
87-68-3	Hexachlorobutadiene	ND	33	16	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	660	31	ug/kg	
67-72-1	Hexachloroethane	ND	160	21	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	ND	33	12	ug/kg	
78-59-1	Isophorone	ND	66	29	ug/kg	
91-57-6	2-Methylnaphthalene	985	66	15	ug/kg	
88-74-4	2-Nitroaniline	ND	160	24	ug/kg	
99-09-2	3-Nitroaniline	ND	160	13	ug/kg	
100-01-6	4-Nitroaniline	ND	160	21	ug/kg	
91-20-3	Naphthalene	173	33	14	ug/kg	
98-95-3	Nitrobenzene	ND	66	15	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	66	20	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	160	22	ug/kg	
85-01-8	Phenanthrene	67.6	33	16	ug/kg	
129-00-0	Pyrene	32.4	33	15	ug/kg	J

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	70%		30-109%
4165-62-2	Phenol-d5	76%		28-108%
118-79-6	2,4,6-Tribromophenol	76%		28-125%
4165-60-0	Nitrobenzene-d5	99%		28-113%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest LabLink@514900 10:37 13-Aug-2009

Report of Analysis

Page 1 of 3

Client Sample ID:	SW-CCI-003(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-4	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	88.4
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	3E21095.D	1	07/30/09	OYA	07/29/09	OP39117	E3E960
Run #2							

Run #	Initial Weight	Final Volume
Run #1	35.2 g	1.0 ml
Run #2		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	160	27	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	160	39	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	160	33	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	160	39	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	640	350	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	640	34	ug/kg	
95-48-7	2-Methylphenol	ND	64	35	ug/kg	
	3&4-Methylphenol	ND	64	43	ug/kg	
88-75-5	2-Nitrophenol	ND	160	34	ug/kg	
100-02-7	4-Nitrophenol	ND	320	41	ug/kg	
87-86-5	Pentachlorophenol	ND	320	41	ug/kg	
108-95-2	Phenol	ND	64	24	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	160	35	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	160	43	ug/kg	
83-32-9	Acenaphthene	ND	32	17	ug/kg	
208-96-8	Acenaphthylene	ND	32	14	ug/kg	
98-86-2	Acetophenone	ND	160	16	ug/kg	
120-12-7	Anthracene	ND	32	14	ug/kg	
1912-24-9	Atrazine	ND	160	20	ug/kg	
56-55-3	Benzo(a)anthracene	ND	32	19	ug/kg	
50-32-8	Benzo(a)pyrene	ND	32	13	ug/kg	
205-99-2	Benzo(b)fluoranthene	ND	32	17	ug/kg	
191-24-2	Benzo(g,h,i)perylene	ND	32	16	ug/kg	
207-08-9	Benzo(k)fluoranthene	ND	32	16	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	64	18	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	64	16	ug/kg	
92-52-4	1,1'-Biphenyl	ND	64	16	ug/kg	
100-52-7	Benzaldehyde	ND	160	94	ug/kg	
91-58-7	2-Chloronaphthalene	ND	64	14	ug/kg	
106-47-8	4-Chloroaniline	ND	160	13	ug/kg	
86-74-8	Carbazole	ND	64	14	ug/kg	
105-60-2	Caprolactam	ND	64	26	ug/kg	

ND = Not detected MDL - Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-003(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-4	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	88.4
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	ND	32	15	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	64	16	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	64	15	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	64	17	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	64	21	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	64	18	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	64	14	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	160	56	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	ND	32	15	ug/kg	
132-64-9	Dibenzofuran	ND	64	15	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	64	20	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	64	14	ug/kg	
84-66-2	Diethyl phthalate	ND	64	14	ug/kg	
131-11-3	Dimethyl phthalate	ND	64	15	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	ND	64	17	ug/kg	
206-44-0	Fluoranthene	ND	32	15	ug/kg	
86-73-7	Fluorene	ND	32	15	ug/kg	
118-74-1	Hexachlorobenzene	ND	64	18	ug/kg	
87-68-3	Hexachlorobutadiene	ND	32	15	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	640	30	ug/kg	
67-72-1	Hexachloroethane	ND	160	21	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	ND	32	12	ug/kg	
78-59-1	Isophorone	ND	64	28	ug/kg	
91-57-6	2-Methylnaphthalene	ND	64	14	ug/kg	
88-74-4	2-Nitroaniline	ND	160	24	ug/kg	
99-09-2	3-Nitroaniline	ND	160	13	ug/kg	
100-01-6	4-Nitroaniline	ND	160	20	ug/kg	
91-20-3	Naphthalene	ND	32	14	ug/kg	
98-95-3	Nitrobenzene	ND	64	14	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	64	20	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	160	22	ug/kg	
85-01-8	Phenanthrene	ND	32	16	ug/kg	
129-00-0	Pyrene	18.8	32	14	ug/kg	J

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	68%		30-109%
4165-62-2	Phenol-d5	72%		28-108%
118-79-6	2,4,6-Tribromophenol	76%		28-125%
4165-60-0	Nitrobenzene-d5	94%		28-113%

ND = Not detected MDL - Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

Accutest LabLink@514900 10:37 13-Aug-2009

Report of Analysis

3.5
3

Client Sample ID:	SW-CCI-004(19')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-5	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	87.3
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	3E21096.D	1	07/30/09	OYA	07/29/09	OP39117	E3E960
Run #2							

Run #	Initial Weight	Final Volume
Run #1	35.0 g	1.0 ml
Run #2		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	160	27	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	160	40	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	160	34	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	160	40	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	650	350	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	650	35	ug/kg	
95-48-7	2-Methylphenol	ND	65	35	ug/kg	
	3&4-Methylphenol	ND	65	44	ug/kg	
88-75-5	2-Nitrophenol	ND	160	34	ug/kg	
100-02-7	4-Nitrophenol	ND	330	42	ug/kg	
87-86-5	Pentachlorophenol	ND	330	42	ug/kg	
108-95-2	Phenol	ND	65	25	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	160	35	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	160	44	ug/kg	
83-32-9	Acenaphthene	ND	33	17	ug/kg	
208-96-8	Acenaphthylene	ND	33	14	ug/kg	
98-86-2	Acetophenone	ND	160	16	ug/kg	
120-12-7	Anthracene	ND	33	15	ug/kg	
1912-24-9	Atrazine	ND	160	21	ug/kg	
56-55-3	Benzo(a)anthracene	ND	33	20	ug/kg	
50-32-8	Benzo(a)pyrene	ND	33	14	ug/kg	
205-99-2	Benzo(b)fluoranthene	ND	33	18	ug/kg	
191-24-2	Benzo(g,h,i)perylene	ND	33	16	ug/kg	
207-08-9	Benzo(k)fluoranthene	ND	33	16	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	65	18	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	65	16	ug/kg	
92-52-4	1,1'-Biphenyl	ND	65	16	ug/kg	
100-52-7	Benzaldehyde	ND	160	96	ug/kg	
91-58-7	2-Chloronaphthalene	ND	65	15	ug/kg	
106-47-8	4-Chloroaniline	ND	160	13	ug/kg	
86-74-8	Carbazole	ND	65	14	ug/kg	
105-60-2	Caprolactam	ND	65	26	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-004(19')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-5	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	87.3
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	ND	33	15	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	65	16	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	65	15	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	65	17	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	65	22	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	65	18	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	65	15	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	160	57	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	ND	33	16	ug/kg	
132-64-9	Dibenzofuran	ND	65	15	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	65	20	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	65	14	ug/kg	
84-66-2	Diethyl phthalate	ND	65	15	ug/kg	
131-11-3	Dimethyl phthalate	ND	65	16	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	82.7	65	17	ug/kg	
206-44-0	Fluoranthene	ND	33	15	ug/kg	
86-73-7	Fluorene	ND	33	15	ug/kg	
118-74-1	Hexachlorobenzene	ND	65	19	ug/kg	
87-68-3	Hexachlorobutadiene	ND	33	16	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	650	31	ug/kg	
67-72-1	Hexachloroethane	ND	160	21	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	ND	33	12	ug/kg	
78-59-1	Isophorone	ND	65	29	ug/kg	
91-57-6	2-Methylnaphthalene	ND	65	15	ug/kg	
88-74-4	2-Nitroaniline	ND	160	24	ug/kg	
99-09-2	3-Nitroaniline	ND	160	13	ug/kg	
100-01-6	4-Nitroaniline	ND	160	20	ug/kg	
91-20-3	Naphthalene	ND	33	14	ug/kg	
98-95-3	Nitrobenzene	ND	65	15	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	65	20	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	160	22	ug/kg	
85-01-8	Phenanthrene	ND	33	16	ug/kg	
129-00-0	Pyrene	16.3	33	14	ug/kg	J

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	63%		30-109%
4165-62-2	Phenol-d5	67%		28-108%
118-79-6	2,4,6-Tribromophenol	66%		28-125%
4165-60-0	Nitrobenzene-d5	86%		28-113%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest LabLink@514900 10:37 13-Aug-2009

Report of Analysis

Page 1 of 3

Client Sample ID:	SW-CCI-005(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-6	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	83.6
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	3E21093.D	1	07/30/09	OYA	07/29/09	OP39117	E3E960
Run #2							

Run #	Initial Weight	Final Volume
Run #1	35.0 g	1.0 ml
Run #2		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	170	28	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	170	41	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	170	36	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	170	42	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	680	370	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	680	36	ug/kg	
95-48-7	2-Methylphenol	ND	68	37	ug/kg	
	3&4-Methylphenol	ND	68	46	ug/kg	
88-75-5	2-Nitrophenol	ND	170	36	ug/kg	
100-02-7	4-Nitrophenol	ND	340	44	ug/kg	
87-86-5	Pentachlorophenol	ND	340	44	ug/kg	
108-95-2	Phenol	ND	68	26	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	170	37	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	170	45	ug/kg	
83-32-9	Acenaphthene	ND	34	18	ug/kg	
208-96-8	Acenaphthylene	ND	34	15	ug/kg	
98-86-2	Acetophenone	ND	170	17	ug/kg	
120-12-7	Anthracene	ND	34	15	ug/kg	
1912-24-9	Atrazine	ND	170	21	ug/kg	
56-55-3	Benzo(a)anthracene	ND	34	21	ug/kg	
50-32-8	Benzo(a)pyrene	ND	34	14	ug/kg	
205-99-2	Benzo(b)fluoranthene	ND	34	19	ug/kg	
191-24-2	Benzo(g,h,i)perylene	ND	34	17	ug/kg	
207-08-9	Benzo(k)fluoranthene	ND	34	17	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	68	19	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	68	17	ug/kg	
92-52-4	1,1'-Biphenyl	ND	68	17	ug/kg	
100-52-7	Benzaldehyde	ND	170	100	ug/kg	
91-58-7	2-Chloronaphthalene	ND	68	15	ug/kg	
106-47-8	4-Chloroaniline	ND	170	14	ug/kg	
86-74-8	Carbazole	ND	68	15	ug/kg	
105-60-2	Caprolactam	ND	68	27	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-005(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-6	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	83.6
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	ND	34	16	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	68	17	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	68	16	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	68	18	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	68	23	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	68	19	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	68	15	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	170	59	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	ND	34	16	ug/kg	
132-64-9	Dibenzofuran	ND	68	16	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	68	21	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	68	15	ug/kg	
84-66-2	Diethyl phthalate	ND	68	15	ug/kg	
131-11-3	Dimethyl phthalate	ND	68	16	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	ND	68	18	ug/kg	
206-44-0	Fluoranthene	ND	34	16	ug/kg	
86-73-7	Fluorene	ND	34	16	ug/kg	
118-74-1	Hexachlorobenzene	ND	68	19	ug/kg	
87-68-3	Hexachlorobutadiene	ND	34	16	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	680	32	ug/kg	
67-72-1	Hexachloroethane	ND	170	22	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	ND	34	13	ug/kg	
78-59-1	Isophorone	ND	68	30	ug/kg	
91-57-6	2-Methylnaphthalene	ND	68	15	ug/kg	
88-74-4	2-Nitroaniline	ND	170	25	ug/kg	
99-09-2	3-Nitroaniline	ND	170	14	ug/kg	
100-01-6	4-Nitroaniline	ND	170	21	ug/kg	
91-20-3	Naphthalene	ND	34	15	ug/kg	
98-95-3	Nitrobenzene	ND	68	15	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	68	21	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	170	23	ug/kg	
85-01-8	Phenanthrene	ND	34	17	ug/kg	
129-00-0	Pyrene	ND	34	15	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	70%		30-109%
4165-62-2	Phenol-d5	74%		28-108%
118-79-6	2,4,6-Tribromophenol	74%		28-125%
4165-60-0	Nitrobenzene-d5	93%		28-113%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest LabLink@514900 10:37 13-Aug-2009

Report of Analysis

Page 1 of 3

Client Sample ID:	SW-CCI-006(19')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-7	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	91.3
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	3E21097.D	1	07/30/09	OYA	07/29/09	OP39117	E3E960
Run #2							

Run #	Initial Weight	Final Volume
Run #1	35.3 g	1.0 ml
Run #2		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	160	26	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	160	38	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	160	32	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	160	38	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	620	340	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	620	33	ug/kg	
95-48-7	2-Methylphenol	ND	62	34	ug/kg	
	3&4-Methylphenol	ND	62	42	ug/kg	
88-75-5	2-Nitrophenol	ND	160	33	ug/kg	
100-02-7	4-Nitrophenol	ND	310	40	ug/kg	
87-86-5	Pentachlorophenol	ND	310	40	ug/kg	
108-95-2	Phenol	ND	62	23	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	160	34	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	160	41	ug/kg	
83-32-9	Acenaphthene	ND	31	16	ug/kg	
208-96-8	Acenaphthylene	ND	31	13	ug/kg	
98-86-2	Acetophenone	ND	160	15	ug/kg	
120-12-7	Anthracene	ND	31	14	ug/kg	
1912-24-9	Atrazine	ND	160	20	ug/kg	
56-55-3	Benzo(a)anthracene	ND	31	19	ug/kg	
50-32-8	Benzo(a)pyrene	ND	31	13	ug/kg	
205-99-2	Benzo(b)fluoranthene	ND	31	17	ug/kg	
191-24-2	Benzo(g,h,i)perylene	ND	31	15	ug/kg	
207-08-9	Benzo(k)fluoranthene	ND	31	16	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	62	17	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	62	15	ug/kg	
92-52-4	1,1'-Biphenyl	ND	62	16	ug/kg	
100-52-7	Benzaldehyde	ND	160	91	ug/kg	
91-58-7	2-Chloronaphthalene	ND	62	14	ug/kg	
106-47-8	4-Chloroaniline	ND	160	13	ug/kg	
86-74-8	Carbazole	ND	62	13	ug/kg	
105-60-2	Caprolactam	ND	62	25	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-006(19')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-7	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	91.3
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	ND	31	14	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	62	15	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	62	15	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	62	16	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	62	21	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	62	17	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	62	14	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	160	54	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	ND	31	15	ug/kg	
132-64-9	Dibenzofuran	ND	62	14	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	62	19	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	62	14	ug/kg	
84-66-2	Diethyl phthalate	ND	62	14	ug/kg	
131-11-3	Dimethyl phthalate	ND	62	15	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	ND	62	16	ug/kg	
206-44-0	Fluoranthene	ND	31	14	ug/kg	
86-73-7	Fluorene	ND	31	15	ug/kg	
118-74-1	Hexachlorobenzene	ND	62	18	ug/kg	
87-68-3	Hexachlorobutadiene	ND	31	15	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	620	29	ug/kg	
67-72-1	Hexachloroethane	ND	160	20	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	ND	31	11	ug/kg	
78-59-1	Isophorone	ND	62	27	ug/kg	
91-57-6	2-Methylnaphthalene	ND	62	14	ug/kg	
88-74-4	2-Nitroaniline	ND	160	23	ug/kg	
99-09-2	3-Nitroaniline	ND	160	13	ug/kg	
100-01-6	4-Nitroaniline	ND	160	19	ug/kg	
91-20-3	Naphthalene	ND	31	13	ug/kg	
98-95-3	Nitrobenzene	ND	62	14	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	62	19	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	160	21	ug/kg	
85-01-8	Phenanthrene	ND	31	15	ug/kg	
129-00-0	Pyrene	ND	31	14	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	69%		30-109%
4165-62-2	Phenol-d5	74%		28-108%
118-79-6	2,4,6-Tribromophenol	78%		28-125%
4165-60-0	Nitrobenzene-d5	93%		28-113%

ND = Not detected MDL - Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

Accutest LabLink@514900 10:37 13-Aug-2009

Report of Analysis

Page 1 of 3

Client Sample ID:	SW-CCI-007(19')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-8	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	84.5
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	3E21098.D	1	07/30/09	OYA	07/29/09	OP39117	E3E960
Run #2							

Run #	Initial Weight	Final Volume
Run #1	35.2 g	1.0 ml
Run #2		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	170	28	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	170	41	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	170	35	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	170	41	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	670	360	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	670	36	ug/kg	
95-48-7	2-Methylphenol	ND	67	36	ug/kg	
	3&4-Methylphenol	ND	67	45	ug/kg	
88-75-5	2-Nitrophenol	ND	170	35	ug/kg	
100-02-7	4-Nitrophenol	ND	340	43	ug/kg	
87-86-5	Pentachlorophenol	ND	340	43	ug/kg	
108-95-2	Phenol	ND	67	25	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	170	36	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	170	45	ug/kg	
83-32-9	Acenaphthene	ND	34	18	ug/kg	
208-96-8	Acenaphthylene	ND	34	14	ug/kg	
98-86-2	Acetophenone	ND	170	16	ug/kg	
120-12-7	Anthracene	ND	34	15	ug/kg	
1912-24-9	Atrazine	ND	170	21	ug/kg	
56-55-3	Benzo(a)anthracene	32.5	34	20	ug/kg	J
50-32-8	Benzo(a)pyrene	26.7	34	14	ug/kg	J
205-99-2	Benzo(b)fluoranthene	22.2	34	18	ug/kg	J
191-24-2	Benzo(g,h,i)perylene	44.3	34	16	ug/kg	
207-08-9	Benzo(k)fluoranthene	ND	34	17	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	67	18	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	67	17	ug/kg	
92-52-4	1,1'-Biphenyl	ND	67	17	ug/kg	
100-52-7	Benzaldehyde	ND	170	98	ug/kg	
91-58-7	2-Chloronaphthalene	ND	67	15	ug/kg	
106-47-8	4-Chloroaniline	ND	170	14	ug/kg	
86-74-8	Carbazole	ND	67	14	ug/kg	
105-60-2	Caprolactam	66.4	67	27	ug/kg	J

ND = Not detected MDL - Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-007(19')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-8	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	84.5
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	24.9	34	15	ug/kg	J
111-91-1	bis(2-Chloroethoxy)methane	ND	67	17	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	67	16	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	67	17	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	67	22	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	67	18	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	67	15	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	170	58	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	ND	34	16	ug/kg	
132-64-9	Dibenzofuran	ND	67	16	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	67	21	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	67	15	ug/kg	
84-66-2	Diethyl phthalate	ND	67	15	ug/kg	
131-11-3	Dimethyl phthalate	ND	67	16	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	341	67	18	ug/kg	
206-44-0	Fluoranthene	27.6	34	15	ug/kg	J
86-73-7	Fluorene	ND	34	16	ug/kg	
118-74-1	Hexachlorobenzene	ND	67	19	ug/kg	
87-68-3	Hexachlorobutadiene	ND	34	16	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	670	32	ug/kg	
67-72-1	Hexachloroethane	ND	170	22	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	19.4	34	12	ug/kg	J
78-59-1	Isophorone	ND	67	30	ug/kg	
91-57-6	2-Methylnaphthalene	ND	67	15	ug/kg	
88-74-4	2-Nitroaniline	ND	170	25	ug/kg	
99-09-2	3-Nitroaniline	ND	170	14	ug/kg	
100-01-6	4-Nitroaniline	ND	170	21	ug/kg	
91-20-3	Naphthalene	ND	34	15	ug/kg	
98-95-3	Nitrobenzene	ND	67	15	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	67	21	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	170	23	ug/kg	
85-01-8	Phenanthrene	ND	34	17	ug/kg	
129-00-0	Pyrene	83.5	34	15	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	62%		30-109%
4165-62-2	Phenol-d5	69%		28-108%
118-79-6	2,4,6-Tribromophenol	70%		28-125%
4165-60-0	Nitrobenzene-d5	85%		28-113%

ND = Not detected MDL - Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

Accutest LabLink@514900 10:37 13-Aug-2009

Report of Analysis

Page 1 of 3

Client Sample ID:	SW-CCI-008(13.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-9	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	85.4
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	3E21099.D	1	07/30/09	OYA	07/29/09	OP39117	E3E960

Run #1	Initial Weight	Final Volume
Run #2	35.2 g	1.0 ml

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	170	28	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	170	40	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	170	35	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	170	41	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	670	360	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	670	35	ug/kg	
95-48-7	2-Methylphenol	ND	67	36	ug/kg	
	3&4-Methylphenol	ND	67	45	ug/kg	
88-75-5	2-Nitrophenol	ND	170	35	ug/kg	
100-02-7	4-Nitrophenol	ND	330	43	ug/kg	
87-86-5	Pentachlorophenol	ND	330	43	ug/kg	
108-95-2	Phenol	ND	67	25	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	170	36	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	170	44	ug/kg	
83-32-9	Acenaphthene	ND	33	18	ug/kg	
208-96-8	Acenaphthylene	ND	33	14	ug/kg	
98-86-2	Acetophenone	ND	170	16	ug/kg	
120-12-7	Anthracene	ND	33	15	ug/kg	
1912-24-9	Atrazine	ND	170	21	ug/kg	
56-55-3	Benzo(a)anthracene	ND	33	20	ug/kg	
50-32-8	Benzo(a)pyrene	ND	33	14	ug/kg	
205-99-2	Benzo(b)fluoranthene	ND	33	18	ug/kg	
191-24-2	Benzo(g,h,i)perylene	ND	33	16	ug/kg	
207-08-9	Benzo(k)fluoranthene	ND	33	17	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	67	18	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	67	17	ug/kg	
92-52-4	1,1'-Biphenyl	ND	67	17	ug/kg	
100-52-7	Benzaldehyde	ND	170	97	ug/kg	
91-58-7	2-Chloronaphthalene	ND	67	15	ug/kg	
106-47-8	4-Chloroaniline	ND	170	14	ug/kg	
86-74-8	Carbazole	ND	67	14	ug/kg	
105-60-2	Caprolactam	ND	67	27	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-008(13.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-9	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	85.4
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	ND	33	15	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	67	17	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	67	16	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	67	17	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	67	22	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	67	18	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	67	15	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	170	58	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	ND	33	16	ug/kg	
132-64-9	Dibenzofuran	ND	67	16	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	67	21	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	67	15	ug/kg	
84-66-2	Diethyl phthalate	ND	67	15	ug/kg	
131-11-3	Dimethyl phthalate	ND	67	16	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	98.2	67	17	ug/kg	
206-44-0	Fluoranthene	ND	33	15	ug/kg	
86-73-7	Fluorene	ND	33	16	ug/kg	
118-74-1	Hexachlorobenzene	ND	67	19	ug/kg	
87-68-3	Hexachlorobutadiene	ND	33	16	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	670	31	ug/kg	
67-72-1	Hexachloroethane	ND	170	21	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	ND	33	12	ug/kg	
78-59-1	Isophorone	ND	67	29	ug/kg	
91-57-6	2-Methylnaphthalene	ND	67	15	ug/kg	
88-74-4	2-Nitroaniline	ND	170	24	ug/kg	
99-09-2	3-Nitroaniline	ND	170	13	ug/kg	
100-01-6	4-Nitroaniline	ND	170	21	ug/kg	
91-20-3	Naphthalene	ND	33	14	ug/kg	
98-95-3	Nitrobenzene	ND	67	15	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	67	21	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	170	22	ug/kg	
85-01-8	Phenanthrene	ND	33	16	ug/kg	
129-00-0	Pyrene	24.6	33	15	ug/kg	J

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	62%		30-109%
4165-62-2	Phenol-d5	70%		28-108%
118-79-6	2,4,6-Tribromophenol	77%		28-125%
4165-60-0	Nitrobenzene-d5	87%		28-113%

ND = Not detected MDL - Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

Accutest LabLink@514900 10:37 13-Aug-2009

Report of Analysis

Page 1 of 3

3.10
3

Client Sample ID:	SW-CCI-009(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-10	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	94.6
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	3E21100.D	1	07/30/09	OYA	07/29/09	OP39117	E3E960

Run #1	Initial Weight	Final Volume
Run #2	35.0 g	1.0 ml

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	150	25	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	150	37	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	150	31	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	150	37	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	600	330	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	600	32	ug/kg	
95-48-7	2-Methylphenol	ND	60	33	ug/kg	
	3&4-Methylphenol	ND	60	40	ug/kg	
88-75-5	2-Nitrophenol	ND	150	32	ug/kg	
100-02-7	4-Nitrophenol	ND	300	39	ug/kg	
87-86-5	Pentachlorophenol	ND	300	39	ug/kg	
108-95-2	Phenol	ND	60	23	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	150	33	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	150	40	ug/kg	
83-32-9	Acenaphthene	ND	30	16	ug/kg	
208-96-8	Acenaphthylene	ND	30	13	ug/kg	
98-86-2	Acetophenone	ND	150	15	ug/kg	
120-12-7	Anthracene	ND	30	14	ug/kg	
1912-24-9	Atrazine	ND	150	19	ug/kg	
56-55-3	Benzo(a)anthracene	ND	30	18	ug/kg	
50-32-8	Benzo(a)pyrene	ND	30	13	ug/kg	
205-99-2	Benzo(b)fluoranthene	ND	30	16	ug/kg	
191-24-2	Benzo(g,h,i)perylene	ND	30	15	ug/kg	
207-08-9	Benzo(k)fluoranthene	ND	30	15	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	60	16	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	60	15	ug/kg	
92-52-4	1,1'-Biphenyl	ND	60	15	ug/kg	
100-52-7	Benzaldehyde	ND	150	88	ug/kg	
91-58-7	2-Chloronaphthalene	ND	60	13	ug/kg	
106-47-8	4-Chloroaniline	ND	150	12	ug/kg	
86-74-8	Carbazole	ND	60	13	ug/kg	
105-60-2	Caprolactam	ND	60	24	ug/kg	

ND = Not detected MDL - Method Detection Limit

RL = Reporting Limit

E = Indicates value exceeds calibration range

J = Indicates an estimated value

B = Indicates analyte found in associated method blank

N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-009(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-10	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	94.6
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	ND	30	14	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	60	15	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	60	14	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	60	16	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	60	20	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	60	16	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	60	14	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	150	52	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	ND	30	14	ug/kg	
132-64-9	Dibenzofuran	ND	60	14	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	60	19	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	60	13	ug/kg	
84-66-2	Diethyl phthalate	ND	60	14	ug/kg	
131-11-3	Dimethyl phthalate	ND	60	14	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	ND	60	16	ug/kg	
206-44-0	Fluoranthene	ND	30	14	ug/kg	
86-73-7	Fluorene	ND	30	14	ug/kg	
118-74-1	Hexachlorobenzene	ND	60	17	ug/kg	
87-68-3	Hexachlorobutadiene	ND	30	14	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	600	29	ug/kg	
67-72-1	Hexachloroethane	ND	150	20	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	ND	30	11	ug/kg	
78-59-1	Isophorone	ND	60	27	ug/kg	
91-57-6	2-Methylnaphthalene	ND	60	14	ug/kg	
88-74-4	2-Nitroaniline	ND	150	22	ug/kg	
99-09-2	3-Nitroaniline	ND	150	12	ug/kg	
100-01-6	4-Nitroaniline	ND	150	19	ug/kg	
91-20-3	Naphthalene	ND	30	13	ug/kg	
98-95-3	Nitrobenzene	ND	60	14	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	60	19	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	150	20	ug/kg	
85-01-8	Phenanthrene	ND	30	15	ug/kg	
129-00-0	Pyrene	ND	30	13	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	58%		30-109%
4165-62-2	Phenol-d5	65%		28-108%
118-79-6	2,4,6-Tribromophenol	63%		28-125%
4165-60-0	Nitrobenzene-d5	83%		28-113%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 3

3.1
3

Client Sample ID:	SW-CCI-011(24.5)	Date Sampled:	08/10/09
Lab Sample ID:	JA25246-1	Date Received:	08/10/09
Matrix:	SO - Soil	Percent Solids:	83.6
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	2M23445.D	1	08/11/09	OYA	08/11/09	OP39365	E2M989
Run #2							

Run #	Initial Weight	Final Volume
Run #1	35.4 g	1.0 ml
Run #2		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	170	28	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	170	41	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	170	35	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	170	41	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	680	360	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	680	36	ug/kg	
95-48-7	2-Methylphenol	ND	68	36	ug/kg	
	3&4-Methylphenol	ND	68	45	ug/kg	
88-75-5	2-Nitrophenol	ND	170	35	ug/kg	
100-02-7	4-Nitrophenol	ND	340	43	ug/kg	
87-86-5	Pentachlorophenol	ND	340	44	ug/kg	
108-95-2	Phenol	ND	68	26	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	170	36	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	170	45	ug/kg	
83-32-9	Acenaphthene	ND	34	18	ug/kg	
208-96-8	Acenaphthylene	ND	34	14	ug/kg	
98-86-2	Acetophenone	ND	170	17	ug/kg	
120-12-7	Anthracene	ND	34	15	ug/kg	
1912-24-9	Atrazine	ND	170	21	ug/kg	
56-55-3	Benzo(a)anthracene	ND	34	20	ug/kg	
50-32-8	Benzo(a)pyrene	ND	34	14	ug/kg	
205-99-2	Benzo(b)fluoranthene	ND	34	18	ug/kg	
191-24-2	Benzo(g,h,i)perylene	ND	34	16	ug/kg	
207-08-9	Benzo(k)fluoranthene	ND	34	17	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	68	18	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	68	17	ug/kg	
92-52-4	1,1'-Biphenyl	ND	68	17	ug/kg	
100-52-7	Benzaldehyde	ND	170	99	ug/kg	
91-58-7	2-Chloronaphthalene	ND	68	15	ug/kg	
106-47-8	4-Chloroaniline	ND	170	14	ug/kg	
86-74-8	Carbazole	ND	68	14	ug/kg	
105-60-2	Caprolactam	ND	68	27	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-011(24.5)	Date Sampled:	08/10/09
Lab Sample ID:	JA25246-1	Date Received:	08/10/09
Matrix:	SO - Soil	Percent Solids:	83.6
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	ND	34	16	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	68	17	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	68	16	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	68	18	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	68	23	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	68	18	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	68	15	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	170	58	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	ND	34	16	ug/kg	
132-64-9	Dibenzofuran	ND	68	16	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	68	21	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	68	15	ug/kg	
84-66-2	Diethyl phthalate	ND	68	15	ug/kg	
131-11-3	Dimethyl phthalate	ND	68	16	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	ND	68	18	ug/kg	
206-44-0	Fluoranthene	ND	34	15	ug/kg	
86-73-7	Fluorene	ND	34	16	ug/kg	
118-74-1	Hexachlorobenzene	ND	68	19	ug/kg	
87-68-3	Hexachlorobutadiene	ND	34	16	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	680	32	ug/kg	
67-72-1	Hexachloroethane	ND	170	22	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	ND	34	13	ug/kg	
78-59-1	Isophorone	ND	68	30	ug/kg	
91-57-6	2-Methylnaphthalene	ND	68	15	ug/kg	
88-74-4	2-Nitroaniline	ND	170	25	ug/kg	
99-09-2	3-Nitroaniline	ND	170	14	ug/kg	
100-01-6	4-Nitroaniline	ND	170	21	ug/kg	
91-20-3	Naphthalene	ND	34	15	ug/kg	
98-95-3	Nitrobenzene	ND	68	15	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	68	21	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	170	23	ug/kg	
85-01-8	Phenanthrene	ND	34	17	ug/kg	
129-00-0	Pyrene	ND	34	15	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	69%		30-109%
4165-62-2	Phenol-d5	70%		28-108%
118-79-6	2,4,6-Tribromophenol	69%		28-125%
4165-60-0	Nitrobenzene-d5	78%		28-113%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

3.1
3

Client Sample ID:	SW-CCI-012	Date Sampled:	08/13/09
Lab Sample ID:	JA25522-1	Date Received:	08/13/09
Matrix:	SO - Soil	Percent Solids:	82.7
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	2M23780.D	1	08/26/09	OYA	08/21/09	OP39537	F2M1000
Run #2 ^a	M67017.D	1	08/31/09	LP	08/28/09	OP39654	EM2487

	Initial Weight	Final Volume
Run #1	35.6 g	1.0 ml
Run #2	35.0 g	1.0 ml

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	170	28	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	170	41	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	170	35	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	170	41	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	680	370	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	680	36	ug/kg	
95-48-7	2-Methylphenol	ND	68	37	ug/kg	
	3&4-Methylphenol	ND	68	46	ug/kg	
88-75-5	2-Nitrophenol	ND	170	36	ug/kg	
100-02-7	4-Nitrophenol	ND	340	43	ug/kg	
87-86-5	Pentachlorophenol	ND	340	44	ug/kg	
108-95-2	Phenol	ND	68	26	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	170	37	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	170	45	ug/kg	
83-32-9	Acenaphthene	ND	34	18	ug/kg	
208-96-8	Acenaphthylene	ND	34	15	ug/kg	
98-86-2	Acetophenone	ND	170	17	ug/kg	
120-12-7	Anthracene	ND	34	15	ug/kg	
1912-24-9	Atrazine	ND	170	21	ug/kg	
56-55-3	Benzo(a)anthracene	ND	34	21	ug/kg	
50-32-8	Benzo(a)pyrene	ND	34	14	ug/kg	
205-99-2	Benzo(b)fluoranthene	ND	34	18	ug/kg	
191-24-2	Benzo(g,h,i)perylene	ND	34	16	ug/kg	
207-08-9	Benzo(k)fluoranthene	ND	34	17	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	68	19	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	68	17	ug/kg	
92-52-4	1,1'-Biphenyl	ND	68	17	ug/kg	
100-52-7	Benzaldehyde	ND	170	99	ug/kg	
91-58-7	2-Chloronaphthalene	ND	68	15	ug/kg	
106-47-8	4-Chloroaniline	ND	170	14	ug/kg	
86-74-8	Carbazole	ND	68	14	ug/kg	
105-60-2	Caprolactam	ND	68	27	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-012	Date Sampled:	08/13/09
Lab Sample ID:	JA25522-1	Date Received:	08/13/09
Matrix:	SO - Soil	Percent Solids:	82.7
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	ND	34	16	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	68	17	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	68	16	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	68	18	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	68	23	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	68	19	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	68	15	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	170	59	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	ND	34	16	ug/kg	
132-64-9	Dibenzofuran	ND	68	16	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	68	21	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	68	15	ug/kg	
84-66-2	Diethyl phthalate	ND	68	15	ug/kg	
131-11-3	Dimethyl phthalate	ND	68	16	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	ND	68	18	ug/kg	
206-44-0	Fluoranthene	ND	34	15	ug/kg	
86-73-7	Fluorene	ND	34	16	ug/kg	
118-74-1	Hexachlorobenzene	ND	68	19	ug/kg	
87-68-3	Hexachlorobutadiene	ND	34	16	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	680	32	ug/kg	
67-72-1	Hexachloroethane	ND	170	22	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	ND	34	13	ug/kg	
78-59-1	Isophorone	ND	68	30	ug/kg	
91-57-6	2-Methylnaphthalene	ND	68	15	ug/kg	
88-74-4	2-Nitroaniline	ND	170	25	ug/kg	
99-09-2	3-Nitroaniline	ND	170	14	ug/kg	
100-01-6	4-Nitroaniline	ND	170	21	ug/kg	
91-20-3	Naphthalene	ND	34	15	ug/kg	
98-95-3	Nitrobenzene	ND	68	15	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	68	21	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	170	23	ug/kg	
85-01-8	Phenanthrene	ND	34	17	ug/kg	
129-00-0	Pyrene	ND	34	15	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	55%	54%	30-109%
4165-62-2	Phenol-d5	60%	53%	28-108%
118-79-6	2,4,6-Tribromophenol	66%	57%	28-125%
4165-60-0	Nitrobenzene-d5	60%	58%	28-113%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 3

Client Sample ID:	SW-CCI-013	Date Sampled:	08/13/09
Lab Sample ID:	JA25522-2	Date Received:	08/13/09
Matrix:	SO - Soil	Percent Solids:	88.2
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	2M23781.D	1	08/26/09	OYA	08/21/09	OP39537	E2M1000
Run #2 ^a	M67018.D	1	08/31/09	LP	08/28/09	OP39654	EM2487

	Initial Weight	Final Volume
Run #1	35.0 g	1.0 ml
Run #2	35.0 g	1.0 ml

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	160	27	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	160	39	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	160	34	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	160	40	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	650	350	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	650	34	ug/kg	
95-48-7	2-Methylphenol	ND	65	35	ug/kg	
	3&4-Methylphenol	ND	65	43	ug/kg	
88-75-5	2-Nitrophenol	ND	160	34	ug/kg	
100-02-7	4-Nitrophenol	ND	320	41	ug/kg	
87-86-5	Pentachlorophenol	ND	320	42	ug/kg	
108-95-2	Phenol	ND	65	24	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	160	35	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	160	43	ug/kg	
83-32-9	Acenaphthene	ND	32	17	ug/kg	
208-96-8	Acenaphthylene	ND	32	14	ug/kg	
98-86-2	Acetophenone	ND	160	16	ug/kg	
120-12-7	Anthracene	ND	32	15	ug/kg	
1912-24-9	Atrazine	ND	160	20	ug/kg	
56-55-3	Benzo(a)anthracene	60.3	32	20	ug/kg	
50-32-8	Benzo(a)pyrene	60.9	32	13	ug/kg	
205-99-2	Benzo(b)fluoranthene	70.4	32	18	ug/kg	
191-24-2	Benzo(g,h,i)perylene	53.0	32	16	ug/kg	
207-08-9	Benzo(k)fluoranthene	48.3	32	16	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	65	18	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	65	16	ug/kg	
92-52-4	1,1'-Biphenyl	ND	65	16	ug/kg	
100-52-7	Benzaldehyde	ND	160	95	ug/kg	
91-58-7	2-Chloronaphthalene	ND	65	14	ug/kg	
106-47-8	4-Chloroaniline	ND	160	13	ug/kg	
86-74-8	Carbazole	ND	65	14	ug/kg	
105-60-2	Caprolactam	ND	65	26	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-013	Date Sampled:	08/13/09
Lab Sample ID:	JA25522-2	Date Received:	08/13/09
Matrix:	SO - Soil	Percent Solids:	88.2
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	72.4	32	15	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	65	16	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	65	15	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	65	17	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	65	22	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	65	18	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	65	15	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	160	56	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	ND	32	15	ug/kg	
132-64-9	Dibenzofuran	ND	65	15	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	65	20	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	65	14	ug/kg	
84-66-2	Diethyl phthalate	ND	65	15	ug/kg	
131-11-3	Dimethyl phthalate	ND	65	15	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	37.9	65	17	ug/kg	J
206-44-0	Fluoranthene	117	32	15	ug/kg	
86-73-7	Fluorene	ND	32	15	ug/kg	
118-74-1	Hexachlorobenzene	ND	65	18	ug/kg	
87-68-3	Hexachlorobutadiene	ND	32	15	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	650	31	ug/kg	
67-72-1	Hexachloroethane	ND	160	21	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	43.1	32	12	ug/kg	
78-59-1	Isophorone	ND	65	28	ug/kg	
91-57-6	2-Methylnaphthalene	ND	65	15	ug/kg	
88-74-4	2-Nitroaniline	ND	160	24	ug/kg	
99-09-2	3-Nitroaniline	ND	160	13	ug/kg	
100-01-6	4-Nitroaniline	ND	160	20	ug/kg	
91-20-3	Naphthalene	ND	32	14	ug/kg	
98-95-3	Nitrobenzene	ND	65	15	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	65	20	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	160	22	ug/kg	
85-01-8	Phenanthrene	53.0	32	16	ug/kg	
129-00-0	Pyrene	108	32	14	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	56%	61%	30-109%
4165-62-2	Phenol-d5	56%	59%	28-108%
118-79-6	2,4,6-Tribromophenol	68%	68%	28-125%
4165-60-0	Nitrobenzene-d5	60%	66%	28-113%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Client Sample ID:	SW-CCI-014	Date Sampled:	08/13/09
Lab Sample ID:	JA25522-3	Date Received:	08/13/09
Matrix:	SO - Soil	Percent Solids:	85.6
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	2M23782.D	1	08/26/09	OYA	08/21/09	OP39537	E2M1000
Run #2 ^a	M67019.D	1	08/31/09	LP	08/28/09	OP39654	EM2487

Run #	Initial Weight	Final Volume
Run #1	35.6 g	1.0 ml
Run #2	35.4 g	1.0 ml

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	160	27	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	160	40	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	160	34	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	160	40	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	660	350	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	660	35	ug/kg	
95-48-7	2-Methylphenol	ND	66	35	ug/kg	
	3&4-Methylphenol	ND	66	44	ug/kg	
88-75-5	2-Nitrophenol	ND	160	34	ug/kg	
100-02-7	4-Nitrophenol	ND	330	42	ug/kg	
87-86-5	Pentachlorophenol	ND	330	42	ug/kg	
108-95-2	Phenol	ND	66	25	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	160	35	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	160	44	ug/kg	
83-32-9	Acenaphthene	ND	33	17	ug/kg	
208-96-8	Acenaphthylene	ND	33	14	ug/kg	
98-86-2	Acetophenone	ND	160	16	ug/kg	
120-12-7	Anthracene	ND	33	15	ug/kg	
1912-24-9	Atrazine	ND	160	21	ug/kg	
56-55-3	Benzo(a)anthracene	73.6	33	20	ug/kg	
50-32-8	Benzo(a)pyrene	71.7	33	14	ug/kg	
205-99-2	Benzo(b)fluoranthene	86.7	33	18	ug/kg	
191-24-2	Benzo(g,h,i)perylene	56.2	33	16	ug/kg	
207-08-9	Benzo(k)fluoranthene	52.2	33	16	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	66	18	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	66	16	ug/kg	
92-52-4	1,1'-Biphenyl	ND	66	17	ug/kg	
100-52-7	Benzaldehyde	ND	160	96	ug/kg	
91-58-7	2-Chloronaphthalene	ND	66	15	ug/kg	
106-47-8	4-Chloroaniline	ND	160	13	ug/kg	
86-74-8	Carbazole	ND	66	14	ug/kg	
105-60-2	Caprolactam	ND	66	26	ug/kg	

ND = Not detected MDL - Method Detection Limit J = Indicates an estimated value
 RL = Reporting Limit B = Indicates analyte found in associated method blank
 E = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-014	Date Sampled:	08/13/09
Lab Sample ID:	JA25522-3	Date Received:	08/13/09
Matrix:	SO - Soil	Percent Solids:	85.6
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	81.3	33	15	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	66	16	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	66	15	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	66	17	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	66	22	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	66	18	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	66	15	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	160	57	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	18.0	33	16	ug/kg	J
132-64-9	Dibenzofuran	ND	66	15	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	66	20	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	66	15	ug/kg	
84-66-2	Diethyl phthalate	ND	66	15	ug/kg	
131-11-3	Dimethyl phthalate	ND	66	16	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	ND	66	17	ug/kg	
206-44-0	Fluoranthene	154	33	15	ug/kg	
86-73-7	Fluorene	ND	33	15	ug/kg	
118-74-1	Hexachlorobenzene	ND	66	19	ug/kg	
87-68-3	Hexachlorobutadiene	ND	33	16	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	660	31	ug/kg	
67-72-1	Hexachloroethane	ND	160	21	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	51.5	33	12	ug/kg	
78-59-1	Isophorone	ND	66	29	ug/kg	
91-57-6	2-Methylnaphthalene	ND	66	15	ug/kg	
88-74-4	2-Nitroaniline	ND	160	24	ug/kg	
99-09-2	3-Nitroaniline	ND	160	13	ug/kg	
100-01-6	4-Nitroaniline	ND	160	21	ug/kg	
91-20-3	Naphthalene	ND	33	14	ug/kg	
98-95-3	Nitrobenzene	ND	66	15	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	66	20	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	160	22	ug/kg	
85-01-8	Phenanthrene	47.8	33	16	ug/kg	
129-00-0	Pyrene	140	33	14	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	66%	64%	30-109%
4165-62-2	Phenol-d5	67%	62%	28-108%
118-79-6	2,4,6-Tribromophenol	79%	72%	28-125%
4165-60-0	Nitrobenzene-d5	71%	70%	28-113%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 3

Client Sample ID:	SW-CCI-015	Date Sampled:	08/14/09
Lab Sample ID:	JA25638-1	Date Received:	08/14/09
Matrix:	SO - Soil	Percent Solids:	81.8
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	M66759.D	1	08/17/09	LP	08/14/09	OP39446	EM2470

Run #1	Initial Weight	Final Volume
Run #2	35.0 g	1.0 ml

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	170	29	ug/kg	
59-50-7	4-Chloro-3-methyl phenol	ND	170	42	ug/kg	
120-83-2	2,4-Dichlorophenol	ND	170	36	ug/kg	
105-67-9	2,4-Dimethylphenol	ND	170	43	ug/kg	
51-28-5	2,4-Dinitrophenol	ND	700	380	ug/kg	
534-52-1	4,6-Dinitro-o-cresol	ND	700	37	ug/kg	
95-48-7	2-Methylphenol	ND	70	38	ug/kg	
	3&4-Methylphenol	ND	70	47	ug/kg	
88-75-5	2-Nitrophenol	ND	170	37	ug/kg	
100-02-7	4-Nitrophenol	ND	350	45	ug/kg	
87-86-5	Pentachlorophenol	ND	350	45	ug/kg	
108-95-2	Phenol	ND	70	26	ug/kg	
95-95-4	2,4,5-Trichlorophenol	ND	170	38	ug/kg	
88-06-2	2,4,6-Trichlorophenol	ND	170	46	ug/kg	
83-32-9	Acenaphthene	ND	35	18	ug/kg	
208-96-8	Acenaphthylene	ND	35	15	ug/kg	
98-86-2	Acetophenone	ND	170	17	ug/kg	
120-12-7	Anthracene	ND	35	16	ug/kg	
1912-24-9	Atrazine	ND	170	22	ug/kg	
56-55-3	Benzo(a)anthracene	ND	35	21	ug/kg	
50-32-8	Benzo(a)pyrene	ND	35	14	ug/kg	
205-99-2	Benzo(b)fluoranthene	ND	35	19	ug/kg	
191-24-2	Benzo(g,h,i)perylene	ND	35	17	ug/kg	
207-08-9	Benzo(k)fluoranthene	ND	35	18	ug/kg	
101-55-3	4-Bromophenyl phenyl ether	ND	70	19	ug/kg	
85-68-7	Butyl benzyl phthalate	ND	70	17	ug/kg	
92-52-4	1,1'-Biphenyl	ND	70	18	ug/kg	
100-52-7	Benzaldehyde	ND	170	100	ug/kg	
91-58-7	2-Chloronaphthalene	ND	70	16	ug/kg	
106-47-8	4-Chloroaniline	ND	170	14	ug/kg	
86-74-8	Carbazole	ND	70	15	ug/kg	
105-60-2	Caprolactam	ND	70	28	ug/kg	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-015	Date Sampled:	08/14/09
Lab Sample ID:	JA25638-1	Date Received:	08/14/09
Matrix:	SO - Soil	Percent Solids:	81.8
Method:	SW846 8270C SW846 3550B		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	ND	35	16	ug/kg	
111-91-1	bis(2-Chloroethoxy)methane	ND	70	17	ug/kg	
111-44-4	bis(2-Chloroethyl)ether	ND	70	16	ug/kg	
108-60-1	bis(2-Chloroisopropyl)ether	ND	70	18	ug/kg	
7005-72-3	4-Chlorophenyl phenyl ether	ND	70	23	ug/kg	
121-14-2	2,4-Dinitrotoluene	ND	70	19	ug/kg	
606-20-2	2,6-Dinitrotoluene	ND	70	16	ug/kg	
91-94-1	3,3'-Dichlorobenzidine	ND	170	60	ug/kg	
53-70-3	Dibenzo(a,h)anthracene	ND	35	17	ug/kg	
132-64-9	Dibenzofuran	ND	70	16	ug/kg	
84-74-2	Di-n-butyl phthalate	ND	70	22	ug/kg	
117-84-0	Di-n-octyl phthalate	ND	70	15	ug/kg	
84-66-2	Diethyl phthalate	ND	70	16	ug/kg	
131-11-3	Dimethyl phthalate	ND	70	17	ug/kg	
117-81-7	bis(2-Ethylhexyl)phthalate	ND	70	18	ug/kg	
206-44-0	Fluoranthene	ND	35	16	ug/kg	
86-73-7	Fluorene	ND	35	16	ug/kg	
118-74-1	Hexachlorobenzene	ND	70	20	ug/kg	
87-68-3	Hexachlorobutadiene	ND	35	17	ug/kg	
77-47-4	Hexachlorocyclopentadiene	ND	700	33	ug/kg	
67-72-1	Hexachloroethane	ND	170	23	ug/kg	
193-39-5	Indeno(1,2,3-cd)pyrene	ND	35	13	ug/kg	
78-59-1	Isophorone	ND	70	31	ug/kg	
91-57-6	2-Methylnaphthalene	ND	70	16	ug/kg	
88-74-4	2-Nitroaniline	ND	170	26	ug/kg	
99-09-2	3-Nitroaniline	ND	170	14	ug/kg	
100-01-6	4-Nitroaniline	ND	170	22	ug/kg	
91-20-3	Naphthalene	ND	35	15	ug/kg	
98-95-3	Nitrobenzene	ND	70	16	ug/kg	
621-64-7	N-Nitroso-di-n-propylamine	ND	70	22	ug/kg	
86-30-6	N-Nitrosodiphenylamine	ND	170	24	ug/kg	
85-01-8	Phenanthrene	ND	35	17	ug/kg	
129-00-0	Pyrene	ND	35	15	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	64%		30-109%
4165-62-2	Phenol-d5	58%		28-108%
118-79-6	2,4,6-Tribromophenol	99%		28-125%
4165-60-0	Nitrobenzene-d5	78%		28-113%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest LabLink@514900 10:37 13-Aug-2009

Report of Analysis

3.1
3

Client Sample ID:	FB072009	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-1	Date Received:	07/21/09
Matrix:	AQ - Field Blank Soil	Percent Solids:	n/a
Method:	SW846 8270C SW846 3510C		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	M66213.D	1	07/24/09	LP	07/23/09	OP39030	EM2441
Run #2							

Run #	Initial Volume	Final Volume
Run #1	1000 ml	1.0 ml
Run #2		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
95-57-8	2-Chlorophenol	ND	5.0	1.1	ug/l	
59-50-7	4-Chloro-3-methyl phenol	ND	5.0	1.1	ug/l	
120-83-2	2,4-Dichlorophenol	ND	5.0	1.2	ug/l	
105-67-9	2,4-Dimethylphenol	ND	5.0	1.7	ug/l	
51-28-5	2,4-Dinitrophenol	ND	20	0.74	ug/l	
534-52-1	4,6-Dinitro-o-cresol	ND	20	0.51	ug/l	
95-48-7	2-Methylphenol	ND	2.0	1.1	ug/l	
	3&4-Methylphenol	ND	2.0	1.0	ug/l	
88-75-5	2-Nitrophenol	ND	5.0	1.2	ug/l	
100-02-7	4-Nitrophenol	ND	10	0.83	ug/l	
87-86-5	Pentachlorophenol	ND	10	0.80	ug/l	
108-95-2	Phenol	ND	2.0	0.58	ug/l	
95-95-4	2,4,5-Trichlorophenol	ND	5.0	1.3	ug/l	
88-06-2	2,4,6-Trichlorophenol	ND	5.0	1.2	ug/l	
83-32-9	Acenaphthene	ND	1.0	0.37	ug/l	
208-96-8	Acenaphthylene	ND	1.0	0.27	ug/l	
98-86-2	Acetophenone	ND	5.0	0.40	ug/l	
120-12-7	Anthracene	ND	1.0	0.16	ug/l	
1912-24-9	Atrazine	ND	5.0	0.39	ug/l	
100-52-7	Benzaldehyde	ND	5.0	0.40	ug/l	
56-55-3	Benzo(a)anthracene	ND	1.0	0.12	ug/l	
50-32-8	Benzo(a)pyrene	ND	1.0	0.095	ug/l	
205-99-2	Benzo(b)fluoranthene	ND	1.0	0.25	ug/l	
191-24-2	Benzo(g,h,i)perylene	ND	1.0	0.12	ug/l	
207-08-9	Benzo(k)fluoranthene	ND	1.0	0.38	ug/l	
101-55-3	4-Bromophenyl phenyl ether	ND	2.0	0.35	ug/l	
85-68-7	Butyl benzyl phthalate	ND	2.0	0.25	ug/l	
92-52-4	1,1'-Biphenyl	ND	2.0	0.42	ug/l	
91-58-7	2-Chloronaphthalene	ND	5.0	0.42	ug/l	
106-47-8	4-Chloroaniline	ND	5.0	0.25	ug/l	
86-74-8	Carbazole	ND	2.0	0.17	ug/l	
105-60-2	Caprolactam	ND	2.0	0.20	ug/l	

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	FB072009	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-1	Date Received:	07/21/09
Matrix:	AQ - Field Blank Soil	Percent Solids:	n/a
Method:	SW846 8270C SW846 3510C		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

ABN TCL List (CLP4.2 list)

CAS No.	Compound	Result	RL	MDL	Units	Q
218-01-9	Chrysene	ND	1.0	0.11	ug/l	
111-91-1	bis(2-Chloroethoxy)methane	ND	2.0	0.25	ug/l	
111-44-4	bis(2-Chloroethyl)ether	ND	2.0	0.31	ug/l	
108-60-1	bis(2-Chloroisopropyl)ether	ND	2.0	0.39	ug/l	
7005-72-3	4-Chlorophenyl phenyl ether	ND	2.0	0.35	ug/l	
121-14-2	2,4-Dinitrotoluene	ND	2.0	0.22	ug/l	
606-20-2	2,6-Dinitrotoluene	ND	2.0	0.33	ug/l	
91-94-1	3,3'-Dichlorobenzidine	ND	5.0	0.30	ug/l	
53-70-3	Dibenzo(a,h)anthracene	ND	1.0	0.15	ug/l	
132-64-9	Dibenzofuran	ND	5.0	0.30	ug/l	
84-74-2	Di-n-butyl phthalate	ND	2.0	0.19	ug/l	
117-84-0	Di-n-octyl phthalate	ND	2.0	0.40	ug/l	
84-66-2	Diethyl phthalate	ND	2.0	0.17	ug/l	
131-11-3	Dimethyl phthalate	ND	2.0	0.23	ug/l	
117-81-7	bis(2-Ethylhexyl)phthalate	ND	2.0	0.33	ug/l	
206-44-0	Fluoranthene	ND	1.0	0.17	ug/l	
86-73-7	Fluorene	ND	1.0	0.27	ug/l	
118-74-1	Hexachlorobenzene	ND	2.0	0.37	ug/l	
87-68-3	Hexachlorobutadiene	ND	1.0	0.37	ug/l	
77-47-4	Hexachlorocyclopentadiene	ND	20	0.67	ug/l	
67-72-1	Hexachloroethane	ND	5.0	0.26	ug/l	
193-39-5	Indeno(1,2,3-cd)pyrene	ND	1.0	0.13	ug/l	
78-59-1	Isophorone	ND	2.0	0.25	ug/l	
91-57-6	2-Methylnaphthalene	ND	2.0	0.66	ug/l	
88-74-4	2-Nitroaniline	ND	5.0	0.24	ug/l	
99-09-2	3-Nitroaniline	ND	5.0	0.29	ug/l	
100-01-6	4-Nitroaniline	ND	5.0	0.18	ug/l	
91-20-3	Naphthalene	ND	1.0	0.43	ug/l	
98-95-3	Nitrobenzene	ND	2.0	0.25	ug/l	
621-64-7	N-Nitroso-di-n-propylamine	ND	2.0	0.44	ug/l	
86-30-6	N-Nitrosodiphenylamine	ND	5.0	0.22	ug/l	
85-01-8	Phenanthrene	ND	1.0	0.21	ug/l	
129-00-0	Pyrene	ND	1.0	0.16	ug/l	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
367-12-4	2-Fluorophenol	47%		13-68%
4165-62-2	Phenol-d5	30%		10-49%
118-79-6	2,4,6-Tribromophenol	107%		37-130%
4165-60-0	Nitrobenzene-d5	95%		25-112%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

3.1
3

Client Sample ID:	B-CCI-02(08)	Date Sampled:	06/30/09
Lab Sample ID:	JA22090-1	Date Received:	06/30/09
Matrix:	SO - Soil	Percent Solids:	83.0
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	3G38572.D	1	07/01/09	TDR	07/01/09	OP38644	G3G1441
Run #2	3G38577.D	2	07/01/09	TDR	07/01/09	OP38644	G3G1441

Run #1	Initial Weight	Final Volume
Run #1	18.0 g	10.0 ml
Run #2	18.0 g	10.0 ml

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.3	0.27	ug/kg	
319-84-6	alpha-BHC	ND	1.3	0.25	ug/kg	
319-85-7	beta-BHC	ND	1.3	0.45	ug/kg	
319-86-8	delta-BHC	ND	1.3	0.24	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.3	0.33	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.3	0.41	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.3	0.30	ug/kg	
60-57-1	Dieldrin	3.5	1.3	0.26	ug/kg	J
72-54-8	4,4'-DDD	6.7	1.3	0.22	ug/kg	
72-55-9	4,4'-DDE	43.8	1.3	0.33	ug/kg	
50-29-3	4,4'-DDT	86.5	2.7	0.62	ug/kg	
72-20-8	Endrin	ND	1.3	0.29	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.3	0.29	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.3	0.31	ug/kg	
959-98-8	Endosulfan-I	ND	1.3	0.27	ug/kg	
33213-65-9	Endosulfan-II	ND	1.3	0.44	ug/kg	
76-44-8	Heptachlor	ND	1.3	0.35	ug/kg	
1024-57-3	Heptachlor epoxide	1.4	1.3	0.27	ug/kg	
72-43-5	Methoxychlor	ND	1.3	0.36	ug/kg	
53494-70-5	Endrin ketone	ND	1.3	0.27	ug/kg	
8001-35-2	Toxaphene	ND	17	13	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	68%	67%	28-138%
877-09-8	Tetrachloro-m-xylene	63%	62%	28-138%
2051-24-3	Decachlorobiphenyl	85%	81%	22-156%
2051-24-3	Decachlorobiphenyl	93%	84%	22-156%

- (a) Reported from 1st signal. %D of check on 2nd signal excess method criteria (15 %) so using for confirmation only.
- (b) Result is from Run# 2

ND = Not detected MDL - Method Detection Limit J = Indicates an estimated value
 RL = Reporting Limit B = Indicates analyte found in associated method blank
 E = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 1

Client Sample ID:	B-CCI-003 (20')	Date Sampled:	07/15/09
Lab Sample ID:	JA23207-1	Date Received:	07/15/09
Matrix:	SO - Soil	Percent Solids:	91.6
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	XX87924.D	1	07/16/09	JSE	07/16/09	OP38917	GXX3531

Run #1	Initial Weight	Final Volume
Run #2	17.1 g	10.0 ml

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.3	0.26	ug/kg	
319-84-6	alpha-BHC	ND	1.3	0.24	ug/kg	
319-85-7	beta-BHC	ND	1.3	0.43	ug/kg	
319-86-8	delta-BHC	ND	1.3	0.23	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.3	0.31	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.3	0.39	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.3	0.28	ug/kg	
60-57-1	Dieldrin	ND	1.3	0.25	ug/kg	
72-54-8	4,4'-DDD	ND	1.3	0.21	ug/kg	
72-55-9	4,4'-DDE	ND	1.3	0.32	ug/kg	
50-29-3	4,4'-DDT	ND	1.3	0.30	ug/kg	
72-20-8	Endrin	ND	1.3	0.27	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.3	0.27	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.3	0.30	ug/kg	
959-98-8	Endosulfan-I	ND	1.3	0.26	ug/kg	
33213-65-9	Endosulfan-II	ND	1.3	0.42	ug/kg	
76-44-8	Heptachlor	ND	1.3	0.34	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.3	0.26	ug/kg	
72-43-5	Methoxychlor	ND	1.3	0.34	ug/kg	
53494-70-5	Endrin ketone	ND	1.3	0.25	ug/kg	
8001-35-2	Toxaphene	ND	16	12	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	64%		28-138%
877-09-8	Tetrachloro-m-xylene	63%		28-138%
2051-24-3	Decachlorobiphenyl	66%		22-156%
2051-24-3	Decachlorobiphenyl	79%		22-156%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 1

Client Sample ID:	B-CCI-004 (20')	Date Sampled:	07/15/09
Lab Sample ID:	JA23207-2	Date Received:	07/15/09
Matrix:	SO - Soil	Percent Solids:	91.6
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	XX87925.D	1	07/16/09	JSE	07/16/09	OP38917	GXX3531
Run #2							

Run #	Initial Weight	Final Volume
Run #1	17.1 g	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.3	0.26	ug/kg	
319-84-6	alpha-BHC	ND	1.3	0.24	ug/kg	
319-85-7	beta-BHC	ND	1.3	0.43	ug/kg	
319-86-8	delta-BHC	ND	1.3	0.23	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.3	0.31	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.3	0.39	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.3	0.28	ug/kg	
60-57-1	Dieldrin	ND	1.3	0.25	ug/kg	
72-54-8	4,4'-DDD	ND	1.3	0.21	ug/kg	
72-55-9	4,4'-DDE	ND	1.3	0.32	ug/kg	
50-29-3	4,4'-DDT	ND	1.3	0.30	ug/kg	
72-20-8	Endrin	ND	1.3	0.27	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.3	0.27	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.3	0.30	ug/kg	
959-98-8	Endosulfan-I	ND	1.3	0.26	ug/kg	
33213-65-9	Endosulfan-II	ND	1.3	0.42	ug/kg	
76-44-8	Heptachlor	ND	1.3	0.34	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.3	0.26	ug/kg	
72-43-5	Methoxychlor	ND	1.3	0.34	ug/kg	
53494-70-5	Endrin ketone	ND	1.3	0.25	ug/kg	
8001-35-2	Toxaphene	ND	16	12	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	73%		28-138%
877-09-8	Tetrachloro-m-xylene	72%		28-138%
2051-24-3	Decachlorobiphenyl	68%		22-156%
2051-24-3	Decachlorobiphenyl	81%		22-156%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 1

Client Sample ID:	B-CCI-006 (15')	Date Sampled:	07/22/09
Lab Sample ID:	JA23843-1	Date Received:	07/23/09
Matrix:	SO - Soil	Percent Solids:	90.6
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	1G46885.D	1	07/24/09	OPM	07/23/09	OP39027	G1G1717
Run #2							

Run #	Initial Weight	Final Volume
Run #1	17.3 g	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.3	0.26	ug/kg	
319-84-6	alpha-BHC	ND	1.3	0.24	ug/kg	
319-85-7	beta-BHC	ND	1.3	0.43	ug/kg	
319-86-8	delta-BHC	ND	1.3	0.23	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.3	0.31	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.3	0.39	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.3	0.28	ug/kg	
60-57-1	Dieldrin	ND	1.3	0.25	ug/kg	
72-54-8	4,4'-DDD	ND	1.3	0.21	ug/kg	
72-55-9	4,4'-DDE	ND	1.3	0.32	ug/kg	
50-29-3	4,4'-DDT	ND	1.3	0.30	ug/kg	
72-20-8	Endrin	ND	1.3	0.27	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.3	0.27	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.3	0.30	ug/kg	
959-98-8	Endosulfan-I	ND	1.3	0.26	ug/kg	
33213-65-9	Endosulfan-II	ND	1.3	0.42	ug/kg	
76-44-8	Heptachlor	ND	1.3	0.34	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.3	0.26	ug/kg	
72-43-5	Methoxychlor	ND	1.3	0.34	ug/kg	
53494-70-5	Endrin ketone	ND	1.3	0.25	ug/kg	
8001-35-2	Toxaphene	ND	16	12	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	101%		28-138%
877-09-8	Tetrachloro-m-xylene	95%		28-138%
2051-24-3	Decachlorobiphenyl	76%		22-156%
2051-24-3	Decachlorobiphenyl	91%		22-156%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 1

Client Sample ID:	B-CCI-007 (20')	Date Sampled:	07/22/09
Lab Sample ID:	JA23843-2	Date Received:	07/23/09
Matrix:	SO - Soil	Percent Solids:	94.5
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	1G46886.D	1	07/24/09	OPM	07/23/09	OP39027	G1G1717
Run #2							

Run #	Initial Weight	Final Volume
Run #1	17.1 g	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.2	0.25	ug/kg	
319-84-6	alpha-BHC	ND	1.2	0.23	ug/kg	
319-85-7	beta-BHC	ND	1.2	0.41	ug/kg	
319-86-8	delta-BHC	ND	1.2	0.23	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.2	0.30	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.2	0.38	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.2	0.28	ug/kg	
60-57-1	Dieldrin	ND	1.2	0.24	ug/kg	
72-54-8	4,4'-DDD	ND	1.2	0.20	ug/kg	
72-55-9	4,4'-DDE	ND	1.2	0.31	ug/kg	
50-29-3	4,4'-DDT	ND	1.2	0.29	ug/kg	
72-20-8	Endrin	ND	1.2	0.27	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.2	0.26	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.2	0.29	ug/kg	
959-98-8	Endosulfan-I	ND	1.2	0.25	ug/kg	
33213-65-9	Endosulfan-II	ND	1.2	0.41	ug/kg	
76-44-8	Heptachlor	ND	1.2	0.33	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.2	0.25	ug/kg	
72-43-5	Methoxychlor	ND	1.2	0.33	ug/kg	
53494-70-5	Endrin ketone	ND	1.2	0.25	ug/kg	
8001-35-2	Toxaphene	ND	15	12	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	70%		28-138%
877-09-8	Tetrachloro-m-xylene	85%		28-138%
2051-24-3	Decachlorobiphenyl	91%		22-156%
2051-24-3	Decachlorobiphenyl	90%		22-156%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

3.5
3

Client Sample ID:	DUPLICATE		Date Sampled:	07/22/09
Lab Sample ID:	JA23843-5		Date Received:	07/23/09
Matrix:	SO - Soil		Percent Solids:	94.2
Method:	SW846 8081A SW846 3545			
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY			

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	1G46889.D	1	07/24/09	OPM	07/23/09	OP39027	G1G1717
Run #2							

Run #	Initial Weight	Final Volume
Run #1	17.3 g	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.2	0.25	ug/kg	
319-84-6	alpha-BHC	ND	1.2	0.23	ug/kg	
319-85-7	beta-BHC	ND	1.2	0.41	ug/kg	
319-86-8	delta-BHC	ND	1.2	0.22	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.2	0.30	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.2	0.38	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.2	0.27	ug/kg	
60-57-1	Dieldrin	ND	1.2	0.24	ug/kg	
72-54-8	4,4'-DDD	ND	1.2	0.20	ug/kg	
72-55-9	4,4'-DDE	ND	1.2	0.31	ug/kg	
50-29-3	4,4'-DDT	ND	1.2	0.28	ug/kg	
72-20-8	Endrin	ND	1.2	0.26	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.2	0.26	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.2	0.29	ug/kg	
959-98-8	Endosulfan-I	ND	1.2	0.25	ug/kg	
33213-65-9	Endosulfan-II	ND	1.2	0.41	ug/kg	
76-44-8	Heptachlor	ND	1.2	0.33	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.2	0.25	ug/kg	
72-43-5	Methoxychlor	ND	1.2	0.33	ug/kg	
53494-70-5	Endrin ketone	ND	1.2	0.24	ug/kg	
8001-35-2	Toxaphene	ND	15	12	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	42%		28-138%
877-09-8	Tetrachloro-m-xylene	61%		28-138%
2051-24-3	Decachlorobiphenyl	46%		22-156%
2051-24-3	Decachlorobiphenyl	57%		22-156%

ND = Not detected MDL - Method Detection Limit J = Indicates an estimated value
 RL = Reporting Limit B = Indicates analyte found in associated method blank
 E = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

3.1
3

Client Sample ID:	B-CCI-010(22')	Date Sampled:	07/24/09
Lab Sample ID:	JA24059-1	Date Received:	07/27/09
Matrix:	SO - Soil	Percent Solids:	82.1
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	1G46961.D	1	07/28/09	OPM	07/28/09	OP39013	G1G1719

Run #1	Initial Weight	Final Volume
Run #2	17.4 g	10.0 ml

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.4	0.28	ug/kg	
319-84-6	alpha-BHC	ND	1.4	0.26	ug/kg	
319-85-7	beta-BHC	ND	1.4	0.47	ug/kg	
319-86-8	delta-BHC	ND	1.4	0.26	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.4	0.34	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.4	0.43	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.4	0.31	ug/kg	
60-57-1	Dieldrin	ND	1.4	0.27	ug/kg	
72-54-8	4,4'-DDD	ND	1.4	0.23	ug/kg	
72-55-9	4,4'-DDE	ND	1.4	0.35	ug/kg	
50-29-3	4,4'-DDT	ND	1.4	0.32	ug/kg	
72-20-8	Endrin	ND	1.4	0.30	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.4	0.30	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.4	0.33	ug/kg	
959-98-8	Endosulfan-I	ND	1.4	0.29	ug/kg	
33213-65-9	Endosulfan-II	ND	1.4	0.46	ug/kg	
76-44-8	Heptachlor	ND	1.4	0.37	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.4	0.28	ug/kg	
72-43-5	Methoxychlor	ND	1.4	0.38	ug/kg	
53494-70-5	Endrin ketone	ND	1.4	0.28	ug/kg	
8001-35-2	Toxaphene	ND	18	13	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	76%		28-138%
877-09-8	Tetrachloro-m-xylene	68%		28-138%
2051-24-3	Decachlorobiphenyl	71%		22-156%
2051-24-3	Decachlorobiphenyl	87%		22-156%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-013(10')	Date Sampled:	07/27/09
Lab Sample ID:	JA24059-3	Date Received:	07/27/09
Matrix:	SO - Soil	Percent Solids:	82.9
Method:	SW846 8081A SW846 3545	Project: Courtlandt Corners I, 364 East 161st Street, Bronx, NY	

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	1G46963.D	1	07/28/09	OPM	07/28/09	OP39013	G1G1719
Run #2							

Run #	Initial Weight	Final Volume
Run #1	17.2 g	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.4	0.28	ug/kg	
319-84-6	alpha-BHC	ND	1.4	0.26	ug/kg	
319-85-7	beta-BHC	ND	1.4	0.47	ug/kg	
319-86-8	delta-BHC	ND	1.4	0.26	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.4	0.34	ug/kg	
5103-71-9	alpha-Chlordane	2.1	1.4	0.43	ug/kg	
5103-74-2	gamma-Chlordane	1.6	1.4	0.31	ug/kg	
60-57-1	Dieldrin	ND	1.4	0.27	ug/kg	
72-54-8	4,4'-DDD	1.9	1.4	0.23	ug/kg	J
72-55-9	4,4'-DDE	8.3	1.4	0.35	ug/kg	
50-29-3	4,4'-DDT	45.0	1.4	0.32	ug/kg	
72-20-8	Endrin	ND	1.4	0.30	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.4	0.30	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.4	0.33	ug/kg	
959-98-8	Endosulfan-I	ND	1.4	0.29	ug/kg	
33213-65-9	Endosulfan-II	ND	1.4	0.46	ug/kg	
76-44-8	Heptachlor	ND	1.4	0.37	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.4	0.28	ug/kg	
72-43-5	Methoxychlor	ND	1.4	0.38	ug/kg	
53494-70-5	Endrin ketone	ND	1.4	0.28	ug/kg	
8001-35-2	Toxaphene	ND	18	13	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	88%		28-138%
877-09-8	Tetrachloro-m-xylene	82%		28-138%
2051-24-3	Decachlorobiphenyl	93%		22-156%
2051-24-3	Decachlorobiphenyl	93%		22-156%

ND = Not detected MDL - Method Detection Limit J = Indicates an estimated value
 RL = Reporting Limit B = Indicates analyte found in associated method blank
 E = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 1

Client Sample ID:	B-CCI-022	Date Sampled:	08/04/09
Lab Sample ID:	JA24692-1	Date Received:	08/04/09
Matrix:	SO - Soil	Percent Solids:	87.0
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	1G47134.D	1	08/05/09	OPM	08/05/09	OP39253	G1G1725
Run #2							

Run #	Initial Weight	Final Volume
Run #1	17.1 g	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.3	0.27	ug/kg	
319-84-6	alpha-BHC	ND	1.3	0.25	ug/kg	
319-85-7	beta-BHC	ND	1.3	0.45	ug/kg	
319-86-8	delta-BHC	ND	1.3	0.25	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.3	0.33	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.3	0.41	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.3	0.30	ug/kg	
60-57-1	Dieldrin	ND	1.3	0.26	ug/kg	
72-54-8	4,4'-DDD	ND	1.3	0.22	ug/kg	
72-55-9	4,4'-DDE	ND	1.3	0.34	ug/kg	
50-29-3	4,4'-DDT	3.1	1.3	0.31	ug/kg	
72-20-8	Endrin	ND	1.3	0.29	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.3	0.29	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.3	0.32	ug/kg	
959-98-8	Endosulfan-I	ND	1.3	0.27	ug/kg	
33213-65-9	Endosulfan-II	ND	1.3	0.44	ug/kg	
76-44-8	Heptachlor	ND	1.3	0.36	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.3	0.27	ug/kg	
72-43-5	Methoxychlor	ND	1.3	0.36	ug/kg	
53494-70-5	Endrin ketone	ND	1.3	0.27	ug/kg	
8001-35-2	Toxaphene	ND	17	13	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	70%		28-138%
877-09-8	Tetrachloro-m-xylene	67%		28-138%
2051-24-3	Decachlorobiphenyl	85%		22-156%
2051-24-3	Decachlorobiphenyl	80%		22-156%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-023	Date Sampled:	08/04/09
Lab Sample ID:	JA24692-2	Date Received:	08/04/09
Matrix:	SO - Soil	Percent Solids:	84.1
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	1G47135.D	1	08/05/09	OPM	08/05/09	OP39253	G1G1725
Run #2							

Run #	Initial Weight	Final Volume
Run #1	17.2 g	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.4	0.28	ug/kg	
319-84-6	alpha-BHC	ND	1.4	0.26	ug/kg	
319-85-7	beta-BHC	ND	1.4	0.46	ug/kg	
319-86-8	delta-BHC	ND	1.4	0.25	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.4	0.34	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.4	0.42	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.4	0.31	ug/kg	
60-57-1	Dieldrin	ND	1.4	0.27	ug/kg	
72-54-8	4,4'-DDD	ND	1.4	0.22	ug/kg	
72-55-9	4,4'-DDE	ND	1.4	0.34	ug/kg	
50-29-3	4,4'-DDT	ND	1.4	0.32	ug/kg	
72-20-8	Endrin	ND	1.4	0.30	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.4	0.29	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.4	0.32	ug/kg	
959-98-8	Endosulfan-I	ND	1.4	0.28	ug/kg	
33213-65-9	Endosulfan-II	ND	1.4	0.46	ug/kg	
76-44-8	Heptachlor	ND	1.4	0.37	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.4	0.28	ug/kg	
72-43-5	Methoxychlor	ND	1.4	0.37	ug/kg	
53494-70-5	Endrin ketone	ND	1.4	0.28	ug/kg	
8001-35-2	Toxaphene	ND	17	13	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	64%		28-138%
877-09-8	Tetrachloro-m-xylene	62%		28-138%
2051-24-3	Decachlorobiphenyl	69%		22-156%
2051-24-3	Decachlorobiphenyl	69%		22-156%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 1

Client Sample ID:	B-CCI-024	Date Sampled:	08/04/09
Lab Sample ID:	JA24692-3	Date Received:	08/04/09
Matrix:	SO - Soil	Percent Solids:	89.0
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	1G47136.D	1	08/05/09	OPM	08/05/09	OP39253	G1G1725
Run #2							

Run #	Initial Weight	Final Volume
Run #1	17.2 g	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.3	0.26	ug/kg	
319-84-6	alpha-BHC	ND	1.3	0.24	ug/kg	
319-85-7	beta-BHC	ND	1.3	0.44	ug/kg	
319-86-8	delta-BHC	ND	1.3	0.24	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.3	0.32	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.3	0.40	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.3	0.29	ug/kg	
60-57-1	Dieldrin	ND	1.3	0.25	ug/kg	
72-54-8	4,4'-DDD	ND	1.3	0.21	ug/kg	
72-55-9	4,4'-DDE	ND	1.3	0.33	ug/kg	
50-29-3	4,4'-DDT	ND	1.3	0.30	ug/kg	
72-20-8	Endrin	ND	1.3	0.28	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.3	0.28	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.3	0.31	ug/kg	
959-98-8	Endosulfan-I	ND	1.3	0.27	ug/kg	
33213-65-9	Endosulfan-II	ND	1.3	0.43	ug/kg	
76-44-8	Heptachlor	ND	1.3	0.35	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.3	0.26	ug/kg	
72-43-5	Methoxychlor	ND	1.3	0.35	ug/kg	
53494-70-5	Endrin ketone	ND	1.3	0.26	ug/kg	
8001-35-2	Toxaphene	ND	16	12	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	86%		28-138%
877-09-8	Tetrachloro-m-xylene	85%		28-138%
2051-24-3	Decachlorobiphenyl	101%		22-156%
2051-24-3	Decachlorobiphenyl	98%		22-156%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

3.4
3

Client Sample ID:	B-CCI-025	Date Sampled:	08/04/09
Lab Sample ID:	JA24692-4	Date Received:	08/04/09
Matrix:	SO - Soil	Percent Solids:	94.7
Method:	SW846 8081A SW846 3545	Project: Courtlandt Corners I, 364 East 161st Street, Bronx, NY	

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	1G47138.D	1	08/05/09	OPM	08/05/09	OP39253	G1G1725
Run #2							

Run #	Initial Weight	Final Volume
Run #1	17.1 g	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.2	0.25	ug/kg	
319-84-6	alpha-BHC	ND	1.2	0.23	ug/kg	
319-85-7	beta-BHC	ND	1.2	0.41	ug/kg	
319-86-8	delta-BHC	ND	1.2	0.23	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.2	0.30	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.2	0.38	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.2	0.27	ug/kg	
60-57-1	Dieldrin	ND	1.2	0.24	ug/kg	
72-54-8	4,4'-DDD	ND	1.2	0.20	ug/kg	
72-55-9	4,4'-DDE	ND	1.2	0.31	ug/kg	
50-29-3	4,4'-DDT	ND	1.2	0.29	ug/kg	
72-20-8	Endrin	ND	1.2	0.27	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.2	0.26	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.2	0.29	ug/kg	
959-98-8	Endosulfan-I	ND	1.2	0.25	ug/kg	
33213-65-9	Endosulfan-II	ND	1.2	0.41	ug/kg	
76-44-8	Heptachlor	ND	1.2	0.33	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.2	0.25	ug/kg	
72-43-5	Methoxychlor	ND	1.2	0.33	ug/kg	
53494-70-5	Endrin ketone	ND	1.2	0.25	ug/kg	
8001-35-2	Toxaphene	ND	15	12	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	53%		28-138%
877-09-8	Tetrachloro-m-xylene	58%		28-138%
2051-24-3	Decachlorobiphenyl	66%		22-156%
2051-24-3	Decachlorobiphenyl	67%		22-156%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

3.1
3

Client Sample ID:	B-CCI-026 (2')	Date Sampled:	08/06/09
Lab Sample ID:	JA25087-1	Date Received:	08/07/09
Matrix:	SO - Soil	Percent Solids:	83.0
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	1G47222.D	1	08/10/09	OPM	08/07/09	OP39322	G1G1728
Run #2							

Run #	Initial Weight	Final Volume
Run #1	17.1 g	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.4	0.28	ug/kg	
319-84-6	alpha-BHC	ND	1.4	0.26	ug/kg	
319-85-7	beta-BHC	ND	1.4	0.47	ug/kg	
319-86-8	delta-BHC	ND	1.4	0.26	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.4	0.35	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.4	0.43	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.4	0.31	ug/kg	
60-57-1	Dieldrin	ND	1.4	0.27	ug/kg	
72-54-8	4,4'-DDD	ND	1.4	0.23	ug/kg	
72-55-9	4,4'-DDE	ND	1.4	0.35	ug/kg	
50-29-3	4,4'-DDT	3.1	1.4	0.33	ug/kg	
72-20-8	Endrin	ND	1.4	0.30	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.4	0.30	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.4	0.33	ug/kg	
959-98-8	Endosulfan-I	ND	1.4	0.29	ug/kg	
33213-65-9	Endosulfan-II	ND	1.4	0.47	ug/kg	
76-44-8	Heptachlor	ND	1.4	0.37	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.4	0.28	ug/kg	
72-43-5	Methoxychlor	ND	1.4	0.38	ug/kg	
53494-70-5	Endrin ketone	ND	1.4	0.28	ug/kg	
8001-35-2	Toxaphene	ND	18	13	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	79%		28-138%
877-09-8	Tetrachloro-m-xylene	64%		28-138%
2051-24-3	Decachlorobiphenyl	67%		22-156%
2051-24-3	Decachlorobiphenyl	80%		22-156%

ND = Not detected MDL - Method Detection Limit J = Indicates an estimated value
 RL = Reporting Limit B = Indicates analyte found in associated method blank
 E = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Report of Analysis

32
3

Client Sample ID:	B-CCI-027 (2')	Date Sampled:	08/06/09
Lab Sample ID:	JA25087-2	Date Received:	08/07/09
Matrix:	SO - Soil	Percent Solids:	88.1
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	1G47223.D	1	08/10/09	OPM	08/07/09	OP39322	G1G1728
Run #2	1G47224.D	10	08/10/09	OPM	08/07/09	OP39322	G1G1728

Run #	Initial Weight	Final Volume
Run #1	17.1 g	10.0 ml
Run #2	17.1 g	10.0 ml

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.3	0.27	ug/kg	
319-84-6	alpha-BHC	ND	1.3	0.24	ug/kg	
319-85-7	beta-BHC	ND	1.3	0.44	ug/kg	
319-86-8	delta-BHC	ND	1.3	0.24	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.3	0.33	ug/kg	
5103-71-9	alpha-Chlordane	14.0	1.3	0.41	ug/kg	
5103-74-2	gamma-Chlordane	15.4	1.3	0.30	ug/kg	
60-57-1	Dieldrin ^a	6.0	1.3	0.26	ug/kg	
72-54-8	4,4'-DDD	12.8	1.3	0.22	ug/kg	
72-55-9	4,4'-DDE	130 ^b	13	3.3	ug/kg	
50-29-3	4,4'-DDT	283 ^b	13	3.1	ug/kg	
72-20-8	Endrin	ND	1.3	0.29	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.3	0.28	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.3	0.31	ug/kg	
959-98-8	Endosulfan-I	ND	1.3	0.27	ug/kg	
33213-65-9	Endosulfan-II	ND	1.3	0.44	ug/kg	
76-44-8	Heptachlor	ND	1.3	0.35	ug/kg	
1024-57-3	Heptachlor epoxide	4.7	1.3	0.27	ug/kg	
72-43-5	Methoxychlor	ND	1.3	0.36	ug/kg	
53494-70-5	Endrin ketone	ND	1.3	0.26	ug/kg	
8001-35-2	Toxaphene	ND	17	13	ug/kg	

444444
5

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	60%	69%	28-138%
877-09-8	Tetrachloro-m-xylene	52%	53%	28-138%
2051-24-3	Decachlorobiphenyl	239% ^c	237% ^c	22-156%
2051-24-3	Decachlorobiphenyl	235% ^c	244% ^c	22-156%

- (a) More than 40 % RPD for detected concentrations between the two GC columns.
- (b) Result is from Run# 2
- (c) Outside control limits due to matrix interference.

ND = Not detected MDL - Method Detection Limit J = Indicates an estimated value
 RL = Reporting Limit B = Indicates analyte found in associated method blank
 E = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 1

Client Sample ID:	B-CCI-028(2')	Date Sampled:	08/11/09
Lab Sample ID:	JA25446-1	Date Received:	08/12/09
Matrix:	SO - Soil	Percent Solids:	79.6
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	3G39359.D	1	08/15/09	TDR	08/13/09	OP39415	G3G1473

Run #1	Initial Weight	Final Volume
Run #2	17.3 g	10.0 ml

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.5	0.29	ug/kg	
319-84-6	alpha-BHC	ND	1.5	0.27	ug/kg	
319-85-7	beta-BHC	ND	1.5	0.49	ug/kg	
319-86-8	delta-BHC	ND	1.5	0.27	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.5	0.36	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.5	0.44	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.5	0.32	ug/kg	
60-57-1	Dieldrin	ND	1.5	0.28	ug/kg	
72-54-8	4,4'-DDD	ND	1.5	0.24	ug/kg	
72-55-9	4,4'-DDE	ND	1.5	0.36	ug/kg	
50-29-3	4,4'-DDT	ND	1.5	0.34	ug/kg	
72-20-8	Endrin	ND	1.5	0.31	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.5	0.31	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.5	0.34	ug/kg	
959-98-8	Endosulfan-I	ND	1.5	0.30	ug/kg	
33213-65-9	Endosulfan-II	ND	1.5	0.48	ug/kg	
76-44-8	Heptachlor	ND	1.5	0.38	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.5	0.29	ug/kg	
72-43-5	Methoxychlor	ND	1.5	0.39	ug/kg	
53494-70-5	Endrin ketone	ND	1.5	0.29	ug/kg	
8001-35-2	Toxaphene	ND	18	14	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	96%		28-138%
877-09-8	Tetrachloro-m-xylene	93%		28-138%
2051-24-3	Decachlorobiphenyl	102%		22-156%
2051-24-3	Decachlorobiphenyl	121%		22-156%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 1

3.2
3

Client Sample ID:	B-CCI-029(20')	Date Sampled:	08/12/09
Lab Sample ID:	JA25446-2	Date Received:	08/12/09
Matrix:	SO - Soil	Percent Solids:	79.4
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	3G39360.D	1	08/15/09	TDR	08/13/09	OP39415	G3G1473
Run #2							

Run #1	Initial Weight	Final Volume
Run #1	17.4 g	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.4	0.29	ug/kg	
319-84-6	alpha-BHC	ND	1.4	0.27	ug/kg	
319-85-7	beta-BHC	ND	1.4	0.48	ug/kg	
319-86-8	delta-BHC	ND	1.4	0.26	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.4	0.36	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.4	0.44	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.4	0.32	ug/kg	
60-57-1	Dieldrin	ND	1.4	0.28	ug/kg	
72-54-8	4,4'-DDD	ND	1.4	0.23	ug/kg	
72-55-9	4,4'-DDE	ND	1.4	0.36	ug/kg	
50-29-3	4,4'-DDT	ND	1.4	0.34	ug/kg	
72-20-8	Endrin	ND	1.4	0.31	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.4	0.31	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.4	0.34	ug/kg	
959-98-8	Endosulfan-I	ND	1.4	0.30	ug/kg	
33213-65-9	Endosulfan-II	ND	1.4	0.48	ug/kg	
76-44-8	Heptachlor	ND	1.4	0.38	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.4	0.29	ug/kg	
72-43-5	Methoxychlor	ND	1.4	0.39	ug/kg	
53494-70-5	Endrin ketone	ND	1.4	0.29	ug/kg	
8001-35-2	Toxaphene	ND	18	14	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	84%		28-138%
877-09-8	Tetrachloro-m-xylene	82%		28-138%
2051-24-3	Decachlorobiphenyl	95%		22-156%
2051-24-3	Decachlorobiphenyl	106%		22-156%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 1

Client Sample ID:	B-CCI-030(4')	Date Sampled:	08/12/09
Lab Sample ID:	JA25446-3	Date Received:	08/12/09
Matrix:	SO - Soil	Percent Solids:	89.1
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	3G39361.D	1	08/15/09	TDR	08/13/09	OP39415	G3G1473
Run #2							

Run #1	Initial Weight	Final Volume
Run #1	17.2 g	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.3	0.26	ug/kg	
319-84-6	alpha-BHC	ND	1.3	0.24	ug/kg	
319-85-7	beta-BHC	ND	1.3	0.44	ug/kg	
319-86-8	delta-BHC	ND	1.3	0.24	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.3	0.32	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.3	0.40	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.3	0.29	ug/kg	
60-57-1	Dieldrin ^a	2.2	1.3	0.25	ug/kg	J
72-54-8	4,4'-DDD	ND	1.3	0.21	ug/kg	
72-55-9	4,4'-DDE	4.6	1.3	0.33	ug/kg	
50-29-3	4,4'-DDT	23.7	1.3	0.30	ug/kg	
72-20-8	Endrin	ND	1.3	0.28	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.3	0.28	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.3	0.31	ug/kg	
959-98-8	Endosulfan-I	ND	1.3	0.27	ug/kg	
33213-65-9	Endosulfan-II	ND	1.3	0.43	ug/kg	
76-44-8	Heptachlor	ND	1.3	0.35	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.3	0.26	ug/kg	
72-43-5	Methoxychlor	ND	1.3	0.35	ug/kg	
53494-70-5	Endrin ketone	3.0	1.3	0.26	ug/kg	
8001-35-2	Toxaphene	ND	16	12	ug/kg	

CAS No.	Surrrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	67%		28-138%
877-09-8	Tetrachloro-m-xylene	65%		28-138%
2051-24-3	Decachlorobiphenyl	104%		22-156%
2051-24-3	Decachlorobiphenyl	108%		22-156%

(a) More than 40 % RPD for detected concentrations between the two GC columns.

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

3.4
3

Client Sample ID:	B-CCI-031(10')	Date Sampled:	08/12/09
Lab Sample ID:	JA25446-4	Date Received:	08/12/09
Matrix:	SO - Soil	Percent Solids:	89.4
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	3G39362.D	1	08/15/09	TDR	08/13/09	OP39415	G3G1473
Run #2							

Run #1	Initial Weight	Final Volume
Run #1	17.0 g	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.3	0.27	ug/kg	
319-84-6	alpha-BHC	ND	1.3	0.24	ug/kg	
319-85-7	beta-BHC	ND	1.3	0.44	ug/kg	
319-86-8	delta-BHC	ND	1.3	0.24	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.3	0.32	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.3	0.40	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.3	0.29	ug/kg	
60-57-1	Dieldrin	ND	1.3	0.26	ug/kg	
72-54-8	4,4'-DDD	ND	1.3	0.21	ug/kg	
72-55-9	4,4'-DDE	ND	1.3	0.33	ug/kg	
50-29-3	4,4'-DDT	ND	1.3	0.30	ug/kg	
72-20-8	Endrin	ND	1.3	0.28	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.3	0.28	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.3	0.31	ug/kg	
959-98-8	Endosulfan-I	ND	1.3	0.27	ug/kg	
33213-65-9	Endosulfan-II	ND	1.3	0.44	ug/kg	
76-44-8	Heptachlor	ND	1.3	0.35	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.3	0.26	ug/kg	
72-43-5	Methoxychlor	ND	1.3	0.35	ug/kg	
53494-70-5	Endrin ketone	ND	1.3	0.26	ug/kg	
8001-35-2	Toxaphene	ND	16	12	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	75%		28-138%
877-09-8	Tetrachloro-m-xylene	67%		28-138%
2051-24-3	Decachlorobiphenyl	109%		22-156%
2051-24-3	Decachlorobiphenyl	96%		22-156%

ND = Not detected MDL - Method Detection Limit J = Indicates an estimated value
 RL = Reporting Limit B = Indicates analyte found in associated method blank
 E = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Accutest LabLink@514900 10:37 13-Aug-2009

Report of Analysis

Page 1 of 1

Client Sample ID:	SW-CCI-001(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-2	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	81.1
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	1G46946.D	1	07/27/09	OPM	07/27/09	OP39078	G1G1718
Run #2							

Run #	Initial Weight	Final Volume
Run #1	17.1 g	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.4	0.29	ug/kg	
319-84-6	alpha-BHC	ND	1.4	0.27	ug/kg	
319-85-7	beta-BHC	ND	1.4	0.48	ug/kg	
319-86-8	delta-BHC	ND	1.4	0.26	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.4	0.35	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.4	0.44	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.4	0.32	ug/kg	
60-57-1	Dieldrin	ND	1.4	0.28	ug/kg	
72-54-8	4,4'-DDD	ND	1.4	0.23	ug/kg	
72-55-9	4,4'-DDE	ND	1.4	0.36	ug/kg	
50-29-3	4,4'-DDT	ND	1.4	0.33	ug/kg	
72-20-8	Endrin	ND	1.4	0.31	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.4	0.31	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.4	0.34	ug/kg	
959-98-8	Endosulfan-I	ND	1.4	0.29	ug/kg	
33213-65-9	Endosulfan-II	ND	1.4	0.48	ug/kg	
76-44-8	Heptachlor	ND	1.4	0.38	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.4	0.29	ug/kg	
72-43-5	Methoxychlor	ND	1.4	0.39	ug/kg	
53494-70-5	Endrin ketone	ND	1.4	0.29	ug/kg	
8001-35-2	Toxaphene	ND	18	14	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	100%		28-138%
877-09-8	Tetrachloro-m-xylene	81%		28-138%
2051-24-3	Decachlorobiphenyl	111%		22-156%
2051-24-3	Decachlorobiphenyl	103%		22-156%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest LabLink@514900 10:37 13-Aug-2009

Report of Analysis

Page 1 of 1

Client Sample ID:	SW-CCI-002(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-3	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	86.1
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	1G46947.D	1	07/27/09	OPM	07/27/09	OP39078	G1G1718
Run #2							

Run #	Initial Weight	Final Volume
Run #1	17.0 g	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.4	0.28	ug/kg	
319-84-6	alpha-BHC	ND	1.4	0.25	ug/kg	
319-85-7	beta-BHC	ND	1.4	0.46	ug/kg	
319-86-8	delta-BHC	ND	1.4	0.25	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.4	0.34	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.4	0.42	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.4	0.30	ug/kg	
60-57-1	Dieldrin	ND	1.4	0.27	ug/kg	
72-54-8	4,4'-DDD	ND	1.4	0.22	ug/kg	
72-55-9	4,4'-DDE	ND	1.4	0.34	ug/kg	
50-29-3	4,4'-DDT	ND	1.4	0.32	ug/kg	
72-20-8	Endrin	ND	1.4	0.29	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.4	0.29	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.4	0.32	ug/kg	
959-98-8	Endosulfan-I	ND	1.4	0.28	ug/kg	
33213-65-9	Endosulfan-II	ND	1.4	0.45	ug/kg	
76-44-8	Heptachlor	ND	1.4	0.36	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.4	0.27	ug/kg	
72-43-5	Methoxychlor	ND	1.4	0.37	ug/kg	
53494-70-5	Endrin ketone	ND	1.4	0.27	ug/kg	
8001-35-2	Toxaphene	ND	17	13	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	98%		28-138%
877-09-8	Tetrachloro-m-xylene	92%		28-138%
2051-24-3	Decachlorobiphenyl	130%		22-156%
2051-24-3	Decachlorobiphenyl	104%		22-156%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest LabLink@514900 10:37 13-Aug-2009

Report of Analysis

Page 1 of 1

Client Sample ID:	SW-CCI-003(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-4	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	88.4
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	3G39039.D	1	07/29/09	TDR	07/28/09	OP39110	G3G1461
Run #2							

Run #1	Initial Weight	Final Volume
Run #1	17.1 g	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.3	0.27	ug/kg	
319-84-6	alpha-BHC	ND	1.3	0.24	ug/kg	
319-85-7	beta-BHC	ND	1.3	0.44	ug/kg	
319-86-8	delta-BHC	ND	1.3	0.24	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.3	0.32	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.3	0.40	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.3	0.29	ug/kg	
60-57-1	Dieldrin	ND	1.3	0.26	ug/kg	
72-54-8	4,4'-DDD	ND	1.3	0.21	ug/kg	
72-55-9	4,4'-DDE	ND	1.3	0.33	ug/kg	
50-29-3	4,4'-DDT	ND	1.3	0.31	ug/kg	
72-20-8	Endrin	ND	1.3	0.28	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.3	0.28	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.3	0.31	ug/kg	
959-98-8	Endosulfan-I	ND	1.3	0.27	ug/kg	
33213-65-9	Endosulfan-II	ND	1.3	0.44	ug/kg	
76-44-8	Heptachlor	ND	1.3	0.35	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.3	0.26	ug/kg	
72-43-5	Methoxychlor	ND	1.3	0.35	ug/kg	
53494-70-5	Endrin ketone	ND	1.3	0.26	ug/kg	
8001-35-2	Toxaphene	ND	17	13	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	91%		28-138%
877-09-8	Tetrachloro-m-xylene	86%		28-138%
2051-24-3	Decachlorobiphenyl	97%		22-156%
2051-24-3	Decachlorobiphenyl	97%		22-156%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

3.5
3

Client Sample ID:	SW-CCI-004(19')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-5	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	87.3
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	1G46981.D	1	07/29/09	OPM	07/27/09	OP39078	G1G1720
Run #2							

Run #	Initial Weight	Final Volume
Run #1	17.1 g	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.3	0.27	ug/kg	
319-84-6	alpha-BHC	ND	1.3	0.25	ug/kg	
319-85-7	beta-BHC	ND	1.3	0.45	ug/kg	
319-86-8	delta-BHC	ND	1.3	0.25	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.3	0.33	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.3	0.41	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.3	0.30	ug/kg	
60-57-1	Dieldrin	ND	1.3	0.26	ug/kg	
72-54-8	4,4'-DDD	ND	1.3	0.22	ug/kg	
72-55-9	4,4'-DDE	ND	1.3	0.33	ug/kg	
50-29-3	4,4'-DDT	ND	1.3	0.31	ug/kg	
72-20-8	Endrin	ND	1.3	0.29	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.3	0.29	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.3	0.31	ug/kg	
959-98-8	Endosulfan-I	ND	1.3	0.27	ug/kg	
33213-65-9	Endosulfan-II	ND	1.3	0.44	ug/kg	
76-44-8	Heptachlor	ND	1.3	0.36	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.3	0.27	ug/kg	
72-43-5	Methoxychlor	ND	1.3	0.36	ug/kg	
53494-70-5	Endrin ketone	ND	1.3	0.27	ug/kg	
8001-35-2	Toxaphene	ND	17	13	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	102%		28-138%
877-09-8	Tetrachloro-m-xylene	100%		28-138%
2051-24-3	Decachlorobiphenyl	112%		22-156%
2051-24-3	Decachlorobiphenyl	116%		22-156%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest LabLink@514900 10:37 13-Aug-2009

Report of Analysis

3.6
3

Client Sample ID:	SW-CCI-005(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-6	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	83.6
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	1G46937.D	1	07/27/09	OPM	07/27/09	OP39078	G1G1718
Run #2							

Run #	Initial Weight	Final Volume
Run #1	17.0 g	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.4	0.28	ug/kg	
319-84-6	alpha-BHC	ND	1.4	0.26	ug/kg	
319-85-7	beta-BHC	ND	1.4	0.47	ug/kg	
319-86-8	delta-BHC	ND	1.4	0.26	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.4	0.35	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.4	0.43	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.4	0.31	ug/kg	
60-57-1	Dieldrin	ND	1.4	0.27	ug/kg	
72-54-8	4,4'-DDD	ND	1.4	0.23	ug/kg	
72-55-9	4,4'-DDE	ND	1.4	0.35	ug/kg	J
50-29-3	4,4'-DDT	ND	1.4	0.33	ug/kg	
72-20-8	Endrin	ND	1.4	0.30	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.4	0.30	ug/kg	J
7421-93-4	Endrin aldehyde	ND	1.4	0.33	ug/kg	J
959-98-8	Endosulfan-I	ND	1.4	0.29	ug/kg	J
33213-65-9	Endosulfan-II	ND	1.4	0.47	ug/kg	
76-44-8	Heptachlor	ND	1.4	0.37	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.4	0.28	ug/kg	
72-43-5	Methoxychlor	ND	1.4	0.38	ug/kg	
53494-70-5	Endrin ketone	ND	1.4	0.28	ug/kg	J
8001-35-2	Toxaphene	ND	18	13	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	129%		28-138%
877-09-8	Tetrachloro-m-xylene	107%		28-138%
2051-24-3	Decachlorobiphenyl	121%		22-156%
2051-24-3	Decachlorobiphenyl	126%		22-156%

ND = Not detected MDL - Method Detection Limit J = Indicates an estimated value
 RL = Reporting Limit B = Indicates analyte found in associated method blank
 E = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Accutest LabLink@514900 10:37 13-Aug-2009

Report of Analysis

Page 1 of 1

Client Sample ID:	SW-CCI-006(19')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-7	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	91.3
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	1G46949.D	1	07/27/09	OPM	07/27/09	OP39078	G1G1718
Run #2							

Run #	Initial Weight	Final Volume
Run #1	17.3 g	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.3	0.26	ug/kg	
319-84-6	alpha-BHC	ND	1.3	0.23	ug/kg	
319-85-7	beta-BHC	ND	1.3	0.42	ug/kg	
319-86-8	delta-BHC	ND	1.3	0.23	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.3	0.31	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.3	0.39	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.3	0.28	ug/kg	
60-57-1	Dieldrin	ND	1.3	0.25	ug/kg	
72-54-8	4,4'-DDD	ND	1.3	0.21	ug/kg	
72-55-9	4,4'-DDE	ND	1.3	0.32	ug/kg	
50-29-3	4,4'-DDT	ND	1.3	0.29	ug/kg	
72-20-8	Endrin	ND	1.3	0.27	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.3	0.27	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.3	0.30	ug/kg	
959-98-8	Endosulfan-I	ND	1.3	0.26	ug/kg	
33213-65-9	Endosulfan-II	ND	1.3	0.42	ug/kg	
76-44-8	Heptachlor	ND	1.3	0.34	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.3	0.25	ug/kg	
72-43-5	Methoxychlor	ND	1.3	0.34	ug/kg	
53494-70-5	Endrin ketone	ND	1.3	0.25	ug/kg	
8001-35-2	Toxaphene	ND	16	12	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	86%		28-138%
877-09-8	Tetrachloro-m-xylene	70%		28-138%
2051-24-3	Decachlorobiphenyl	75%		22-156%
2051-24-3	Decachlorobiphenyl	83%		22-156%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-007(19')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-8	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	84.5
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	1G46950.D	1	07/27/09	OPM	07/27/09	OP39078	G1G1718
Run #2							

Run #	Initial Weight	Final Volume
Run #1	17.0 g	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.4	0.28	ug/kg	
319-84-6	alpha-BHC	ND	1.4	0.26	ug/kg	
319-85-7	beta-BHC	ND	1.4	0.47	ug/kg	
319-86-8	delta-BHC	ND	1.4	0.25	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.4	0.34	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.4	0.43	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.4	0.31	ug/kg	
60-57-1	Dieldrin	ND	1.4	0.27	ug/kg	
72-54-8	4,4'-DDD ^a	2.3	1.4	0.23	ug/kg	J
72-55-9	4,4'-DDE	ND	1.4	0.35	ug/kg	
50-29-3	4,4'-DDT	ND	1.4	0.32	ug/kg	
72-20-8	Endrin	ND	1.4	0.30	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.4	0.30	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.4	0.33	ug/kg	
959-98-8	Endosulfan-I	ND	1.4	0.28	ug/kg	
33213-65-9	Endosulfan-II	ND	1.4	0.46	ug/kg	
76-44-8	Heptachlor	ND	1.4	0.37	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.4	0.28	ug/kg	
72-43-5	Methoxychlor	ND	1.4	0.37	ug/kg	
53494-70-5	Endrin ketone	ND	1.4	0.28	ug/kg	
8001-35-2	Toxaphene	ND	17	13	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	99%		28-138%
877-09-8	Tetrachloro-m-xylene	92%		28-138%
2051-24-3	Decachlorobiphenyl	77%		22-156%
2051-24-3	Decachlorobiphenyl	88%		22-156%

(a) More than 40 % RPD for detected concentrations between the two GC columns.

ND = Not detected MDL - Method Detection Limit J = Indicates an estimated value
 RL = Reporting Limit B = Indicates analyte found in associated method blank
 E = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Report of Analysis

3.9
3

Client Sample ID:	SW-CCI-008(13.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-9	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	85.4
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #1	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #2	1G46968.D	1	07/28/09	OPM	07/27/09	OP39078	G1G1719

Run #1	Initial Weight	Final Volume
Run #2	17.1 g	10.0 ml

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.4	0.28	ug/kg	
319-84-6	alpha-BHC	ND	1.4	0.25	ug/kg	
319-85-7	beta-BHC	ND	1.4	0.46	ug/kg	
319-86-8	delta-BHC	ND	1.4	0.25	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.4	0.34	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.4	0.42	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.4	0.30	ug/kg	
60-57-1	Dieldrin	ND	1.4	0.27	ug/kg	
72-54-8	4,4'-DDD	ND	1.4	0.22	ug/kg	
72-55-9	4,4'-DDE	ND	1.4	0.34	ug/kg	
50-29-3	4,4'-DDT	ND	1.4	0.32	ug/kg	
72-20-8	Endrin	ND	1.4	0.29	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.4	0.29	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.4	0.32	ug/kg	
959-98-8	Endosulfan-I	ND	1.4	0.28	ug/kg	
33213-65-9	Endosulfan-II	ND	1.4	0.45	ug/kg	
76-44-8	Heptachlor	ND	1.4	0.36	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.4	0.27	ug/kg	
72-43-5	Methoxychlor	ND	1.4	0.37	ug/kg	
53494-70-5	Endrin ketone	ND	1.4	0.27	ug/kg	
8001-35-2	Toxaphene	ND	17	13	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	79%		28-138%
877-09-8	Tetrachloro-m-xylene	79%		28-138%
2051-24-3	Decachlorobiphenyl	87%		22-156%
2051-24-3	Decachlorobiphenyl	93%		22-156%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	SW-CCI-009(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-10	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	94.6
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	1G46982.D	1	07/29/09	OPM	07/27/09	OP39078	G1G1720
Run #2							

Run #	Initial Weight	Final Volume
Run #1	17.1 g	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.2	0.25	ug/kg	
319-84-6	alpha-BHC	ND	1.2	0.23	ug/kg	
319-85-7	beta-BHC	ND	1.2	0.41	ug/kg	
319-86-8	delta-BHC	ND	1.2	0.23	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.2	0.30	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.2	0.38	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.2	0.28	ug/kg	
60-57-1	Dieldrin	ND	1.2	0.24	ug/kg	
72-54-8	4,4'-DDD	ND	1.2	0.20	ug/kg	
72-55-9	4,4'-DDE	ND	1.2	0.31	ug/kg	
50-29-3	4,4'-DDT	ND	1.2	0.29	ug/kg	
72-20-8	Endrin	ND	1.2	0.27	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.2	0.26	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.2	0.29	ug/kg	
959-98-8	Endosulfan-I	ND	1.2	0.25	ug/kg	
33213-65-9	Endosulfan-II	ND	1.2	0.41	ug/kg	
76-44-8	Heptachlor	ND	1.2	0.33	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.2	0.25	ug/kg	
72-43-5	Methoxychlor	ND	1.2	0.33	ug/kg	
53494-70-5	Endrin ketone	ND	1.2	0.25	ug/kg	
8001-35-2	Toxaphene	ND	15	12	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	62%		28-138%
877-09-8	Tetrachloro-m-xylene	60%		28-138%
2051-24-3	Decachlorobiphenyl	55%		22-156%
2051-24-3	Decachlorobiphenyl	69%		22-156%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 1

Client Sample ID:	SW-CCI-011(24.5)	Date Sampled:	08/10/09
Lab Sample ID:	JA25246-1	Date Received:	08/10/09
Matrix:	SO - Soil	Percent Solids:	83.6
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	3G39248.D	1	08/11/09	TDR	08/11/09	OP39322	G3G1469
Run #2							

Run #	Initial Weight	Final Volume
Run #1	17.2 g	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.4	0.28	ug/kg	
319-84-6	alpha-BHC	ND	1.4	0.26	ug/kg	
319-85-7	beta-BHC	ND	1.4	0.47	ug/kg	
319-86-8	delta-BHC	ND	1.4	0.25	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.4	0.34	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.4	0.43	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.4	0.31	ug/kg	
60-57-1	Dieldrin	ND	1.4	0.27	ug/kg	
72-54-8	4,4'-DDD	ND	1.4	0.23	ug/kg	
72-55-9	4,4'-DDE	ND	1.4	0.35	ug/kg	
50-29-3	4,4'-DDT	ND	1.4	0.32	ug/kg	
72-20-8	Endrin	ND	1.4	0.30	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.4	0.30	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.4	0.33	ug/kg	
959-98-8	Endosulfan-I	ND	1.4	0.28	ug/kg	
33213-65-9	Endosulfan-II	ND	1.4	0.46	ug/kg	
76-44-8	Heptachlor	ND	1.4	0.37	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.4	0.28	ug/kg	
72-43-5	Methoxychlor	ND	1.4	0.37	ug/kg	
53494-70-5	Endrin ketone	ND	1.4	0.28	ug/kg	
8001-35-2	Toxaphene	ND	17	13	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	83%		28-138%
877-09-8	Tetrachloro-m-xylene	77%		28-138%
2051-24-3	Decachlorobiphenyl	82%		22-156%
2051-24-3	Decachlorobiphenyl	106%		22-156%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 1

Client Sample ID:	SW-CCI-012	Date Sampled:	08/13/09
Lab Sample ID:	JA25522-1	Date Received:	08/13/09
Matrix:	SO - Soil	Percent Solids:	82.7
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	3G39385.D	1	08/17/09	OPM	08/15/09	OP39456	G3G1474
Run #2							

Run #	Initial Weight	Final Volume
Run #1	17.0 g	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.4	0.29	ug/kg	
319-84-6	alpha-BHC	ND	1.4	0.26	ug/kg	
319-85-7	beta-BHC	ND	1.4	0.48	ug/kg	
319-86-8	delta-BHC	ND	1.4	0.26	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.4	0.35	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.4	0.44	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.4	0.32	ug/kg	
60-57-1	Dieldrin	ND	1.4	0.28	ug/kg	
72-54-8	4,4'-DDD	ND	1.4	0.23	ug/kg	
72-55-9	4,4'-DDE	ND	1.4	0.35	ug/kg	
50-29-3	4,4'-DDT	3.0	1.4	0.33	ug/kg	J
72-20-8	Endrin	ND	1.4	0.31	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.4	0.30	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.4	0.33	ug/kg	
959-98-8	Endosulfan-I	ND	1.4	0.29	ug/kg	
33213-65-9	Endosulfan-II	ND	1.4	0.47	ug/kg	
76-44-8	Heptachlor	ND	1.4	0.38	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.4	0.28	ug/kg	
72-43-5	Methoxychlor	ND	1.4	0.38	ug/kg	
53494-70-5	Endrin ketone	ND	1.4	0.28	ug/kg	
8001-35-2	Toxaphene	ND	18	13	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	76%		28-138%
877-09-8	Tetrachloro-m-xylene	80%		28-138%
2051-24-3	Decachlorobiphenyl	95%		22-156%
2051-24-3	Decachlorobiphenyl	108%		22-156%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Client Sample ID:	SW-CCI-013	Date Sampled:	08/13/09
Lab Sample ID:	JA25522-2	Date Received:	08/13/09
Matrix:	SO - Soil	Percent Solids:	88.2
Method:	SW846 8081A SW846 3545	Project: Courtlandt Corners I, 364 East 161st Street, Bronx, NY	

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	3G39386.D	1	08/17/09	OPM	08/15/09	OP39456	G3G1474
Run #2							

Run #	Initial Weight	Final Volume
Run #1	17.2 g	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.3	0.27	ug/kg	
319-84-6	alpha-BHIC	ND	1.3	0.24	ug/kg	
319-85-7	beta-BHC	ND	1.3	0.44	ug/kg	
319-86-8	delta-BHC	ND	1.3	0.24	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.3	0.32	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.3	0.40	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.3	0.29	ug/kg	
60-57-1	Dieldrin	ND	1.3	0.26	ug/kg	
72-54-8	4,4'-DDD	ND	1.3	0.21	ug/kg	
72-55-9	4,4'-DDE	ND	1.3	0.33	ug/kg	
50-29-3	4,4'-DDT	4.9	1.3	0.31	ug/kg	
72-20-8	Endrin	ND	1.3	0.28	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.3	0.28	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.3	0.31	ug/kg	
959-98-8	Endosulfan-I	ND	1.3	0.27	ug/kg	
33213-65-9	Endosulfan-II	ND	1.3	0.44	ug/kg	
76-44-8	Heptachlor	ND	1.3	0.35	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.3	0.26	ug/kg	
72-43-5	Methoxychlor	ND	1.3	0.35	ug/kg	
53494-70-5	Endrin ketone	ND	1.3	0.26	ug/kg	
8001-35-2	Toxaphene	ND	16	12	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	77%		28-138%
877-09-8	Tetrachloro-m-xylene	83%		28-138%
2051-24-3	Decachlorobiphenyl	84%		22-156%
2051-24-3	Decachlorobiphenyl	95%		22-156%

ND = Not detected MDL - Method Detection Limit J = Indicates an estimated value
 RL = Reporting Limit B = Indicates analyte found in associated method blank
 E = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 1

Client Sample ID:	SW-CCI-014	Date Sampled:	08/13/09
Lab Sample ID:	JA25522-3	Date Received:	08/13/09
Matrix:	SO - Soil	Percent Solids:	85.6
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	3G39387.D	1	08/17/09	OPM	08/15/09	OP39456	G3G1474
Run #2							

Run #	Initial Weight	Final Volume
Run #1	17.3 g	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.4	0.27	ug/kg	
319-84-6	alpha-BHC	ND	1.4	0.25	ug/kg	
319-85-7	beta-BHC	ND	1.4	0.45	ug/kg	
319-86-8	delta-BHC	ND	1.4	0.25	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.4	0.33	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.4	0.41	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.4	0.30	ug/kg	
60-57-1	Dieldrin	ND	1.4	0.26	ug/kg	
72-54-8	4,4'-DDD	ND	1.4	0.22	ug/kg	
72-55-9	4,4'-DDE	ND	1.4	0.34	ug/kg	
50-29-3	4,4'-DDT	ND	1.4	0.31	ug/kg	
72-20-8	Endrin	ND	1.4	0.29	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.4	0.29	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.4	0.32	ug/kg	
959-98-8	Endosulfan-I	ND	1.4	0.28	ug/kg	
33213-65-9	Endosulfan-II	ND	1.4	0.45	ug/kg	
76-44-8	Heptachlor	ND	1.4	0.36	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.4	0.27	ug/kg	
72-43-5	Methoxychlor	ND	1.4	0.36	ug/kg	
53494-70-5	Endrin ketone	ND	1.4	0.27	ug/kg	
8001-35-2	Toxaphene	ND	17	13	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	77%		28-138%
877-09-8	Tetrachloro-m-xylene	77%		28-138%
2051-24-3	Decachlorobiphenyl	82%		22-156%
2051-24-3	Decachlorobiphenyl	96%		22-156%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest Laboratories

Report of Analysis

Page 1 of 1

Client Sample ID:	SW-CCI-015	Date Sampled:	08/14/09
Lab Sample ID:	JA25638-1	Date Received:	08/14/09
Matrix:	SO - Soil	Percent Solids:	81.8
Method:	SW846 8081A SW846 3545		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	3G39376.D	1	08/15/09	TDR	08/14/09	OP39343	G3G1473
Run #2							

Run #	Initial Weight	Final Volume
Run #1	17.2 g	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	1.4	0.29	ug/kg	
319-84-6	alpha-BHC	ND	1.4	0.26	ug/kg	
319-85-7	beta-BHC	ND	1.4	0.48	ug/kg	
319-86-8	delta-BHC	ND	1.4	0.26	ug/kg	
58-89-9	gamma-BHC (Lindane)	ND	1.4	0.35	ug/kg	
5103-71-9	alpha-Chlordane	ND	1.4	0.43	ug/kg	
5103-74-2	gamma-Chlordane	ND	1.4	0.32	ug/kg	
60-57-1	Dieldrin	ND	1.4	0.28	ug/kg	
72-54-8	4,4'-DDD	ND	1.4	0.23	ug/kg	
72-55-9	4,4'-DDE	ND	1.4	0.35	ug/kg	
50-29-3	4,4'-DDT	ND	1.4	0.33	ug/kg	
72-20-8	Endrin	ND	1.4	0.31	ug/kg	
1031-07-8	Endosulfan sulfate	ND	1.4	0.30	ug/kg	
7421-93-4	Endrin aldehyde	ND	1.4	0.33	ug/kg	
959-98-8	Endosulfan-I	ND	1.4	0.29	ug/kg	
33213-65-9	Endosulfan-II	ND	1.4	0.47	ug/kg	
76-44-8	Heptachlor	ND	1.4	0.38	ug/kg	
1024-57-3	Heptachlor epoxide	ND	1.4	0.28	ug/kg	
72-43-5	Methoxychlor	ND	1.4	0.38	ug/kg	
53494-70-5	Endrin ketone	ND	1.4	0.28	ug/kg	
8001-35-2	Toxaphene	ND	18	13	ug/kg	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	79%		28-138%
877-09-8	Tetrachloro-m-xylene	81%		28-138%
2051-24-3	Decachlorobiphenyl	93%		22-156%
2051-24-3	Decachlorobiphenyl	109%		22-156%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Accutest LabLink@514900 10:37 13-Aug-2009

Report of Analysis

Page 1 of 1

Client Sample ID:	FB072009	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-1	Date Received:	07/21/09
Matrix:	AQ - Field Blank Soil	Percent Solids:	n/a
Method:	SW846 8081A SW846 3510C		
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	3G38960.D	1	07/25/09	TDR	07/24/09	OP39062	G3G1458
Run #2							

Run #	Initial Volume	Final Volume
Run #1	960 ml	10.0 ml
Run #2		

Pesticide TCL List

CAS No.	Compound	Result	RL	MDL	Units	Q
309-00-2	Aldrin	ND	0.021	0.0018	ug/l	
319-84-6	alpha-BHC	ND	0.021	0.0010	ug/l	
319-85-7	beta-BHC	ND	0.021	0.0037	ug/l	
319-86-8	delta-BHC	ND	0.021	0.0036	ug/l	
58-89-9	gamma-BHC (Lindane)	ND	0.021	0.0012	ug/l	
5103-71-9	alpha-Chlordane	ND	0.021	0.0020	ug/l	
5103-74-2	gamma-Chlordane	ND	0.021	0.0025	ug/l	
60-57-1	Dieldrin	ND	0.021	0.0014	ug/l	
72-54-8	4,4'-DDD	ND	0.021	0.0048	ug/l	
72-55-9	4,4'-DDE	ND	0.021	0.0014	ug/l	
50-29-3	4,4'-DDT	ND	0.021	0.0036	ug/l	
72-20-8	Endrin	ND	0.021	0.0032	ug/l	
1031-07-8	Endosulfan sulfate	ND	0.021	0.0027	ug/l	
7421-93-4	Endrin aldehyde	ND	0.021	0.0039	ug/l	
53494-70-5	Endrin ketone	ND	0.021	0.0027	ug/l	
959-98-8	Endosulfan-I	ND	0.021	0.0018	ug/l	
33213-65-9	Endosulfan-II	ND	0.021	0.0036	ug/l	
76-44-8	Heptachlor	ND	0.021	0.0021	ug/l	
1024-57-3	Heptachlor epoxide	ND	0.021	0.0016	ug/l	
72-43-5	Methoxychlor	ND	0.021	0.0071	ug/l	
8001-35-2	Toxaphene	ND	0.26	0.26	ug/l	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
877-09-8	Tetrachloro-m-xylene	88%		30-137%
877-09-8	Tetrachloro-m-xylene	93%		30-137%
2051-24-3	Decachlorobiphenyl	46%		10-137%
2051-24-3	Decachlorobiphenyl	47%		10-137%

ND = Not detected MDL - Method Detection Limit
 RL = Reporting Limit
 E = Indicates value exceeds calibration range

J = Indicates an estimated value
 B = Indicates analyte found in associated method blank
 N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	B-CCI-02(08)	Date Sampled:	06/30/09
Lab Sample ID:	JA22090-1	Date Received:	06/30/09
Matrix:	SO - Soil	Percent Solids:	83.0
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	14400	23	mg/kg	1	07/01/09	07/01/09	ND SW846 6010B ²	SW846 3050B ³
Antimony	< 2.3	J 2.3	mg/kg	1	07/01/09	07/01/09	ND SW846 6010B ²	SW846 3050B ³
Arsenic	3.3	2.3	mg/kg	1	07/01/09	07/01/09	ND SW846 6010B ²	SW846 3050B ³
Barium	117	23	mg/kg	1	07/01/09	07/01/09	ND SW846 6010B ²	SW846 3050B ³
Beryllium	0.71	0.58	mg/kg	1	07/01/09	07/01/09	ND SW846 6010B ²	SW846 3050B ³
Cadmium	< 0.58	0.58	mg/kg	1	07/01/09	07/01/09	ND SW846 6010B ²	SW846 3050B ³
Calcium	7570	580	mg/kg	1	07/01/09	07/01/09	ND SW846 6010B ²	SW846 3050B ³
Chromium	21.9	1.2	mg/kg	1	07/01/09	07/01/09	ND SW846 6010B ²	SW846 3050B ³
Cobalt	10.3	5.8	mg/kg	1	07/01/09	07/01/09	ND SW846 6010B ²	SW846 3050B ³
Copper	27.4	2.9	mg/kg	1	07/01/09	07/01/09	ND SW846 6010B ²	SW846 3050B ³
Iron	19400	12	mg/kg	1	07/01/09	07/01/09	ND SW846 6010B ²	SW846 3050B ³
Lead	106	2.3	mg/kg	1	07/01/09	07/01/09	ND SW846 6010B ²	SW846 3050B ³
Magnesium	7960	580	mg/kg	1	07/01/09	07/01/09	ND SW846 6010B ²	SW846 3050B ³
Manganese	453	1.7	mg/kg	1	07/01/09	07/01/09	ND SW846 6010B ²	SW846 3050B ³
Mercury	0.17	0.036	mg/kg	1	07/01/09	07/01/09	JW SW846 7471A ¹	SW846 7471A ⁴
Nickel	17.2	4.6	mg/kg	1	07/01/09	07/01/09	ND SW846 6010B ²	SW846 3050B ³
Potassium	1830	J 1200	mg/kg	1	07/01/09	07/01/09	ND SW846 6010B ²	SW846 3050B ³
Selenium	< 2.3	2.3	mg/kg	1	07/01/09	07/01/09	ND SW846 6010B ²	SW846 3050B ³
Silver	< 1.2	1.2	mg/kg	1	07/01/09	07/01/09	ND SW846 6010B ²	SW846 3050B ³
Sodium	< 1200	1200	mg/kg	1	07/01/09	07/01/09	ND SW846 6010B ²	SW846 3050B ³
Thallium	< 1.2	1.2	mg/kg	1	07/01/09	07/01/09	ND SW846 6010B ²	SW846 3050B ³
Vanadium	30.0	5.8	mg/kg	1	07/01/09	07/01/09	ND SW846 6010B ²	SW846 3050B ³
Zinc	143	2.3	mg/kg	1	07/01/09	07/01/09	ND SW846 6010B ²	SW846 3050B ³

- (1) Instrument QC Batch: MA22767
(2) Instrument QC Batch: MA22771
(3) Prep QC Batch: MP48700
(4) Prep QC Batch: MP48712

RI. = Reporting Limit

Report of Analysis

3.1
3

Client Sample ID: B-CCI-003 (20')	Date Sampled: 07/15/09
Lab Sample ID: JA23207-1	Date Received: 07/15/09
Matrix: SO - Soil	Percent Solids: 91.6
Project: Courtlandt Corners I, 364 East 161st Street, Bronx, NY	

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	6160	21	mg/kg	1	07/16/09	07/16/09 GT	SW846 6010B ²	SW846 3050B ⁶
Antimony	< 2.1	J 2.1	mg/kg	1	07/16/09	07/16/09 GT	SW846 6010B ²	SW846 3050B ⁶
Arsenic	< 2.1	2.1	mg/kg	1	07/16/09	07/16/09 GT	SW846 6010B ²	SW846 3050B ⁶
Barium	32.2	21	mg/kg	1	07/16/09	07/16/09 GT	SW846 6010B ²	SW846 3050B ⁶
Beryllium	0.56	J 0.52	mg/kg	1	07/16/09	07/17/09 ND	SW846 6010B ³	SW846 3050B ⁶
Cadmium	< 0.52	0.52	mg/kg	1	07/16/09	07/16/09 GT	SW846 6010B ²	SW846 3050B ⁶
Calcium	90700	2600	mg/kg	5	07/16/09	07/17/09 ND	SW846 6010B ⁴	SW846 3050B ⁶
Chromium	11.2	1.0	mg/kg	1	07/16/09	07/17/09 ND	SW846 6010B ³	SW846 3050B ⁶
Cobalt	< 5.2	5.2	mg/kg	1	07/16/09	07/16/09 GT	SW846 6010B ²	SW846 3050B ⁶
Copper	10.1	2.6	mg/kg	1	07/16/09	07/16/09 GT	SW846 6010B ²	SW846 3050B ⁶
Iron	10800	J 15	mg/kg	1	07/16/09	07/16/09 GT	SW846 6010B ²	SW846 3050B ⁶
Lead	4.9	2.1	mg/kg	1	07/16/09	07/16/09 GT	SW846 6010B ²	SW846 3050B ⁶
Magnesium	51600	520	mg/kg	1	07/16/09	07/16/09 GT	SW846 6010B ²	SW846 3050B ⁶
Manganese	278	J 1.6	mg/kg	1	07/16/09	07/16/09 GT	SW846 6010B ²	SW846 3050B ⁶
Mercury	< 0.036	0.036	mg/kg	1	07/15/09	07/16/09 TG	SW846 7471A ¹	SW846 7471A ⁵
Nickel	9.6	J 4.2	mg/kg	1	07/16/09	07/16/09 GT	SW846 6010B ²	SW846 3050B ⁶
Potassium	1390	1000	mg/kg	1	07/16/09	07/16/09 GT	SW846 6010B ²	SW846 3050B ⁶
Selenium	< 2.1	2.1	mg/kg	1	07/16/09	07/16/09 GT	SW846 6010B ²	SW846 3050B ⁶
Silver	< 1.0	1.0	mg/kg	1	07/16/09	07/16/09 GT	SW846 6010B ²	SW846 3050B ⁶
Sodium	< 1000	1000	mg/kg	1	07/16/09	07/16/09 GT	SW846 6010B ²	SW846 3050B ⁶
Thallium	< 1.0	1.0	mg/kg	1	07/16/09	07/16/09 GT	SW846 6010B ²	SW846 3050B ⁶
Vanadium	15.5	5.2	mg/kg	1	07/16/09	07/16/09 GT	SW846 6010B ²	SW846 3050B ⁶
Zinc	33.8	J 2.1	mg/kg	1	07/16/09	07/16/09 GT	SW846 6010B ²	SW846 3050B ⁶

- (1) Instrument QC Batch: MA22829
- (2) Instrument QC Batch: MA22831
- (3) Instrument QC Batch: MA22834
- (4) Instrument QC Batch: MA22837
- (5) Prep QC Batch: MP48764
- (6) Prep QC Batch: MP48880

RL = Reporting Limit

Report of Analysis

Client Sample ID:	B-CCI-004 (20')	Date Sampled:	07/15/09
Lab Sample ID:	JA23207-2	Date Received:	07/15/09
Matrix:	SO - Soil	Percent Solids:	91.6
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum ^a	1680	210	mg/kg	10	07/16/09	07/17/09	ND	SW846 6010B ³
Antimony	< 2.1	J 2.1	mg/kg	1	07/16/09	07/16/09	GT	SW846 6010B ²
Arsenic	< 2.1	2.1	mg/kg	1	07/16/09	07/16/09	GT	SW846 6010B ²
Barium	< 21	21	mg/kg	1	07/16/09	07/16/09	GT	SW846 6010B ²
Beryllium	< 0.52	J 0.52	mg/kg	1	07/16/09	07/17/09	ND	SW846 6010B ³
Cadmium	< 0.52	0.52	mg/kg	1	07/16/09	07/16/09	GT	SW846 6010B ²
Calcium	273000	5200	mg/kg	10	07/16/09	07/17/09	ND	SW846 6010B ³
Chromium	2.6	1.0	mg/kg	1	07/16/09	07/17/09	ND	SW846 6010B ³
Cobalt	< 5.2	5.2	mg/kg	1	07/16/09	07/16/09	GT	SW846 6010B ²
Copper	< 2.6	2.6	mg/kg	1	07/16/09	07/16/09	GT	SW846 6010B ²
Iron	4230	J 15	mg/kg	1	07/16/09	07/16/09	GT	SW846 6010B ²
Lead	< 2.1	2.1	mg/kg	1	07/16/09	07/16/09	GT	SW846 6010B ²
Magnesium	173000	2600	mg/kg	5	07/16/09	07/17/09	ND	SW846 6010B ³
Manganese	192	J 1.6	mg/kg	1	07/16/09	07/16/09	GT	SW846 6010B ²
Mercury	< 0.034	0.034	mg/kg	1	07/15/09	07/16/09	TG	SW846 7471A ¹
Nickel	< 4.2	J 4.2	mg/kg	1	07/16/09	07/16/09	GT	SW846 6010B ²
Potassium	< 1000	1000	mg/kg	1	07/16/09	07/16/09	GT	SW846 6010B ²
Selenium	< 2.1	2.1	mg/kg	1	07/16/09	07/16/09	GT	SW846 6010B ²
Silver	< 1.0	1.0	mg/kg	1	07/16/09	07/16/09	GT	SW846 6010B ²
Sodium	< 1000	1000	mg/kg	1	07/16/09	07/16/09	GT	SW846 6010B ²
Thallium	< 1.0	1.0	mg/kg	1	07/16/09	07/16/09	GT	SW846 6010B ²
Vanadium	< 5.2	5.2	mg/kg	1	07/16/09	07/16/09	GT	SW846 6010B ²
Zinc	8.2	J 2.1	mg/kg	1	07/16/09	07/16/09	GT	SW846 6010B ²

- (1) Instrument QC Batch: MA22829
 (2) Instrument QC Batch: MA22831
 (3) Instrument QC Batch: MA22834
 (4) Prep QC Batch: MP48764
 (5) Prep QC Batch: MP48880

^a Elevated detection limit due to dilution required for high interfering element.

RL = Reporting Limit

Report of Analysis

Client Sample ID:	B-CCI-006 (15')	Date Sampled:	07/22/09
Lab Sample ID:	JA23843-1	Date Received:	07/23/09
Matrix:	SO - Soil	Percent Solids:	90.6
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	5090	22	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ²
Antimony	< 2.2	J 2.2	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ²
Arsenic	< 2.2	2.2	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ²
Barium	36.1	22	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ²
Beryllium	< 0.55	0.55	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ²
Cadmium	< 0.55	0.55	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ²
Calcium	1980	J 550	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ²
Chromium	26.1	1.1	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ²
Cobalt	9.9	5.5	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ²
Copper	11.9	2.8	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ²
Iron	11300	11	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ²
Lead	< 2.2	2.2	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ²
Magnesium	2600	550	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ²
Manganese	263	1.7	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ²
Mercury	< 0.033	0.033	mg/kg	1	07/24/09	07/24/09	JW	SW846 7471A ¹
Nickel	12.4	4.4	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ²
Potassium	1400	1100	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ²
Selenium	< 2.2	2.2	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ²
Silver	< 1.1	1.1	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ²
Sodium	< 1100	1100	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ²
Thallium	< 1.1	1.1	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ²
Vanadium	21.0	5.5	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ²
Zinc	24.2	2.2	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ²

- (1) Instrument QC Batch: MA22865
- (2) Instrument QC Batch: MA22869
- (3) Prep QC Batch: MP48970
- (4) Prep QC Batch: MP48971

RL = Reporting Limit

Report of Analysis

Client Sample ID:	B-CCI-007 (20')	Date Sampled:	07/22/09
Lab Sample ID:	JA23843-2	Date Received:	07/23/09
Matrix:	SO - Soil	Percent Solids:	94.5
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	6850	22	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ² SW846 3050B ⁴
Antimony	< 2.2	J 2.2	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ² SW846 3050B ⁴
Arsenic	< 2.2	2.2	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ² SW846 3050B ⁴
Barium	36.7	22	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ² SW846 3050B ⁴
Beryllium	< 0.56	0.56	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ² SW846 3050B ⁴
Cadmium	< 0.56	0.56	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ² SW846 3050B ⁴
Calcium	3600	J 560	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ² SW846 3050B ⁴
Chromium	27.2	1.1	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ² SW846 3050B ⁴
Cobalt	7.1	5.6	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ² SW846 3050B ⁴
Copper	17.3	2.8	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ² SW846 3050B ⁴
Iron	13500	11	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ² SW846 3050B ⁴
Lead	6.5	2.2	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ² SW846 3050B ⁴
Magnesium	5150	560	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ² SW846 3050B ⁴
Manganese	176	1.7	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ² SW846 3050B ⁴
Mercury	< 0.031	0.031	mg/kg	1	07/24/09	07/24/09	JW	SW846 7471A ¹ SW846 7471A ³
Nickel	17.4	4.5	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ² SW846 3050B ⁴
Potassium	1930	1100	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ² SW846 3050B ⁴
Selenium	< 2.2	2.2	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ² SW846 3050B ⁴
Silver	< 1.1	1.1	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ² SW846 3050B ⁴
Sodium	< 1100	1100	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ² SW846 3050B ⁴
Thallium	< 1.1	1.1	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ² SW846 3050B ⁴
Vanadium	24.4	5.6	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ² SW846 3050B ⁴
Zinc	36.0	2.2	mg/kg	1	07/24/09	07/24/09	ND	SW846 6010B ² SW846 3050B ⁴

(1) Instrument QC Batch: MA22865

(2) Instrument QC Batch: MA22869

(3) Prep QC Batch: MP48970

(4) Prep QC Batch: MP48971

RL = Reporting Limit

Report of Analysis

3.5
3

Client Sample ID:	DUPLICATE	Date Sampled:	07/22/09
Lab Sample ID:	JA23843-5	Date Received:	07/23/09
Matrix:	SO - Soil	Percent Solids:	94.2
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	6410	20	mg/kg	1	07/24/09	07/24/09 ND	SW846 6010B ²	SW846 3050B ⁴
Antimony	< 2.0	J 2.0	mg/kg	1	07/24/09	07/24/09 ND	SW846 6010B ²	SW846 3050B ⁴
Arsenic	< 2.0	2.0	mg/kg	1	07/24/09	07/24/09 ND	SW846 6010B ²	SW846 3050B ⁴
Barium	35.1	20	mg/kg	1	07/24/09	07/24/09 ND	SW846 6010B ²	SW846 3050B ⁴
Beryllium	< 0.51	0.51	mg/kg	1	07/24/09	07/24/09 ND	SW846 6010B ²	SW846 3050B ⁴
Cadmium	< 0.51	0.51	mg/kg	1	07/24/09	07/24/09 ND	SW846 6010B ²	SW846 3050B ⁴
Calcium	5050	J 510	mg/kg	1	07/24/09	07/24/09 ND	SW846 6010B ²	SW846 3050B ⁴
Chromium	24.8	1.0	mg/kg	1	07/24/09	07/24/09 ND	SW846 6010B ²	SW846 3050B ⁴
Cobalt	6.5	5.1	mg/kg	1	07/24/09	07/24/09 ND	SW846 6010B ²	SW846 3050B ⁴
Copper	15.7	2.5	mg/kg	1	07/24/09	07/24/09 ND	SW846 6010B ²	SW846 3050B ⁴
Iron	12500	10	mg/kg	1	07/24/09	07/24/09 ND	SW846 6010B ²	SW846 3050B ⁴
Lead	7.2	2.0	mg/kg	1	07/24/09	07/24/09 ND	SW846 6010B ²	SW846 3050B ⁴
Magnesium	5750	510	mg/kg	1	07/24/09	07/24/09 ND	SW846 6010B ²	SW846 3050B ⁴
Manganese	171	1.5	mg/kg	1	07/24/09	07/24/09 ND	SW846 6010B ²	SW846 3050B ⁴
Mercury	< 0.034	0.034	mg/kg	1	07/24/09	07/24/09 JW	SW846 7471A ¹	SW846 7471A ³
Nickel	16.1	4.0	mg/kg	1	07/24/09	07/24/09 ND	SW846 6010B ²	SW846 3050B ⁴
Potassium	1820	1000	mg/kg	1	07/24/09	07/24/09 ND	SW846 6010B ²	SW846 3050B ⁴
Selenium	< 2.0	2.0	mg/kg	1	07/24/09	07/24/09 ND	SW846 6010B ²	SW846 3050B ⁴
Silver	< 1.0	1.0	mg/kg	1	07/24/09	07/24/09 ND	SW846 6010B ²	SW846 3050B ⁴
Sodium	< 1000	1000	mg/kg	1	07/24/09	07/24/09 ND	SW846 6010B ²	SW846 3050B ⁴
Thallium	< 1.0	1.0	mg/kg	1	07/24/09	07/24/09 ND	SW846 6010B ²	SW846 3050B ⁴
Vanadium	23.3	5.1	mg/kg	1	07/24/09	07/24/09 ND	SW846 6010B ²	SW846 3050B ⁴
Zinc	34.2	2.0	mg/kg	1	07/24/09	07/24/09 ND	SW846 6010B ²	SW846 3050B ⁴

- (1) Instrument QC Batch: MA22865
- (2) Instrument QC Batch: MA22869
- (3) Prep QC Batch: MP48970
- (4) Prep QC Batch: MP48971

RL = Reporting Limit

Report of Analysis

Client Sample ID:	B-CCI-010(22')	Date Sampled:	07/24/09
Lab Sample ID:	JA24059-1	Date Received:	07/27/09
Matrix:	SO - Soil	Percent Solids:	82.1
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	3410	J 23	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Antimony ^a	< 12	J 12	mg/kg	5	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Arsenic	< 2.3	J 2.3	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Barium	< 23	J 23	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Beryllium	< 0.59	J 0.59	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Cadmium	< 0.59	J 0.59	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Calcium	159000	J 2900	mg/kg	5	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Chromium	6.7	J 1.2	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Cobalt	< 5.9	J 5.9	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Copper	9.8	J 2.9	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Iron	6280	J 12	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Lead ^a	< 12	J 12	mg/kg	5	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Magnesium	98500	J 590	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Manganese	224	J 1.8	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Mercury	< 0.038	J 0.038	mg/kg	1	07/28/09	07/28/09	JW SW846 7471A ¹	SW846 7471A ³
Nickel	6.1	J 4.7	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Potassium	< 1200	J 1200	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Selenium ^a	< 12	J 12	mg/kg	5	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Silver	< 1.2	J 1.2	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Sodium	< 1200	J 1200	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Thallium ^a	< 5.9	J 5.9	mg/kg	5	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Vanadium	10	J 5.9	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Zinc	13.2	J 2.3	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴

(1) Instrument QC Batch: MA22876

(2) Instrument QC Batch: MA22881

(3) Prep QC Batch: MP49006

(4) Prep QC Batch: MP49014

(a) Elevated detection limit due to dilution required for high interfering element.

RL = Reporting Limit

Report of Analysis

Client Sample ID:	B-CCI-013(10')	Date Sampled:	07/27/09
Lab Sample ID:	JA24059-3	Date Received:	07/27/09
Matrix:	SO - Soil	Percent Solids:	82.9
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	14400	J 23	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Antimony	< 2.3	J 2.3	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Arsenic	6.2	J 2.3	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Barium	192	J 23	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Beryllium	0.71	J 0.59	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Cadmium	0.87	J 0.59	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Calcium	4320	J 590	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Chromium	26.1	J 1.2	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Cobalt	11.5	J 5.9	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Copper	43.8	J 2.9	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Iron	21900	J 12	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Lead	443	J 2.3	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Magnesium	3660	J 590	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Manganese	611	J 1.8	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Mercury	0.31	J 0.038	mg/kg	1	07/28/09	07/28/09	JW SW846 7471A ¹	SW846 7471A ³
Nickel	21.5	J 4.7	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Potassium	1740	J 1200	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Selenium	< 2.3	J 2.3	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Silver	< 1.2	J 1.2	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Sodium	< 1200	J 1200	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Thallium	< 1.2	J 1.2	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Vanadium	42.5	J 5.9	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴
Zinc	339	J 2.3	mg/kg	1	07/28/09	07/28/09	ND SW846 6010B ²	SW846 3050B ⁴

(1) Instrument QC Batch: MA22876

(2) Instrument QC Batch: MA22881

(3) Prep QC Batch: MP49006

(4) Prep QC Batch: MP49014

RL = Reporting Limit

Report of Analysis

Client Sample ID: B-CCI-022	Date Sampled: 08/04/09
Lab Sample ID: JA24692-1	Date Received: 08/04/09
Matrix: SO - Soil	Percent Solids: 87.0
Project: Courtlandt Corners I, 364 East 161st Street, Bronx, NY	

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	12500	23	mg/kg	1	08/05/09	08/05/09	ND SW846 6010B ²	SW846 3050B ⁴
Antimony	< 2.3	J 2.3	mg/kg	1	08/05/09	08/05/09	ND SW846 6010B ²	SW846 3050B ⁴
Arsenic	3.4	J 2.3	mg/kg	1	08/05/09	08/05/09	ND SW846 6010B ²	SW846 3050B ⁴
Barium	146	23	mg/kg	1	08/05/09	08/05/09	ND SW846 6010B ²	SW846 3050B ⁴
Beryllium	< 0.57	J 0.57	mg/kg	1	08/05/09	08/05/09	ND SW846 6010B ²	SW846 3050B ⁴
Cadmium	< 0.57	J 0.57	mg/kg	1	08/05/09	08/05/09	ND SW846 6010B ²	SW846 3050B ⁴
Calcium	24000	J 570	mg/kg	1	08/05/09	08/05/09	ND SW846 6010B ²	SW846 3050B ⁴
Chromium	25.8	J 1.1	mg/kg	1	08/05/09	08/05/09	ND SW846 6010B ²	SW846 3050B ⁴
Cobalt	10.6	J 5.7	mg/kg	1	08/05/09	08/05/09	ND SW846 6010B ²	SW846 3050B ⁴
Copper	28.1	J 2.9	mg/kg	1	08/05/09	08/05/09	ND SW846 6010B ²	SW846 3050B ⁴
Iron	19400	11	mg/kg	1	08/05/09	08/05/09	ND SW846 6010B ²	SW846 3050B ⁴
Lead	169	2.3	mg/kg	1	08/05/09	08/05/09	ND SW846 6010B ²	SW846 3050B ⁴
Magnesium	16700	570	mg/kg	1	08/05/09	08/05/09	ND SW846 6010B ²	SW846 3050B ⁴
Manganese	443	1.7	mg/kg	1	08/05/09	08/05/09	ND SW846 6010B ²	SW846 3050B ⁴
Mercury	0.28	J 0.035	mg/kg	1	08/05/09	08/05/09	JW SW846 7471A ¹	SW846 7471A ³
Nickel	18.9	J 4.6	mg/kg	1	08/05/09	08/05/09	ND SW846 6010B ²	SW846 3050B ⁴
Potassium	2880	1100	mg/kg	1	08/05/09	08/05/09	ND SW846 6010B ²	SW846 3050B ⁴
Selenium	< 2.3	J 2.3	mg/kg	1	08/05/09	08/05/09	ND SW846 6010B ²	SW846 3050B ⁴
Silver	< 1.1	1.1	mg/kg	1	08/05/09	08/05/09	ND SW846 6010B ²	SW846 3050B ⁴
Sodium	< 1100	1100	mg/kg	1	08/05/09	08/05/09	ND SW846 6010B ²	SW846 3050B ⁴
Thallium	< 1.1	1.1	mg/kg	1	08/05/09	08/05/09	ND SW846 6010B ²	SW846 3050B ⁴
Vanadium	34.2	5.7	mg/kg	1	08/05/09	08/05/09	ND SW846 6010B ²	SW846 3050B ⁴
Zinc	99.4	2.3	mg/kg	1	08/05/09	08/05/09	ND SW846 6010B ²	SW846 3050B ⁴

- (1) Instrument QC Batch: MA22922
(2) Instrument QC Batch: MA22923
(3) Prep QC Batch: MP49100
(4) Prep QC Batch: MP49104

RL = Reporting Limit

Report of Analysis

3.2
3

Client Sample ID: B-CCI-023	Date Sampled: 08/04/09
Lab Sample ID: JA24692-2	Date Received: 08/04/09
Matrix: SO - Soil	Percent Solids: 84.1
Project: Courtlandt Corners I, 364 East 161st Street, Bronx, NY	

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	14600	23	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Antimony	< 2.3	J 2.3	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Arsenic	< 2.3	2.3	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Barium	75.1	23	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Beryllium	< 0.57	J 0.57	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Cadmium	< 0.57	J 0.57	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Calcium	1940	J 570	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Chromium	24.3	1.1	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Cobalt	8.8	J 5.7	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Copper	15.6	2.9	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Iron	17400	11	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Lead	23.8	2.3	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Magnesium	4270	570	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Manganese	405	1.7	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Mercury	0.047	J 0.035	mg/kg	1	08/05/09	08/05/09 JW	SW846 7471A ¹	SW846 7471A ³
Nickel	16.4	J 4.6	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Potassium	1240	1100	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Selenium	< 2.3	2.3	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Silver	< 1.1	1.1	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Sodium	< 1100	1100	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Thallium	< 1.1	1.1	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Vanadium	30.7	5.7	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Zinc	45.9	2.3	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴

- (1) Instrument QC Batch: MA22922
- (2) Instrument QC Batch: MA22923
- (3) Prep QC Batch: MP49100
- (4) Prep QC Batch: MP49104

RL = Reporting Limit

Report of Analysis

3.3
3

Client Sample ID: B-CCI-024	Date Sampled: 08/04/09
Lab Sample ID: JA24692-3	Date Received: 08/04/09
Matrix: SO - Soil	Percent Solids: 89.0
Project: Courtlandt Corners I, 364 East 161st Street, Bronx, NY	

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	6590	22	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Antimony	< 2.2	J 2.2	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Arsenic	< 2.2	2.2	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Barium	30.6	22	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Beryllium	< 0.55	J 0.55	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Cadmium	< 0.55	0.55	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Calcium	12900	J 550	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Chromium	18.8	1.1	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Cobalt	6.3	J 5.5	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Copper	14.1	2.7	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Iron	12200	11	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Lead	5.0	2.2	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Magnesium	9350	550	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Manganese	302	1.6	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Mercury	< 0.037	0.037	mg/kg	1	08/05/09	08/05/09 JW	SW846 7471A ¹	SW846 7471A ³
Nickel	13.4	J 4.4	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Potassium	1710	1100	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Selenium	< 2.2	2.2	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Silver	< 1.1	1.1	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Sodium	< 1100	1100	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Thallium	< 1.1	1.1	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Vanadium	22.3	5.5	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴
Zinc	28.0	2.2	mg/kg	1	08/05/09	08/05/09 ND	SW846 6010B ²	SW846 3050B ⁴

- (1) Instrument QC Batch: MA22922
- (2) Instrument QC Batch: MA22923
- (3) Prep QC Batch: MP49100
- (4) Prep QC Batch: MP49104

RL = Reporting Limit

Report of Analysis

3.4
3

Client Sample ID:	B-CCI-025	Date Sampled:	08/04/09
Lab Sample ID:	JA24692-4	Date Received:	08/04/09
Matrix:	SO - Soil	Percent Solids:	94.7
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	4610	21	mg/kg	1	08/05/09	08/05/09	ND	SW846 6010B ²
Antimony	< 2.1	J 2.1	mg/kg	1	08/05/09	08/05/09	ND	SW846 6010B ²
Arsenic	< 2.1	2.1	mg/kg	1	08/05/09	08/05/09	ND	SW846 6010B ²
Barium	21.0	21	mg/kg	1	08/05/09	08/05/09	ND	SW846 6010B ²
Beryllium	< 0.53	J 0.53	mg/kg	1	08/05/09	08/05/09	ND	SW846 6010B ²
Cadmium	< 0.53	0.53	mg/kg	1	08/05/09	08/05/09	ND	SW846 6010B ²
Calcium	14800	J 530	mg/kg	1	08/05/09	08/05/09	ND	SW846 6010B ²
Chromium	15.3	1.1	mg/kg	1	08/05/09	08/05/09	ND	SW846 6010B ²
Cobalt	< 5.3	J 5.3	mg/kg	1	08/05/09	08/05/09	ND	SW846 6010B ²
Copper	10.4	2.7	mg/kg	1	08/05/09	08/05/09	ND	SW846 6010B ²
Iron	9130	11	mg/kg	1	08/05/09	08/05/09	ND	SW846 6010B ²
Lead	10.7	2.1	mg/kg	1	08/05/09	08/05/09	ND	SW846 6010B ²
Magnesium	9880	530	mg/kg	1	08/05/09	08/05/09	ND	SW846 6010B ²
Manganese	201	1.6	mg/kg	1	08/05/09	08/05/09	ND	SW846 6010B ²
Mercury	< 0.031	0.031	mg/kg	1	08/05/09	08/05/09	JW	SW846 7471A ¹
Nickel	10.8	J 4.3	mg/kg	1	08/05/09	08/05/09	ND	SW846 6010B ²
Potassium	1180	1100	mg/kg	1	08/05/09	08/05/09	ND	SW846 6010B ²
Selenium	< 2.1	2.1	mg/kg	1	08/05/09	08/05/09	ND	SW846 6010B ²
Silver	< 1.1	1.1	mg/kg	1	08/05/09	08/05/09	ND	SW846 6010B ²
Sodium	< 1100	1100	mg/kg	1	08/05/09	08/05/09	ND	SW846 6010B ²
Thallium	< 1.1	1.1	mg/kg	1	08/05/09	08/05/09	ND	SW846 6010B ²
Vanadium	15.7	5.3	mg/kg	1	08/05/09	08/05/09	ND	SW846 6010B ²
Zinc	20.3	2.1	mg/kg	1	08/05/09	08/05/09	ND	SW846 6010B ²

- (1) Instrument QC Batch: MA22922
- (2) Instrument QC Batch: MA22923
- (3) Prep QC Batch: MP49100
- (4) Prep QC Batch: MP49104

RL = Reporting Limit

Report of Analysis

Client Sample ID:	B-CCI-026 (2')	Date Sampled:	08/06/09
Lab Sample ID:	JA25087-1	Date Received:	08/07/09
Matrix:	SO - Soil	Percent Solids:	83.0
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	10000	J 23	mg/kg	1	08/10/09	08/10/09 GT	SW846 6010B ²	SW846 3050B ⁴
Antimony	< 2.3	J 2.3	mg/kg	1	08/10/09	08/10/09 GT	SW846 6010B ²	SW846 3050B ⁴
Arsenic	5.1	2.3	mg/kg	1	08/10/09	08/10/09 GT	SW846 6010B ²	SW846 3050B ⁴
Barium	234	23	mg/kg	1	08/10/09	08/10/09 GT	SW846 6010B ²	SW846 3050B ⁴
Beryllium	< 0.58	0.58	mg/kg	1	08/10/09	08/10/09 GT	SW846 6010B ²	SW846 3050B ⁴
Cadmium	0.63	0.58	mg/kg	1	08/10/09	08/10/09 GT	SW846 6010B ²	SW846 3050B ⁴
Calcium	24500	580	mg/kg	1	08/10/09	08/10/09 GT	SW846 6010B ²	SW846 3050B ⁴
Chromium	20.5	J 1.2	mg/kg	1	08/10/09	08/10/09 GT	SW846 6010B ²	SW846 3050B ⁴
Cobalt	9.0	5.8	mg/kg	1	08/10/09	08/10/09 GT	SW846 6010B ²	SW846 3050B ⁴
Copper	52.1	2.9	mg/kg	1	08/10/09	08/10/09 GT	SW846 6010B ²	SW846 3050B ⁴
Iron	18200	12	mg/kg	1	08/10/09	08/10/09 GT	SW846 6010B ²	SW846 3050B ⁴
Lead	349	2.3	mg/kg	1	08/10/09	08/10/09 GT	SW846 6010B ²	SW846 3050B ⁴
Magnesium	9350	580	mg/kg	1	08/10/09	08/10/09 GT	SW846 6010B ²	SW846 3050B ⁴
Manganese	300	1.7	mg/kg	1	08/10/09	08/10/09 GT	SW846 6010B ²	SW846 3050B ⁴
Mercury	0.46	0.036	mg/kg	1	08/10/09	08/10/09 JW	SW846 7471A ¹	SW846 7471A ³
Nickel	18.2	J 4.6	mg/kg	1	08/10/09	08/10/09 GT	SW846 6010B ²	SW846 3050B ⁴
Potassium	2000	J 1200	mg/kg	1	08/10/09	08/10/09 GT	SW846 6010B ²	SW846 3050B ⁴
Selenium	< 2.3	2.3	mg/kg	1	08/10/09	08/10/09 GT	SW846 6010B ²	SW846 3050B ⁴
Silver	< 1.2	1.2	mg/kg	1	08/10/09	08/10/09 GT	SW846 6010B ²	SW846 3050B ⁴
Sodium	< 1200	1200	mg/kg	1	08/10/09	08/10/09 GT	SW846 6010B ²	SW846 3050B ⁴
Thallium	< 1.2	1.2	mg/kg	1	08/10/09	08/10/09 GT	SW846 6010B ²	SW846 3050B ⁴
Vanadium	29.4	J 5.8	mg/kg	1	08/10/09	08/10/09 GT	SW846 6010B ²	SW846 3050B ⁴
Zinc	270	J 2.3	mg/kg	1	08/10/09	08/10/09 GT	SW846 6010B ²	SW846 3050B ⁴

- (1) Instrument QC Batch: MA22941
- (2) Instrument QC Batch: MA22944
- (3) Prep QC Batch: MP49159
- (4) Prep QC Batch: MP49161

RL = Reporting Limit

Report of Analysis

Client Sample ID:	B-CCI-027 (2')	Date Sampled:	08/06/09
Lab Sample ID:	JA25087-2	Date Received:	08/07/09
Matrix:	SO - Soil	Percent Solids:	88.1
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	10400	J 22	mg/kg	1	08/10/09	08/10/09	GT SW846 6010B ²	SW846 3050B ⁴
Antimony	< 2.2	J 2.2	mg/kg	1	08/10/09	08/10/09	GT SW846 6010B ²	SW846 3050B ⁴
Arsenic	4.3	2.2	mg/kg	1	08/10/09	08/10/09	GT SW846 6010B ²	SW846 3050B ⁴
Barium	221	22	mg/kg	1	08/10/09	08/10/09	GT SW846 6010B ²	SW846 3050B ⁴
Beryllium	< 0.54	0.54	mg/kg	1	08/10/09	08/10/09	GT SW846 6010B ²	SW846 3050B ⁴
Cadmium	1.6	0.54	mg/kg	1	08/10/09	08/10/09	GT SW846 6010B ²	SW846 3050B ⁴
Calcium	6000	540	mg/kg	1	08/10/09	08/10/09	GT SW846 6010B ²	SW846 3050B ⁴
Chromium	21.7	J 1.1	mg/kg	1	08/10/09	08/10/09	GT SW846 6010B ²	SW846 3050B ⁴
Cobalt	7.5	5.4	mg/kg	1	08/10/09	08/10/09	GT SW846 6010B ²	SW846 3050B ⁴
Copper	34.7	2.7	mg/kg	1	08/10/09	08/10/09	GT SW846 6010B ²	SW846 3050B ⁴
Iron	16000	11	mg/kg	1	08/10/09	08/10/09	GT SW846 6010B ²	SW846 3050B ⁴
Lead	276	2.2	mg/kg	1	08/10/09	08/10/09	GT SW846 6010B ²	SW846 3050B ⁴
Magnesium	3240	540	mg/kg	1	08/10/09	08/10/09	GT SW846 6010B ²	SW846 3050B ⁴
Manganese	290	1.6	mg/kg	1	08/10/09	08/10/09	GT SW846 6010B ²	SW846 3050B ⁴
Mercury	0.51	0.036	mg/kg	1	08/10/09	08/10/09	JW SW846 7471A ¹	SW846 7471A ³
Nickel	25.5	J 4.3	mg/kg	1	08/10/09	08/10/09	GT SW846 6010B ²	SW846 3050B ⁴
Potassium	1660	J 1100	mg/kg	1	08/10/09	08/10/09	GT SW846 6010B ²	SW846 3050B ⁴
Selenium	< 2.2	2.2	mg/kg	1	08/10/09	08/10/09	GT SW846 6010B ²	SW846 3050B ⁴
Silver	< 1.1	1.1	mg/kg	1	08/10/09	08/10/09	GT SW846 6010B ²	SW846 3050B ⁴
Sodium	< 1100	1100	mg/kg	1	08/10/09	08/10/09	GT SW846 6010B ²	SW846 3050B ⁴
Thallium	< 1.1	1.1	mg/kg	1	08/10/09	08/10/09	GT SW846 6010B ²	SW846 3050B ⁴
Vanadium	27.6	J 5.4	mg/kg	1	08/10/09	08/10/09	GT SW846 6010B ²	SW846 3050B ⁴
Zinc	318	J 2.2	mg/kg	1	08/10/09	08/10/09	GT SW846 6010B ²	SW846 3050B ⁴

- (1) Instrument QC Batch: MA22941
(2) Instrument QC Batch: MA22944
(3) Prep QC Batch: MP49159
(4) Prep QC Batch: MP49161

RL = Reporting Limit

Report of Analysis

3.1
3

Client Sample ID: B-CCI-028(2')	Date Sampled: 08/11/09
Lab Sample ID: JA25446-1	Date Received: 08/12/09
Matrix: SO - Soil	Percent Solids: 79.6
Project: Courtlandt Corners I, 364 East 161st Street, Bronx, NY	

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	4890	26	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Antimony ^a	< 5.2	J 5.2	mg/kg	2	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Arsenic	3.2	2.6	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Barium	66.9	26	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Beryllium	< 0.65	0.65	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Cadmium	< 0.65	0.65	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Calcium	92500	J 1300	mg/kg	2	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Chromium	8.9	1.3	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Cobalt	< 6.5	6.5	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Copper	14.1	3.3	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Iron	6750	13	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Lead ^a	53.9	5.2	mg/kg	2	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Magnesium	46800	650	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Manganese	282	2.0	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Mercury	0.060	0.041	mg/kg	1	08/13/09	08/13/09 JW	SW846 7471A ¹	SW846 7471A ⁴
Nickel	9.3	5.2	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Potassium	1740	1300	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Selenium ^a	< 5.2	J 5.2	mg/kg	2	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Silver	< 1.3	1.3	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Sodium	< 1300	1300	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Thallium ^a	< 2.6	J 2.6	mg/kg	2	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Vanadium	11.1	6.5	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Zinc	34.5	J 2.6	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³

- (1) Instrument QC Batch: MA22964
- (2) Instrument QC Batch: MA22966
- (3) Prep QC Batch: MP49224
- (4) Prep QC Batch: MP49225

(a) Elevated detection limit due to dilution required for high interfering element.

RL = Reporting Limit

Report of Analysis

32
3

Client Sample ID:	B-CCI-029(20')	Date Sampled:	08/12/09
Lab Sample ID:	JA25446-2	Date Received:	08/12/09
Matrix:	SO - Soil	Percent Solids:	79.4
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	10600	24	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Antimony	< 2.4	J 2.4	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Arsenic	< 2.4	2.4	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Barium	62.6	24	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Beryllium	< 0.60	0.60	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Cadmium	< 0.60	J 0.60	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Calcium	17000	J 600	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Chromium	24.6	1.2	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Cobalt	8.2	J 6.0	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Copper	21.0	J 3.0	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Iron	15400	J 12	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Lead	20.8	2.4	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Magnesium	12300	J 600	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Manganese	292	J 1.8	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Mercury	< 0.038	0.038	mg/kg	1	08/13/09	08/13/09	JW SW846 7471A ¹	SW846 7471A ⁴
Nickel	20.3	J 4.8	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Potassium	2250	J 1200	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Selenium	< 2.4	2.4	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Silver	< 1.2	1.2	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Sodium	< 1200	1200	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Thallium	< 1.2	1.2	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Vanadium	28.6	J 6.0	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Zinc	58.0	J 2.4	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³

- (1) Instrument QC Batch: MA22964
- (2) Instrument QC Batch: MA22966
- (3) Prep QC Batch: MP49224
- (4) Prep QC Batch: MP49225

RL = Reporting Limit

Report of Analysis



Client Sample ID: B-CCI-030(4')	Date Sampled: 08/12/09
Lab Sample ID: JA25446-3	Date Received: 08/12/09
Matrix: SO - Soil	Percent Solids: 89.1
Project: Courtlandt Corners I, 364 East 161st Street, Bronx, NY	

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	11400	22	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Antimony	< 2.2	J 2.2	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Arsenic	3.8	2.2	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Barium	271	22	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Beryllium	< 0.56	0.56	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Cadmium	< 0.56	0.56	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Calcium	27500	J 560	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Chromium	24.8	1.1	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Cobalt	8.7	J 5.6	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Copper	36.9	J 2.8	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Iron	17600	J 11	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Lead	320	2.2	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Magnesium	14800	J 560	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Manganese	299	J 1.7	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Mercury	0.36	0.035	mg/kg	1	08/13/09	08/13/09 JW	SW846 7471A ¹	SW846 7471A ⁴
Nickel	22.0	J 4.4	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Potassium	2540	J 1100	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Selenium	< 2.2	2.2	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Silver	< 1.1	1.1	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Sodium	< 1100	1100	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Thallium	< 1.1	1.1	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Vanadium	75.6	J 5.6	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³
Zinc	531	J 2.2	mg/kg	1	08/13/09	08/13/09 GT	SW846 6010B ²	SW846 3050B ³

- (1) Instrument QC Batch: MA22964
- (2) Instrument QC Batch: MA22966
- (3) Prep QC Batch: MP49224
- (4) Prep QC Batch: MP49225

RL = Reporting Limit

Report of Analysis

3.4
3

Client Sample ID:	B-CCI-031(10')	Date Sampled:	08/12/09
Lab Sample ID:	JA25446-4	Date Received:	08/12/09
Matrix:	SO - Soil	Percent Solids:	89.4
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	11000	21	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Antimony	<2.1	J 2.1	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Arsenic	<2.1	J 2.1	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Barium	51.7	21	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Beryllium	<0.53	0.53	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Cadmium	<0.53	0.53	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Calcium	28200	J 530	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Chromium	19.0	J 1.1	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Cobalt	9.9	J 5.3	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Copper	22.4	J 2.7	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Iron	18000	J 11	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Lead	12.6	J 2.1	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Magnesium	22500	J 530	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Manganese	453	J 1.6	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Mercury	0.085	0.034	mg/kg	1	08/13/09	08/13/09	JW SW846 7471A ¹	SW846 7471A ⁴
Nickel	16.6	J 4.3	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Potassium	2320	J 1100	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Selenium	<2.1	J 2.1	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Silver	<1.1	J 1.1	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Sodium	<1100	J 1100	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Thallium	<1.1	J 1.1	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Vanadium	27.5	J 5.3	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³
Zinc	51.6	J 2.1	mg/kg	1	08/13/09	08/13/09	GT SW846 6010B ²	SW846 3050B ³

- (1) Instrument QC Batch: MA22964
- (2) Instrument QC Batch: MA22966
- (3) Prep QC Batch: MP49224
- (4) Prep QC Batch: MP49225

RL = Reporting Limit

Report of Analysis

Client Sample ID:	SW-CCI-001(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-2	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	81.1
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	13300	25	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Antimony	< 2.5	2.5	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Arsenic	3.6	2.5	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Barium	90.9	25	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Beryllium	0.86	0.62	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Cadmium	< 0.62	0.62	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Calcium	11300	620	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Chromium	37.3	1.2	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Cobalt	12.0	6.2	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Copper	35.2	3.1	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Iron	27800	12	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Lead	9.0	2.5	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Magnesium	11800	620	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Manganese	477	1.9	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Mercury	< 0.038	0.038	mg/kg	1	07/23/09	07/24/09 JW	SW846 7471A ¹	SW846 7471A ³
Nickel	29.3	5.0	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Potassium	3880	1200	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Selenium	< 2.5	2.5	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Silver	< 1.2	1.2	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Sodium	< 1200	1200	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Thallium	< 1.2	1.2	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Vanadium	43.7	6.2	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Zinc	68.6	2.5	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴

- (1) Instrument QC Batch: MA22865
(2) Instrument QC Batch: MA22888
(3) Prep QC Batch: MP48970
(4) Prep QC Batch: MP49013

RL = Reporting Limit

Report of Analysis



Client Sample ID:	SW-CCI-002(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-3	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	86.1
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	10500	22	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Antimony	< 2.2	J 2.2	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Arsenic	< 2.2	2.2	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Barium	59.3	22	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Beryllium	0.63	0.56	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Cadmium	< 0.56	0.56	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Calcium	12400	560	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Chromium	25.4	1.1	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Cobalt	8.0	5.6	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Copper	23.5	2.8	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Iron	19000	11	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Lead	7.4	2.2	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Magnesium	11500	560	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Manganese	303	J 1.7	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Mercury	< 0.038	0.038	mg/kg	1	07/23/09	07/24/09 JW	SW846 7471A ¹	SW846 7471A ³
Nickel	19.8	4.5	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Potassium	2720	1100	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Selenium	< 2.2	2.2	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Silver	< 1.1	1.1	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Sodium	< 1100	1100	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Thallium	< 1.1	1.1	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Vanadium	28.3	5.6	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Zinc	53.6	2.2	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴

- (1) Instrument QC Batch: MA22865
- (2) Instrument QC Batch: MA22888
- (3) Prep QC Batch: MP48970
- (4) Prep QC Batch: MP49013

RL = Reporting Limit

Report of Analysis

3.4
3

Client Sample ID: SW-CCI-003(18.5')	Date Sampled: 07/20/09
Lab Sample ID: JA23638-4	Date Received: 07/21/09
Matrix: SO - Soil	Percent Solids: 88.4
Project: Courtlandt Corners I, 364 East 161st Street, Bronx, NY	

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	6150	23	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Antimony	< 2.3	J 2.3	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Arsenic	< 2.3	2.3	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Barium	26.5	23	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Beryllium	< 0.57	0.57	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Cadmium	< 0.57	0.57	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Calcium	4240	570	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Chromium	24.3	1.1	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Cobalt	6.0	5.7	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Copper	18.3	2.8	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Iron	12300	11	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Lead	5.5	2.3	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Magnesium	4480	570	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Manganese	257	J 1.7	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Mercury	< 0.037	0.037	mg/kg	1	07/23/09	07/24/09 JW	SW846 7471A ¹	SW846 7471A ³
Nickel	15.2	4.5	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Potassium	1130	1100	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Selenium	< 2.3	2.3	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Silver	< 1.1	1.1	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Sodium	< 1100	1100	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Thallium	< 1.1	1.1	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Vanadium	22.0	5.7	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Zinc	30.7	2.3	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴

- (1) Instrument QC Batch: MA22865
- (2) Instrument QC Batch: MA22888
- (3) Prep QC Batch: MP48970
- (4) Prep QC Batch: MP49013

RL = Reporting Limit

Report of Analysis

3.5
3

Client Sample ID: SW-CCI-004(19')	Date Sampled: 07/20/09
Lab Sample ID: JA23638-5	Date Received: 07/21/09
Matrix: SO - Soil	Percent Solids: 87.3
Project: Courtlandt Corners I, 364 East 161st Street, Bronx, NY	

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	8240	23	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Antimony	< 2.3	J 2.3	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Arsenic	< 2.3	2.3	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Barium	48.4	23	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Beryllium	0.79	0.58	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Cadmium	< 0.58	0.58	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Calcium	84600	1200	mg/kg	2	07/28/09	08/03/09 VC	SW846 6010B ³	SW846 3050B ⁵
Chromium	17.5	1.2	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Cobalt	6.1	5.8	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Copper	19.2	2.9	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Iron	13800	12	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Lead ^a	7.2	4.6	mg/kg	2	07/28/09	08/03/09 VC	SW846 6010B ³	SW846 3050B ⁵
Magnesium	48400	580	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Manganese	295	J 1.7	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Mercury	< 0.035	0.035	mg/kg	1	07/23/09	07/24/09 JW	SW846 7471A ¹	SW846 7471A ⁴
Nickel	13.3	4.6	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Potassium	2290	1200	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Selenium ^a	< 4.6	4.6	mg/kg	2	07/28/09	08/03/09 VC	SW846 6010B ³	SW846 3050B ⁵
Silver	< 1.2	1.2	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Sodium	< 1200	1200	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Thallium	< 1.2	1.2	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Vanadium	24.0	5.8	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Zinc	45.9	2.3	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵

- (1) Instrument QC Batch: MA22865
- (2) Instrument QC Batch: MA22888
- (3) Instrument QC Batch: MA22904
- (4) Prep QC Batch: MP48970
- (5) Prep QC Batch: MP49013

(a) Elevated detection limit due to dilution required for high interfering element.

RL = Reporting Limit

Report of Analysis

3.6
3

Client Sample ID: SW-CCI-005(18.5')	Date Sampled: 07/20/09
Lab Sample ID: JA23638-6	Date Received: 07/21/09
Matrix: SO - Soil	Percent Solids: 83.6
Project: Courtlandt Corners I, 364 East 161st Street, Bronx, NY	

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	11300	24	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Antimony	< 2.4	J 2.4	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Arsenic	< 2.4	2.4	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Barium	83.7	24	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Beryllium	0.82	0.59	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Cadmium	< 0.59	0.59	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Calcium	34300	590	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Chromium	24.1	1.2	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Cobalt	9.9	5.9	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Copper	24.4	3.0	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Iron	19200	24	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Lead	6.3	2.4	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Magnesium	19100	590	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Manganese	317	J 1.8	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Mercury	< 0.036	0.036	mg/kg	1	07/23/09	07/24/09 JW	SW846 7471A ¹	SW846 7471A ³
Nickel	21.3	4.7	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Potassium	3730	1200	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Selenium	< 2.4	2.4	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Silver	< 1.2	1.2	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Sodium	< 1200	1200	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Thallium	< 1.2	1.2	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Vanadium	33.1	5.9	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Zinc	62.6	2.4	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴

- (1) Instrument QC Batch: MA22865
- (2) Instrument QC Batch: MA22882
- (3) Prep QC Batch: MP48970
- (4) Prep QC Batch: MP49013

RL = Reporting Limit

Report of Analysis

Client Sample ID:	SW-CCI-006(19')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-7	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	91.3
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	4840	21	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Antimony	< 2.1	J 2.1	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Arsenic	< 2.1	2.1	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Barium	< 21	21	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Beryllium	0.65	0.53	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Cadmium	< 0.53	0.53	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Calcium	96100	1600	mg/kg	3	07/28/09	08/03/09 VC	SW846 6010B ³	SW846 3050B ⁵
Chromium	10.4	1.1	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Cobalt	< 5.3	5.3	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Copper	13.8	2.6	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Iron	9480	11	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Lead ^a	< 6.3	6.3	mg/kg	3	07/28/09	08/03/09 VC	SW846 6010B ³	SW846 3050B ⁵
Magnesium	58300	530	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Manganese	227	J 1.6	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Mercury	< 0.032	0.032	mg/kg	1	07/23/09	07/24/09 JW	SW846 7471A ¹	SW846 7471A ⁴
Nickel	9.8	4.2	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Potassium	< 1100	1100	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Selenium ^a	< 6.3	6.3	mg/kg	3	07/28/09	08/03/09 VC	SW846 6010B ³	SW846 3050B ⁵
Silver	< 1.1	1.1	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Sodium	< 1100	1100	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Thallium	< 1.1	1.1	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Vanadium	14.0	5.3	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵
Zinc	38.5	2.1	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁵

(1) Instrument QC Batch: MA22865

(2) Instrument QC Batch: MA22888

(3) Instrument QC Batch: MA22904

(4) Prep QC Batch: MP48970

(5) Prep QC Batch: MP49013

(a) Elevated detection limit due to dilution required for high interfering element.

RL = Reporting Limit

Report of Analysis

Client Sample ID:	SW-CCI-007(19')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-8	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	84.5
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	14600	24	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Antimony	< 2.4	J 2.4	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Arsenic	< 2.4	2.4	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Barium	79.3	24	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Beryllium	1.8	0.60	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Cadmium	< 0.60	0.60	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Calcium	28700	600	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Chromium	30.5	1.2	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Cobalt	10.6	6.0	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Copper	32.4	3.0	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Iron	24800	12	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Lead	8.7	2.4	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Magnesium	18600	600	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Manganese	311	J 1.8	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Mercury	< 0.036	0.036	mg/kg	1	07/23/09	07/24/09 JW	SW846 7471A ¹	SW846 7471A ³
Nickel	24.8	4.8	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Potassium	3470	1200	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Selenium	< 2.4	2.4	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Silver	< 1.2	1.2	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Sodium	< 1200	1200	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Thallium	< 1.2	1.2	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Vanadium	40.9	6.0	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Zinc	72.4	2.4	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴

(1) Instrument QC Batch: MA22865

(2) Instrument QC Batch: MA22888

(3) Prep QC Batch: MP48970

(4) Prep QC Batch: MP49013

RL = Reporting Limit

Report of Analysis



Client Sample ID:	SW-CCI-008(13.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-9	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	85.4
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	6710	23	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Antimony	< 2.3	J 2.3	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Arsenic	< 2.3	2.3	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Barium	33.8	23	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Beryllium	< 0.57	0.57	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Cadmium	< 0.57	0.57	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Calcium	6410	570	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Chromium	30.7	1.1	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Cobalt	7.5	5.7	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Copper	14.6	2.9	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Iron	13600	11	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Lead	5.7	2.3	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Magnesium	4760	570	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Manganese	417	J 1.7	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Mercury	< 0.039	0.039	mg/kg	1	07/23/09	07/24/09 JW	SW846 7471A ¹	SW846 7471A ³
Nickel	13.4	4.6	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Potassium	1680	1100	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Selenium	< 2.3	2.3	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Silver	< 1.1	1.1	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Sodium	< 1100	1100	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Thallium	< 1.1	1.1	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Vanadium	26.1	5.7	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴
Zinc	34.9	2.3	mg/kg	1	07/28/09	07/29/09 VC	SW846 6010B ²	SW846 3050B ⁴

- (1) Instrument QC Batch: MA22865
- (2) Instrument QC Batch: MA22888
- (3) Prep QC Batch: MP48970
- (4) Prep QC Batch: MP49013

RL = Reporting Limit

Report of Analysis

Client Sample ID:	SW-CCI-009(18.5')	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-10	Date Received:	07/21/09
Matrix:	SO - Soil	Percent Solids:	94.6
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method	
Aluminum	2690	21	mg/kg	1	07/28/09	07/29/09	VC	SW846 6010B ²	SW846 3050B ⁵
Antimony	< 2.1	J 2.1	mg/kg	1	07/28/09	07/29/09	VC	SW846 6010B ²	SW846 3050B ⁵
Arsenic	< 2.1	2.1	mg/kg	1	07/28/09	07/29/09	VC	SW846 6010B ²	SW846 3050B ⁵
Barium	< 21	21	mg/kg	1	07/28/09	07/29/09	VC	SW846 6010B ²	SW846 3050B ⁵
Beryllium	< 0.53	0.53	mg/kg	1	07/28/09	07/29/09	VC	SW846 6010B ²	SW846 3050B ⁵
Cadmium	< 0.53	0.53	mg/kg	1	07/28/09	07/29/09	VC	SW846 6010B ²	SW846 3050B ⁵
Calcium	152000	2700	mg/kg	5	07/28/09	08/03/09	VC	SW846 6010B ³	SW846 3050B ⁵
Chromium	6.5	1.1	mg/kg	1	07/28/09	07/29/09	VC	SW846 6010B ²	SW846 3050B ⁵
Cobalt	< 5.3	5.3	mg/kg	1	07/28/09	07/29/09	VC	SW846 6010B ²	SW846 3050B ⁵
Copper	11.2	2.7	mg/kg	1	07/28/09	07/29/09	VC	SW846 6010B ²	SW846 3050B ⁵
Iron	6090	11	mg/kg	1	07/28/09	07/29/09	VC	SW846 6010B ²	SW846 3050B ⁵
Lead ^a	< 11	11	mg/kg	5	07/28/09	08/03/09	VC	SW846 6010B ³	SW846 3050B ⁵
Magnesium	87200	530	mg/kg	1	07/28/09	07/29/09	VC	SW846 6010B ²	SW846 3050B ⁵
Manganese	166	J 1.6	mg/kg	1	07/28/09	07/29/09	VC	SW846 6010B ²	SW846 3050B ⁵
Mercury	< 0.033	0.033	mg/kg	1	07/23/09	07/24/09	JW	SW846 7471A ¹	SW846 7471A ⁴
Nickel	5.6	4.3	mg/kg	1	07/28/09	07/29/09	VC	SW846 6010B ²	SW846 3050B ⁵
Potassium	< 1100	1100	mg/kg	1	07/28/09	07/29/09	VC	SW846 6010B ²	SW846 3050B ⁵
Selenium ^a	< 11	11	mg/kg	5	07/28/09	08/03/09	VC	SW846 6010B ³	SW846 3050B ⁵
Silver	< 1.1	1.1	mg/kg	1	07/28/09	07/29/09	VC	SW846 6010B ²	SW846 3050B ⁵
Sodium	< 1100	1100	mg/kg	1	07/28/09	07/29/09	VC	SW846 6010B ²	SW846 3050B ⁵
Thallium	< 1.1	1.1	mg/kg	1	07/28/09	07/29/09	VC	SW846 6010B ²	SW846 3050B ⁵
Vanadium	8.4	5.3	mg/kg	1	07/28/09	07/29/09	VC	SW846 6010B ²	SW846 3050B ⁵
Zinc	19.0	2.1	mg/kg	1	07/28/09	07/29/09	VC	SW846 6010B ²	SW846 3050B ⁵

- (1) Instrument QC Batch: MA22865
- (2) Instrument QC Batch: MA22888
- (3) Instrument QC Batch: MA22904
- (4) Prep QC Batch: MP48970
- (5) Prep QC Batch: MP49013

(a) Elevated detection limit due to dilution required for high interfering element.

RL = Reporting Limit

Report of Analysis

Client Sample ID:	SW-CCI-011(24.5)	Date Sampled:	08/10/09
Lab Sample ID:	JA25246-1	Date Received:	08/10/09
Matrix:	SO - Soil	Percent Solids:	83.6
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	7890	25	mg/kg	1	08/11/09	08/11/09	GT	SW846 6010B ²
Antimony	< 2.5	2.5	mg/kg	1	J 08/11/09	08/11/09	GT	SW846 6010B ²
Arsenic	< 2.5	2.5	mg/kg	1	08/11/09	08/11/09	GT	SW846 6010B ²
Barium	34.5	25	mg/kg	1	08/11/09	08/11/09	GT	SW846 6010B ²
Beryllium	< 0.62	0.62	mg/kg	1	08/11/09	08/11/09	GT	SW846 6010B ²
Cadmium	< 0.62	0.62	mg/kg	1	08/11/09	08/11/09	GT	SW846 6010B ²
Calcium	53700	1200	mg/kg	2	J 08/11/09	08/11/09	GT	SW846 6010B ²
Chromium	16.0	1.2	mg/kg	1	J 08/11/09	08/11/09	GT	SW846 6010B ²
Cobalt	7.2	6.2	mg/kg	1	08/11/09	08/11/09	GT	SW846 6010B ²
Copper	13.0	3.1	mg/kg	1	08/11/09	08/11/09	GT	SW846 6010B ²
Iron	12900	12	mg/kg	1	08/11/09	08/11/09	GT	SW846 6010B ²
Lead ^a	< 4.9	4.9	mg/kg	2	J 08/11/09	08/11/09	GT	SW846 6010B ²
Magnesium	32300	620	mg/kg	1	08/11/09	08/11/09	GT	SW846 6010B ²
Manganese	195	1.8	mg/kg	1	J 08/11/09	08/11/09	GT	SW846 6010B ²
Mercury	< 0.036	0.036	mg/kg	1	08/11/09	08/11/09	JW	SW846 7471A ¹
Nickel	11.0	4.9	mg/kg	1	08/11/09	08/11/09	GT	SW846 6010B ²
Potassium	2280	1200	mg/kg	1	08/11/09	08/11/09	GT	SW846 6010B ²
Selenium ^a	< 4.9	4.9	mg/kg	2	J 08/11/09	08/11/09	GT	SW846 6010B ²
Silver	< 1.2	1.2	mg/kg	1	08/11/09	08/11/09	GT	SW846 6010B ²
Sodium	< 1200	1200	mg/kg	1	08/11/09	08/11/09	GT	SW846 6010B ²
Thallium ^a	< 2.5	2.5	mg/kg	2	J 08/11/09	08/11/09	GT	SW846 6010B ²
Vanadium	19.5	6.2	mg/kg	1	08/11/09	08/11/09	GT	SW846 6010B ²
Zinc	57.3	2.5	mg/kg	1	08/11/09	08/11/09	GT	SW846 6010B ²

(1) Instrument QC Batch: MA22948

(2) Instrument QC Batch: MA22949

(3) Prep QC Batch: MP49173

(4) Prep QC Batch: MP49180

(a) Elevated detection limit due to dilution required for high interfering element.

RL = Reporting Limit

Report of Analysis

3.1
3

Client Sample ID:	SW-CCI-012	Date Sampled:	08/13/09
Lab Sample ID:	JA25522-1	Date Received:	08/13/09
Matrix:	SO - Soil	Percent Solids:	82.7
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	12900	J 24	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Antimony	< 2.4	J 2.4	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Arsenic	3.0	2.4	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Barium	105	24	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Beryllium	< 0.61	0.61	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Cadmium	< 0.61	0.61	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Calcium	11500	610	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Chromium	32.4	1.2	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Cobalt	10.6	6.1	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Copper	49.0	3.0	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Iron	19500	J 12	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Lead	92.8	2.4	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Magnesium	10400	J 610	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Manganese	321	J 1.8	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Mercury	< 0.036	0.036	mg/kg	1	08/21/09	08/24/09 JW	SW846 7471A ¹	SW846 7471A ⁴
Nickel	21.0	4.8	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Potassium	3040	J 1200	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Selenium	< 2.4	2.4	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Silver	< 1.2	1.2	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Sodium	< 1200	1200	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Thallium	< 1.2	1.2	mg/kg	1	08/24/09	08/28/09 VC	SW846 6010B ³	SW846 3050B ⁵
Vanadium	35.0	6.1	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Zinc	83.7	2.4	mg/kg	1	08/24/09	08/28/09 VC	SW846 6010B ³	SW846 3050B ⁵

- (1) Instrument QC Batch: MA23012
- (2) Instrument QC Batch: MA23026
- (3) Instrument QC Batch: MA23038
- (4) Prep QC Batch: MP49355
- (5) Prep QC Batch: MP49363

RL = Reporting Limit

Report of Analysis

Client Sample ID:	SW-CCI-013	Date Sampled:	08/13/09
Lab Sample ID:	JA25522-2	Date Received:	08/13/09
Matrix:	SO - Soil	Percent Solids:	88.2
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	8430	J 22	mg/kg	1	08/24/09	08/26/09	ND SW846 6010B ²	SW846 3050B ⁵
Antimony	< 2.2	J 2.2	mg/kg	1	08/24/09	08/26/09	ND SW846 6010B ²	SW846 3050B ⁵
Arsenic	4.7	2.2	mg/kg	1	08/24/09	08/26/09	ND SW846 6010B ²	SW846 3050B ⁵
Barium	112	22	mg/kg	1	08/24/09	08/26/09	ND SW846 6010B ²	SW846 3050B ⁵
Beryllium	< 0.55	0.55	mg/kg	1	08/24/09	08/26/09	ND SW846 6010B ²	SW846 3050B ⁵
Cadmium	< 0.55	0.55	mg/kg	1	08/24/09	08/26/09	ND SW846 6010B ²	SW846 3050B ⁵
Calcium	23000	550	mg/kg	1	08/24/09	08/26/09	ND SW846 6010B ²	SW846 3050B ⁵
Chromium	19.7	1.1	mg/kg	1	08/24/09	08/26/09	ND SW846 6010B ²	SW846 3050B ⁵
Cobalt	6.1	5.5	mg/kg	1	08/24/09	08/26/09	ND SW846 6010B ²	SW846 3050B ⁵
Copper	36.6	2.8	mg/kg	1	08/24/09	08/26/09	ND SW846 6010B ²	SW846 3050B ⁵
Iron	15000	J 11	mg/kg	1	08/24/09	08/26/09	ND SW846 6010B ²	SW846 3050B ⁵
Lead	186	2.2	mg/kg	1	08/24/09	08/26/09	ND SW846 6010B ²	SW846 3050B ⁵
Magnesium	10700	J 550	mg/kg	1	08/24/09	08/26/09	ND SW846 6010B ²	SW846 3050B ⁵
Manganese	272	J 1.7	mg/kg	1	08/24/09	08/26/09	ND SW846 6010B ²	SW846 3050B ⁵
Mercury	0.24	0.037	mg/kg	1	08/21/09	08/24/09	JW SW846 7471A ¹	SW846 7471A ⁴
Nickel	12.8	4.4	mg/kg	1	08/24/09	08/26/09	ND SW846 6010B ²	SW846 3050B ⁵
Potassium	1590	J 1100	mg/kg	1	08/24/09	08/26/09	ND SW846 6010B ²	SW846 3050B ⁵
Selenium	2.2	2.2	mg/kg	1	08/24/09	08/26/09	ND SW846 6010B ²	SW846 3050B ⁵
Silver	< 1.1	1.1	mg/kg	1	08/24/09	08/26/09	ND SW846 6010B ²	SW846 3050B ⁵
Sodium	< 1100	1100	mg/kg	1	08/24/09	08/26/09	ND SW846 6010B ²	SW846 3050B ⁵
Thallium	< 1.1	1.1	mg/kg	1	08/24/09	08/26/09	ND SW846 6010B ²	SW846 3050B ⁵
Vanadium	24.0	5.5	mg/kg	1	08/24/09	08/26/09	ND SW846 6010B ²	SW846 3050B ⁵
Zinc	119	2.2	mg/kg	1	08/24/09	08/28/09	VC SW846 6010B ³	SW846 3050B ⁵

- (1) Instrument QC Batch: MA23012
- (2) Instrument QC Batch: MA23026
- (3) Instrument QC Batch: MA23038
- (4) Prep QC Batch: MP49355
- (5) Prep QC Batch: MP49363

RL = Reporting Limit

Report of Analysis

Client Sample ID:	SW-CCI-014	Date Sampled:	08/13/09
Lab Sample ID:	JA25522-3	Date Received:	08/13/09
Matrix:	SO - Soil	Percent Solids:	85.6
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	12400	J 23	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Antimony	< 2.3	J 2.3	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Arsenic	5.7	2.3	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Barium	141	23	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Beryllium	< 0.58	0.58	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Cadmium	0.85	0.58	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Calcium	21600	580	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Chromium	28.4	1.2	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Cobalt	9.0	5.8	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Copper	44.4	2.9	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Iron	18700	J 12	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Lead	633	2.3	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Magnesium	12500	J 580	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Manganese	445	J 1.8	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Mercury	0.32	0.037	mg/kg	1	08/21/09	08/24/09 JW	SW846 7471A ¹	SW846 7471A ⁴
Nickel	17.4	4.7	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Potassium	1400	J 1200	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Selenium	< 2.3	2.3	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Silver	< 1.2	1.2	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Sodium	< 1200	1200	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Thallium	< 1.2	1.2	mg/kg	1	08/24/09	08/28/09 VC	SW846 6010B ³	SW846 3050B ⁵
Vanadium	26.8	5.8	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵
Zinc	360	2.3	mg/kg	1	08/24/09	08/26/09 ND	SW846 6010B ²	SW846 3050B ⁵

- (1) Instrument QC Batch: MA23012
- (2) Instrument QC Batch: MA23026
- (3) Instrument QC Batch: MA23038
- (4) Prep QC Batch: MP49355
- (5) Prep QC Batch: MP49363

RL = Reporting Limit

Report of Analysis

Client Sample ID:	SW-CCI-015	Date Sampled:	08/14/09
Lab Sample ID:	JA25638-1	Date Received:	08/14/09
Matrix:	SO - Soil	Percent Solids:	81.8
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	12900	J 25	mg/kg	1	08/17/09	08/17/09	ND SW846 6010B ²	SW846 3050B ⁴
Antimony	< 2.5	J 2.5	mg/kg	1	08/17/09	08/17/09	ND SW846 6010B ²	SW846 3050B ⁴
Arsenic	< 2.5	2.5	mg/kg	1	08/17/09	08/17/09	ND SW846 6010B ²	SW846 3050B ⁴
Barium	79.2	25	mg/kg	1	08/17/09	08/17/09	ND SW846 6010B ²	SW846 3050B ⁴
Beryllium	< 0.63	0.63	mg/kg	1	08/17/09	08/17/09	ND SW846 6010B ²	SW846 3050B ⁴
Cadmium	< 0.63	0.63	mg/kg	1	08/17/09	08/17/09	ND SW846 6010B ²	SW846 3050B ⁴
Calcium	12600	630	mg/kg	1	08/17/09	08/17/09	ND SW846 6010B ²	SW846 3050B ⁴
Chromium	29.1	1.3	mg/kg	1	08/17/09	08/17/09	ND SW846 6010B ²	SW846 3050B ⁴
Cobalt	11.6	6.3	mg/kg	1	08/17/09	08/17/09	ND SW846 6010B ²	SW846 3050B ⁴
Copper	28.8	3.2	mg/kg	1	08/17/09	08/17/09	ND SW846 6010B ²	SW846 3050B ⁴
Iron	19000	13	mg/kg	1	08/17/09	08/17/09	ND SW846 6010B ²	SW846 3050B ⁴
Lead	7.3	2.5	mg/kg	1	08/17/09	08/17/09	ND SW846 6010B ²	SW846 3050B ⁴
Magnesium	12400	630	mg/kg	1	08/17/09	08/17/09	ND SW846 6010B ²	SW846 3050B ⁴
Manganese	375	1.9	mg/kg	1	08/17/09	08/17/09	ND SW846 6010B ²	SW846 3050B ⁴
Mercury	< 0.039	0.039	mg/kg	1	08/15/09	08/15/09	JW SW846 7471A ¹	SW846 7471A ³
Nickel	25.1	5.0	mg/kg	1	08/17/09	08/17/09	ND SW846 6010B ²	SW846 3050B ⁴
Potassium	3950	1300	mg/kg	1	08/17/09	08/17/09	ND SW846 6010B ²	SW846 3050B ⁴
Selenium	< 2.5	2.5	mg/kg	1	08/17/09	08/17/09	ND SW846 6010B ²	SW846 3050B ⁴
Silver	< 1.3	1.3	mg/kg	1	08/17/09	08/17/09	ND SW846 6010B ²	SW846 3050B ⁴
Sodium	< 1300	1300	mg/kg	1	08/17/09	08/17/09	ND SW846 6010B ²	SW846 3050B ⁴
Thallium	< 1.3	1.3	mg/kg	1	08/17/09	08/17/09	ND SW846 6010B ²	SW846 3050B ⁴
Vanadium	34.9	6.3	mg/kg	1	08/17/09	08/17/09	ND SW846 6010B ²	SW846 3050B ⁴
Zinc	61.0	2.5	mg/kg	1	08/17/09	08/17/09	ND SW846 6010B ²	SW846 3050B ⁴

(1) Instrument QC Batch: MA22975

(2) Instrument QC Batch: MA22979

(3) Prep QC Batch: MP49257

(4) Prep QC Batch: MP49260

RL = Reporting Limit

Report of Analysis

Client Sample ID:	FB072009	Date Sampled:	07/20/09
Lab Sample ID:	JA23638-1	Date Received:	07/21/09
Matrix:	AQ - Field Blank Soil	Percent Solids:	n/a
Project:	Courtlandt Corners I, 364 East 161st Street, Bronx, NY		

Total Metals Analysis

Analyte	Result	RL	Units	DF	Prep	Analyzed By	Method	Prep Method
Aluminum	< 200	200	ug/l	1	07/28/09	07/29/09 GT	SW846 6010B ²	SW846 3010A ⁴
Antimony	< 6.0	6.0	ug/l	1	07/28/09	07/29/09 GT	SW846 6010B ²	SW846 3010A ⁴
Arsenic	< 3.0	3.0	ug/l	1	07/28/09	07/29/09 GT	SW846 6010B ²	SW846 3010A ⁴
Barium	< 200	200	ug/l	1	07/28/09	07/29/09 GT	SW846 6010B ²	SW846 3010A ⁴
Beryllium	< 1.0	1.0	ug/l	1	07/28/09	07/29/09 GT	SW846 6010B ²	SW846 3010A ⁴
Cadmium	< 3.0	3.0	ug/l	1	07/28/09	07/29/09 GT	SW846 6010B ²	SW846 3010A ⁴
Calcium	< 5000	5000	ug/l	1	07/28/09	07/29/09 GT	SW846 6010B ²	SW846 3010A ⁴
Chromium	< 10	10	ug/l	1	07/28/09	07/29/09 GT	SW846 6010B ²	SW846 3010A ⁴
Cobalt	< 50	50	ug/l	1	07/28/09	07/29/09 GT	SW846 6010B ²	SW846 3010A ⁴
Copper	< 10	10	ug/l	1	07/28/09	07/29/09 GT	SW846 6010B ²	SW846 3010A ⁴
Iron	< 100	100	ug/l	1	07/28/09	07/29/09 GT	SW846 6010B ²	SW846 3010A ⁴
Lead	< 3.0	3.0	ug/l	1	07/28/09	07/29/09 GT	SW846 6010B ²	SW846 3010A ⁴
Magnesium	< 5000	5000	ug/l	1	07/28/09	07/29/09 GT	SW846 6010B ²	SW846 3010A ⁴
Manganese	< 15	15	ug/l	1	07/28/09	07/29/09 GT	SW846 6010B ²	SW846 3010A ⁴
Mercury	< 0.20	0.20	ug/l	1	07/23/09	07/23/09 JW	SW846 7470A ¹	SW846 7470A ³
Nickel	< 10	10	ug/l	1	07/28/09	07/29/09 GT	SW846 6010B ²	SW846 3010A ⁴
Potassium	< 10000	10000	ug/l	1	07/28/09	07/29/09 GT	SW846 6010B ²	SW846 3010A ⁴
Selenium	< 10	10	ug/l	1	07/28/09	07/29/09 GT	SW846 6010B ²	SW846 3010A ⁴
Silver	< 10	10	ug/l	1	07/28/09	07/29/09 GT	SW846 6010B ²	SW846 3010A ⁴
Sodium	< 10000	10000	ug/l	1	07/28/09	07/29/09 GT	SW846 6010B ²	SW846 3010A ⁴
Thallium	< 2.0	2.0	ug/l	1	07/28/09	07/29/09 GT	SW846 6010B ²	SW846 3010A ⁴
Vanadium	< 50	50	ug/l	1	07/28/09	07/29/09 GT	SW846 6010B ²	SW846 3010A ⁴
Zinc	< 20	20	ug/l	1	07/28/09	07/29/09 GT	SW846 6010B ²	SW846 3010A ⁴

(1) Instrument QC Batch: MA22860

(2) Instrument QC Batch: MA22889

(3) Prep QC Batch: MP48955

(4) Prep QC Batch: MP49016

RL = Reporting Limit