

**FINAL PROJECT REPORT FOR THE
DECONTAMINATION AND RELEASE
OF THE BAUSCH & LOMB BATCH ROOM**

**BAUSCH & LOMB
ROCHESTER, NEW YORK**

January, 1995

NES Document Number 82A8629

Prepared and submitted by:
NES, Inc./IES
44 Shelter Rock Road
Danbury, CT 06810


**FINAL PROJECT REPORT FOR THE
DECONTAMINATION AND RELEASE
OF THE BAUSCH & LOMB BATCH ROOM**

**BAUSCH & LOMB
ROCHESTER, NEW YORK**

January, 1995

NES Document Number 82A8629

**Prepared and submitted by:
NES, Inc./IES
44 Shelter Rock Road
Danbury, CT 06810**

Prepared by: 
Dwight L. Vann
Radiological Engineer


1/17/95
Date

Review by: 
Brandon Graber
Site Supervisor

1/17/95
Date

Review by: 
Joe Darman
Project Manager

1/17/95
Date

Approved by: 
Dennis W. Reisenweaver
Department Manager

1/17/95
Date

TABLE OF CONTENTS

1.	EXECUTIVE SUMMARY	1
2.	WORK PLAN SUMMARY	1
3.	SCHEDULE	4
4.	RADIOLOGICAL PROTECTION	4
4.1	Radiation Work Permits	4
4.2	Contamination Control	4
4.3	Airborne Radioactivity	4
5.	RELEASE CRITERIA	5
6.	QUALITY ASSURANCE	6
7.	DECOMMISSIONING OPERATIONS	6
7.1	Mobilization	6
7.2	Initial Conditions	6
7.2.1	Initial surveys	6
7.2.2	Initial survey results summary	9
7.3	Setup Radiologically Controlled Areas (RCAs)	10
7.4	Concrete Decontamination	11
7.5	Other Decontamination	11
7.6	Air Sampling	12
7.7	Final Release Survey	12
7.8	Preparation/Shipment of Waste	12
7.8.1	Waste characterization	13
7.9	Demobilization	14
8.	FINAL RELEASE SURVEY	14
8.1	Final Survey Approach	14
8.1.1	Surface scan	15
8.1.2	Direct measurements	15
8.1.3	Removable contamination measurements	16
8.1.4	Dose rate measurements	16
8.2	Survey Instrumentation	16
8.2.1	MDA's	17
8.2.2	Interpretation of survey results	18
8.3	Summary of Results	18
8.3.1	Unaffected areas	18

8.3.2 Affected Area	21
9. REFERENCES	23

ATTACHMENT 1 SURVEY SHEETS

LIST OF TABLES

Table 1: Acceptable Surface Contamination Levels	5
Table 2: Bird Cage	9
Table 3: South Hood	9
Table 4: Middle Hood	10
Table 5: Bulk Sample Results	10
Table 6: Final Survey Instrumentation	17
Table 7: Summary of Results From Unaffected Area Surveys	19
Table 7a: MDA's for Unaffected Area Surveys	19
Table 7b: True Mean Activity Upper Limit at the 95% Confidence Level for the Unaffected Area Survey Units	20
Table 7c: Summary of Dose Rate Measurements for Unaffected Areas	20
Table 8: Summary of Survey Results From Affected Area (Batch Room)	21
Table 8a: MDA's for Affected Area Surveys	21
Table 8b: True Mean Activity Upper Limit at the 95% Confidence Level for the Affected Area Survey Units	22
Table 8c: Summary of Dose Rate Measurements for Affected Area	22

LIST OF FIGURES

Figure 1: Glass Plant	2
Figure 2: Batch Room	3

1. EXECUTIVE SUMMARY

NES/IES was awarded the contract to perform the decontamination and final release survey of the Bausch & Lomb Glass Plant Batch Room in December, 1994. Mobilization began on December 5, 1994, and the task was completed on December 15, 1994. Along with the decontamination of the Batch Room, four unaffected areas were surveyed for unrestricted release. Figure 1 shows the location of the Batch Room within the Bausch & Lomb Glass Plant.

The objective was to remove and dispose of contaminated materials from the Batch Room, and release the unaffected areas so the Glass Plant could be released for unrestricted use. At the end of the decontamination activities, a final release survey was performed within the Batch Room to verify that the decontamination efforts were successful (see Figure 2 for a diagram of the Batch Room). All building surface decontamination efforts were successful as evidenced by the final release survey. Surface contamination release level results complied with the "Guidelines for Decontamination of Facilities and Equipment Prior to Release for Unrestricted Use or Termination of Licenses for Byproduct, Source or Special Nuclear Material," USNRC, August 1987; and Table 5 of the New York Department of Labor Industrial Code Rule 38 "Ionizing Radiation Protection," June 29, 1994.

No personnel exposure to ionizing radiation in excess of 10% of state standards were noted during this work. Therefore 12 NYCRR Part 38.28, "Records" requirements are not applicable, as monitoring was not required pursuant to 12 NYCRR Part 38.24, "Personnel Monitoring."


Radioactive waste was packaged and awaits disposal at Envirocare of Utah, Inc, pending sample analysis. The waste consisted of thorium contaminated concrete, sheet metal, miscellaneous debris and radioactive potassium.

2. WORK PLAN SUMMARY

The purpose of the work plan was to provide a logical and safe method for the decontamination of the Batch Room at the Bausch & Lomb Glass Plant located in Rochester, New York.


The goal was to decontaminate the facility to below the unrestricted release criteria. The final release survey was performed in accordance with NUREG/CR-5849, "Manual for Conducting Radiological Surveys in Support of License Termination," June, 1992.

Figure 1: Glass Plant



* Not drawn to scale.

Figure 2: Batch Room



* Not drawn to scale

The project was divided into seven tasks to ensure the requirements of Bausch & Lomb were met in the most cost effective manner. These tasks were:

- Task 1 - Engineering
- Task 2 - Mobilization
- Task 3 - Perform Initial Survey
- Task 4 - Scabble Floor
- Task 5 - Final Survey
- Task 6 - Demobilization
- Task 7 - Final Report

3. SCHEDULE

NES/IES began mobilization and initial surveys on December 5, 1994. Decontamination activities began December 8, 1994. The final survey was started on December 12, 1994 and was completed December 14, 1994. Demobilization was completed on December 15, 1994.

4. RADIOLOGICAL PROTECTION

4.1 Radiation Work Permits

All decommissioning work within established Radiologically Controlled Areas (RCA's) was performed under the issuance of a Radiation Work Permit (RWP) prepared by the site supervisor. The RWP described the radiological conditions under which work in the RCA was to be performed. The RWP required the use of engineering controls and protective clothing, as necessary, to ensure that the work was accomplished in a radiologically safe manner while maintaining personnel radiation exposure as low as reasonably achievable (ALARA).

4.2 Contamination Control

RCA's were established where decontamination activities could generate airborne radioactivity or surface contamination. RCA's were isolated from the general work areas through the use of radiation barrier rope and warning signs. Prior to leaving an RCA, all personnel surveyed themselves with survey meters.

4.3 Airborne Radioactivity

Air samples were taken on a routine basis during all survey and decontamination activities within the Batch Room. The air samples were used to verify that the decontamination activities were not creating airborne radioactivity.

Airborne concentrations did not approach $2\text{E-}13 \mu\text{Ci/ml}$ which would have required the use of respiratory protection.

5. RELEASE CRITERIA

Release criteria established for the decontamination of the Bausch & Lomb Batch Room is in accordance with the levels presented in USNRC draft Regulatory Guide "Guidelines for the Decontamination of Facilities and Equipment Prior to Release for Unrestricted Use or Termination of Licenses for Byproduct, Source and Special Nuclear Material," May, 1987, and Table 5 of the New York Department of Labor Industrial Code Rule 38, "Ionizing Radiation Protection," June 29, 1994. The criteria from the USNRC guidance is stated below:

Table 1: Acceptable Surface Contamination Levels

NUCLIDE ^a	AVERAGE ^{b c}	MAXIMUM ^{b d}	REMOVABLE ^{b e}
U-nat, U-235, U-238, and associated decay products	5,000 dpm α /100 cm ²	15,000 dpm α /100 cm ²	1,000 dpm α /100 cm ²
Th-nat, Th-232, Sr-90, Ra-223, Ra-224, U-232, I-126, I-131, I-133	1,000 dpm/100 cm ²	3,000 dpm/100 cm ²	200 dpm/100 cm ²
Beta-gamma emitters (nuclides with decay modes other than alpha emission or spontaneous fission) except Sr-90 and others noted above.	5,000 dpm β - γ /100 cm ²	15,000 dpm β - γ /100 cm ²	1,000 dpm β - γ /100 cm ²

- ^a Where surface contamination by both alpha and beta/gamma emitting nuclides exists, the limits established for alpha and beta/gamma emitting nuclides should apply independently.
- ^b As used in this table, dpm (disintegrations per minute) means the rate of emission by radioactive material as determined by correcting the counts per minute observed by an appropriate detector for background, efficiency, and geometric factors associated with the instrumentation.
- ^c Measurements of average contaminant should not be averaged over more than 1 square meter. For objects of less surface area, the average should be derived for each such object.
- ^d The maximum contamination level applies to an area of not more than 100 cm².
- ^e The amount of removable radioactive material per 100 cm² of surface area should be determined by wiping that area with dry filter or soft absorbent paper, applying moderate pressure, and assessing the amount of radioactive material on the wipe with an appropriate instrument of known efficiency. When removable contamination on objects of less surface area is determined, the pertinent levels should be reduced proportionally and the entire surface should be wiped.

In addition to the above guidance, dose rates did not exceed background by more than $5 \mu\text{Rem/hr}$, measured at 1 meter from building surfaces.

6. QUALITY ASSURANCE

The NES/IES Program Management Plan (document #82A8626) was implemented for the duration of the project. The elements of this plan include controlled procedures for performing all decommissioning activities, daily instrument performance checks, data review including routine surveys, radiation work permits, and the use of properly calibrated instrumentation.

7. DECOMMISSIONING OPERATIONS

7.1 Mobilization

NES personnel arrived at the Bausch & Lomb site on December 5, 1994 and set up an office and count room adjacent to the Batch Room. Material and equipment were transported from the Danbury office to the Bausch & Lomb site. NES personnel familiarized themselves with the project area by conducting a walk-through of the Glass Plant.

All NES personnel were qualified radiation workers in accordance with 10 CFR 19.12, and received a general safety briefing from the Site Supervisor.

7.2 Initial Conditions

During the decontamination and survey activities, the building was being prepared for demolition. Most of the useful equipment and material had been removed. All material and equipment still located within the building was considered scrap. Utilities such as electric, gas, phone, and water were still being supplied at the start of the decontamination activities. Access to the site was limited to personnel directly related to the demolition of the Glass Plant.

A large hole in the roof was located at the northern end of the Batch Room. Part of the floor area was directly exposed to snow and rain as a result. Heat was not available in the Batch Room.

7.2.1 Initial surveys

Baseline surveys were performed to determine background radioactivity levels, and proper posting requirements. Background dose rates were determined to be $5 \mu\text{Rem/hr}$. Five areas were identified by Bausch & Lomb for initial surveying

including: the Batch Room, R&D Area, Furnace Area, Storage Room, and the Process Line. These areas were selected based on Bausch & Lomb's understanding of the site operation. Based on operating history and prior survey results, the Batch Room was classified as an affected area, while the remaining four areas were classified as unaffected areas. The Batch Room was designated as a Radiologically Controlled Area (RCA), and "Radioactive Material" signs were posted at all entrances to the room.

Survey measurements for surface activity consisted of a combination of surface scans, direct (i.e., fixed plus removable) contamination measurements, and removable surface contamination measurements. The surface scan is used to identify the presence of elevated direct radiation that might indicate activity or hot spots. The surface scan was conducted for all nuclides potentially present, i.e., thorium isotopes and their associated daughters. Systematic measurements of direct and removable surface contamination were performed to quantify the levels of activity.

Minimum Detectable Activities (MDA's) were calculated for both the fixed contamination survey instrumentation and the smear counter. See Attachment 1 for MDA calculations and results.

Air sampling was performed in the Batch Room for each day of decontamination activities. See Attachment 1 for Minimum Detectable Count Rate (MDCR)/Minimum Detectable Concentration (MDC) calculations and air sample results.

Surface Scan

A 100% surface scan survey was performed for the floor and lower wall surfaces of affected areas. A 10% scan was performed for the floor and lower wall surfaces of unaffected areas. A ratemeter/scaler connected to a Geiger-Muller (GM) pancake probe was used to perform the surface scan for beta/gamma emitters and a ZnS detector was used to scan for alpha emitters. The surface scan was performed by keeping the detector as close to the surface as possible and moving the detector at a slow speed (i.e., speed did not exceed one detector width per second for alpha and beta/gamma radiations). Areas of elevated surface activity were noted by changes in the audible signal from the survey meter, rather than by observing fluctuations in the analog/digital meter reading.

Areas of elevated activity identified by the surface scan were marked on the surface with colored chalk by circling the elevated areas to establish boundaries. Direct measurements were taken to determine the extent of residual activity.

Scans of the Glass Plant unaffected areas led to the discovery of two areas with radioactivity above background. The first area was the R&D area, located in the southwest corner of the building. The radioactivity in this area was limited to steel trays containing a crystalline form of KNO_3 . After the trays were transferred to the Batch Room, the R&D area was resurveyed as an unaffected area. The second area was the Furnace Area, located in the middle of the northern half of the Glass Plant. One of the furnaces had traces of radioactivity in the fire brick. The radioactivity in the firebrick was attributed to naturally occurring radioactive material (NORM) and not the result of licensed operations. Furthermore, the furnace area scan did not identify any areas greater than the release criteria. For this reason, the firebrick was left in place and the furnace area was surveyed as an unaffected area. The storage and process line were surveyed as unaffected areas and no residual activity above the release criteria was found.

Direct Measurements

For all building surfaces, i.e., floors, lower and upper walls, and ceilings, random measurement locations for each surface were collected. One 1 minute direct beta-gamma and one 1 minute direct alpha surface contamination reading were taken from areas with the highest elevated activity identified during the surface scan. The Ludlum 2220 meter and 44-9 GM pancake probe were used to perform the beta-gamma direct measurements, while the Ludlum 2220 meter and AC-3 alpha scintillation probe were used to perform the alpha direct measurements.

Removable Contamination Measurements

Smears for removable contamination were collected at random locations throughout the Glass Plant. The smears were obtained by wiping an area of approximately 100 cm^2 using a dry filter paper. Survey results indicate removable contamination levels below USNRC and state guidelines.

Dose Rate Measurements

Dose rate measurements were obtained using a Bicron MicroRem meter. The dose rates in the Glass Plant were $< 5 \mu\text{Rem/hr}$ above background at all surveyed locations within the building.

Bulk Samples

Bulk samples were collected from the vacuum cleaner dust and the KNO_3 trays to determine the nature and extent of the contamination.

7.2.2 Initial survey results summary

The following tables summarize the results from the initial surveys that had indications of activity level greater than USNRC and state release criteria. See Attachment 1 for initial survey documents including layouts.

Batch Room

Table 2: Bird Cage

Location	Direct β - γ	Direct α	Removable β - γ , α
Platform	5,182	< 69	NA
Platform	8,347	< 69	NA
Platform	7,647	< 69	NA

Note: Activity results are in dpm/100 cm²

Table 3: South Hood

Location	Direct β - γ	Direct α	Removable β - γ , α
Center Roof	17,600	2,430	NA
Right Roof	12,100	4,440	NA
Sifter Basket	19,700	1,050	< 32, < 8
Sifter Basket	30,900	1,590	< 32, < 8

Note: Activity results are in dpm/100 cm²

Table 4: Middle Hood

Grid Number	Direct β - γ	Direct α	Removable β - γ , α
Right Exhaust	30,100	1,200	< 32, < 8
Right Roof	25,200	6,420	< 32, < 8
Right Wall	21,476	990	NA
Sifter Side	140,000	42,000	89, 52
Floor	4,200	3,810	< 32, < 8
Floor	9,604	2,250	< 32, < 8
Floor	23,156	2,010	< 32, 35
Floor	15,680	450	NA

Note: Activity results are in dpm/100 cm²

Table 5: Bulk Sample Results

Sample Source	Type	Location	Results
Vacuum Cleaner	Dust, Chipped Concrete	Batch Room	55 pCi/g Th-232 35 pCi/g K-40
KNO ₃ Trays	Crystalline, powdered	Batch Room (Removed from R&D Area)	361 pCi/g K-40

7.3 Setup Radiologically Controlled Areas (RCAs)

The Batch Room was maintained as a radiologically controlled area (RCA) during the project. The entrance to the RCA was the northern sliding door. The western sliding door was roped off and posted. This door was kept shut during the project, and only used before and after decontamination activities. A 4'x4' hole in the western wall was roped and posted until it was nailed shut with a wood panel. Personnel were required to frisk out at the RCA entrance, and a waste container was available for potentially contaminated clothing and supplies.

7.4 Concrete Decontamination

This task required the issuance of a Radiation Work Permit (RWP) prior to task commencement. The RWP included worker protective clothing requirements, health physics technician (HP) job coverage and air sampling requirements during concrete decontamination operations.

The fixed floor contamination within the Batch Room was removed using a chipping hammer with a scabbling attachment. The scabbler worked by driving steel points into the surface of the concrete at very high speeds. Generated debris was collected by the vacuum cleaner during the scabbling process. Dust was controlled by wetting the surface to be scabbled and keeping the vacuum hose close to the scabbler during operation. A High Efficiency Particulate Air (HEPA) vacuum was used to collect any remaining debris on the floor.

The effectiveness of the decontamination was determined through the use of field survey instrumentation. Specifically, as the concrete was scabbled, the freshly exposed concrete surface was scanned for the presence of further contamination. Alpha probes (e.g., Eberline AC-3) were used for detection of thorium isotopes, while beta/gamma probes (e.g., Ludlum 44-9 GM pancake probes) allowed detection of K-40 or thorium daughters in the contaminated concrete. Any remaining "hot spots" were identified and further remediated to below release levels.

The scabbled concrete debris was placed in a disposal container (a B-25 strong tight container). Personnel were required to wear respiratory protection while emptying the vacuum cleaner in precaution of airborne contamination being generated. The collected dust and debris was carefully deposited into a bag, sealed and placed into a disposal container.

7.5 Other Decontamination

Other surfaces additional to concrete required decontamination. The bird cage had a contaminated, crusty substance which was loosened and vacuumed up. Contaminated wood along the southern wall was cut out with a reciprocating saw. Contaminated sheet metal from the south and middle hoods was cut out with a reciprocating saw. Contaminated ventilation from the middle hood was segmented with the reciprocating saw to accommodate the limited space of the B-25 box. The contaminated sifter basket from the south hood was removed from the sifter. The entire sifter from the middle hood was disassembled. All waste generated from these activities was disposed of into a B-25 strong tight box. The trays of KNO_3 were collected and disposed of separately in a 55 gallon drum.

These tasks also required the issuance of a Radiation Work Permit (RWP) prior to task commencement. The RWP included the worker protective clothing requirements, HP coverage and air sampling requirements.

Contamination surveys were performed as necessary outside the radiologically controlled areas. These surveys were conducted to verify that contamination had not spread from the RCA during decommissioning operations. These surveys aided in the substantiation that administrative and engineering controls implemented during decontamination operations were adequate. No contamination outside of the RCA was detected by the routine surveys.

7.6 Air Sampling

Air samples were collected over the course of the decommissioning project. Air samples with initial results near or exceeding $2\text{E-}13 \mu\text{Ci/ml}$ for alpha emitters were recounted within 24 hours to allow for decay of short-lived naturally-occurring radionuclides. No long lived activity was detected. The air samples collected during decontamination activities were all less than the allowable derived air concentration for alpha emitters. Air sample results are contained in Attachment 1.

7.7 Final Release Survey

At the completion of decontamination efforts, a final release survey was performed in accordance with NUREG/CR-5849, "Manual for Conducting Radiological Surveys in Support of License Termination," 1992. The survey objective was the unrestricted release of the Glass Plant. Section 8 of this report contains survey methodology and results.

7.8 Preparation/Shipment of Waste

This task included packaging of the radioactive waste. Contaminated waste consisted of concrete, wood, sheet metal, protective clothing, decontamination material (i.e., HEPA filter, rags, etc.). The waste containers were surveyed and the results documented (see Attachment 1). Two representative samples of the waste have been sent to a Utah state approved laboratory to be analyzed for acceptance at the Envirocare facility. The containers are stored on-site until they are approved for final shipment to Envirocare of Utah Inc. All non-radioactive waste will remain on-site for disposal by Bausch & Lomb.

7.8.1 Waste characterization

Initial analysis of the two wastes generated (i.e. concrete removed and KNO_3) is as follows:

K-40	240 lbs	x	361 pCi/g	x	454 g/lb	=	3.94×10^7 pCi	(55 gallon drum)
Th-nat	1196 lbs	x	50 pCi/g	x	454 g/lb	=	2.71×10^7 pCi	(B-25 Box)
K-40	1196 lbs	x	35 pCi/g	x	454 g/lb	=	1.90×10^7 pCi	(B-25 Box)

For a total of:

27.1 μCi of Th-nat, and 19.0 μCi of K-40 in the B-25 Box (46.1 μCi total)
68.2 μCi of K-40 in the 55 gallon drum.

Both totals are below the exempt quantities in 12 NYCRR Part 38.41, Table 4. Thus, per 12 NYCRR Part 38.25 Section c.4.i, labeling of waste containers is not required. Furthermore, the quantity of K-40 in Bausch & Lomb's possession is within the allowable limits as established in 12 NYCRR Part 38, Table 1, Exemption 28.

The total specific activity of the thorium is determined using the following equation:

$$\text{Curies/gram} = \frac{(N) (1.873e-11)}{T_{1/2}} \quad (1)$$

where,

$T_{1/2}$ = Half life (4.45×10^{17} seconds)
N = Number of atoms per gram (2.6×10^{21} atoms per gram)

For a total of 0.11 $\mu\text{Ci/g}$ of Th-232.

With a total of 27.1 μCi of Th-232 in the waste container, and a specific activity of 0.11 $\mu\text{Ci/g}$, the total weight of Th-232 is 246.4 grams (8.8 ounces). This is below the 15 pounds of Th-232 allowed under 12 NYCRR Part 38, Table 3, item d. Similar calculations were performed for Th-228 resulting in an insignificant contribution to the total source material possessed by Bausch & Lomb.

7.9 Demobilization

All equipment was packed up for removal back to the Danbury office on December 15, 1994. The site was cleaned of all rubbish generated by the work activities. A site closeout meeting was held between NES and Bausch & Lomb to discuss the project and to transfer responsibility of open items to Bausch & Lomb (waste disposal).

8. FINAL RELEASE SURVEY

The survey procedures contained herein are consistent with the recommendations of draft NUREG/CR-5849, "Manual for Conducting Radiological Surveys in Support of License Termination."

8.1 Final Survey Approach

The initial site characterization served as a guide to effectively classify the Bausch & Lomb Glass Plant into affected and unaffected areas. Specifically, the affected area was taken as the area containing residual activity above USNRC and state release criteria, i.e., the Batch Room.

A survey unit was defined for the affected area according to the guidance presented in NUREG/CR-5849. The Batch Room constituted one survey unit (see Figure 1 for room location). A grid system was established within the room to facilitate the systematic selection of survey locations and to provide a mechanism for referencing a survey measurement to a specific location for possible further remediation.

The basic grid system consisted of dividing the floors and 2 meters up the lower walls into 1 meter grids. The upper walls and the ceilings were not gridded. Survey measurements performed on ungridded surfaces were referenced to a floor grid or to a prominent building feature.

The remaining areas (i.e., R&D Area, Furnace Area, Storage Room, and Process Line) were classified as unaffected areas and the initial survey served as the final release survey. Each area was defined as a survey unit for purposes of the survey. A minimum of 10% of the floor and lower walls were surveyed in each area. A minimum of thirty randomly selected spots from each area were surveyed for total and removable activity. Dose rate measurements were taken at 1 meter from the floor in all of the unaffected areas.

8.1.1 Surface scan

Initially, a 100% surface scan survey was performed for the floor and lower wall surfaces of affected areas and 10% for unaffected areas. A gas proportional floor monitor was used to initially survey the floor for elevated activity and hot spots. A ratemeter/scaler connected to a GM pancake probe was used to perform further scans for beta/gamma emitters and a ZnS detector was used to scan for alpha emitters.

Locations of areas of elevated activity identified by the surface scan were marked on the surface with chalk and documented.

8.1.2 Direct measurements

One direct beta-gamma and one direct alpha surface contamination reading was taken from the center of each floor and lower wall grid in affected areas and 30 random locations in unaffected areas. Each direct measurement was one minute in duration. The Ludlum 2220 rate meter/scaler and 44-9 GM pancake probe, were used to perform the beta-gamma direct measurements, while the Ludlum 2220 rate meter/scaler and AC-3 alpha scintillation probe were used to perform the alpha direct measurements.

Additional measurements were performed in the vicinity of direct measurements that indicated the presence of residual activity exceeding 60% of the guideline value (i.e., 3000 dpm/100 cm² for beta/gamma emitters and 600 dpm/100 cm² for alpha emitters). An estimate of the elevated activity area was performed and recorded. Five additional measurements were performed at random locations within the grid. The average residual activity value reported on the survey form was calculated by:

$$x_m = \left[\frac{\sum x_i}{n} \right] \quad (2)$$

where,

x_m = mean residual activity for the 1 m² grid,

x_i = residual activity of each direct measurement,

n = number of residual activity measurements (= 5).

The elevated residual activity area was remediated if the mean residual activity for the grid exceeded the guideline value.

For the ceiling of affected areas, 32 points were surveyed. The points correlated to floor grids. At each measurement location a scan of the immediate area was performed to identify the presence of any elevated residual activity, followed by a direct measurement. No contamination was identified in the ceiling area.

8.1.3 Removable contamination measurements

A smear for removable contamination was collected at each location of direct surface activity measurement. Each smear was counted by a Ludlum 2220 rate meter/scaler with 44-9 GM pancake probe for beta-gamma measurements, a Ludlum 2220 rate meter/scaler with AC-3 alpha scintillation probe for alpha measurements, or a Canberra automated smear counter. The results of the removable surface contamination survey were documented on the appropriate survey form. Refer to the respective room survey documentation found in Attachment 1 for removable contamination survey results.

8.1.4 Dose rate measurements

Dose rate measurements were obtained at 1 meter from floor and lower wall surfaces. These measurements were performed at a frequency of one (1) systematic measurement per grid. The Bicron MicroRem meter was used to obtain the dose rate measurements.

Dose rate measurements were documented and are included with the survey documentation contained in Attachment 1. Dose rates from all areas were within background levels and ranged from 4 to 9 $\mu\text{Rem/hr}$.

8.2 Survey Instrumentation

The following table lists the instrumentation used during the final release survey, its primary use, calibration date, calibration due date and serial number. Calibration data sheets are also included in Attachment 1.

Table 6: Final Survey Instrumentation

Instrument	Use	Calibration Date	Calibration Due Date	Serial Number Meter/Probe	Efficiency
Ludlum 2220 ratemeter/scaler w/AC-3 probe	Alpha direct measurements and scan surveys	10/17/94	10/17/95	50067 / 712582	0.0667
Ludlum 2220 ratemeter/scaler w/44-9 probe	Beta-gamma direct measurements and scan surveys	10/17/94	10/17/95	52836 / PRO68918	0.239
Ludlum 2220 ratemeter/scaler w/44-9 probe	Beta-gamma direct measurements and scan surveys	10/17/94	10/17/95	50061 / PRO66761	0.238
Ludlum 2220 ratemeter/scaler w/43-37 probe	Floor monitor for beta-gamma measurements	8/20/94	8/20/95	50062 / PRO68422	0.103
Canberra Series 20 MCA w/HPGe detector	Gamma spectroscopy sample analysis	As Needed	As Needed	989997	NA
Canberra 2504	Alpha/Beta/Gamma Smear Counter	As Needed	As Needed	11924461	0.342 α 0.392 $\beta\gamma$
Bicron Micro-Rem	Dose rate surveys	10/14/94	10/14/95	B218L	NA

8.2.1 MDA's

Minimum detectable activities (MDA's) were calculated for both the hand-held survey instrumentation (e.g., Ludlum 2220 with the 44-9 GM pancake probe and Ludlum 2220 meter with the AC-3 alpha scintillation probe). The MDA was calculated by the following equation (reference 4):

$$MDA = \frac{\frac{2.71}{T_s} + 3.29 \sqrt{\frac{R_b}{T_b} + \frac{R_b}{T_s}}}{(efficiency) \left(\frac{probe\ area}{100\ cm^2} \right)}, \quad (3)$$

where,

R_b = Background counting rate (cpm),
 T_b = Background count time (min), and
 T_s = Sample count time (min).

The MDA for the Ludlum 2220 and 44-9 GM pancake probe was calculated in the same units as the fixed contamination results (i.e., dpm/100 cm²). For MDA results see attachment 1.

8.2.2 Interpretation of survey results

Direct contamination readings were converted to dpm/100 cm² using the following formula (probe area for the Ludlum 44-9 GM pancake probe was 15 cm²):

$$\text{dpm/100 cm}^2 = \frac{\text{gross cpm} - \text{background cpm}}{(\text{efficiency}) \left(\frac{\text{probe area}}{100 \text{ cm}^2} \right)} \quad (4)$$

The average background count rate for the direct survey instrumentation was determined by a series of ten 1 minute counts. Each direct measurement of fixed contamination was 1 minute in duration. If a direct measurement resulted in a value less than the calculated minimum detectable activity for the survey instrumentation, this was recorded as less than the value for the MDA (i.e, if the MDA for beta-gamma activity was 940, a value measured below MDA was recorded as < 940).

8.3 Summary of Results

The survey results presented here are a summary of the actual survey data presented in Attachment 1.

8.3.1 Unaffected areas

The following table provides a summary of the survey data from the final release survey of the unaffected areas of the Glass Plant (activity levels below MDA are reported as less than the given value for that MDA):

Table 7: Summary of Results From Unaffected Area Surveys

Location	Activity (dpm/100cm ²)				Dose (μRem/hour)
	Direct		Removable		
	βγ	α	βγ	α	
R&D Area	< 1360-3360	< 69	< 940	< 69	4-5
Process Line	< 1360	< 69	< 940	< 69	5-6
Storage Room	< 1360-2520	< 69	< 940	< 69	4-6
Furnace Area	< 1366-3420	< 69-150	< 940	< 69	4-5

Table 7a: MDA's for Unaffected Area Surveys

Location	Activity (dpm/100cm ²)			
	Direct		Removable	
	$\beta\gamma$	α	$\beta\gamma$	α
R&D Area	1360	69	940	69
Process Line	1360	69	940	69
Storage Room	1360	69	940	69
Furnace Area	1360	69	940	69

Table 7b: True Mean Activity Upper Limit at the 95% Confidence Level for the Unaffected Area Survey Units

Location	Activity (dpm/100cm ²)											
	Direct						Removable					
	$\beta\gamma$			α			$\beta\gamma$			α		
	n_s	x_{ave}	μ_α	n_s	x_{ave}	μ_α	n_s	x_{ave}	μ_α	n_s	x_{ave}	μ_α
R&D Area	30	1504	1648	30	69	69	30	940	940	30	69	69
Process Line	30	1360	1360	30	69	69	30	940	940	30	69	69
Storage Room	30	1409	1477	30	69	69	30	940	940	30	69	69
Furnace Area	30	1657	1854	30	69	69	30	940	940	30	69	69

Where: n_s = number of measurements within a survey unit
 x_{ave} = calculated mean for a survey unit
 μ_α = true mean activity upper limit at the 95% confidence level

Table 7c: Summary of Dose Rate Measurements for Unaffected Areas

Location	Dose Rate (μ Rem/hr)			
	n_s	x_{ave}	s_x	μ_α
R&D Area	30	4.8	0.4	4.9
Process Line	30	5.2	0.4	5.3
Storage Room	30	5.0	0.5	5.2
Furnace Area	30	4.9	0.3	5.0

Where: n_s = number of measurements within a survey unit
 x_{ave} = calculated mean for a survey unit
 s_x = standard deviation
 μ_α = true mean activity upper limit at the 95% confidence level

8.3.2 Affected Area

The following table provides a summary of the survey data from the final release survey of the Batch Room (activity levels below MDA are reported as less than the given value for that MDA):

Table 8: Summary of Survey Results From Affected Area (Batch Room)

Location	Activity (dpm/100cm ²)				Dose (μRem/hour)
	Direct		Removable		
	βγ	α	βγ	α	
North Wall	< 1366	< 69	< 940	< 69	5-6
South Wall	< 1366	< 69-276	< 940	< 69	6-8
West Wall	< 1366	< 69	< 940	< 69	5-6
East Wall	< 1366	< 69-120	< 940	< 69	5-9
Floor	< 1366-3506	< 69-600	< 940	< 69	5-8
Overheads	< 1366	< 69	< 940	< 69	4

Table 8a: MDA's for Affected Area Surveys

Location	Activity (dpm/100cm ²)			
	Direct		Removable	
	$\beta\gamma$	α	$\beta\gamma$	α
North Wall	1366	69	940	69
South Wall	1366	69	940	69
West Wall	1366	69	940	69
East Wall	1366	69	940	69
Floor	1366	69	940	69
Overheads	1366	69	940	69

Table 8b: True Mean Activity Upper Limit at the 95% Confidence Level for the Affected Area Survey Units

Location	Activity (dpm/100cm ²)											
	Direct						Removable					
	$\beta\gamma$			α			$\beta\gamma$			α		
	n_s	\bar{x}_{ave}	μ_α	n_s	\bar{x}_{ave}	μ_α	n_s	\bar{x}_{ave}	μ_α	n_s	\bar{x}_{ave}	μ_α
North Wall	18	1366	1366	18	69	69	18	940	940	18	69	69
South Wall	18	1366	1366	18	86	107	18	940	940	18	69	69
West Wall	42	1366	1366	42	69	69	42	940	940	42	69	69
East Wall	42	1366	1366	42	70	72	42	940	940	42	69	69
Floor	189	1451	1491	189	104	116	189	940	940	189	69	69
Overheads	32	1366	1366	32	69	69	32	940	940	32	69	69

Where: n_s = number of measurements within a survey unit
 \bar{x}_{ave} = calculated mean for a survey unit
 μ_α = true mean activity upper limit at the 95% confidence level

Table 8c: Summary of Dose Rate Measurements for Affected Area

Location	Dose Rate (μ Rem/hr)			
	n_s	\bar{x}_{ave}	s_x	μ_α
North Wall	18	5.2	0.4	5.3
South Wall	18	6.6	0.6	6.9
West Wall	42	5.3	0.5	5.4
East Wall	42	7.2	1.2	7.6
Floor	189	5.8	0.8	5.9
Overheads	32	4.0	0.0	4.0

Where: n_s = number of measurements within a survey unit
 \bar{x}_{ave} = calculated mean for a survey unit
 s_x = standard deviation
 μ_α = true mean activity upper limit at the 95% confidence level

9. REFERENCES

1. USNRC "Guidelines for Decontamination of Facilities and Equipment Prior to Release for Unrestricted Use or Termination of Licenses for Byproduct, Source or Special Nuclear Material," May 1987.
2. NUREG/CR-5849. "Manual for Conducting Radiological Surveys in Support of License Termination," 1992.
3. NES Procedure 82A8006, "Radiation Worker Handbook and Training Manual."
4. Strom, Daniel J. And Stansbury, Paul S.; "Minimum Detectable Activity When Background is Counted Longer than the Sample," Health Physics 63(3):360-361, 1992.
5. New York Department of Labor Industrial Code Rule 38, "Ionizing Radiation Protection," June 29, 1994.17

Attachment 1

Survey Sheets

ATTACHMENT CONTENTS

Section 1 - Initial Surveys

- Bird Cage
- South Hood
- South Hood Sifter
- Middle Hood

Section 2 - Unaffected Area Surveys

- R&D Area
- Process Line
- Storage Room
- Furnace Area

Section 3 - Affected Area Surveys

- Batch Room North Wall
- Batch Room South Wall
- Batch Room West Wall
- Batch Room East Wall
- Batch Room Floor
- Batch Room Overheads
- Batch Room Miscellaneous Equipment
 - Hood #1
 - Hood #2
 - Mixer #1
 - Mixer #2
- Batch Room Scale
- Batch Room Hopper
- Batch Room Bird Cage

Section 4 - Batch Samples

- Analysis of 55 Gallon Drum and B-25 Box Contents

Section 5 - Calibration Documents

Section 6 - Air Sample Results

Section 7 - Waste Container Surveys

Section 1

Initial Surveys

Bird Cage
South Hood
South Hood Sifter
Middle Hood

MODEL Bicron
SERIAL # BZ18L
CAL DUE 10-14-95
EFFICIENCY NA
TYPE Micro Rem
BKG 5 mRem/hr
MDA NA

5'

HOOPER

* <69α
* 4458B

FILL PORT

#5
* <69α
* 541B

* <69α #5
* 5182B

* <69α
* 8347B

* <69α
* 7647B

under pallet
* <69α * <69α
* <1360B * <1360B

WOODEN PALLET

UNDER PAL

* <69α #5
* 1652B * <69α
* <1360B

* <69α
* <1360B

* <69α
* 1821B

* <69α #5
* <1360B

PLATFORM

N

Background has been subtracted from fixed readings

Background has been subtracted from fixed readings

NOTE: Smear results in dpm/100 cm² unless otherwise noted. (1) denotes smear location. An * followed by a number in dpm (i.e., * 100 dpm) denotes direct probe readings in dpm/100 cm². # denotes dose rates in ~~mR/hr~~ *mrem/hr @ 1m*

PURPOSE: Initial Survey

[illegible]

SURVEYOR D. L. Vann
 REVIEWER B. D. D.

DATE 12/7/94
 TIME 1030


SURVEY METERS

MODEL L-2220
 SERIAL # 52836
 CAL DUE 10/17/95
 EFFICIENCY 0.239
 TYPE 44-9
 BKG 98
 MDA 1360 dpm/100cm²

MODEL L-2220
 SERIAL # 50067
 CAL DUE 10/17/95
 EFFICIENCY 0.0667
 TYPE AC-3
 BKG 0
 MDA 69 dpm/100 cm²

MODEL Bicron
 SERIAL # B218L
 CAL DUE 10/14/95
 EFFICIENCY NA
 TYPE Micro Rem
 BKG 5 nRem/hr
 MDA NA

LOCATION: South Hood, Bausch & Lomb



NOTE: Smear results in dpm/100 cm² unless otherwise noted. (1) denotes smear location. An * followed by a number in dpm (i.e., * 100 dpm) denotes direct probe readings in dpm/100 cm². # denotes dose rates in ~~mR/hr~~ nRem/hr @ 1m. indicates survey performed on roof.

PURPOSE: Initial Survey

SMEAR	dpm/100 cm ²	
#	By	α
34	< 32	< 8
35	< 32	< 8
36	< 32	< 8
37	< 32	< 8
38	< 32	< 8
39	< 32	< 8
40	< 32	< 8
41	< 32	< 8
42	< 32	< 8
Surveys performed on Canberra CAL DUE 1-1-95		
2 eff.	34.20	%
8/8 eff.	39.40	%
2 Bkg	0 cpm	
8/8 Bkg	8 cpm	
2 MDA	8 dpm/100 cm ²	
8/8 MDA	32 dpm/100 cm ²	
NA		

~~MODEL _____~~
~~SERIAL # _____~~
~~CAL DUE _____~~
~~EFFICIENCY _____~~
~~TYPE NA _____~~
~~BKG _____~~
~~MDA _____~~

Hand-drawn diagram of a two-tiered cylindrical structure. The top tier is labeled with a circled '33' and the text '* < 69 x' and '* < 1360 B'. A vertical line with a small circle at the top passes through the center of the top tier. The bottom tier is labeled with a circled '32' and the text '* 1050 x' and '* 19,700 B'. The bottom tier is also labeled with a circled '31' and the text '* 1590 x' and '* 39,900 B'. The bottom tier is shaded with diagonal lines.

PURPOSE: Initial Survey

[illegible]


SURVEYOR D. L. Morris for Kevin Graczyk DATE 12-7-94
 REVIEWER B. D. D. TIME 1125

SURVEY METERS

MODEL Ludlum 2220
 SERIAL # 52836
 CAL DUE 10/17/95
 EFFICIENCY 0.239
 TYPE 44-9
 BKG 98 cpm
 MDA 1360 dpm/100cm²

MODEL Ludlum 2220
 SERIAL # 50067
 CAL DUE 10/17/95
 EFFICIENCY 0.0667
 TYPE AC-3
 BKG 0 cpm
 MDA 69 dpm/100cm²

MODEL Bicron
 SERIAL # B218L
 CAL DUE 10/14/95
 EFFICIENCY NA
 TYPE Micro Rem
 BKG 5 nRem/hr
 MDA NA

LOCATION: Bausch & Lomb middle Hood


NOTE: Smear results in dpm/100 cm² unless otherwise noted. (1) denotes smear location. An * followed by a number in dpm (i.e., * 100 dpm) denotes direct probe readings in dpm/100 cm². # denotes dose rates in ~~nRm~~ nRem/hr @ 1m

PURPOSE: Initial Survey

SMEAR	dpm/100 cm ²	
#	By	α
1	< 32	11
2	< 32	9
3	< 32	< 8
4	< 32	< 8
5	< 32	< 8
6	< 32	< 8
7	< 32	< 8
8	< 32	26
9	< 32	< 8
10	< 32	14
11	< 32	< 8
12	< 32	< 8
13	< 32	< 8
14	< 32	< 8
15	< 32	< 8
16	< 32	49
17	< 32	< 8
18	< 32	< 8
19	< 32	< 8
20	< 32	< 8

SURVEYOR D.L. Vann for Kevin Graczyk DATE 12-7-94
 REVIEWER _____ TIME 1125

SURVEY METERS

MODEL Ludlum 2220SERIAL # 52836CAL DUE 10/17/95EFFICIENCY 0.239TYPE 44-9BKG 98 cpmMDA 1360 dpm/100cm²MODEL Ludlum 2220SERIAL # 50067CAL DUE 10/17/95EFFICIENCY 0.0667TYPE AC-3BKG 0 cpmMDA 69 dpm/100cm²MODEL BicronSERIAL # B2/8LCAL DUE 10/14/95EFFICIENCY NATYPE Micro RemBKG 5 nRem/hrMDA NALOCATION: Bausch + Lomb middle Hood CONTINUOUS

SMEAR	dpm/100 cm ²	
#	By	α
21	< 32	< 8
22	< 32	< 8
23	< 32	9
24	< 32	< 8
25	< 32	< 8
26	51	43
27	< 32	35
28	< 32	29
29	89	52
30	< 32	9
Smear Survey counted		
on Canberra 11924461		
Δ eff.	34.20	%
Δ Bkg.	0 cpm	CAL DUE 1-1-95
B eff.	39.40	%
B Bkg	8 cpm	
Δ MDA	8 dpm/100 cm ²	
B MDA	32 dpm/100 cm ²	
	N	A

NOTE: Smear results in dpm/100 cm² unless otherwise noted. (1) denotes smear location. An * followed by a number in dpm (i.e., * 100 dpm) denotes direct probe readings in dpm/100 cm². # denotes dose rates in mR/hr.

PURPOSE: Initial Survey

MDA CALCULATION SHEET

METER 2220 ^W/44-9 SERIAL #: 52836TS: 1 TB: 1 RB: 98cpm EFF: 0.239 PROBE SIZE: 15 cm²

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{98}{1} + \frac{98}{1} \right)^{1/2}$$

$$\frac{0.239}{1} \left(\frac{15}{100} \text{ cm}^2 \right)$$

$$MDA = \underline{1360 \text{ dpm}/100 \text{ cm}^2}$$

METER 2220 ^W/AC-3 SERIAL #: 50067TS: 1 TB: 1 RB: 0cpm EFF: 0.0667 PROBE SIZE: 59 cm²

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{0}{1} + \frac{0}{1} \right)^{1/2}$$

$$\frac{0.0667}{1} \left(\frac{59}{100} \text{ cm}^2 \right)$$

$$MDA = \underline{69 \text{ dpm}/100 \text{ cm}^2}$$

METER CAUDERA SERIAL #: 11924461TS: 1 TB: 10 RB: 0cpm EFF: 0.392 PROBE SIZE: N/A

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{0}{1} + \frac{0}{10} \right)^{1/2}$$

$$\frac{0.392}{1} \left(\frac{N/A}{100} \text{ cm}^2 \right)$$

$$\alpha \quad MDA = \underline{8 \text{ dpm}/100 \text{ cm}^2}$$

METER CAUDERA SERIAL #: 11924461TS: 1 TB: 10 RB: 8cpm EFF: 0.392 PROBE SIZE: N/A

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{8}{1} + \frac{8}{10} \right)^{1/2}$$

$$\frac{0.392}{1} \left(\frac{N/A}{100} \text{ cm}^2 \right)$$

$$\beta \quad MDA = \underline{32 \text{ dpm}/100 \text{ cm}^2}$$

Technician: Brenden Guber Date: 12-7-94Reviewed by: D. L. Vann Date: 12-7-94

Section 2

Unaffected Area Surveys

**R&D Area
Process Line
Storage Room
Furnace Area**

* KNO_3 containing K-40 was removed from the area and brought to the batch room.

REMOVABLE CONTAMINATION

DATE: 12/7/94LOCATION: Bausch + Lomb R+D AreaCOUNTER: Ludlum 2220 S#50067
Model NumberEFFICIENCY: 0.0667 N/A
Alpha Beta/Gamma

DPM/100cm ²			DPM/100cm ²			DPM/100cm ²		
LOCATION	βγ	α	LOCATION	βγ	α	LOCATION	βγ	α
1		< 69	21		< 69			
2		< 69	22		< 69			
3		< 69	23		< 69			
4		< 69	24		< 69			
5		< 69	25		< 69			
6		< 69	26	N/A	< 69			
7		< 69	27		< 69			
8		< 69	28		< 69			
9		< 69	29		< 69			
10		< 69	30		< 69			
11	N/A	< 69					N/A	
12		< 69						
13		< 69						
14		< 69						
15		< 69		N/A				
16		< 69						
17		< 69						
18		< 69						
19		< 69						
20		< 69						

ALPHA BKG COUNT RATE: 0 cpmBETA/GAMMA BKG COUNT RATE: N/Aα MDA (dpm/100cm²) 69βγ MDA (dpm/100cm²) N/ASURVEYOR D.L. Vann for Kevin GraczykSUPERVISOR B. Guler

REMOVABLE CONTAMINATION

DATE: 12-7-94LOCATION: Bausch + Lomb R + DCOUNTER: Ludlum 222050061 DLU
~~53856~~
Number 1-2-16EFFICIENCY: N/A

0,239

~~0,238~~ DLU 1-2-96

Model

Number

Alpha

Beta/Gamma

DPM/100cm ²			DPM/100cm ²			DPM/100cm ²		
LOCATION	By	α	LOCATION	By	α	LOCATION	By	α
1	< 940		21	< 940				
2	< 940		22	< 940				
3	< 940		23	< 940				
4	< 940		24	< 940				
5	< 940		25	< 940	N			
6	< 940		26	< 940	A			
7	< 940		27	< 940				
8	< 940		28	< 940				
9	< 940		29	< 940				
10	< 940	N	30	< 940			N	
11	< 940	A					A	
12	< 940							
13	< 940							
14	< 940							
15	< 940				N			
16	< 940				A			
17	< 940							
18	< 940							
19	< 940							
20	< 940							

ALPHA BKG COUNT RATE: N/Aα MDA (dpm/100cm²)N/ABy MDA (dpm/100cm²)940BETA/GAMMA BKG COUNT RATE: 44 cpm

SURVEYOR

P. L. Ugon

SUPERVISOR

Bruno Lab

AVERAGE CONTAMINATION SURVEY RESULTS

DATE: 12/7/94 LOCATION: Bausch & Lomb R&D
 INSTRUMENT: Ludlum 2220 S# 50067 PROBE: AC-3 EFF: 0.0667
 MODEL NUMBER MODEL

GRID NUMBER	dpm/100cm ²	GRID NUMBER	dpm/100cm ²
1	< 69 ^{BUG 1-13-15} 69	20	< 69
2	< 69	21	< 69
3	< 69	22	< 69
4	< 69	23	< 69
5	< 69	24	< 69
6	< 69	25	< 69
7	< 69	26	< 69
8	< 69	27	< 69
9	< 69	28	< 69
10	< 69	29	< 69
11	< 69	30	< 69
12	< 69		
13	< 69		
14	< 69		
15	< 69		
16	< 69		
17	< 69		
18	< 69		
19	< 69		

Average Background counts per minute 0 cpm
 MDA (dpm/100cm²) 69

*BASED ON AVERAGE OF 5 DIRECT MEASUREMENTS
 (INITIAL DIRECT MEASUREMENT WAS GREATER THAN 60% OF GUIDELINE VALUE)

$\frac{\text{dpm}}{100\text{cm}^2} = \frac{(\text{gross count rate}) - (\text{background count rate})}{(\text{efficiency}) \left(\frac{\text{probe area}}{100\text{cm}^2} \right)}$

SURVEYOR D. L. Vroom for Kevin Graczyk
 SUPERVISOR B. Dule

AVERAGE CONTAMINATION SURVEY RESULTS

Page 5 of 7

DATE: 12/7/94 LOCATION: Bunsch & Lumb R&D Area
 INSTRUMENT: Ludlum 2220 S# 52836 PROBE: 44-9 EFF: 0.239
 MODEL NUMBER MODEL

GRID NUMBER	dpm/100cm ²	GRID NUMBER	dpm/100cm ²
1	< 1360	20	< 1360
2	2800	21	< 1360
3	< 1360	22	< 1360
4	< 1360	23	3360
5	< 1360	24	< 1360
6	< 1360	25	< 1360
7	< 1360	26	< 1360
8	< 1360	27	< 1360
9	< 1360	28	< 1360
10	< 1360	29	< 1360
11	< 1360	30	< 1360
12	< 1360		
13	< 1360		
14	< 1360		
15	< 1360		
16	< 1360		
17	< 1360		
18	< 1360		
19	2240		

Average Background counts per minute 98
 MDA (dpm/100cm²) 1360

*BASED ON AVERAGE OF 5 DIRECT MEASUREMENTS
 (INITIAL DIRECT MEASUREMENT WAS GREATER THAN 60% OF GUIDELINE VALUE)

$$\frac{\text{dpm}}{100\text{cm}^2} = \frac{(\text{gross count rate}) - (\text{background count rate})}{(\text{efficiency}) \quad (\text{probe area})}$$

100cm²

D. L. Vann for
 K. Granger
 SURVEYOR
 SUPERVISOR B. Granger

EXPOSURE RATE SURVEY RESULTS

DATE: 12-7-94LOCATION: Bruckh & Lomb R&D AREAINSTRUMENT: Bicron B218LCAL DUE: ^{BIG 1-13-95}
~~10-14-95~~ 10-14-95

GRID NUMBER	EXPOSURE RATE $\mu\text{R/h}$		GRID NUMBER	EXPOSURE RATE $\mu\text{R/h}$	
	1cm	1m		1cm	1m
1		5 uR/hr	18		5 uR/hr
2		5 uR/hr	19		5 uR/hr
3		5 uR/hr	20		5 uR/hr
4		5 uR/hr	21		5 uR/hr
5		5 uR/hr	22		5 uR/hr
6		4 uR/hr	23		5 uR/hr
7		4 uR/hr	24	N A	5 uR/hr
8		4 uR/hr	25		5 uR/hr
9	N A	4 uR/hr	26		5 uR/hr
10		5 uR/hr	27		5 uR/hr
11		5 uR/hr	28		4 uR/hr
12		5 uR/hr	29		4 uR/hr ₁₀₀
13		5 uR/hr	30		4 uR/hr
14		5 uR/hr			
15		5 uR/hr		N	
16		5 uR/hr		A	
17		5 uR/hr			

MEASUREMENTS TAKEN AT 1cm AND/OR 1m ABOVE THE FLOOR.

BKG: 5 uR/hr

SURVEYOR

SUPERVISOR

Kir R. Dwyer
Bruce L. Lutz

MDA CALCULATION SHEET

METER 2220 w/44-9 SERIAL #: 52836TS: 1 TB: 1 RB: 98 cpm EFF: 0.239 PROBE SIZE: 15 cm²

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{98}{1} + \frac{98}{1} \right)^{1/2}$$

$$0.239 \quad (15 / 100 \text{ cm}^2)$$

$$MDA = \underline{1360 \text{ dpm}/100 \text{ cm}^2}$$

METER 2220 w/Ac-3 SERIAL #: 50067TS: 1 TB: 1 RB: 0 cpm EFF: 0.0667 PROBE SIZE: 59 cm²

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{0}{1} + \frac{0}{1} \right)^{1/2}$$

$$0.0667 \quad (59 / 100 \text{ cm}^2)$$

$$MDA = \underline{69 \text{ dpm}/100 \text{ cm}^2}$$

METER 2220 w/44-9 SERIAL #: 50061TS: 1 TB: 1 RB: 44 cpm EFF: 0.238 PROBE SIZE: 15 cm²

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{44}{1} + \frac{44}{1} \right)^{1/2}$$

$$0.238 \quad (15 / 100 \text{ cm}^2)$$

$$MDA = \underline{940 \text{ dpm}/100 \text{ cm}^2}$$

Removable
Only

METER _____ SERIAL #: _____

TS: _____ TB: _____ RB: _____ EFF: _____ PROBE SIZE: _____

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{\quad}{1} + \frac{\quad}{1} \right)^{1/2}$$

$$(\quad / 100 \text{ cm}^2)$$

$$MDA = \underline{\quad}$$

Technician: D. L. Vann Date: 12/7/94Reviewed by: B. Gule Date: 12/7/94

Bausch & Lomb Release Survey Data

Survey Unit: R&D Area

Activity (dpm/100cm²)

<u>Location Number</u>	Removable α			Removable β - γ		
	<u>Activity</u>	<u>Uncertainty *</u>	<u>MDA</u>	<u>Activity</u>	<u>Uncertainty *</u>	<u>MDA</u>
1	69	16	69	940	64	940
2	69	16	69	940	64	940
3	69	16	69	940	64	940
4	69	16	69	940	64	940
5	69	16	69	940	64	940
6	69	16	69	940	64	940
7	69	16	69	940	64	940
8	69	16	69	940	64	940
9	69	16	69	940	64	940
10	69	16	69	940	64	940
11	69	16	69	940	64	940
12	69	16	69	940	64	940
13	69	16	69	940	64	940
14	69	16	69	940	64	940
15	69	16	69	940	64	940
16	69	16	69	940	64	940
17	69	16	69	940	64	940
18	69	16	69	940	64	940
19	69	16	69	940	64	940
20	69	16	69	940	64	940
21	69	16	69	940	64	940
22	69	16	69	940	64	940
23	69	16	69	940	64	940
24	69	16	69	940	64	940
25	69	16	69	940	64	940
26	69	16	69	940	64	940
27	69	16	69	940	64	940
28	69	16	69	940	64	940
29	69	16	69	940	64	940
30	69	16	69	940	64	940

* Uncertainties represent the 95% Confidence level, based on counting statistics.
When survey data indicated less than MDA, the MDA value was used in the calculation.

Bausch & Lomb Release Survey Data

Survey Unit: R&D Area

Activity (dpm/100cm²)

Location Number	Direct α			Direct β - γ		
	Activity	Uncertainty *	MDA	Activity	Uncertainty *	MDA
1	69	16	69	1360	79	1360
2	69	16	69	2800	109	1360
3	69	16	69	1360	79	1360
4	69	16	69	1360	79	1360
5	69	16	69	1360	79	1360
6	69	16	69	1360	79	1360
7	69	16	69	1360	79	1360
8	69	16	69	1360	79	1360
9	69	16	69	1360	79	1360
10	69	16	69	1360	79	1360
11	69	16	69	1360	79	1360
12	69	16	69	1360	79	1360
13	69	16	69	1360	79	1360
14	69	16	69	1360	79	1360
15	69	16	69	1360	79	1360
16	69	16	69	1360	79	1360
17	69	16	69	1360	79	1360
18	69	16	69	1360	79	1360
19	69	16	69	2240	98	1360
20	69	16	69	1360	79	1360
21	69	16	69	1360	79	1360
22	69	16	69	1360	79	1360
23	69	16	69	3360	118	1360
24	69	16	69	1360	79	1360
25	69	16	69	1360	79	1360
26	69	16	69	1360	79	1360
27	69	16	69	1360	79	1360
28	69	16	69	1360	79	1360
29	69	16	69	1360	79	1360
30	69	16	69	1360	79	1360

* Uncertainties represent the 95% Confidence level, based on counting statistics.
When survey data indicated less than MDA, the MDA value was used in the calculation.

NINETY-FIVE PERCENT CONFIDENCE CALCULATION SHEET

Bausch + Lomb

Survey Unit R+D Area Date 12-7-94

Meter Endium 2220 Serial # 50067

Probe AC-3 Serial # 712582

MDA 69 dpm/100cm² Survey Type Removable α

Guideline Value 200 (dpm/100cm²)

Average Measurement Level

$$x_{ave} = \frac{1}{n_s} \sum_{i=1}^{n_s} x_i$$

where:

x_{ave} = calculated mean for a survey unit,
 n_s = number of measurements within a survey unit,
 x_i = systematic and random measurements at point (i)

$\sum x_i$ 2070

n_s 30

x_{ave} = 69

Standard Deviation

$$S_x = \sqrt{\frac{\sum_{i=1}^{n_s} (x_{ave} - x_i)^2}{n_s - 1}}$$

where:

S_x = standard deviation
 n_s = number of measurements within a survey unit
 x_i = systematic and random measurements at point (i)
 x_{ave} = calculated mean for a survey unit

x_{ave} = 69 n_s = 30

$\sum (x_{ave} - x_i)^2$ = 0 S_x = 0

$$\mu_{\alpha} = x_{ave} + t_{1-\alpha, df} \frac{s_x}{\sqrt{n_s}}$$

where:

μ_{α} = value compared guideline value to determine 95% Confidence Level
 x_{ave} = calculated mean for a survey unit,
 $t_{1-\alpha, df}$ = 95% confidence level for n-1 degrees of freedom (from NUREG/CR-5849, Table B-1, "Factors for Comparison of Survey Data")
 s_x = standard deviation of measurements in a survey unit, and
 n_s = number of measurements within a survey unit used to determine x_{ave} and s_x .

x_{ave} = 69
 $t_{1-\alpha, df}$ = 1.699
 s_x = 0
 n_s = 30
 μ_{α} = 69

NOTE: When determining x_{ave} , only minimum detectable activity (MDA) values were available for some measurements locations; the MDA values were therefore used as actual activity levels for the purpose of performing this calculation.

The site specific guideline values for this project are 5000 dpm/100 cm² fixed plus removable and 1000 dpm/100 cm² removable for beta-gamma emitting nuclides, and 1000 dpm/100 cm² fixed plus removable and 200 for alpha emitting nuclides. If μ_{α} is less than the specific guideline value, the data for this survey unit satisfies the guideline at the 95% confidence level. If μ_{α} is greater than the specific guideline value, further remediation of the survey unit is necessary.

Individual Completing Form: D. L. Vann Date: 12-7-94

Reviewed By: R. K. H. Date: 1/10/95

NINETY-FIVE PERCENT CONFIDENCE CALCULATION SHEET

Survey Unit Bausch & Lomb R & D Area Date 12-7-94

Meter Ludlum 2220 Serial # 50067

Probe AC-3 Serial # 712582

MDA 69 dpm/100cm² Survey Type Direct α

Guideline Value 1000 (dpm/100cm²)

Average Measurement Level

$$x_{ave} = \frac{1}{n_s} \sum_{i=1}^{n_s} x_i$$

where:

x_{ave} = calculated mean for a survey unit,
 n_s = number of measurements within a survey unit,
 x_i = systematic and random measurements at point (i)

Σx_i 2070

n_s 30

x_{ave} = 69

Standard Deviation

$$S_x = \sqrt{\frac{\sum_{i=1}^{n_s} (x_{ave} - x_i)^2}{n_s - 1}}$$

where:

S_x = standard deviation
 n_s = number of measurements within a survey unit
 x_i = systematic and random measurements at point (i)
 x_{ave} = calculated mean for a survey unit

x_{ave} = 69 n_s = 30

$\Sigma (x_{ave} - x_i)^2$ = 0 S_x = 0

$$\mu_{\alpha} = x_{ave} + t_{1-\alpha, df} \frac{s_x}{\sqrt{n_s}}$$

where:

μ_{α} = value compared guideline value to determine 95 % Confidence Level
 x_{ave} = calculated mean for a survey unit,
 $t_{1-\alpha, df}$ = 95 % confidence level for n-1 degrees of freedom (from NUREG/CR-5849, Table B-1, "Factors for Comparison of Survey Data")
 s_x = standard deviation of measurements in a survey unit, and
 n_s = number of measurements within a survey unit used to determine x_{ave} and s_x .

x_{ave} = 69
 $t_{1-\alpha, df}$ = 1.699
 s_x = 0
 n_s = 30
 μ_{α} = 69

NOTE: When determining x_{ave} , only minimum detectable activity (MDA) values were available for some measurements locations; the MDA values were therefore used as actual activity levels for the purpose of performing this calculation.

The site specific guideline values for this project are 5000 dpm/100 cm² fixed plus removable and 1000 dpm/100 cm² removable for beta-gamma emitting nuclides, and 1000 dpm/100 cm² fixed plus removable and 200 for alpha emitting nuclides. If μ_{α} is less than the specific guideline value, the data for this survey unit satisfies the guideline at the 95 % confidence level. If μ_{α} is greater than the specific guideline value, further remediation of the survey unit is necessary.

Individual Completing Form: O. L. Vann Date: 12-7-94

Reviewed By: R. B. Jett Date: 1/10/95

NINETY-FIVE PERCENT CONFIDENCE CALCULATION SHEET

Survey Unit Bausch + Lomb R+D Area Date 12-7-94

Meter Lndlum 2220 Serial # 50061

Probe 44-9 Serial # PR066761

MDA 940 dpm/100cm² Survey Type Removable B-Y

Guideline Value 1000 (dpm/100cm²)

Average Measurement Level

$$x_{ave} = \frac{1}{n_s} \sum_{i=1}^{n_s} x_i$$

where:

x_{ave} = calculated mean for a survey unit,
 n_s = number of measurements within a survey unit,
 x_i = systematic and random measurements at point (i)

$\sum x_i$ 28,200

n_s 30

x_{ave} = 940

Standard Deviation

$$S_x = \sqrt{\frac{\sum_{i=1}^{n_s} (x_{ave} - x_i)^2}{n_s - 1}}$$

where:

S_x = standard deviation
 n_s = number of measurements within a survey unit
 x_i = systematic and random measurements at point (i)
 x_{ave} = calculated mean for a survey unit

x_{ave} = 940 n_s = 30

$\sum (x_{ave} - x_i)^2$ = 0 S_x = 0

$$\mu_{\alpha} = x_{ave} + t_{1-\alpha, df} \frac{s_x}{\sqrt{n_s}}$$

where:

μ_{α} = value compared guideline value to determine 95 % Confidence Level
 x_{ave} = calculated mean for a survey unit,
 $t_{1-\alpha, df}$ = 95 % confidence level for n-1 degrees of freedom (from NUREG/CR-5849, Table B-1, "Factors for Comparison of Survey Data")
 s_x = standard deviation of measurements in a survey unit, and
 n_s = number of measurements within a survey unit used to determine x_{ave} and s_x .

$$x_{ave} = \underline{940}$$

$$t_{1-\alpha, df} = \underline{1.699}$$

$$s_x = \underline{0}$$

$$n_s = \underline{30}$$

$$\mu_{\alpha} = \underline{940}$$

NOTE: When determining x_{ave} , only minimum detectable activity (MDA) values were available for some measurements locations; the MDA values were therefore used as actual activity levels for the purpose of performing this calculation.

The site specific guideline values for this project are 5000 dpm/100 cm² fixed plus removable and 1000 dpm/100 cm² removable for beta-gamma emitting nuclides, and 1000 dpm/100 cm² fixed plus removable and 200 for alpha emitting nuclides. If μ_{α} is less than the specific guideline value, the data for this survey unit satisfies the guideline at the 95 % confidence level. If μ_{α} is greater than the specific guideline value, further remediation of the survey unit is necessary.

Individual Completing Form: D. L. Vann Date: 12-7-94

Reviewed By: R. Kuyt Date: 1/10/95

NINETY-FIVE PERCENT CONFIDENCE CALCULATION SHEET

Ban Sch + Lumb

Survey Unit R+D Area Date 12-7-94

Meter Lundum 2220 Serial # 52836

Probe 44-9 Serial # PRO 68918

MDA 1360 dpm/100cm² Survey Type Direct B-2

Guideline Value 5000 (dpm/100cm²)

Average Measurement Level

$$x_{ave} = \frac{1}{n_s} \sum_{i=1}^{n_s} x_i$$

where:

x_{ave} = calculated mean for a survey unit,
 n_s = number of measurements within a survey unit,
 x_i = systematic and random measurements at point (i)

Σx_i 45,120

n_s 30

x_{ave} = 1504

Standard Deviation

$$S_x = \sqrt{\frac{\sum_{i=1}^{n_s} (x_{ave} - x_i)^2}{n_s - 1}}$$

where:

S_x = standard deviation
 n_s = number of measurements within a survey unit
 x_i = systematic and random measurements at point (i)
 x_{ave} = calculated mean for a survey unit

x_{ave} = 1504 n_s = 30

$\Sigma (x_{ave} - x_i)^2$ = 6,225,920 S_x = 463,3

$$\mu_{\alpha} = x_{ave} + t_{1-\alpha, df} \frac{s_x}{\sqrt{n_s}}$$

where:

μ_{α} = value compared guideline value to determine 95 % Confidence Level
 x_{ave} = calculated mean for a survey unit,
 $t_{1-\alpha, df}$ = 95 % confidence level for n-1 degrees of freedom (from NUREG/CR-5849, Table B-1, "Factors for Comparison of Survey Data")
 s_x = standard deviation of measurements in a survey unit, and
 n_s = number of measurements within a survey unit used to determine x_{ave} and s_x .

$$x_{ave} = 1504$$

$$t_{1-\alpha, df} = 1.699$$

$$s_x = 463.3$$

$$n_s = 30$$

$$\mu_{\alpha} = 1647.7$$

NOTE: When determining x_{ave} , only minimum detectable activity (MDA) values were available for some measurements locations; the MDA values were therefore used as actual activity levels for the purpose of performing this calculation.

The site specific guideline values for this project are 5000 dpm/100 cm² fixed plus removable and 1000 dpm/100 cm² removable for beta-gamma emitting nuclides, and 1000 dpm/100 cm² fixed plus removable and 200 for alpha emitting nuclides. If μ_{α} is less than the specific guideline value, the data for this survey unit satisfies the guideline at the 95 % confidence level. If μ_{α} is greater than the specific guideline value, further remediation of the survey unit is necessary.

Individual Completing Form: D. L. Vann Date: 12-7-94

Reviewed By: R. Kipt Date: 1/10/95

NINETY-FIVE PERCENT CONFIDENCE CALCULATION SHEET

Survey Unit Bansch + Lomb R + D Area Date 12-7-94

Meter Bicron MicroRem Serial # BZ18L

Probe NA Serial # NA

MDA NA Survey Type Exposure

Guideline Value* 10 mRem/hr (~~dpm/100cm²~~) BIG 1-13-95

*includes BKG

Average Measurement Level

$$x_{ave} = \frac{1}{n_s} \sum_{i=1}^{n_s} x_i$$

where:

x_{ave} = calculated mean for a survey unit,
 n_s = number of measurements within a survey unit,
 x_i = systematic and random measurements at point (i)

$\sum x_i$ 143

n_s 30

x_{ave} = 4.8

Standard Deviation

$$S_x = \sqrt{\frac{\sum_{i=1}^{n_s} (x_{ave} - x_i)^2}{n_s - 1}}$$

where:

S_x = standard deviation
 n_s = number of measurements within a survey unit
 x_i = systematic and random measurements at point (i)
 x_{ave} = calculated mean for a survey unit

x_{ave} = 4.8 n_s = 30

$\sum (x_{ave} - x_i)^2$ = 5.4 S_x = 0.4

$$\mu_{\alpha} = x_{ave} + t_{1-\alpha, df} \frac{s_x}{\sqrt{n_s}}$$

where:

μ_{α} = value compared guideline value to determine 95% Confidence Level
 x_{ave} = calculated mean for a survey unit,
 $t_{1-\alpha, df}$ = 95% confidence level for n-1 degrees of freedom (from NUREG/CR-5849, Table B-1, "Factors for Comparison of Survey Data")
 s_x = standard deviation of measurements in a survey unit, and
 n_s = number of measurements within a survey unit used to determine x_{ave} and s_x .

$$x_{ave} = 4.8$$

$$t_{1-\alpha, df} = 1.699$$

$$s_x = 0.4$$

$$n_s = 30$$

$$\mu_{\alpha} = 4.9$$

NOTE: When determining x_{ave} , only minimum detectable activity (MDA) values were available for some measurements locations; the MDA values were therefore used as actual activity levels for the purpose of performing this calculation.

The site specific guideline values for this project are 5000 dpm/100 cm² fixed plus removable and 1000 dpm/100 cm² removable for beta-gamma emitting nuclides, and 1000 dpm/100 cm² fixed plus removable and 200 for alpha emitting nuclides. If μ_{α} is less than the specific guideline value, the data for this survey unit satisfies the guideline at the 95% confidence level. If μ_{α} is greater than the specific guideline value, further remediation of the survey unit is necessary.

Individual Completing Form: D. L. Vann Date: 12-7-94

Reviewed By: R. Bight Date: 1/10/95

[illegible]

AVERAGE CONTAMINATION SURVEY RESULTS

Page 2 of 7

DATE: 12/7/94 LOCATION: Bausch + Lomb Process Line
 INSTRUMENT: Endium 2220 S# 50067 PROBE: AC-3 EFF: 0.0667
 MODEL NUMBER MODEL

GRID NUMBER	dpm/100cm ²	GRID NUMBER	dpm/100cm ²
1	< 69	20	< 69
2	< 69	21	< 69
3	< 69	22	< 69
4	< 69	23	< 69
5	< 69	24	< 69
6	< 69	25	< 69
7	< 69	26	< 69
8	< 69	27	< 69
9	< 69	28	< 69
10	< 69	29	< 69
11	< 69	30	< 69
12	< 69		
13	< 69		
14	< 69		
15	< 69		
16	< 69		
17	< 69		
18	< 69		
19	< 69		

Average Background counts per minute 0
 MDA (dpm/100cm²) 69

*BASED ON AVERAGE OF 5 DIRECT MEASUREMENTS
 (INITIAL DIRECT MEASUREMENT WAS GREATER THAN 60% OF GUIDELINE VALUE)

$$\frac{\text{dpm}}{100\text{cm}^2} = \frac{(\text{gross count rate}) - (\text{background count rate})}{(\text{efficiency}) \left(\frac{\text{probe area}}{100\text{cm}^2} \right)}$$

D. L. Vann for
 SURVEYOR Kevin Graczyk
 SUPERVISOR B. Gule

AVERAGE CONTAMINATION SURVEY RESULTS

Page 5 of 7

DATE: 12/7/94 LOCATION: Bausch & Lomb Process Line
 INSTRUMENT: Ludlum 2220 S# 52836 PROBE: 44-9 EFF: 0.239
 MODEL NUMBER MODEL

GRID NUMBER	dpm/100cm ²	GRID NUMBER	dpm/100cm ²
1	< 1360	20	< 1360
2	< 1360	21	< 1360
3	< 1360	22	< 1360
4	< 1360	23	< 1360
5	< 1360	24	< 1360
6	< 1360	25	< 1360
7	< 1360	26	< 1360
8	< 1360	27	< 1360
9	< 1360	28	< 1360
10	< 1360	29	< 1360
11	< 1360	30	< 1360
12	< 1360		
13	< 1360		
14	< 1360		
15	< 1360		
16	< 1360		
17	< 1360		
18	< 1360		
19	< 1360		

Average Background counts per minute 98
 MDA (dpm/100cm²) 1360

*BASED ON AVERAGE OF 5 DIRECT MEASUREMENTS
 (INITIAL DIRECT MEASUREMENT WAS GREATER THAN 60% OF GUIDELINE VALUE)

$$\frac{\text{dpm}}{100\text{cm}^2} = \frac{(\text{gross count rate}) - (\text{background count rate})}{(\text{efficiency}) \left(\frac{\text{probe area}}{100\text{cm}^2} \right)}$$

D. L. Vann for
 SURVEYOR Kevin Graczyk
 SUPERVISOR B. [Signature]

REMOVABLE CONTAMINATION

DATE: 12/7/94LOCATION: Bausch & Lomb Process LineCOUNTER: Ludlum 2220 S# 50067
Model NumberEFFICIENCY: 0.0667 NA
Alpha Beta/Gamma

DPM/100cm ²			DPM/100cm ²			DPM/100cm ²		
LOCATION	βγ	α	LOCATION	βγ	α	LOCATION	βγ	α
1		< 69	21		< 69			
2		< 69	22		< 69			
3		< 69	23		< 69			
4		< 69	24		< 69			
5		< 69	25		< 69			
6		< 69	26	NA	< 69			
7		< 69	27		< 69			
8		< 69	28		< 69			
9		< 69	29		< 69			
10		< 69	30		< 69			
11	NA	< 69					NA	
12		< 69						
13		< 69						
14		< 69						
15		< 69		NA				
16		< 69		A				
17		< 69						
18		< 69						
19		< 69						
20		< 69						

ALPHA BKG COUNT RATE: 0 cpmBETAGAMMA BKG COUNT RATE: NAα MDA (dpm/100cm²) 69βγ MDA (dpm/100cm²) NASURVEYOR P.L. Vann for Kevin GraczykSUPERVISOR B. Dub

REMOVABLE CONTAMINATION

DATE: 12/7/94LOCATION: Bensch & Lamb Process LineCOUNTER: Luc/um 2220 # 50061
Model NumberEFFICIENCY: NA 0.238
Alpha Beta/Gamma

DPM/100cm ²			DPM/100cm ²			DPM/100cm ²		
LOCATION	βγ	α	LOCATION	βγ	α	LOCATION	βγ	α
1	< 940		21	< 940				
2	< 940		22	< 940				
3	< 940		23	< 940				
4	< 940		24	< 940				
5	< 940		25	< 940	N			
6	< 940		26	< 940	A			
7	< 940		27	< 940				
8	< 940		28	< 940				
9	< 940		29	< 940				
10	< 940	N	30	< 940		N		
11	< 940	A				A		
12	< 940							
13	< 940							
14	< 940							
15	< 940				N			
16	< 940				A			
17	< 940							
18	< 940							
19	< 940							
20	< 940							

ALPHA BKG COUNT RATE: NABETA/GAMMA BKG COUNT RATE: 44 cpmα MDA (dpm/100cm²) NAβγ MDA (dpm/100cm²) 940SURVEYOR B. DubSUPERVISOR D. J. Vann

EXPOSURE RATE SURVEY RESULTS

DATE: 12-7-94LOCATION: Bausch and Lomb Process LineINSTRUMENT: Bicron B 218LCAL DUE: 10-14-95

EXPOSURE RATE $\mu\text{R/h}$			EXPOSURE RATE $\mu\text{R/h}$		
GRID NUMBER	1cm	1m	GRID NUMBER	1cm	1m
X <u>RED</u>					
1		5 uR/hr	18		6 uR/hr
2		5 uR/hr	19		6 uR/hr
3		5 uR/hr	20		6 uR/hr
4		5 uR/hr	21		5 uR/hr
5		5 uR/hr	22		5 uR/hr
6		5 uR/hr	23		5 uR/hr
7		5 uR/hr	24	N/A	5 uR/hr
8		5 uR/hr	25		5 uR/hr
9	N/A	5 uR/hr	26		5 uR/hr
10		5 uR/hr	27		5 uR/hr
11		5 uR/hr	28		5 uR/hr
12		5 uR/hr	29		6 uR/hr
13		5 uR/hr	30	N/A	6 uR/hr
14		5 uR/hr			
15		5 uR/hr			
16		5 uR/hr			
17	6 uR/hr	5 uR/hr			

MEASUREMENTS TAKEN AT 1cm AND/OR 1m ABOVE THE FLOOR.

BKG: 5 uR/hr

SURVEYOR

SUPERVISOR

V. R. [Signature]
Bernard [Signature]

MDA CALCULATION SHEETMETER 2220 w/449 SERIAL #: 52836TS: 1 TB: 1 RB: 98 cpm EFF: 0.239 PROBE SIZE: 15 cm²

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{98}{1} + \frac{98}{1} \right)^{1/2}$$

$$\frac{0.239}{1} \left(\frac{15}{100} \text{ cm}^2 \right)$$

$$MDA = \underline{1360 \text{ dpm}/100 \text{ cm}^2}$$

METER 2220 w/449-3 SERIAL #: 50067TS: 1 TB: 1 RB: 0 cpm EFF: 0.0667 PROBE SIZE: 59 cm²

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{0}{1} + \frac{0}{1} \right)^{1/2}$$

$$\frac{0.0667}{1} \left(\frac{59}{100} \text{ cm}^2 \right)$$

$$MDA = \underline{69 \text{ dpm}/100 \text{ cm}^2}$$

METER 2220 w/449 SERIAL #: 50061TS: 1 TB: 1 RB: 44 EFF: 0.238 PROBE SIZE: 15 cm²

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{44}{1} + \frac{44}{1} \right)^{1/2}$$

$$\frac{0.238}{1} \left(\frac{15}{100} \text{ cm}^2 \right)$$

$$MDA = \underline{940 \text{ dpm}/100 \text{ cm}^2}$$

Removable
only

METER _____ SERIAL #: _____

TS: _____ TB: _____ RB: _____ EFF: _____ PROBE SIZE: _____

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{\quad}{1} + \frac{\quad}{1} \right)^{1/2}$$

$$\left(\frac{\quad}{100} \text{ cm}^2 \right)$$

$$MDA = \underline{\quad}$$

Technician: D. L. VannDate: 12/7/94Reviewed by: EB DuleDate: 12/7/94

Bausch & Lomb Release Survey Data

Survey Unit: Process Line

Activity (dpm/100cm²)

Location Number	Removable α			Removable β - γ		
	Activity	Uncertainty *	MDA	Activity	Uncertainty *	MDA
1	69	16	69	940	64	940
2	69	16	69	940	64	940
3	69	16	69	940	64	940
4	69	16	69	940	64	940
5	69	16	69	940	64	940
6	69	16	69	940	64	940
7	69	16	69	940	64	940
8	69	16	69	940	64	940
9	69	16	69	940	64	940
10	69	16	69	940	64	940
11	69	16	69	940	64	940
12	69	16	69	940	64	940
13	69	16	69	940	64	940
14	69	16	69	940	64	940
15	69	16	69	940	64	940
16	69	16	69	940	64	940
17	69	16	69	940	64	940
18	69	16	69	940	64	940
19	69	16	69	940	64	940
20	69	16	69	940	64	940
21	69	16	69	940	64	940
22	69	16	69	940	64	940
23	69	16	69	940	64	940
24	69	16	69	940	64	940
25	69	16	69	940	64	940
26	69	16	69	940	64	940
27	69	16	69	940	64	940
28	69	16	69	940	64	940
29	69	16	69	940	64	940
30	69	16	69	940	64	940

* Uncertainties represent the 95% Confidence level, based on counting statistics.
When survey data indicated less than MDA, the MDA value was used in the calculation.

Bausch & Lomb Release Survey Data

Survey Unit:

Process Line

Activity (dpm/100cm²)

Location Number	Direct α			Direct β - γ		
	Activity	Uncertainty *	MDA	Activity	Uncertainty *	MDA
1	69	16	69	1360	79	1360
2	69	16	69	1360	79	1360
3	69	16	69	1360	79	1360
4	69	16	69	1360	79	1360
5	69	16	69	1360	79	1360
6	69	16	69	1360	79	1360
7	69	16	69	1360	79	1360
8	69	16	69	1360	79	1360
9	69	16	69	1360	79	1360
10	69	16	69	1360	79	1360
11	69	16	69	1360	79	1360
12	69	16	69	1360	79	1360
13	69	16	69	1360	79	1360
14	69	16	69	1360	79	1360
15	69	16	69	1360	79	1360
16	69	16	69	1360	79	1360
17	69	16	69	1360	79	1360
18	69	16	69	1360	79	1360
19	69	16	69	1360	79	1360
20	69	16	69	1360	79	1360
21	69	16	69	1360	79	1360
22	69	16	69	1360	79	1360
23	69	16	69	1360	79	1360
24	69	16	69	1360	79	1360
25	69	16	69	1360	79	1360
26	69	16	69	1360	79	1360
27	69	16	69	1360	79	1360
28	69	16	69	1360	79	1360
29	69	16	69	1360	79	1360
30	69	16	69	1360	79	1360

* Uncertainties represent the 95% Confidence level, based on counting statistics.
When survey data indicated less than MDA, the MDA value was used in the calculation.

NINETY-FIVE PERCENT CONFIDENCE CALCULATION SHEET

Survey Unit Process Line Date 12-7-94

Meter End/um 2220 Serial # 52836

Probe 44-9 Serial # PRO 68918

MDA 1360 Survey Type Direct B-8

Guideline Value 5000 (dpm/100cm²)

Average Measurement Level

$$x_{ave} = \frac{1}{n_s} \sum_{i=1}^{n_s} x_i$$

where:

x_{ave} = calculated mean for a survey unit,
 n_s = number of measurements within a survey unit,
 x_i = systematic and random measurements at point (i)

$\sum x_i$ 40,800

n_s 30

x_{ave} = 1360

Standard Deviation

$$S_x = \sqrt{\frac{\sum_{i=1}^{n_s} (x_{ave} - x_i)^2}{n_s - 1}}$$

where:

S_x = standard deviation
 n_s = number of measurements within a survey unit
 x_i = systematic and random measurements at point (i)
 x_{ave} = calculated mean for a survey unit

x_{ave} = 1360 n_s = 30

$\sum (x_{ave} - x_i)^2$ = 0 S_x = 0

$$\mu_{\alpha} = x_{ave} + t_{1-\alpha, df} \frac{s_x}{\sqrt{n_s}}$$

where:

μ_{α} = value compared guideline value to determine 95 % Confidence Level
 x_{ave} = calculated mean for a survey unit,
 $t_{1-\alpha, df}$ = 95 % confidence level for n-1 degrees of freedom (from NUREG/CR-5849, Table B-1, "Factors for Comparison of Survey Data")
 s_x = standard deviation of measurements in a survey unit, and
 n_s = number of measurements within a survey unit used to determine x_{ave} and s_x .

$$x_{ave} = \underline{1360}$$

$$t_{1-\alpha, df} = \underline{1.699}$$

$$s_x = \underline{0}$$

$$n_s = \underline{30}$$

$$\mu_{\alpha} = \underline{1360}$$

NOTE: When determining x_{ave} , only minimum detectable activity (MDA) values were available for some measurements locations; the MDA values were therefore used as actual activity levels for the purpose of performing this calculation.

The site specific guideline values for this project are 5000 dpm/100 cm² fixed plus removable and 1000 dpm/100 cm² removable for beta-gamma emitting nuclides, and 1000 dpm/100 cm² fixed plus removable and 200 for alpha emitting nuclides. If μ_{α} is less than the specific guideline value, the data for this survey unit satisfies the guideline at the 95 % confidence level. If μ_{α} is greater than the specific guideline value, further remediation of the survey unit is necessary.

Individual Completing Form: D. L. Vann Date: 12-7-94

Reviewed By: R. Kight Date: 1/10/95

NINETY-FIVE PERCENT CONFIDENCE CALCULATION SHEET

Survey Unit Process Line Date 12-7-94

Meter Ludlum 2220 Serial # 50067

Probe AC-3 Serial # 712582

MDA 69 dpm/100cm² Survey Type Direct α

Guideline Value 1000 (dpm/100cm²)

Average Measurement Level

$$x_{ave} = \frac{1}{n_s} \sum_{i=1}^{n_s} x_i$$

where:

x_{ave} = calculated mean for a survey unit,
 n_s = number of measurements within a survey unit,
 x_i = systematic and random measurements at point (i)

Σx_i 2070

n_s 30

x_{ave} = 69

Standard Deviation

$$S_x = \sqrt{\frac{\sum_{i=1}^{n_s} (x_{ave} - x_i)^2}{n_s - 1}}$$

where:

S_x = standard deviation
 n_s = number of measurements within a survey unit
 x_i = systematic and random measurements at point (i)
 x_{ave} = calculated mean for a survey unit

x_{ave} = 69 n_s = 30

$\Sigma (x_{ave} - x_i)^2$ = 0 S_x = 0

$$\mu_{\alpha} = x_{ave} + t_{1-\alpha, df} \frac{s_x}{\sqrt{n_s}}$$

where:

μ_{α} = value compared guideline value to determine 95% Confidence Level
 x_{ave} = calculated mean for a survey unit,
 $t_{1-\alpha, df}$ = 95% confidence level for n-1 degrees of freedom (from NUREG/CR-5849, Table B-1, "Factors for Comparison of Survey Data")
 s_x = standard deviation of measurements in a survey unit, and
 n_s = number of measurements within a survey unit used to determine x_{ave} and s_x .

x_{ave} = 69
 $t_{1-\alpha, df}$ = 1.699
 s_x = 0
 n_s = 30
 μ_{α} = 69

NOTE: When determining x_{ave} , only minimum detectable activity (MDA) values were available for some measurements locations; the MDA values were therefore used as actual activity levels for the purpose of performing this calculation.

The site specific guideline values for this project are 5000 dpm/100 cm² fixed plus removable and 1000 dpm/100 cm² removable for beta-gamma emitting nuclides, and 1000 dpm/100 cm² fixed plus removable and 200 for alpha emitting nuclides. If μ_{α} is less than the specific guideline value, the data for this survey unit satisfies the guideline at the 95% confidence level. If μ_{α} is greater than the specific guideline value, further remediation of the survey unit is necessary.

Individual Completing Form: D. L. Vann Date: 12-7-94

Reviewed By: R. Kjilt Date: 1/10/95

NINETY-FIVE PERCENT CONFIDENCE CALCULATION SHEET

Survey Unit Process Line Date 12-7-94

Meter Ludlum 2220 Serial # 50061

Probe 44-9 Serial # PRO 66761

MDA 940 dpm/100cm² Survey Type Removable B-8

Guideline Value 1000 (dpm/100cm²)

Average Measurement Level

$$x_{ave} = \frac{1}{n_s} \sum_{i=1}^{n_s} x_i$$

where:

x_{ave} = calculated mean for a survey unit,
 n_s = number of measurements within a survey unit,
 x_i = systematic and random measurements at point (i)

$\sum x_i$ 28,200

n_s 30

x_{ave} = 940

Standard Deviation

$$S_x = \sqrt{\frac{\sum_{i=1}^{n_s} (x_{ave} - x_i)^2}{n_s - 1}}$$

where:

S_x = standard deviation
 n_s = number of measurements within a survey unit
 x_i = systematic and random measurements at point (i)
 x_{ave} = calculated mean for a survey unit

x_{ave} = 940 n_s = 30

$\sum (x_{ave} - x_i)^2$ = 0 S_x = 0

$$\mu_{\alpha} = x_{ave} + t_{1-\alpha, df} \frac{s_x}{\sqrt{n_s}}$$

where:

μ_{α} = value compared guideline value to determine 95 % Confidence Level
 x_{ave} = calculated mean for a survey unit,
 $t_{1-\alpha, df}$ = 95 % confidence level for n-1 degrees of freedom (from NUREG/CR-5849, Table B-1, "Factors for Comparison of Survey Data")
 s_x = standard deviation of measurements in a survey unit, and
 n_s = number of measurements within a survey unit used to determine x_{ave} and s_x .

x_{ave} = 940
 $t_{1-\alpha, df}$ = 1.699
 s_x = 0
 n_s = 30
 μ_{α} = 940

NOTE: When determining x_{ave} , only minimum detectable activity (MDA) values were available for some measurements locations; the MDA values were therefore used as actual activity levels for the purpose of performing this calculation.

The site specific guideline values for this project are 5000 dpm/100 cm² fixed plus removable and 1000 dpm/100 cm² removable for beta-gamma emitting nuclides, and 1000 dpm/100 cm² fixed plus removable and 200 for alpha emitting nuclides. If μ_{α} is less than the specific guideline value, the data for this survey unit satisfies the guideline at the 95 % confidence level. If μ_{α} is greater than the specific guideline value, further remediation of the survey unit is necessary.

Individual Completing Form: D. L. Vann Date: 12-7-94

Reviewed By: R. H. J. Date: 1/10/95

NINETY-FIVE PERCENT CONFIDENCE CALCULATION SHEET

Bausch & Lomb

Survey Unit process line Date 12-7-94

Meter Ludlum 2220 Serial # 50067

Probe AC-3 Serial # 712582

MDA 69 dpm/100cm² Survey Type Removable ☒

Guideline Value 200 (dpm/100cm²)

Average Measurement Level

$$x_{ave} = \frac{1}{n_s} \sum_{i=1}^{n_s} x_i$$

where:

x_{ave} = calculated mean for a survey unit,
 n_s = number of measurements within a survey unit,
 x_i = systematic and random measurements at point (i)

$\sum x_i$ 2070

n_s 30

x_{ave} = 69

Standard Deviation

$$S_x = \sqrt{\frac{\sum_{i=1}^{n_s} (x_{ave} - x_i)^2}{n_s - 1}}$$

where:

S_x = standard deviation
 n_s = number of measurements within a survey unit
 x_i = systematic and random measurements at point (i)
 x_{ave} = calculated mean for a survey unit

x_{ave} = 69 n_s = 30

$\sum (x_{ave} - x_i)^2$ = 0 S_x = 0

$$\mu_{\alpha} = x_{ave} + t_{1-\alpha, df} \frac{s_x}{\sqrt{n_s}}$$

where:

μ_{α} = value compared guideline value to determine 95 % Confidence Level
 x_{ave} = calculated mean for a survey unit,
 $t_{1-\alpha, df}$ = 95 % confidence level for n-1 degrees of freedom (from NUREG/CR-5849, Table B-1, "Factors for Comparison of Survey Data")
 s_x = standard deviation of measurements in a survey unit, and
 n_s = number of measurements within a survey unit used to determine x_{ave} and s_x .

x_{ave} = 69
 $t_{1-\alpha, df}$ = 1.699
 s_x = 0
 n_s = 30
 μ_{α} = 69

NOTE: When determining x_{ave} , only minimum detectable activity (MDA) values were available for some measurements locations; the MDA values were therefore used as actual activity levels for the purpose of performing this calculation.

The site specific guideline values for this project are 5000 dpm/100 cm² fixed plus removable and 1000 dpm/100 cm² removable for beta-gamma emitting nuclides, and 1000 dpm/100 cm² fixed plus removable and 200 for alpha emitting nuclides. If μ_{α} is less than the specific guideline value, the data for this survey unit satisfies the guideline at the 95 % confidence level. If μ_{α} is greater than the specific guideline value, further remediation of the survey unit is necessary.

Individual Completing Form: D. L. Vann Date: 12-7-94

Reviewed By: R. V. J. H. Date: 1/10/95

NINETY-FIVE PERCENT CONFIDENCE CALCULATION SHEET

Survey Unit Process Line Date 12-7-94

Meter Bilon microRem Serial # B 218 L

Probe NA Serial # NA

MDA NA Survey Type Exposure

Guideline Value 10 nRem/hr (dpm/100cm²)

Average Measurement Level

$$x_{ave} = \frac{1}{n_s} \sum_{i=1}^{n_s} x_i$$

where:

x_{ave} = calculated mean for a survey unit,
 n_s = number of measurements within a survey unit,
 x_i = systematic and random measurements at point (i)

Σx_i 155

n_s 30

x_{ave} = 5.2

Standard Deviation

$$S_x = \sqrt{\frac{\sum_{i=1}^{n_s} (x_{ave} - x_i)^2}{n_s - 1}}$$

where:

S_x = standard deviation
 n_s = number of measurements within a survey unit
 x_i = systematic and random measurements at point (i).
 x_{ave} = calculated mean for a survey unit

x_{ave} = 5.2 n_s = 30

$\Sigma (x_{ave} - x_i)^2$ = 4.2 S_x = 0.4

$$\mu_{\alpha} = x_{ave} + t_{1-\alpha, df} \frac{s_x}{\sqrt{n_s}}$$

where:

μ_{α} = value compared guideline value to determine 95 % Confidence Level
 x_{ave} = calculated mean for a survey unit,
 $t_{1-\alpha, df}$ = 95 % confidence level for n-1 degrees of freedom (from NUREG/CR-5849, Table B-1, "Factors for Comparison of Survey Data")
 s_x = standard deviation of measurements in a survey unit, and
 n_s = number of measurements within a survey unit used to determine x_{ave} and s_x .

$$x_{ave} = \underline{5.2}$$

$$t_{1-\alpha, df} = \underline{1.699}$$

$$s_x = \underline{0.4}$$

$$n_s = \underline{30}$$

$$\mu_{\alpha} = \underline{5.3}$$

NOTE: When determining x_{ave} , only minimum detectable activity (MDA) values were available for some measurements locations; the MDA values were therefore used as actual activity levels for the purpose of performing this calculation.

The site specific guideline values for this project are 5000 dpm/100 cm² fixed plus removable and 1000 dpm/100 cm² removable for beta-gamma emitting nuclides, and 1000 dpm/100 cm² fixed plus removable and 200 for alpha emitting nuclides. If μ_{α} is less than the specific guideline value, the data for this survey unit satisfies the guideline at the 95 % confidence level. If μ_{α} is greater than the specific guideline value, further remediation of the survey unit is necessary.

Individual Completing Form: D. L. Vann Date: 12-7-94

Reviewed By: L. Kight Date: 1/10/95

[illegible]

REMOVABLE CONTAMINATION

DATE: 12-7-94LOCATION: Bausch & Lomb Storage RoomCOUNTER: Endium 2220 50067
Model NumberEFFICIENCY: 0.0667 NA
Alpha Beta/Gamma

DPM/100cm ²			DPM/100cm ²			DPM/100cm ²		
LOCATION	$\beta\gamma$	α	LOCATION	$\beta\gamma$	α	LOCATION	$\beta\gamma$	α
1		< 69	21		< 69			
2		< 69	22		< 69			
3		< 69	23		< 69			
4		< 69	24		< 69			
5		< 69	25	NA	< 69			
6		< 69	26		< 69			
7		< 69	27		< 69			
8		< 69	28		< 69			
9		< 69	29		< 69			
10		< 69	30		< 69			
11	NA	< 69					NA	
12		< 69						
13		< 69						
14		< 69						
15		< 69		NA				
16		< 69						
17		< 69						
18		< 69						
19		< 69						
20		< 69						

ALPHA BKG COUNT RATE: 0 cpmBETA/GAMMA BKG COUNT RATE: NA α MDA (dpm/100cm²) 69 $\beta\gamma$ MDA (dpm/100cm²) NASURVEYOR D. L. Vurn
for Kevin GraczykSUPERVISOR B. Dule

REMOVABLE CONTAMINATION

DATE: 12-7-94LOCATION: Bqusch + Lamb Storage RoomCOUNTER: Luc/Lum 222050061 ^{plu. 10}
~~52836~~
NumberEFFICIENCY: N/A
Alpha0.238
Beta/Gamma

DPM/100cm ²			DPM/100cm ²			DPM/100cm ²		
LOCATION	$\beta\gamma$	α	LOCATION	$\beta\gamma$	α	LOCATION	$\beta\gamma$	α
1	< 940		21	< 940				
2	< 940		22	< 940				
3	< 940		23	< 940				
4	< 940		24	< 940				
5	< 940		25	< 940	N			
6	< 940		26	< 940	A			
7	< 940		27	< 940				
8	< 940		28	< 940				
9	< 940		29	< 940				
10	< 940	N	30	< 940		N		
11	< 940	A				A		
12	< 940							
13	< 940							
14	< 940							
15	< 940				N			
16	< 940				A			
17	< 940							
18	< 940							
19	< 940							
20	< 940							

ALPHA BKG COUNT RATE: N/A α MDA (dpm/100cm²) N/A $\beta\gamma$ MDA (dpm/100cm²) 940BETAGAMMA BKG COUNT RATE: 44 cpmSURVEYOR D. L. VannSUPERVISOR Bronckh

AVERAGE CONTAMINATION SURVEY RESULTS

Page 4 of 7

DATE: 12-7-94

LOCATION: Bausch & Lomb storage area

INSTRUMENT: Ludlum 2220 50067
MODEL NUMBER

PROBE: AC-3 EFF: 0.0667
MODEL

GRID NUMBER	dpm/100cm ²	GRID NUMBER	dpm/100cm ²
1	< 69	20	< 69
2	< 69	21	< 69
3	< 69	22	< 69
4	< 69	23	< 69
5	< 69	24	< 69
6	< 69	25	< 69
7	< 69	26	< 69
8	< 69	27	< 69
9	< 69	28	< 69
10	< 69	29	< 69
11	< 69	30	< 69
12	< 69		
13	< 69		
14	< 69		
15	< 69		
16	< 69		
17	< 69		
18	< 69		
19	< 69		

Average Background counts per minute 0 cpm
MDA (dpm/100cm²) 69

*BASED ON AVERAGE OF 5 DIRECT MEASUREMENTS
(INITIAL DIRECT MEASUREMENT WAS GREATER THAN 60% OF GUIDELINE VALUE)

$$\frac{\text{dpm}}{100\text{cm}^2} = \frac{(\text{gross count rate}) - (\text{background count rate})}{(\text{efficiency}) (\text{probe area})}$$

100cm² 100cm²

SURVEYOR P. L. Vonn for Kevin Graczyk
SUPERVISOR B. Dub

AVERAGE CONTAMINATION SURVEY RESULTS

Page 5 of 7

DATE: 12-7-94

LOCATION: Bausch & Lomb Storage area

INSTRUMENT: Endium 2220 52836
MODEL NUMBER

PROBE: 44-9 EFF: 0.239
MODEL

GRID NUMBER	dpm/100cm ²	GRID NUMBER	dpm/100cm ²
1	< 1360	20	< 1360
2	< 1360	21	< 1360
3	< 1360	22	< 1360
4	< 1360	23	< 1360
5	< 1360	24	< 1360
6	< 1360	25	2520
7	< 1360	26	< 1360
8	< 1360	27	< 1360
9	< 1360	28	< 1360
10	< 1360	29	< 1360
11	1680	30	< 1360
12	< 1360		
13	< 1360		
14	< 1360		
15	< 1360		
16	< 1360		
17	< 1360		
18	< 1360		
19	< 1360		

Average Background counts per minute 98

MDA (dpm/100cm²) 1360

*BASED ON AVERAGE OF 5 DIRECT MEASUREMENTS
(INITIAL DIRECT MEASUREMENT WAS GREATER THAN 60% OF GUIDELINE VALUE)

$$\frac{\text{dpm}}{100\text{cm}^2} = \frac{(\text{gross count rate}) - (\text{background count rate})}{(\text{efficiency}) \left(\frac{\text{probe area}}{100\text{cm}^2} \right)}$$

SURVEYOR D. L. Vann for Kevin Graczyk

SUPERVISOR B. Duler

EXPOSURE RATE SURVEY RESULTS

DATE: 12-7-94LOCATION: Bausch & Lomb Storage AreaINSTRUMENT: Bicron 13218LCAL DUE: 10-14-95

GRID NUMBER	EXPOSURE RATE $\mu\text{R/h}$		GRID NUMBER	EXPOSURE RATE $\mu\text{R/h}$	
	1cm	1m		1cm	1m
1		5uR/hr	18		5uR/hr
2		5uR/hr	19		5uR/hr
3		5uR/hr	20		5uR/hr
4		5uR/hr	21		5uR/hr
5		5uR/hr	22		5uR/hr
6		5uR/hr	23		5uR/hr
7		5uR/hr	24	N/A	5uR/hr
8		5uR/hr	25		5uR/hr
9	N/A	5uR/hr	26		5uR/hr
10		5uR/hr	27		4uR/hr
11		6uR/hr	28		4uR/hr
12		6uR/hr	29		4uR/hr
13		5uR/hr	30		4uR/hr
14		5uR/hr			
15		5uR/hr		N	
16		6uR/hr		A	
17		6uR/hr			

MEASUREMENTS TAKEN AT 1cm AND/OR 1m ABOVE THE FLOOR.

BKG: 5uR/hrSURVEYOR: D. L. VaneSUPERVISOR: Bureau Sub

MDA CALCULATION SHEET

METER 2220 w/449 SERIAL #: 52836TS: 1 TB: 1 RB: 98 cpm EFF: 0.239 PROBE SIZE: 15 cm²

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{98}{1} + \frac{98}{1} \right)^{1/2}$$

$$0.239 \quad (15 / 100 \text{ cm}^2)$$

$$MDA = \underline{1360 \text{ dpm}/100 \text{ cm}^2}$$

METER 2220 w/AC-3 SERIAL #: 50067TS: 1 TB: 1 RB: 0 cpm EFF: 0.0667 PROBE SIZE: 59 cm²

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{0}{1} + \frac{0}{1} \right)^{1/2}$$

$$0.0667 \quad (59 / 100 \text{ cm}^2)$$

$$MDA = \underline{69 \text{ dpm}/100 \text{ cm}^2}$$

METER 2220 w/449 SERIAL #: 50061TS: 1 TB: 1 RB: 44 cpm EFF: 0.238 PROBE SIZE: 15 cm²

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{44}{1} + \frac{44}{1} \right)^{1/2}$$

$$0.238 \quad (15 / 100 \text{ cm}^2)$$

$$MDA = \underline{940 \text{ dpm}/100 \text{ cm}^2}$$

Removable
only

METER _____ SERIAL #: _____

TS: _____ TB: _____ RB: _____ EFF: _____ PROBE SIZE: _____

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{\quad}{1} + \frac{\quad}{1} \right)^{1/2}$$

$$\quad (\quad / 100 \text{ cm}^2)$$

$$MDA = \underline{\quad}$$

Technician: D. L. Vann Date: 12-7-94Reviewed by: B. Dula Date: 12-7-94

Bausch & Lomb Release Survey Data

Survey Unit: Storage Area

Activity (dpm/100cm²)

Location Number	Removable α			Removable β - γ		
	Activity	Uncertainty *	MDA	Activity	Uncertainty *	MDA
1	69	16	69	940	64	940
2	69	16	69	940	64	940
3	69	16	69	940	64	940
4	69	16	69	940	64	940
5	69	16	69	940	64	940
6	69	16	69	940	64	940
7	69	16	69	940	64	940
8	69	16	69	940	64	940
9	69	16	69	940	64	940
10	69	16	69	940	64	940
11	69	16	69	940	64	940
12	69	16	69	940	64	940
13	69	16	69	940	64	940
14	69	16	69	940	64	940
15	69	16	69	940	64	940
16	69	16	69	940	64	940
17	69	16	69	940	64	940
18	69	16	69	940	64	940
19	69	16	69	940	64	940
20	69	16	69	940	64	940
21	69	16	69	940	64	940
22	69	16	69	940	64	940
23	69	16	69	940	64	940
24	69	16	69	940	64	940
25	69	16	69	940	64	940
26	69	16	69	940	64	940
27	69	16	69	940	64	940
28	69	16	69	940	64	940
29	69	16	69	940	64	940
30	69	16	69	940	64	940

* Uncertainties represent the 95% Confidence level, based on counting statistics.
When survey data indicated less than MDA, the MDA value was used in the calculation.

Bausch & Lomb Release Survey Data

Survey Unit: Storage Area

Activity (dpm/100cm²)

<u>Location Number</u>	<u>Direct α</u>			<u>Direct β-γ</u>		
	<u>Activity</u>	<u>Uncertainty *</u>	<u>MDA</u>	<u>Activity</u>	<u>Uncertainty *</u>	<u>MDA</u>
1	69	16	69	1360	79	1360
2	69	16	69	1360	79	1360
3	69	16	69	1360	79	1360
4	69	16	69	1360	79	1360
5	69	16	69	1360	79	1360
6	69	16	69	1360	79	1360
7	69	16	69	1360	79	1360
8	69	16	69	1360	79	1360
9	69	16	69	1360	79	1360
10	69	16	69	1360	79	1360
11	69	16	69	1680	87	1360
12	69	16	69	1360	79	1360
13	69	16	69	1360	79	1360
14	69	16	69	1360	79	1360
15	69	16	69	1360	79	1360
16	69	16	69	1360	79	1360
17	69	16	69	1360	79	1360
18	69	16	69	1360	79	1360
19	69	16	69	1360	79	1360
20	69	16	69	1360	79	1360
21	69	16	69	1360	79	1360
22	69	16	69	1360	79	1360
23	69	16	69	1360	79	1360
24	69	16	69	1360	79	1360
25	69	16	69	2520	104	1360
26	69	16	69	1360	79	1360
27	69	16	69	1360	79	1360
28	69	16	69	1360	79	1360
29	69	16	69	1360	79	1360
30	69	16	69	1360	79	1360

* Uncertainties represent the 95% Confidence level, based on counting statistics.
When survey data indicated less than MDA, the MDA value was used in the calculation.

NINETY-FIVE PERCENT CONFIDENCE CALCULATION SHEET

Survey Unit Storage Room Date 12-7-94

Meter Ludlum 2220 Serial # 50067

Probe AC-3 Serial # 712582

MDA 69 dpm/100cm² Survey Type Removable α

Guideline Value 200 (dpm/100cm²)

Average Measurement Level

$$x_{ave} = \frac{1}{n_s} \sum_{i=1}^{n_s} x_i$$

where:

x_{ave} = calculated mean for a survey unit,
 n_s = number of measurements within a survey unit,
 x_i = systematic and random measurements at point (i)

$\sum x_i$ 2070

n_s 30

x_{ave} = 69

Standard Deviation

$$S_x = \sqrt{\frac{\sum_{i=1}^{n_s} (x_{ave} - x_i)^2}{n_s - 1}}$$

where:

S_x = standard deviation
 n_s = number of measurements within a survey unit
 x_i = systematic and random measurements at point (i)
 x_{ave} = calculated mean for a survey unit

x_{ave} = 69 n_s = 30

$\sum (x_{ave} - x_i)^2$ = 0 S_x = 0

$$\mu_{\alpha} = x_{ave} + t_{1-\alpha, df} \frac{s_x}{\sqrt{n_s}}$$

where:

μ_{α} = value compared guideline value to determine 95 % Confidence Level
 x_{ave} = calculated mean for a survey unit,
 $t_{1-\alpha, df}$ = 95 % confidence level for n-1 degrees of freedom (from NUREG/CR-5849, Table B-1, "Factors for Comparison of Survey Data")
 s_x = standard deviation of measurements in a survey unit, and
 n_s = number of measurements within a survey unit used to determine x_{ave} and s_x .

x_{ave} = 69
 $t_{1-\alpha, df}$ = 1.699
 s_x = 0
 n_s = 30
 μ_{α} = 69

NOTE: When determining x_{ave} , only minimum detectable activity (MDA) values were available for some measurements locations; the MDA values were therefore used as actual activity levels for the purpose of performing this calculation.

The site specific guideline values for this project are 5000 dpm/100 cm² fixed plus removable and 1000 dpm/100 cm² removable for beta-gamma emitting nuclides, and 1000 dpm/100 cm² fixed plus removable and 200 for alpha emitting nuclides. If μ_{α} is less than the specific guideline value, the data for this survey unit satisfies the guideline at the 95 % confidence level. If μ_{α} is greater than the specific guideline value, further remediation of the survey unit is necessary.

Individual Completing Form: D. L. Vann Date: 12-7-94

Reviewed By: R. Bight Date: 1/10/95

NINETY-FIVE PERCENT CONFIDENCE CALCULATION SHEET

Survey Unit Storage Room Date 12-7-94

Meter Ludlum 2220 Serial # 50061

Probe 44-9 Serial # PRO66761

MDA 940 dpm/100cm² Survey Type Removable B-8

Guideline Value 1000 (dpm/100cm²)

Average Measurement Level

$$x_{ave} = \frac{1}{n_s} \sum_{i=1}^{n_s} x_i$$

where:

x_{ave} = calculated mean for a survey unit,
 n_s = number of measurements within a survey unit,
 x_i = systematic and random measurements at point (i)

$\sum x_i$ 28,200

n_s 30

x_{ave} = 940

Standard Deviation

$$S_x = \sqrt{\frac{\sum_{i=1}^{n_s} (x_{ave} - x_i)^2}{n_s - 1}}$$

where:

S_x = standard deviation
 n_s = number of measurements within a survey unit
 x_i = systematic and random measurements at point (i)
 x_{ave} = calculated mean for a survey unit

x_{ave} = 940 n_s = 30

$\sum (x_{ave} - x_i)^2$ = 0 S_x = 0

$$\mu_{\alpha} = x_{ave} + t_{1-\alpha, df} \frac{s_x}{\sqrt{n_s}}$$

where:

μ_{α} = value compared guideline value to determine 95% Confidence Level
 x_{ave} = calculated mean for a survey unit,
 $t_{1-\alpha, df}$ = 95% confidence level for n-1 degrees of freedom (from NUREG/CR-5849, Table B-1, "Factors for Comparison of Survey Data")
 s_x = standard deviation of measurements in a survey unit, and
 n_s = number of measurements within a survey unit used to determine x_{ave} and s_x .

x_{ave} = 940
 $t_{1-\alpha, df}$ = 1.699
 s_x = 0
 n_s = 30
 μ_{α} = 940

NOTE: When determining x_{ave} , only minimum detectable activity (MDA) values were available for some measurements locations; the MDA values were therefore used as actual activity levels for the purpose of performing this calculation.

The site specific guideline values for this project are 5000 dpm/100 cm² fixed plus removable and 1000 dpm/100 cm² removable for beta-gamma emitting nuclides, and 1000 dpm/100 cm² fixed plus removable and 200 for alpha emitting nuclides. If μ_{α} is less than the specific guideline value, the data for this survey unit satisfies the guideline at the 95% confidence level. If μ_{α} is greater than the specific guideline value, further remediation of the survey unit is necessary.

Individual Completing Form: D. L. Vann Date: 12-7-94

Reviewed By: R. Kight Date: 1/10/95

NINETY-FIVE PERCENT CONFIDENCE CALCULATION SHEET

Survey Unit Storage Area Date 12-7-94

Meter Endium 2220 Serial # 50067

Probe AC-3 Serial # 712582

MDA 69 dpm/100cm² Survey Type Direct &

Guideline Value 1000 (dpm/100cm²)

Average Measurement Level

$$x_{ave} = \frac{1}{n_i} \sum_{i=1}^{n_i} x_i$$

where:

x_{ave} = calculated mean for a survey unit,
 n_i = number of measurements within a survey unit,
 x_i = systematic and random measurements at point (i)

$\sum x_i$ 2070

n_i 30

x_{ave} = 69

Standard Deviation

$$S_x = \sqrt{\frac{\sum_{i=1}^{n_i} (x_{ave} - x_i)^2}{n_i - 1}}$$

where:

S_x = standard deviation
 n_i = number of measurements within a survey unit
 x_i = systematic and random measurements at point (i)
 x_{ave} = calculated mean for a survey unit

x_{ave} = 69 n_i = 30

$\sum (x_{ave} - x_i)^2$ = 0 s_x = 0

$$\mu_{\alpha} = x_{ave} + t_{1-\alpha, df} \frac{s_x}{\sqrt{n_s}}$$

where:

μ_{α} = value compared guideline value to determine 95 % Confidence Level
 x_{ave} = calculated mean for a survey unit,
 $t_{1-\alpha, df}$ = 95 % confidence level for n-1 degrees of freedom (from NUREG/CR-5849, Table B-1, "Factors for Comparison of Survey Data")
 s_x = standard deviation of measurements in a survey unit, and
 n_s = number of measurements within a survey unit used to determine x_{ave} and s_x .

x_{ave} = 69
 $t_{1-\alpha, df}$ = 1.699
 s_x = 0
 n_s = 30
 μ_{α} = 69

NOTE: When determining x_{ave} , only minimum detectable activity (MDA) values were available for some measurements locations; the MDA values were therefore used as actual activity levels for the purpose of performing this calculation.

The site specific guideline values for this project are 5000 dpm/100 cm² fixed plus removable and 1000 dpm/100 cm² removable for beta-gamma emitting nuclides, and 1000 dpm/100 cm² fixed plus removable and 200 for alpha emitting nuclides. If μ_{α} is less than the specific guideline value, the data for this survey unit satisfies the guideline at the 95 % confidence level. If μ_{α} is greater than the specific guideline value, further remediation of the survey unit is necessary.

Individual Completing Form: D. L. Vann Date: 12-7-94

Reviewed By: R. Hjelt Date: 1/10/95

NINETY-FIVE PERCENT CONFIDENCE CALCULATION SHEET

Survey Unit Storage Area Date 12-7-94

Meter Lucilon 2220 Serial # 52836

Probe 44-9 Serial # PRO 68918

MDA 1360 dpm/100cm² Survey Type Direct B-8

Guideline Value 5000 (dpm/100cm²)

Average Measurement Level

$$x_{ave} = \frac{1}{n_s} \sum_{i=1}^{n_s} x_i$$

where:

x_{ave} = calculated mean for a survey unit,
 n_s = number of measurements within a survey unit,
 x_i = systematic and random measurements at point (i)

$\sum x_i$ 42,280

n_s 30

x_{ave} = 1409.3

Standard Deviation

$$S_x = \sqrt{\frac{\sum_{i=1}^{n_s} (x_{ave} - x_i)^2}{n_s - 1}}$$

where:

S_x = standard deviation
 n_s = number of measurements within a survey unit
 x_i = systematic and random measurements at point (i)
 x_{ave} = calculated mean for a survey unit

x_{ave} = 1409.3 n_s = 30

$\sum (x_{ave} - x_i)^2$ = 1,374,987 S_x = 217.7

$$\mu_{\alpha} = x_{ave} + t_{1-\alpha, df} \frac{s_x}{\sqrt{n_s}}$$

where:

μ_{α} = value compared guideline value to determine 95% Confidence Level
 x_{ave} = calculated mean for a survey unit,
 $t_{1-\alpha, df}$ = 95% confidence level for n-1 degrees of freedom (from NUREG/CR-5849, Table B-1, "Factors for Comparison of Survey Data")
 s_x = standard deviation of measurements in a survey unit, and
 n_s = number of measurements within a survey unit used to determine x_{ave} and s_x .

$$x_{ave} = \underline{1409.3}$$

$$t_{1-\alpha, df} = \underline{1.699}$$

$$s_x = \underline{217.7}$$

$$n_s = \underline{30}$$

$$\mu_{\alpha} = \underline{1476.8}$$

NOTE: When determining x_{ave} , only minimum detectable activity (MDA) values were available for some measurements locations; the MDA values were therefore used as actual activity levels for the purpose of performing this calculation.

The site specific guideline values for this project are 5000 dpm/100 cm² fixed plus removable and 1000 dpm/100 cm² removable for beta-gamma emitting nuclides, and 1000 dpm/100 cm² fixed plus removable and 200 for alpha emitting nuclides. If μ_{α} is less than the specific guideline value, the data for this survey unit satisfies the guideline at the 95% confidence level. If μ_{α} is greater than the specific guideline value, further remediation of the survey unit is necessary.

Individual Completing Form: D. L. Vann Date: 12-7-94

Reviewed By: R. Bight Date: 1/10/95

NINETY-FIVE PERCENT CONFIDENCE CALCULATION SHEET

Survey Unit Storage Area Date 12-7-94

Meter Bicron MicroRem Serial # B218L

Probe NA Serial # NA

MDA NA Survey Type Exposure

Guideline Value 10 mRem/hr (~~dpm/100cm²~~)

Average Measurement Level

$$x_{ave} = \frac{1}{n_s} \sum_{i=1}^{n_s} x_i$$

where:

x_{ave} = calculated mean for a survey unit,
 n_s = number of measurements within a survey unit,
 x_i = systematic and random measurements at point (i)

$\sum x_i$ 150

n_s 30

x_{ave} = 5

Standard Deviation

$$S_x = \sqrt{\frac{\sum_{i=1}^{n_s} (x_{ave} - x_i)^2}{n_s - 1}}$$

where:

S_x = standard deviation
 n_s = number of measurements within a survey unit
 x_i = systematic and random measurements at point (i)
 x_{ave} = calculated mean for a survey unit

x_{ave} = 5 n_s = 30

$\sum (x_{ave} - x_i)^2$ = 8 S_x = 0.5

$$\mu_{\alpha} = x_{ave} + t_{1-\alpha, df} \frac{s_x}{\sqrt{n_s}}$$

where:

μ_{α} = value compared guideline value to determine 95 % Confidence Level
 x_{ave} = calculated mean for a survey unit,
 $t_{1-\alpha, df}$ = 95 % confidence level for n-1 degrees of freedom (from NUREG/CR-5849, Table B-1, "Factors for Comparison of Survey Data")
 s_x = standard deviation of measurements in a survey unit, and
 n_s = number of measurements within a survey unit used to determine x_{ave} and s_x .

x_{ave} = 5
 $t_{1-\alpha, df}$ = 1.699
 s_x = 0.5
 n_s = 30
 μ_{α} = 5.2

NOTE: When determining x_{ave} , only minimum detectable activity (MDA) values were available for some measurements locations; the MDA values were therefore used as actual activity levels for the purpose of performing this calculation.

The site specific guideline values for this project are 5000 dpm/100 cm² fixed plus removable and 1000 dpm/100 cm² removable for beta-gamma emitting nuclides, and 1000 dpm/100 cm² fixed plus removable and 200 for alpha emitting nuclides. If μ_{α} is less than the specific guideline value, the data for this survey unit satisfies the guideline at the 95 % confidence level. If μ_{α} is greater than the specific guideline value, further remediation of the survey unit is necessary.

Individual Completing Form: D. L. Vann Date: 12-7-94

Reviewed By: R. Kight Date: 4/13/95

[illegible]

REMOVABLE CONTAMINATION

DATE: 12-7-94LOCATION: Bausch & Lomb Furnace AreaCOUNTER: Ludlum 2220 50067
Model NumberEFFICIENCY: 0.0667 NA
Alpha Beta/Gamma

DPM/100cm ²			DPM/100cm ²			DPM/100cm ²		
LOCATION	βγ	α	LOCATION	βγ	α	LOCATION	βγ	α
1		< 69	21		< 69			
2		< 69	22		< 69			
3		< 69	23		< 69			
4		< 69	24		< 69			
5		< 69	25	NA	< 69			
6		< 69	26		< 69			
7		< 69	27		< 69			
8		< 69	28		< 69			
9		< 69	29		< 69			
10		< 69	30		< 69			
11	NA	< 69					NA	
12		< 69						
13		< 69						
14		< 69						
15		< 69		NA				
16		< 69						
17		< 69						
18		< 69						
19		< 69						
20		< 69						

ALPHA BKG COUNT RATE: 0 cpmBETA/GAMMA BKG COUNT RATE: NAα MDA (dpm/100cm²) 69βγ MDA (dpm/100cm²) NASURVEYOR D. L. Vann for Kevin GraczykSUPERVISOR B. Dula

REMOVABLE CONTAMINATION

DATE: 12-7-94 LOCATION: Bausch + Lomb Furnace Area
 COUNTER: Ludlum 2220 ⁵⁰⁰⁶¹~~52836~~ ^{OLV}_{ab} EFFICIENCY: N/A 0.238
 Model Number Alpha Beta/Gamma

DPM/100cm ²			DPM/100cm ²			DPM/100cm ²		
LOCATION	$\beta\gamma$	α	LOCATION	$\beta\gamma$	α	LOCATION	$\beta\gamma$	α
1	<940		21	<940				
2	<940		22	<940				
3	<940		23	<940				
4	<940		24	<940				
5	<940		25	<940				
6	<940		26	<940				
7	<940		27	<940				
8	<940		28	<940				
9	<940	N/A	29	<940	N/A			
10	<940		30	<940				
11	<940						N/A	
12	<940							
13	<940							
14	<940							
15	<940			N/A				
16	<940							
17	<940							
18	<940							
19	<940							
20	<940							

ALPHA BKG COUNT RATE: N/A
 BETA/GAMMA BKG COUNT RATE: 44 cpm

α MDA (dpm/100cm²) N/A
 $\beta\gamma$ MDA (dpm/100cm²) 940

SURVEYOR

SUPERVISOR

AVERAGE CONTAMINATION SURVEY RESULTS

DATE: 12-7-94 LOCATION: Bausch + Lomb Furnace Area
 INSTRUMENT: Ludlum 2220 50067 PROBE: AC-3 EFF: 0.0667
 MODEL NUMBER MODEL

GRID NUMBER	dpm/100cm ²	GRID NUMBER	dpm/100cm ²
1	< 69	20	< 69
2	< 69	21	< 69
3	< 69	22	< 69
4	< 69	23	< 69
5	< 69	24	< 69
6	< 69	25	< 69
7	< 69	26	< 69
8	< 69	27	< 69
9	< 69	28	< 69
10	< 69	29	< 69
11	< 69	30	< 69
12	< 69		
13	< 69		
14	< 69		
15	< 69		
16	< 69		
17	< 69		
18	< 69		
19	< 69		

Average Background counts per minute 0
 MDA (dpm/100cm²) 69

*BASED ON AVERAGE OF 5 DIRECT MEASUREMENTS
 (INITIAL DIRECT MEASUREMENT WAS GREATER THAN 60% OF GUIDELINE VALUE)

$$\frac{\text{dpm}}{100\text{cm}^2} = \frac{(\text{gross count rate}) - (\text{background count rate})}{(\text{efficiency}) \left(\frac{\text{probe area}}{100\text{cm}^2} \right)}$$

SURVEYOR D. L. Jung for Kevin Graczyk
 SUPERVISOR B. Dahn

AVERAGE CONTAMINATION SURVEY RESULTS

Page 5 of 8

DATE: 12-7-94

LOCATION: Bausch & Lomb Furnace Area

INSTRUMENT: Ludlum 2220 52836
MODEL NUMBER

PROBE: 44-9 EFF: 0.239
MODEL

GRID NUMBER	dpm/100cm ²	GRID NUMBER	dpm/100cm ²
1	< 1360	20	< 1360
2	< 1360	21	< 1360
3	< 1360	22	2800
4	< 1360	23	< 1360
5	< 1360	24	< 1360
6	< 1360	25	< 1360
7	< 1360	26	< 1360
8	< 1360	27	< 1360
9	< 1360	28	2520
10	< 1360	29	< 1360
11	3360	30	< 1360
12	< 1360	N A	< 1360 ^{OLV} 1-2-95
13	< 1360		
14	2240		
15	2800		
16	3360		
17	< 1360		
18	< 1360		
19	< 1360		

Average Background counts per minute 98

MDA (dpm/100cm²) 1360

*BASED ON AVERAGE OF 5 DIRECT MEASUREMENTS
(INITIAL DIRECT MEASUREMENT WAS GREATER THAN 60% OF GUIDELINE VALUE)

$$\frac{\text{dpm}}{100\text{cm}^2} = \frac{(\text{gross count rate}) - (\text{background count rate})}{(\text{efficiency}) \quad (\text{probe area})}$$

100cm²

SURVEYOR D. L. Van der
Kevin Graczyk
SUPERVISOR B. D. ...

EXPOSURE RATE SURVEY RESULTS

DATE: 12-7-94LOCATION: Bausch & Lomb Furnace AreaINSTRUMENT: Bicron B218LCAL DUE: 10-14-95

GRID NUMBER	EXPOSURE RATE $\mu\text{R/h}$		GRID NUMBER	EXPOSURE RATE $\mu\text{R/h}$	
	1cm	1m		1cm	1m
1		5uR/hr	18		5uR/hr
2		5uR/hr	19		5uR/hr
3		5uR/hr	20		5uR/hr
4		5uR/hr	21		5uR/hr
5		5uR/hr	22		5uR/hr
6		5uR/hr	23		5uR/hr
7		5uR/hr	24	N/A	5uR/hr
8		4uR/hr	25		5uR/hr
9	N/A	4uR/hr	26		5uR/hr
10		4uR/hr	27		5uR/hr
11		5uR/hr	28		5uR/hr
12		5uR/hr	29		5uR/hr
13		5uR/hr	30		5uR/hr
14		5uR/hr			
15		5uR/hr		N	
16		5uR/hr		A	
17		5uR/hr			

MEASUREMENTS TAKEN AT 1cm AND/OR 1m ABOVE THE FLOOR.

BKG: 5uR/hrSURVEYOR K. R. [Signature]
SUPERVISOR Bronson [Signature]

SURVEYOR D. L. Vann
 REVIEWER B. [Signature]

DATE 12-8-94
 TIME 1000


SURVEY METERS

MODEL Ludlum 2220
 SERIAL # 50061
 CAL DUE 10-17-95
 EFFICIENCY 0.238
 TYPE 44-9
 BKG 98 cpm
 MDA 1366 dpm/100cm²

MODEL Ludlum 2220
 SERIAL # 50067
 CAL DUE 10-17-95
 EFFICIENCY 0.0667
 TYPE AC-3
 BKG 0 cpm
 MDA 69 dpm/100cm²

MODEL Bicron
 SERIAL # B218L
 CAL DUE 10-14-95
 EFFICIENCY NA
 TYPE Micro Rem
 BKG SmR/hr
 MDA NA

LOCATION: Bausch & Lomb Furnace Area (Platform level 2)



NOTE: Smear results in dpm/100 cm² unless otherwise noted. (1) denotes smear location. An * followed by a number in dpm (i.e., * 100 dpm) denotes direct probe readings in dpm/100 cm². # denotes dose rates in mR/hr.

PURPOSE: Unaffected area release survey

SMEAR	dpm/100 cm ²	
#	βγ	α
1	< 940	< 69
2	< 940	< 69
3	< 940	< 69
4	< 940	< 69
5	< 940	< 69
6	< 940	< 69
7	< 940	< 69
8	< 940	< 69
9	< 940	< 69
10	< 940	< 69
11	< 940	< 69
12	< 940	< 69
13	< 940	< 69
14	< 940	< 69
15	< 940	< 69
16	< 940	< 69
17	< 940	< 69
18	< 940	< 69
19	< 940	< 69
20	< 940	< 69

MDA CALCULATION SHEETMETER 2220 w/44-9 SERIAL #: 52836TS: 1 TB: 1 RB: 98 cpm EFF: 0.239 PROBE SIZE: 15 cm²

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{98}{1} + \frac{98}{1} \right)^{1/2}$$

$$0.239 (15 / 100 \text{ cm}^2)$$

$$MDA = \underline{1360 \text{ dpm}/100 \text{ cm}^2}$$

METER 2220 w/44-9 SERIAL #: 50061TS: 1 TB: 1 RB: 98 cpm EFF: 0.238 PROBE SIZE: 15 cm²

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{98}{1} + \frac{98}{1} \right)^{1/2}$$

$$0.238 (15 / 100 \text{ cm}^2)$$

$$MDA = \underline{1366 \text{ dpm}/100 \text{ cm}^2}$$

METER 2220 w/Ac-3 SERIAL #: 50067TS: 1 TB: 1 RB: 0 cpm EFF: 0.0667 PROBE SIZE: 59 cm²

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{0}{1} + \frac{0}{1} \right)^{1/2}$$

$$0.0667 (59 / 100 \text{ cm}^2)$$

$$MDA = \underline{69 \text{ dpm}/100 \text{ cm}^2}$$

METER 2220 w/44-9 SERIAL #: 50061TS: 1 TB: 1 RB: 44 cpm EFF: 0.238 PROBE SIZE: 15 cm²

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{44}{1} + \frac{44}{1} \right)^{1/2}$$

$$0.238 (15 / 100 \text{ cm}^2)$$

$$MDA = \underline{940 \text{ dpm}/100 \text{ cm}^2}$$

Removable
onlyTechnician: D. L. VannDate: 12-7-94Reviewed by: B. GuleDate: 12-7-94

Bausch & Lomb Release Survey Data

Survey Unit: Furnace Area

Activity (dpm/100cm²)

Location Number	Removable α			Removable β - γ		
	Activity	Uncertainty *	MDA	Activity	Uncertainty *	MDA
1	69	16	69	940	64	940
2	69	16	69	940	64	940
3	69	16	69	940	64	940
4	69	16	69	940	64	940
5	69	16	69	940	64	940
6	69	16	69	940	64	940
7	69	16	69	940	64	940
8	69	16	69	940	64	940
9	69	16	69	940	64	940
10	69	16	69	940	64	940
11	69	16	69	940	64	940
12	69	16	69	940	64	940
13	69	16	69	940	64	940
14	69	16	69	940	64	940
15	69	16	69	940	64	940
16	69	16	69	940	64	940
17	69	16	69	940	64	940
18	69	16	69	940	64	940
19	69	16	69	940	64	940
20	69	16	69	940	64	940
21	69	16	69	940	64	940
22	69	16	69	940	64	940
23	69	16	69	940	64	940
24	69	16	69	940	64	940
25	69	16	69	940	64	940
26	69	16	69	940	64	940
27	69	16	69	940	64	940
28	69	16	69	940	64	940
29	69	16	69	940	64	940
30	69	16	69	940	64	940

* Uncertainties represent the 95% Confidence level, based on counting statistics.
When survey data indicated less than MDA, the MDA value was used in the calculation.

Bausch & Lomb Release Survey Data

Survey Unit: Furnace Area

Activity (dpm/100cm²)

Location Number	Direct α			Direct β - γ		
	Activity	Uncertainty *	MDA	Activity	Uncertainty *	MDA
1	69	16	69	1360	79	1360
2	69	16	69	1360	79	1360
3	69	16	69	1360	79	1360
4	69	16	69	1360	79	1360
5	69	16	69	1360	79	1360
6	69	16	69	1360	79	1360
7	69	16	69	1360	79	1360
8	69	16	69	1360	79	1360
9	69	16	69	1360	79	1360
10	69	16	69	1360	79	1360
11	69	16	69	3360	118	1360
12	69	16	69	1360	79	1360
13	69	16	69	1360	79	1360
14	69	16	69	2240	98	1360
15	69	16	69	2800	109	1360
16	69	16	69	3360	118	1360
17	69	16	69	1360	79	1360
18	69	16	69	1360	79	1360
19	69	16	69	1360	79	1360
20	69	16	69	1360	79	1360
21	69	16	69	1360	79	1360
22	69	16	69	2800	109	1360
23	69	16	69	1360	79	1360
24	69	16	69	1360	79	1360
25	69	16	69	1360	79	1360
26	69	16	69	1360	79	1360
27	69	16	69	1360	79	1360
28	69	16	69	2520	104	1360
29	69	16	69	1360	79	1360
30	69	16	69	1360	79	1360

* Uncertainties represent the 95% Confidence level, based on counting statistics.
When survey data indicated less than MDA, the MDA value was used in the calculation.

NINETY-FIVE PERCENT CONFIDENCE CALCULATION SHEET

level 1
Survey Unit furnace Area Date 12-7-94

Meter Lucdm 2220 Serial # 50067

Probe Ac-3 Serial # 712582

MDA 69 dpm/100cm² Survey Type Removable X

Guideline Value 200 (dpm/100cm²)

Average Measurement Level

$$x_{ave} = \frac{1}{n_s} \sum_{i=1}^{n_s} x_i$$

where:

x_{ave} = calculated mean for a survey unit,
 n_s = number of measurements within a survey unit,
 x_i = systematic and random measurements at point (i)

$\sum x_i$ 2070

n_s 30

x_{ave} = 69

Standard Deviation

$$S_x = \sqrt{\frac{\sum_{i=1}^{n_s} (x_{ave} - x_i)^2}{n_s - 1}}$$

where:

S_x = standard deviation
 n_s = number of measurements within a survey unit
 x_i = systematic and random measurements at point (i)
 x_{ave} = calculated mean for a survey unit

x_{ave} = 69 n_s = 30

$\sum (x_{ave} - x_i)^2$ = 0 S_x = 0

$$\mu_{\alpha} = x_{ave} + t_{1-\alpha,df} \frac{s_x}{\sqrt{n_s}}$$

where:

μ_{α} = value compared guideline value to determine 95% Confidence Level
 x_{ave} = calculated mean for a survey unit,
 $t_{1-\alpha,df}$ = 95% confidence level for n-1 degrees of freedom (from NUREG/CR-5849, Table B-1, "Factors for Comparison of Survey Data")
 s_x = standard deviation of measurements in a survey unit, and
 n_s = number of measurements within a survey unit used to determine \bar{x} and s .

x_{ave} = 69
 $t_{1-\alpha,df}$ = 1.740
 s_x = 0
 n_s = 18
 μ_{α} = 69

NOTE: When determining x_{ave} , only minimum detectable activity (MDA) values were available for some measurements locations; the MDA values were therefore used as actual activity levels for the purpose of performing this calculation.

The site specific guideline values for this project are 5000 dpm/100 cm² fixed plus removable and 1000 dpm/100 cm² removable for beta-gamma emitting nuclides, and 1000 dpm/100 cm² fixed plus removable and 200 dpm/100 cm² removable for alpha emitting nuclides. If μ_{α} is less than the specific guideline value, the data for this survey unit satisfies the guideline at the 95% confidence level. If μ_{α} is greater than the specific guideline value, further remediation of the survey unit is necessary.

Individual Completing Form: J. Allen Date: 1-10-95

Reviewed By: R. Kight Date: 1/13/95

NINETY-FIVE PERCENT CONFIDENCE CALCULATION SHEET

Survey Unit Batch Room N. Wall Date 12-11-94

Meter Ludlum 2220 Serial # 50061

Probe 44-9 Serial # PRO 66761

MDA 940 (dpm/100cm²) Survey Type removable β-γ

Guideline Value 1000 (dpm/100cm²)

Average Measurement Level

$$x_{ave} = \frac{1}{n_s} \sum_{i=1}^{n_s} x_i$$

where:

x_{ave} = calculated mean for a survey unit,
 n_s = number of measurements within a survey unit,
 x_i = systematic and random measurements at point (i)

$\sum x_i$ 16920

n_s 18

x_{ave} = 940

Standard Deviation

$$S_x = \sqrt{\frac{\sum_{i=1}^{n_s} (x_{ave} - x_i)^2}{n_s - 1}}$$

where:

S_x = standard deviation
 n_s = number of measurements within a survey unit
 x_i = systematic and random measurements at point (i)
 x_{ave} = calculated mean for a survey unit

x_{ave} = 940

n_s = 18

$\sum (x_{ave} - x_i)^2$ = 0

S_x = 0

$$\mu_{\alpha} = x_{ave} + t_{1-\alpha,df} \frac{s_x}{\sqrt{n_s}}$$

where:

μ_{α} = value compared guideline value to determine 95% Confidence Level
 x_{ave} = calculated mean for a survey unit,
 $t_{1-\alpha,df}$ = 95% confidence level for n-1 degrees of freedom (from NUREG/CR-5849, Table B-1, "Factors for Comparison of Survey Data")
 s_x = standard deviation of measurements in a survey unit, and
 n_s = number of measurements within a survey unit used to determine \bar{x} and s .

x_{ave} = 940

$t_{1-\alpha,df}$ = 1.740

s_x = 0

n_s = 18

μ_{α} = 940

NOTE: When determining x_{ave} , only minimum detectable activity (MDA) values were available for some measurements locations; the MDA values were therefore used as actual activity levels for the purpose of performing this calculation.

The site specific guideline values for this project are 5000 dpm/100 cm² fixed plus removable and 1000 dpm/100 cm² removable for beta-gamma emitting nuclides, and 1000 dpm/100 cm² fixed plus removable and 200 dpm/100 cm² removable for alpha emitting nuclides. If μ_{α} is less than the specific guideline value, the data for this survey unit satisfies the guideline at the 95% confidence level. If μ_{α} is greater than the specific guideline value, further remediation of the survey unit is necessary.

Individual Completing Form: S. Adam Date: 1-10-95

Reviewed By: R. Kight Date: 1/13/95

NINETY-FIVE PERCENT CONFIDENCE CALCULATION SHEET

Survey Unit Batch Room N. Wall Date 12-11-94

Meter Cudlum 2220 Serial # 50067

Probe AC-3 Serial # 712582

MDA 69 (dpm/100cm²) Survey Type direct

Guideline Value 1000 (dpm/100cm²)

Average Measurement Level

$$x_{ave} = \frac{1}{n_s} \sum_{i=1}^{n_s} x_i$$

where:

x_{ave} = calculated mean for a survey unit,
 n_s = number of measurements within a survey unit,
 x_i = systematic and random measurements at point (i)

$\sum x_i$ 1242 n_s 18 x_{ave} = 69

Standard Deviation

$$S_x = \sqrt{\frac{\sum_{i=1}^{n_s} (x_{ave} - x_i)^2}{n_s - 1}}$$

where:

S_x = standard deviation
 n_s = number of measurements within a survey unit
 x_i = systematic and random measurements at point (i)
 x_{ave} = calculated mean for a survey unit

x_{ave} = 69 n_s = 18

$\sum (x_{ave} - x_i)^2$ = 0 S_x = 0

$$\mu_{\alpha} = x_{ave} + t_{1-\alpha,df} \frac{s_x}{\sqrt{n_s}}$$

where:

μ_{α} = value compared guideline value to determine 95% Confidence Level
 x_{ave} = calculated mean for a survey unit,
 $t_{1-\alpha,df}$ = 95% confidence level for n-1 degrees of freedom (from NUREG/CR-5849, Table B-1, "Factors for Comparison of Survey Data")
 s_x = standard deviation of measurements in a survey unit, and
 n_s = number of measurements within a survey unit used to determine \bar{x} and s .

x_{ave} = 69

$t_{1-\alpha,df}$ = 1.740

s_x = 0

n_s = 18

μ_{α} = 69

NOTE: When determining x_{ave} , only minimum detectable activity (MDA) values were available for some measurements locations; the MDA values were therefore used as actual activity levels for the purpose of performing this calculation.

The site specific guideline values for this project are 5000 dpm/100 cm² fixed plus removable and 1000 dpm/100 cm² removable for beta-gamma emitting nuclides, and 1000 dpm/100 cm² fixed plus removable and 200 dpm/100 cm² removable for alpha emitting nuclides. If μ_{α} is less than the specific guideline value, the data for this survey unit satisfies the guideline at the 95% confidence level. If μ_{α} is greater than the specific guideline value, further remediation of the survey unit is necessary.

Individual Completing Form: J. Shum Date: 1/10/95

Reviewed By: R. Kight Date: 1/13/95

NINETY-FIVE PERCENT CONFIDENCE CALCULATION SHEET

Survey Unit Batch Room N. Wall Date 12/11/94

Meter Ludlum 2220 Serial # 50061

Probe 44-9 Serial # PRO 66761

MDA 1366 (dpm/100cm²) Survey Type direct B-7

Guideline Value 5000 (dpm/100cm²)

Average Measurement Level

$$x_{ave} = \frac{1}{n_s} \sum_{i=1}^{n_s} x_i$$

where:

x_{ave} = calculated mean for a survey unit,
 n_s = number of measurements within a survey unit,
 x_i = systematic and random measurements at point (i)

Σx_i 24588

n_s 18

x_{ave} = 1366

Standard Deviation

$$S_x = \sqrt{\frac{\sum_{i=1}^{n_s} (x_{ave} - x_i)^2}{n_s - 1}}$$

where:

S_x = standard deviation
 n_s = number of measurements within a survey unit
 x_i = systematic and random measurements at point (i)
 x_{ave} = calculated mean for a survey unit

x_{ave} = 1366

n_s = 18

$\Sigma (x_{ave} - x_i)^2$ = 0

S_x = 0

$$\mu_{\alpha} = x_{ave} + t_{1-\alpha,df} \frac{s_x}{\sqrt{n_s}}$$

where:

μ_{α} = value compared guideline value to determine 95% Confidence Level
 x_{ave} = calculated mean for a survey unit,
 $t_{1-\alpha,df}$ = 95% confidence level for n-1 degrees of freedom (from NUREG/CR-5849, Table B-1, "Factors for Comparison of Survey Data")
 s_x = standard deviation of measurements in a survey unit, and
 n_s = number of measurements within a survey unit used to determine \bar{x} and s .

x_{ave} = 1366

$t_{1-\alpha,df}$ = 1.740

s_x = 0

n_s = 18

μ_{α} = 1366

NOTE: When determining x_{ave} , only minimum detectable activity (MDA) values were available for some measurements locations; the MDA values were therefore used as actual activity levels for the purpose of performing this calculation.

The site specific guideline values for this project are 5000 dpm/100 cm² fixed plus removable and 1000 dpm/100 cm² removable for beta-gamma emitting nuclides, and 1000 dpm/100 cm² fixed plus removable and 200 dpm/100 cm² removable for alpha emitting nuclides. If μ_{α} is less than the specific guideline value, the data for this survey unit satisfies the guideline at the 95% confidence level. If μ_{α} is greater than the specific guideline value, further remediation of the survey unit is necessary.

Individual Completing Form: J. Ahoo Date: 1/16/95

Reviewed By: P. Kijit Date: 1/13/95

NINETY-FIVE PERCENT CONFIDENCE CALCULATION SHEET

Survey Unit Batch Room N. Wall Date 12-11-94

Meter Bicron Serial # B218L

Probe NA Serial # NA

MDA 14A (dpm/100cm²) Survey Type Exposure

Guideline Value 10 mRem/hr (~~dpm/100cm²~~) 1-13-95

Average Measurement Level

$$x_{ave} = \frac{1}{n_s} \sum_{i=1}^{n_s} x_i$$

where:

x_{ave} = calculated mean for a survey unit,
 n_s = number of measurements within a survey unit,
 x_i = systematic and random measurements at point (i)

$\sum x_i$ 93

n_s 18

x_{ave} = 5.2

Standard Deviation

$$S_x = \sqrt{\frac{\sum_{i=1}^{n_s} (x_{ave} - x_i)^2}{n_s - 1}}$$

where:

S_x = standard deviation
 n_s = number of measurements within a survey unit
 x_i = systematic and random measurements at point (i)
 x_{ave} = calculated mean for a survey unit

x_{ave} = 5.2

n_s = 18

$\sum (x_{ave} - x_i)^2$ = 2.5

S_x = 0.4

$$\mu_{\alpha} = x_{ave} + t_{1-\alpha,df} \frac{s_x}{\sqrt{n_s}}$$

where:

μ_{α} = value compared guideline value to determine 95 % Confidence Level
 x_{ave} = calculated mean for a survey unit,
 $t_{1-\alpha,df}$ = 95% confidence level for n-1 degrees of freedom (from NUREG/CR-5849, Table B-1, "Factors for Comparison of Survey Data")
 s_x = standard deviation of measurements in a survey unit, and
 n_s = number of measurements within a survey unit used to determine \bar{x} and s .

x_{ave} = 5.2
 $t_{1-\alpha,df}$ = 1.740
 s_x = 0.4
 n_s = 18
 μ_{α} = 5.3

NOTE: When determining x_{ave} , only minimum detectable activity (MDA) values were available for some measurements locations; the MDA values were therefore used as actual activity levels for the purpose of performing this calculation.

The site specific guideline values for this project are 5000 dpm/100 cm² fixed plus removable and 1000 dpm/100 cm² removable for beta-gamma emitting nuclides, and 1000 dpm/100 cm² fixed plus removable and 200 dpm/100 cm² removable for alpha emitting nuclides. If μ_{α} is less than the specific guideline value, the data for this survey unit satisfies the guideline at the 95% confidence level. If μ_{α} is greater than the specific guideline value, further remediation of the survey unit is necessary.

Individual Completing Form: D. L. Vann Date: 12-11-94

Reviewed By: R. Bight Date: 1/13/95

REMOVABLE CONTAMINATION

DATE: 12-11-94

End/In m	2220	50061
COUNTER: End/In m	2220	50067
Model		Number

LOCATION: Bausch + Lomb Batch Room
South wall

EFFICIENCY: 0.0467 0.238
Alpha Beta/Gamma

<u>DPM/100cm²</u>			<u>DPM/100cm²</u>			<u>DPM/100cm²</u>		
<u>LOCATION</u>	<u>βγ</u>	<u>α</u>	<u>LOCATION</u>	<u>βγ</u>	<u>α</u>	<u>LOCATION</u>	<u>βγ</u>	<u>α</u>
A1	< 940	< 69						
A2	< 940	< 69						
A3	< 940	< 69						
A4	< 940	< 69						
A5	< 940	< 69						
A6	< 940	< 69						
A7	< 940	< 69						
A8	< 940	< 69						
A9	< 940	< 69						
B1	< 940	< 69	N A			N A		
B2	< 940	< 69						
B3	< 940	< 69						
B4	< 940	< 69						
B5	< 940	< 69						
B6	< 940	< 69						
B7	< 940	< 69						
B8	< 940	< 69						
B9	< 940	< 69						
	N A							

ALPHA BKG COUNT RATE: 0 cpm

BETA/GAMMA BKG COUNT RATE: 4/4 cpm

α MDA (dpm/100cm²) 69

By MDA (dpm/100cm²) 940

SURVEYOR D. L. Vann

SUPERVISOR B. Gula

AVERAGE CONTAMINATION SURVEY RESULTS

DATE: 12-11-94

LOCATION: Bausch + Lomb batch Room South wing

INSTRUMENT: Ludlum 2220 50067
MODEL NUMBER

PROBE: AC-3 EFF: 0.0667
MODEL

GRID NUMBER	dpm/100cm2	GRID NUMBER	dpm/100cm2
A1	< 69		
A2	< 69		
A3	< 69		
A4	* 276		
A5	< 69		
A6	* 164		
A7	< 69		
A8	< 69		
A9	< 69		
B1	< 69		
B2	< 69		
B3	< 69		
B4	< 69		
B5	< 69		
B6	< 69		
B7	< 69		
B8	< 69		
B9	< 69		
N A			

Average Background counts per minute 0

MDA (dpm/100cm²) 69

*BASED ON AVERAGE OF 5 DIRECT MEASUREMENTS
(INITIAL DIRECT MEASUREMENT WAS GREATER THAN 60% OF GUIDELINE VALUE)

$$\frac{\text{dpm}}{100\text{cm}^2} = \frac{(\text{gross count rate}) - (\text{background count rate})}{(\text{efficiency}) \times \frac{(\text{probe area})}{100\text{cm}^2}}$$

D.L. Vann for
SURVEYOR Kevin Graczyk
SUPERVISOR B. G. [Signature]

AVERAGE CONTAMINATION SURVEY RESULTS

DATE: 12-11-94LOCATION: Bausch + Lomb Patch Room South wallINSTRUMENT: Ludlum 2200 50061
MODEL NUMBERPROBE: 44-9 EFF: 0.238
MODEL

GRID NUMBER	dpm/100cm ²	GRID NUMBER	dpm/100cm ²
A1	<1366		
A2	<1366		
A3	<1366		
A4	* <1366		
A5	* <1366		
A6	* <1366		
A7	<1366		
A8	* <1366		
A9	<1366		
B1	<1366		
B2	<1366		
B3	<1366		
B4	<1366		
B5	<1366		
B6	<1366		
B7	<1366		
B8	<1366		
B9	<1366		

Average Background counts per minute 98MDA (dpm/100cm²) 1366

*BASED ON AVERAGE OF 5 DIRECT MEASUREMENTS
(INITIAL DIRECT MEASUREMENT WAS GREATER THAN 60% OF GUIDELINE VALUE)

$$\frac{\text{dpm}}{100\text{cm}^2} = \frac{(\text{gross count rate}) - (\text{background count rate})}{(\text{efficiency}) (\text{probe area})}$$

100cm²

SURVEYOR Kim R. [Signature]SUPERVISOR Brandon [Signature]

EXPOSURE RATE SURVEY RESULTS

DATE: 12-11-94LOCATION: Bausch & Lomb Batch Room South WallINSTRUMENT: Bicron B218LCAL DUE: 10-14-95



EXPOSURE RATE			EXPOSURE RATE		
GRID NUMBER		$\mu R/h$	GRID NUMBER		$\mu R/h$
A1	1cm	7 ^{plv} $\mu R/hr$		1cm	1m
A2		7 $\mu R/hr$			
A3		7 $\mu R/hr$			
A4		7 $\mu R/hr$			
A5		7 $\mu R/hr$			
A6		8 $\mu R/hr$			
A7		7 $\mu R/hr$			
A8		6 $\mu R/hr$			
A9	N/A	6 $\mu R/hr$	N/A		
B1		6 $\mu R/hr$	A		
B2		6 $\mu R/hr$			
B3		6 $\mu R/hr$			
B4		7 $\mu R/hr$			
B5		7 $\mu R/hr$			
B6		7 $\mu R/hr$			
B7		6 $\mu R/hr$			
B8		6 $\mu R/hr$			
B9		6 $\mu R/hr$			

MEASUREMENTS TAKEN AT 1cm AND/OR 1m ABOVE THE FLOOR.

BKG: 7 $\mu R/hr$

SURVEYOR

SUPERVISOR

MDA CALCULATION SHEET

METER 2220 w/44-9 SERIAL #: 50061TS: 1 TB: 1 RB: 98 cpm EFF: 0.238 PROBE SIZE: 15 cm²

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{98}{1} + \frac{98}{1} \right)^{1/2}$$

$$0.238 (15 / 100 \text{ cm}^2)$$

$$MDA = \underline{1366 \text{ dpm}/100 \text{ cm}^2}$$

METER 2220 w/Hc-3 SERIAL #: 50067TS: 1 TB: 1 RB: 0 cpm EFF: 0.0667 PROBE SIZE: 59 cm²

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{0}{1} + \frac{0}{1} \right)^{1/2}$$

$$0.0667 (59 / 100 \text{ cm}^2)$$

$$MDA = \underline{69 \text{ dpm}/100 \text{ cm}^2}$$

METER 2220 w/44-9 SERIAL #: 50061TS: 1 TB: 1 RB: 44 cpm EFF: 0.238 PROBE SIZE: 15 cm²

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{44}{1} + \frac{44}{1} \right)^{1/2}$$

$$0.238 (15 / 100 \text{ cm}^2)$$

$$MDA = \underline{940 \text{ dpm}/100 \text{ cm}^2}$$

Removable
only

METER _____ SERIAL #: _____

TS: _____ TB: _____ RB: _____ EFF: _____ PROBE SIZE: _____

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{\quad}{1} + \frac{\quad}{1} \right)^{1/2}$$

$$(\quad / 100 \text{ cm}^2)$$

$$MDA = \underline{\quad}$$

Technician: D. L. Vann Date: 12-11-94Reviewed by: B. Bul Date: 12-11-94

Bausch & Lomb Release Survey Data

Survey Unit: Batch Room South Wall

Activity (dpm/100cm²)

<u>Grid No.</u>	Removable α			Removable β - γ		
	<u>Activity</u>	<u>Uncertainty *</u>	<u>MDA</u>	<u>Activity</u>	<u>Uncertainty *</u>	<u>MDA</u>
A- 1	69	16	69	940	64	940
A- 2	69	16	69	940	64	940
A- 3	69	16	69	940	64	940
A- 4	69	16	69	940	64	940
A- 5	69	16	69	940	64	940
A- 6	69	16	69	940	64	940
A- 7	69	16	69	940	64	940
A- 8	69	16	69	940	64	940
A- 9	69	16	69	940	64	940
B- 1	69	16	69	940	64	940
B- 2	69	16	69	940	64	940
B- 3	69	16	69	940	64	940
B- 4	69	16	69	940	64	940
B- 5	69	16	69	940	64	940
B- 6	69	16	69	940	64	940
B- 7	69	16	69	940	64	940
B- 8	69	16	69	940	64	940
B- 9	69	16	69	940	64	940

* Uncertainties represent the 95% Confidence level, based on counting statistics.
When survey data indicated less than MDA, the MDA value was used in the calculation.

Bausch & Lomb Release Survey Data

Survey Unit: Batch Room South Wall

Activity (dpm/100cm²)

<u>Grid No.</u>	<u>Direct α</u>			<u>Direct β-γ</u>		
	<u>Activity</u>	<u>Uncertainty *</u>	<u>MDA</u>	<u>Activity</u>	<u>Uncertainty *</u>	<u>MDA</u>
A- 1	69	16	69	1366	79	1366
A- 2	69	16	69	1366	79	1366
A- 3	69	16	69	1366	79	1366
A- 4	276	33	69	1366	79	1366
A- 5	69	16	69	1366	79	1366
A- 6	164	25	69	1366	79	1366
A- 7	69	16	69	1366	79	1366
A- 8	69	16	69	1366	79	1366
A- 9	69	16	69	1366	79	1366
B- 1	69	16	69	1366	79	1366
B- 2	69	16	69	1366	79	1366
B- 3	69	16	69	1366	79	1366
B- 4	69	16	69	1366	79	1366
B- 5	69	16	69	1366	79	1366
B- 6	69	16	69	1366	79	1366
B- 7	69	16	69	1366	79	1366
B- 8	69	16	69	1366	79	1366
B- 9	69	16	69	1366	79	1366

* Uncertainties represent the 95% Confidence level, based on counting statistics.
When survey data indicated less than MDA, the MDA value was used in the calculation.

$$\mu_{\alpha} = x_{ave} + t_{1-\alpha, df} \frac{s_x}{\sqrt{n_s}}$$

where:

μ_{α} = value compared guideline value to determine 95 % Confidence Level
 x_{ave} = calculated mean for a survey unit,
 $t_{1-\alpha, df}$ = 95 % confidence level for n-1 degrees of freedom (from NUREG/CR-5849, Table B-1, "Factors for Comparison of Survey Data")
 s_x = standard deviation of measurements in a survey unit, and
 n_s = number of measurements within a survey unit used to determine x_{ave} and s_x .

x_{ave} = 69
 $t_{1-\alpha, df}$ = 1.699
 s_x = 0
 n_s = 30
 μ_{α} = 69

NOTE: When determining x_{ave} , only minimum detectable activity (MDA) values were available for some measurements locations; the MDA values were therefore used as actual activity levels for the purpose of performing this calculation.

The site specific guideline values for this project are 5000 dpm/100 cm² fixed plus removable and 1000 dpm/100 cm² removable for beta-gamma emitting nuclides, and 1000 dpm/100 cm² fixed plus removable and 200 for alpha emitting nuclides. If μ_{α} is less than the specific guideline value, the data for this survey unit satisfies the guideline at the 95 % confidence level. If μ_{α} is greater than the specific guideline value, further remediation of the survey unit is necessary.

Individual Completing Form: D. L. Vann Date: 12-7-94

Reviewed By: R. Kuyt Date: 1/13/95

NINETY-FIVE PERCENT CONFIDENCE CALCULATION SHEET

Survey Unit Level 1 Furnace Area Date 12-7-94
 Meter Ludlum 2220 Serial # 50061
 Probe 44-9 Serial # PRO 66761
 MDA 940 (dpm/100cm²) Survey Type Removable B-8
 Guideline Value 1000 (dpm/100cm²)

Average Measurement Level

$$x_{ave} = \frac{1}{n_s} \sum_{i=1}^{n_s} x_i$$

where:

x_{ave} = calculated mean for a survey unit,
 n_s = number of measurements within a survey unit,
 x_i = systematic and random measurements at point (i)

$\sum x_i$ 28,200

n_s 30

x_{ave} = 940

Standard Deviation

$$S_x = \sqrt{\frac{\sum_{i=1}^{n_s} (x_{ave} - x_i)^2}{n_s - 1}}$$

where:

S_x = standard deviation
 n_s = number of measurements within a survey unit
 x_i = systematic and random measurements at point (i)
 x_{ave} = calculated mean for a survey unit

x_{ave} = 940

n_s = 30

$\sum (x_{ave} - x_i)^2$ = 0

S_x = 0

$$\mu_{\alpha} = x_{ave} + t_{1-\alpha,df} \frac{s_x}{\sqrt{n_s}}$$

where:

μ_{α} = value compared guideline value to determine 95% Confidence Level
 x_{ave} = calculated mean for a survey unit,
 $t_{1-\alpha,df}$ = 95% confidence level for n-1 degrees of freedom (from NUREG/CR-5849, Table B-1, "Factors for Comparison of Survey Data")
 s_x = standard deviation of measurements in a survey unit, and
 n_s = number of measurements within a survey unit used to determine x_{ave} and s_x .

x_{ave} = 940
 $t_{1-\alpha,df}$ = 1.699
 s_x = 0
 n_s = 30
 μ_{α} = 940

NOTE: When determining x_{ave} , only minimum detectable activity (MDA) values were available for some measurements locations; the MDA values were therefore used as actual activity levels for the purpose of performing this calculation.

The site specific guideline values for this project are 5000 dpm/100 cm² fixed plus removable and 1000 dpm/100 cm² removable for beta-gamma emitting nuclides, and 1000 dpm/100 cm² fixed plus removable and 200 dpm/100 cm² removable for alpha emitting nuclides. If μ_{α} is less than the specific guideline value, the data for this survey unit satisfies the guideline at the 95% confidence level. If μ_{α} is greater than the specific guideline value, further remediation of the survey unit is necessary.

Individual Completing Form: D. L. Vann Date: 12-7-94

Reviewed By: R. K. LA Date: 1/13/95

NINETY-FIVE PERCENT CONFIDENCE CALCULATION SHEET

Survey Unit level 1 furnace Area Date 12-7-94

Meter Lucas 2220 Serial # 50067

Probe AC-3 Serial # 712582

MDA 69 (dpm/100cm²) Survey Type Direct &

Guideline Value 1000 (dpm/100cm²)

Average Measurement Level

$$x_{ave} = \frac{1}{n_s} \sum_{i=1}^{n_s} x_i$$

where:

x_{ave} = calculated mean for a survey unit,
 n_s = number of measurements within a survey unit,
 x_i = systematic and random measurements at point (i)

Σx_i 2070

n_s 30

x_{ave} = 69

Standard Deviation

$$S_x = \sqrt{\frac{\sum_{i=1}^{n_s} (x_{ave} - x_i)^2}{n_s - 1}}$$

where:

S_x = standard deviation
 n_s = number of measurements within a survey unit
 x_i = systematic and random measurements at point (i)
 x_{ave} = calculated mean for a survey unit

x_{ave} = 69

n_s = 30

$\Sigma (x_{ave} - x_i)^2$ = 0

S_x = 0

$$\mu_{\alpha} = x_{ave} + t_{1-\alpha,df} \frac{s_x}{\sqrt{n_s}}$$

where:

μ_{α} = value compared guideline value to determine 95% Confidence Level
 x_{ave} = calculated mean for a survey unit,
 $t_{1-\alpha,df}$ = 95% confidence level for n-1 degrees of freedom (from NUREG/CR-5849, Table B-1, "Factors for Comparison of Survey Data")
 s_x = standard deviation of measurements in a survey unit, and
 n_s = number of measurements within a survey unit used to determine \bar{x} and s .

$$x_{ave} = \underline{69}$$

$$t_{1-\alpha,df} = \underline{1.699}$$

$$s_x = \underline{0}$$

$$n_s = \underline{30}$$

$$\mu_{\alpha} = \underline{69}$$

NOTE: When determining x_{ave} , only minimum detectable activity (MDA) values were available for some measurements locations; the MDA values were therefore used as actual activity levels for the purpose of performing this calculation.

The site specific guideline values for this project are 5000 dpm/100 cm² fixed plus removable and 1000 dpm/100 cm² removable for beta-gamma emitting nuclides, and 1000 dpm/100 cm² fixed plus removable and 200 dpm/100 cm² removable for alpha emitting nuclides. If μ_{α} is less than the specific guideline value, the data for this survey unit satisfies the guideline at the 95% confidence level. If μ_{α} is greater than the specific guideline value, further remediation of the survey unit is necessary.

Individual Completing Form: D. L. Vann Date: 12-7-94

Reviewed By: R. Hight Date: 1/13/95

NINETY-FIVE PERCENT CONFIDENCE CALCULATION SHEET

Survey Unit level 1 Furnace Area Date 12-7-94
 Meter Ludlum 2220 Serial # 52836
 Probe 44-9 Serial # PRO 68918
 MDA 1360 (dpm/100cm²) Survey Type Direct B-8
 Guideline Value 5000 (dpm/100cm²)

Average Measurement Level

$$x_{ave} = \frac{1}{n_s} \sum_{i=1}^{n_s} x_i$$

where:

x_{ave} = calculated mean for a survey unit,
 n_s = number of measurements within a survey unit,
 x_i = systematic and random measurements at point (i)

$\sum x_i$ 49720

n_s 30

x_{ave} = 1657.3

Standard Deviation

$$S_x = \sqrt{\frac{\sum_{i=1}^{n_s} (x_{ave} - x_i)^2}{n_s - 1}}$$

where:

S_x = standard deviation
 n_s = number of measurements within a survey unit
 x_i = systematic and random measurements at point (i)
 x_{ave} = calculated mean for a survey unit

x_{ave} = 1657.3

n_s = 30

$\sum (x_{ave} - x_i)^2$ = 11,614,987

s_x = 632.9

$$\mu_{\alpha} = x_{ave} + t_{1-\alpha,df} \frac{s_x}{\sqrt{n_s}}$$

where:

μ_{α} = value compared guideline value to determine 95% Confidence Level
 x_{ave} = calculated mean for a survey unit,
 $t_{1-\alpha,df}$ = 95% confidence level for n-1 degrees of freedom (from NUREG/CR-5849, Table B-1, "Factors for Comparison of Survey Data")
 s_x = standard deviation of measurements in a survey unit, and
 n_s = number of measurements within a survey unit used to determine x_{ave} and s_x .

$$x_{ave} = \underline{1657.3}$$

$$t_{1-\alpha,df} = \underline{1.699}$$

$$s_x = \underline{632.9}$$

$$n_s = \underline{30}$$

$$\mu_{\alpha} = \underline{1853.6}$$

NOTE: When determining x_{ave} , only minimum detectable activity (MDA) values were available for some measurements locations; the MDA values were therefore used as actual activity levels for the purpose of performing this calculation.

The site specific guideline values for this project are 5000 dpm/100 cm² fixed plus removable and 1000 dpm/100 cm² removable for beta-gamma emitting nuclides, and 1000 dpm/100 cm² fixed plus removable and 200 dpm/100 cm² removable for alpha emitting nuclides. If μ_{α} is less than the specific guideline value, the data for this survey unit satisfies the guideline at the 95% confidence level. If μ_{α} is greater than the specific guideline value, further remediation of the survey unit is necessary.

Individual Completing Form: D. L. Vann Date: 12-7-94

Reviewed By: R. K. L. T. Date: 1/13/95

NINETY-FIVE PERCENT CONFIDENCE CALCULATION SHEET

Survey Unit furnace Area Date 12-7-74
 Meter Bicron Serial # B2182
 Probe NA Serial # NA
 MDA NA (dpm/100cm²) Survey Type Exposure
 Guideline Value 10 mRem/hr (~~dpm/100cm²~~)

Average Measurement Level

$$x_{ave} = \frac{1}{n_s} \sum_{i=1}^{n_s} x_i$$

where:

x_{ave} = calculated mean for a survey unit,
 n_s = number of measurements within a survey unit,
 x_i = systematic and random measurements at point (i)

$\sum x_i$ 147 n_s 30 x_{ave} = 4.9

Standard Deviation

$$S_x = \sqrt{\frac{\sum_{i=1}^{n_s} (x_{ave} - x_i)^2}{n_s - 1}}$$

where:

S_x = standard deviation
 n_s = number of measurements within a survey unit
 x_i = systematic and random measurements at point (i)
 x_{ave} = calculated mean for a survey unit

x_{ave} = 4.9 n_s = 30
 $\sum (x_{ave} - x_i)^2$ = 2.7 S_x = 0.3

$$\mu_{\alpha} = x_{ave} + t_{1-\alpha,df} \frac{s_x}{\sqrt{n_s}}$$

where:

μ_{α} = value compared guideline value to determine 95% Confidence Level
 x_{ave} = calculated mean for a survey unit,
 $t_{1-\alpha,df}$ = 95% confidence level for n-1 degrees of freedom (from NUREG/CR-5849, Table B-1, "Factors for Comparison of Survey Data")
 s_x = standard deviation of measurements in a survey unit, and
 n_s = number of measurements within a survey unit used to determine \bar{x}_e and s_x .

x_{ave} = 4.9
 $t_{1-\alpha,df}$ = 1.699
 s_x = 0.3
 n_s = 30
 μ_{α} = 5.0

NOTE: When determining x_{ave} , only minimum detectable activity (MDA) values were available for some measurements locations; the MDA values were therefore used as actual activity levels for the purpose of performing this calculation.

The site specific guideline values for this project are 5000 dpm/100 cm² fixed plus removable and 1000 dpm/100 cm² removable for beta-gamma emitting nuclides, and 1000 dpm/100 cm² fixed plus removable and 200 dpm/100 cm² removable for alpha emitting nuclides. If μ_{α} is less than the specific guideline value, the data for this survey unit satisfies the guideline at the 95% confidence level. If μ_{α} is greater than the specific guideline value, further remediation of the survey unit is necessary.

Individual Completing Form: D. L. Vann Date: 12-7-94

Reviewed By: R. Kijl Date: 1/13/95

Section 3

Affected Area Surveys

Batch Room North Wall
Batch Room South Wall
Batch Room West Wall
Batch Room East Wall
Batch Room Floor
Batch Room Overheads
Batch Room Miscellaneous Equipment
Batch Room Scale
Batch Room Hopper
Batch Room Bird Cage

REMOVABLE CONTAMINATION

DATE: 12-11-94

LOCATION: Bensch & Lomb Room North Wall

 COUNTER: Lucium 2220 50061 $\beta\gamma$
 Model Number

 EFFICIENCY: 0.0667 0.238
 Alpha Beta/Gamma

DPM/100cm ²			DPM/100cm ²			DPM/100cm ²		
LOCATION	$\beta\gamma$	α	LOCATION	$\beta\gamma$	α	LOCATION	$\beta\gamma$	α
A1	< 940	< 69						
A2	< 940	< 69						
A3	< 940	< 69						
A4	< 940	< 69						
A5	< 940	< 69						
A6	< 940	< 69						
A7	< 940	< 69						
A8	< 940	< 69						
A9	< 940	< 69						
B1	< 940	< 69						
B2	< 940	< 69						
B3	< 940	< 69						
B4	< 940	< 69						
B5	< 940	< 69						
B6	< 940	< 69						
B7	< 940	< 69						
B8	< 940	< 69						
B9	< 940	< 69						
N								
A								

ALPHA BKG COUNT RATE: 0 cpm

 α MDA (dpm/100cm²) 69

BETA/GAMMA BKG COUNT RATE: 44 cpm

 $\beta\gamma$ MDA (dpm/100cm²) 940

SURVEYOR

SUPERVISOR D. L. Van

AVERAGE CONTAMINATION SURVEY RESULTS

Page 3 of 6

DATE: 12-11-94

LOCATION: Bunsch + Lomb Butch Room
North Wall

INSTRUMENT: Ludlum 2220 50067
MODEL NUMBER

PROBE: Ac-3 EFF: 0.0667
MODEL

GRID NUMBER	dpm/100cm ²	GRID NUMBER	dpm/100cm ²
A1	< 69		
A2	< 69		
A3	< 69		
A4	< 69		
A5	< 69		
A6	< 69		
A7	< 69		
A8	< 69		
A9	< 69		
B1	< 69		
B2	< 69		
B3	< 69		
B4	< 69		
B5	< 69		
B6	< 69		
B7	< 69		
B8	< 69		
B9	< 69		

Average Background counts per minute 0
MDA (dpm/100cm²) 69

*BASED ON AVERAGE OF 5 DIRECT MEASUREMENTS
(INITIAL DIRECT MEASUREMENT WAS GREATER THAN 60% OF GUIDELINE VALUE)

$$\frac{\text{dpm}}{100\text{cm}^2} = \frac{(\text{gross count rate}) - (\text{background count rate})}{(\text{efficiency}) \left(\frac{\text{probe area}}{100\text{cm}^2} \right)}$$

D. L. Vann for
SURVEYOR Kevin Graczyk
SUPERVISOR B. Dule

AVERAGE CONTAMINATION SURVEY RESULTS

DATE: 12-11-94LOCATION: Bausch + Lomb Batch Room North WallINSTRUMENT: Ludlum 2220 650061
MODEL NUMBERPROBE: 44-9 EFF: 0.238
MODEL

GRID NUMBER	dpm/100cm ²	GRID NUMBER	dpm/100cm ²
A1	< 1366		
A2	< 1366		
A3	< 1366		
A4	< 1366		
A5	< 1366		
A6	< 1366		
A7	< 1366		
A8	< 1366		
A9	< 1366		
B1	< 1366		
B2	< 1366		
B3	< 1366		
B4	< 1366		
B5	< 1366		
B6	< 1366		
B7	< 1366		
B8	< 1366		
B9	< 1366		

Average Background counts per minute 98 CPMMDA (dpm/100cm²) 1366

*BASED ON AVERAGE OF 5 DIRECT MEASUREMENTS
(INITIAL DIRECT MEASUREMENT WAS GREATER THAN 60% OF GUIDELINE VALUE)

$$\frac{\text{dpm}}{100\text{cm}^2} = \frac{(\text{gross count rate}) - (\text{background count rate})}{(\text{efficiency}) \quad (\text{probe area})}$$

100cm²

SURVEYOR

SUPERVISOR

[Signature]
[Signature]

EXPOSURE RATE SURVEY RESULTS

DATE: 12-11-94LOCATION: Bausch + Lomb Batch Room NorthwellINSTRUMENT: Bicron B218LCAL DUE: 10-14-95

GRID NUMBER	EXPOSURE RATE $\mu\text{R/h}$		GRID NUMBER	EXPOSURE RATE $\mu\text{R/h}$	
A1	1cm	5 ^{div} $\mu\text{R/hr}$		1cm	1m
A2		5 $\mu\text{R/hr}$			
A3		5 $\mu\text{R/hr}$			
A4		5 $\mu\text{R/hr}$			
A5		5 $\mu\text{R/hr}$			
A6		5 $\mu\text{R/hr}$			
A7		5 $\mu\text{R/hr}$			
A8		6 $\mu\text{R/hr}$		N/A	
A9		6 $\mu\text{R/hr}$			
B1	N/A	6 $\mu\text{R/hr}$			
B2		5 $\mu\text{R/hr}$			
B3		5 $\mu\text{R/hr}$			
B4		5 $\mu\text{R/hr}$			
B5		5 $\mu\text{R/hr}$			
B6		5 $\mu\text{R/hr}$			
B7		5 $\mu\text{R/hr}$			
B8		5 $\mu\text{R/hr}$			
B9		6 $\mu\text{R/hr}$			

MEASUREMENTS TAKEN AT 1cm AND/OR 1m ABOVE THE FLOOR.

BKG: 5 $\mu\text{R/hr}$

SURVEYOR

SUPERVISOR



MDA CALCULATION SHEETMETER 2220 w/449 SERIAL #: 50061TS: 1 TB: 1 RB: 98 cpm EFF: 0.238 PROBE SIZE: 15 cm²

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{98}{1} + \frac{98}{1} \right)^{1/2}$$

$$0.238 (15 / 100 \text{ cm}^2)$$

$$MDA = \underline{1366 \text{ dpm} / 100 \text{ cm}^2}$$

METER 2220 w/449 SERIAL #: 50067TS: 1 TB: 1 RB: 0 cpm EFF: 0.0667 PROBE SIZE: 59 cm²

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{0}{1} + \frac{0}{1} \right)^{1/2}$$

$$0.0667 (59 / 100 \text{ cm}^2)$$

$$MDA = \underline{69 \text{ dpm} / 100 \text{ cm}^2}$$

METER 2220 w/449 SERIAL #: 50061TS: 1 TB: 1 RB: 44 cpm EFF: 0.238 PROBE SIZE: 15 cm²

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{44}{1} + \frac{44}{1} \right)^{1/2}$$

$$0.238 (15 / 100 \text{ cm}^2)$$

$$MDA = \underline{940 \text{ dpm} / 100 \text{ cm}^2}$$

Removable
only

METER _____ SERIAL #: _____

TS: _____ TB: _____ RB: _____ EFF: _____ PROBE SIZE: _____

$$MDA = \frac{2.71}{1} + 3.29 \left(\frac{\quad}{1} + \frac{\quad}{1} \right)^{1/2}$$

$$(\quad / 100 \text{ cm}^2)$$

$$MDA = \underline{\hspace{2cm}}$$

Technician: D. L. Vann Date: 12-11-94Reviewed by: B. D. D. D. Date: 12-11-94

Bausch & Lomb Release Survey Data

Survey Unit: Batch Room North Wall

Activity (dpm/100cm²)

<u>Grid No.</u>	Removable α			Removable β - γ		
	<u>Activity</u>	<u>Uncertainty *</u>	<u>MDA</u>	<u>Activity</u>	<u>Uncertainty *</u>	<u>MDA</u>
A- 1	69	16	69	940	64	940
A- 2	69	16	69	940	64	940
A- 3	69	16	69	940	64	940
A- 4	69	16	69	940	64	940
A- 5	69	16	69	940	64	940
A- 6	69	16	69	940	64	940
A- 7	69	16	69	940	64	940
A- 8	69	16	69	940	64	940
A- 9	69	16	69	940	64	940
B- 1	69	16	69	940	64	940
B- 2	69	16	69	940	64	940
B- 3	69	16	69	940	64	940
B- 4	69	16	69	940	64	940
B- 5	69	16	69	940	64	940
B- 6	69	16	69	940	64	940
B- 7	69	16	69	940	64	940
B- 8	69	16	69	940	64	940
B- 9	69	16	69	940	64	940

* Uncertainties represent the 95% Confidence level, based on counting statistics.
When survey data indicated less than MDA, the MDA value was used in the calculation.

Bausch & Lomb Release Survey Data

Survey Unit: Batch Room North Wall

Activity (dpm/100cm²)

<u>Grid No.</u>	Direct α			Direct β - γ		
	<u>Activity</u>	<u>Uncertainty *</u>	<u>MDA</u>	<u>Activity</u>	<u>Uncertainty *</u>	<u>MDA</u>
A- 1	69	16	69	1366	79	1366
A- 2	69	16	69	1366	79	1366
A- 3	69	16	69	1366	79	1366
A- 4	69	16	69	1366	79	1366
A- 5	69	16	69	1366	79	1366
A- 6	69	16	69	1366	79	1366
A- 7	69	16	69	1366	79	1366
A- 8	69	16	69	1366	79	1366
A- 9	69	16	69	1366	79	1366
B- 1	69	16	69	1366	79	1366
B- 2	69	16	69	1366	79	1366
B- 3	69	16	69	1366	79	1366
B- 4	69	16	69	1366	79	1366
B- 5	69	16	69	1366	79	1366
B- 6	69	16	69	1366	79	1366
B- 7	69	16	69	1366	79	1366
B- 8	69	16	69	1366	79	1366
B- 9	69	16	69	1366	79	1366

* Uncertainties represent the 95% Confidence level, based on counting statistics.
When survey data indicated less than MDA, the MDA value was used in the calculation.

NINETY-FIVE PERCENT CONFIDENCE CALCULATION SHEET

Survey Unit Batch Room N. Wall Date 12-11-94

Meter Ludlum 2220 Serial # 50067

Probe AC-3 Serial # 712582

MDA 69 (dpm/100cm²) Survey Type α removable

Guideline Value 200 (dpm/100cm²)

Average Measurement Level

$$x_{ave} = \frac{1}{n_s} \sum_{i=1}^{n_s} x_i$$

where:

x_{ave} = calculated mean for a survey unit,
 n_s = number of measurements within a survey unit,
 x_i = systematic and random measurements at point (i)

Σx_i 1242

n_s 18

x_{ave} = 69

Standard Deviation

$$S_x = \sqrt{\frac{\sum_{i=1}^{n_s} (x_{ave} - x_i)^2}{n_s - 1}}$$

where:

S_x = standard deviation
 n_s = number of measurements within a survey unit
 x_i = systematic and random measurements at point (i)
 x_{ave} = calculated mean for a survey unit

x_{ave} = 69

n_s = 18

$\Sigma (x_{ave} - x_i)^2$ = 0

s_x = 0