

5A
VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: AMRO Environmental Laboratories Contract: _____
 Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039
 Lab File ID: C:\HPCHEM\1\DATA\08 BFB Injection Date: 08/06/05
 Instrument ID: V-3 BFB Injection Time: 5:51
 GC Column: HP-624 ID: 0.20 (mm) Heated Purge: (Y/N) N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	17.7
75	30.0 - 60.0% of mass 95	40.4
95	Base peak, 100% relative abundance	100.0
96	5.0 - 9.0% of mass 95	6.9
173	Less than 2.0% of mass 174	0.0 (0.0)1
174	50.0 - 120.0% of mass 95	90.1
175	5.0 - 9.0% of mass 174	6.3 (7.0)1
176	95.0 - 101.0% of mass 174	87.8 (97.4)1
177	5.0 - 9.0% of mass 176	6.4 (7.2)2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

	EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
01	8260 ccv 25ppb	8260 ccv 25ppb	EM1\DATA\080605\G	08/06/05	6:29
02	lcsf-08/06/05	lcsf-08/06/05	EM1\DATA\080605\G	08/06/05	7:14
03	mb-08/06/05	mb-08/06/05	EM1\DATA\080605\G	08/06/05	8:21
04	ZZZZZ	0508028-05A	EM1\DATA\080605\G	08/06/05	8:55
05	ZZZZZ	0508042-04A	EM1\DATA\080605\G	08/06/05	9:31
06	ZZZZZ	0508040-04A	EM1\DATA\080605\G	08/06/05	10:05
07	ZZZZZ	0508042-01A	EM1\DATA\080605\G	08/06/05	10:43
08	ZZZZZ	0508042-02A	EM1\DATA\080605\G	08/06/05	11:18
09	ZZZZZ	0508042-03A	EM1\DATA\080605\G	08/06/05	11:53
10	ZZZZZ	0508040-02A	EM1\DATA\080605\G	08/06/05	12:27
11	ZZZZZ	0508040-03A	EM1\DATA\080605\G	08/06/05	13:01
12	ZZZZZ	0508040-01A	EM1\DATA\080605\G	08/06/05	13:35
13	0508040-04Amsf	0508040-04Amsf	EM1\DATA\080605\G	08/06/05	14:10
14	0508040-04Amsdf	0508040-04Amsdf	EM1\DATA\080605\G	08/06/05	14:45
15	MW-9D	0508039-01A	EM1\DATA\080605\G	08/06/05	15:54
16	MW-5A	0508039-02A	EM1\DATA\080605\G	08/06/05	16:30
17	MW-8B	0508039-03A	EM1\DATA\080605\G	08/06/05	17:05
18	MW-8C	0508039-04A	EM1\DATA\080605\G	08/06/05	17:39

5A
VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: AMRO Environmental Laboratories Contract: _____
 Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039
 Lab File ID: C:\HPCHEM\1\DATA\08 BFB Injection Date: 08/09/05
 Instrument ID: V-3 BFB Injection Time: 8:38
 GC Column: HP-624 ID: 0.20 (mm) Heated Purge: (Y/N) N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	15.2
75	30.0 - 60.0% of mass 95	44.0
95	Base peak, 100% relative abundance	100.0
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.0 (0.0)1
174	50.0 - 120.0% of mass 95	94.7
175	5.0 - 9.0% of mass 174	5.3 (5.6)1
176	95.0 - 101.0% of mass 174	91.8 (96.9)1
177	5.0 - 9.0% of mass 176	5.9 (6.5)2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

	EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
01	8260 ccv 25ppb	8260 ccv 25ppb	EM1\DATA\080905\G	08/09/05	9:45
02	lcsf-08/09/05	lcsf-08/09/05	EM1\DATA\080905\G	08/09/05	10:38
03	mb-08/09/05	mb-08/09/05	EM1\DATA\080905\G	08/09/05	12:32
04	MW-8A	0508039-05A	EM1\DATA\080905\G	08/09/05	13:07
05	MW-8A	0508039-05A	EM1\DATA\080905\G	08/09/05	13:41
06	ZZZZZ	0508040-01A	EM1\DATA\080905\G	08/09/05	14:15
07	MW-5D	0508039-06A	EM1\DATA\080905\G	08/09/05	15:23
08	MW-5B	0508039-08A	EM1\DATA\080905\G	08/09/05	16:31
09	MW-9A	0508039-11A	EM1\DATA\080905\G	08/09/05	18:13
10	MW-8D	0508039-14A	EM1\DATA\080905\G	08/09/05	18:47
11	MW-8D	0508039-14Amsf	EM1\DATA\080905\G	08/09/05	19:21
12	MW-8D	0508039-14Amsdf	EM1\DATA\080905\G	08/09/05	19:55

5A
VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: AMRO Environmental Laboratories Contract: _____
 Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039
 Lab File ID: C:\HPCHEM\1\DATA\08 BFB Injection Date: 08/10/05
 Instrument ID: V-3 BFB Injection Time: 6:46
 GC Column: HP-624 ID: 0.20 (mm) Heated Purge: (Y/N) N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	16.5
75	30.0 - 60.0% of mass 95	40.6
95	Base peak, 100% relative abundance	100.0
96	5.0 - 9.0% of mass 95	8.2
173	Less than 2.0% of mass 174	0.0 (0.0)1
174	50.0 - 120.0% of mass 95	95.4
175	5.0 - 9.0% of mass 174	7.3 (7.7)1
176	95.0 - 101.0% of mass 174	91.6 (96.1)1
177	5.0 - 9.0% of mass 176	7.1 (7.7)2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

	EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
01	8260 ccv 25ppb	8260 ccv 25ppb	EM1\DATA\081005\G	08/10/05	7:38
02	lcsf-08/10/05	lcsf-08/10/05	EM1\DATA\081005\G	08/10/05	8:12
03	mb-08/10/05	mb-08/10/05	EM1\DATA\081005\G	08/10/05	9:55
04	MW-5C	0508039-07A	EM1\DATA\081005\G	08/10/05	10:32
05	MW-7B	0508039-09A	EM1\DATA\081005\G	08/10/05	11:06
06	MW-7C	0508039-10A	EM1\DATA\081005\G	08/10/05	11:40
07	MW-5B	0508039-08A	EM1\DATA\081005\G	08/10/05	12:14
08	MW-5C	0508039-07Amsf	EM1\DATA\081005\G	08/10/05	12:49
09	MW-5C	0508039-07Amsdf	EM1\DATA\081005\G	08/10/05	13:24
10	MW-9B	0508039-12A	EM1\DATA\081005\G	08/10/05	14:33
11	MW-9C	0508039-13A	EM1\DATA\081005\G	08/10/05	15:08
12	MW-2B	0508039-15A	EM1\DATA\081005\G	08/10/05	15:43
13	MW-1	0508039-17A	EM1\DATA\081005\G	08/10/05	16:53
14	MW-11C	0508039-18A	EM1\DATA\081005\G	08/10/05	17:28
15	MW-6A	0508039-19A	EM1\DATA\081005\G	08/10/05	18:03
16	MW-6B	0508039-20A	EM1\DATA\081005\G	08/10/05	18:37
17	ZZZZZ	0508063-01A	EM1\DATA\081005\G	08/10/05	20:54
18	ZZZZZ	0508063-02A	EM1\DATA\081005\G	08/10/05	21:28
19	ZZZZZ	0508063-03A	EM1\DATA\081005\G	08/10/05	22:02
20	ZZZZZ	0508063-04A	EM1\DATA\081005\G	08/10/05	22:37

5A

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: AMRO Environmental Laboratories Contract: _____
Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039
Lab File ID: C:\HPCHEM\1\DATA\08 BFB Injection Date: 08/10/05
Instrument ID: V-3 BFB Injection Time: 6:46
GC Column: HP-624 ID: 0.20 (mm) Heated Purge: (Y/N) N

5A

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: AMRO Environmental Laboratories Contract: _____
 Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039
 Lab File ID: C:\HPCHEM\1\DATA\08 BFB Injection Date: 08/11/05
 Instrument ID: V-3 BFB Injection Time: 6:00
 GC Column: HP-624 ID: 0.20 (mm) Heated Purge: (Y/N) N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	17.9
75	30.0 - 60.0% of mass 95	39.9
95	Base peak, 100% relative abundance	100.0
96	5.0 - 9.0% of mass 95	6.4
173	Less than 2.0% of mass 174	0.0 (0.0)1
174	50.0 - 120.0% of mass 95	88.7
175	5.0 - 9.0% of mass 174	6.6 (7.5)1
176	95.0 - 101.0% of mass 174	85.7 (96.6)1
177	5.0 - 9.0% of mass 176	5.8 (6.8)2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

	EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
01	8260 ccv 25ppb	8260 ccv 25ppb	EM1\DATA\081105\G	08/11/05	7:13
02	lcsf-08/11/05	lcsf-08/11/05	EM1\DATA\081105\G	08/11/05	7:48
03	mb-08/11/05	mb-08/11/05	EM1\DATA\081105\G	08/11/05	8:57
04	MW-2A	0508039-16A	EM1\DATA\081105\G	08/11/05	9:31
05	MW-6C	0508039-21A	EM1\DATA\081105\G	08/11/05	10:05
06	MW-11D	0508039-23A	EM1\DATA\081105\G	08/11/05	11:14
07	MW-3	0508039-24A	EM1\DATA\081105\G	08/11/05	11:48
08	MW-12C	0508039-25A	EM1\DATA\081105\G	08/11/05	12:23
09	MW-7A	0508039-27A	EM1\DATA\081105\G	08/11/05	13:32
10	MW-2A	0508039-16Amsf	EM1\DATA\081105\G	08/11/05	15:16
11	MW-2A	0508039-16Amsdf	EM1\DATA\081105\G	08/11/05	15:50
12	MW-4	0508039-22A	EM1\DATA\081105\G	08/11/05	17:35
13	ZZZZZ	0508081-04A	EM1\DATA\081105\G	08/11/05	22:07
14	ZZZZZ	0508081-02A	EM1\DATA\081105\G	08/11/05	22:41
15	ZZZZZ	0508081-03A	EM1\DATA\081105\G	08/11/05	23:16
16	ZZZZZ	0508081-01A	EM1\DATA\081105\G	08/12/05	0:25

5A
VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: AMRO Environmental Laboratories Contract: _____
 Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039
 Lab File ID: C:\HPCHEM\1\DATA\08 BFB Injection Date: 08/12/05
 Instrument ID: V-3 BFB Injection Time: 6:08
 GC Column: HP-624 ID: 0.20 (mm) Heated Purge: (Y/N) N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	16.7
75	30.0 - 60.0% of mass 95	42.7
95	Base peak, 100% relative abundance	100.0
96	5.0 - 9.0% of mass 95	7.0
173	Less than 2.0% of mass 174	0.0 (0.0)1
174	50.0 - 120.0% of mass 95	89.6
175	5.0 - 9.0% of mass 174	5.7 (6.4)1
176	95.0 - 101.0% of mass 174	87.0 (97.1)1
177	5.0 - 9.0% of mass 176	6.2 (7.1)2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

	EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
01	8260 ccv 25ppb	8260 ccv 25ppb	EM1\DATA\081205\G	08/12/05	7:12
02	lcsf-08/12/05	lcsf-08/12/05	EM1\DATA\081205\G	08/12/05	7:49
03	mb-08/12/05	mb-08/12/05	EM1\DATA\081205\G	08/12/05	8:58
04	MW-6D	0508039-28A	EM1\DATA\081205\G	08/12/05	9:32
05	MW-12D	0508039-26A	EM1\DATA\081205\G	08/12/05	10:06
06	ZZZZZ	0508081-01A	EM1\DATA\081205\G	08/12/05	10:41
07	MW-2B	0508039-15A	EM1\DATA\081205\G	08/12/05	11:15
08	MW-12C	0508039-25A	EM1\DATA\081205\G	08/12/05	12:16
09	MW-6B	0508039-20A	EM1\DATA\081205\G	08/12/05	12:50
10	MW-7A	0508039-27A	EM1\DATA\081205\G	08/12/05	13:24
11	ZZZZZ	0508081-02A	EM1\DATA\081205\G	08/12/05	13:58
12	MW-6D	0508039-28Amsf	EM1\DATA\081205\G	08/12/05	15:11
13	MW-6D	0508039-28Amsdf	EM1\DATA\081205\G	08/12/05	16:19
14	ZZZZZ	0508058-03A	EM1\DATA\081205\G	08/12/05	17:28
15	ZZZZZ	0508058-04A	EM1\DATA\081205\G	08/12/05	18:02

LAB FILE ID:	0.5 PPB	<u>G1581.D</u>	1.0 PPB	<u>G1583.D</u>	2.0 PPB	<u>G1584.D</u>
5.0 PPB	10 PPB =	<u>G1586.D</u>	25 PPB =	<u>G1588.D</u>	50 PPB =	<u>G1589.D</u>
100 PPB	300 PPB	<u>G1595.D</u>	200 PPB	<u>G1593.D</u>		

COMPOUND	0.5 PPB	1.0 PPB	2.0 PPB	5.0 PPB	10 PPB	25 PPB	50 PPB	100 PPB	300 PPB	200 PPB	RRF	%
												RSD
1,4-Dioxane					0.003	0.003	0.003	0.002	0.002	0.003	0.003	19.365
Acrylonitrile	0.152	0.151	0.135	0.131	0.138	0.116	0.142	0.136	0.131	0.133	0.137	7.639
Cyclohexanone				0.001	0.001			0.002	0.001		0.001	40.000
Dichlorofluoromethane		0.513	0.439	0.428	0.437	0.378	0.350	0.416	0.389	0.408	0.418	11.070
Diisopropyl ether	1.395	1.238	1.031	1.047	1.035	0.956	0.892	0.983	0.961	1.015	1.055	14.201
Ethyl Tertiary Butyl Ether	0.829	0.688	0.661	0.654	0.664	0.580	0.544	0.573	0.527	0.566	0.629	14.375
Freon-113	0.119	0.106	0.146	0.146	0.151	0.134	0.123	0.151	0.135	0.149	0.136	11.449
Iodomethane	0.576	0.540	0.497	0.501	0.513	0.464	0.432	0.515	0.482	0.518	0.504	7.885
Methyl methacrylate	0.289	0.304	0.264	0.273	0.318	0.323	0.260	0.251	0.247	0.245	0.277	10.573
Tertiary Amyl Methyl Ether	0.910	0.808	0.720	0.719	0.747	0.657	0.625	0.642	0.578	0.616	0.702	14.356
Tertiary Butanol	0.029	0.030	0.028	0.030	0.032	0.029	0.029	0.025	0.020	0.025	0.028	12.512
Vinyl acetate				0.012	0.012	0.003	0.023	0.025	0.026	0.025	0.018	49.897
Dichlorodifluoromethane		0.074	0.168	0.181	0.179	0.143	0.138	0.161	0.139	0.158	0.149	21.737
Chloromethane	*	0.468	0.372	0.346	0.342	0.271	0.267	0.312	0.297	0.316	0.332	18.493
Vinyl chloride	*	0.327	0.277	0.290	0.291	0.239	0.228	0.277	0.263	0.276	0.284	14.900
Chloroethane		0.205	0.201	0.206	0.209	0.171	0.159	0.190	0.162	0.177	0.187	10.610
Bromomethane		0.265	0.230	0.294	0.252	0.204	0.181	0.219	0.170	0.186	0.222	18.822
Trichlorofluoromethane	0.216	0.251	0.276	0.273	0.276	0.229	0.215	0.256	0.229	0.255	0.248	9.660
Diethyl ether	0.196	0.199	0.188	0.188	0.192	0.182	0.168	0.186	0.182	0.187	0.187	4.597
Acetone			0.065	0.064	0.060	0.052	0.055	0.052	0.046	0.047	0.055	13.169

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: AMRO Environmental Laboratories Contract: _____

Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039

Instrument ID: V-3 Calibration Date(s): 07/26/05 07/26/05

Heated Purge: (Y/N) N Calibration Times: 13:05 21:11

GC Column: HP-624 ID: 0.20 (mm)

LAB FILE ID:		0.5 PPB		1.0 PPB		2.0 PPB		5.0 PPB		10 PPB		25 PPB		50 PPB		100 PPB		300 PPB		200 PPB		RRF		% RSD	
COMPOUND		0.5 PPB	1.0 PPB	2.0 PPB	5.0 PPB	10 PPB	25 PPB	50 PPB	100 PPB	300 PPB	200 PPB	RRF	% RSD												
1,1-Dichloroethene	*	0.326	0.310	0.232	0.241	0.238	0.203	0.190	0.231	0.221	0.239	0.243	17.656												
Carbon disulfide		1.076	0.943	0.846	0.835	0.852	0.722	0.694	0.858	0.832	0.884	0.854	12.466												
Methylene chloride			0.475	0.422	0.369	0.369	0.341	0.314	0.359	0.363	0.373	0.376	12.453												
Methyl tert-butyl ether		0.549	0.527	0.475	0.447	0.461	0.424	0.409	0.424	0.397	0.423	0.454	11.127												
trans-1,2-Dichloroethene		0.368	0.296	0.276	0.275	0.281	0.252	0.233	0.272	0.266	0.276	0.280	12.682												
1,1-Dichloroethane	*	0.648	0.565	0.549	0.524	0.505	0.467	0.422	0.491	0.487	0.505	0.516	11.891												
2-Butanone			0.040	0.053	0.055	0.064	0.045	0.050	0.047	0.042	0.042	0.049	15.850												
2,2-Dichloropropane		0.274	0.338	0.254	0.258	0.229	0.188	0.164	0.189	0.174	0.187	0.226	24.533												
cis-1,2-Dichloroethene		0.367	0.380	0.312	0.303	0.304	0.278	0.254	0.288	0.292	0.299	0.308	12.481												
Chloroform	*	0.516	0.452	0.470	0.435	0.472	0.406	0.384	0.442	0.417	0.434	0.443	8.467												
Tetrahydrofuran			0.178	0.153	0.134	0.109	0.100	0.095	0.090	0.086	0.088	0.115	28.615												
Bromochloromethane		0.216	0.191	0.148	0.177	0.170	0.164	0.155	0.170	0.166	0.175	0.173	11.038												
1,1,1-Trichloroethane		0.289	0.270	0.247	0.258	0.272	0.244	0.232	0.275	0.252	0.263	0.260	6.527												
1,1-Dichloropropene		0.489	0.392	0.365	0.378	0.361	0.331	0.311	0.362	0.330	0.350	0.367	13.409												
Carbon tetrachloride		0.257	0.273	0.218	0.237	0.237	0.210	0.197	0.232	0.207	0.221	0.229	10.205												
1,2-Dichloroethane		0.387	0.329	0.351	0.328	0.352	0.314	0.306	0.326	0.317	0.331	0.334	7.075												
Benzene		1.422	1.287	1.205	1.188	1.226	1.080	1.008	1.151	1.104	1.156	1.183	9.744												
Trichloroethene		0.351	0.303	0.291	0.278	0.293	0.258	0.249	0.281	0.275	0.283	0.286	9.716												
1,2-Dichloropropane	*	0.395	0.331	0.332	0.320	0.338	0.308	0.293	0.323	0.323	0.332	0.330	8.059												
Bromodichloromethane		0.471	0.369	0.320	0.321	0.353	0.319	0.319	0.354	0.359	0.365	0.355	12.856												

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: AMRO Environmental Laboratories Contract: _____Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039Instrument ID: V-3 Calibration Date(s): 07/26/05 07/26/05Heated Purge: (Y/N) N Calibration Times: 13:05 21:11GC Column: HP-624 ID: 0.20 (mm)

LAB FILE ID:	0.5 PPB	1.0 PPB	2.0 PPB	5.0 PPB	10 PPB	25 PPB	50 PPB	100 PPB	300 PPB	200 PPB	RRF	% RSD
<u>5.0 PPB G1585.D</u>	<u>10 PPB = G1586.D</u>	<u>25 PPB = G1588.D</u>	<u>50 PPB = G1589.D</u>									
<u>100 PPB G1591.D</u>	<u>300 PPB G1595.D</u>											
COMPOUND	0.5 PPB	1.0 PPB	2.0 PPB	5.0 PPB	10 PPB	25 PPB	50 PPB	100 PPB	300 PPB	200 PPB	RRF	% RSD
Dibromomethane	0.240	0.228	0.207	0.219	0.238	0.212	0.211	0.224	0.217	0.220	0.222	4.994
4-Methyl-2-pentanone		0.049	0.035	0.051	0.057	0.052	0.055	0.052	0.050	0.050	0.050	12.400
cis-1,3-Dichloropropene	0.578	0.530	0.435	0.484	0.495	0.463	0.457	0.494	0.502	0.506	0.494	8.110
Toluene	* 0.731	0.711	0.646	0.659	0.708	0.649	0.593	0.676	0.673	0.696	0.674	5.945
trans-1,3-Dichloropropene	0.496	0.452	0.365	0.391	0.419	0.393	0.399	0.425	0.428	0.437	0.421	8.760
1,1,2-Trichloroethane	0.371	0.359	0.282	0.301	0.306	0.280	0.281	0.289	0.281	0.285	0.304	11.108
1,2-Dibromoethane	0.351	0.370	0.321	0.315	0.344	0.316	0.316	0.325	0.320	0.327	0.331	5.585
2-Hexanone	0.299	0.286	0.224	0.257	0.249	0.213	0.236	0.211	0.205	0.203	0.238	14.247
1,3-Dichloropropane	0.991	0.802	0.743	0.693	0.715	0.694	0.667	0.679	0.683	0.694	0.736	13.285
Tetrachloroethene	0.384	0.326	0.330	0.311	0.316	0.295	0.275	0.309	0.305	0.318	0.317	8.946
Dibromochloromethane	0.484	0.411	0.407	0.416	0.431	0.412	0.422	0.435	0.452	0.450	0.432	5.596
Chlorobenzene	* 1.345	1.126	1.022	1.044	1.070	0.992	0.961	1.027	1.019	1.046	1.065	10.116
1,1,1,2-Tetrachloroethane	0.347	0.349	0.328	0.323	0.342	0.310	0.310	0.329	0.335	0.338	0.331	4.181
Ethylbenzene	* 1.862	1.642	1.460	1.516	1.584	1.449	1.408	1.562	1.564	1.610	1.566	8.192
m,p-Xylene	0.726	0.631	0.589	0.583	0.629	0.578	0.554	0.608		0.619	0.613	8.073
o-Xylene	0.716	0.631	0.585	0.550	0.576	0.553	0.544	0.580	0.583	0.594	0.591	8.553
Styrene	1.229	1.146	1.037	1.111	1.095	0.989	1.042	1.115	1.124	1.144	1.103	6.132
Bromoforn	* 0.263	0.290	0.254	0.294	0.299	0.274	0.302	0.301	0.305	0.309	0.289	6.544
Isopropylbenzene	2.987	2.775	2.398	2.420	2.492	2.134	2.108	2.458	2.598	2.561	2.493	10.628
1,1,2,2-Tetrachloroethane	* 1.241	1.205	1.071	1.246	1.267	1.141	1.116	1.094	1.115	1.094	1.159	6.326

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: AMRO Environmental Laboratories Contract: _____

Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039

Instrument ID: V-3 Calibration Date(s): 07/26/05 07/26/05

Heated Purge: (Y/N) N Calibration Times: 13:05 21:11

GC Column: HP-624 ID: 0.20 (mm)

LAB FILE ID:	0.5 PPB	1.0 PPB	2.0 PPB	5.0 PPB	10 PPB	25 PPB	50 PPB	100 PPB	300 PPB	200 PPB	RRF	% RSD
5.0 PPB <u>G1585.D</u>	0.920	0.906	0.864	0.918	0.916	0.824	0.827	0.782	0.775	0.771	0.850	7.324
100 PPB <u>G1591.D</u>	0.895	1.018	0.936	0.926	0.978	0.886	0.858	0.901	0.924	0.944	0.927	5.011
	3.689	3.591	3.063	3.187	3.126	2.525	2.711	3.131	3.237	3.341	3.160	11.169
	2.392	2.168	2.003	2.038	2.003	1.833	1.694	1.911	1.965	1.975	1.998	9.357
	2.849	2.670	2.285	2.404	2.325	2.124	2.015	2.221	2.277	2.321	2.349	10.478
	2.492	2.179	2.025	2.084	2.139	1.995	1.820	2.103	2.147	2.156	2.114	8.042
	2.198	1.760	1.621	1.590	1.608	1.557	1.387	1.626	1.804	1.811	1.696	12.816
	2.588	2.330	2.040	2.179	2.195	2.050	1.927	2.154	2.245	2.275	2.198	8.309
	3.175	2.689	2.432	2.441	2.534	2.158	2.177	2.559	2.795	2.767	2.573	11.810
	2.794	2.327	2.063	2.134	2.184	2.095	1.913	2.163	2.301	2.307	2.228	10.588
	1.925	1.777	1.574	1.636	1.682	1.541	1.457	1.581	1.580	1.603	1.636	8.093
	2.002	1.855	1.704	1.742	1.752	1.592	1.497	1.597	1.574	1.607	1.692	8.965
	2.781	2.081	1.913	1.924	1.998	1.775	1.752	1.993	2.136	2.139	2.049	14.140
	1.799	1.579	1.480	1.582	1.581	1.475	1.422	1.489	1.507	1.513	1.543	6.769
	0.190	0.115	0.092	0.142	0.141	0.129	0.146	0.137	0.135	0.133	0.136	18.200
	1.124	1.218	0.981	1.008	1.026	0.965	0.887	0.935	0.975	0.974	1.009	9.495
	0.421	0.368	0.352	0.366	0.362	0.335	0.303	0.334	0.374	0.372	0.359	8.699
		3.530	2.807	3.077	3.050	2.883	2.896	2.747	2.885	2.849	2.969	7.922
	1.380	1.030	0.965	0.987	0.998	0.945	0.909	0.927	0.966	0.952	1.006	13.519
	0.267	0.275	0.281	0.281	0.285	0.275	0.274	0.288	0.263	0.269	0.276	2.910

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: AMRO Environmental Laboratories Contract: _____

Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039

Instrument ID: V-3 Calibration Date(s): 07/26/05 07/26/05

Heated Purge: (Y/N) N Calibration Times: 13:05 21:11

GC Column: HP-624 ID: 0.20 (mm)

LAB FILE ID:	0.5 PPB	<u>G1581.D</u>	1.0 PPB	<u>G1583.D</u>	2.0 PPB	<u>G1584.D</u>						
5.0 PPB	<u>G1585.D</u>	10 PPB =	<u>G1586.D</u>	25 PPB =	<u>G1588.D</u>	50 PPB = <u>G1589.D</u>						
100 PPB	<u>G1591.D</u>	300 PPB	<u>G1595.D</u>	200 PPB	<u>G1593.D</u>							
COMPOUND	0.5 PPB	1.0 PPB	2.0 PPB	5.0 PPB	10 PPB	25 PPB	50 PPB	100 PPB	300 PPB	200 PPB	RRF	% RSD
1,2-Dichloroethane-d4	0.244	0.253	0.260	0.259	0.257	0.255	0.253	0.255	0.245	0.256	0.254	2.111
Toluene-d8	0.872	0.874	0.891	0.889	0.901	0.882	0.908	0.903	0.892	0.899	0.891	1.366
4-Bromofluorobenzene	0.395	0.400	0.405	0.401	0.405	0.412	0.423	0.417	0.413	0.414	0.409	2.129

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: AMRO Environmental Laboratories Contract: _____Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039Instrument ID: V-3 Calibration Date: 08/06/05 Time: 6:29Lab File ID: C:\HPCHEM\1\DATA\080605 Init. Calib. Date(s): 07/26/05 07/26/05EPA Sample No.: 8260 ccv 25ppb Init. Calib. Times: 13:05 21:11Heated Purge: (Y/N) NGC Column: HP-624 ID: 0.20 (mm)

COMPOUND	IC RRF	CCV RRF	MIN RRF	%D	MAX %D
Dichlorodifluoromethane	0.149	0.181		21.477	
Chloromethane	0.332	0.355	0.100	6.82	
Vinyl chloride	0.284	0.305		7.243	20.0
Chloroethane	0.187	0.209		11.964	
Bromomethane	0.222	0.245		10.195	
Trichlorofluoromethane	0.248	0.271		9.451	
Diethyl ether	0.187	0.193		3.319	
Acetone	0.055	0.078		41.497	
1,1-Dichloroethene	0.243	0.259		6.541	20.0
Carbon disulfide	0.854	0.929		8.757	
Methylene chloride	0.376	0.383		1.832	
Methyl tert-butyl ether	0.454	0.437		-3.66	
trans-1,2-Dichloroethene	0.279	0.280		0.179	
1,1-Dichloroethane	0.516	0.495	0.100	-4.126	
2-Butanone	0.049	0.067		37.671	
2,2-Dichloropropane	0.225	0.231		2.439	
cis-1,2-Dichloroethene	0.308	0.297		-3.477	
Chloroform	0.443	0.473		6.82	20.0
Tetrahydrofuran	0.115	0.087		-24.201	
Bromochloromethane	0.173	0.182		5.081	
1,1,1-Trichloroethane	0.260	0.280		7.61	
1,1-Dichloropropene	0.367	0.347		-5.424	
Carbon tetrachloride	0.229	0.243		6.16	
1,2-Dichloroethane	0.334	0.340		1.766	
Benzene	1.183	1.129		-4.54	
Trichloroethene	0.286	0.277		-3.215	
1,2-Dichloropropane	0.330	0.316		-4.097	20.0
Bromodichloromethane	0.355	0.351		-1.127	
Dibromomethane	0.222	0.214		-3.43	
4-Methyl-2-pentanone	0.050	0.052		3.769	
cis-1,3-Dichloropropene	0.494	0.467		-5.542	
Toluene	0.674	0.667		-1.068	20.0
trans-1,3-Dichloropropene	0.420	0.403		-4.162	
1,1,2-Trichloroethane	0.304	0.288		-5.107	
1,2-Dibromoethane	0.330	0.315		-4.69	

All other compounds must meet a minimum RRF of 0.010.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: AMRO Environmental Laboratories Contract: _____Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039Instrument ID: V-3 Calibration Date: 08/06/05 Time: 6:29Lab File ID: C:\HPCHEM\1\DATA\080605 Init. Calib. Date(s): 07/26/05 07/26/05EPA Sample No.: 8260 ccv 25ppb Init. Calib. Times: 13:05 21:11Heated Purge: (Y/N) NGC Column: HP-624 ID: 0.20 (mm)

COMPOUND	IC RRF	CCV RRF	MIN RRF	%D	MAX %D
2-Hexanone	0.238	0.282		18.338	
1,3-Dichloropropane	0.736	0.663		-9.931	
Tetrachloroethene	0.317	0.310		-2.177	
Dibromochloromethane	0.432	0.425		-1.62	
Chlorobenzene	1.065	1.032	0.300	-3.117	
1,1,1,2-Tetrachloroethane	0.331	0.328		-0.936	
Ethylbenzene	1.566	1.478		-5.601	20.0
m,p-Xylene	0.613	0.584		-4.731	
o-Xylene	0.591	0.555		-6.123	
Styrene	1.103	1.072		-2.828	
Bromoform	0.289	0.276	0.100	-4.531	
Isopropylbenzene	2.493	2.352		-5.66	
1,1,2,2-Tetrachloroethane	1.159	1.036	0.300	-10.613	
1,2,3-Trichloropropane	0.850	0.758		-10.855	
Bromobenzene	0.927	0.876		-5.461	
n-Propylbenzene	3.160	3.099		-1.933	
2-Chlorotoluene	1.998	1.854		-7.216	
4-Chlorotoluene	2.349	2.183		-7.071	
1,3,5-Trimethylbenzene	2.114	2.049		-3.075	
tert-Butylbenzene	1.696	1.582		-6.733	
1,2,4-Trimethylbenzene	2.198	2.020		-8.111	
sec-Butylbenzene	2.573	2.432		-5.469	
4-Isopropyltoluene	2.228	2.101		-5.704	
1,3-Dichlorobenzene	1.636	1.538		-5.967	
1,4-Dichlorobenzene	1.692	1.621		-4.208	
n-Butylbenzene	2.049	1.884		-8.062	
1,2-Dichlorobenzene	1.543	1.459		-5.426	
1,2-Dibromo-3-chloropropane	0.136	0.122		-10.294	
1,2,4-Trichlorobenzene	1.009	0.890		-11.82	
Hexachlorobutadiene	0.359	0.353		-1.589	
Naphthalene	2.969	2.405		-19.005	
1,2,3-Trichlorobenzene	1.006	0.880		-12.516	

All other compounds must meet a minimum RRF of 0.010.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: AMRO Environmental Laboratories Contract: _____Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039Instrument ID: V-3 Calibration Date: 08/09/05 Time: 9:45Lab File ID: C:\HPCHEM\1\DATA\080905 Init. Calib. Date(s): 07/26/05 07/26/05EPA Sample No.: 8260 ccv 25ppb Init. Calib. Times: 13:05 21:11Heated Purge: (Y/N) NGC Column: HP-624 ID: 0.20 (mm)

COMPOUND	IC RRF	CCV RRF	MIN RRF	%D %D	MAX %D
Dichlorodifluoromethane	0.149	0.189		26.846	
Chloromethane	0.332	0.374	0.100	12.538	
Vinyl chloride	0.284	0.312		9.705	20.0
Chloroethane	0.187	0.214		14.643	
Bromomethane	0.222	0.246		10.645	
Trichlorofluoromethane	0.248	0.277		11.874	
Diethyl ether	0.187	0.206		10.278	
Acetone	0.055	0.074		34.24	
1,1-Dichloroethene	0.243	0.268		10.243	20.0
Carbon disulfide	0.854	0.955		11.801	
Methylene chloride	0.376	0.404		7.415	
Methyl tert-butyl ether	0.454	0.491		8.245	
trans-1,2-Dichloroethene	0.279	0.295		5.546	
1,1-Dichloroethane	0.516	0.527	0.100	2.072	
2-Butanone	0.049	0.065		33.562	
2,2-Dichloropropane	0.225	0.248		9.978	
cis-1,2-Dichloroethene	0.308	0.318		3.347	
Chloroform	0.443	0.472		6.594	20.0
Tetrahydrofuran	0.115	0.091		-20.716	
Bromochloromethane	0.173	0.193		11.432	
1,1,1-Trichloroethane	0.260	0.274		5.304	
1,1-Dichloropropene	0.367	0.350		-4.606	
Carbon tetrachloride	0.229	0.234		2.228	
1,2-Dichloroethane	0.334	0.358		7.154	
Benzene	1.183	1.187		0.364	
Trichloroethene	0.286	0.289		0.978	
1,2-Dichloropropane	0.330	0.339		2.883	20.0
Bromodichloromethane	0.355	0.383		7.887	
Dibromomethane	0.222	0.232		4.693	
4-Methyl-2-pentanone	0.050	0.051		1.774	
cis-1,3-Dichloropropene	0.494	0.520		5.178	
Toluene	0.674	0.694		2.937	20.0
trans-1,3-Dichloropropene	0.420	0.450		7.015	
1,1,2-Trichloroethane	0.304	0.312		2.801	
1,2-Dibromoethane	0.330	0.340		2.874	

All other compounds must meet a minimum RRF of 0.010.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: AMRO Environmental Laboratories Contract: _____Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039Instrument ID: V-3 Calibration Date: 08/09/05 Time: 9:45Lab File ID: C:\HPCHEM\1\DATA\080905 Init. Calib. Date(s): 07/26/05 07/26/05EPA Sample No.: 8260 ccv 25ppb Init. Calib. Times: 13:05 21:11Heated Purge: (Y/N) NGC Column: HP-624 ID: 0.20 (mm)

COMPOUND	IC RRF	CCV RRF	MIN RRF	%D	MAX %D
2-Hexanone	0.238	0.275		15.401	
1,3-Dichloropropane	0.736	0.702		-4.633	
Tetrachloroethene	0.317	0.322		1.609	
Dibromochloromethane	0.432	0.471		9.028	
Chlorobenzene	1.065	1.082	0.300	1.577	
1,1,1,2-Tetrachloroethane	0.331	0.351		6.01	
Ethylbenzene	1.566	1.574		0.53	20.0
m,p-Xylene	0.613	0.628		2.447	
o-Xylene	0.591	0.603		1.996	
Styrene	1.103	1.146		3.88	
Bromoform	0.289	0.305	0.100	5.5	
Isopropylbenzene	2.493	2.458		-1.408	
1,1,2,2-Tetrachloroethane	1.159	1.129	0.300	-2.588	
1,2,3-Trichloropropane	0.850	0.838		-1.447	
Bromobenzene	0.927	0.968		4.468	
n-Propylbenzene	3.160	3.217		1.801	
2-Chlorotoluene	1.998	1.965		-1.661	
4-Chlorotoluene	2.349	2.350		0.038	
1,3,5-Trimethylbenzene	2.114	2.184		3.311	
tert-Butylbenzene	1.696	1.685		-0.66	
1,2,4-Trimethylbenzene	2.198	2.245		2.124	
sec-Butylbenzene	2.573	2.620		1.839	
4-Isopropyltoluene	2.228	2.305		3.451	
1,3-Dichlorobenzene	1.636	1.682		2.837	
1,4-Dichlorobenzene	1.692	1.757		3.829	
n-Butylbenzene	2.049	2.075		1.259	
1,2-Dichlorobenzene	1.543	1.622		5.14	
1,2-Dibromo-3-chloropropane	0.136	0.130		-4.412	
1,2,4-Trichlorobenzene	1.009	1.030		2.051	
Hexachlorobutadiene	0.359	0.385		7.332	
Naphthalene	2.969	2.950		-0.651	
1,2,3-Trichlorobenzene	1.006	1.022		1.601	

All other compounds must meet a minimum RRF of 0.010.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: AMRO Environmental Laboratories Contract: _____Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039Instrument ID: V-3 Calibration Date: 08/10/05 Time: 7:38Lab File ID: E:\HPCHEM\1\DATA\081005 Init. Calib. Date(s): 07/26/05 07/26/05EPA Sample No.: 8260 ccv 25ppb Init. Calib. Times: 13:05 21:11Heated Purge: (Y/N) NGC Column: HP-624 ID: 0.20 (mm)

COMPOUND	IC RRF	CCV RRF	MIN RRF	%D	MAX %D
1,1,1,2-Tetrachloroethane	0.331	0.352		6.312	
1,1,1-Trichloroethane	0.260	0.300		15.296	
1,1,2,2-Tetrachloroethane	1.159	1.093	0.300	-5.695	
1,1,2-Trichloroethane	0.304	0.308		1.483	
1,1-Dichloroethane	0.516	0.534	0.100	3.428	
1,1-Dichloroethene	0.243	0.263		8.186	20.0
1,1-Dichloropropene	0.367	0.363		-1.063	
1,2,3-Trichlorobenzene	1.006	0.867		-13.809	
1,2,3-Trichloropropane	0.850	0.786		-7.562	
1,2,4-Trichlorobenzene	1.009	0.924		-8.451	
1,2,4-Trimethylbenzene	2.198	2.132		-3.016	
1,2-Dibromo-3-chloropropane	0.136	0.130		-4.412	
1,2-Dibromoethane	0.330	0.336		1.664	
1,2-Dichlorobenzene	1.543	1.497		-2.962	
1,2-Dichloroethane	0.334	0.370		10.745	
1,2-Dichloropropane	0.330	0.320		-2.883	20.0
1,3,5-Trimethylbenzene	2.114	2.058		-2.649	
1,3-Dichlorobenzene	1.636	1.559		-4.683	
1,3-Dichloropropane	0.736	0.678		-7.893	
1,4-Dichlorobenzene	1.692	1.654		-2.257	
1,4-Dioxane	0.003	0.003		12.5	
2,2-Dichloropropane	0.225	0.257		13.969	
2-Butanone	0.049	0.073		50	
2-Chlorotoluene	1.998	1.840		-7.917	
2-Hexanone	0.238	0.311		30.508	
4-Chlorotoluene	2.349	2.206		-6.092	
4-Isopropyltoluene	2.228	2.115		-5.076	
4-Methyl-2-pentanone	0.050	0.057		13.747	
Acetone	0.055	0.086		56.009	
Acrylonitrile	0.137	0.149		9.158	
Benzene	1.183	1.169		-1.158	
Bromobenzene	0.927	0.903		-2.547	
Bromochloromethane	0.173	0.191		10.277	
Bromodichloromethane	0.355	0.380		7.042	
Bromoform	0.289	0.298	0.100	3.079	

All other compounds must meet a minimum RRF of 0.010.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: AMRO Environmental Laboratories Contract: _____Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039Instrument ID: V-3 Calibration Date: 08/10/05 Time: 7:38Lab File ID: E:\HPCHEM\1\DATA\081005 Init. Calib. Date(s): 07/26/05 07/26/05EPA Sample No.: 8260 ccv 25ppb Init. Calib. Times: 13:05 21:11Heated Purge: (Y/N) NGC Column: HP-624 ID: 0.20 (mm)

COMPOUND	IC RRF	CCV RRF	MIN RRF	%D	MAX %D
Bromomethane	0.222	0.252		13.343	
Carbon disulfide	0.854	0.936		9.576	
Carbon tetrachloride	0.229	0.261		14.024	
Chlorobenzene	1.065	1.057	0.300	-0.77	
Chloroethane	0.187	0.224		20	
Chloroform	0.443	0.474		7.046	20.0
Chloromethane	0.332	0.356	0.100	7.121	
cis-1,2-Dichloroethene	0.308	0.321		4.322	
cis-1,3-Dichloropropene	0.494	0.489		-1.092	
Dibromochloromethane	0.432	0.454		5.093	
Dibromomethane	0.222	0.231		4.242	
Dichlorodifluoromethane	0.149	0.180		20.805	
Dichlorofluoromethane	0.418	0.481		15.194	
Diethyl ether	0.187	0.205		9.743	
Diisopropyl ether	1.055	0.980		-7.135	
Ethyl Tertiary Butyl Ether	0.629	0.652		3.723	
Ethylbenzene	1.566	1.512		-3.43	20.0
Freon-113	0.136	0.150		10.294	
Hexachlorobutadiene	0.359	0.356		-0.753	
Iodomethane	0.504	0.577		14.53	
Isopropylbenzene	2.493	2.317		-7.063	
m,p-Xylene	0.613	0.604		-1.468	
Methyl tert-butyl ether	0.454	0.471		3.836	
Methylene chloride	0.376	0.421		11.935	
n-Butylbenzene	2.049	1.909		-6.842	
n-Propylbenzene	3.160	3.007		-4.845	
Naphthalene	2.969	2.572		-13.381	
o-Xylene	0.591	0.572		-3.248	
sec-Butylbenzene	2.573	2.364		-8.112	
Styrene	1.103	1.118		1.342	
tert-Butylbenzene	1.696	1.558		-8.148	
Tertiary Amyl Methyl Ether	0.702	0.685		-2.449	
Tertiary Butanol	0.028	0.026		-6.137	
Tetrachloroethene	0.317	0.326		2.872	
Tetrahydrofuran	0.115	0.095		-17.231	

All other compounds must meet a minimum RRF of 0.010.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: AMRO Environmental Laboratories Contract: _____Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039Instrument ID: V-3 Calibration Date: 08/10/05 Time: 7:38Lab File ID: E:\HPCHEM\1\DATA\081005 Init. Calib. Date(s): 07/26/05 07/26/05EPA Sample No.: 8260 ccv 25ppb Init. Calib. Times: 13:05 21:11Heated Purge: (Y/N) NGC Column: HP-624 ID: 0.20 (mm)

COMPOUND	IC RRF	CCV RRF	MIN RRF	%D	MAX %D
Toluene	0.674	0.688		2.047	20.0
trans-1,2-Dichloroethene	0.279	0.304		8.766	
trans-1,3-Dichloropropene	0.420	0.431		2.497	
Trichloroethene	0.286	0.288		0.629	
Trichlorofluoromethane	0.248	0.284		14.701	
Vinyl acetate	0.018	0.023		27.778	
Vinyl chloride	0.284	0.299		5.134	20.0

All other compounds must meet a minimum RRF of 0.010.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: AMRO Environmental Laboratories Contract: _____Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039Instrument ID: V-3 Calibration Date: 08/11/05 Time: 7:13Lab File ID: C:\HPCHEM\1\DATA\081105 Init. Calib. Date(s): 07/26/05 07/26/05EPA Sample No.: 8260 ccv 25ppb Init. Calib. Times: 13:05 21:11Heated Purge: (Y/N) NGC Column: HP-624 ID: 0.20 (mm)

COMPOUND	IC RRF	CCV RRF	MIN RRF	%D	MAX %D
1,1,1,2-Tetrachloroethane	0.331	0.337		1.782	
1,1,1-Trichloroethane	0.260	0.286		9.915	
1,1,2,2-Tetrachloroethane	1.159	1.007	0.300	-13.115	
1,1,2-Trichloroethane	0.304	0.294		-3.13	
1,1-Dichloroethane	0.516	0.508	0.100	-1.608	
1,1-Dichloroethene	0.243	0.253		4.072	20.0
1,1-Dichloropropene	0.367	0.344		-6.241	
1,2,3-Trichlorobenzene	1.006	0.922		-8.341	
1,2,3-Trichloropropane	0.850	0.755		-11.208	
1,2,4-Trichlorobenzene	1.009	0.933		-7.56	
1,2,4-Trimethylbenzene	2.198	2.067		-5.973	
1,2-Dibromo-3-chloropropane	0.136	0.120		-11.765	
1,2-Dibromoethane	0.330	0.318		-3.782	
1,2-Dichlorobenzene	1.543	1.480		-4.064	
1,2-Dichloroethane	0.334	0.359		7.453	
1,2-Dichloropropane	0.330	0.302		-8.346	20.0
1,3,5-Trimethylbenzene	2.114	2.038		-3.595	
1,3-Dichlorobenzene	1.636	1.542		-5.723	
1,3-Dichloropropane	0.736	0.653		-11.289	
1,4-Dichlorobenzene	1.692	1.646		-2.73	
1,4-Dioxane	0.003	0.002		-25	
2,2-Dichloropropane	0.225	0.242		7.317	
2-Butanone	0.049	0.060		23.288	
2-Chlorotoluene	1.998	1.799		-9.969	
2-Hexanone	0.238	0.252		5.749	
4-Chlorotoluene	2.349	2.135		-9.114	
4-Isopropyltoluene	2.228	2.110		-5.3	
4-Methyl-2-pentanone	0.050	0.046		-8.204	
Acetone	0.055	0.069		25.17	
Acrylonitrile	0.137	0.121		-11.355	
Benzene	1.183	1.089		-7.923	
Bromobenzene	0.927	0.884		-4.597	
Bromochloromethane	0.173	0.186		7.39	
Bromodichloromethane	0.355	0.374		5.352	
Bromoform	0.289	0.295	0.100	2.041	

All other compounds must meet a minimum RRF of 0.010.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: AMRO Environmental Laboratories Contract: _____Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039Instrument ID: V-3 Calibration Date: 08/11/05 Time: 7:13Lab File ID: C:\HPCHEM\1\DATA\081105 Init. Calib. Date(s): 07/26/05 07/26/05EPA Sample No.: 8260 ccv 25ppb Init. Calib. Times: 13:05 21:11Heated Purge: (Y/N) NGC Column: HP-624 ID: 0.20 (mm)

COMPOUND	IC RRF	CCV RRF	MIN RRF	%D	MAX %D
Bromomethane	0.222	0.236		6.147	
Carbon disulfide	0.854	0.926		8.406	
Carbon tetrachloride	0.229	0.252		10.092	
Chlorobenzene	1.065	1.026	0.300	-3.68	
Chloroethane	0.187	0.200		7.143	
Chloroform	0.443	0.465		5.014	20.0
Chloromethane	0.332	0.325	0.100	-2.207	
cis-1,2-Dichloroethene	0.308	0.305		-0.877	
cis-1,3-Dichloropropene	0.494	0.465		-5.947	
Dibromochloromethane	0.432	0.455		5.324	
Dibromomethane	0.222	0.224		1.083	
Dichlorodifluoromethane	0.149	0.187		25.503	
Dichlorofluoromethane	0.418	0.458		9.686	
Diethyl ether	0.187	0.188		0.642	
Diisopropyl ether	1.055	0.916		-13.2	
Ethyl Tertiary Butyl Ether	0.629	0.628		-0.095	
Ethylbenzene	1.566	1.486		-5.09	20.0
Freon-113	0.136	0.154		13.235	
Hexachlorobutadiene	0.359	0.386		7.611	
Iodomethane	0.504	0.567		12.545	
Isopropylbenzene	2.493	2.272		-8.868	
m,p-Xylene	0.613	0.594		-3.1	
Methyl tert-butyl ether	0.454	0.441		-2.778	
Methylene chloride	0.376	0.382		1.566	
n-Butylbenzene	2.049	1.896		-7.476	
n-Propylbenzene	3.160	2.931		-7.25	
Naphthalene	2.969	2.526		-14.93	
o-Xylene	0.591	0.562		-4.939	
sec-Butylbenzene	2.573	2.367		-7.995	
Styrene	1.103	1.075		-2.556	
tert-Butylbenzene	1.696	1.577		-7.027	
Tertiary Amyl Methyl Ether	0.702	0.639		-9	
Tertiary Butanol	0.028	0.021		-24.188	
Tetrachloroethene	0.317	0.324		2.24	
Tetrahydrofuran	0.115	0.070		-39.013	

All other compounds must meet a minimum RRF of 0.010.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: AMRO Environmental Laboratories Contract: _____Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039Instrument ID: V-3 Calibration Date: 08/11/05 Time: 7:13Lab File ID: C:\HPCHEM\1\DATA\081105 Init. Calib. Date(s): 07/26/05 07/26/05EPA Sample No.: 8260 ccv 25ppb Init. Calib. Times: 13:05 21:11Heated Purge: (Y/N) NGC Column: HP-624 ID: 0.20 (mm)

COMPOUND	IC RRF	CCV RRF	MIN RRF	%D	MAX %D
Toluene	0.674	0.646		-4.183	20.0
trans-1,2-Dichloroethene	0.279	0.286		2.326	
trans-1,3-Dichloropropene	0.420	0.403		-4.162	
Trichloroethene	0.286	0.279		-2.516	
Trichlorofluoromethane	0.248	0.301		21.567	
Vinyl acetate	0.018	0.024		33.333	
Vinyl chloride	0.284	0.269		-5.415	20.0

NOT ANALYZED
TB

All other compounds must meet a minimum RRF of 0.010.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: AMRO Environmental Laboratories Contract: _____Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039Instrument ID: V-3 Calibration Date: 08/12/05 Time: 7:12Lab File ID: C:\HPCHEM\1\DATA\081205 Init. Calib. Date(s): 07/26/05 07/26/05EPA Sample No.: 8260 ccv 25ppb Init. Calib. Times: 13:05 21:11Heated Purge: (Y/N) NGC Column: HP-624 ID: 0.20 (mm)

COMPOUND	IC RRF	CCV RRF	MIN RRF	%D	MAX %D
1,1,1,2-Tetrachloroethane	0.331	0.000		-100	
1,1,1-Trichloroethane	0.260	0.000		-100	
1,1,2,2-Tetrachloroethane	1.159	0.000	0.300	-100	
1,1,2-Trichloroethane	0.304	0.000		-100	
1,1-Dichloroethane	0.516	0.000	0.100	-100	
1,1-Dichloroethene	0.243	0.000		-100	20.0
1,1-Dichloropropene	0.367	0.000		-100	
1,2,3-Trichlorobenzene	1.006	0.000		-100	
1,2,3-Trichloropropane	0.850	0.000		-100	
1,2,4-Trichlorobenzene	1.009	0.000		-100	
1,2,4-Trimethylbenzene	2.198	0.000		-100	
1,2-Dibromo-3-chloropropane	0.136	0.000		-100	
1,2-Dibromoethane	0.330	0.000		-100	
1,2-Dichlorobenzene	1.543	0.000		-100	
1,2-Dichloroethane	0.334	0.000		-100	
1,2-Dichloropropane	0.330	0.000		-100	20.0
1,3,5-Trimethylbenzene	2.114	0.000		-100	
1,3-Dichlorobenzene	1.636	0.000		-100	
1,3-Dichloropropane	0.736	0.000		-100	
1,4-Dichlorobenzene	1.692	0.000		-100	
1,4-Dioxane	0.003	0.000		-100	
2,2-Dichloropropane	0.225	0.000		-100	
2-Butanone	0.049	0.000		-100	
2-Chlorotoluene	1.998	0.000		-100	
2-Hexanone	0.238	0.000		-100	
4-Chlorotoluene	2.349	0.000		-100	
4-Isopropyltoluene	2.228	0.000		-100	
4-Methyl-2-pentanone	0.050	0.000		-100	
Acetone	0.055	0.000		-100	
Acrylonitrile	0.137	0.000		-100	
Benzene	1.183	0.000		-100	
Bromobenzene	0.927	0.000		-100	
Bromochloromethane	0.173	0.000		-100	
Bromodichloromethane	0.355	0.000		-100	
Bromoform	0.289	0.000	0.100	-100	

*verified
for
new
data
TS*

All other compounds must meet a minimum RRF of 0.010.

7B
VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: AMRO Environmental Laboratories Contract: _____
 Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039
 Instrument ID: V-3 Calibration Date: 08/12/05 Time: 7:12
 Lab File ID: C:\HPCHEM\1\DATA\081205 Init. Calib. Date(s): 07/26/05 07/26/05
 EPA Sample No.: 8260 ccv 25ppb Init. Calib. Times: 13:05 21:11
 Heated Purge: (Y/N) N
 GC Column: HP-624 ID: 0.20 (mm)

COMPOUND	IC RRF	CCV RRF	MIN RRF	%D	MAX %D
Bromomethane	0.222	0.000		-100	
Carbon disulfide	0.854	0.000		-100	
Carbon tetrachloride	0.229	0.000		-100	
Chlorobenzene	1.065	0.000	0.300	-100	
Chloroethane	0.187	0.000		-100	
Chloroform	0.443	0.000		-100	20.0
Chloromethane	0.332	0.000	0.100	-100	
cis-1,2-Dichloroethene	0.308	0.000		-100	
cis-1,3-Dichloropropene	0.494	0.000		-100	
Dibromochloromethane	0.432	0.000		-100	
Dibromomethane	0.222	0.000		-100	
Dichlorodifluoromethane	0.149	0.000		-100	
Dichlorofluoromethane	0.418	0.000		-100	
Diethyl ether	0.187	0.000		-100	
Diisopropyl ether	1.055	0.000		-100	
Ethyl Tertiary Butyl Ether	0.629	0.000		-100	
Ethylbenzene	1.566	0.000		-100	20.0
Freon-113	0.136	0.000		-100	
Hexachlorobutadiene	0.359	0.000		-100	
Iodomethane	0.504	0.000		-100	
Isopropylbenzene	2.493	0.000		-100	
m,p-Xylene	0.613	0.000		-100	
Methyl tert-butyl ether	0.454	0.000		-100	
Methylene chloride	0.376	0.000		-100	
n-Butylbenzene	2.049	0.000		-100	
n-Propylbenzene	3.160	0.000		-100	
Naphthalene	2.969	0.000		-100	
o-Xylene	0.591	0.000		-100	
sec-Butylbenzene	2.573	0.000		-100	
Styrene	1.103	0.000		-100	
tert-Butylbenzene	1.696	0.000		-100	
Tertiary Amyl Methyl Ether	0.702	0.000		-100	
Tertiary Butanol	0.028	0.000		-100	
Tetrachloroethene	0.317	0.000		-100	
Tetrahydrofuran	0.115	0.000		-100	

All other compounds must meet a minimum RRF of 0.010.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: AMRO Environmental Laboratories Contract: _____
Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039
Instrument ID: V-3 Calibration Date: 08/12/05 Time: 7:12
Lab File ID: C:\HPCHEM\1\DATA\081205 Init. Calib. Date(s): 07/26/05 07/26/05
EPA Sample No.: 8260 ccv 25ppb Init. Calib. Times: 13:05 21:11
Heated Purge: (Y/N) N
GC Column: HP-624 ID: 0.20 (mm)

COMPOUND	IC RRF	CCV RRF	MIN RRF	%D	MAX %D
Toluene	0.674	0.000		-100	20.0
trans-1,2-Dichloroethene	0.279	0.000		-100	
trans-1,3-Dichloropropene	0.420	0.000		-100	
Trichloroethene	0.286	0.000		-100	
Trichlorofluoromethane	0.248	0.000		-100	
Vinyl acetate	0.018	0.000		-100	
Vinyl chloride	0.284	0.000		-100	20.0

All other compounds must meet a minimum RRF of 0.010.

8A
VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: AMRO Environmental Laboratories Contract: _____

Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039

Lab File ID (Standard): C:\HPCHEM\1\DATA\08060 Date Analyzed: 08/06/05

EPA Sample No. (VSTD050##) 8260 ccv 25ppb Time Analyzed: 6:29

Instrument ID: V-3 Heated Purge: (Y/N) N

GC Column: HP-624 ID: 0.20 (mm)

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	1107292	8.13	839905	13.88	424341	18.39
UPPER LIMIT	2214584	8.63	1679810	14.38	848682	18.89
LOWER LIMIT	553646	7.63	419953	13.38	212171	17.89
EPA SAMPLE						
01 lcsf-08/06/05	1086199	8.130	808019	13.880	404876	18.390
02 lcsf-08/06/05	1086199	8.130	808019	13.880	404876	18.390
03 mb-08/06/05	1043290	8.120	795080	13.870	382598	18.380
04 mb-08/06/05	1043290	8.120	795080	13.870	382598	18.380
05 ZZZZZ	1017875	8.130	774692	13.870	378911	18.390
06 ZZZZZ	974446	8.130	740605	13.860	348335	18.380
07 ZZZZZ	984582	8.130	733175	13.860	350676	18.380
08 ZZZZZ	974107	8.120	739436	13.870	359665	18.380
09 ZZZZZ	1005524	8.130	750399	13.870	349371	18.380
10 ZZZZZ	973075	8.130	745051	13.870	339744	18.390
11 ZZZZZ	960094	8.140	723175	13.870	342125	18.380
12 ZZZZZ	971620	8.130	721872	13.870	340515	18.390
13 ZZZZZ	969212	8.130	740209	13.870	343575	18.380
14 0508040-04Amsf	1011488	8.140	753309	13.870	386279	18.380
15 0508040-04Amsdf	994826	8.120	754835	13.870	386711	18.380
16 MW-9D	973981	8.120	746347	13.870	343540	18.380
17 MW-5A	940305	8.130	725053	13.860	337295	18.380
18 MW-8B	976620	8.120	736256	13.860	346247	18.380
19 MW-8C	984376	8.120	745294	13.850	345092	18.380

IS1 = Fluorobenzene
IS2 = Chlorobenzene-d5
IS3 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

page 1 of 5

FORM VIII VOA

OLM04.2

8A
VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: AMRO Environmental Laboratories Contract: _____

Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039

Lab File ID (Standard): C:\HPCHEM\1\DATA\08090 Date Analyzed: 08/09/05

EPA Sample No. (VSTD050##) 8260 ccv 25ppb Time Analyzed: 9:45

Instrument ID: V-3 Heated Purge: (Y/N) N

GC Column: HP-624 ID: 0.20 (mm)

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	1046362	8.13	805551	13.88	406619	18.39
UPPER LIMIT	2092724	8.63	1611102	14.38	813238	18.89
LOWER LIMIT	523181	7.63	402776	13.38	203310	17.89
EPA SAMPLE						
01 lcsf-08/09/05	1115301	8.140	853545	13.870	414640	18.400
02 mb-08/09/05	1050111	8.140	795984	13.880	371590	18.400
03 MW-8A	1035159	8.150	776195	13.880	364911	18.400
04 MW-8A	990948	8.150	739434	13.880	343668	18.410
05 ZZZZZ	1003359	8.140	756772	13.880	344611	18.400
06 MW-5D	980445	8.150	750516	13.880	349977	18.410
07 MW-5B	944909	8.150	719562	13.880	340972	18.390
08 MW-9A	921864	8.130	697906	13.880	332707	18.390
09 MW-8D	908175	8.150	690835	13.880	327895	18.410
10 MW-8D	934131	8.150	739407	13.880	371530	18.410
11 MW-8D	994877	8.150	749313	13.890	386067	18.410

IS1 = Fluorobenzene
IS2 = Chlorobenzene-d5
IS3 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

page 2 of 5

FORM VIII VOA

OLM04.2

8A
VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: AMRO Environmental Laboratories Contract: _____

Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039

Lab File ID (Standard): E:\HPCHEM\1\DATA\08100 Date Analyzed: 08/10/05

EPA Sample No. (VSTD050##) 8260 ccv 25ppb Time Analyzed: 7:38

Instrument ID: V-3 Heated Purge: (Y/N) N

GC Column: HP-624 ID: 0.20 (mm)

	AREA #	RT #	AREA #	RT #	AREA #	RT #
12 HOUR STD	0	0	0	0	0	0
UPPER LIMIT	0	0.5	0	0.5	0	0.5
LOWER LIMIT	0	-0.5	0	-0.5	0	-0.5
EPA SAMPLE						
01 lcsf-08/10/05	966541	8.150	737713	13.880	373449	18.410
02 mb-08/10/05	974426	8.150	747410	13.880	341089	18.410
03 MW-5C	944825	8.150	721388	13.890	334415	18.400
04 MW-7B	915186	8.150	682868	13.890	319664	18.400
05 MW-7C	888850	8.150	675801	13.880	312705	18.410
06 MW-5B	878778	8.150	683182	13.880	328827	18.410
07 MW-5C	930462	8.150	709299	13.880	365071	18.410
08 MW-5C	989639	8.150	760400	13.880	384970	18.410
09 MW-9B	917385	8.150	687074	13.880	322328	18.400
10 MW-9C	885793	8.150	683573	13.880	312396	18.410
11 MW-2B	858303	8.150	666808	13.880	300620	18.410
12 MW-1	892038	8.150	697681	13.880	317688	18.400
13 MW-11C	882942	8.150	672589	13.880	318549	18.410
14 MW-6A	854783	8.150	671101	13.880	307766	18.400
15 MW-6B	860434	8.150	687350	13.880	316446	18.410
16 ZZZZZ						
17 ZZZZZ						
18 ZZZZZ						
19 ZZZZZ						

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

page 3 of 5

FORM VIII VOA

OLM04.2

8A
VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: AMRO Environmental Laboratories Contract: _____

Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039

Lab File ID (Standard): C:\HPCHEM\1\DATA\08110 Date Analyzed: 08/11/05

EPA Sample No. (VSTD050##) 8260 ccv 25ppb Time Analyzed: 7:13

Instrument ID: V-3 Heated Purge: (Y/N) N

GC Column: HP-624 ID: 0.20 (mm)

		AREA #	RT #	AREA #	RT #	AREA #	RT #
	12 HOUR STD	0	0	0	0	0	0
	UPPER LIMIT	0	0.5	0	0.5	0	0.5
	LOWER LIMIT	0	-0.5	0	-0.5	0	-0.5
	EPA SAMPLE						
01	lcsf-08/11/05	891748	8.160	667813	13.890	341275	18.420
02	lcsf-08/11/05			667813	13.890	341275	18.420
03	mb-08/11/05	918137	8.170	682975	13.900	314486	18.410
04	mb-08/11/05	918137	8.170	682975	13.900	314486	18.410
05	MW-2A	856019	8.150	635485	13.900	294722	18.410
06	MW-6C	838628	8.150	631989	13.900	284849	18.410
07	MW-11D	827716	8.150	611974	13.890	283940	18.410
08	MW-3	857241	8.150	643547	13.880	299356	18.410
09	MW-12C	839588	8.150	662701	13.890	295132	18.410
10	MW-7A	803551	8.150	622440	13.880	280918	18.410
11	MW-2A	934950	8.150	730293	13.880	366576	18.410
12	MW-2A	992045	8.150	753100	13.880	386962	18.410
13	MW-4	856497	8.150	642484	13.880	290636	18.410
14	ZZZZZ	820241	8.150	609269	13.890	273252	18.410
15	ZZZZZ	794612	8.150	597872	13.890	268712	18.410
16	ZZZZZ	812784	8.150	607756	13.880	272620	18.410
17	ZZZZZ	840851	8.150	616308	13.880	273124	18.410

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

page 4 of 5

FORM VIII VOA

OLM04.2

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: AMRO Environmental Laboratories Contract: _____Lab Code: AMRO Case No.: LBG NY SAS No.: _____ SDG No.: 0508039Lab File ID (Standard): C:\HPCHEM\1\DATA\08120 Date Analyzed: 08/12/05EPA Sample No. (VSTD050##) 8260 ccv 25ppb Time Analyzed: 7:12Instrument ID: V-3 Heated Purge: (Y/N) NGC Column: HP-624 ID: 0.20 (mm)

		AREA #	RT #	AREA #	RT #	AREA #	RT #
	12 HOUR STD	0	0	0	0	0	0
	UPPER LIMIT	0	0.5	0	0.5	0	0.5
	LOWER LIMIT	0	-0.5	0	-0.5	0	-0.5
	EPA SAMPLE						
01	lcsf-08/12/05	871502	8.150	662778	13.880	339181	18.400
02	lcsf-08/12/05			662778	13.880	339181	18.400
03	mb-08/12/05	956667	8.140	708416	13.890	311437	18.400
04	mb-08/12/05	956667	8.140	708416	13.890	311437	18.400
05	MW-6D	810067	8.150	622658	13.890	285027	18.400
06	MW-12D	870972	8.140	648583	13.880	288722	18.410
07	ZZZZZ	821655	8.150	617201	13.880	291244	18.410
08	MW-2B	797702	8.150	598956	13.890	261344	18.410
09	MW-12C	920124	8.150	714772	13.890	363964	18.410
10	MW-6B	858356	8.150	653581	13.880	326278	18.410
11	MW-7A	637017	8.140	471086	13.880	207127	18.400
12	ZZZZZ	566941	8.150	432096	13.880	191984	18.400
13	MW-6D	866088	8.130	655703	13.880	323818	18.390
14	MW-6D	832190	8.150	634851	13.880	323753	18.410
15	ZZZZZ	803536	8.160	599091	13.890	282937	18.420
16	ZZZZZ	752584	8.160	556669	13.890	254119	18.420

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

GC/MS TUNE RESULTS

BFB

Data File : C:\HPCHEM\1\DATA\072605\G1579.D

Vial: 100

Acq On : 26 Jul 05 11:51 am

Operator: kty

Sample : [bfb tune 50 ng]

Inst : voa 3


Misc :

Multiplr: 1.00000

MS Integration Params: rteint.p

Method : C:\HPCHEM\1\METHODS\8A_05_26.M (RTE Integrator)

Title : 8260 DB-624 col 25m x 0.2mm V-3



Spectrum Information: Average of 6.627 to 6.725 min.

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.6	2964	PASS
75	95	30	60	39.7	7096	PASS
95	95	100	100	100.0	17852	PASS
96	95	5	9	6.8	1213	PASS
173	174	0.00	2	0.3	40	PASS
174	95	50	120	87.6	15647	PASS
175	174	5	9	7.5	1166	PASS
176	174	95	101	97.9	15319	PASS
177	176	5	9	5.9	911	PASS

BFB

Data File : C:\HPCHEM\1\DATA\080605\G1703.D

Acq On : 6 Aug 05 5:51 am

Sample : [bfb tune 50 ng]

Misc :

Vial: 100

Operator: kty


Inst : voa 3

Multiplr: 1.00000

MS Integration Params: rteint.p

Method : C:\HPCHEM\1\METHODS\8A_07_26.M (RTE Integrator)

Title : 8260 DB-624 col 25m x 0.2mm V-3



Spectrum Information: Average of 6.616 to 6.689 min.

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.7	3560	PASS
75	95	30	60	40.4	8120	PASS
95	95	100	100	100.0	20080	PASS
96	95	5	9	6.9	1378	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	90.1	18102	PASS
175	174	5	9	7.0	1261	PASS
176	174	95	101	97.4	17628	PASS
177	176	5	9	7.2	1278	PASS

BFB

Data File : C:\HPCHEM\1\DATA\080905\G1731.D

Vial: 100

Acq On : 9 Aug 05 8:38 am

Operator: kty

Sample : [bfb tune 50 ng]

Inst : voa 3


Misc :

Multiplr: 1.00000

MS Integration Params: rteint.p

Method : C:\HPCHEM\1\METHODS\8A_07_26.M (RTE Integrator)

Title : 8260 DB-624 col 25m x 0.2mm V-3



Spectrum Information: Average of 6.624 to 6.689 min.

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.2	3205	PASS
75	95	30	60	44.0	9286	PASS
95	95	100	100	100.0	21128	PASS
96	95	5	9	6.6	1385	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	94.7	20004	PASS
175	174	5	9	5.6	1123	PASS
176	174	95	101	96.9	19388	PASS
177	176	5	9	6.5	1254	PASS

kty 08/10/05

BFB

Data File : C:\HPCHEM\1\DATA\081005\G1754.D

Vial: 100

Acq On : 10 Aug 05 6:46 am

Operator: kty

Sample : [bfb tune 50 ng]

Inst : voa 3


Misc :

Multiplr: 1.00000

MS Integration Params: rteint.p

Method : C:\HPCHEM\1\METHODS\8A_07_26.M (RTE Integrator)

Title : 8260 DB-624 col 25m x 0.2mm V-3



Spectrum Information: Average of 6.624 to 6.706 min.

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.5	2634	PASS
75	95	30	60	40.6	6478	PASS
95	95	100	100	100.0	15945	PASS
96	95	5	9	8.2	1310	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	95.4	15204	PASS
175	174	5	9	7.7	1165	PASS
176	174	95	101	96.1	14612	PASS
177	176	5	9	7.7	1131	PASS

Key 08/10/05

BFB

Data File : C:\HPCHEM\1\DATA\081105\G1783.D

Vial: 100

Acq On : 11 Aug 05 6:00 am

Operator: kty

Sample : [bfb tune 50 ng]

Inst : voa 3


Misc :

Multiplr: 1.00000

MS Integration Params: rteint.p

Method : C:\HPCHEM\1\METHODS\8A_07_26.M (RTE Integrator)

Title : 8260 DB-624 col 25m x 0.2mm V-3



Spectrum Information: Average of 6.638 to 6.720 min.

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.9	2719	PASS
75	95	30	60	39.9	6060	PASS
95	95	100	100	100.0	15202	PASS
96	95	5	9	6.4	970	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	88.7	13487	PASS
175	174	5	9	7.5	1008	PASS
176	174	95	101	96.6	13023	PASS
177	176	5	9	6.8	884	PASS

Aug 08/11/05

BFB

Data File : C:\HPCHEM\1\DATA\081205\G1819.D

Acq On : 12 Aug 05 6:08 am

Sample : [bfb tune 50 ng]

Misc :

MS Integration Params: rteint.p

Method : C:\HPCHEM\1\METHODS\8A_07_26.M (RTE Integrator)


Title : 8260 DB-624 col 25m x 0.2mm V-3

Vial: 100

Operator: kty

Inst : voa 3

Multiplr: 1.00000



Spectrum Information: Average of 6.632 to 6.689 min.

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.7	3037	PASS
75	95	30	60	42.7	7761	PASS
95	95	100	100	100.0	18196	PASS
96	95	5	9	7.0	1276	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	89.6	16309	PASS
175	174	5	9	6.4	1043	PASS
176	174	95	101	97.1	15836	PASS
177	176	5	9	7.1	1129	PASS

Handwritten signature
08/12/05

METHOD BLANK

Data File : C:\HPCHEM\1\DATA\080605\G1707.D

Vial: 100

Acq On : 6 Aug 05 8:21 am

Operator: kty

Sample : [mb-08/06/05]df=1MF=5UG{8260_WX}

Inst : voa 3

Misc : xtr08/06/05 samp=5ml fv=5ml

Multiplr: 1.00000

MS Integration Params: rteint.p

Quant Time: Aug 6 12:58 2005 Quant Results File: 8A_07_26.RES

Quant Method : C:\HPCHEM\1\METHODS\8A_07_26.M (RTE Integrator)

Title : 8260 DB-624 col 25m x 0.2mm V-3

Last Update : Wed Jul 27 12:23:07 2005

Response via : Initial Calibration

DataAcq Meth : 8A_07_26

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	8.12	96	1043290	25.00	ug/l	-0.03
58) Chlorobenzene-d5	13.87	117	795080	25.00	ug/l	-0.03
71) 1,4-Dichlorobenzene-d4	18.38	152	382598	25.00	ug/l	-0.05

System Monitoring Compounds

32) Dibromofluoromethane	6.85	113	289699	25.18	ug/l	-0.03
Spiked Amount	25.000	Range	85 - 116	Recovery	=	100.72%
35) 1,2-Dichloroethane-d4	7.47	65	285071	26.93	ug/l	-0.03
Spiked Amount	25.000	Range	77 - 127	Recovery	=	107.72%
52) Toluene-d8	11.15	98	933796	25.11	ug/l	-0.03
Spiked Amount	25.000	Range	86 - 114	Recovery	=	100.44%
59) 4-Bromofluorobenzene	16.16	174	314701	24.23	ug/l	-0.03
Spiked Amount	25.000	Range	79 - 117	Recovery	=	96.92%
94) 2,5-Dibromotoluene	0.00	250	0	0.00	ug/l	
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.00%#

Target Compounds

92) Naphthalene	22.78	128	24124	0.53	ug/l	Qvalue 97
-----------------	-------	-----	-------	------	------	-----------

Handwritten signature and date:
08/06/05

Quantitation Report

Data File : C:\HPCHEM\1\DATA\080605\G1707.D

Vial: 100

Acq On : 6 Aug 05 8:21 am

Operator: kty

Sample : [mb-08/06/05]df=1MF=5UG{8260_WX}

Inst : voa 3

Misc : xtr08/06/05 samp=5ml fv=5ml

Multiplr: 1.00000

MS Integration Params: rteint.p


Quant Time: Aug 6 12:58 2005 Quant Results File: 8A_07_26.RES

Method : C:\HPCHEM\1\METHODS\8A_07_26.M (RTE Integrator)

Title : 8260 DB-624 col 25m x 0.2mm V-3

Last Update : Wed Jul 27 12:23:07 2005

Response via : Initial Calibration



Data File : C:\HPCHEM\1\DATA\080605\G1707.D

Vial: 100

Acq On : 6 Aug 05 8:21 am

Operator: kty

Sample : [mb-08/06/05]df=1MF=5UG{8260_WX}

Inst : voa 3

Misc : xtr08/06/05 samp=5ml fv=5ml

Multiplr: 1.00000

MS Integration Params: rteint.p

Quant Time: Aug 6 8:50 2005 Quant Results File: 8A_07_26.RES

Quant Method : C:\HPCHEM\1\METHODS\8A_07_26.M (RTE Integrator)

Title : 8260 DB-624 col 25m x 0.2mm V-3

Last Update : Wed Jul 27 12:23:07 2005

Response via : Initial Calibration

DataAcq Meth : 8A_07_26

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	8.12	96	1043290	25.00	ug/l	-0.03
58) Chlorobenzene-d5	13.87	117	795080	25.00	ug/l	-0.03
71) 1,4-Dichlorobenzene-d4	18.38	152	382598	25.00	ug/l	-0.05

System Monitoring Compounds

32) Dibromofluoromethane	6.85	113	289699	25.18	ug/l	-0.03
Spiked Amount	25.000	Range	85 - 116	Recovery	=	100.72%
35) 1,2-Dichloroethane-d4	7.47	65	285071	26.93	ug/l	-0.03
Spiked Amount	25.000	Range	77 - 127	Recovery	=	107.72%
52) Toluene-d8	11.15	98	933796	25.11	ug/l	-0.03
Spiked Amount	25.000	Range	86 - 114	Recovery	=	100.44%
59) 4-Bromofluorobenzene	16.16	174	314701	24.23	ug/l	-0.03
Spiked Amount	25.000	Range	79 - 117	Recovery	=	96.92%
94) 2,5-Dibromotoluene	0.00	250	0	0.00	ug/l	
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.00%#

Target Compounds

					Qvalue	
6) Bromomethane	1.94	94	1694	-5.21	ug/l	# 11
17) Acrolein	2.94	56	517	30.29	ug/l	# 69
27) 2,2-Dichloropropane	5.93	77	1646	-1.59	ug/l	# 45
30) Tetrahydrofuran	6.52	42	411	-1.38	ug/l	# 37
92) Naphthalene	22.78	128	24124	0.53	ug/l	97

 (#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : C:\HPCHEM\1\DATA\080605\G1707.D

Vial: 100

Acq On : 6 Aug 05 8:21 am

Operator: kty

Sample : [mb-08/06/05]df=1MF=5UG{8260_WX}

Inst : voa 3

Misc : xtr08/06/05 samp=5ml fv=5ml

Multiplr: 1.00000

MS Integration Params: rteint.p


Quant Time: Aug 6 8:50 2005 Quant Results File: 8A_07_26.RES

Method : C:\HPCHEM\1\METHODS\8A_07_26.M (RTE Integrator)

Title : 8260 DB-624 col 25m x 0.2mm V-3

Last Update : Wed Jul 27 12:23:07 2005




Response via : Initial Calibration



Quantitation Report

Data File : C:\HPCHEM\1\DATA\080605\G1707.D Vial: 100
 Acq On : 6 Aug 05 8:21 am Operator: kty
 Sample : [mb-08/06/05]df=1MF=5UG{8260_WX} Inst : voa 3
 Misc : xtr08/06/05 samp=5ml fv=5ml Multiplr: 1.00000
 Quant Time: Aug 6 12:58 2005 Quant Results-File: temp.res

Method : C:\HPCHEM\1\METHODS\8A_07_26.M (RTE Integrator)
 Title : 8260 DB-624 col 25m x 0.2mm V-3
 Last Update : Wed Jul 27 12:23:07 2005
 Response via : Multiple Level Calibration



(92) Naphthalene (v)

22.78min 0.53ug/l

response 24124

Ion	Exp%	Act%
128.00	100	100
129.00	10.70	9.44
0.00	0.00	0.00
0.00	0.00	0.00

Data File : C:\HPCHEM\1\DATA\080905\G1737.D Vial: 100
Acq On : 9 Aug 05 12:32 pm Operator: kty
Sample : [mb-08/09/05]df=1MF=5UG{8260_WX} Inst : voa 3
Misc : xtr08/09/05 samp=5ml fv=5ml Multiplr: 1.00000
MS Integration Params: rteint.p
Quant Time: Aug 10 9:23 2005 Quant Results File: 8A_07_26.RES

Quant Method : C:\HPCHEM\1\METHODS\8A_07_26.M (RTE Integrator)
Title : 8260 DB-624 col 25m x 0.2mm V-3
Last Update : Wed Jul 27 12:23:07 2005
Response via : Initial Calibration
DataAcq Meth : 8A_07_26

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	8.14	96	1050111	25.00	ug/l	-0.02
58) Chlorobenzene-d5	13.88	117	795984	25.00	ug/l	-0.02
71) 1,4-Dichlorobenzene-d4	18.40	152	371590	25.00	ug/l	-0.03

System Monitoring Compounds

32) Dibromofluoromethane	6.87	113	311255	26.88	ug/l	-0.02
Spiked Amount	25.000	Range	85 - 116	Recovery	=	107.52%
35) 1,2-Dichloroethane-d4	7.48	65	272736	25.60	ug/l	-0.02
Spiked Amount	25.000	Range	77 - 127	Recovery	=	102.40%
52) Toluene-d8	11.16	98	931158	24.88	ug/l	-0.02
Spiked Amount	25.000	Range	86 - 114	Recovery	=	99.52%
59) 4-Bromofluorobenzene	16.16	174	305921	23.52	ug/l	-0.03
Spiked Amount	25.000	Range	79 - 117	Recovery	=	94.08%
94) 2,5-Dibromotoluene	0.00	250	0	0.00	ug/l	
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.00%#

Target Compounds


Qvalue

Handwritten signature
08/10/05

Quantitation Report

Data File : C:\HPCHEM\1\DATA\080905\G1737.D Vial: 100
Acq On : 9 Aug 05 12:32 pm Operator: kty
Sample : [mb-08/09/05]df=1MF=5UG(8260_WX) Inst : voa 3
Misc : xtr08/09/05 samp=5ml fv=5ml Multiplr: 1.00000
MS Integration Params: rteint.p
Quant Time: Aug 10 9:23 2005 Quant Results File: 8A_07_26.RES

Method : C:\HPCHEM\1\METHODS\8A_07_26.M (RTE Integrator)
Title : 8260 DB-624 col 25m x 0.2mm V-3
Last Update : Wed Jul 27 12:23:07 2005
Response via : Initial Calibration



Data File : C:\HPCHEM\1\DATA\080905\G1737.D Vial: 100
Acq On : 9 Aug 05 12:32 pm Operator: kty
Sample : [mb-08/09/05]df=1MF=5UG{8260_WX} Inst : voa 3
Misc : xtr08/09/05 samp=5ml fv=5ml Multiplr: 1.00000
MS Integration Params: rteint.p
Quant Time: Aug 9 13:01 2005 Quant Results File: 8A_07_26.RES

Quant Method : C:\HPCHEM\1\METHODS\8A_07_26.M (RTE Integrator)
Title : 8260 DB-624 col 25m x 0.2mm V-3
Last Update : Wed Jul 27 12:23:07 2005
Response via : Initial Calibration
DataAcq Meth : 8A_07_26

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	8.14	96	1050111	25.00	ug/l	-0.02
58) Chlorobenzene-d5	13.88	117	795984	25.00	ug/l	-0.02
71) 1,4-Dichlorobenzene-d4	18.40	152	371590	25.00	ug/l	-0.03

System Monitoring Compounds

32) Dibromofluoromethane	6.87	113	311255	26.88	ug/l	-0.02
Spiked Amount	25.000	Range	85 - 116	Recovery	=	107.52%
35) 1,2-Dichloroethane-d4	7.48	65	272736	25.60	ug/l	-0.02
Spiked Amount	25.000	Range	77 - 127	Recovery	=	102.40%
52) Toluene-d8	11.16	98	931158	24.88	ug/l	-0.02
Spiked Amount	25.000	Range	86 - 114	Recovery	=	99.52%
59) 4-Bromofluorobenzene	16.16	174	305921	23.52	ug/l	-0.03
Spiked Amount	25.000	Range	79 - 117	Recovery	=	94.08%
94) 2,5-Dibromotoluene	0.00	250	0	0.00	ug/l	
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.00%#

Target Compounds

					Qvalue
6) Bromomethane	2.03	94	2193	-5.14 ug/l #	11
27) 2,2-Dichloropropane	5.86	77	889	-1.70 ug/l #	45
30) Tetrahydrofuran	6.57	42	494	-1.36 ug/l #	37

Quantitation Report

Data File : C:\HPCHEM\1\DATA\080905\G1737.D
Acq On : 9 Aug 05 12:32 pm
Sample : [mb-08/09/05]df=1MF=5UG{8260_WX}
Misc : xtr08/09/05 samp=5ml fv=5ml
MS Integration Params: rteint.p

Vial: 100
Operator: kty
Inst : voa 3
Multiplr: 1.00000

Quant Time: Aug 9 13:01 2005 Quant Results File: 8A_07_26.RES

Method : C:\HPCHEM\1\METHODS\8A_07_26.M (RTE Integrator)
Title : 8260 DB-624 col 25m x 0.2mm V-3
Last Update : Wed Jul 27 12:23:07 2005
Response via : Initial Calibration

