

Norlite Corporation

2003.01.17 - HW
ARAT Protocol
FILE
10.11.16
Copies to LEONE ✓
BRZNER ✓



628 SO. SARATOGA ST.
P.O. BOX 694
COHOES, NY 12047
PHONE (518) 235-0401
FAX (518) 235-0233

February 3, 2003

Mr. William J. Clarke
Regional Permit Administrator
New York State Department of Environmental Conservation
Region 4
1150 North Westcott Road
Schenectady, NY 12306-2014

Certified: 7001-0320-0001-8803-2014

Mr. Parag Amin, PE
New York State Department of Environmental Conservation
Hazardous Waste Bureau
625 Broadway
Albany, NY 12233-7251

Certified: 7001-0320-0001-8803-1987

Dear Sirs:

Enclosed you will find a copy of Norlite's test protocol for its 2003 Annual Relative Accuracy Test Audit (RATA) per 373 Permit Module VII.D.3. Norlite anticipates the field test to be scheduled for mid March 2003.


If you have any questions, please feel free to contact me directly.

Sincerely,

Timothy F. Lachell
Plant Manager

cc: Ralph Tommasino
Dave Ostaszewski (ENSR)

File:c:\wp60\text\tim\02030301.wpd





360 Old Colony Road • Suite 1
Norton, MA 02766
(508) 226-6700 • Fax (508) 226-6778


PRETEST PROTOCOL
NORLITE CORPORATION
CONTINUOUS EMISSION MONITORING SYSTEM
RELATIVE ACCURACY TEST AUDIT
1ST QUARTER 2003

Source Designation:
Norlite Corporation
Rotary Kilns No. 1 & 2
628 South Saratoga Street
Cohoes, New York


Concerning:
CEMS Relative Accuracy Test Audit
O₂, CO CISCO CEMS

Prepared for:
Norlite Corporation
628 South Saratoga Street
Cohoes, New York

Prepared by:
CEMServices Inc.
360 Old Colony Road
Norton, Massachusetts 02766


Sean M. MacKay
Manager, Environmental Services

1/17/2003
Date



*Emissions Monitoring Protocol - Norlite Corporation Cohoes, New York
CEMS Data Accuracy Assessment Test Program - 1st Quarter, 2003*

TABLE OF CONTENTS

	Page
1. INTRODUCTION	1
2. FACILITY DESCRIPTION	3
A. General	
B. Test Location	
C. Facility Continuous Emission Monitoring System	
3. REFERENCE METHOD TEST PROCEDURES	6
A. Relative Accuracy Test Audit	
B. CEM Calibration Procedures - EPA Test Method 6C	
C. Carbon Monoxide - EPA Test Method 10	
D. Oxygen - EPA Test Method 3A	
E. Calibration Drift, Calibration Error and Response Time Tests-Pt 266, App.IX, PS 2	
F. Stratification Test	
4. REFERENCE METHOD TEST EQUIPMENT	10
A. Mobile CEM Laboratory	
B. Calibration Gases	
5. QUALITY CONTROL PROCEDURES	12
A. General	

LIST OF TABLES

1-1 POLLUTANTS AND TEST METHODOLOGIES	1
1-2 EMISSION LIMITS, ALLOWABLE RA'S & PERFORMANCE SPECIFICATIONS	1
2-1 FACILITY CEMS (KILNS 1 & 2)	5
3-1 PS 2 CEM TRAVERSE POINT LOCATIONS (KILNS 1 & 2)	7
3-2 KILN No. 1 & 2 STRATIFICATION CHECK TRAVERSE POINT LOCATIONS	9
4-1 REFERENCE METHOD ANALYZERS	11
4-2 REFERENCE METHOD CALIBRATION GASES - RATA	11
4-3 CALIBRATION GASES - CALIBRATION ERROR TEST	11

LIST OF FIGURES

2-1 KILN 1 STACK SCHEMATIC	4
2-2 KILN 2 STACK SCHEMATIC	5

APPENDICES

- A - Definition of Abbreviations**
- B - Sample of Calculation Sheets**
- C - Sample of Field Data Sheets**
- D - Sample of Reference Method Calibration Sheets**
- E - Reference Method Sample Train Schematics**
- F - Outline for PST Report**

**Emissions Monitoring Protocol - Norlite Corporation Cohoes, New York
CEMS Data Accuracy Assessment Test Program - 1st Quarter, 2003**

1. INTRODUCTION

CEM Solutions of Hudson, New York and CEMServices of Norton, Massachusetts have been retained by the Norlite Corporation of Cohoes, New York to conduct a Continuous Emissions Monitoring System (CEMS) data accuracy assessment program on the primary and backup CEMS installed on Kilns No. 1 and 2 at the facility. The program is tentatively scheduled for March 2003.

The Norlite Corporation operates two dry process rotary kilns (Kiln No. 1 and Kiln No. 2). Each of these kilns has an independent CEMS to monitor flue gas emissions. In addition, there are backup CEMS that monitor each kiln. The backup systems are identical to the primary systems and operate independently with the exception of sharing the data acquisition system with their respective primary systems. These CEMS monitor the Oxygen (O₂) and Carbon Monoxide (CO) in each kiln's exhaust gas.

The objective of the data accuracy assessment test program is to determine the calibration error, response time and relative accuracy (RA) of these primary and backup CEMS as required by the Methods Manual for Compliance with the Boiler/Industrial Furnace (BIF) Regulations (40 CFR 266), Appendix IX, Section 2.0 - Performance Specifications for CEMS. The RA's will be determined by Relative Accuracy Test Audits (RATA's).

Table 1-1 indicates the air contaminants to be tested and the test methodologies to be used during the emissions test program. Table 1-2 indicates the emission limits and the allowable RA's for the facility.

**TABLE 1-1
POLLUTANTS AND TEST METHODOLOGIES**

Constituents	Test Methods
Oxygen (O ₂)	EPA Test Method 3A
Carbon Monoxide (CO)	EPA Test Method 10

**TABLE 1-2
EMISSION LIMITS, ALLOWABLE RA'S AND PERFORMANCE SPECIFICATIONS**

Parameter	Emissions Limit	Allowable RA	Calibration Drift 24 Hour	Calibration Error	Response Time
CO	100 ppmv @ 7% O ₂	10 % *	< 6 ppm low range < 90 ppm high range	< 10 ppm low range < 150 ppm high range	< 2 min (both ranges)
O ₂	N/A	N/A	< 0.5 %	< 0.5 %	< 2 min

* 10 % of RM average or within 10 ppm of RM average, whichever is less restrictive.

**Emissions Monitoring Protocol - Norlite Corporation Cohoes, New York
CEMS Data Accuracy Assessment Test Program - 1st Quarter, 2003**

1. INTRODUCTION (continued)

As specified in 40 CFR 266, CEMServices will conduct Reference Method (RM) tests to acquire emissions data for comparison to data generated by the facility's CEMS. Each RATA will consist of nine to twelve, twenty-one (21) minute test runs. This data will be used to calculate the Relative Accuracy (RA) of each facility monitor for the following units: O₂ (%dv), CO (ppmdv @ 7% O₂). The RA of each facility monitor will be calculated to the 100th decimal place using the following equation:

$$RA = \{ |d| + |CC| \} \times 100 / RM$$
$$CC = t \times S_d / (n)^{0.5}$$

Where:

RA = Relative Accuracy, percent

|d| = Mean absolute value of the differences between the CEMS and RM values

|CC| = Absolute value of the 2.5 percent error confidence coefficient

RM = Average Reference Method value or emission standard (or permit limit)

t = student t-value (2.5 percent error, one-tailed)

S_d = Standard deviation of the differences between the CEMS and RM values

n = number of data points (9)

Testing will be performed while the kilns are operating at normal operating load waste feed and/or shale feed conditions.

Sean M. MacKay, CEMServices Manager of Environmental Services, will be the Project Director for the RATA portions of this test program. A qualified engineer will assist him. Mark Krizar, CEM Solutions Project Engineer, will be responsible for the calibration error tests, the response time tests, and the collection of facility CEM and operating data. Mr. Tim Lachell of the Norlite Corporation is the facility contact and can be reached at:

*Mr. Tim Lachell
Norlite Corporation
P.O. Box 694
Cohoes, New York*

(518) 235-0401

2. FACILITY DESCRIPTION

A. General

The Norlite facility produces an expanded shale lightweight aggregate in two dry process rotary kilns which are capable of firing waste liquid low grade fuel (LLGF), fuel oil, natural gas or a combination of these fuels.

Kiln No. 1 is manufactured by Fuller-Taylor and is steel shell lined with approximately six-inch refractory brick. The burn zone extends approximately 30 feet from the burner end of the kiln. The burning zone gas temperature is maintained at 2,200 °F to 3,000 °F. Kiln No. 2 is manufactured by Allis-Chalmers and is nearly identical in design, engineering, and constructions as Kiln No. 1.

Raw materials are quarried on-site, sized and transported to the kiln via trucks and a conveyor system. The raw product is introduced to the kiln at the feed (back) end. Calcination of the product occurs at a product temperature of 1,700 °F to 2,000 °F. The shale is then heated to the point of incipient fusion where it is in a semi-plastic state to expand internal gases, thereby creating voids. The cooled vitreous clinker is then discharged and stockpiled.

The rated capacity of the kilns are approximately 27 tons per hour (tph) clinker, however each unit is permitted for 22 tph maximum. Typically, 2.5 MMBtu are required to produce one ton of clinker at maximum capacity. Heat is supplied by firing waste liquid low grade fuel (LLGF), fuel oil, natural gas or a combination of these fuels. All fuel is injected countercurrent to the product flow through the kiln by burners at the discharge (front) end of the kiln. Fuel firing rates typically range from 4 to 10 gallons per minute (gpm), depending upon the fuel energy content.

Each kiln has thermocouples mounted at the kiln gas exit and at the inlet to the baghouse for monitoring process temperature. LGF feed is monitored continuously with a Micromotion Doppler flow meter. Typical kiln exhaust gas and material retention times are 15 seconds and 45 minutes, respectively. Draft for each kiln is supplied by a Barrons 400 HP induced draft fan rated at 53,000 ACFM at 450 °F. Secondary combustion air is supplied by ford draft clinker cooler fans rated at a total of 25,000 CFM. The clinker cooler at the front end of the kiln preheats the secondary combustion air.

B. Test Locations

All RM testing will take place at the sampling locations of each kiln. These sampling locations are in close proximity to the facility CEM sample probes. The diameter of the steel ducts at these locations is four feet. The sample ports for the primary CEMS on Kiln No. 1 are located approximately 6.5 feet (1.6 diameters) downstream of a bend in the duct exiting the baghouse and approximately 9.5 feet (2.4 diameters) upstream of an expansion joint prior to the ID fan. The sample ports for the backup CEMS sample probes are located approximately 9.3 feet (2.3 diameters) of a bend in the duct exiting the baghouse and approximately 6.7 feet (1.7 diameters) upstream of an expansion joint prior to the ID fan.

2. FACILITY DESCRIPTION (continued)

The sample ports for the primary and backup CEMS sample probes on Kiln No. 2 are located approximately 11 feet (2.8 diameters) downstream of the exit of the bag house outlet duct and 5.75 feet (1.4 diameters) upstream of a 45 degree bend in the duct leading to the ID fan. The primary and backup probes are on opposite sides of the duct. The figures below and on the following page are a schematic of the Kiln 1 and 2 locations.




FIGURE 2-1 KILN 1 STACK SCHEMATIC

2. FACILITY DESCRIPTION (continued)




FIGURE 2-2 KILN 2 STACK SCHEMATIC

C. Facility CEMS

As a condition of the New York State Part 373 Hazardous Waste Permit, Norlite is required to operate a CEMS to monitor O₂, and CO in each kiln's exhaust gas. Norlite's permit requires that they operate the CEMS in accordance with the applicable standards defined in 40 CFR 264, and 266. CEMS calibration checks are conducted daily using NIST traceable gas standards. In addition to the annual RATA, quarterly cylinder gas audits are conducted on each CEMS.

The extractive CEMS at Norlite was manufactured by Custom Instrumentation Services Corporation of Englewood, CO. The system measures and records dry undiluted concentrations of CO and O₂ on the two kilns. To comply with the data availability requirements, two points in each stack are sampled with two sets of analyzers (CO and O₂) on each stack. From this data, one database is created which, under normal operating conditions, has values for every minute in the hour. Table 2-1 is a description of the facility's monitors:

**TABLE 2-1
FACILITY CEMS (KILNS 1 & 2)**

Parameter	Manufacturer	Operating Principle	Ranges
O ₂	Siemens/Oxymat 5E	Paramagnetic	0-25 %
CO	Siemens/Ultramat 5E	Infrared	0-200, 0-3000 ppm (dual)

3. REFERENCE METHOD TEST PROCEDURES

A. Relative Accuracy Test Audit

The data accuracy assessment of the CEMS will be conducted using a Relative Accuracy Test Audit. This audit will consist of comparing data generated by the facility CEMS to data acquired simultaneously using US EPA reference test methods. Nine (9) to twelve (12), twenty-one (21) minute test runs will be conducted. All testing will take place at the stack sampling locations.

The clock time of the data acquisition system in the CEMServices mobile laboratory will be set exactly to the facility CEMS time. The start and stop time of each test run will be documented on the facility's data acquisition system. During each test run, all CEM and process operation data will be printed out. Load conditions will be well above fifty (50) percent of rated maximum during all RATA testing.

B. CEM Calibration Procedures - EPA Test Method 6C

Method 6C specifies the calibration procedures and requirements for the instrumentation methods (as well as the requirements for the determination of Sulfur Dioxide emissions from stationary sources). Before any testing is conducted, the analytic range of all test analyzers will be set up so that expected source emissions are at least thirty (30) percent of this range and do not exceed this range. Once this range is determined, calibration gases will be chosen within this range. Only gases prepared according to EPA Protocol #1 will be used.

Analyzer calibration error checks will be conducted by challenging each analyzer with a zero, mid, and high range gas. Analyzer responses to these gases will be within two (2) percent of the instrument's analytical range. Before and after each test run a sampling system bias check will be conducted on each monitor. This check consists of introducing the calibration gases at the base of the sampling probe thus allowing the gases to travel through the entire sampling system including the condensers used to remove moisture content. The analyzer responses to this check will be recorded onto field calibration sheets. All system bias check responses will be within five (5) percent of the instruments analytical range, when compared to the analyzer calibration error check conducted initially.

The sampling system bias check conducted prior to each test run will be compared to the sampling system bias check conducted at the completion of that same run. Differences between the two bias checks constitute the upscale and zero calibration drifts. All calculated calibration drifts will be below three (3) percent of the analytical range of the analyzer.

Once the initial system bias check is conducted the system will be put into the sample mode and data acquisition will be initiated. The probe will be positioned at the first traverse point. The probe will be 5/8" stainless steel tube that will be traversed at 16.7%, 50.0%, and 83.3% of the stack diameter as follows:

**Emissions Monitoring Protocol - Norlite Corporation Cohoes, New York
CEMS Data Accuracy Assessment Test Program - 1st Quarter, 2003**

3. REFERENCE METHOD TEST PROCEDURES (continued)

**TABLE 3-1
PS2 CEM RM TRAVERSE POINT LOCATIONS (KILNS 1 & 2)**

Traverse Point #	% of Diameter	Probe Mark (inches)
1	16.7	8.0
2	50.0	24.0
3	83.3	40.0

Voltage output from each monitor will be documented by a Strawberry Tree Data Shuttle. The data shuttle sends all signals via a RS-232 cable to a computer for data archiving. Data points will be logged every two (2) seconds and averaged in one minute increments during each test run. At the test run completion, one-minute averages will be automatically printed out and a calibration will be performed. Calibration values will be logged in the program and printed out with calculated bias and drift percentages.

Results from the initial and final system bias checks will be used to adjust the raw run average to correct it for any deviations due to the system bias. This data will be printed out in a format that lists the analyzers response to the gas, the bias, the drift, the raw run average, and the corrected run average.

C. Carbon Monoxide - EPA Test Method 10

Method 10 is used for the determination of Carbon Monoxide emissions from stationary sources using instrumental analyzer procedures. Because Method 10 procedures are much less stringent than Method 6C, all calibration procedures and requirements will be conducted according to EPA Test Method 6C to improve data quality.

A Thermo Environmental Model 48 Gas Filter Correlation (GFC) analyzer will be used to continuously sample the CO concentrations in the gas stream. GFC spectroscopy is based on the comparison of the infrared (IR) absorption spectrum of the measured gas to that of other gases in the sample being analyzed. This technique is implemented by using a high concentration sample of the measured gas (i.e. CO) as a filter for the infrared radiation transmitted through the analyzer. Radiation from an IR source is chopped and passed through a gas filter alternating between CO and N2 due to rotation of the filter wheel. The radiation then passes through an interference filter and on to an absorption cell. The IR radiation exits the sample cell and falls on to an IR detector. The CO gas filter produces a reference beam which cannot be further attenuated by CO in the sample cell. The N2 side of the filter wheel is transparent to the IR radiation and thus produces a measure beam which is partially absorbed by CO in the cell. The chopped detector signal is modulated by the alternation between the two gas filters with is amplified and related to the concentration of CO in the sample cell. Other gases, which absorb the reference and measure beams equally, do not cause modulation of the detector signal leaving the GFC responding specifically to CO. With this in mind, there is no need for an ascarite trap (ascarite scrubs out CO2).

**Emissions Monitoring Protocol - Norlite Corporation Cohoes, New York
CEMS Data Accuracy Assessment Test Program - 1st Quarter, 2003**

3. REFERENCE METHOD TEST PROCEDURES (continued)

An interference check is performed on the CO analyzer (and all other analyzers) every six months in accordance with the interference check procedures described in EPA Method 20. In addition, an interference check is performed during each field use in the form of the calibration error test. Since the calibration gases go to all analyzers during the test, the CO response to nitrogen can be compared to responses to the mid and high-level O₂, CO₂ blends. If the CO₂ gas is interfering it will bias the CO result which should otherwise be reading near zero. All interference checks will be below the required allowable limit of 1% described in 40 CFR 266, Appendix IX, Section 2.1 (the limit is 2% in Method 20 of 40 CFR 60).

D. Oxygen - EPA Test Method 3A

Method 3A is used for the determination of Oxygen emissions from stationary sources using instrumental analyzer procedures. All calibration procedures and requirements for this instrumentation method are identical to those found in EPA Test Method 6C.

O₂ content in the effluent will be determined by a Teledyne Model 326A monitor which utilizes a micro-fuel cell that consumes O₂ from the atmosphere surrounding the measurement probe. The consumption of O₂ generates a proportional electrical current. This current is then amplified and provides a signal output of 0-1 V DC which corresponds to a full scale range of 0-25 % O₂.

E. Calibration Drift, Calibration Error and Response Time Tests - Pt. 266, Appendix IX, PS 2

A calibration drift (CD) test will be conducted during the RATA test. While the facility is operating at normal conditions, the CD will be determined at twenty-four hour intervals for seven consecutive days (no adjustments will be made during this period). For this test two calibration gases (zero and span) are introduced into the entire system. The CEMS response is recorded and subtracted from the gas value. The span gas values for the CO monitors will be 100-180 ppm for the low range (0-200 ppm) and 1500-2700 ppm for the high range (0-3000 ppm). For the O₂ monitors the span gas value will be 12.5-22.5 %. The calibration drift may not exceed 3 % of span

The calibration error (CE) test will consist of challenging the facility CEMS with EPA Protocol 1 gases at three measurement points within both ranges of the analyzers corresponding to zero (0 to 20 % of span), mid (30-40 % of span) and high (70-80 % of span) level gases. The mean difference between the CEMS response to these gases and the actual gas values may not exceed 5 % of span.

Response time of the facility CEMS measurement system will be determined by first introducing zero gas into the system at the calibration valve until the reading is stable (no change greater than 1 % of full scale for 30 sec); then the valve will be switched to sample to monitor the stack effluent until a stable reading is obtained. This will be the upscale response time. Next, the high level gas will be introduced to the system. Once the system has stabilized at this point, the system will be switched to sample mode to monitor the effluent until a stable reading is obtained. This will be the downscale response time. This procedure will be repeated three times at each range and the response time will be the "greater" of the times. Emissions data during the test for both kilns will be obtained and included in the final report

**Emissions Monitoring Protocol - Norlite Corporation Cohoes, New York
CEMS Data Accuracy Assessment Test Program - 1st Quarter, 2003**

3. REFERENCE METHOD TEST PROCEDURES (continued)

F. Stratification Check

In accordance with 40 CFR 60, Part 266, Appendix IX, section 2.1.3, a stratification check will be performed. Stratification is defined as a difference in excess of 10 percent between the average concentration in the stack and the concentration at any point more than 1 meter from the stack wall. A dual probe system will be used to measure the average effluent concentration (facility CEMS) while measurements at each traverse point (RM CEMS) are made. Only one port per cross section is available for each stack, therefore a one port, eight-point traverse will be conducted on each kiln. The stratification test will most likely be performed as part of the first RATA run on each kiln. The following table is the stratification check traverse point locations for both kilns.

**TABLE 3-2
KILN No. 1 & 2 STRATIFICATION CHECK TRAVERSE POINT LOCATIONS**

Traverse Point	% Diameter	Location (inches)
1	3.2	1.5
2	10.5	5.0
3	19.4	9.3
4	32.3	15.5
5	67.7	32.5
6	80.6	38.7
7	89.5	43.0
8	96.8	46.5

4. REFERENCE METHOD TEST EQUIPMENT

A. Mobile CEM Laboratory

All reference method test methods described in Section 3 will be conducted using the CEMServices mobile CEM laboratory. This laboratory consists of all analyzers and support equipment used to conduct the CEM sampling during this test program. The following is a description of each item that makes up the entire system:

Sample Probe - A stainless steel probe will be used for this test program. The probe has a filter at the end of it to remove particulate matter. The other end contains a heated "flood chamber" to allow either sample or system calibration.

Particulate Filter - This in-stack filter is a Labyrinth Systems 5 micron sintered stainless steel design.

Calibration Valve Assembly - This assembly consists of a Labyrinth Systems heated "flood chamber". Sample and calibration modes are actuated from a valve located in the CEM lab. The flood chamber consists of "T" valve assembly that connects the sample line, the probe, and the calibration line. During calibration, gas flows up the calibration line to the "flood chamber" and down the sample line. Excess gas is discharged out the probe cleaning the filter. This chamber will be heated to 250 degrees F, with a self-regulating Omega heat bar.

Heated Sample Line - The heated sample line is ten (10) feet long and transports the gas sample from the CEM probe to the moisture removal system. This jumper is temperature self regulating and will maintain a temperature of 250 degrees F.

Moisture Removal System - This system continuously removes moisture from the sample gas while maintaining minimal contact between the condensate and the sample gas. CEMServices uses an ice bath condenser consisting of three (3) stainless steel heat exchangers which are continuously drained of condensate by two (2) peristaltic pumps. The inlet to the system is connected to the heated sample line and the outlet will be connected to the sample transport line.

Sample Transport Line - 3/8-inch OD Teflon tubing will be used to transport the gas sample from the moisture removal system to the mobile laboratory. Approximately one hundred and fifty (150) feet of tubing will be used.

Sample Pump - A dual headed diaphragm pump will be used to transport the gas sample through the system to the sample gas manifold. This pump is manufactured by Air Dimension and all parts coming into contact with the gas stream are either Teflon or stainless steel.

Sample Gas Manifold - This manifold consists of a series of valves and adjustable rotameters capable of setting and maintaining the desired backpressure and flow rate to the analyzers during both sampling and calibration.

**Emissions Monitoring Protocol - Norlite Corporation Cohoes, New York
CEMS Data Accuracy Assessment Test Program - 1st Quarter, 2003**

4. REFERENCE METHOD TEST EQUIPMENT (continued)

Sample Gas Analyzers - CEMServices used the following analyzers to complete this test program:

**TABLE 4-1
REFERENCE METHOD ANALYZERS**

Gas	Manufacturer	Model	Serial #	Range
O2	Teledyne	326A	153157	0-25 %
CO	Thermo Electron	48	48-40156-262	0-200 PPM

Data Recorder - All voltage outputs from the analyzers are sent to a Strawberry Tree Data Shuttle. This shuttle logged data a two second intervals. Data from the shuttle is sent to a computer where a Strawberry Tree data acquisition program lists instantaneous concentration values for each parameter. At the conclusion of each run, one-minute averages are printed out and a calibration is initiated through the program. The calibration data is used to correct the raw averages for system bias and drift.

C. Calibration Gases

All calibration gases used in this test program will be prepared according to EPA Protocol #1. As per EPA Test Method 6C for all O2 testing, all mid calibration gas values will be between 40-60 % of the analytical range of the analyzer, and all high calibration gases will be between 80-100 %. As per Method 10 for all CO testing the mid calibration gas will be approximately 30 % of the analytical range of the analyzer, and the high calibration gas will be approximately 60 %. The zero calibrations for all analyzers will be conducted using pre-purified grade Nitrogen.

**TABLE 4-2
REFERENCE METHOD CALIBRATION GASES - RATA**

Gas/Range	Allowable Values	Cal Point
O2 0-25 %	10-15 20-25	Mid High
CO 0-200 PPM	approx. 60 approx. 120 approx. span	Mid High Span

**TABLE 4-3
CALIBRATION GASES - CALIBRATION ERROR TEST**

Measurement Point	CO ppm		O2 %
	Low Range	High Range	
1	0-40	0-600	0-2
2	60-80	900-1200	8-10
3	140-160	2100-2400	14-16

5. QUALITY CONTROL PROCEDURES

A. General

Throughout all phases of this test program strict attention will be given to all testing to provide the highest quality of results possible.

All of CEMServices test equipment is of the highest quality available and undergoes routine maintenance to ensure top operating condition. This includes instruments, sample conditioners, sample lines, and probes.

Sampling will be conducted by trained personnel with extensive experience in source sampling. All sampling and analysis will be conducted in strict accordance with EPA test procedures (where available). The quality control procedures found in the EPA Quality Assurance Handbook for Air Pollution Measurement Systems will be adhered to as well.

Analyzer calibrations will be performed at the beginning of each test day. System calibrations will be performed before and after each test run through the entire sampling system.

As mentioned in Section 3, an ascarite trap is not needed for the GFC CO analyzers. However, an interference check is performed on the CO analyzer (and all other analyzers) every six months in accordance with the interference check procedures described in EPA Method 20. In addition, an interference check is performed during each field use in the form of the calibration error test. Since the calibration gases go to all analyzers during the test, the CO response to nitrogen can be compared to responses to the mid and high-level O₂, CO₂ blends. If the CO₂ gas is interfering it will bias the CO result which should otherwise be reading near zero.

All calculations will be conducted in strict accordance with the equations found in the individual Methods. Calculations will be conducted on a computer and the input data will be checked by a person other than the original calculator to ensure that it is correct.

The entire staff of CEMServices is thoroughly familiar with all test methods used in this program and has extensive experience in source emission monitoring.

DEFINITION OF ABBREVIATIONS

ACFM	Flowrate reported in actual cubic feet per minute.
An	Area of the nozzle, cross-sectional, in square feet.
As	Area of the stack in square feet.
BWO	Water vapor in gas stream, proportional by volume.
CC	Percent error confidence coefficient (one tailed).
Cd	Conversion calibration for concentration (PPMdv to lbs/SCF)
Cgas	Final emissions data reported by CEMS, adjusted for calibration drift. Reported as ppm dry, proportional by volume.
Cm	Average CEM response to initial and final span gas system calibration.
Cma	Concentration of the calibration gases.
Co	Average CEM response to initial and final zero gas system calibration.
Craw	Raw emissions data reported by the CEMS, uncorrected for calibration drift.
Cwet	Final emissions data reported by CEMS, adjusted for calibration drift and water vapor. Reported as ppm wet, proportional by volume.
% CO	Percent of carbon monoxide in the flue gas.
% CO₂	Percent of carbon dioxide in the flue gas.
Cp	Pitot tube coefficient.
Cs	The concentration in the stack in pounds per standard cubic foot.
Cs'	The concentration in the stack in grains per standard cubic foot.
Cs' @ 12%	The concentration in the stack in grains per dry standard cubic feet corrected to 12% CO ₂ .
DELTA H	The pressure differential across orifice meter, reported in inches of H ₂ O.
DELTA H(ABS)	The pressure differential across orifice meter, absolute conditions in inches of mercury.
Dn (IN)	Diameter of the nozzle in inches.
DGM IN	Temperature of the dry gas meter inlet, reported in degrees Fahrenheit.
DGM OUT	Temperature of the dry gas meter outlet, reported in degrees Fahrenheit.
Ds (FT)	Diameter of the stack in feet.
DSCFH	Dry standard cubic feet per hour.
DSCFM	Dry standard cubic feet per minute.
DSCMH	Dry standard cubic meters per hour.
E	Emission rate in pounds per million Btu using F Factor of fuel burned.
END METER	The dry gas meter reading at the end of the test.
F FACTOR	The theoretical amount of air in dry standard cubic feet (DSCF) needed to combust a million Btu's worth of fuel.
GR/BHP-HR	Grams per brake horsepower hour.
IMP(FIN)	Final volume of absorbing solution in impinger.
IMP(INT)	Initial volume of absorbing solution in impinger.
INT METER	The dry gas meter reading at the beginning of the test.
% ISO	Variation of sampling from isokinetic conditions.
LB/HR	Pounds per hour.
LB/MMBTU	Pounds per million British Thermal Unit.
LB/SCF	Pounds per standard cubic foot.
Md (DRY)	The dry molecular weight of the flue gas in pounds per pound mole.
MI	Volume in milliliters.
Mg/M3	Milligrams per cubic meter.
Mn	Total particulate found in sample minus the acetone residue (blank). Reported in milligrams.
Ms (WET)	Wet or actual molecular weight of the flue gas in pounds per pound mole.
MW	Molecular weight
% N2	The percent of nitrogen in the flue gas.
NO. PTS	Number of traverse points.
% O2	% oxygen in the flue gas.
P BAR	Barometric pressure at test location.
PIT COEFF	Pitot tube coefficient (S Type=.84, standard=.99).
PPM	Parts per million.



DEFINITION OF ABBREVIATIONS

PPMdv	Parts per million - dry volume.
PPMwv	Parts per million - wet volume.
P STK	Static pressure of the stack in inches of water.
PMR	The pollutant mass rate in pounds per hour.
PS (ABS)	Absolute stack pressure in inches of mercury.
Pstd	Standard absolute pressure, (29.92 in. Hg).
Qs	The volumetric flow rate of the flue gas in dry standard cubic feet per hour.
RA	Relative accuracy.
RATA	Relative accuracy test audit.
RM	Reference Method.
Sd	Emission standard (allowable emission rate).
SQ ROOT	The square root of each velocity head measurement (Delta P).
SQRT DELTA P	The average of the square roots of the measured pressure drops.
Stack Temp	The temperature of the stack in degrees (°F) Fahrenheit.
TM (°F)	Average temperature of the dry gas meter in degrees Fahrenheit.
TM (°R)	Average temperature of the dry gas meter in degrees Rankine.
TS (°R)	The temperature of the stack in degrees Rankine.
VEL HEAD	The pressure drop measured across the pitot tubes.
VI (TOT)	The amount of water collected in the impingers in milliliters.
VM (CF)	The volume sampled through the dry gas meter in cubic feet.
VM STD	Volume sampled through the dry gas meter corrected to standard conditions.
VOC	Volatile organic compounds
VS	Velocity of the stack gas in feet per second.
VW STD	The amount of moisture collected, corrected to standard conditions.
Y	Dry gas meter calibration factor.



EFFLUENT GAS CALIBRATION CORRECTION

FACILITY:
 UNIT:
 DATE:

RUN ID#:
 START:
 END:

O2 CO2 CO

Cma = ACTUAL CAL GA

INIT ZERO
 FINAL ZERO

Co = AVG ZERO
 INIT UPSCALE
 FINAL UPSCALE

Cm = AVG UPSCALE

Craw = RAW RM DATA

Cgas = ADJ RM DATA

$$C_{gas} = (C_{RAW} - C_o) * \frac{C_{ma}}{(C_m - C_o)}$$

$$CO @ 3\% O_2 = C_{gas} * ((20.9-3)/(20.9 - \% O_2))$$

RELATIVE ACCURACY CALCULATION SHEET -

PLANT:
 LOCATION:
 TEST DATE:
 PARAMETER:
 UNITS:

RUN NAME

RUN #	1	2	3	4	5	6	7	8	9
-------	---	---	---	---	---	---	---	---	---

RM DATA

PLANT DATA

DIFFERENCE

DIFF SQRD

D = ARITH. MEAN

SUM DIFF

(SUM DIF)SQRD

SUM(DIF SQRD)

N = NUMBER OF RUNS

Sd = STANDARD DEV.

T.975 = T VALUE

CC = $T.975 * Sd / \text{SQRT}(N)$

CC = CONFIDENCE COE

PLANT AVERAGE =

RM = AVG RM DATA

ABSOLUTE DIFF. =

RA = $(\text{ABS}(D) + \text{ABS}(CC)) / \text{RM} * 100$

Sd = STANDARD =

RA = REL ACCURACY =

SYSTEM CALIBRATION BIAS & DRIFT DATA

FACILITY: _____ UNIT: _____

DATE: _____ RUN TIME: _____

RUN: _____

CALIBRATION SPANS:

O2	CO2	CO
PCT	PCT	PPM

GAS	Analyzer cal. response	Initial		Final		Drift % span
		System bias response	System cal. bias % span	System bias response	System cal. bias % span	
O2 ZERO						
O2 UPSCALE						
CO2 ZERO						
CO2 UPSCALE						
CO ZERO						
CO UPSCALE						

System Bias = (System cal response - Analyzer cal response) / Span * 100

Drift = (Final system cal response - Initial system cal response) / Span * 100

System Bias may not exceed 5% / Drift may not exceed 3%



360 Old Colony Road, Suite 1, Norton, MA 02766

ANALYZER CALIBRATION DATA

FACILITY: _____ UNIT: _____

DATE: _____ OPERATOR: _____

START TIME: _____

CALIBRATION SPANS:

O2	CO2	CO
PCT	PCT	PPM

	GAS	Cylinder value	Analyzer cal. response	Absolute difference	Difference % span	Cylinder number
O2	zero					
	mid					
	high					
CO2	zero					
	mid					
	high					
CO	zero					
	mid					
	high					

Difference % Span = Absolute difference / Span * 100

Note - Difference % Span may not exceed 2%

RESPONSE TIME

FACILITY : _____ UNIT : _____

DATE : _____ OPERATOR: _____

Span Gas Concentration : _____

Analyzer Span Concentration: _____

Upscale:

- 1 _____ seconds
- 2 _____ seconds
- 3 _____ seconds

Average Upscale Response: _____ Seconds.

Downscale:

- 1 _____ seconds
- 2 _____ seconds
- 3 _____ seconds

Average Downscale Response: _____ Seconds.

Average System Response: _____ Seconds.

Slower Average Time: _____ Seconds.

INTERFERENCE RESPONSE TEST

Date: _____

CEM System: _____

Calibration Span & Gas Values:

	CO	SO ₂	O ₂	CO ₂	NOx
Span					
Gas Value					

Analyzer Type: _____


Serial #: _____

Test Gas Type	Concentration (ppm)	Analyzer Response (ppm)	% of Span


$$\% \text{ of Span} = \frac{\text{Analyzer Response}}{\text{Instrument Span}} \times 100$$

Interference may not exceed 2% of span.





Schematic of EPA Method 3A Sampling Train



Schematic of EPA Method 10 Sampling Train

*Emissions Monitoring Protocol - Norlite Corporation Cohoes, New York
CEMS Data Accuracy Assessment Test Program - 1st Quarter, 2003*

OUTLINE FOR PST REPORT

	Page
1. INTRODUCTION	1
2. SUMMARY OF RESULTS	3
3. FACILITY DESCRIPTION	5
A. General	
B. Test Location	
C. Facility CEMS	
4. REFERENCE METHOD TEST PROCEDURES	7
A. Relative Accuracy Test Audit	
B. CEM Calibration Procedures - EPA Test Method 6C	
C. Oxygen and Carbon Dioxide - EPA Test Method 3A	
D. Carbon Monoxide - EPA Test Method 10	
E. Calibration Drift, Calibration Error & Response Time Tests - Pt 266, App. IX, PS 2	
5. REFERENCE METHOD TEST EQUIPMENT	11
A. Mobile CEM Laboratory	
B. Calibration Gases	
6. QUALITY CONTROL PROCEDURES	14
A. General	

LIST OF TABLES

1-1 POLLUTANTS AND TEST METHODOLOGIES	1
1-2 EMISSION LIMITS AND ALLOWABLE RA'S	1
2-1 KILN NO. 1A RESULTS	3
2-2 KILN NO. 1 B RESULTS	3
2-3 KILN NO. 2 A RESULTS	4
2-4 KILN NO. 2 B RESULTS	4
3-1 FACILITY MONITORS	6
4-1 CEM RM TRAVERSE POINT LOCATIONS	8
5-1 REFERENCE METHOD ANALYZERS	12
5-2 REFERENCE METHOD EPA PROTOCOL #1 CALIBRATION GASES	13

APPENDICES

- A - Relative Accuracy Calculation Sheets (Kiln No.1 and No.2)
- B - Calibration Drift Test Results (Kiln No.1 and No.2)
- C - Calibration Test Error Results (Kiln No.1 and No.2)
- D - Response Time Test Results (Kiln No.1 and No.2)
- E - Reference Method CEM Raw Run and Calibration Data (Kiln No.1 and No.2)
- F - Facility CEM and Operating Data (Kiln No.1 and No.2)
- G - Reference Method Equipment Calibration Sheets
- H - Calibration Gas Certificates of Analysis
- I - Definition of Abbreviations

